



**Department  
of Health**

**Wadsworth  
Center**

# **New York State Biomonitoring Program for Trace Elements**

## **Event #1, 2023**

### **Trace Elements in Whole Blood, Urine, and Serum**

## **April, 2023**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



**Event #1, 2023:  
Trace Elements in Whole Blood, Urine, and Serum**

4/7/2023

Dear Laboratory Director,

This report summarizes performance for the first biomonitoring proficiency test (PT) event of 2023 for Trace Elements in Whole Blood, Urine, and Serum. One of the key goals of this PT program is to achieve harmonization of biomonitoring data for trace elements. In this report, we summarize the responses to our recent survey request. Please refer to the attachment at the end of the report for more details.

**Target Value Assignment and Performance Evaluation:**

For these PT materials, target values have been assigned for a limited number of trace elements that are gradable under criteria set by the NYS DOH Biomonitoring PT program. See assay-specific narratives for details. Data for additional trace elements are reported and are included here in order to characterize the PT materials more completely. Participant data and descriptive statistics are provided for educational purposes. No target value or acceptable range is implied.

Where the data permit, robust statistics were used to assign target values based on Algorithm A as defined by ISO 13528:2005E *Statistical methods for use in proficiency testing by inter-laboratory comparisons* [1]. Acceptable ranges for the graded elements are based on consensus criteria and/or those set by the NYS DOH's PT program. For example, some are fixed based on US regulatory guidelines (Pb, Cd) while for other elements the criteria are based on a consensus of the Network of PT scheme organizers for trace elements in occupational and environmental laboratory medicine [2]. Quality specifications are element and matrix specific; full details are provided under each element specific narrative.

A confidential, three-digit code number assigned by PT program staff identifies all laboratory participants.

Samples for the next PT event (Event #2, 2023) will be shipped May 17, 2023. Comments about this report may be directed to [trel@health.ny.gov](mailto:trel@health.ny.gov).

Sincerely,

Patrick J. Parsons, PhD  
Chief, Inorganic and Nuclear Chemistry,  
Division of Environmental Sciences  
Wadsworth Center

Kayla Mehigan  
Coordinator, Biomonitoring PT Program,  
Division of Environmental Sciences  
Wadsworth Center



**Department  
of Health**

**Wadsworth  
Center**

**Event #1, 2023**

**Trace Elements in  
Whole Blood**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



**Event #1, 2023:  
Trace Elements in Whole Blood**

**PT Materials**

Human whole blood was purchased from Zen-Bio, Inc. and preserved with K<sub>2</sub>EDTA. The company certifies that this material was "non-reactive" for HBsAg, HBV DNA, HIV-1,2 Ab, HIV-1 RNA, HCV Ab, HCV RNA, and STS. Units of whole blood were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with arsenic (As), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), barium (Ba), beryllium (Be), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), titanium (Ti), thallium (Tl), uranium (U), vanadium (V), tungsten (W), and zinc (Zn). Whole blood samples were homogenized overnight prior to aliquoting 2-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories

**Graded Elements**

Seven elements in whole blood are formally graded: As, Cd, Co, Cr, Hg, Mn, and Pb. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

**Additional Elements**

An additional 25 elements were reported by at least one participant: Ag, Al, Ba, Be, Bi, Cs, Cu, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, Tl, U, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #1, 2023: Summary Statistics

Whole Blood As (µg/L)					
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	18.9	31.7	8.8	4.55	1.72
<b>Upper Limit</b>	24.9	38.0	14.8	10.55	7.72
<b>Lower Limit</b>	12.9	25.4	2.8	0.00	0.00
<b>Arithmetic SD (s)</b>	0.5	0.5	0.2	0.19	0.11
<b>Arithmetic RSD (%)</b>	2.6	1.6	2.7	4.2	6.4
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

The acceptable range is based on quality specifications:  $\pm 6$  µg/L or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6$  µg/L at concentrations less than or equal to 30 µg/L. These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #1, 2023: Performance of Participating Laboratories






Whole Blood As (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
	<b>Target</b>	<b>18.9</b>	<b>31.7</b>	<b>8.8</b>	<b>4.55</b>	<b>1.72</b>
103	ICP-MS/MS	18.4	31.5	8.37	4.33	1.56
110	ICP-MS/MS	19.1	31.4	8.84	4.38	1.72
147	ICP-MS	18.2	31.4	8.91	4.70	1.90
264	ICP-MS	19.24	31.44	8.96	4.74	1.73
293	DRC/CC-ICP-MS	19.55	32.7	8.63	4.71	1.7
391	ICP-MS	*7.331 ↓	*26.0	*4.48	<0.000	<0.000
597	ICP-MS/MS	18.9	31.9	8.99	4.44	1.72

Based on the grading criteria for As in Whole Blood, 97% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Whole Blood As



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±6 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±6 µg/L at concentrations less than or equal to 30 µg/L.



## Results for Event #1, 2023: Summary Statistics

Whole Blood Cd (µg/L)					
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Target (Robust Mean (x*))	1.57	6.07	9.8	0.78	3.32
Upper Limit	2.57	7.07	11.3	1.78	4.32
Lower Limit	0.57	5.07	8.3	0.00	2.32
Robust SD (s*)	0.05	0.16	0.5	0.03	0.10
Robust RSD (%)	3.2	2.6	5.1	4.1	3.0
Number of Sample Measurements (N)	11	12	12	10	12
Standard Uncertainty (u)	0.02	0.06	0.2	0.01	0.04

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.7 \mu\text{g/L}$ . These quality specifications are based on those used by US OSHA for occupational exposure.





### Results for Event #1, 2023: Performance of Participating Laboratories






Whole Blood Cd (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Target		1.57	6.07	9.8	0.78	3.32
103	ICP-MS/MS	1.54	5.96	9.91	0.779	3.33
107	ICP-MS/MS	1.585	6.075	9.784	0.770	3.366
110	ICP-MS	1.56	6.13	10.0	0.80	3.36
116	ICP-MS/MS	1.76	6.27	10.6	<1.50	3.64
147	ICP-MS	1.64	6.62	10.6	0.900	3.81
264	ICP-MS	1.59	5.94	10.18	0.79	3.31
293	DRC/CC-ICP-MS	1.56	5.70	9.360	0.8	3.16
391	ICP-MS	<0.000 ↓	5.95	8.117 ↓	<0.000	0.77 ↓
597	ICP-MS/MS	1.53	6.05	9.61	0.751	3.27
605	ICP-MS	1.51	6.10	9.50	0.755	3.28
606	ICP-MS/MS	1.42	6.04	9.53	0.711	3.26
686	ICP-MS	1.61	6.20	9.84	0.764	3.50

Based on the grading criteria for Cd in Whole Blood, 95% of results were satisfactory, with 1 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Whole Blood Cd



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.7 \mu\text{g/L}$ .



## Results for Event #1, 2023: Summary Statistics

	Whole Blood Co (µg/L)				
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	0.58	2.47	15.2	5.0	1.11
<b>Upper Limit</b>	2.08	3.97	18.2	6.5	2.61
<b>Lower Limit</b>	0.00	0.97	12.2	3.5	0.00
<b>Arithmetic SD (s)</b>	0.04	0.17	0.6	0.4	0.07
<b>Arithmetic RSD (%)</b>	7.2	6.9	3.9	8.0	6.3
<b>Number of Sample Measurements (N)</b>	8	8	7	8	7

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #1, 2023: Performance of Participating Laboratories






Whole Blood Co (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
	<b>Target</b>	<b>0.58</b>	<b>2.47</b>	<b>15.2</b>	<b>5.0</b>	<b>1.11</b>
103	ICP-MS/MS	0.593	2.48	15.0	5.01	1.13
110	ICP-MS/MS	0.58	2.50	15.5	5.12	1.14
147	ICP-MS	0.660	2.73	16.0	5.70	1.21
255	ICP-MS	0.53	2.4	14	4.7	1
264	ICP-MS	0.61	2.52	15.11	5.13	1.07
293	DRC/CC-ICP-MS	0.58	2.44	15.57	5.01	1.14
391	ICP-MS	0.53	2.12	*18.72 ↑	4.12	*0.53
597	ICP-MS/MS	0.580	2.54	15.1	5.07	1.09

Based on the grading criteria for Co in Whole Blood, 98% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Whole Blood Co



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories. Gray area = acceptable range based on quality specifications:

±1.5 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 7.5 µg/L.



## Results for Event #1, 2023: Summary Statistics

Whole Blood Cr ( $\mu\text{g/L}$ )					
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	9.1	2.9	1.23	0.68	5.08
<b>Upper Limit</b>	11.1	4.9	3.23	2.68	7.08
<b>Lower Limit</b>	7.1	0.9	0.00	0.00	3.08
<b>Arithmetic SD (s)</b>	1.0	0.5	0.22	0.14	0.22
<b>Arithmetic RSD (%)</b>	11	17	18	21	4.3
<b>Number of Sample Measurements (N)</b>	8	7	6	5	7

The acceptable range is based on quality specifications:  
 $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #1, 2023: Performance of Participating Laboratories






Whole Blood Cr (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Target		9.1	2.9	1.23	0.68	5.08
103	ICP-MS/MS	7.74	3.64	<1.50	<1.50	4.81
110	ICP-MS/MS	9.47	2.53	1.13	0.70	5.12
147	DRC/CC-ICP-MS	10.3	2.99	1.02	0.759	5.41
255	ICP-MS	9.5	2.5	1.2	<1.0	4.9
264	ICP-MS	9.75	3.25	1.66	0.83	5.30
293	DRC/CC-ICP-MS	9.45	2.49	1.21	0.63	5.05
391	ICP-MS	7.29	*6.75 ↑	*3.64 ↑	*1.84	*6.15
597	ICP-MS/MS	9.59	2.58	1.15	0.469	4.95

Based on the grading criteria for Cr in Whole Blood, 95% of results were satisfactory, with 1 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Whole Blood Cr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±2 µg/L at concentrations less than or equal to 10 µg/L.





## Results for Event #1, 2023: Summary Statistics

Whole Blood Hg (µg/L)					
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Target (Robust Mean (x*))	0.99	0.56	6.5	22.4	2.86
Upper Limit	3.99	3.56	9.5	29.1	5.86
Lower Limit	0.00	0.00	3.5	15.7	0.00
Robust SD (s*)	0.12	0.06	0.6	1.9	0.23
Robust RSD (%)	12	11	9.2	8.5	8.0
Number of Sample Measurements (N)	10	11	12	12	12
Standard Uncertainty (u)	0.05	0.02	0.2	0.7	0.08

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #1, 2023: Performance of Participating Laboratories






Whole Blood Hg (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
	<b>Target</b>	<b>0.99</b>	<b>0.56</b>	<b>6.5</b>	<b>22.4</b>	<b>2.86</b>
103	ICP-MS/MS	1.04	0.759	7.11	22.6	2.92
107	ICP-MS/MS	0.94	0.56	6.75	22.62	2.73
110	ICP-MS	0.86	0.43	6.39	21.3	2.75
116	ICP-MS/MS	<1.50	<1.50	6.90	24.1	2.90
147	ICP-MS	0.991	0.564	6.58	23.7	2.85
264	ICP-MS	1.19	0.69	5.58	22.21	2.56
293	DRC/CC-ICP-MS	0.7	0.44	4.5	16.81	2.12
391	CV-AAS	0.0	0.22	1.7 ↓	5.03 ↓	22.29 ↑
597	ICP-MS/MS	0.909	0.573	6.16	21.0	2.70
605	ICP-MS	1.03	0.570	6.69	23.1	2.96
606	ICP-MS/MS	1.15	0.576	7.23	26.0	3.48
686	ICP-MS	0.982	0.530	6.72	23.4	2.95

Based on the grading criteria for Hg in Whole Blood, 95% of results were satisfactory, with 1 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Whole Blood Hg



**Legend:**

○ HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±3 µg/L or ±30% around the target value, whichever is greater; thus, it is fixed at ±3 µg/L at concentrations less than or equal to 10 µg/L.



## Results for Event #1, 2023: Summary Statistics

Whole Blood Mn (µg/L)					
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	18.4	23.0	17.6	26.6	13
<b>Upper Limit</b>	21.5	26.9	20.6	31.1	16
<b>Lower Limit</b>	15.3	19.1	14.6	22.1	10
<b>Arithmetic SD (s)</b>	2.3	2.2	2.1	2.4	3
<b>Arithmetic RSD (%)</b>	13	9.6	12	9.0	22
<b>Number of Sample Measurements (N)</b>	8	8	8	8	8

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ . These quality specifications were recently proposed by a network of Trace Element PT program organizers (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry Laboratory Medicine 2016; 54(12): 1921-1928).



### Results for Event #1, 2023: Performance of Participating Laboratories






Whole Blood Mn (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
	<b>Target</b>	<b>18.4</b>	<b>23.0</b>	<b>17.6</b>	<b>26.6</b>	<b>13</b>
107	ICP-MS/MS	18.91	22.88	17.10	26.93	11.77
110	ICP-MS	16.3	22.0	16.6	25.9	11.0
147	ICP-MS	20.6	26.6	20.4	31.4 ↑	14.6
264	ICP-MS	21.87 ↑	25.80	20.88 ↑	28.08	15.85
293	DRC/CC-ICP-MS	14.9 ↓	20.04	14.7	24.0	9.77 ↓
391	ICP-MS	19.6	22.17	18.2	24.1	17.69 ↑
597	ICP-MS/MS	16.8	21.7	16.3	25.5	11.0
606	ICP-MS/MS	18.1	22.6	16.9	26.6	11.8

Based on the grading criteria for Mn in Whole Blood, 85% of results were satisfactory, with 2 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Whole Blood Mn



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ .



## Results for Event #1, 2023: Summary Statistics

Whole Blood Pb (µg/dL)					
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
<b>Target (Robust Mean (x*))</b>	21.9	0.66	2.22	0.96	4.14
<b>Upper Limit</b>	24.1	2.66	4.22	2.96	6.14
<b>Lower Limit</b>	19.7	0.00	0.22	0.00	2.14
<b>Robust SD (s*)</b>	0.4	0.03	0.07	0.12	0.13
<b>Robust RSD (%)</b>	1.8	5.1	3.2	12	3.1
<b>Number of Sample Measurements (N)</b>	13	7	12	8	12
<b>Standard Uncertainty (u)</b>	0.2	NA	0.02	NA	0.05

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/dL}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ . These quality specifications are recommended by the Clinical Laboratory Standards Institute (CLSI, C40-A2) and have been proposed for use in proficiency testing programs approved under CLIA by the Centers for Medicare and Medicaid Services (CMS) in the USA. (<https://clsi.org/standards/products/clinical-chemistry-and-toxicology/documents/c40/>)

An arithmetic mean, SD, RSD and n are provided for samples BE23-02 and BE23-04.



### Results for Event #1, 2023: Performance of Participating Laboratories

Whole Blood Pb (µg/dL)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
	Target	21.9	0.66	2.22	0.96	4.14
103	ICP-MS/MS	21.5	0.689	2.19	0.963	4.07
107	ICP-MS/MS	21.999	0.660	2.172	0.938	4.034
110	ICP-MS	21.7	0.67	2.18	0.91	4.12
116	ICP-MS/MS	22.4	<3.00	2.27	<3.00	4.30
147	ICP-MS	22.4	0.700	2.32	1.13	4.27
264	ICP-MS	21.82	0.61	2.17	0.85	4.15
293	DRC/CC-ICP-MS	22.33	0.62	2.27	0.8	4.14
343	ASV-LeadCare	18.7 ↓	<3.3	<3.3	<3.3	<3.3
391	ETAAS-Z	22.47	*1.22	3.26	*1.5	4.87
597	ICP-MS/MS	21.4	0.656	2.15	0.935	4.03
605	ICP-MS	22.0	<1.00	2.28	1.13	4.13
606	ICP-MS/MS	21.9	<1.00	2.15	<1.00	3.92
686	ICP-MS	22.0	<1.00	2.20	<1.00	4.16






Based on the grading criteria for Pb in Whole Blood, 98% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.





# Results for Event #1, 2023: Summary Figures

## Whole Blood Pb



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/dL}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ .



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Mo (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
103	ICP-MS/MS	2.34	<1.50	3.71	7.80	<1.50
110	ICP-MS/MS	2.71	0.64	3.85	8.06	1.11
147	ICP-MS	2.63	0.676	4.06	8.40	1.04
264	ICP-MS	<0.01	<0.01	<0.01	*1.07	<0.01
442	DRC/CC-ICP-MS	2.43	0.546	3.61	7.59	0.916
597	ICP-MS/MS	2.45	0.524	3.74	7.66	0.882

### Summary Statistics






	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	2.51	0.60	3.79	7.9	0.99
<b>Arithmetic SD (s)</b>	0.15	0.07	0.17	0.3	0.11
<b>Arithmetic RSD (%)</b>	6.0	12	4.5	4.2	11
<b>Number of Sample Measurements (N)</b>	5	4	5	5	4

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Whole Blood Mo



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Sb (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
103	ICP-MS/MS	2.90	*6.17	0.391	0.752	1.61
110	ICP-MS/MS	3.53	8.01	0.54	0.99	2.21
147	ICP-MS	3.82	8.51	0.564	1.08	2.39
264	ICP-MS	3.62	8.21	0.43	0.85	2.44
293	DRC/CC-ICP-MS	3.2	7.6	0.5	1.0	2.0
442	DRC/CC-ICP-MS	3.79	8.37	0.544	1.08	2.35
597	ICP-MS/MS	3.58	8.35	0.538	1.02	2.23

### Summary Statistics






	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	3.5	8.2	0.50	0.97	2.2
Arithmetic SD (s)	0.3	0.3	0.07	0.12	0.3
Arithmetic RSD (%)	9.5	4.0	14	12	13
Number of Sample Measurements (N)	7	6	7	7	7

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Whole Blood Sb



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Se (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
103	ICP-MS/MS	145	119	238	331	168
107	ICP-MS/MS	154.6	121.5	247.8	346.3	179.3
110	ICP-MS/MS	147	115	244	316	178
147	ICP-MS	145	116	242	339	172
264	ICP-MS	160.8	124.9	261.1	350.8	179.5
293	DRC/CC-ICP-MS	146	115	216	319	162
597	ICP-MS/MS	145	116	245	322	175

### Summary Statistics






	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	149	118	242	332	173
Arithmetic SD (s)	6	4	14	14	7
Arithmetic RSD (%)	4.0	3.2	5.8	4.2	4.0
Number of Sample Measurements (N)	7	7	7	7	7

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Summary Figures

### Whole Blood Se



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood TI (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
103	ICP-MS/MS	0.680	0.117	1.12	0.369	3.28
110	ICP-MS/MS	0.70	0.12	1.13	0.36	3.45
147	ICP-MS	0.726	0.132	1.17	0.413	3.52
264	ICP-MS	0.67	0.11	1.11	0.33	3.44
293	DRC/CC-ICP-MS	0.68	0.12	1.100	0.35	3.36
597	ICP-MS/MS	0.665	0.117	1.10	0.345	3.31

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	0.687	0.119	1.12	0.36	3.39
<b>Arithmetic SD (s)</b>	0.023	0.007	0.03	0.03	0.09
<b>Arithmetic RSD (%)</b>	3.3	5.9	2.3	8.0	2.7
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6






\*Denotes a statistical Outlier.





# Results for Event #1, 2023: Summary Figures

## Whole Blood TI



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Ba (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	1.54	6.29	2.83	1.89	3.91
147	ICP-MS	1.34	6.51	2.98	2.77	3.80
597	ICP-MS/MS	1.32	5.99	2.50	1.60	3.88

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	1.40	6.3	2.8	2.1	3.86
Arithmetic SD (s)	0.12	0.3	0.3	0.6	0.06
Arithmetic RSD (%)	8.6	4.8	9.0	29	1.6
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Whole Blood Be (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	5.68	0.64	2.43	4.44	1.00
147	ICP-MS	5.32	<0.991	2.25	4.27	1.04
597	ICP-MS/MS	5.48	0.655	2.25	4.12	0.917
Summary Statistics						
		BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )		5.5	0.647	2.31	4.28	0.99
Arithmetic SD (s)		0.2	0.011	0.10	0.16	0.06
Arithmetic RSD (%)		3.3	1.7	4.3	3.7	6.1
Number of Sample Measurements (N)		3	2	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Bi (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
147	ICP-MS	<0.0334	<0.0334	0.374	<0.0334	0.362
597	ICP-MS/MS	<0.0299	<0.0299	0.350	<0.0299	0.364

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	NA	NA	0.36	NA	0.363
Arithmetic SD (s)	NA	NA	0.02	NA	0.001
Arithmetic RSD (%)	NA	NA	4.7	NA	0.39
Number of Sample Measurements (N)	NA	NA	2	NA	2

\*Denotes a statistical Outlier.

Statistical data was not calculated for BE23-01, BE23-02 and BE23-04 based on a lack of consensus among participating labs.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Cs (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	1.08	1.65	1.24	1.66	1.23
147	ICP-MS	1.05	1.61	1.16	1.62	1.20
597	ICP-MS/MS	1.05	1.61	1.19	1.62	1.22

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	1.06	1.62	1.20	1.63	1.217
Arithmetic SD (s)	0.02	0.02	0.04	0.02	0.015
Arithmetic RSD (%)	1.6	1.4	3.3	1.4	1.2
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Cu (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	933	1540	860	721	1840
147	ICP-MS	978	1645	915	807	1970
597	ICP-MS/MS	940	1590	869	728	1870

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	950	1590	880	750	1890
Arithmetic SD (s)	24	50	30	50	70
Arithmetic RSD (%)	2.5	3.1	3.4	6.7	3.7
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Whole Blood Ni (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
103	ICP-MS/MS	2.15	<1.50	11.2	4.98	1.62
110	ICP-MS/MS	2.79	1.20	12.3	5.22	2.03
147	ICP-MS	2.63	0.975	12.5	5.40	1.77
597	ICP-MS/MS	2.51	0.833	11.8	4.63	1.64
Summary Statistics						
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05	
Arithmetic Mean ( $\bar{x}$ )	2.5	1.0	12.0	5.1	1.8	
Arithmetic SD (s)	0.3	0.2	0.6	0.3	0.2	
Arithmetic RSD (%)	11	19	5.0	5.9	11	
Number of Sample Measurements (N)	4	3	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Whole Blood Pt ( $\mu\text{g/L}$ )						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	0.474	5.52	2.11	5.08	0.822
293	DRC/CC-ICP-MS	0.44	4.85	1.67	4.20	0.69

Summary Statistics						
	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05	
Arithmetic Mean ( $\bar{x}$ )	0.46	5.2	1.9	4.6	0.76	
Arithmetic SD (s)	0.02	0.5	0.3	0.6	0.09	
Arithmetic RSD (%)	4.3	9.6	16	13	12	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.





## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Sn ( $\mu\text{g/L}$ )

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	5.83	0.75	1.67	2.95	0.36
147	ICP-MS	5.86	0.757	1.71	3.12	0.365
597	ICP-MS/MS	5.59	0.68	1.55	2.76	0.312

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	5.76	0.73	1.64	2.9	0.35
Arithmetic SD (s)	0.15	0.04	0.08	0.2	0.03
Arithmetic RSD (%)	2.6	5.5	4.9	6.1	8.6
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Whole Blood Sr ( $\mu\text{g/L}$ )						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
103	ICP-MS/MS	19.4	22.4	31.3	23.1	31.5
147	ICP-MS	19.8	23.4	33.2	24.4	32.9
597	ICP-MS/MS	20.2	24.1	32.5	23.4	32.6
Summary Statistics						
		BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )		19.8	23.3	32.3	23.6	32.3
Arithmetic SD (s)		0.4	0.9	1.0	0.7	0.7
Arithmetic RSD (%)		2.0	3.9	3.1	3.0	2.2
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Ti (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
200	DRC/CC-ICP-MS	5.9	4.1	10.9	7.3	6.1
442	ICP-MS/MS	3.35	1.20	8.83	5.75	2.40
597	ICP-MS/MS	4.31	2.32	9.04	6.40	3.30

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	4.5	NA	9.6	6.5	NA
Arithmetic SD (s)	1.3	NA	1.1	0.8	NA
Arithmetic RSD (%)	29	NA	11	12	NA
Number of Sample Measurements (N)	3	NA	3	3	NA

\*Denotes a statistical Outlier.

Statistical data was not calculated for BE23-02 and BE23-05 based on a lack of consensus among participating labs.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Whole Blood U (µg/L)						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
103	ICP-MS/MS	<0.0500	0.142	<0.0500	0.110	0.0927
110	ICP-MS/MS	0.0284	0.158	0.0207	0.123	0.0966
147	ICP-MS	0.0279	0.152	0.0229	0.124	0.0914
597	ICP-MS/MS	0.0285	0.161	0.0202	0.115	0.0938
Summary Statistics						
		BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )		0.0283	0.153	0.0213	0.118	0.094
Arithmetic SD (s)		0.0003	0.008	0.0014	0.007	0.002
Arithmetic RSD (%)		1.1	5.2	6.6	5.9	2.4
Number of Sample Measurements (N)		3	4	3	4	4

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood V (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	0.11	1.47	4.87	0.63	2.20
147	DRC/CC-ICP-MS	0.0878	1.48	4.77	0.622	2.34
597	ICP-MS/MS	0.0983	1.42	4.78	0.580	2.18

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	0.099	1.46	4.81	0.61	2.24
Arithmetic SD (s)	0.011	0.03	0.06	0.03	0.09
Arithmetic RSD (%)	11	2.1	1.2	4.9	4.0
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Whole Blood W ( $\mu\text{g/L}$ )						
Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	2.29	0.26	0.54	1.25	3.30
200	ICP-MS	2.87	0.30	0.700	1.53	4.27
597	ICP-MS/MS	2.28	0.225	0.504	1.29	3.26
Summary Statistics						
		BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )		2.5	0.26	0.58	1.36	3.6
Arithmetic SD (s)		0.3	0.04	0.10	0.15	0.6
Arithmetic RSD (%)		12	15	17	11	17
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Whole Blood Zn (µg/L)

Lab Code	Method	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
110	ICP-MS/MS	5630	4760	6990	4720	6960
147	ICP-MS	5660	5000	7190	5137	7124
597	ICP-MS/MS	5520	4830	6930	4700	6870

### Summary Statistics

	BE23-01	BE23-02	BE23-03	BE23-04	BE23-05
Arithmetic Mean ( $\bar{x}$ )	5600	4860	7040	4850	6980
Arithmetic SD (s)	70	120	140	250	130
Arithmetic RSD (%)	1.3	2.5	2.0	5.2	1.9
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



Results for Event #1, 2023:
Additional Elements in Whole Blood

Whole Blood Ag (µg/L)

Table with 7 columns: Lab Code, Method, BE23-01, BE23-02, BE23-03, BE23-04, BE23-05. Row 1: 147, ICP-MS, 0.167, <0.151, <0.151, <0.151, <0.151

Whole Blood Al (µg/L)

Table with 7 columns: Lab Code, Method, BE23-01, BE23-02, BE23-03, BE23-04, BE23-05. Row 1: 147, ICP-MS, <4.86, <4.86, <4.86, <4.86, <4.86. Row 2: 597, ICP-MS/MS, 9.38, <3.88, 6.17, 7.67, 10.3

Whole Blood I (µg/L)

Table with 7 columns: Lab Code, Method, BE23-01, BE23-02, BE23-03, BE23-04, BE23-05. Row 1: 147, ICP-MS, 26.5, 24.4, 20.3, 26.1, 20.0

Whole Blood Li (µg/L)

Table with 7 columns: Lab Code, Method, BE23-01, BE23-02, BE23-03, BE23-04, BE23-05. Row 1: 147, ICP-MS, 0.784, 0.819, 0.854, 0.826, 0.763

Whole Blood Mg (µg/L)

Table with 7 columns: Lab Code, Method, BE23-01, BE23-02, BE23-03, BE23-04, BE23-05. Row 1: 597, ICP-MS/MS, 27000, 27800, 34400, 27200, 34500

Whole Blood Te (µg/L)

Table with 7 columns: Lab Code, Method, BE23-01, BE23-02, BE23-03, BE23-04, BE23-05. Row 1: 110, ICP-MS/MS, 0.03, 0.01, 0.04, 0.02, 0.03. Row 2: 147, ICP-MS, <0.0561, <0.0561, <0.0561, <0.0561, <0.0561

Whole Blood Th (µg/L)

Table with 7 columns: Lab Code, Method, BE23-01, BE23-02, BE23-03, BE23-04, BE23-05. Row 1: 147, ICP-MS, <0.0255, <0.0255, <0.0255, <0.0255, <0.0255. Row 2: 597, ICP-MS/MS, 0.0157, <0.00741, <0.00741, <0.00741, <0.00741





**Department  
of Health**

**Wadsworth  
Center**

**Event #1, 2023**

**Trace Elements in  
Urine**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #1, 2023: Trace Elements in Urine

### PT Materials

Urine was collected from volunteer donors into polyethylene containers and stored at 4°C. Following collection, urine was acidified to 1% (v/v) with nitric acid and mixed with a sulfamic acid solution (stock solution contained 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100) to a final concentration of 1% (v/v) to stabilize Hg. Urine was stored frozen at -80°C pending further preparation. The urine was thawed at room temperature and precipitated salts removed by centrifugation. Urine supernatants were combined into five separate pools. Each urine pool was supplemented with arsenic (As), barium (Ba), beryllium (Be), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), thallium (Tl), uranium (U), aluminum (Al), cesium (Cs), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), strontium (Sr), tellurium (Te), titanium (Ti), vanadium (V), tungsten (W), and zinc (Zn). PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Eleven elements in urine are formally graded: As, Ba, Be, Cd, Co, Cr, Hg, Mn, Pb, Tl, and U. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 22 elements were reported by at least one participant: Ag, Al, Bi, Cs, Cu, Fe, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #1, 2023: Summary Statistics

	Urine As (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	22.8	7.6	34.9	44	1.2
<b>Upper Limit</b>	28.8	13.6	41.9	53	7.2
<b>Lower Limit</b>	16.8	1.6	27.9	35	0.0
<b>Robust SD (s*)</b>	1.6	0.6	2.6	4	0.2
<b>Robust RSD (%)</b>	7.0	7.9	7.4	9.3	20
<b>Number of Sample Measurements (N)</b>	15	15	15	15	10
<b>Standard Uncertainty (u)</b>	0.5	0.2	0.8	1	0.1

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2023: Performance of Participating Laboratories






Urine As (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
	Target	22.8	7.6	34.9	44	1.2
103	ICP-MS/MS	21.6	7.02	32.4	40.8	1.06
107	DRC/CC-ICP-MS	21.87	6.81	33.82	36.98	0.96
110	DRC/CC-ICP-MS	24.0	7.87	37.1	46.9	1.31
116	ICP-MS/MS	25.4	7.51	35.3	45.2	<5.00
147	ICP-MS	21.8	7.18	33.9	42.1	0.876
220	DRC/CC-ICP-MS	24.7	8.60	38.5	49.0	<2.0
264	ICP-MS	23.55	8.13	36.16	45.28	1.47
293	DRC/CC-ICP-MS	21.32	7.11	32.13	40.63	1.07
391	ICP-MS	38.401 ↑	19.848 ↑	55.662 ↑	72.918 ↑	8.199 ↑
399	DRC/CC-ICP-MS	21.6	7.10	32.7	40.4	1.15
442	ICP-MS/MS	21.6	7.33	32.7	41	1.21
597	ICP-MS/MS	22.6	7.93	36.6	45.1	1.36
605	ICP-MS	22.9	7.96	35.2	45.9	<2.00
606	ICP-MS/MS	23.9	7.97	35.9	47.4	<2.00
686	DRC/CC-ICP-MS	19.9	7.61	31.9	41.4	<6.00

Based on the grading criteria for As in Urine, 93% of results were satisfactory, with 1 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine As



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±6 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±6 µg/L at concentrations less than or equal to 30 µg/L.



### Results for Event #1, 2023: Summary Statistics

	Urine Ba (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	0.69	5.2	7.1	1.10	2.87
<b>Upper Limit</b>	1.69	6.3	8.5	2.10	3.87
<b>Lower Limit</b>	0.00	4.2	5.7	0.10	1.87
<b>Robust SD (s*)</b>	0.04	0.4	0.4	0.07	0.20
<b>Robust RSD (%)</b>	5.8	7.5	5.6	6.4	7.0
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.02	0.1	0.1	0.03	0.07

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



## Results for Event #1, 2023: Performance of Participating Laboratories






		Urine Ba (µg/L)				
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Target		0.69	5.2	7.1	1.10	2.87
107	ICP-MS	0.625	4.819	6.776	1.064	2.692
110	ICP-MS	0.68	5.57	7.64	1.19	3.12
116	ICP-MS/MS	0.969	4.78	6.89	0.984	2.6
147	ICP-MS	0.761	5.32	7.45	1.12	3.08
220	ICP-MS	0.71	5.54	7.65	1.21	3.10
264	ICP-MS	0.64	4.82	6.65	1.07	2.68
399	ICP-MS/MS	0.712	5.20	7.12	1.13	2.86
597	ICP-MS/MS	0.668	4.89	6.78	1.07	2.83
605	ICP-MS	0.700	5.34	7.18	1.12	2.89
606	ICP-MS/MS	0.484	5.75	6.92	0.994	2.85
686	ICP-MS	0.690	5.27	7.18	1.13	2.89

Based on the grading criteria for Ba in Urine, 100% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine Ba



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 5 µg/L.





## Results for Event #1, 2023: Summary Statistics

	Urine Be (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	1.44	5.20	0.81	0.60	3.35
<b>Upper Limit</b>	2.44	6.24	1.81	1.60	4.35
<b>Lower Limit</b>	0.44	4.16	0.00	0.00	2.35
<b>Robust SD (s*)</b>	0.04	0.23	0.03	0.03	0.18
<b>Robust RSD (%)</b>	2.8	4.4	3.7	4.7	5.4
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.02	0.09	0.01	0.01	0.07

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



## Results for Event #1, 2023: Performance of Participating Laboratories






		Urine Be (µg/L)				
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Target		1.44	5.20	0.81	0.60	3.35
107	ICP-MS	1.415	4.886	0.798	0.571	3.123
110	ICP-MS	1.45	5.24	0.81	0.61	3.35
116	ICP-MS/MS	1.41	4.91	0.743	0.576	3.12
147	ICP-MS	1.57	5.46	0.823	0.587	3.57
220	ICP-MS	1.53	5.40	0.87	0.65	3.52
264	ICP-MS	1.42	5.06	0.80	0.57	3.25
293	ICP-MS	1.41	5.31	0.78	0.59	3.28
399	ICP-MS/MS	1.37	5.00	0.798	0.605	3.21
597	ICP-MS/MS	1.42	5.25	0.805	0.576	3.41
605	ICP-MS	1.51	5.42	0.847	0.637	3.49
686	ICP-MS	1.47	5.26	0.843	0.631	3.50

Based on the grading criteria for Be in Urine, 100% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine Be



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



## Results for Event #1, 2023: Summary Statistics

	Urine Cd (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	1.74	0.388	3.43	0.75	2.63
<b>Upper Limit</b>	2.74	1.388	4.43	1.75	3.63
<b>Lower Limit</b>	0.74	0.000	2.43	0.00	1.63
<b>Robust SD (s*)</b>	0.06	0.022	0.12	0.04	0.10
<b>Robust RSD (%)</b>	3.4	5.7	3.5	5.8	3.8
<b>Number of Sample Measurements (N)</b>	16	15	16	15	16
<b>Standard Uncertainty (u)</b>	0.02	0.007	0.04	0.01	0.03

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2023: Performance of Participating Laboratories






		Urine Cd (µg/L)				
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Target		1.74	0.388	3.43	0.75	2.63
103	ICP-MS/MS	1.74	0.388	3.57	0.741	2.69
107	DRC/CC-ICP-MS	1.758	0.366	3.279	0.699	2.496
110	ICP-MS	1.73	0.389	3.27	0.735	2.53
116	ICP-MS/MS	2.31	0.386	3.70	0.776	2.75
147	ICP-MS	1.73	0.361	3.38	0.714	2.68
220	ICP-MS	1.76	0.40	3.42	0.79	2.65
264	ICP-MS	1.83	0.36	3.65	0.81	2.75
293	DRC/CC-ICP-MS	1.74	0.38	3.43	0.73	2.57
324	ICP-MS	1.754	<1	3.305	<1	2.510
391	ICP-MS	1.675	0.538	3.353	0.844	2.581
399	DRC/CC-ICP-MS	1.69	0.384	3.46	0.734	2.65
442	ICP-MS/MS	1.58	0.412	3.55	0.734	2.6
597	ICP-MS/MS	1.65	0.367	3.41	0.696	2.55
605	ICP-MS	1.87	0.401	3.49	0.841	2.85
606	ICP-MS/MS	1.67	0.419	3.35	0.772	2.62
686	ICP-MS	1.74	0.391	3.43	0.747	2.63

Based on the grading criteria for Cd in Urine, 100% of results were satisfactory, with 0 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine Cd



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 6.6 µg/L.



## Results for Event #1, 2023: Summary Statistics

	Urine Co (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	1.64	0.80	5.02	1.04	2.98
<b>Upper Limit</b>	3.14	2.30	6.52	2.54	4.48
<b>Lower Limit</b>	0.14	0.00	3.52	0.00	1.48
<b>Robust SD (s*)</b>	0.08	0.07	0.24	0.12	0.18
<b>Robust RSD (%)</b>	4.9	8.8	4.8	12	6.0
<b>Number of Sample Measurements (N)</b>	13	12	13	13	13
<b>Standard Uncertainty (u)</b>	0.03	0.03	0.08	0.04	0.06

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #1, 2023: Performance of Participating Laboratories

		Urine Co (µg/L)				
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Target		1.64	0.80	5.02	1.04	2.98
103	ICP-MS/MS	1.53	0.722	4.76	0.935	2.78
107	ICP-MS	1.567	0.731	4.815	0.918	2.810
110	ICP-MS	1.70	0.86	5.16	1.14	3.02
147	ICP-MS	1.62	0.784	4.95	0.984	2.92
220	ICP-MS	1.78	0.92	5.69	1.26	3.32
264	ICP-MS	1.69	0.81	5.16	1.04	3.00
293	DRC/CC-ICP-MS	1.67	0.78	5.03	0.98	3.03
324	ICP-MS	1.582	<1	4.759	1.095	2.786
391	ICP-MS	1.628	0.984	4.817	1.102	3.664
399	DRC/CC-ICP-MS	1.53	0.740	4.87	0.918	2.79
597	ICP-MS/MS	1.61	0.777	5.03	0.914	2.90
605	ICP-MS	1.70	0.890	5.25	1.13	3.04
606	ICP-MS/MS	1.69	0.764	5.26	1.14	3.11






Based on the grading criteria for Co in Urine, 100% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.





# Results for Event #1, 2023: Summary Figures

## Urine Co



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1.5 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 10 µg/L.



### Results for Event #1, 2023: Summary Statistics

	Urine Cr (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	3.4	0.52	5.6	2.3	3.5
<b>Upper Limit</b>	6.4	3.52	8.6	5.3	6.5
<b>Lower Limit</b>	0.4	0.00	2.6	0.0	0.5
<b>Robust SD (s*)</b>	0.4	0.18	0.5	0.4	0.3
<b>Robust RSD (%)</b>	11	35	8.9	17	9.1
<b>Number of Sample Measurements (N)</b>	11	8	11	11	11
<b>Standard Uncertainty (u)</b>	0.1	NA	0.2	0.2	0.1

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $15 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers

An arithmetic mean, SD, RSD and n are provided for sample UE23-02.



### Results for Event #1, 2023: Performance of Participating Laboratories






		Urine Cr (µg/L)				
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Target		3.4	0.52	5.6	2.3	3.5
103	ICP-MS/MS	2.77	<0.600	4.39	2.03	3.07
107	DRC/CC-ICP-MS	3.05	0.23	5.16	1.65	3.24
110	DRC/CC-ICP-MS	4.18	*1.07	6.49	3.26	4.23
116	ICP-MS/MS	3.59	0.360	5.59	2.06	3.27
147	DRC/CC-ICP-MS	3.64	0.460	5.98	2.19	3.72
264	ICP-MS	3.66	0.68	5.96	2.59	3.64
293	DRC/CC-ICP-MS	3.31	0.66	5.42	2.33	3.54
324	ICP-MS	3.685	*1.104	5.702	2.762	3.732
391	ICP-MS	2.502	0.781	5.261	2.51	3.712
597	ICP-MS/MS	3.29	0.499	5.43	2.04	3.01
605	ICP-MS	3.56	0.462	5.73	2.35	3.52

Based on the grading criteria for Cr in Urine, 100% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine Cr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±3 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±3 µg/L at concentrations less than or equal to 15 µg/L.



## Results for Event #1, 2023: Summary Statistics

	Urine Hg (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	4.5	2.9	11.0	19.6	1.05
<b>Upper Limit</b>	7.5	5.9	14.3	25.5	4.05
<b>Lower Limit</b>	1.5	0.0	7.7	13.7	0.00
<b>Robust SD (s*)</b>	0.5	0.3	1.1	2.3	0.23
<b>Robust RSD (%)</b>	11	12	10	12	22
<b>Number of Sample Measurements (N)</b>	13	13	13	13	12
<b>Standard Uncertainty (u)</b>	0.2	0.1	0.4	0.8	0.08

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2023: Performance of Participating Laboratories






		Urine Hg (µg/L)				
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Target		4.5	2.9	11.0	19.6	1.05
103	ICP-MS/MS	4.57	2.91	10.7	19.9	1.04
107	DRC/CC-ICP-MS	4.04	4.48	15.66 ↑	16.56	4.10 ↑
110	ICP-MS	4.33	2.63	10.1	18.3	0.808
116	ICP-MS/MS	3.71	2.26	8.78	16.7	0.761
147	ICP-MS	4.37	2.75	10.5	18.2	1.02
264	ICP-MS	10.30 ↑	3.73	11.87	22.42	1.24
293	DRC/CC-ICP-MS	4.07	2.67	9.92	18.51	1.13
391	ICP-MS	5.45	3.082	13.473	21.965	0.824
442	ICP-MS/MS	4.55	2.91	10.4	19.5	0.862
597	ICP-MS/MS	4.42	2.63	10.5	18.5	1.12
605	ICP-MS	4.78	2.74	11.5	22.1	<1.00
606	ICP-MS/MS	4.83	2.90	11.7	20.5	1.29
686	ICP-MS	4.90	3.17	11.6	21.7	1.09

Based on the grading criteria for Hg in Urine, 95% of results were satisfactory, with 1 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine Hg



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±3 µg/L or ±30% around the target value, whichever is greater; thus, it is fixed at ±3 µg/L at concentrations less than or equal to 10 µg/L.



## Results for Event #1, 2023: Summary Statistics

	Urine Mn (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	1.08	6.2	0.53	2.91	3.9
<b>Upper Limit</b>	1.63	7.8	1.08	3.64	4.9
<b>Lower Limit</b>	0.53	4.7	0.00	2.18	2.9
<b>Robust SD (s*)</b>	0.13	0.4	0.07	0.22	0.3
<b>Robust RSD (%)</b>	12	5.8	13	7.6	7.9
<b>Number of Sample Measurements (N)</b>	14	14	13	14	14
<b>Standard Uncertainty (u)</b>	0.04	0.1	0.02	0.07	0.1

The acceptable range is based on quality specifications:  $\pm 0.55 \mu\text{g/L}$  or  $\pm 25\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.55 \mu\text{g/L}$  at concentrations less than or equal to  $2.2 \mu\text{g/L}$ . Quality specifications for Mn are consistent with those used by other External Quality Assessment Schemes for trace elements. (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry and Laboratory Medicine.2016; 54(12): 1921-1928).





### Results for Event #1, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine Mn (µg/L)				
		UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
	<b>Target</b>	<b>1.08</b>	<b>6.2</b>	<b>0.53</b>	<b>2.91</b>	<b>3.9</b>
103	ICP-MS/MS	0.974	5.77	0.468	2.63	3.67
107	DRC/CC-ICP-MS	0.973	5.731	0.479	2.677	3.518
110	DRC/CC-ICP-MS	1.08	6.46	0.53	2.93	4.02
116	ICP-MS/MS	2.72 ↑	6.45	0.575	2.87	4.36
147	DRC/CC-ICP-MS	1.15	6.65	0.582	3.18	4.29
220	DRC/CC-ICP-MS	1.06	6.43	0.56	2.96	4.03
264	ICP-MS	1.07	6.36	0.55	2.87	4.02
293	DRC/CC-ICP-MS	0.96	6.06	0.43	2.75	3.79
324	ICP-MS	1.069	6.253	<1	3.005	3.875
391	ICP-MS	1.499	4.232 ↓	0.176	3.703 ↑	2.997
399	DRC/CC-ICP-MS	0.923	5.77	0.536	2.66	3.67
597	ICP-MS/MS	1.07	6.05	0.522	2.89	3.89
605	ICP-MS	1.06	6.62	0.748	2.97	3.93
606	ICP-MS/MS	2.03 ↑	6.29	0.604	4.40 ↑	4.92 ↑

Based on the grading criteria for Mn in Urine, 91% of results were satisfactory, with 2 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine Mn



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 0.55 \mu\text{g/L}$  or  $\pm 25\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.55 \mu\text{g/L}$  at concentrations less than or equal to  $2.2 \mu\text{g/L}$ .



## Results for Event #1, 2023: Summary Statistics

	Urine Pb (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	1.85	7.75	0.47	4.08	11.1
<b>Upper Limit</b>	2.85	9.30	1.47	5.08	13.3
<b>Lower Limit</b>	0.85	6.20	0.00	3.08	8.9
<b>Robust SD (s*)</b>	0.07	0.22	0.06	0.13	0.4
<b>Robust RSD (%)</b>	3.8	2.8	13	3.2	3.5
<b>Number of Sample Measurements (N)</b>	16	16	15	16	16
<b>Standard Uncertainty (u)</b>	0.02	0.07	0.02	0.04	0.1

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2023: Performance of Participating Laboratories






		Urine Pb (µg/L)				
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Target		1.85	7.75	0.47	4.08	11.1
103	ICP-MS/MS	1.88	7.72	0.438	4.21	11.2
107	ICP-MS	1.874	7.834	0.472	4.065	11.416
110	ICP-MS	1.78	7.59	0.43	3.93	10.7
116	ICP-MS/MS	1.80	7.27	0.415	3.93	10.2
147	ICP-MS	1.87	7.77	0.450	4.06	10.9
220	ICP-MS	1.81	7.90	0.39	4.18	11.4
264	ICP-MS	2.07	8.15	0.50	4.16	11.53
293	DRC/CC-ICP-MS	1.85	7.68	0.86	3.96	11.1
324	ICP-MS	1.875	7.766	<1	4.112	10.756
391	ICP-MS	1.682	7.083	0.538	3.773	10.133
399	ICP-MS/MS	1.83	7.71	0.523	4.07	11.1
442	ICP-MS/MS	1.94	7.89	0.497	4.18	11.3
597	ICP-MS/MS	1.78	7.60	0.453	3.99	10.9
605	ICP-MS	1.93	8.28	0.457	4.28	11.7
606	ICP-MS/MS	1.78	7.61	0.409	4.08	11.4
686	ICP-MS	1.89	7.95	0.518	4.17	11.3

Based on the grading criteria for Pb in Urine, 100% of results were satisfactory, with 0 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine Pb



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



## Results for Event #1, 2023: Summary Statistics

	Urine TI (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	3.85	0.232	1.60	0.76	3.05
<b>Upper Limit</b>	4.62	0.432	1.92	0.96	3.66
<b>Lower Limit</b>	3.08	0.032	1.28	0.56	2.44
<b>Robust SD (s*)</b>	0.09	0.025	0.06	0.03	0.12
<b>Robust RSD (%)</b>	2.3	11	3.8	4.6	3.9
<b>Number of Sample Measurements (N)</b>	14	14	14	14	14
<b>Standard Uncertainty (u)</b>	0.03	0.008	0.02	0.01	0.04

The acceptable range is based on quality specifications:  $\pm 0.2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.2 \mu\text{g/L}$  at concentrations less than or equal to  $1 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine TI (µg/L)				
		UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
	<b>Target</b>	<b>3.85</b>	<b>0.232</b>	<b>1.60</b>	<b>0.76</b>	<b>3.05</b>
103	ICP-MS/MS	3.86	0.219	1.62	0.787	3.07
107	ICP-MS	3.872	0.226	1.583	0.762	2.996
110	ICP-MS	3.81	0.233	1.57	0.751	3.03
116	ICP-MS/MS	3.49	0.384	1.54	0.650	2.68
147	ICP-MS	4.27	0.258	1.74	0.816	3.41
220	ICP-MS	3.83	0.20	1.58	0.74	3.00
264	ICP-MS	4.00	0.24	1.66	0.79	3.18
293	DRC/CC-ICP-MS	4.05	0.25	1.6	0.78	3.12
391	ICP-MS	3.431	0.206	1.434	0.662	2.696
399	ICP-MS/MS	3.83	0.230	1.57	0.738	3.03
597	ICP-MS/MS	3.88	0.245	1.64	0.767	3.16
605	ICP-MS	4.04	0.253	1.70	0.777	3.12
606	ICP-MS/MS	3.78	0.226	1.61	0.759	2.96
686	ICP-MS	3.81	0.148	1.52	0.677	3.05

Based on the grading criteria for TI in Urine, 100% of results were satisfactory, with 0 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine TI



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±0.2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.2 µg/L at concentrations less than or equal to 1 µg/L.





### Results for Event #1, 2023: Summary Statistics

	Urine U (µg/L)				
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Target (Robust Mean (x*))</b>	0.0217	0.154	0.005	0.127	0.100
<b>Upper Limit</b>	0.0517	0.185	0.035	0.157	0.130
<b>Lower Limit</b>	0.0000	0.123	0.000	0.097	0.070
<b>Robust SD (s*)</b>	0.0020	0.008	0.001	0.009	0.008
<b>Robust RSD (%)</b>	9.2	5.2	9.8	7.1	8.4
<b>Number of Sample Measurements (N)</b>	14	14	6	14	14
<b>Standard Uncertainty (u)</b>	0.0007	0.003	NA	0.003	0.003

The acceptable range is based on quality specifications: ±0.03 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.03 µg/L at concentrations less than or equal to 0.15 µg/L. These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

An arithmetic mean, SD, RSD and n are provided for sample UE23-03.



Results for Event #1, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine U (µg/L)				
		UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
	<b>Target</b>	<b>0.0217</b>	<b>0.154</b>	<b>0.005</b>	<b>0.127</b>	<b>0.100</b>
103	ICP-MS/MS	0.0205	0.154	<0.0200	0.123	0.0937
107	ICP-MS	0.138 ↑	1.350 ↑	*4.656 ↑	0.670 ↑	2.026 ↑
110	ICP-MS	0.0200	0.156	0.0051	0.123	0.100
116	ICP-MS/MS	0.0230	0.147	<0.0150	0.131	0.0930
147	ICP-MS	0.0203	0.150	<0.00809	0.118	0.0971
220	ICP-MS	0.021	0.153	0.0056	0.125	0.103
264	ICP-MS	0.02	0.14	<0.01	0.12	0.09
324	ICP-MS	0.019	0.143	0.005	0.115	0.086
391	ICP-MS	0.034	0.169	*0.028	0.151	0.135 ↑
399	ICP-MS/MS	0.023	0.176	0.006	0.140	0.110
597	ICP-MS/MS	0.0228	0.152	0.00478	0.131	0.101
605	ICP-MS	0.022	0.158	0.006	0.130	0.101
606	ICP-MS/MS	0.021	0.153	<0.005	0.121	0.099
686	ICP-MS	0.0219	0.159	<0.0150	0.126	0.102

Based on the grading criteria for U in Urine, 91% of results were satisfactory, with 1 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Urine U



### Legend:

○ HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±0.03 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.03 µg/L at concentrations less than or equal to 0.15 µg/L.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Urine AI (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
147	ICP-MS	<13.8	<13.8	<13.8	17.6	20.9
264	ICP-MS	7.07	9.78	8.47	19.68	20.75
293	DRC/CC-ICP-MS	7.82	12.95	8.9	19.43	17.27
324	ICP-MS	8.220	11.281	9.977	19.607	19.759
391	ICP-MS	11.823	12.567	11.434	19.151	20.788
597	ICP-MS/MS	4.94	6.38	<4.03	17.6	17.8

### Summary Statistics






	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	8	11	9.7	18.8	19.5
<b>Arithmetic SD (s)</b>	2	3	1.3	1.0	1.6
<b>Arithmetic RSD (%)</b>	31	25	13	5.3	8.2
<b>Number of Sample Measurements (N)</b>	5	5	4	6	6

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Urine AI



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics






Urine Cs (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
107	ICP-MS	3.35	8.26	1.63	2.82	10.81
110	ICP-MS	3.69	9.40	1.83	3.16	12.1
147	ICP-MS	3.51	8.73	1.75	2.86	11.1
220	ICP-MS	3.61	9.14	1.85	3.12	11.8
264	ICP-MS	3.30	8.41	1.66	2.87	10.70
399	ICP-MS/MS	3.46	8.86	1.73	2.90	11.1
597	ICP-MS/MS	3.26	8.37	1.67	2.87	10.7
605	ICP-MS	3.69	9.06	1.79	3.07	11.9
606	ICP-MS/MS	3.35	8.47	1.69	2.88	10.9
Summary Statistics						
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05	
Arithmetic Mean ( $\bar{x}$ )	3.47	8.7	1.73	2.95	11.2	
Arithmetic SD (s)	0.17	0.4	0.08	0.13	0.5	
Arithmetic RSD (%)	4.9	4.6	4.6	4.4	4.5	
Number of Sample Measurements (N)	9	9	9	9	9	

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Urine Cs



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2023: Laboratory Data and Summary Statistics

Urine Cu (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
110	ICP-MS	27.2	76.1	39.2	*16.5	*15.8
116	ICP-MS/MS	20.1	67.5	33.0	5.54	7.86
147	ICP-MS	28.9	83.2	43.2	*18.7	*17.2
264	ICP-MS	26.84	77.54	37.17	6.39	9.50
293	DRC/CC-ICP-MS	19.07	64.84	31.15	5.09	7.63
324	ICP-MS	20.729	71.618	33.826	6.018	8.821
391	ICP-MS	16.141	58.472	28.009	4.36	7.63
597	ICP-MS/MS	21.4	70.6	34.8	6.93	9.48

Summary Statistics						
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05	
Arithmetic Mean ( $\bar{x}$ )	23	71	35	5.7	8.5	
Arithmetic SD (s)	5	8	5	0.9	0.9	
Arithmetic RSD (%)	22	11	14	16	11	
Number of Sample Measurements (N)	8	8	8	6	6	






\*Denotes a statistical Outlier.





# Results for Event #1, 2023: Summary Figures

## Urine Cu



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Urine Mo (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
103	ICP-MS/MS	91.7	16.5	61.4	38.6	48.9
107	ICP-MS	89.84	16.37	60.29	38.90	45.32
110	ICP-MS	99.5	18.4	65.8	42.5	52.0
147	ICP-MS	91.0	16.8	61.5	38.1	47.1
220	ICP-MS	96.0	17.8	65.6	41.6	49.9
264	ICP-MS	80.18	17.09	53.75	37.29	40.05
293	DRC/CC-ICP-MS	91.36	17.18	59.61	37.5	47.64
324	ICP-MS	92.770	16.928	61.732	38.058	46.782
399	ICP-MS/MS	95.3	17.4	64.2	40.2	48.6
597	ICP-MS/MS	87.9	16.1	60.4	37.2	46.4
605	ICP-MS	97.1	18.2	65.2	41.7	49.1
606	ICP-MS/MS	94.2	17.3	62.4	39.6	49.0

### Summary Statistics






	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Robust Mean (x*)</b>	93	17.2	62	39.2	47.9
<b>Robust SD (s*)</b>	4	0.8	3	2.0	1.9
<b>Robust RSD (%)</b>	4.4	4.7	4.5	5.1	4.0
<b>Number of Sample Measurements (N)</b>	12	12	12	12	12
<b>Standard Uncertainty (u)</b>	1	0.3	1	0.7	0.7

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Urine Mo



### Legend:

- HHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Urine Ni (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
103	ICP-MS/MS	4.71	3.23	5.90	1.55	7.71
107	DRC/CC-ICP-MS	4.99	3.17	6.83	1.41	8.40
110	ICP-MS	5.93	4.11	8.02	4.49	9.21
147	ICP-MS	5.24	3.42	6.69	1.64	8.57
264	ICP-MS	5.45	3.45	6.95	1.89	8.76
293	DRC/CC-ICP-MS	5.23	3.36	6.84	2.67	8.23
324	ICP-MS	5.210	3.640	6.266	2.519	7.894
391	ICP-MS	8.191	4.91	6.631	2.219	12.202
442	DRC/CC-ICP-MS	5.3	3.31	6.96	1.71	8.03
597	ICP-MS/MS	5.26	3.29	6.52	1.94	8.41
605	ICP-MS	5.14	3.35	6.88	1.66	8.61

### Summary Statistics






	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Robust Mean (x*)</b>	5.25	3.39	6.77	2.0	8.5
<b>Robust SD (s*)</b>	0.18	0.15	0.23	0.5	0.5
<b>Robust RSD (%)</b>	3.4	4.4	3.4	25	5.9
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.07	0.06	0.09	0.2	0.2

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Urine Ni



### Legend:

● HHEAR Labs ● Other Labs

Horizontal purple line = robust mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics






Urine Pt (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
107	ICP-MS	0.2953	2.0123	0.6916	1.0894	3.0068
110	ICP-MS	0.308	2.01	0.681	1.10	3.10
220	ICP-MS	*0.25	*1.64	*0.57	0.93	2.55
264	ICP-MS	0.32	1.99	0.67	1.06	2.60
293	DRC/CC-ICP-MS	0.3	2.07	0.65	1.02	2.89
399	ICP-MS/MS	0.293	2.02	0.668	1.08	2.97
605	ICP-MS	0.312	2.07	0.685	1.11	2.91
Summary Statistics						
		UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.305	2.03	0.674	1.06	2.86
<b>Arithmetic SD (s)</b>		0.010	0.03	0.015	0.06	0.21
<b>Arithmetic RSD (%)</b>		3.3	1.6	2.2	5.7	7.3
<b>Number of Sample Measurements (N)</b>		6	6	6	7	7

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Urine Pt



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Urine Sb (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
103	ICP-MS/MS	2.18	1.20	0.855	0.380	0.582
107	ICP-MS	2.861	1.494	0.899	0.390	0.555
110	ICP-MS	2.96	1.59	0.896	0.420	0.566
147	ICP-MS	2.87	1.57	0.933	0.400	0.558
220	ICP-MS	2.85	1.66	0.96	0.45	0.52
264	ICP-MS	2.39	1.34	0.80	0.34	0.45
293	DRC/CC-ICP-MS	2.8	1.62	0.89	0.41	0.62
324	ICP-MS	2.940	1.605	<1	<1	<1
399	ICP-MS/MS	2.90	1.55	0.929	0.403	0.567
597	ICP-MS/MS	2.69	1.51	0.863	0.420	0.555
605	ICP-MS	2.91	1.67	0.807	<0.800	<0.800
606	ICP-MS/MS	2.90	1.61	0.940	0.433	0.526

### Summary Statistics

	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Robust Mean (x*)</b>	2.85	1.57	0.89	0.41	0.55
<b>Robust SD (s*)</b>	0.09	0.08	0.06	0.02	0.03
<b>Robust RSD (%)</b>	3.2	5.1	6.7	6.1	5.4
<b>Number of Sample Measurements (N)</b>	12	12	11	10	10
<b>Standard Uncertainty (u)</b>	0.03	0.03	0.02	0.01	0.01






\*Denotes a statistical Outlier.





# Results for Event #1, 2023: Summary Figures

## Urine Sb



### Legend:

○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = robust mean of all laboratories.  
 Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Urine Se (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
103	ICP-MS/MS	74.7	205	160	292	79.5
110	DRC/CC-ICP-MS	75.7	217	167	324	88.7
147	ICP-MS	75.5	209	159	289	78.8
293	DRC/CC-ICP-MS	64.77	186.41	141.39	265.4	72.67
597	ICP-MS/MS	78.8	224	176	314	88.6

### Summary Statistics






	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Arithmetic Mean ( $\bar{x}$ )	74	208	161	297	82
Arithmetic SD (s)	5	14	13	23	7
Arithmetic RSD (%)	6.8	6.7	8.1	7.7	8.5
Number of Sample Measurements (N)	5	5	5	5	5

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Summary Figures

### Urine Se



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics






Urine Sn (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
107	ICP-MS	0.30	4.00	0.64	1.62	0.75
110	ICP-MS	0.30	4.37	0.61	2.57	0.85
147	ICP-MS	0.277	4.24	0.481	1.78	0.716
220	ICP-MS	<0.4	4.31	<0.4	2.01	0.62
264	ICP-MS	*0.15	3.49	0.21	0.90	0.46
324	ICP-MS	<1	4.272	<1	2.288	<1
399	ICP-MS/MS	0.318	5.08	0.590	2.13	0.867
597	ICP-MS/MS	0.267	4.48	0.448	1.75	0.738
605	ICP-MS	<0.900	4.63	<0.900	2.42	<0.900
Summary Statistics						
		UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Arithmetic Mean ( $\bar{x}$ )		0.292	4.3	0.50	1.9	0.71
Arithmetic SD (s)		0.020	0.4	0.16	0.5	0.14
Arithmetic RSD (%)		6.8	10	32	26	20
Number of Sample Measurements (N)		5	9	6	9	7

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Urine Sn



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Urine Sr (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
103	ICP-MS/MS	129	28.0	192	96.6	87.7
107	ICP-MS	121.6	26.9	181.7	95.0	81.0
200	ICP-MS	141	30.9	206	109	95.1
220	ICP-MS	134	30.9	206	106	92.4
264	ICP-MS	118.9	26.54	178.9	93.95	81.46
399	DRC/CC-ICP-MS	127	28.4	194	98.2	87.6
597	ICP-MS/MS	123	27.6	188	94.4	86.4
605	ICP-MS	136	30.9	203	106	92.1

### Summary Statistics






	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Arithmetic Mean ( $\bar{x}$ )	129	28.8	194	100	88
Arithmetic SD (s)	8	1.9	11	6	5
Arithmetic RSD (%)	6.2	6.6	5.7	6.0	5.7
Number of Sample Measurements (N)	8	8	8	8	8

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Urine Sr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Urine V (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
116	ICP-MS/MS	*0.610	0.627	2.27	4.14	0.909
147	DRC/CC-ICP-MS	0.429	0.648	1.98	3.56	0.903
293	DRC/CC-ICP-MS	0.44	0.67	2.05	3.75	*0.98
597	ICP-MS/MS	0.424	0.630	2.11	3.70	0.915
605	ICP-MS	<0.300	0.657	2.27	4.23	0.890
Summary Statistics						
		UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.431	0.646	2.14	3.9	0.904
<b>Arithmetic SD (s)</b>		0.008	0.018	0.13	0.3	0.011
<b>Arithmetic RSD (%)</b>		1.9	2.8	6.1	7.5	1.2
<b>Number of Sample Measurements (N)</b>		3	5	5	5	4






\*Denotes a statistical Outlier.





# Results for Event #1, 2023: Summary Figures

## Urine V



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Urine W (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
107	ICP-MS	0.138	1.350	4.656	0.670	2.026
110	ICP-MS	0.159	1.46	4.90	0.712	2.18
147	ICP-MS	<0.171	1.29	4.45	0.668	1.95
200	ICP-MS	0.13	1.24	3.93	0.66	2.2
220	ICP-MS	0.16	1.4	4.87	0.71	2.08
264	ICP-MS	0.10	1.31	4.47	0.70	1.95
324	ICP-MS	<1	1.385	4.578	<1	2.004
399	ICP-MS/MS	0.142	1.44	4.72	0.683	2.05
597	ICP-MS/MS	0.178	1.40	4.82	0.678	2.11
605	ICP-MS	<0.180	1.47	4.92	0.721	2.16
606	ICP-MS/MS	0.181	1.42	4.67	0.581	3.87

Summary Statistics					
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Robust Mean (x*)</b>	0.15	1.38	4.67	0.68	2.09
<b>Robust SD (s*)</b>	0.03	0.07	0.22	0.03	0.12
<b>Robust RSD (%)</b>	18	5.1	4.7	4.4	5.7
<b>Number of Sample Measurements (N)</b>	8	11	11	10	11
<b>Standard Uncertainty (u)</b>	NA	0.03	0.08	0.01	0.04






\*Denotes a statistical Outlier.

An arithmetic mean, SD, RSD and n are provided for sample UE23-01.



# Results for Event #1, 2023: Summary Figures

## Urine W



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = robust mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2023: Laboratory Data and Summary Statistics






Urine Zn (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
110	ICP-MS	116	180	103	288	157
147	ICP-MS	112	199	99.3	269	156
264	ICP-MS	129.22	184.88	102.10	289.60	163.00
293	DRC/CC-ICP-MS	112.42	175.16	98.69	267.32	147.06
324	ICP-MS	102.527	169.753	93.781	275.639	149.796
391	ICP-MS	93.45	157.485	92.679	272.654	155.035
597	ICP-MS/MS	104	165	94.4	274	151
Summary Statistics						
		UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		110	176	98	277	154
<b>Arithmetic SD (s)</b>		11	14	4	9	5
<b>Arithmetic RSD (%)</b>		10	8.0	4.1	3.2	3.2
<b>Number of Sample Measurements (N)</b>		7	7	7	7	7

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Summary Figures

### Urine Zn



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Urine Te (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
110	ICP-MS	1.16	0.141	0.323	2.06	0.659
147	ICP-MS	0.902	0.115	0.223	1.65	0.569

### Summary Statistics

	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
Arithmetic Mean ( $\bar{x}$ )	1.0	0.13	0.27	1.9	0.61
Arithmetic SD (s)	0.2	0.02	0.07	0.3	0.06
Arithmetic RSD (%)	17	14	26	16	9.8
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Urine Ti (µg/L)						
Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
442	ICP-MS/MS	3.3	1.2	8.37	3.91	1.51
597	ICP-MS/MS	3.91	<1.90	9.80	6.77	3.27

Summary Statistics						
	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05	
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	3.6	NA	9.1	NA	NA	
<b>Arithmetic SD (s)</b>	0.4	NA	1.0	NA	NA	
<b>Arithmetic RSD (%)</b>	11	NA	11	NA	NA	
<b>Number of Sample Measurements (N)</b>	2	NA	2	NA	NA	

\*Denotes a statistical Outlier.

Statistical data was not calculated for UE23-02, UE23-04 and UE23-05 based on a lack of consensus among participating labs.



## Results for Event #1, 2023: Additional Elements in Urine

### Urine Ag (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
147	ICP-MS	<0.151	<0.151	<0.151	<0.151	<0.151

### Urine Bi (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
147	ICP-MS	<0.0794	<0.0794	<0.0794	<0.0794	<0.0794
264	ICP-MS	<0.01	<0.01	<0.01	<0.01	<0.01
597	ICP-MS/MS	<0.0244	<0.0244	<0.0244	<0.0244	<0.0244

### Urine Fe (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
324	ICP-MS	7.192	9.094	4.095	11.419	7.396

### Urine I (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
147	ICP-MS	33.2	38.4	76.2	146	138

### Urine Li (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
147	ICP-MS	8.19	9.44	15.6	27.3	9.99

### Urine Mg (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
597	ICP-MS/MS	15200	17500	27700	52000	25300

### Urine Th (µg/L)

Lab Code	Method	UE23-01	UE23-02	UE23-03	UE23-04	UE23-05
147	ICP-MS	<0.102	<0.102	<0.102	<0.102	<0.102
597	ICP-MS/MS	0.0227	0.0155	0.0158	0.0325	0.0188





**Department  
of Health**

**Wadsworth  
Center**

**Event #1, 2023**

**Trace Elements in  
Serum**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #1, 2023: Trace Elements in Serum

### PT Materials

Test materials were prepared from human serum obtained from Zen-Bio, Inc. The company certifies that these materials were tested by FDA approved methods and found to be negative for HIV 1Z2 and HIV-1 RNA, and non-reactive to HBsAg, HCV3 and STS. Units of serum were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with aluminum (Al), cobalt (Co), chromium (Cr), copper (Cu), selenium (Se), zinc (Zn), arsenic (As), beryllium (Be), cadmium (Cd), mercury (Hg), manganese (Mn), molybdenum (Mo), nickel (Ni), lead (Pb), platinum (Pt), antimony (Sb), tin (Sn), strontium (Sr), titanium (Ti), thallium (Tl), uranium (U), vanadium (V) and tungsten (W). PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Six elements in serum are formally graded: Al, Co, Cr, Cu, Se, and Zn. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 25 were reported by at least one participant: As, Ba, Be, Bi, Cd, Cs, Fe, Hg, I, Li, Mg, Mn, Mo, Ni, Pb, Pt, Sb, Sn, Sr, Th, Ti, Tl, U, V, and W. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #1, 2023: Summary Statistics

	Serum AI (µg/L)				
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	77	8.4	42	16.3	27.9
<b>Upper Limit</b>	92	13.4	50	21.3	33.5
<b>Lower Limit</b>	62	3.4	34	11.3	22.3
<b>Arithmetic SD (s)</b>	5	1.7	4	1.4	2.2
<b>Arithmetic RSD (%)</b>	6.5	20	9.5	8.7	7.9
<b>Number of Sample Measurements (N)</b>	6	5	6	5	5

The acceptable range is based on quality specifications:  $\pm 5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 5 \mu\text{g/L}$  at concentrations less than or equal to  $25 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2023: Performance of Participating Laboratories






		Serum AI (µg/L)				
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
	<b>Target</b>	<b>77</b>	<b>8.4</b>	<b>42</b>	<b>16.3</b>	<b>27.9</b>
147	ETAAS-Z	76.1	8.34	43.7	17.2	30.5
264	ICP-MS	79.49	7.43	39.48	14.79	27.36
293	DRC/CC-ICP-MS	72.58	9.14	40.86	17.74	29.03
391	ETAAS-Z	83.20	*24.2 ↑	49.23	*28.4 ↑	*42.3 ↑
485	HR-ICP-MS	81.1	10.7	39.2	17.1	27.9
597	ICP-MS/MS	69.6	6.19	36.6	14.8	24.6

Based on the grading criteria for AI in Serum, 90% of results were satisfactory, with 1 of the 6 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Serum AI



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±5 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±5 µg/L at concentrations less than or equal to 25 µg/L.



## Results for Event #1, 2023: Summary Statistics

	Serum Co (µg/L)				
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	0.473	9.9	1.00	3.28	3.89
<b>Upper Limit</b>	1.973	11.4	2.50	4.78	5.39
<b>Lower Limit</b>	0.000	8.4	0.00	1.78	2.39
<b>Arithmetic SD (s)</b>	0.020	0.4	0.02	0.14	0.18
<b>Arithmetic RSD (%)</b>	4.2	3.8	1.8	4.3	4.6
<b>Number of Sample Measurements (N)</b>	7	7	6	7	7

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



## Results for Event #1, 2023: Performance of Participating Laboratories






		Serum Co ( $\mu\text{g/L}$ )				
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
	<b>Target</b>	<b>0.473</b>	<b>9.9</b>	<b>1.00</b>	<b>3.28</b>	<b>3.89</b>
103	ICP-MS/MS	0.456	9.67	0.969	3.23	3.83
110	ICP-MS/MS	0.47	10.1	1.01	3.41	3.96
147	DRC/CC-ICP-MS	0.441	9.13	*0.878	3.02	3.54
264	ICP-MS	0.48	10.19	1.01	3.37	4.00
293	DRC/CC-ICP-MS	0.5	10.05	1.01	3.36	3.99
485	HR-ICP-MS	0.476	10.2	1.02	3.37	4.11
597	ICP-MS/MS	0.491	9.90	1.01	3.20	3.83

Based on the grading criteria for Co in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Serum Co



**Legend:**  
○ HHEAR Labs    ● Other Labs  
Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
±1.5 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 10 µg/L.





## Results for Event #1, 2023: Summary Statistics

	Serum Cr (µg/L)				
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	4.9	1.33	0.68	2.27	9.2
<b>Upper Limit</b>	6.9	3.33	2.68	4.27	11.2
<b>Lower Limit</b>	2.9	0.00	0.00	0.27	7.2
<b>Arithmetic SD (s)</b>	0.3	0.11	0.09	0.11	0.6
<b>Arithmetic RSD (%)</b>	6.7	8.3	13	4.8	6.5
<b>Number of Sample Measurements (N)</b>	7	7	6	6	7

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



## Results for Event #1, 2023: Performance of Participating Laboratories






		Serum Cr (µg/L)				
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
	<b>Target</b>	<b>4.9</b>	<b>1.33</b>	<b>0.68</b>	<b>2.27</b>	<b>9.2</b>
103	ICP-MS/MS	4.34	1.29	0.584	2.13	8.40
110	ICP-MS/MS	4.96	1.21	*3.40 ↑	2.30	9.23
147	DRC/CC-ICP-MS	4.69	1.20	0.577	2.20	8.68
264	ICP-MS	5.42	1.44	0.79	*2.80	9.94
293	DRC/CC-ICP-MS	5.08	1.29	0.72	2.45	9.64
485	HR-ICP-MS	5.00	1.43	0.666	2.28	9.44
597	ICP-MS/MS	4.91	1.43	0.730	2.28	8.81

Based on the grading criteria for Cr in Serum, 97% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Serum Cr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #1, 2023: Summary Statistics

	Serum Cu (µg/L)				
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	1210	790	1530	1260	910
<b>Upper Limit</b>	1390	910	1760	1450	1050
<b>Lower Limit</b>	1030	670	1300	1070	770
<b>Arithmetic SD (s)</b>	50	40	70	70	70
<b>Arithmetic RSD (%)</b>	4.1	5.1	4.6	5.6	7.7
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #1, 2023: Performance of Participating Laboratories






		Serum Cu (µg/L)					
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
	Target	1210	790	1530	1260	910	
107	DRC/CC-ICP-MS	1254	833	1636	1375	1056 ↑	
110	ICP-MS/MS	1230	808	1560	1270	895	
147	DRC/CC-ICP-MS	1144	743	1442	1201	858	
264	ICP-MS	1249	825	1580	1299	934	
293	DRC/CC-ICP-MS	1189	801	1519	1259	896	
436	FAAS	1143	730	1429	1143	826	
597	ICP-MS/MS	1240	799	1560	1260	901	

Based on the grading criteria for Cu in Serum, 97% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #1, 2023: Summary Figures

### Serum Cu



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ .



## Results for Event #1, 2023: Summary Statistics

	Serum Se ( $\mu\text{g/L}$ )				
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	130	281	98	208	160
<b>Upper Limit</b>	156	337	118	250	192
<b>Lower Limit</b>	104	225	78	166	128
<b>Arithmetic SD (s)</b>	6	11	3	8	7
<b>Arithmetic RSD (%)</b>	4.6	3.9	3.4	3.8	4.4
<b>Number of Sample Measurements (N)</b>	8	8	8	8	8

The acceptable range is based on quality specifications:  
 $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #1, 2023: Performance of Participating Laboratories

		Serum Se ( $\mu\text{g/L}$ )				
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
Target		130	281	98	208	160
103	ICP-MS/MS	137	293	100	210	159
107	DRC/CC-ICP-MS	127.3	275.3	95.4	204.2	154.9
110	ICP-MS/MS	131	283	97.9	206	157
147	DRC/CC-ICP-MS	123	260	94.0	194	151
264	ICP-MS	135.92	283.51	101.04	208.42	160.37
293	DRC/CC-ICP-MS	135	293	103	216	163
436	A-7	122	284	94	220	174
597	ICP-MS/MS	132	278	97.6	203	158






Based on the grading criteria for Se in Serum, 100% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.





# Results for Event #1, 2023: Summary Figures

## Serum Se



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #1, 2023: Summary Statistics

	Serum Zn (µg/L)				
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	581	742	849	702	749
<b>Upper Limit</b>	668	853	976	807	861
<b>Lower Limit</b>	494	631	722	597	637
<b>Arithmetic SD (s)</b>	20	24	35	21	31
<b>Arithmetic RSD (%)</b>	3.4	3.2	4.1	3.0	4.1
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

The acceptable range is based on quality specifications:  $\pm 15 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 15 \mu\text{g/L}$  at concentrations less than or equal to  $100 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #1, 2023: Performance of Participating Laboratories






		Serum Zn (µg/L)				
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
	<b>Target</b>	<b>581</b>	<b>742</b>	<b>849</b>	<b>702</b>	<b>749</b>
107	DRC/CC-ICP-MS	608	749	864	673	801
110	ICP-MS/MS	593	770	873	729	759
147	DRC/CC-ICP-MS	552	699	784	680	706
264	ICP-MS	576	739	846	708	742
293	DRC/CC-ICP-MS	569	745	843	706	739
597	ICP-MS/MS	590	750	881	715	747

Based on the grading criteria for Zn in Serum, 100% of results were satisfactory, with 0 of the 6 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2023: Summary Figures

## Serum Zn



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±15 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±15 µg/L at concentrations less than or equal to 100 µg/L.



### Results for Event #1, 2023: Laboratory Data and Summary Statistics






Serum Mn (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	8.63	11.0	2.84	1.43	5.68
110	ICP-MS/MS	8.8	11.2	3.0	1.5	5.7
147	DRC/CC-ICP-MS	8.30	9.84	2.72	1.37	5.31
264	ICP-MS	9.10	11.40	*3.48	*2.07	6.25
293	DRC/CC-ICP-MS	8.640	11.19	2.92	1.37	5.620
597	ICP-MS/MS	8.67	10.7	2.88	1.35	5.46
Summary Statistics						
		SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		8.7	10.9	2.87	1.40	5.7
<b>Arithmetic SD (s)</b>		0.3	0.6	0.10	0.06	0.3
<b>Arithmetic RSD (%)</b>		3.0	5.5	3.5	4.3	5.6
<b>Number of Sample Measurements (N)</b>		6	6	5	5	6

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Serum Mn



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics






Serum Mo (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	0.410	3.73	2.35	5.48	1.34
110	ICP-MS/MS	0.50	3.85	2.35	5.52	1.45
147	DRC/CC-ICP-MS	0.512	3.61	2.29	5.34	1.45
293	DRC/CC-ICP-MS	*0.920	3.800	2.360	5.350	1.620
485	HR-ICP-MS	<1	3.86	2.47	*6.14	1.39
597	ICP-MS/MS	0.499	3.67	2.28	5.18	1.32
Summary Statistics						
		SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.48	3.75	2.35	5.37	1.43
<b>Arithmetic SD (s)</b>		0.05	0.10	0.07	0.13	0.11
<b>Arithmetic RSD (%)</b>		10	2.7	3.0	2.4	7.7
<b>Number of Sample Measurements (N)</b>		4	6	6	5	6

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Serum Mo



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.





## Results for Event #1, 2023: Laboratory Data and Summary Statistics






Serum Ni (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	3.81	1.50	7.60	5.31	3.03
110	ICP-MS/MS	4.1	1.8	8.0	5.5	3.2
147	DRC/CC-ICP-MS	3.69	1.54	7.22	5.10	2.94
293	DRC/CC-ICP-MS	3.99	1.77	7.98	5.67	3.42
442	DRC/CC-ICP-MS	3.85	1.62	8.03	5.72	3.16
485	HR-ICP-MS	3.96	1.58	7.84	5.54	3.20
597	ICP-MS/MS	*3.07	*0.799	6.83	4.51	2.42
Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	3.90	1.64	7.6	5.3	3.1	
<b>Arithmetic SD (s)</b>	0.15	0.12	0.5	0.4	0.3	
<b>Arithmetic RSD (%)</b>	3.8	7.3	6.6	7.5	10	
<b>Number of Sample Measurements (N)</b>	6	6	7	7	7	

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Serum Ni



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics






Serum V (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
110	ICP-MS/MS	0.81	2.05	0.54	1.43	3.57
147	DRC/CC-ICP-MS	0.684	1.86	0.455	1.31	3.36
293	DRC/CC-ICP-MS	0.75	1.98	0.51	1.51	3.7
485	HR-ICP-MS	0.728	1.99	0.497	1.40	3.60
597	ICP-MS/MS	0.754	1.98	0.502	1.48	3.55
Summary Statistics						
		SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.75	1.97	0.50	1.43	3.56
<b>Arithmetic SD (s)</b>		0.05	0.07	0.03	0.08	0.12
<b>Arithmetic RSD (%)</b>		6.7	3.6	6.2	5.6	3.4
<b>Number of Sample Measurements (N)</b>		5	5	5	5	5

\*Denotes a statistical Outlier.



# Results for Event #1, 2023: Summary Figures

## Serum V



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum As ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	1.01	8.28	3.37	14.3	4.87
110	ICP-MS/MS	1.02	8.16	3.14	14.2	4.85
147	DRC/CC-ICP-MS	1.03	7.29	3.05	12.6	4.40
597	ICP-MS/MS	1.07	8.03	3.29	13.8	4.78
Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	1.03	7.9	3.21	13.7	4.7	
Arithmetic SD (s)	0.03	0.4	0.14	0.8	0.2	
Arithmetic RSD (%)	2.5	5.1	4.4	5.8	4.7	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Ba (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
110	ICP-MS/MS	3.36	0.70	2.67	1.70	2.15
147	ICP-MS	3.10	0.556	2.45	1.58	1.81
597	ICP-MS/MS	3.27	0.649	2.88	1.59	2.00
Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	3.24	0.64	2.7	1.62	2.0	
Arithmetic SD (s)	0.13	0.07	0.2	0.07	0.2	
Arithmetic RSD (%)	4.0	11	8.2	4.3	8.5	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Be (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
110	ICP-MS/MS	2.08	4.36	0.96	3.36	0.45
147	ICP-MS	1.97	3.77	0.937	2.75	0.488
293	ICP-MS	2.09	4.05	0.94	3.17	0.48
597	ICP-MS/MS	2.21	4.42	1.01	3.29	0.504
Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	2.09	4.2	0.96	3.1	0.48	
Arithmetic SD (s)	0.10	0.3	0.03	0.3	0.02	
Arithmetic RSD (%)	4.8	7.1	3.1	8.6	4.8	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Cd (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	7.76	4.20	0.935	2.61	0.227
110	ICP-MS/MS	7.54	4.25	0.93	2.60	0.23
147	ICP-MS	7.14	4.06	0.905	2.47	0.223
597	ICP-MS/MS	7.25	3.95	0.891	2.51	0.229
Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	7.4	4.12	0.92	2.55	0.227	
Arithmetic SD (s)	0.3	0.14	0.02	0.07	0.003	
Arithmetic RSD (%)	3.8	3.4	2.3	2.7	1.3	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.





## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Cs (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
110	ICP-MS/MS	0.64	0.74	1.02	1.93	0.66
597	ICP-MS/MS	0.665	0.746	1.04	2.00	0.661

Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	0.65	0.743	1.03	1.97	0.661	
Arithmetic SD (s)	0.02	0.004	0.01	0.05	0.001	
Arithmetic RSD (%)	2.8	0.54	1.4	2.5	0.11	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Hg (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	2.43	4.42	0.960	3.34	0.659
110	ICP-MS/MS	2.49	4.24	0.95	3.03	0.61
597	ICP-MS/MS	2.23	4.09	0.983	2.94	0.679
Summary Statistics						
		SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
Arithmetic Mean ( $\bar{x}$ )		2.38	4.3	0.96	3.1	0.65
Arithmetic SD (s)		0.14	0.2	0.02	0.2	0.04
Arithmetic RSD (%)		5.9	4.0	1.8	6.8	6.2
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Serum I (µg/L)

Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
147	ICP-MS	57.8	55.2	56.6	38.1	38.5
442	ICP-MS	59.2	58.2	61.1	40.0	40.7

### Summary Statistics

	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
Arithmetic Mean ( $\bar{x}$ )	58.5	57	59	39.0	40
Arithmetic SD (s)	1.0	2	3	1.3	2
Arithmetic RSD (%)	1.7	3.7	5.1	3.3	4.0
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Mg (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
264	ICP-MS	19653.0	20616.0	18844.0	18920.0	20850.0
597	ICP-MS/MS	19800	20200	18800	18400	20300

Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	19730	20400	18820	18700	20600	
<b>Arithmetic SD (s)</b>	100	300	30	400	400	
<b>Arithmetic RSD (%)</b>	0.51	1.5	0.16	2.1	1.9	
<b>Number of Sample Measurements (N)</b>	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Pb (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	0.512	1.88	2.69	13.5	5.26
110	ICP-MS/MS	0.60	1.98	2.81	13.3	5.23
597	ICP-MS/MS	0.612	1.92	2.69	12.7	4.90
Summary Statistics						
		SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
Arithmetic Mean ( $\bar{x}$ )		0.57	1.93	2.73	13.2	5.1
Arithmetic SD (s)		0.05	0.05	0.07	0.4	0.2
Arithmetic RSD (%)		8.8	2.6	2.6	3.0	3.9
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Pt ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
110	ICP-MS/MS	1.19	0.132	0.377		0.740
264	ICP-MS	1.21	0.09	0.34	0.40	0.77
293	DRC/CC-ICP-MS	1.17	0.13	0.38	0.42	0.80

Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	1.19	0.12	0.37	0.41	0.77	
Arithmetic SD (s)	0.02	0.02	0.02	0.01	0.03	
Arithmetic RSD (%)	1.7	21	6.0	3.4	3.9	
Number of Sample Measurements (N)	3	3	3	2	3	

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Sb (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	3.98	0.819	3.29	2.26	6.34
110	ICP-MS/MS	4.80	0.97	4.19	2.46	7.22
147	ICP-MS	4.33	0.885	3.69	2.22	6.60
597	ICP-MS/MS	4.63	0.941	4.00	2.51	6.76

Summary Statistics					
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
Arithmetic Mean ( $\bar{x}$ )	4.4	0.90	3.8	2.36	6.7
Arithmetic SD (s)	0.4	0.07	0.4	0.14	0.4
Arithmetic RSD (%)	9.1	7.8	11	5.9	6.0
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Sn ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
110	ICP-MS/MS	6.61	2.69	3.53	5.02	0.62
597	ICP-MS/MS	6.16	2.42	3.29	4.35	0.529

Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	6.4	2.6	3.4	4.7	0.57	
Arithmetic SD (s)	0.3	0.2	0.2	0.5	0.06	
Arithmetic RSD (%)	4.7	7.4	5.0	11	11	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.





## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Serum Sr (µg/L)

Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	98.2	92.2	49.8	92.5	69.3
200	ICP-MS	*129.0	*122.0	*63.4	*124.0	*90.2
597	ICP-MS/MS	98.0	92.3	50.7	92.6	69.7

### Summary Statistics

	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
Arithmetic Mean ( $\bar{x}$ )	98.1	92.3	50.3	92.6	69.5
Arithmetic SD (s)	0.1	0.1	0.6	0.1	0.3
Arithmetic RSD (%)	0.14	0.080	1.3	0.080	0.41
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum Ti (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
200	DRC/CC-ICP-MS	5.2	4.6	2.6	10.1	6.0
442	ICP-MS/MS	5.25	3.61	1.34	10	7.95
485	HR-ICP-MS	4.63	3.36	0.870	8.94	7.15
597	ICP-MS/MS	*7.55	*6.55	4.10	*12.5	*10.5
Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	5.0	3.9	NA	9.68	7.0	
Arithmetic SD (s)	0.3	0.7	NA	0.64	1.0	
Arithmetic RSD (%)	6.9	17	NA	6.6	14	
Number of Sample Measurements (N)	3	3	NA	3	3	

\*Denotes a statistical Outlier.

Statistical data was not calculated for SE23-04 based on a lack of consensus among participating labs.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum TI (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	0.882	2.26	1.62	0.344	0.696
110	ICP-MS/MS	0.91	2.28	1.67	0.34	0.70
147	ICP-MS	0.836	2.13	1.57	0.341	0.677
597	ICP-MS/MS	0.804	2.01	1.51	0.319	0.624
Summary Statistics						
	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05	
Arithmetic Mean ( $\bar{x}$ )	0.86	2.17	1.59	0.336	0.67	
Arithmetic SD (s)	0.05	0.13	0.07	0.011	0.03	
Arithmetic RSD (%)	5.8	6.0	4.4	3.3	4.5	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

Serum U (µg/L)						
Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
103	ICP-MS/MS	0.139	0.109	0.155	0.0703	0.184
110	ICP-MS/MS	0.139	0.107	0.153	0.068	0.188
597	ICP-MS/MS	0.138	0.105	0.134	0.0669	0.177
Summary Statistics						
		SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
Arithmetic Mean ( $\bar{x}$ )		0.1387	0.107	0.147	0.068	0.183
Arithmetic SD (s)		0.0006	0.002	0.012	0.002	0.006
Arithmetic RSD (%)		0.43	1.9	8.2	2.5	3.3
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Laboratory Data and Summary Statistics

### Serum W (µg/L)

Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
110	ICP-MS/MS	4.72	1.03	2.89	1.29	0.28
597	ICP-MS/MS	4.42	0.900	2.72	1.18	0.229

### Summary Statistics

	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
Arithmetic Mean ( $\bar{x}$ )	4.6	0.96	2.81	1.24	0.25
Arithmetic SD (s)	0.2	0.09	0.12	0.08	0.04
Arithmetic RSD (%)	4.3	9.4	4.3	6.5	16
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #1, 2023: Additional Elements in Serum

### Serum Bi (µg/L)

Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
147	ICP-MS	<0.0397	<0.0397	<0.0397	<0.0397	<0.0397
597	ICP-MS/MS	<0.0300	<0.0300	<0.0300	<0.0300	<0.0300

### Serum Fe (µg/L)

Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
264	ICP-MS	1228.00	852.00	1187.00	974.00	1486.00

### Serum Li (µg/L)

Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
147	ICP-MS	0.749	0.495	0.557	0.715	0.902

### Serum Th (µg/L)

Lab Code	Method	SE23-01	SE23-02	SE23-03	SE23-04	SE23-05
597	ICP-MS/MS	<0.00488	<0.00488	<0.00488	<0.00488	<0.00488



## References

1. ISO/FDIS-13528 (2005) Statistical methods for use in proficiency testing by interlaboratory comparisons. International Organization for Standardization, Geneva.
2. Taylor A, Angerer J, Arnaud J, Claeys F, Jones RL, Mazarrasa O, Mairiaux E, Menditto A, Parsons PJ, Patriarca M, Pineau A, Valkonen S, Weber J-P, Weykamp C. Occupational and environmental laboratory medicine: A network of EQAS organisers. Accreditation and Quality Assurance. 2006;11(8-9):435-9. PubMed PMID: 086NJ-0011.

## Summary of Participant Survey Responses

Thank you for taking the time to complete our 2022 Participant Survey. We appreciate your feedback and suggestions to improve the quality of the program.

Several participants expressed an interest in expanding the current panel to include arsenic species in urine, specifically Monomethylarsonic acid (MMA), Dimethylarsinic acid (DMA), Arsenobetaine (AsB), Arsenocholine (AsC) and inorganic arsenic species (i.e.,  $\Sigma \text{As}^{3+}$  and  $\text{As}^{5+}$ ). It was interesting to see a recent paper published on the determination of arsenic species in the NYS PT urine samples (1). While the PT program does not currently supplement As species into the urine pools, some were measurable nonetheless. The arsenic species determined were endogenous. Several participants suggested including Hg species to the whole blood panel.

Based on these suggestions, we plan to evaluate the feasibility of including speciation measurements as a one-time educational event. Educational events are not intended to be graded as confidence in assigned values is limited by the number of participants that report results. If your laboratory is interested in participating in an educational event for urine As species and/or blood Hg species, then please reach out to us at [trel@health.ny.gov](mailto:trel@health.ny.gov).

1. Quarles CD, Sullivan P, Bohlim N, Saetveit N. Rapid automated total arsenic and arsenic speciation by inductively coupled plasma mass spectrometry. *Journal of Analytical Atomic Spectrometry*. 2022;37(6):1240-6. doi: 10.1039/D2JA00055E.