

**September 2010 *Chlamydia* & *Neisseria gonorrhoeae* Results**  
**Number of participating laboratories: 32**

**Grade distribution**

100%	32
90-99%	1
80-89%	1
60-69	1

Sample	Correct responses		
	Chlamydia Direct Detection	GC Direct Detection	GC Culture
A	Positive	Positive	Positive
B	Negative	Negative	Negative
C	Positive	Negative	Negative
D	Positive	Negative	Positive
E	Negative	Positive	Positive

<i>Chlamydia trachomatis</i> Direct Detection			
Sample	# Labs reporting		
	A, C, D	B, E	
Method	+	-	+
Gen-Probe Aptima Combo 2	13	13	
BD ProbeTec ET CT or CT/GC	7	7	
Digene Hybrid Capture hc2 CT/GC	5	4	1
Wampole Clearview Chlamydia	2	2	
Roche Diagnostics COBAS AMPLICOR CT/NG	2	2	
User-Developed Real-Time PCR	1	1	
Quidel QuickVue Chlamydia	1	1	
Gen-Probe PACE 2 CT or CT/GC	1	1	

<i>Neisseria gonorrhoeae</i> Direct Detection							
Sample	# Labs using method						
	A, E	B		C		D	
Method	+	-	+	-	+	-	+
Gen-Probe Aptima Combo 2	13	11	2	12	1	13	
BD ProbeTec ET CT/GC	7	7		7		7	
Digene Hybrid Capture hc2 CT/GC	5	0	5	4	1	4	1
Roche Diagnostics COBAS AMPLICOR CT/NG	2	2		2		2	
User-Developed Real-Time PCR	1	1		1		1	
Gen-Probe PACE 2 CT or CT/GC	1	1		1		1	

<i>Neisseria gonorrhoeae</i> Culture	
Method	# Labs using method
bioMerieux API NH	1
Boule Diagnostics Phadebact Monoclonal GC	1
Jembec Thayer Martin Chocolate Agar	2