

September 2010 Throat Culture Results
Number of participating laboratories: 25

Grade distribution

100% 22
80-89% 3

Correct Responses:

- A – Positive
- B – Positive
- C – Negative
- D – Negative
- E – Positive

Method	# Labs reporting						
	Sample						
	A		B		C	D	E
	+	-	+	-	-	-	+
Bacitracin disk only	8		7	1	8	8	8
Selective Strep Agar	4	1	5		5	5	5
Bacitracin disk and Murex Streptex	2		2		2	2	2
Remel PathoDx	2		2		2	2	2
Acceava Rapid Strep A Kit	2		2		2	2	2
Bacitracin disk and Remel PathoDx	1	1	2		2	2	2
BD BBL Streptocard	1		1		1	1	1
Mckesson Strep A Cassette	1		1		1	1	1
Remel Streptex	1		1		1	1	1
Not given	1		1		1	1	1