

**New York State Department of Health
Wadsworth Center**

Proficiency Testing Program

26-Apr-06

**DIAGNOSTIC IMMUNOLOGY &
HUMAN IMMUNODEFICIENCY VIRUS
SUMMARY ANALYSIS**

**Proficiency Test Event
26-Apr-06**

**Diagnostic Immunology & Human Immunodeficiency Virus
Summary Report**

Steven Bush M.S., Wendy Archinal Ph.D., and Susan Wong Ph.D.

Table of Contents

Review of Grading Policy	Page 3
Determination of Acceptable Responses	Page 4
Qualitative / Quantitative Test Analytes	Page 5 - 29
Antinuclear Ab	Page 5 - 6
Antistreptolysin O	Page 7 - 9
Cytomegalovirus Ab	Page 10 - 11
Hepatitis B Core Ab	Page 12
Hepatitis B Surface Ag	Page 13
Hepatitis Be Ag	Page 14
Hepatitis C Ab	Page 14
HIV Ab Screening/ Confirmation	Page 15
HIV p24 Ag	Page 15
HTLV1 Ab	Page 16
Infectious Mononucleosis	Page 17 - 18
Lyme Disease (<u>Borrelia burgdorferi</u>) Ab	Page 19 - 20
Rheumatoid Factor	Page 21 - 23
Rubella IgG Ab	Page 24 - 25
Rubella IgM	Page 26
Syphilis-Reagin Ab	Page 27 - 28
Syphilis-Treponemal Ab	Page 29
Quantitative Test Analytes	Page 30 - 31
Alpha-1 Antitrypsin	Page 30
Complement C'3	Page 30
Complement C'4	Page 30
Immunoglobulin A	Page 31
Immunoglobulin E	Page 31
Immunoglobulin G	Page 31
Immunoglobulin M	Page 31
Acceptable Responses	Page 32 - 33

The data summarized in this report were tabulated from test results and accompanying information submitted by laboratories that participated in the April 26, 2006 Diagnostic Immunology and Human Immunodeficiency Virus proficiency test events. Participants are encouraged to review the data and to compare results and test kit performances.

Laboratories were evaluated on the basis of their responses for each analyte and on overall performance for all the analytes tested in the permit category. Appropriate responses were determined by participant consensus requiring 80% agreement in each test.

Qualitative/quantitative results were graded in relation to results given by participants for specific test kits. When the number of participants that used a specific test kit was less than 10, results were graded considering results given for the method used. Target values and acceptable ranges were determined as indicated in Page 5.

Grading Criteria:

- ⊖ When both qualitative and quantitative results were reported ten points were deducted for each incorrect result. When only qualitative OR quantitative results were reported twenty points were deducted for each incorrect result.
- ⊖ For **Diagnostic Services** failure to attain an overall testing score of at least 80% is unsatisfactory performance
- ⊖ For **Donor Services** failure to attain an overall testing score of 100% is unsatisfactory
- ⊖ For **HIV** failure to attain an overall testing score of 100% is unsatisfactory performance.
- ⊖ Laboratories failing two out of three consecutive proficiency test events for an analyte or for the permit category will fail the proficiency testing program for the analyte or for the permit category and may be required to cease patient testing for that analyte/category.

Summary Tables

Test kit manufacturer names are in *italics*. In some tables, test kits are grouped under test methods shown in bold letters. In all tables, test methods and test kit manufacturer names are listed in alphabetic order. Only the testing systems used by 10 or more laboratories are listed in this report.

For qualitative tests, results are summarized as the number of laboratories that reported a test sample as reactive to the number that reported it as non-reactive. In addition, where test results depend on a quantitative value (e.g. titer, IU/ml) the values reported are given in separate tables. They are expressed, where applicable, as the Mean \pm S.D. when ten or more laboratories reported data.

For quantitative tests, values reported variously as mg/dl, IU/ml, ratio, etc. are given as the Mean \pm S.D. when ten or more laboratories reported results. Titers are given as endpoint titers.

Disclaimer

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

Determination of Acceptable Responses

Analyte or Test	Criteria
Alpha-1 Antitrypsin	Target value \pm 3 S.D.
Antinuclear Antibody	Target value \pm 2 dilutions or positive or negative
Antistreptolysin O	Target value \pm 2 dilutions or positive or negative
Complement C'3, C'4	Target value \pm 3 S.D.
Cytomegalovirus Antibody	Reactive / nonreactive
Hepatitis (HbsAg, anti-HBc, HBeAg, and HCAb)	Reactive / nonreactive
HIV 1 Ab, Ag	Reactive / nonreactive
HTLV 1 Ab (EIA Ab, WB)	Reactive / nonreactive
Lyme Disease Ab, WB IgG, IgM	Reactive / nonreactive
Immunoglobulin A, E, M	Target value \pm 3 S.D.
Immunoglobulin G	Target value \pm 25 %
Infectious Mononucleosis	Target value \pm 2 dilutions or positive or negative
Rheumatoid Factor	Target value \pm 2 dilutions or positive or negative
Rubella Ab, IgM	Target value \pm 2 dilutions or positive or negative or Immune or nonimmune
Syphilis Reagin Antibody	Target value \pm 1 dilution
Syphilis Treponemal Antibody	Reactive / nonreactive

Antinuclear Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	46				47				48				49				50			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	56	4	7%	52	93%	56	100%	0	0%	56	100%	0	0%	56	100%	0	0%	0	0%	56	100%
<i>Bio-Rad</i>	15		0%	15	100%	15	100%		0%	15	100%		0%	15	100%		0%		0%	15	100%
<i>Diamedix</i>	11	1	9%	10	91%	11	100%		0%	11	100%		0%	11	100%		0%		0%	11	100%
<i>Others</i>	30	3	10%	27	90%	30	100%		0%	30	100%		0%	30	100%		0%		0%	30	100%
IFA	76	1	1%	75	99%	75	99%	1	1%	76	100%	0	0%	76	100%	0	0%	0	0%	76	100%
<i>Bio-Rad</i>	21		0%	21	100%	21	100%		0%	21	100%		0%	21	100%		0%		0%	21	100%
<i>The Binding Site</i>	13		0%	13	100%	12	92%	1	8%	13	100%		0%	13	100%		0%		0%	13	100%
<i>Wampole/Zeus</i>	23	1	4%	22	96%	23	100%		0%	23	100%		0%	23	100%		0%		0%	23	100%
<i>Others</i>	19		0%	19	100%	19	100%		0%	19	100%		0%	19	100%		0%		0%	19	100%
Other Methods	10		0%	10	100%	10	100%		0%	10	100%		0%	10	100%		0%	1	10%	9	90%
Analyte Total	142	5	4%	137	96%	141	99%	1	1%	142	100%	0	0%	142	100%	0	0%	1	1%	141	99%

Note: Of the 115 laboratories reporting staining patterns: 93% found test sample 46 to stain Homogenous, and 95% found test sample 50 to stain Speckled.

Antinuclear Antibody

The number of laboratories that reported titers is listed for positive test samples 46 and 50. Only testing systems with 10 or more laboratories reporting titers are listed in this table.

Method	No. Labs	Sample 46						Sample 50							
		80	160	320	640	1280	2560	Titers							
Manufacturer		80	160	320	640	1280	2560	80	160	320	640	1280	2560	5120	10240
IFA															
<i>Bio-Rad</i>	20	4	4	9	2			1	1	6	7	3			
<i>Wampole/ Zeus</i>	21	3	10	3	1	2	1		4	2	5	5	2	1	1
<i>The Binding site</i>	13	2	3	4	4					1	4	6	1		

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers.

Antistreptolysin O

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	16				17				18				19				20			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Hemagglutination	10	10	100%	0	0%	0	0%	10	100%	10	100%	0	0%	0	0%	10	100%	10	100%	0	0%
<i>Wampole /Zues</i>	10	10	100%		0%		0%	10	100%	10	100%		0%		0%	10	100%	10	100%		0%
Latex Agglutination	86	86	100%	0	0%	1	1%	85	99%	86	100%	0	0%	0	0%	86	100%	86	100%	0	0%
<i>Behring</i>	21	21	100%		0%		0%	21	100%	21	100%		0%		0%	21	100%	21	100%		0%
<i>Fisher</i>	26	26	100%		0%		0%	26	100%	26	100%		0%		0%	26	100%	26	100%		0%
<i>Remel</i>	14	14	100%		0%		0%	14	100%	14	100%		0%		0%	14	100%	14	100%		0%
<i>Others</i>	25	25	100%		0%	1	4%	24	96%	25	100%		0%		0%	25	100%	25	100%		0%
Nephelometry	28	28	100%	0	0%	0	0%	28	100%	28	100%	0	0%	0	0%	28	100%	28	100%	0	0%
<i>Beckman</i>	16	16	100%		0%		0%	16	100%	16	100%		0%		0%	16	100%	16	100%		0%
<i>Behring</i>	12	12	100%		0%		0%	12	100%	12	100%		0%		0%	12	100%	12	100%		0%
Turbidimetry	17	17	100%	0	0%	0	0%	17	100%	16	94%	1	6%	0	0%	17	100%	17	100%	0	0%
<i>Roche Diagnostics</i>	12	12	100%		0%		0%	12	100%	11	92%	1	8%		0%	12	100%	12	100%		0%
<i>Others</i>	5	5	100%		0%		0%	5	100%	5	100%		0%		0%	5	100%	5	100%		0%
Other Methods	2	2	100%	0	0%	0	0%	2	100%	2	100%	0	0%	0	0%	2	100%	2	100%	0	0%
Analyte Total	143	143	100%	0	0%	1	1%	142	99%	142	99%	1	1%	0	0%	143	100%	143	100%	0	0%

Antistreptolysin O Latex Agglutination Procedures

The number of laboratories that reported titers is listed for positive test samples 17 and 19. Only testing systems with 10 or more laboratories reporting titers are listed in this table.

Method <i>Manufacturer</i>	No. Labs	Sample 17 Titers					Sample 19 Titers				
		100	200	400	800	1600	100	200	400	800	1600
<i>Biokit/ Fisher</i>	16		2	12	2			12	4		
<i>Dade Behring</i> ^[1]	18			12	2			6	8		
<i>Remel</i>	12		1	11			1	9	2		

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers.

Antistreptolysin O

Results are summarized for positive test samples 17 and 19. The Mean values \pm S.D. are given where 10 or more laboratories reported quantitative results. Outlier values are omitted.

Method	No.	IU/ml	
<i>Manufacturer</i>	Labs	Sample 17	Sample 19
Nephelometry			
<i>Dade Behring Neph.</i>	13	436 \pm 34	553 \pm 35
<i>Beckman Coulter Immage</i>	10	688 \pm 69	751 \pm 44
Turbidimetry			
<i>Roche Diagnostics Cobas</i>	10	698 \pm 81	787 \pm 36

Cytomegalovirus Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	11*				12				13				14				15			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	38	7	18%	27	71%	38	100%	1	3%	1	3%	37	97%	38	100%	0	0%	38	100%	0	0%
<i>Wampole /Zeus</i>	17	2	12%	13	76%	17	100%		0%		0%	17	100%	17	100%		0%	17	100%		0%
<i>Others</i>	21	5	24%	14	67%	21	100%	1	5%	1	5%	20	95%	21	100%		0%	21	100%		0%
ELFA	21	0	0%	21	100%	21	100%	0	0%	0	0%	21	100%	21	100%	0	0%	21	100%	0	0%
<i>bioMérieux Vidas</i>	21		0%	21	100%	21	100%		0%		0%	21	100%	21	100%		0%	21	100%		0%
Chemiluminescence	15	4	27%	9	60%	15	100%	0	0%	1	7%	14	93%	15	100%	0	0%	15	100%	0	0%
<i>Diagnostic Products</i>	15	4	27%	9	60%	15	100%		0%	1	7%	14	93%	15	100%		0%	15	100%		0%
Hemagglutination	12	0	0%	12	100%	12	100%	0	0%	0	0%	12	100%	12	100%	0	0%	12	100%	0	0%
<i>Olympus</i>	12		0%	12	100%	12	100%		0%		0%	12	100%	12	100%		0%	12	100%		0%
Other Methods	12		0%	12	100%	12	100%		0%	1	8%	11	92%	11	92%	1	8%	12	100%		0%
Analyte Total	98	11	11%	81	83%	98	100%	1	1%	3	3%	95	97%	97	99%	1	1%	98	100%	0	0%

* Test sample #11 was not authenticated - A consensus agreement cannot be reached among participants, by regulation requirement, the sample cannot be graded (scored) and all participating laboratories get credit for this sample. 7 laboratories across multiple systems reported an equivocal or indeterminate result for sample #11.

Cytomegalovirus Antibody

Results are summarized for positive test samples 11 and 13. For the procedure indicated, the Mean values \pm S.D. are given where 10 or more laboratories reported quantitative results.

<i>Manufacturer</i>	No. Labs	Units	Sample Number	
			11	13
EIA				
<i>Wampole/ Zeus Scientific</i>	12	Ratio	1.12 \pm 0.19	2.58 \pm 0.20
ELFA				
<i>bioMerieux Vidas</i>	14	AU/ml	6.38 \pm 0.65	57 \pm 6.6

Hepatitis B Core Antibody

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method	No. Labs	6				7				8				9				10			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Chemiluminescence	68	0	0%	68	100%	68	100%	0	0%	0	0%	68	100%	1	1%	67	99%	0	0%	68	100%
<i>Bayer</i>	18		0%	18	100%	18	100%		0%		0%	18	100%		0%	18	100%		0%	18	100%
<i>Diagnostic Products</i>	24		0%	24	100%	24	100%		0%		0%	24	100%	1	4%	23	96%		0%	24	100%
<i>Ortho</i>	26		0%	26	100%	26	100%		0%		0%	26	100%		0%	26	100%		0%	26	100%
EIA	85	1	1%	84	99%	85	100%	0	0%	0	0%	85	100%	0	0%	85	100%	0	0%	85	100%
<i>Abbott</i>	63	1	2%	62	98%	63	100%		0%		0%	63	100%		0%	63	100%		0%	63	100%
<i>Ortho</i>	12		0%	12	100%	12	100%		0%		0%	12	100%		0%	12	100%		0%	12	100%
<i>Other</i>	10		0%	10	100%	10	100%		0%		0%	10	100%		0%	10	100%		0%	10	100%
Other Methods	2		0%	2	100%	2	100%		0%		0%	2	100%		0%	2	100%		0%	2	100%
Analyte Total	155	1	1%	154	99%	155	100%	0	0%	0	0%	155	100%	1	1%	154	99%	0	0%	155	100%

Hepatitis B Surface Antigen

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	6				7				8				9				10			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Chemiluminescence	86	86	100%	0	0%	80	93%	4	5%	0	0%	86	100%	86	100%	0	0%	1	1%	85	99%
<i>Bayer</i>	23	23	100%		0%	23	100%		0%		0%	23	100%	23	100%		0%		0%	23	100%
<i>Diagnostic Products</i>	28	28	100%		0%	28	100%		0%		0%	28	100%	28	100%		0%		0%	28	100%
<i>Ortho</i> ^[1]	35	35	100%		0%	29	83%	4	11%		0%	35	100%	35	100%		0%	1	3%	34	97%
EIA	116	116	100%	0	0%	116	100%	0	0%	0	0%	116	100%	116	100%	0	0%	1	1%	115	99%
<i>Abbott</i>	83	83	100%		0%	83	100%		0%		0%	83	100%	83	100%		0%	1	1%	82	99%
<i>Bio-Rad</i>	10	10	100%		0%	10	100%		0%		0%	10	100%	10	100%		0%		0%	10	100%
<i>Roche</i>	10	10	100%		0%	10	100%		0%		0%	10	100%	10	100%		0%		0%	10	100%
<i>Other</i>	13	13	100%		0%	13	100%		0%		0%	13	100%	13	100%		0%		0%	13	100%
Other Methods	3	3	100%		0%	3	100%		0%		0%	3	100%	3	100%		0%		0%	3	100%
Analyte Total	205	205	100%	0	0%	199	97%	4	2%	0	0%	205	100%	205	100%	0	0%	2	1%	203	99%

^[1] An equivocal or borderline result on sample #7 was reported by 2 labs.

Hepatitis Be Antigen

Participant Results/ Sample Number																							
R = Reactive/ Positive; N = Non-Reactive/ Negative																							
Method	No. Labs	66				67				68				69				70					
		N		%		R		%		N		%		R		%		N		%		R	
EIA	31	31	100%	0	0%	0	0%	31	100%	31	100%	0	0%	1	3%	30	97%	31	100%	0	0%		
<i>DiaSorin</i>	30	30	100%		0%		0%	30	100%	30	100%		0%		0%	30	100%	30	100%		0%		
<i>Other</i>	1	1	100%		0%		0%	1	100%	1	100%		0%	1	100%		0%	1	100%		0%		
Other Methods	4	4	100%		0%		0%	4	100%	4	100%		0%		0%	4	100%	4	100%		0%		
Analyte Total	35	35	100%	0	0%	0	0%	35	100%	35	100%	0	0%	1	3%	34	97%	35	100%	0	0%		

Hepatitis C Antibody

Participant Results/ Sample Number																							
R = Reactive/ Positive; N = Non-Reactive/ Negative																							
Method	No. Labs	71				72				73				74				75					
		N		%		R		%		N		%		R		%		N		%		R	
Chemiluminescence	62	62	100%	0	0%	62	100%	0	0%	0	0%	61	98%	62	100%	0	0%	0	0%	62	100%		
<i>Bayer</i>	27	27	100%		0%	27	100%		0%		0%	27	100%	27	100%		0%		0%	27	100%		
<i>Ortho</i> ^[1]	35	35	100%		0%	35	100%		0%		0%	34	97%	35	100%		0%		0%	35	100%		
EIA	102	100	98%	2	2%	100	98%	2	2%	0	0%	102	100%	101	99%	1	1%	0	0%	102	100%		
<i>Abbott</i>	75	74	99%	1	1%	73	97%	2	3%		0%	75	100%	74	99%	1	1%		0%	75	100%		
<i>Ortho</i>	27	26	96%	1	4%	27	100%		0%		0%	27	100%	27	100%		0%		0%	27	100%		
Other Methods	1	1	100%		0%	1	100%		0%		0%	1	100%	1	100%		0%		0%	1	100%		
Analyte Total	165	163	99%	2	1%	163	99%	2	1%	0	0%	164	99%	164	99%	1	1%	0	0%	165	100%		

^[1] An indeterminate result on sample #73 was reported by 1 labs.

HIV Antibody

Method <i>Manufacturer</i>		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
		31				32				33				34				35			
No. Labs	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	
EIA	129	129	100%	0	0%	0	0%	129	100%	129	100%	0	0%	0	0%	129	100%	129	100%	0	0%
<i>Abbott</i>	68	68	100%		0%		0%	68	100%	68	100%		0%		0%	68	100%	68	100%		0%
<i>bioMérieux</i>	17	17	100%		0%		0%	17	100%	17	100%		0%		0%	17	100%	17	100%		0%
<i>Bio-Rad</i>	44	44	100%		0%		0%	44	100%	44	100%		0%		0%	44	100%	44	100%		0%
Rapid EIA	95	95	100%	0	0%	0	0%	95	100%	95	100%	0	0%	0	0%	95	100%	95	100%	0	0%
<i>Orasure</i>	95	95	100%		0%		0%	95	100%	95	100%		0%		0%	95	100%	95	100%		0%
Rapid Immunoassay	39	39	100%	0	0%	0	0%	39	100%	39	100%	0	0%	0	0%	39	100%	39	100%	0	0%
<i>Medmira</i>	26	26	100%		0%		0%	26	100%	26	100%		0%		0%	26	100%	26	100%		0%
<i>Trinity</i>	10	10	100%		0%		0%	10	100%	10	100%		0%		0%	10	100%	10	100%		0%
<i>Other</i>	3	3	100%		0%		0%	3	100%	3	100%		0%		0%	3	100%	3	100%		0%
Western Blot	40	40	100%	0	0%	0	0%	40	100%	40	100%	0	0%	0	0%	40	100%	40	100%	0	0%
<i>Bio-Rad [1]</i>	35	35	100%		0%		0%	35	100%	35	100%		0%		0%	35	100%	35	100%		0%
<i>Other</i>	5	5	100%		0%		0%	5	100%	5	100%		0%		0%	5	100%	5	100%		0%
Other Methods	6	6	100%	0	0%	0	0%	6	100%	6	100%	0	0%	0	0%	6	100%	6	100%	0	0%
Analyte Total	309	309	100%	0	0%	0	0%	309	100%	309	100%	0	0%	0	0%	309	100%	309	100%	0	0%

HIV p24 Antigen

Method <i>Manufacturer</i>		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
		61				62				63				64				65			
No. Labs	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	
EIA	4	4	100%	0	0%	4	100%	0	0%	4	100%	0	0%	0	0%	4	100%	4	100%	0	0%
<i>Beckman Coulter</i>	4	4	100%		0%	4	100%		0%	4	100%		0%		0%	4	100%	4	100%		0%
Analyte Total	4	4	100%	0	0%	4	100%	0	0%	4	100%	0	0%	0	0%	4	100%	4	100%	0	0%

HTLV Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	36				37				38				39				40			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	46	46	100%	0	0%	0	0%	46	100%	46	100%	0	0%	46	100%	0	0%	0	0%	46	100%
<i>Abbott</i>	29	29	100%		0%		0%	29	100%	29	100%		0%	29	100%		0%		0%	29	100%
<i>bioMérieux</i>	17	17	100%		0%		0%	17	100%	17	100%		0%	17	100%		0%		0%	17	100%
Analyte Total	46	46	100%	0	0%	0	0%	46	100%	46	100%	0	0%	46	100%	0	0%	0	0%	46	100%

Infectious Mononucleosis

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method	No. Labs	26				27				28				29				30			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
CICA	29	0	0%	29	100%	29	100%	0	0%	29	100%	0	0%	0	0%	29	100%	29	100%	0	0%
<i>Cardinal Health SP</i>	11		0%	11	100%	11	100%		0%	11	100%		0%		0%	11	100%	11	100%		0%
<i>Others</i>	18		0%	18	100%	18	100%		0%	18	100%		0%		0%	18	100%	18	100%		0%
Hemagglutination	49	1	2%	48	98%	49	100%	0	0%	49	100%	0	0%	0	0%	49	100%	36	73%	13	27%
<i>Fisher</i>	15		0%	15	100%	15	100%		0%	15	100%		0%		0%	15	100%	4	27%	11	73%
<i>Wampole/Zeus</i>	25	1	4%	24	96%	25	100%		0%	25	100%		0%		0%	25	100%	24	96%	1	4%
<i>Others</i>	9		0%	9	100%	9	100%		0%	9	100%		0%		0%	9	100%	8	89%	1	11%
Latex Agglutination	152	1	1%	151	99%	151	99%	1	1%	152	100%	0	0%	1	1%	151	99%	151	99%	1	1%
<i>Fisher</i>	53		0%	53	100%	53	100%		0%	53	100%		0%		0%	53	100%	53	100%		0%
<i>Remel</i>	22		0%	22	100%	22	100%		0%	22	100%		0%		0%	22	100%	22	100%		0%
<i>Wampole/Zeus</i>	55		0%	55	100%	55	100%		0%	55	100%		0%		0%	55	100%	55	100%		0%
<i>Others</i>	22	1	5%	21	95%	21	95%	1	5%	22	100%		0%	1	5%	21	95%	21	95%	1	5%
Solid Phase IA	57	0	0%	57	100%	57	100%	0	0%	57	100%	0	0%	0	0%	57	100%	57	100%	0	0%
<i>Seradyn</i>	21		0%	21	100%	21	100%		0%	21	100%		0%		0%	21	100%	21	100%		0%
<i>Wampole/Zeus</i>	27		0%	27	100%	27	100%		0%	27	100%		0%		0%	27	100%	27	100%		0%
<i>Others</i>	9		0%	9	100%	9	100%		0%	9	100%		0%		0%	9	100%	9	100%		0%
Other Methods	3		0%	3	100%	3	100%		0%	3	100%		0%		0%	3	100%	3	100%		0%
Analyte Total	290	2	1%	288	99%	289	100%	1	0%	290	100%	0	0%	1	0%	289	100%	276	95%	14	5%

Note: The expected result for sample #30 is negative. Laboratories who reported any other result than negative for this sample should critically examine their protocol and test kit/method.

Infectious Mononucleosis Latex Agglutination Procedures

The number of laboratories that reported titers are summarized for positive test samples 26 and 29. The dilution schemes laboratories used are represented by the letter A and B. Only methods with 10 or more laboratories reporting titers are listed in this table.

Method	No. Labs	Sample 26 Titers							Sample 29 Titers					
		A	2	4	8	16	32	64	128	2	4	8	16	32
Latex Agglutination	19	A	2	1	2	6	1		2	1	3	5	2	1
		B	1	2	2					4				

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers.

Lyme Disease Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	41				42				43				44				45			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	66	66	100%	0	0%	66	100%	0	0%	65	98%	1	2%	0	0%	66	100%	1	2%	65	98%
<i>Immunitics</i>	17	17	100%		0%	17	100%		0%	17	100%		0%		0%	17	100%		0%	17	100%
<i>MarDx</i>	12	12	100%		0%	12	100%		0%	12	100%		0%		0%	12	100%		0%	12	100%
<i>Wampole /Zeus</i>	30	30	100%		0%	30	100%		0%	29	97%	1	3%		0%	30	100%	1	3%	29	97%
<i>Others</i>	7	7	100%		0%	7	100%		0%	7	100%		0%		0%	7	100%		0%	7	100%
ELFA	30	30	100%	0	0%	30	100%	0	0%	30	100%	0	0%	0	0%	30	100%	0	0%	30	100%
<i>bioMérieux</i>	30	30	100%		0%	30	100%		0%	30	100%		0%		0%	30	100%		0%	30	100%
Other Methods	8	8	100%		0%	8	100%		0%	8	100%		0%		0%	8	100%		0%	8	100%
Analyte Total	104	104	100%	0	0%	104	100%	0	0%	103	99%	1	1%	0	0%	104	100%	1	1%	103	99%

Lyme Western Blot IgG

		Participant Results/ Sample Number																															
		R = Reactive/ Positive; N = Non-Reactive/ Negative; E = Equivocal/ Indeterminate																															
Method	No. Labs	41						42						43						44						45							
		N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%		
Western Blot - IgG	26	26	100%	0	0%	0	0%	26	100%	0	0%	0	0%	26	100%	0	0%	0	0%	0	0%	0	0%	0	0%	26	100%	1	4%	0	0%	25	96%
<i>MarDx</i>	24	24	100%		0%		0%	24	100%		0%		0%	24	100%		0%		0%		0%		0%		0%	24	100%	1	4%		0%	23	96%
<i>Other</i>	2	2	100%		0%		0%	2	100%		0%		0%	2	100%		0%		0%		0%		0%		0%	2	100%		0%		0%	2	100%
Other Methods	4	4	100%		0%		0%	4	100%		0%		0%	4	100%		0%		0%		0%		0%		0%	4	100%		0%		0%	4	100%
Analyte Total	30	30	100%	0	0%	0	0%	30	100%	0	0%	0	0%	30	100%	0	0%	0	0%	0	0%	0	0%	0	0%	30	100%	1	3%	0	0%	29	97%

Lyme Western Blot IgM

		Participant Results/ Sample Number																													
		R = Reactive/ Positive; N = Non-Reactive/ Negative; E = Equivocal/ Indeterminate																													
Method	No. Labs	41						42						43						44*						45					
		N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%
Western Blot - IgM	26	26	100%	0	0%	0	0%	26	100%	0	0%	0	0%	26	100%	0	0%	0	0%	11	42%	1	4%	14	54%	24	92%	2	8%	0	0%
<i>MarDx</i>	24	24	100%		0%		0%	24	100%		0%		0%	24	100%		0%		0%	9	38%	1	4%	14	58%	22	92%	2	8%		0%
<i>Other</i>	2	2	100%		0%		0%	2	100%		0%		0%	2	100%		0%		0%	2	100%		0%		0%	2	100%		0%		0%
Other Methods	3	3	100%		0%		0%	3	100%		0%		0%	3	100%		0%		0%	2	67%		0%	1	33%	3	100%		0%		0%
Analyte Total	29	29	100%	0	0%	0	0%	29	100%	0	0%	0	0%	29	100%	0	0%	0	0%	13	45%	1	3%	15	52%	27	93%	2	7%	0	0%

* Test sample #44 was not authenticated - A consensus agreement cannot be reached among participants, by regulation requirement, the sample cannot be graded (scored) and all participating laboratories get credit for this sample.

Rheumatoid Factor

Method <i>Manufacturer</i>		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
		26				27				28				29				30			
No. Labs	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	
EIA	11	0	0%	11	100%	11	100%	0	0%	0	0%	11	100%	11	100%	0	0%	11	100%	0	0%
<i>Others</i>	11		0%	11	100%	11	100%		0%		0%	11	100%	11	100%		0%	11	100%		0%
Latex Agglutination	128	0	0%	128	100%	127	99%	1	1%	0	0%	128	100%	125	98%	3	2%	127	99%	1	1%
<i>Becton Dickinson</i>	19		0%	19	100%	19	100%		0%		0%	19	100%	19	100%		0%	19	100%		0%
<i>Fisher</i>	47		0%	47	100%	47	100%		0%		0%	47	100%	47	100%		0%	47	100%		0%
<i>Seradyn</i>	12		0%	12	100%	12	100%		0%		0%	12	100%	12	100%		0%	12	100%		0%
<i>Wampole/Zeus</i>	13		0%	13	100%	12	92%	1	8%		0%	13	100%	11	85%	2	15%	12	92%	1	8%
<i>Others</i>	37		0%	37	100%	37	100%		0%		0%	37	100%	36	97%	1	3%	37	100%		0%
Nephelometry	43	0	0%	43	100%	43	100%	0	0%	0	0%	43	100%	43	100%	0	0%	38	88%	0	0%
<i>Beckman Coulter^[1]</i>	22		0%	22	100%	22	100%		0%		0%	22	100%	22	100%		0%	17	77%		0%
<i>Behring</i>	21		0%	21	100%	21	100%		0%		0%	21	100%	21	100%		0%	21	100%		0%
Turbidimetry	36	1	3%	35	97%	35	97%	1	3%	1	3%	35	97%	36	100%	0	0%	36	100%	0	0%
<i>Beckman Coulter</i>	10	1	10%	9	90%	9	90%	1	10%	1	10%	9	90%	10	100%		0%	10	100%		0%
<i>Roche</i>	22		0%	22	100%	22	100%		0%		0%	22	100%	22	100%		0%	22	100%		0%
<i>Others</i>	4		0%	4	100%	4	100%		0%		0%	4	100%	4	100%		0%	4	100%		0%
Other Methods	6		0%	6	100%	6	100%		0%		0%	6	100%	6	100%		0%	6	100%		0%
Analyte Total	224	1	0%	223	100%	222	99%	2	1%	1	0%	223	100%	221	99%	3	1%	218	97%	1	0%

^[1] Of 6 labs running the Beckman Coulter Array, 5 of them reported an "unstable reaction" on sample #30.

Rheumatoid Factor Latex Agglutination Procedure

The number of laboratories that reported titers is listed for positive test samples 26 and 28. The dilution schemes laboratories used are represented by the letter A (IU/ml) and B (titer), testing systems with 10 or more laboratories reporting titers are listed in this table.

<i>Manufacturer</i>	No. Labs	A B	Sample 26 Titers					Sample 28 Titers						
			160 16	320 32	640 64	1280 128	2560 256	5120 512	80 8	160 16	320 32	640 64	1280 128	2560 256
<i>Fisher</i>	36	A B		3 1	1 3	11 9	3		1	7 3	11 8	1		1
<i>Becton Dickson</i>	15	A B	2	4	6	3		2	5	6	1	1		

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers.

Rheumatoid Factor

Results are summarized for positive test samples 26 and 28. The Mean values \pm S.D. are given where 10 or more laboratories reported quantitative results. Outlier values are omitted.

Method <i>Manufacturer</i>	No. Labs	IU/ml Sample 26	IU/ml Sample 28
Nephelometry <i>Beckman Coulter IMMAGE</i>	16	1212 \pm 66	652 \pm 82
Turbidimetry <i>Roche Diag. Cobas</i>	13	694 \pm 58	326 \pm 16

Rubella IgG Antibody

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method	No. Labs	11				12				13				14				15			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Chemiluminescence	59	0	0%	59	100%	59	100%	0	0%	1	2%	58	98%	59	100%	0	0%	59	100%	0	0%
<i>Bayer</i>	27		0%	27	100%	27	100%		0%		0%	27	100%	27	100%		0%	27	100%		0%
<i>Beckman</i>	11		0%	11	100%	11	100%		0%		0%	11	100%	11	100%		0%	11	100%		0%
<i>Diagnostic Products</i>	21		0%	21	100%	21	100%		0%	1	5%	20	95%	21	100%		0%	21	100%		0%
EIA	63	2	3%	61	97%	63	100%	0	0%	1	2%	62	98%	63	100%	0	0%	63	100%	0	0%
<i>Abbott</i>	27	1	4%	26	96%	27	100%		0%		0%	27	100%	27	100%		0%	27	100%		0%
<i>Diamedix</i>	14	1	7%	13	93%	14	100%		0%	1	7%	13	93%	14	100%		0%	14	100%		0%
<i>Wampole/Zeus</i>	12		0%	12	100%	12	100%		0%		0%	12	100%	12	100%		0%	12	100%		0%
<i>Others</i>	10		0%	10	100%	10	100%		0%		0%	10	100%	10	100%		0%	10	100%		0%
ELFA	34	0	0%	34	100%	33	97%	1	3%	0	0%	34	100%	33	97%	1	3%	34	100%	0	0%
<i>bioMérieux Vidas</i>	34		0%	34	100%	33	97%	1	3%		0%	34	100%	33	97%	1	3%	34	100%		0%
Latex Agglutination	38	1	3%	37	97%	38	100%	0	0%	0	0%	38	100%	38	100%	0	0%	38	100%	0	0%
<i>Fisher</i>	15	1	7%	14	93%	15	100%		0%		0%	15	100%	15	100%		0%	15	100%		0%
<i>Others</i>	23		0%	23	100%	23	100%		0%		0%	23	100%	23	100%		0%	23	100%		0%
Other Methods	8		0%	8	100%	8	100%		0%		0%	8	100%	8	100%		0%	8	100%		0%
Analyte Total	202	3	1%	199	99%	201	100%	1	0%	2	1%	200	99%	201	100%	1	0%	202	100%	0	0%

Rubella IgG Antibody

Results are summarized for positive test samples 11 and 13. The Mean values \pm S.D. are given where 10 or more laboratories reported quantitative results. Outlier values are omitted.

Method	No. <i>Manufacturer</i>	Labs	Unit	Sample 11	Sample 13
Chemiluminescence					
	<i>Bayer</i>	22	IU/ml	110 \pm 13	194 \pm 22
	<i>Diagnostic Products</i>	18	IU/ml	15 \pm 2.3	75 \pm 9
EIA					
	<i>Abbott AxSYM</i>	11	IU/ml	19 \pm 2.5	70 \pm 4.3

Rubella IgM Specific

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method <i>Manufacturer</i>	No. Labs	56				57				58				59				60			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	14	14	100%	0	0%	0	0%	14	100%	14	100%	0	0%	14	100%	0	0%	12	86%	1	7%
<i>Other</i> ^[1]	14	14	100%		0%		0%	14	100%	14	100%		0%	14	100%		0%	12	86%	1	7%
Other Methods	7	7	100%		0%		0%	7	100%	7	100%		0%	7	100%		0%	7	100%		0%
Analyte Total	21	21	100%	0	0%	0	0%	21	100%	21	100%	0	0%	21	100%	0	0%	19	90%	1	5%

^[1] An equivocal or borderline result on sample #60 was reported by 1 lab.

Syphilis - Reagin Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	1				2				3*				4				5			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
RPR	276	2	1%	274	99%	275	100%	1	0%	234	85%	42	15%	275	100%	1	0%	1	0%	275	100%
<i>ASI</i>	36		0%	36	100%	36	100%		0%	32	89%	4	11%	36	100%		0%		0%	36	100%
<i>Becton Dickinson</i>	140		0%	140	100%	140	100%		0%	112	80%	28	20%	140	100%		0%		0%	140	100%
<i>Fisher</i>	48	2	4%	46	96%	47	98%	1	2%	41	85%	7	15%	47	98%	1	2%	1	2%	47	98%
<i>True Medix</i>	16		0%	16	100%	16	100%		0%	16	100%		0%	16	100%		0%		0%	16	100%
<i>Wampole/Zeus</i>	22		0%	22	100%	22	100%		0%	22	100%		0%	22	100%		0%		0%	22	100%
<i>Others</i>	14		0%	14	100%	14	100%		0%	11	79%	3	21%	14	100%		0%		0%	14	100%
Other Methods	1		0%	1	100%	1	100%		0%	1	100%		0%	1	100%		0%		0%	1	100%
Analyte Total	277	2	1%	275	99%	276	100%	1	0%	235	85%	42	15%	276	100%	1	0%	1	0%	276	100%

* Test sample #3 was not authenticated - A consensus agreement cannot be reached among participants, by regulation requirement, the sample cannot be graded (scored) and all participating laboratories get credit for this sample.

Syphilis - Reagin Antibody

RPR Procedures

The number of laboratories that reported titers is listed for positive test samples 1 and 5 for the RPR procedure. Only testing systems with 10 or more laboratories reporting titers are listed in this table.

Method <i>Manufacturer</i>	No. Labs	Sample 1 Titers						Sample 5 Titers					
		2	4	8	16	32	64	2	4	8	16	32	64
<i>Becton Dickinson</i>	126		17	80	28		1	2	36	82	4	1	
<i>Fisher</i>	47		11	23	8	1			22	19	16		
<i>Wampole/ Zeus</i>	22	1	3	16	2			1	12	8			
<i>ASI</i>	24		7	14	3				14	10			
<i>True Medix</i>	16	1	6	6	3				7	7	2		

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers.

Syphilis - Treponemal Antibody

Method <i>Manufacturer</i>		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
		1				2				3*				4				5			
No. Labs	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	
EIA	10	0	0%	10	100%	10	100%	0	0%	10	100%	0	0%	10	100%	0	0%	0	0%	10	100%
<i>Other</i>	10		0%	10	100%	10	100%		0%	10	100%		0%	10	100%		0%		0%	10	100%
Gel. Part. Agglut.	34	0	0%	34	100%	34	100%	0	0%	33	97%	1	3%	33	97%	1	3%	1	3%	33	97%
<i>Fujirebio</i>	34		0%	34	100%	34	100%		0%	33	97%	1	3%	33	97%	1	3%	1	3%	33	97%
IFA	26	0	0%	26	100%	26	100%	0	0%	25	96%	1	4%	26	100%	0	0%	0	0%	26	100%
<i>Wampole/Zeus</i>	26		0%	26	100%	26	100%		0%	25	96%	1	4%	26	100%		0%		0%	26	100%
MHA	12	0	0%	12	100%	12	100%	0	0%	2	17%	10	83%	12	100%	0	0%	0	0%	12	100%
<i>Olympus</i>	12		0%	12	100%	12	100%		0%	2	17%	10	83%	12	100%		0%		0%	12	100%
Other Methods	7		0%	7	100%	7	100%		0%	7	100%		0%	7	100%		0%		0%	7	100%
Analyte Total	89	0	0%	89	100%	89	100%	0	0%	77	87%	12	13%	88	99%	1	1%	1	1%	88	99%

* Test sample #3 was not authenticated - A consensus agreement cannot be reached among participants, by regulation requirement, the sample cannot be graded (scored) and all participating laboratories get credit for this sample.

AAT, C'3, and C'4

The Mean mg/dl \pm S.D. is given where 10 or more laboratories reported values. Outlier values are omitted.

Analytes	Sample NO.					
	No. Labs	76	77	78	79	80
Alpha-1-Antitrypsin						
Nephelometry/Beckman Coulter IMMAGE	12	12 \pm 1.5	412 \pm 20	415 \pm 14	< 10	169 \pm 7.1
Nephelometry/Behring Nephelometer	19	18 \pm 1.1	490 \pm 59	483 \pm 54	< 17	167 \pm 8.5
Nephelometry/ Total	33	16 \pm 3.1	451 \pm 49	444 \pm 33	< 17	168 \pm 8.7
Turbidimetry/ Total	7	< 36	438 \pm 20	445 \pm 33	< 36	153 \pm 6.4
Complement C'3						
Nephelometry/Beckman Coulter IMMAGE	20	45 \pm 1.7	437 \pm 24	434 \pm 17	50 \pm 2.1	194 \pm 11
Nephelometry/Behring Nephelometer	18	46 \pm 3.5	491 \pm 27	495 \pm 29	52 \pm 3.8	203 \pm 12
Turbidimetry/Roche Cobas Integra	12	44 \pm 2.0	477 \pm 16	472 \pm 22	48 \pm 1.7	195 \pm 6.6
Nephelometry/ Total	54	45 \pm 1.8	460 \pm 35	461 \pm 37	50 \pm 2.4	195 \pm 7.3
Turbidimetry/ Total	42	46 \pm 3.5	461 \pm 46	470 \pm 33	50 \pm 3.6	191 \pm 12
Complement C'4						
Nephelometry/Beckman Coulter IMMAGE	20	5.1 \pm 0.2	118 \pm 5.3	118 \pm 5.3	4.9 \pm 0.2	34 \pm 2.0
Nephelometry/Behring Nephelometer	16	4.8 \pm 0.2	100 \pm 6.6	102 \pm 8.6	4.3 \pm 0.4	30 \pm 1.7
Turbidimetry/Roche Cobas Integra	11	< 6	99 \pm 2.6	100 \pm 2.5	< 6	30 \pm 1.0
Nephelometry/ Total	54	< 10	107 \pm 11	108 \pm 11	< 10	32 \pm 2.6
Turbidimetry/ Total	40	< 16	100 \pm 11	101 \pm 11	< 16	29 \pm 2.3

IgA, IgE, IgG, and IgM

The Mean mg/dl (IU/ml for IgE) \pm S.D. is given for IgA, IgE, and IgM and Mean mg/dl \pm 25% is given for IgG where 10 or more laboratories reported values. Outlier values are omitted.

Analytes (Unit)	No.	Sample NO.					
		Labs	81	82	83	84	85
Immunoglobulin A (mg/dl)							
Nephelometry/Beckman Coulter Array	11		751 \pm 31	281 \pm 13	270 \pm 7.6	104 \pm 3.4	595 \pm 9.5
Nephelometry/Beckman Coulter IMMAGE	21		708 \pm 23	279 \pm 6.7	257 \pm 12	99 \pm 3.5	613 \pm 25
Nephelometry/Behring Nephelometer	22		729 \pm 31	279 \pm 10	274 \pm 6.5	101 \pm 5.2	633 \pm 25
Turbidimetry/Roche Cobas Integra	13		842 \pm 60	263 \pm 5.2	240 \pm 4.2	90 \pm 3.4	559 \pm 34
Nephelometry/ Total	67		723 \pm 37	277 \pm 10	267 \pm 13	100 \pm 4.5	614 \pm 34
Turbidimetry/ Total	46		737 \pm 92	271 \pm 11	244 \pm 11	93 \pm 6.2	571 \pm 49
Immunoglobulin E (IU/ml)							
Chemiluminescence/Bayer	10		849 \pm 77	282 \pm 17	292 \pm 18	21 \pm 1.3	434 \pm 30
Chemiluminescence/Diag. Prod. Co.	29		811 \pm 65	276 \pm 13	290 \pm 23	28 \pm 1.9	405 \pm 25
FEIA/Pharmacia	10		778 \pm 114	265 \pm 41	275 \pm 34	30 \pm 2.1	427 \pm 26
Chemiluminescence/ Total	44		824 \pm 75	275 \pm 17	287 \pm 23	26 \pm 3.4	413 \pm 31
FEIA/ Total	17		780 \pm 89	268 \pm 32	276 \pm 28	30 \pm 3.1	415 \pm 37
Nephelometry/ Total	14		857 \pm 67	270 \pm 13	295 \pm 28	25 \pm 2.9	428 \pm 33
Immunoglobulin G (mg/dl)							
Nephelometry/Beckman Coulter Array	10		1366 \pm 62	1400 \pm 53	1153 \pm 46	619 \pm 25	2342 \pm 151
Nephelometry/Beckman Coulter IMMAGE	21		1399 \pm 53	1449 \pm 48	1198 \pm 42	630 \pm 32	2425 \pm 87
Nephelometry/Behring Nephelometer	23		1514 \pm 73	1550 \pm 85	1258 \pm 51	671 \pm 26	3051 \pm 166
Turbidimetry/Roche Cobas Integra	13		1354 \pm 30	1433 \pm 31	1130 \pm 21	608 \pm 11	2552 \pm 66
Nephelometry/ Total	67		1445 \pm 86	1488 \pm 88	1214 \pm 65	647 \pm 35	2700 \pm 348
Turbidimetry/ Total	45		1350 \pm 74	1413 \pm 66	1130 \pm 54	592 \pm 28	2468 \pm 200
Immunoglobulin M (mg/dl)							
Nephelometry/Beckman Coulter Array	10		173 \pm 8.2	92 \pm 4.5	771 \pm 39	47 \pm 1.2	233 \pm 13
Nephelometry/Beckman Coulter IMMAGE	21		153 \pm 6.9	93 \pm 3.6	687 \pm 56	43 \pm 1.7	214 \pm 6.7
Nephelometry/Behring Nephelometer	22		234 \pm 17	95 \pm 3.9	840 \pm 52	43 \pm 2.8	315 \pm 19
Turbidimetry/Roche Cobas Integra	14		144 \pm 2.9	90 \pm 3.4	901 \pm 19	36 \pm 2.0	176 \pm 4.1
Nephelometry/ Total	65		198 \pm 49	93 \pm 3.3	770 \pm 83	44 \pm 3.0	260 \pm 49
Turbidimetry/ Total	45		143 \pm 5.7	92 \pm 5.6	819 \pm 89	38 \pm 3.0	197 \pm 27

Acceptable Response (April 26, 2006 PT Event)
Quantitative Tests Results (Acceptable Range)

Analytes	Sample NO.				
Method/ Manufacture					
Alpha-1-Antitrypsin	76	77	78	79	80
Nephelometry/Beckman Coulter Immage	7 - 17	352 - 472	373 - 457	< 10	147 - 191
Nephelometry/Dade Behring Neph.	14 - 22	313 - 666	322 - 645	< 17	141 - 193
Nephelometry/ Total	6 - 26	303 - 598	343 - 544	< 17	142 - 195
Turbidimetry/ Total	< 36	378 - 497	347 - 544	< 36	133 - 172
Complement C'3	76	77	78	79	80
Nephelometry/Beckman Coulter Immage	40 - 51	366 - 509	384 - 484	43 - 57	161 - 226
Nephelometry/Dade Behring Neph.	35 - 56	411 - 572	406- 583	40- 64	168 - 239
Turbidimetry/Roche Cobas Integra	37 - 50	428 - 525	403 - 540	42 - 54	175 - 215
Nephelometry/ Total	39 - 51	356 - 564	351 - 571	42 - 58	172 - 217
Turbidimetry/ Total	35 - 57	323 - 600	371 - 571	39 - 61	157 - 227
Complement C'4	76	77	78	79	80
Nephelometry/Beckman Coulter Immage	4 - 6	102 - 134	102 - 134	4 - 6	28 - 41
Nephelometry/Dade Behring Neph.	4 - 6	80 - 120	75 - 127	3 - 6	24 - 36
Turbidimetry/Roche Cobas Integra	< 6	91 - 107	92 - 107	< 6	27 - 31
Nephelometry/ Total	< 10	72 - 142	74 - 141	< 10	23 - 40
Turbidimetry/ Total	< 16	67 - 133	68 - 133	< 16	22 - 37
Immunoglobulin A	81	82	83	84	85
Nephelometry/Beckman Coulter Array	656 - 846	242 - 319	246 - 293	93 - 115	566 - 625
Nephelometry/Beckman Coulter Immage	640 - 776	259 - 300	219 - 294	88 - 110	537 - 688
Nephelometry/Dade Behring Neph .	635 - 823	248 - 309	255 - 294	85 - 117	557 - 709
Turbidimetry/Roche Cobas Integra	662 - 1021	247 - 279	227 - 252	79 - 100	456 - 662
Nephelometry/ Total	613 - 834	247 - 308	228 - 306	87 - 114	511 - 717
Turbidimetry/ Total	462 - 1013	237 - 305	210 - 277	74 - 112	422 - 720
Immunoglobulin E	81	82	83	84	85
Chemiluminescence/Bayer Advia Centaur	618 - 1081	232 - 333	237 - 345	17 - 26	344 - 524
Chemiluminescence/Diag.Prod. Immulite	616 - 1005	237 - 314	222 - 357	22 - 33	329 - 480
FEIA/Pharmacia Immunocap	437 - 1120	141 - 389	172 - 378	24 - 37	349 - 505
Nephelometry/Dade Behring Neph.	719 - 1050	220 - 315	215 - 365	20 - 30	357 - 545
Chemiluminescence/ Total	598 - 1049	222 - 328	218 - 355	15 - 36	321 - 506
FEIA/ Total	511 - 1049	172 - 365	191 - 360	20 - 39	305 - 525
Nephelometry/ Total	656 - 1057	230 - 310	211 - 380	15 - 33	328 - 528
Immunoglobulin G	81	82	83	84	85
Nephelometry/Beckman Coulter Array	1024 - 1708	1050 - 1750	864 - 1442	464 - 774	1756 - 2927
Nephelometry/Beckman Coulter Immage	1049 - 1749	1086 - 1811	898 - 1498	472 - 787	1819 - 3032
Nephelometry/Dade Behring Neph .	1135 - 1892	1162 - 1938	943 - 1573	503 - 839	2288 - 3814
Turbidimetry/Roche Cobas Integra	1015 - 1693	1074 - 1792	848 - 1413	456 - 761	1913 - 3190
Nephelometry/ Total	1083 - 1806	1116 - 1859	911 - 1516	485 - 809	2025 - 3376
Turbidimetry/ Total	1012 - 1688	1059 - 1766	847 - 1412	443 - 740	1851 - 3086
Immunoglobulin M	81	82	83	84	85
Nephelometry/Beckman Coulter Array	148 - 198	77 - 106	654 - 887	43 - 52	192 - 273
Nephelometry/Beckman Coulter Immage	132 - 174	82 - 104	518 - 856	38 - 50	193 - 234
Nephelometry/Dade Behring Neph .	181 - 286	83 - 107	683 - 998	34 - 52	259 - 371
Turbidimetry/Roche Cobas Integra	134 - 153	79 - 100	845 - 957	30 - 43	163 - 189
Nephelometry/ Total	51 - 345	83 - 103	520 - 1021	34 - 53	114 - 407
Turbidimetry/ Total	126 - 161	74 - 109	551 - 1087	29 - 48	115 - 280

**Acceptable Response (April 26, 2006 PT Event)
Qualitative / Quantitative Tests Results**

Analytes	Sample NO.				
	1	2	3	4	5
Syphilis - Reagin Ab	R	N	N*	N	R
<i>RPR Titer</i>	4 - 16				2 - 16
Syphilis - Treponemal	R	N	N*	N	R
	6	7	8	9	10
HBcAb	R	N	R	R	R
HBsAg	N	N	R	N	R
	11	12	13	14	15
CMV	R*	N	R	N	N
Rubella Ab	R	N	R	N	N
	16	17	18	19	20
ASO	N	R	N	R	N
<i>Latex Agglutination as IU/ml</i>		100 - 1600		100 - 1600	
	26	27	28	29	30
Infectious Mono.	R	N	N	R	N
<i>Latex Agglutination Titer</i>	4 - 64			2 - 64	
Rheumatoid Factor	R	N	R	N	N
<i>Latex Agglutination Titer</i>	16 - 512		8 - 256		
<i>Latex Agglutination as IU/ml</i>	160 - 5120		80 - 2560		
	31	32	33	34	35
HIV Ab Screening/Confirmation	N	R	N	R	N
	36	37	38	39	40
HTLV 1 Ab	N	R	N	N	R
	41	42	43	44	45
LYME Disease Ab	N	N	N	R	R
LYME Disease Ab WB IgG	N	N	N	R	R
LYME Disease Ab WB IgM	N	N	N	R*	N
	46	47	48	49	50
ANA	R	N	N	N	R
<i>EIA Titer</i>	40 - 640				160 - 2560
<i>IFA Titer</i>	40 - 1280				160 - 5120
	56	57	58	59	60
Rubella IgM	N	R	N	N	N
	61	62	63	64	65
HIV p24 Ag	N	N	N	R	N
	66	67	68	69	70
HBeAg	N	R	N	R	N
	71	72	73	74	75
Hepatitis C Ab	N	N	R	N	R

Note: R = Reactive/ Positive; I = Indeterminate; N = Non-Reactive/ Negative

* Test sample was not authenticated - because a consensus of 80% agreement was not reached. All participating laboratories received credit for this sample.