

New York State Department of Health
Wadsworth Center

Proficiency Testing Program

13-May-09

**DIAGNOSTIC IMMUNOLOGY &
HUMAN IMMUNODEFICIENCY VIRUS
SUMMARY ANALYSIS**

**Proficiency Test Event
13-May-09**

**Diagnostic Immunology & Human Immunodeficiency Virus
Summary Report**

Steven Bush M.S., Susan Wong Ph.D., and Mary Marchewka

Table of Contents

Review of Grading Policy	Page 3
Determination of Acceptable Responses	Page 4
Qualitative / Quantitative Test Analytes	Page 5 - 27
Antinuclear Ab	Page 5 - 6
Antistreptolysin O	Page 7 - 9
Cytomegalovirus Ab	Page 10
Hepatitis B Core Ab	Page 11
Hepatitis B Surface Ag Screening/ Confirmation	Page 12
Hepatitis Be Ag	Page 13
Hepatitis C Ab Screening/ Confirmation	Page 13
HIV Ab Screening/ Confirmation	Page 14
HTLV1 Ab	Page 15
Infectious Mononucleosis	Page 16
Lyme Disease (<u>Borrelia burgdorferi</u>) Ab	Page 17 - 18
Rheumatoid Factor	Page 19 - 21
Rubella Ab	Page 22 - 23
Rubella IgM	Page 24
Syphilis-Reagin	Page 25 - 26
Syphilis-Treponemal Ab	Page 27
Quantitative Test Analytes	Page 28 - 29
Alpha-1 Antitrypsin	Page 28
Complement C'3	Page 28
Complement C'4	Page 28
Immunoglobulin A	Page 29
Immunoglobulin E	Page 29
Immunoglobulin G	Page 29
Immunoglobulin M	Page 29
Acceptable Responses	Page 30 - 31

The data summarized in this report were tabulated from test results and accompanying information submitted by laboratories that participated in the May 13, 2009 Diagnostic Immunology and Human Immunodeficiency Virus proficiency test events. Participants are encouraged to review the data and to compare results and test kit performances.

Laboratories were evaluated on the basis of their responses for each analyte and on overall performance for all the analytes tested in the permit category. Appropriate responses were determined by participant consensus requiring 80% agreement in each test.

Qualitative/quantitative results were graded in relation to results given by participants for specific test kits. When the number of participants that used a specific test kit was less than 10, results were graded considering results given for the method used. Target values and acceptable ranges were determined as indicated in Page 4.

Grading Criteria:

- ⊕ For each separate analyte where results were reported, qualitative or quantitative, twenty points were deducted for each incorrect answer. For Syphilis-Reagin (RPR), where both qualitative and quantitative results are reported under one analyte, ten points were deducted for each incorrect quantitative or qualitative result. Titering of positive Syphilis-Reagin samples is mandatory for all Diagnostic Services laboratories who perform this test, unless given an exemption. Failure to titer the positive samples to the endpoint will result in failure for the Syphilis-Reagin analyte.
- ⊕ For **Diagnostic Services** failure to attain an overall testing score of at least 80% is unsatisfactory performance
- ⊕ For **Donor Services** failure to attain an overall testing score of 100% is unsatisfactory
- ⊕ For **HIV** failure to attain an overall testing score of 100% is unsatisfactory performance.
- ⊕ Laboratories failing two out of three consecutive proficiency test events for an analyte or for the permit category will fail the proficiency testing program for the analyte or for the permit category and may be required to cease patient testing for that analyte/category.

Summary Tables

Test kit manufacturer names are in *italics*. In some tables, test kits are grouped under test methods shown in bold letters. In all tables, test methods and test kit manufacturer names are listed in alphabetic order. Only the testing systems used by 10 or more laboratories are listed in this report.

For qualitative tests, results are summarized as the number of laboratories that reported a test sample as reactive to the number that reported it as non-reactive. In addition, where test results depend on a quantitative value (e.g. titer, IU/ml) the values reported are given in separate tables. They are expressed, where applicable, as the Mean \pm S.D. when ten or more laboratories reported data.

For quantitative tests, values reported variously as mg/dl, IU/ml, etc. are given as the Mean \pm S.D. when ten or more laboratories reported results. Titers are given as endpoint titers.

Disclaimer

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

Determination of Acceptable Responses

Analyte or Test	Criteria
Alpha-1 Antitrypsin	Target value \pm 3 S.D.
Antinuclear Antibody	Positive or negative
Antinuclear Antibody Quantitative (IFA systems only)	Target value \pm 2 dilutions
Antistreptolysin O	Positive or negative
Antistreptolysin O Quantitative	Target value \pm 2 dilutions or Target value \pm 3 S.D.
Complement C'3, C'4	Target value \pm 3 S.D.
Cytomegalovirus Antibody	Positive or negative
Hepatitis (HbsAg, anti-HBc, HBeAg, and HCAb)	Reactive or nonreactive
HIV 1 Ab, Ag	Reactive or nonreactive
HTLV 1 Ab	Positive or negative
Lyme Disease Ab, WB IgG, IgM	Positive or negative
Immunoglobulin A, E, M	Target value \pm 3 S.D.
Immunoglobulin G	Target value \pm 25 %
Infectious Mononucleosis	Positive or negative
Rheumatoid Factor	Positive or negative
Rheumatoid Factor Quantitative	Target value \pm 2 dilutions or Target value \pm 3 S.D.
Rubella Ab, IgM	Positive or negative or Immune or nonimmune
Rubella Ab Quantitative	Target value \pm 3 S.D.
Syphilis Reagin Antibody	Reactive or nonreactive Target value \pm 1 dilution
Syphilis Treponemal Antibody	Reactive or nonreactive

Antinuclear Antibody

Participant Results/ Sample Number																						
R = Reactive/ Positive; N = Non-Reactive/ Negative																						
Method	No. Labs	46 R				47 R				48 N				49 N				50 N				
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	
EIA	38		0%	38	100%		0%	38	100%	38	100%		0%	38	100%		0%	38	100%		0%	
IFA	63	0	0%	63	100%	2	3%	61	97%	63	100%	0	0%	63	100%	0	0%	63	100%	0	0%	
<i>Bio-Rad</i>	19		0%	19	100%	1	5%	18	95%	19	100%		0%	19	100%		0%	19	100%		0%	
<i>Wampole/Zeus</i>	16		0%	16	100%		0%	16	100%	16	100%		0%	16	100%		0%	16	100%		0%	
<i>Others</i>	28		0%	28	100%	1	4%	27	96%	28	100%		0%	28	100%		0%	28	100%		0%	
Multiplexed Bead¹	17		0%	17	100%		0%	16	94%	17	100%		0%	17	100%		0%	17	100%		0%	
Other Methods	7	1	14%	6	86%	1	14%	6	86%	6	86%	1	14%	6	86%	1	14%	6	86%	1	14%	
Analyte Total	125	1	1%	124	99%	3	2%	121	97%	124	99%	1	1%	124	99%	1	1%	124	99%	1	1%	

¹ One laboratory reported an equivocal result on sample #47.

Antinuclear Antibody Quantitative (IFA systems only)

The number of laboratories that reported titers is listed for positive test samples 46 and 47. Only testing systems with 10 or more laboratories reporting titers are listed in this table.

Method <i>Manufacturer</i>	No. Labs	Sample 46 Titer								Sample 47 Titer							
		40	80	160	320	640	1280	2560	5120	40	80	160	320	640	1280	2560	5120
IFA <i>Total</i>	81			5	15	29	15	7	3	1	15	34	20	14	5		
<i>Bio-Rad</i>	22			4	5	9	2			1	6	12	2				
<i>Immuno</i>	10				1	2	2					2	2	5			
<i>The Binding site</i>	16				2	6	3	2	1		1	3	6	2	2		
<i>Wampole/ Zeus</i>	19				3	5	2	2	1		4	8	3	2	2		

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers.

Antinuclear Antibody Staining Patterns

Staining Pattern	Sample 46		Sample 47	
	#	%	#	%
<i>Homogenous</i>			91	96%
<i>Nucleolar</i>				
<i>Peripheral</i>				
<i>Speckled</i>	94	99%	1	1%

The reporting of a staining pattern here is for informational purposes only and is not used for grading.

Antistreptolysin O

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	16 N				17 N				18 R				19 N				20 R			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Latex Agglutination	68	68	100%	0	0%	68	100%	0	0%	1	1%	67	99%	68	100%	0	0%	1	1%	67	99%
<i>Dade Behring</i>	17	17	100%		0%	17	100%		0%		0%	17	100%	17	100%		0%		0%	17	100%
<i>Fisher</i>	16	16	100%		0%	16	100%		0%		0%	16	100%	16	100%		0%		0%	16	100%
<i>Remel</i>	15	15	100%		0%	15	100%		0%		0%	15	100%	15	100%		0%		0%	15	100%
<i>Others</i>	20	20	100%		0%	20	100%		0%	1	5%	19	95%	20	100%		0%	1	5%	19	95%
Hemagglutination	10	10	100%	0	0%	10	100%	0	0%	0	0%	10	100%	10	100%	0	0%	0	0%	10	100%
<i>Wampole/Zeus</i>	10	10	100%		0%	10	100%		0%		0%	10	100%	10	100%		0%		0%	10	100%
Other Methods	15	15	100%		0%	15	100%		0%		0%	15	100%	15	100%		0%		0%	15	100%
Analyte Total	93	93	100%	0	0%	93	100%	0	0%	1	1%	92	99%	93	100%	0	0%	1	1%	92	99%

Antistreptolysin O Quantitative Latex Agglutination Procedures

The number of laboratories that reported titers is listed for positive test samples 18 and 20. Only testing systems with 10 or more laboratories reporting titers are listed in this table.

Method <i>Manufacturer</i>	No. Labs	Sample 18 Titer						Sample 20 Titer					
		200	400	800	1600	3200	6400	200	400	800	1600	3200	6400
Latex Total	54		31	7				30	9				
<i>Dade Behring</i>	15		6	2				7	2				
<i>Fisher</i>	13		9	3				10	2				
<i>Remel</i>	13		11	1				10	2				

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers.

Antistreptolysin O Quantitative
Nephelometry & Turbidimetry Procedures

Results are summarized for positive test samples 18 and 20. The Mean values \pm S.D. are given where 10 or more laboratories reported quantitative results. Outlier values are omitted.

Method	Manufacturer	No. Labs	Unit	Sample 18	Sample 20
Nephelometry	Total	16	IU/ml	524 \pm 126	544 \pm 139
Turbidimetry	Total	36	IU/ml	663 \pm 127	668 \pm 128
	<i>Roche Diagnostics Cobas</i>	16	IU/ml	749 \pm 29	751 \pm 30

Cytomegalovirus Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	11 R				12 R				13 N				14 N				15 N			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	37	0	0%	37	100%	0	0%	36	97%	37	100%	0	0%	37	100%	0	0%	36	97%	0	0%
<i>Wampole /Zeus</i>	18		0%	18	100%		0%	18	100%	18	100%		0%	18	100%		0%	18	100%		0%
<i>Others ¹</i>	19		0%	19	100%		0%	18	95%	19	100%		0%	19	100%		0%	18	95%		0%
ELFA	19	0	0%	19	100%	0	0%	19	100%	19	100%	0	0%	19	100%	0	0%	19	100%	0	0%
<i>bioMérieux Vidas</i>	19		0%	19	100%		0%	19	100%	19	100%		0%	19	100%		0%	19	100%		0%
Chemiluminescence	12	0	0%	12	100%	0	0%	12	100%	12	100%	0	0%	12	100%	0	0%	12	100%	0	0%
<i>Diagnostic Products</i>	10		0%	10	100%		0%	10	100%	10	100%		0%	10	100%		0%	10	100%		0%
<i>Others</i>	2		0%	2	100%		0%	2	100%	2	100%		0%	2	100%		0%	2	100%		0%
Hemagglutination	11	0	0%	11	100%	0	0%	11	100%	11	100%	0	0%	11	100%	0	0%	11	100%	0	0%
<i>Olympus</i>	11		0%	11	100%		0%	11	100%	11	100%		0%	11	100%		0%	11	100%		0%
Other Methods	9		0%	9	100%		0%	9	100%	9	100%		0%	9	100%		0%	9	100%		0%
Analyte Total	88	0	0%	88	100%	0	0%	87	99%	88	100%	0	0%	88	100%	0	0%	87	99%	0	0%

¹ One lab reported an equivocal on sample #12 and one lab reported an equivocal on sample #15.

Hepatitis B Core Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	6 N				7 N				8 N				9 R				10 R			
		<i>Manufacturer</i>	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R
Chemiluminescence	103	103	100%	0	0%	103	100%	0	0%	103	100%	0	0%	0	0%	103	100%	2	2%	101	98%
<i>Abbott</i>	13	13	100%		0%	13	100%		0%	13	100%		0%		0%	13	100%		0%	13	100%
<i>Bayer</i>	40	40	100%		0%	40	100%		0%	40	100%		0%		0%	40	100%		0%	40	100%
<i>Diagnostic Products</i>	18	18	100%		0%	18	100%		0%	18	100%		0%		0%	18	100%	1	6%	17	94%
<i>Ortho</i>	32	32	100%		0%	32	100%		0%	32	100%		0%		0%	32	100%	1	3%	31	97%
EIA	44	44	100%	0	0%	44	100%	0	0%	44	100%	0	0%	0	0%	44	100%	1	2%	43	98%
<i>Abbott AxSYM</i>	23	23	100%		0%	23	100%		0%	23	100%		0%		0%	23	100%	1	4%	22	96%
<i>DiaSorin</i>	10	10	100%		0%	10	100%		0%	10	100%		0%		0%	10	100%		0%	10	100%
<i>Ortho</i>	9	9	100%		0%	9	100%		0%	9	100%		0%		0%	9	100%		0%	9	100%
<i>Other</i>	2	2	100%		0%	2	100%		0%	2	100%		0%		0%	2	100%		0%	2	100%
Analyte Total	147	147	100%	0	0%	147	100%	0	0%	147	100%	0	0%	0	0%	147	100%	3	2%	144	98%

Hepatitis B Surface Antigen

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method	No. Labs	6 N				7 N				8 N				9 R				10 N			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Chemiluminescence	137	137	100%	0	0%	137	100%	0	0%	137	100%	0	0%	0	0%	137	100%	134	98%	3	2%
<i>Abbott</i>	13	13	100%		0%	13	100%		0%	13	100%		0%		0%	13	100%	13	100%		0%
<i>Bayer</i>	53	53	100%		0%	53	100%		0%	53	100%		0%		0%	53	100%	53	100%		0%
<i>Diagnostic Products</i>	19	19	100%		0%	19	100%		0%	19	100%		0%		0%	19	100%	17	89%	2	11%
<i>Ortho</i>	42	42	100%		0%	42	100%		0%	42	100%		0%		0%	42	100%	41	98%	1	2%
<i>Roche</i>	10	10	100%		0%	10	100%		0%	10	100%		0%		0%	10	100%	10	100%		0%
EIA	48	47	98%	1	2%	48	100%	0	0%	48	100%	0	0%	0	0%	48	100%	48	100%	0	0%
<i>Abbott AxSYM</i>	26	25	96%	1	4%	26	100%		0%	26	100%		0%		0%	26	100%	26	100%		0%
<i>Bio-Rad</i>	15	15	100%		0%	15	100%		0%	15	100%		0%		0%	15	100%	15	100%		0%
<i>Other</i>	7	7	100%		0%	7	100%		0%	7	100%		0%		0%	7	100%	7	100%		0%
Confirmation	125	124	99%	1	1%	125	100%	0	0%	125	100%	0	0%	0	0%	125	100%	125	100%	0	0%
<i>Abbott AxSYM</i>	13	12	92%	1	8%	13	100%		0%	13	100%		0%		0%	13	100%	13	100%		0%
<i>Bayer/Siemens</i>	39	39	100%		0%	39	100%		0%	39	100%		0%		0%	39	100%	39	100%		0%
<i>Diagnostic Products</i>	10	10	100%		0%	10	100%		0%	10	100%		0%		0%	10	100%	10	100%		0%
<i>Ortho</i>	29	29	100%		0%	29	100%		0%	29	100%		0%		0%	29	100%	29	100%		0%
<i>Other</i>	34	34	100%		0%	34	100%		0%	34	100%		0%		0%	34	100%	34	100%		0%
Other Methods	6	6	100%		0%	6	100%		0%	6	100%		0%		0%	6	100%	6	100%		0%
Analyte Total	316	314	99%	2	1%	316	100%	0	0%	316	100%	0	0%	0	0%	316	100%	313	99%	3	1%

Note: If you do not have enough volume to complete testing, please call before the replacement deadline. If you call too late to have additional sample volume shipped and your reported results are different than the consensus results, your result will be marked incorrect.

Hepatitis Be Antigen

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	66 R				67 N				68 R				69 N				70 N			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	26	0	0%	26	100%	26	100%	0	0%	0	0%	26	100%	26	100%	0	0%	26	100%	0	0%
<i>DiaSorin</i>	25		0%	25	100%	25	100%		0%		0%	25	100%	25	100%		0%	25	100%		0%
<i>Other</i>	1		0%	1	100%	1	100%		0%		0%	1	100%	1	100%		0%	1	100%		0%
Other Methods	4		0%	4	100%	4	100%		0%		0%	4	100%	4	100%		0%	4	100%		0%
Analyte Total	30	0	0%	30	100%	30	100%	0	0%	0	0%	30	100%	30	100%	0	0%	30	100%	0	0%

Hepatitis C Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	71 N				72 N				73 R				74 N				75 R			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Chemiluminescence	100	100	100%	0	0%	100	100%	0	0%	0	0%	100	100%	100	100%	0	0%	0	0%	100	100%
<i>Bayer</i>	54	54	100%		0%	54	100%		0%		0%	54	100%	54	100%		0%		0%	54	100%
<i>Ortho</i>	40	40	100%		0%	40	100%		0%		0%	40	100%	40	100%		0%		0%	40	100%
<i>Other</i>	6	6	100%		0%	6	100%		0%		0%	6	100%	6	100%		0%		0%	6	100%
EIA	62	62	100%	0	0%	62	100%	0	0%	0	0%	62	100%	62	100%	0	0%	0	0%	62	100%
<i>Abbott</i>	40	40	100%		0%	40	100%		0%		0%	40	100%	40	100%		0%		0%	40	100%
<i>Ortho</i>	22	22	100%		0%	22	100%		0%		0%	22	100%	22	100%		0%		0%	22	100%
Confirmation	26	26	100%	0	0%	26	100%	0	0%	0	0%	26	100%	26	100%	0	0%	0	0%	26	100%
<i>Chiron</i>	22	22	100%		0%	22	100%		0%		0%	22	100%	22	100%		0%		0%	22	100%
<i>Other</i>	4	4	100%		0%	4	100%		0%		0%	4	100%	4	100%		0%		0%	4	100%
Other Methods	8	8	100%		0%	8	100%		0%		0%	8	100%	8	100%		0%		0%	8	100%
Analyte Total	196	196	100%	0	0%	196	100%	0	0%	0	0%	196	100%	196	100%	0	0%	0	0%	196	100%

HIV Antibody

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method	No. Labs	31 R				32 N				33 N				34 N				35 R			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Chemiluminescence	39	0	0%	39	100%	39	100%	0	0%	39	100%	0	0%	39	100%	0	0%	0	0%	39	100%
<i>Bayer</i>	26		0%	26	100%	26	100%		0%	26	100%		0%	26	100%		0%		0%	26	100%
<i>Ortho</i>	13		0%	13	100%	13	100%		0%	13	100%		0%	13	100%		0%		0%	13	100%
EIA	89	0	0%	89	100%	89	100%	0	0%	89	100%	0	0%	89	100%	0	0%	0	0%	89	100%
<i>Abbott</i>	39		0%	39	100%	39	100%		0%	39	100%		0%	39	100%		0%		0%	39	100%
<i>Bio-Rad</i>	50		0%	50	100%	50	100%		0%	50	100%		0%	50	100%		0%		0%	50	100%
Rapid EIA	91	0	0%	91	100%	91	100%	0	0%	91	100%	0	0%	91	100%	0	0%	0	0%	91	100%
<i>Orasure</i>	91		0%	91	100%	91	100%		0%	91	100%		0%	91	100%		0%		0%	91	100%
Rapid Immunoassay	62	0	0%	62	100%	47	76%	0	0%	47	76%	0	0%	46	74%	1	2%	1	2%	61	98%
<i>Inverness</i>	18		0%	18	100%	18	100%		0%	18	100%		0%	18	100%		0%		0%	18	100%
<i>Medmira</i>	15		0%	15	100%	15	100%		0%	15	100%		0%	15	100%		0%		0%	15	100%
<i>Trinity</i> ¹	25		0%	25	100%	10	40%		0%	10	40%		0%	9	36%	1	4%	1	4%	24	96%
<i>Other</i>	4		0%	4	100%	4	100%		0%	4	100%		0%	4	100%		0%		0%	4	100%
Western Blot	42	0	0%	42	100%	42	100%	0	0%	42	100%	0	0%	42	100%	0	0%	0	0%	42	100%
<i>Bio-Rad</i>	40		0%	40	100%	40	100%		0%	40	100%		0%	40	100%		0%		0%	40	100%
<i>Others</i>	2		0%	2	100%	2	100%		0%	2	100%		0%	2	100%		0%		0%	2	100%
Other Methods	6		0%	6	100%	6	100%		0%	6	100%		0%	6	100%		0%		0%	6	100%
Analyte Total	329	0	0%	329	100%	314	95%	0	0%	314	95%	0	0%	313	95%	1	0%	1	0%	328	100%

¹ 15 laboratories reported an "invalid" test on negative samples #32,33, and 34. Due to control line issue.

HTLV Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	36 N				37 R				38 N				39 N				40 R			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	23	23	100%	0	0%	0	0%	23	100%	23	100%	0	0%	23	100%	0	0%	0	0%	22	96%
<i>Abbott</i> ¹	22	22	100%		0%		0%	22	100%	22	100%		0%	22	100%		0%		0%	21	95%
<i>Other</i>	1	1	100%		0%		0%	1	100%	1	100%		0%	1	100%		0%		0%	1	100%
Abbott Prism	13	13	100%		0%		0%	13	100%	13	100%		0%	12	92%	1	8%		0%	13	100%
Analyte Total	36	36	100%	0	0%	0	0%	36	100%	36	100%	0	0%	35	97%	1	3%	0	0%	35	97%

¹ One laboratory reported an equivocal on sample #40.

Infectious Mononucleosis

		Participant Results/ Sample Number																					
		R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method <i>Manufacturer</i>	No. Labs	26 N				27 N				28 R				29 R				30 N					
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%		
CICA	17	17	100%	0	0%	17	100%	0	0%	0	0%	17	100%	0	0%	17	100%	0	0%	17	100%	0	0%
<i>Others</i>	17	17	100%		0%	17	100%		0%		0%	17	100%		0%	17	100%		0%	17	100%		0%
Hemagglutination	44	43	98%	1	2%	43	98%	1	2%	0	0%	44	100%	0	0%	44	100%	41	93%	3	7%		
<i>Fisher</i>	16	15	94%	1	6%	15	94%	1	6%		0%	16	100%		0%	16	100%	14	88%	2	13%		
<i>Wampole/Zeus</i>	20	20	100%		0%	20	100%		0%		0%	20	100%		0%	20	100%	19	95%	1	5%		
<i>Others</i>	8	8	100%		0%	8	100%		0%		0%	8	100%		0%	8	100%	8	100%		0%		
Latex Agglutination	139	137	99%	1	1%	136	98%	3	2%	2	1%	137	99%	2	1%	137	99%	139	100%	0	0%		
<i>Fisher</i> ¹	44	43	98%		0%	43	98%	1	2%		0%	44	100%		0%	44	100%	44	100%		0%		
<i>Remel</i>	25	25	100%		0%	25	100%		0%		0%	25	100%		0%	25	100%	25	100%		0%		
<i>Wampole/Zeus</i>	51	51	100%		0%	50	98%	1	2%	1	2%	50	98%		0%	51	100%	51	100%		0%		
<i>Others</i>	19	18	95%	1	5%	18	95%	1	5%	1	5%	18	95%	2	11%	17	89%	19	100%		0%		
Solid Phase IA	61	61	100%	0	0%	61	100%	0	0%	0	0%	61	100%	0	0%	61	100%	61	100%	0	0%		
<i>Inverness</i>	30	30	100%		0%	30	100%		0%		0%	30	100%		0%	30	100%	30	100%		0%		
<i>Seradyn</i>	25	25	100%		0%	25	100%		0%		0%	25	100%		0%	25	100%	25	100%		0%		
<i>Others</i>	6	6	100%		0%	6	100%		0%		0%	6	100%		0%	6	100%	6	100%		0%		
Other Methods	4	4	100%	0	0%	4	100%		0%		0%	4	100%		0%	4	100%	4	100%		0%		
Analyte Total	265	262	99%	2	1%	261	98%	4	2%	2	1%	263	99%	2	1%	263	99%	262	99%	3	1%		

¹ One laboratory reported "not applicable" on sample #26.

Lyme Disease Antibody

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	41 R				42 N				43 N				44 N				45 R			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	57	0	0%	57	100%	57	100%	0	0%	57	100%	0	0%	57	100%	0	0%	0	0%	57	100%
<i>Immunitics</i>	16		0%	16	100%	16	100%		0%	16	100%		0%	16	100%		0%		0%	16	100%
<i>Wampole /Zeus</i>	29		0%	29	100%	29	100%		0%	29	100%		0%	29	100%		0%		0%	29	100%
<i>Others</i>	12		0%	12	100%	12	100%		0%	12	100%		0%	12	100%		0%		0%	12	100%
ELFA	29	0	0%	29	100%	29	100%	0	0%	29	100%	0	0%	29	100%	0	0%	0	0%	29	100%
<i>bioMérieux</i>	29		0%	29	100%	29	100%		0%	29	100%		0%	29	100%		0%		0%	29	100%
Other Methods	7		0%	7	100%	7	100%		0%	7	100%		0%	7	100%		0%		0%	7	100%
Analyte Total	93	0	0%	93	100%	93	100%	0	0%	93	100%	0	0%	93	100%	0	0%	0	0%	93	100%

Lyme Western Blot IgG

		Participant Results/ Sample Number																													
		R = Reactive/ Positive; N = Non-Reactive/ Negative; E = Equivocal/ Indeterminate																													
Method	No. Labs	41 R						42 N						43 N						44 N						45 R					
		N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%
Western Blot - IgG	27	0	0%	0	0%	27	100%	27	100%	0	0%	0	0%	27	100%	0	0%	0	0%	27	100%	0	0%	0	0%	1	4%	0	0%	26	96%
<i>MarDx</i>	24		0%		0%	24	100%	24	100%		0%		0%	24	100%		0%		0%	24	100%		0%		0%	1	4%		0%	23	96%
<i>Other</i>	3		0%		0%	3	100%	3	100%		0%		0%	3	100%		0%		0%	3	100%		0%		0%		0%		0%	3	100%
Other Methods	3		0%		0%	3	100%	3	100%		0%		0%	3	100%		0%		0%	3	100%		0%		0%		0%		0%	3	100%
Analyte Total	30	0	0%	0	0%	30	100%	30	100%	0	0%	0	0%	30	100%	0	0%	0	0%	30	100%	0	0%	0	0%	1	3%	0	0%	29	97%

Lyme Western Blot IgM

		Participant Results/ Sample Number																													
		R = Reactive/ Positive; N = Non-Reactive/ Negative; E = Equivocal/ Indeterminate																													
Method	No. Labs	41 R*						42 N						43 N						44 N						45 N*					
		N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%	N	%	E	%	R	%
Western Blot - IgM	27	15	56%	2	7%	10	37%	27	100%	0	0%	0	0%	27	100%	0	0%	0	0%	27	100%	0	0%	0	0%	20	74%	3	11%	4	15%
<i>MarDx</i>	24	14	58%	2	8%	8	33%	24	100%		0%		0%	24	100%		0%		0%	24	100%		0%		0%	18	75%	3	13%	3	13%
<i>Other</i>	3	1	33%		0%	2	67%	3	100%		0%		0%	3	100%		0%		0%	3	100%		0%		0%	2	67%		0%	1	33%
Other Methods	2		0%		0%	2	100%	2	100%		0%		0%	2	100%		0%		0%	2	100%		0%		0%	1	50%		0%	1	50%
Analyte Total	29	15	52%	2	7%	12	41%	29	100%	0	0%	0	0%	29	100%	0	0%	0	0%	29	100%	0	0%	0	0%	21	72%	3	10%	5	17%

* Samples #41 and 45 are not authenticated, the expected result is positive. When a consensus agreement cannot be reached among participants, by regulation requirements, the sample cannot be graded (scored) and all participating laboratories get credit for this sample.

Rheumatoid Factor

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	26 N				27 N				28 R				29 R				30 N			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Latex Agglutination	96	94	98%	2	2%	95	99%	1	1%	3	3%	93	97%	1	1%	95	99%	95	99%	1	1%
<i>Becton Dickinson</i>	14	14	100%		0%	14	100%		0%		0%	14	100%		0%	14	100%	14	100%		0%
<i>Fisher</i>	39	39	100%		0%	39	100%		0%	2	5%	37	95%		0%	39	100%	38	97%	1	3%
<i>Wampole/Zeus</i>	11	10	91%	1	9%	11	100%		0%		0%	11	100%		0%	11	100%	11	100%		0%
<i>Others</i>	32	31	97%	1	3%	31	97%	1	3%	1	3%	31	97%	1	3%	31	97%	32	100%		0%
Nephelometry	14	14	100%	0	0%	14	100%	0	0%	0	0%	14	100%	0	0%	14	100%	14	100%	0	0%
<i>Others</i>	14	14	100%		0%	14	100%		0%		0%	14	100%		0%	14	100%	14	100%		0%
Turbidimetry	17	17	100%	0	0%	17	100%	0	0%	0	0%	17	100%	0	0%	17	100%	17	100%	0	0%
<i>Others</i>	17	17	100%		0%	17	100%		0%		0%	17	100%		0%	17	100%	17	100%		0%
Other Methods	11	10	91%	1	9%	10	91%	1	9%		0%	11	100%		0%	11	100%	10	91%	1	9%
Analyte Total	138	135	98%	3	2%	136	99%	2	1%	3	2%	135	98%	1	1%	137	99%	136	99%	2	1%

Rheumatoid Factor Quantitative Latex Agglutination Procedure

The number of laboratories that reported titers is listed for positive test samples 28 and 29. The dilution schemes laboratories used are represented by the letter A and B, testing systems with 10 or more laboratories reporting titers are listed in this table.

<i>Manufacturer</i>	No. Labs	A B	Sample 28 Titer								Sample 29 Titer							
			10 1	20 2	40 4	80 8	160 16	320 32	640 64	1280 128	10 1	20 2	40 4	80 8	160 16	320 32	640 64	1280 128
Total	73	A B			5 2	11 1	11 10	11 8	2 2			2 2	9 9	12 9	16 9	3 2		1
<i>Becton Dickenson</i>	11	A B			1	7	2	1			1	6	2	2				
<i>Fisher</i>	31	A B			1 1	2	6 6	7 7				2 1	6 6	8 6		1		

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers.

Rheumatoid Factor Quantitative

Nephelometry & Turbidimetry Procedures

Results are summarized for positive test samples 28 and 29. The Mean values \pm S.D. are given where 10 or more laboratories reported quantitative results. Outlier values are omitted.

Method <i>Manufacturer</i>	No. Labs	Unit	Sample 28	Sample 29
Nephelometry Total	25	IU/ml	220 \pm 41	126 \pm 28
<i>Beckman Coulter IMMAGE</i>	11	IU/ml	263 \pm 10	156 \pm 7.2
<i>Dade Behring Nephelometer</i>	14	IU/ml	188 \pm 20	105 \pm 15
Turbidimetry Total	67	IU/ml	163 \pm 76	110 \pm 45
<i>Beckman Unicel DxC</i>	10	IU/ml	294 \pm 22	174 \pm 7.7
<i>Roche Diag. Cobas</i>	20	IU/ml	113 \pm 4.4	81 \pm 3.5
<i>Roche Hitachi Modular (all models)</i>	11	IU/ml	111 \pm 1.7	79 \pm 2.4

Rubella Antibody

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method	No. Labs	11 R				12 R				13 N				14 N				15 N			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
Chemiluminescence	55	0	0%	55	100%	2	4%	48	87%	55	100%	0	0%	55	100%	0	0%	55	100%	0	0%
<i>Bayer</i>	25		0%	25	100%		0%	25	100%	25	100%		0%	25	100%		0%	25	100%		0%
<i>Beckman</i> ¹	13		0%	13	100%	2	15%	6	46%	13	100%		0%	13	100%		0%	13	100%		0%
<i>Diagnostic Products</i>	12		0%	12	100%		0%	12	100%	12	100%		0%	12	100%		0%	12	100%		0%
<i>Others</i>	5		0%	5	100%		0%	5	100%	5	100%		0%	5	100%		0%	5	100%		0%
EIA	53	0	0%	53	100%	0	0%	51	96%	53	100%	0	0%	53	100%	0	0%	53	100%	0	0%
<i>Abbott</i> ²	19		0%	19	100%		0%	18	95%	19	100%		0%	19	100%		0%	19	100%		0%
<i>Wampole/Zeus</i>	16		0%	16	100%		0%	16	100%	16	100%		0%	16	100%		0%	16	100%		0%
<i>Others</i> ³	18		0%	18	100%		0%	17	94%	18	100%		0%	18	100%		0%	18	100%		0%
Latex Agglutination	31	0	0%	31	100%	1	3%	30	97%	31	100%	0	0%	31	100%	0	0%	30	97%	1	3%
<i>Fisher</i>	19		0%	19	100%	1	5%	18	95%	19	100%		0%	19	100%		0%	18	95%	1	5%
<i>Others</i>	12		0%	12	100%		0%	12	100%	12	100%		0%	12	100%		0%	12	100%		0%
Other Methods	14		0%	14	100%		0%	14	100%	14	100%		0%	14	100%		0%	14	100%		0%
Analyte Total	153	0	0%	153	100%	3	2%	143	93%	153	100%	0	0%	153	100%	0	0%	152	99%	1	1%

¹ Five laboratories reported an equivocal result on sample #12

² One laboratory reported an equivocal result on sample #12

³ One laboratory reported an equivocal result on sample #12

Rubella Antibody Quantitative

Results are summarized for positive test samples 11 and 12. The Mean values \pm S.D. are given where 10 or more laboratories reported quantitative results. Outlier values are omitted.

Method	No.			
<i>Manufacturer</i>	Labs	Unit	Sample 11	Sample 12
Chemiluminescence Total	26	IU/ml	51 \pm 7.2	15 \pm 2.1
<i>Diagnostic Products</i>	16	IU/ml	49 \pm 4.2	15 \pm 1.3
<i>Bayer(not included in total) *</i>	25	IU/ml	496 \pm 19	93 \pm 11
EIA Total	17	IU/ml	45 \pm 11	20 \pm 5.6

* Results from Bayer Advia systems consistently run higher than all others.

Rubella IgM Specific

Participant Results/ Sample Number																					
R = Reactive/ Positive; N = Non-Reactive/ Negative																					
Method	No. Labs	56 N				57 N				58 N				59 R				60 N			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
<i>Manufacturer</i>																					
EIA	7	7	100%		0%	7	100%		0%	7	100%		0%		0%	7	100%	6	86%	1	14%
Chemiluminescence	9	9	100%		0%	9	100%		0%	9	100%		0%		0%	9	100%	9	100%		0%
Analyte Total	16	16	100%	0	0%	16	100%	0	0%	16	100%	0	0%	0	0%	16	100%	15	94%	1	6%

Syphilis - Reagin Antibody

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method	No. Labs	1 N				2 R				3 N				4 R				5 N			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
RPR	245	245	100%	0	0%	0	0%	245	100%	245	100%	0	0%	0	0%	245	100%	245	100%	0	0%
<i>ASI</i>	38	38	100%		0%		0%	38	100%	38	100%		0%		0%	38	100%	38	100%		0%
<i>Becton Dickenson</i>	117	117	100%		0%		0%	117	100%	117	100%		0%		0%	117	100%	117	100%		0%
<i>Fisher</i>	42	42	100%		0%		0%	42	100%	42	100%		0%		0%	42	100%	42	100%		0%
<i>True Medix</i>	16	16	100%		0%		0%	16	100%	16	100%		0%		0%	16	100%	16	100%		0%
<i>Wampole/Zeus</i>	20	20	100%		0%		0%	20	100%	20	100%		0%		0%	20	100%	20	100%		0%
<i>Others</i>	12	12	100%		0%		0%	12	100%	12	100%		0%		0%	12	100%	12	100%		0%
Analyte Total	245	245	100%	0	0%	0	0%	245	100%	245	100%	0	0%	0	0%	245	100%	245	100%	0	0%

Syphilis - Reagin Antibody

RPR Procedures

The number of laboratories that reported titers is listed for positive test samples 2 and 4 for the RPR procedure. Only testing systems with 10 or more laboratories reporting titers are listed in this table.

Method Manufacturer	No. Labs	Sample 2 Titer						Sample 4 Titer					
		1	2	4	8	16	32	1	2	4	8	16	32
<i>Total</i>	222	63	134	24		1		3	85	113	19	2	
<i>ASI</i>	26	12	11	3				1	15	8	2		
<i>Becton Dickenson</i>	110	22	75	13					24	72	13	1	
<i>Fisher</i>	41	15	23	3				1	22	16	2		
<i>True Medix</i>	16	5	7	3		1		1	8	4	2	1	
<i>Wampole/ Zeus</i>	20	6	13	1					10	10			

Note: The number of labs reporting specific titers may not add up to the total number of labs for that system because some labs are not reporting endpoint titers, or they reported the sample as nonreactive.

Syphilis - Treponemal Antibody

		Participant Results/ Sample Number																			
		R = Reactive/ Positive; N = Non-Reactive/ Negative																			
Method	No. Labs	1 N				2 R				3 N				4 R				5 N			
		N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%	N	%	R	%
EIA	18	18	100%	0	0%	0	0%	18	100%	18	100%	0	0%	0	0%	18	100%	18	100%	0	0%
<i>Others</i>	18	18	100%		0%		0%	18	100%	18	100%		0%		0%	18	100%	18	100%		0%
Gel. Part. Agglut.	30	30	100%	0	0%	0	0%	30	100%	30	100%	0	0%	0	0%	30	100%	30	100%	0	0%
<i>Fujirebio</i>	30	30	100%		0%		0%	30	100%	30	100%		0%		0%	30	100%	30	100%		0%
FTA	23	23	100%	0	0%	0	0%	23	100%	22	96%	1	4%	0	0%	23	100%	23	100%	0	0%
<i>Wampole/Zeus</i>	18	18	100%		0%		0%	18	100%	18	100%		0%		0%	18	100%	18	100%		0%
<i>Others</i>	5	5	100%		0%		0%	5	100%	4	80%	1	20%		0%	5	100%	5	100%		0%
Other Methods	13	5	38%	8	62%		0%	13	100%	5	38%	8	62%		0%	13	100%	5	38%	8	62%
Analyte Total	84	76	90%	8	10%	0	0%	84	100%	75	89%	9	11%	0	0%	84	100%	76	90%	8	10%

AAT, C'3, and C'4

The Mean mg/dl \pm S.D. is given where 10 or more laboratories reported values. Outlier values are omitted.

Analytes	No.	Sample NO.					
		Labs	76	77	78	79	80
Alpha-1-Antitrypsin							
Nephelometry/ <i>Behring Nephelometer</i>	18		95 \pm 5.9	36 \pm 1.4	16 \pm 0.7	6.1 \pm 0.1	345 \pm 25
Nephelometry/ <i>Total</i>	25		92 \pm 7.4	36 \pm 1.6	18 \pm 2.3	6.1 \pm 0.1	330 \pm 33
All Method/ <i>Total</i>	36		91 \pm 8.8	36 \pm 2.2	18 \pm 1.8	< 10	320 \pm 32
Complement C'3			76	77	78	79	80
Nephelometry/ <i>Beckman Coulter IMMAGE</i>	14		75 \pm 3.4	34 \pm 2.2	23 \pm 1.4	29 \pm 2.0	276 \pm 15
Nephelometry/ <i>Behring Nephelometer</i>	11		83 \pm 5.2	36 \pm 2.1	24 \pm 0.9	30 \pm 2.1	285 \pm 19
Turbidimetry/ <i>Roche Cobas Integra</i>	15		73 \pm 2.85	35 \pm 2.6	22 \pm 0.8	29 \pm 1.2	285 \pm 5.1
Turbidimetry/ <i>Roche/Hitachi Modular</i>	12		71 \pm 1.8	36 \pm 1.1	24 \pm 0.9	31 \pm 0.7	285 \pm 6.2
Nephelometry/ <i>Total</i>	29		79 \pm 5.4	35 \pm 1.6	24 \pm 1.1	30 \pm 2.1	279 \pm 17
Turbidimetry/ <i>Total</i>	65		72 \pm 3.9	35 \pm 2.7	23 \pm 6.1	30 \pm 2.3	283 \pm 14
All Method/ <i>Total</i>	100		74 \pm 5.2	35 \pm 2.6	24 \pm 1.9	30 \pm 2.2	282 \pm 15
Complement C'4			76	77	78	79	80
Nephelometry/ <i>Beckman Coulter IMMAGE</i>	14		14 \pm 0.5	8.0 \pm 0.6	4.4 \pm 0.4	3.5 \pm 0.3	27 \pm 1.1
Nephelometry/ <i>Behring Nephelometer</i>	11		13 \pm 0.7	7.6 \pm 0.5	4.0 \pm 0.1	3.3 \pm 0.4	27 \pm 1.9
Turbidimetry/ <i>Roche Cobas Integra</i>	16		12 \pm 0.9	6.8 \pm 0.8	3.5 \pm 0.5	3.2 \pm 0.5	28 \pm 0.9
Turbidimetry/ <i>Roche/Hitachi Modular</i>	10		11 \pm 0.2	6.1 \pm 0.1	3.2 \pm 0.4	3.0 \pm 0.1	25 \pm 0.4
Nephelometry/ <i>Total</i>	29		14 \pm 0.8	7.9 \pm 0.6	4.2 \pm 0.4	3.5 \pm 0.3	27 \pm 1.7
Turbidimetry/ <i>Total</i>	63		12 \pm 1.2	6.8 \pm 0.8	3.5 \pm 0.6	3.2 \pm 0.5	27 \pm 3.3
All Method/ <i>Total</i>	98		13 \pm 1.3	7.2 \pm 0.9	3.8 \pm 0.6	3.2 \pm 0.5	27 \pm 2.8

IgA, IgE, IgG, and IgM

The Mean mg/dl (IU/ml for IgE) \pm S.D. is given for IgA, IgE, and IgM and Mean mg/dl \pm 25% is given for IgG where 10 or more laboratories reported values. Outlier values are omitted.

Analytes (Unit)	No.	Sample NO.				
		Labs	81	82	83	84
Immunoglobulin A (mg/dl)						
Nephelometry/Beckman Coulter IMMAGE	17	76 \pm 3.9	109 \pm 6.1	226 \pm 11	181 \pm 10	50 \pm 3.2
Nephelometry/Behring Nephelometer	15	86 \pm 3.8	120 \pm 7.9	267 \pm 16	209 \pm 8.4	55 \pm 4.2
Turbidimetry/ Beckman Coulter Synchron	11	85 \pm 4.3	118 \pm 1.8	244 \pm 7.0	183 \pm 7.4	62 \pm 4.7
Turbidimetry/ Roche Cobas Integra	16	76 \pm 2.4	114 \pm 5.0	240 \pm 6.1	200 \pm 5.9	50 \pm 2.3
Turbidimetry/ Roche/Hitachi Modular	10	72 \pm 8.2	104 \pm 3.6	222 \pm 3.9	184 \pm 4.8	51 \pm 1.1
Nephelometry/ Total	39	81 \pm 6.5	114 \pm 8.4	247 \pm 24	195 \pm 18	52 \pm 4.7
Turbidimetry/ Total	63	80 \pm 7.0	114 \pm 7.1	236 \pm 13	190 \pm 12	54 \pm 6.0
All Method/ Total	108	80 \pm 7.0	115 \pm 7.8	241 \pm 18	193 \pm 15	53 \pm 5.3
Immunoglobulin E (IU/ml)						
Chemiluminescence/Bayer Advia	11	51 \pm 3.0	647 \pm 50	206 \pm 9.5	33 \pm 2.7	258 \pm 16
Chemiluminescence/Diag. Prod. Co.	26	57 \pm 6.0	728 \pm 95	221 \pm 14	40 \pm 3.2	277 \pm 31
FEIA/ Pharmacia Immunocap	13	50 \pm 3.8	702 \pm 77	195 \pm 17	35 \pm 2.4	245 \pm 23
Chemiluminescence/ Total	46	55 \pm 5.7	702 \pm 87	217 \pm 13	38 \pm 4.6	269 \pm 24
FEIA/ Total	18	51 \pm 3.6	708 \pm 71	198 \pm 16	35 \pm 2.2	247 \pm 21
All Method/ Total	78	53 \pm 5.5	707 \pm 90	206 \pm 21	37 \pm 4.5	259 \pm 28
Immunoglobulin G (mg/dl)						
Nephelometry/Beckman Coulter IMMAGE	16	407 \pm 16	847 \pm 36	1282 \pm 55	966 \pm 34	323 \pm 12
Nephelometry/Behring Nephelometer	15	425 \pm 25	866 \pm 34	1363 \pm 74	1002 \pm 28	328 \pm 16
Turbidimetry/ Beckman Coulter Synchron	10	394 \pm 16	805 \pm 18	1287 \pm 10	926 \pm 14	305 \pm 15
Turbidimetry/Roche Cobas Integra	16	398 \pm 11	826 \pm 26	1282 \pm 36	954 \pm 20	307 \pm 13
Turbidimetry/Roche/Hitachi Modular	10	361 \pm 17	777 \pm 14	1185 \pm 29	898 \pm 23	294 \pm 13
Nephelometry/ Total	38	416 \pm 21	853 \pm 36	1315 \pm 73	987 \pm 39	323 \pm 15
Turbidimetry/ Total	60	391 \pm 21	807 \pm 32	1257 \pm 53	925 \pm 38	307 \pm 14
All Method/ Total	105	401 \pm 24	825 \pm 41	1278 \pm 63	947 \pm 52	313 \pm 17
Immunoglobulin M (mg/dl)						
Nephelometry/Beckman Coulter IMMAGE	16	37 \pm 2.2	41 \pm 2.1	125 \pm 6.1	31 \pm 1.6	69 \pm 1.8
Nephelometry/Behring Nephelometer	16	38 \pm 2.7	37 \pm 1.6	140 \pm 6.1	30 \pm 2.5	67 \pm 1.8
Turbidimetry/ Beckman Coulter Synchron	10	39 \pm 1.4	43 \pm 1.2	119 \pm 2.1	31 \pm 1.2	64 \pm 2.1
Turbidimetry/Roche Cobas Integra	16	36 \pm 3.3	37 \pm 3.9	120 \pm 3.8	27 \pm 3.1	66 \pm 2.7
Turbidimetry/Roche/Hitachi Modular	10	36 \pm 4.1	38 \pm 4.7	118 \pm 4.2	29 \pm 2.9	68 \pm 2.7
Nephelometry/ Total	37	38 \pm 2.5	39 \pm 3.7	133 \pm 10	31 \pm 2.3	67 \pm 4.0
Turbidimetry/ Total	60	37 \pm 3.7	41 \pm 5.2	120 \pm 6.3	30 \pm 4.4	67 \pm 4.6
All Method/ Total	104	38 \pm 3.2	40 \pm 4.9	127 \pm 12	30 \pm 2.9	67 \pm 3.7

Acceptable Response (May 13, 2009 PT Event)
Quantitative Tests Results (Acceptable Range) - For groups of 10 labs or more.

Analytes	Sample NO.				
Method/ Manufacture					
Alpha-1-Antitrypsin	76	77	78	79	80
Nephelometry/Dade Behring Neph.	76 - 113	32 - 41	14 - 19	5 - 7	271 - 419
Nephelometry/ Total	69 - 115	31 - 42	10 - 25	5 - 7	231 - 429
All Method/ Total	64 - 118	28 - 43	12 - 23	< 10	223 - 418
Complement C'3	76	77	78	79	80
Nephelometry/Beckman Coulter Image	65 - 86	27 - 41	18 - 28	23 - 36	230 - 322
Nephelometry/Dade Behring Neph.	67 - 99	29 - 43	21 - 27	24 - 37	229 - 342
Turbidimetry/Roche Cobas Integra	64 - 82	26 - 43	20 - 25	25 - 33	269 - 301
Turbidimetry/Roche HitachiModular	65 - 77	32 - 39	21 - 28	28 - 33	266 - 304
Nephelometry/ Total	62 - 96	30 - 40	20 - 28	23 - 36	228 - 330
Turbidimetry/ Total	60 - 84	27 - 44	15 - 31	22 - 37	241 - 324
All Method/ Total	58 - 90	27 - 44	18 - 30	23 - 37	237 - 327
Complement C'4	76	77	78	79	80
Nephelometry/Beckman Coulter Image	12 - 16	6 - 10	3 - 6	2 - 5	23 - 31
Nephelometry/Dade Behring Neph.	10 - 16	6 - 10	3 - 5	2 - 5	21 - 33
Turbidimetry/Roche Cobas Integra	9 - 15	4 - 10	2 - 6	1 - 5	25 - 31
Turbidimetry/Roche HitachiModular	10 - 12	5 - 7	2 - 5	2 - 4	24 - 27
Nephelometry/ Total	11 - 17	6 - 10	3 - 6	2 - 5	22 - 33
Turbidimetry/ Total	8 - 16	4 - 10	1 - 6	1 - 5	17 - 37
All Method/ Total	8 - 17	5 - 10	1 - 6	1 - 5	18 - 36
Immunoglobulin A	81	82	83	84	85
Nephelometry/Beckman Coulter Image	63 - 88	90 - 128	191 - 260	150 - 212	40 - 60
Nephelometry/Dade Behring Neph.	74 - 97	95 - 144	218 - 316	183 - 234	42 - 68
Turbidimetry/ Beckman Coulter Synchron.	72 - 98	113 - 124	223 - 266	160 - 205	47 - 76
Turbidimetry/ Roche Cobas Integra	68 - 83	99 - 130	221 - 258	181 - 218	43 - 58
Turbidimetry/ Roche/Hitachi Modular	47 - 97	93 - 115	210 - 234	169 - 199	47 - 55
Nephelometry/ Total	60 - 101	89 - 140	176 - 319	142 - 248	37 - 66
Turbidimetry/ Total	58 - 101	92 - 136	197 - 275	154 - 226	35 - 72
All Method/ Total	59 - 102	91 - 138	187 - 294	148 - 237	36 - 69
Immunoglobulin E	81	82	83	84	85
Chemiluminescence/Bayer Advia	42 - 61	497 - 797	177 - 235	25 - 42	210 - 305
Chemiluminescence/Diag.Prod. Immulite	38 - 76	441 - 1014	180 - 262	30 - 50	183 - 372
FEIA/Pharmacia Immuncap	38 - 62	470 - 935	144 - 246	27 - 43	176 - 315
Chemiluminescence/ Total	37 - 72	439 - 965	177 - 256	24 - 52	196 - 342
FEIA/ Total	39 - 62	494 - 922	149 - 247	28 - 42	184 - 310
All Method/ Total	36 - 70	437 - 977	142 - 270	23 - 51	175 - 344
Immunoglobulin G	81	82	83	84	85
Nephelometry/Beckman Coulter Image	305 - 509	635 - 1059	961 - 1603	724 - 1207	241 - 404
Nephelometry/Dade Behring Neph.	319 - 532	649 - 1083	1022 - 1704	751 - 1253	246 - 411
Turbidimetry/ Beckman Coulter Synchron.	295 - 493	603 - 1007	965 - 1609	694 - 1158	228 - 381
Turbidimetry/Roche Cobas Integra	297 - 497	619 - 1033	961 - 1603	715 - 1193	230 - 385
Turbidimetry/ Roche/Hitachi Modular	270 - 451	582 - 972	889 - 1482	673 - 1124	220 - 368
Nephelometry/ Total	312 - 521	639 - 1067	986 - 1645	740 - 1234	242 - 405
Turbidimetry/ Total	293 - 489	604 - 1009	942 - 1571	693 - 1157	229 - 384
All Method/ Total	300 - 502	618 - 1032	958 - 1598	710 - 1184	235 - 392
Immunoglobulin M	81	82	83	84	85
Nephelometry/Beckman Coulter Image	30 - 45	34 - 48	107 - 144	26 - 36	63 - 75
Nephelometry/Dade Behring Neph.	30 - 47	31 - 42	122 - 159	22 - 38	61 - 73
Turbidimetry/ Beckman Coulter Synchron.	34 - 43	39 - 48	112 - 125	27 - 35	57 - 71
Turbidimetry/Roche Cobas Integra	26 - 47	25 - 49	108 - 132	17 - 37	58 - 75
Turbidimetry/ Roche/Hitachi Modular	24 - 49	23 - 53	105 - 131	20 - 38	59 - 76
Nephelometry/ Total	30 - 46	27 - 50	103 - 164	24 - 38	55 - 80
Turbidimetry/ Total	26 - 49	25 - 57	100 - 139	16 - 44	53 - 81
All Method/ Total	28 - 48	25 - 55	91 - 163	21 - 39	55 - 79

**Acceptable Response (May 13, 2009 PT Event)
Qualitative / Quantitative Tests Results**

Analytes	Sample NO.				
	1	2	3	4	5
Syphilis - Reagin	N	R	N	R	N
<i>RPR Titer</i>	< 1	1 - 4	< 1	2 - 8	< 1
Syphilis - Treponemal	N	R	N	R	N
	6	7	8	9	10
HBcAb	N	N	N	R	R
HBsAg	N	N	N	R	N
HBsAg Confirmation	N	N	N	R	N
	11	12	13	14	15
CMV	R	R	N	N	N
Rubella Ab	R	R	N	N	N
Rubella Ab Quantitative <i>EIA IU/ml</i>	11 - 79	8 - 37	< 8	< 8	< 8
<i>Chemiluminescent IU/ml</i>	29 - 73	8 - 22	< 10	< 10	< 10
<i>Advia Centaur IU/ml</i>	440 - 552	58 - 128	< 5	< 5	< 5
	16	17	18	19	20
ASO	N	N	R	N	R
ASO Quantitative <i>Latex IU/ml</i>	< 200	< 200	200 - 1600	< 200	200 - 1600
<i>Nephelometry IU/ml</i>	< 100	< 100	147 - 902	< 100	127 - 961
<i>Turbidimetry IU/ml</i>	< 100	< 100	281 - 1045	< 100	283 - 1052
	26	27	28	29	30
Infectious Mono.	N	N	R	R	N
Rheumatoid Factor	N	N	R	R	N
RF Quantitative					
<i>Latex (Dilution Scheme A)</i>	< 20	< 20	40 - 640	20 - 640	< 20
<i>Latex (Dilution Scheme B)</i>	< 2	< 2	4 - 64	2 - 64	< 2
<i>Nephelometry IU/ml</i>	< 20	< 20	96 - 342	40 - 212	< 20
<i>Turbidimetry IU/ml</i>	< 20	< 20	99 - 362	70 - 197	< 20
	31	32	33	34	35
HIV Ab Screening/Confirmation	R	N	N	N	R
	36	37	38	39	40
HTLV 1 Ab	N	R	N	N	R
	41	42	43	44	45
LYME Disease Ab	R	N	N	N	R
LYME Disease Ab WB IgG	R	N	N	N	R
LYME Disease Ab WB IgM	R *	N	N	N	N *
	46	47	48	49	50
ANA	R	R	N	N	N
<i>IFA Titer</i>	160 - 2560	40 - 1280	< 40	< 40	< 40
	56	57	58	59	60
Rubella IgM	N	N	N	R	N
	66	67	68	69	70
Hepatitis Be Ag	R	N	R	N	N
	71	72	73	74	75
Hepatitis C Ab	N	N	R	N	R
Hepatitis C Ab Confirmation	N	N	R	N	R

Note: R = Reactive/ Positive; I = Indeterminate; N = Non-Reactive/ Negative

* This test sample was not authenticated, because a consensus of 80% agreement was not reached. All participating laboratories received credit for this sample.