

Microscopy Results Breakdown March 2008

		SL 0811	SL 0812	SL 0813	SL 0814	SL 0815
Smears Only (n=42)	Negative	2		40	40	
	Positive	40	42	2	2	42
Restricted (n=37)	Negative	2		35	35	
	Positive	35	37	2	2	37
Restricted /-S (n=1)	Negative				1	
	Positive	1	1	1		1
General (n=19)	Negative			19	19	
	Positive	19	19			19
General /-S (n=21)	Negative			21	21	
	Positive	21	21			21

Highlighted areas are false-positive results.

Highlighted areas are false-negative results.

SL 0811: 116/120 labs reported smear positive; 97%

SL 0812: 120/120 labs reported smear positive; 100%

SL 0813: 115/120 labs reported smear negative; 96%

SL 0814: 116/120 labs reported smear negative; 97%

SL 0815: 120/120 labs reported smear positive; 100%

Please note that four NYS permitted laboratories did not participate in this event and are not included in this breakdown.