


Staining Methodology Breakdown March 2008



Of 13 Total errors, laboratories using Carbol Fuchsin reported 8, 5 of which were false positive, and 3 were false negative.

The 5 remaining errors were reported by laboratories using Fluorochrome methods, 4 of which were false positive and 1 false negative.