



**Department
of Health**

**Wadsworth
Center**

TRACE ELEMENTS IN WHOLE BLOOD

Proficiency Test Report

Event #2, 2015

July 2nd, 2015



ANDREW M. CUOMO
Governor

Department of Health

HOWARD A. ZUCKER, M.D., J.D.
Commissioner

SALLY DRESLIN, M.S., R.N.
Executive Deputy Commissioner

July 2, 2015

Trace Elements in Whole Blood Event #2, 2015

Dear Laboratory Director:

Results from the second proficiency test (PT) event in 2015 for Trace Elements in Whole Blood have been tabulated and summarized. Target values for Arsenic, Cadmium, Mercury and Lead in whole blood have been established along with acceptable ranges. Results are graded using element-specific criteria as indicated in each narrative section. A laboratory with an unacceptable significant analytical bias relative to the target value will be expected to investigate the source of the error. A confidential three-digit code number assigned by the PT program identifies participant laboratories. The data for blood lead were previously reported in the Blood Lead PT Report issued July 2nd, 2015, and are reproduced here for completeness.

PT Materials

Test materials for the second event were prepared from caprine (goat) whole blood obtained from animals dosed with lead acetate to create physiologically bound lead (Pb). A total of five blood pools were supplemented with arsenic (as inorganic As^{3+}), cadmium (as Cd^{2+}) and mercury (as inorganic Hg^{2+}). In addition to As, Cd, Pb and Hg, blood pools were supplemented with the trace elements manganese (Mn), thallium (Tl), tin (Sn), titanium (Ti), nickel (Ni), cobalt (Co), chromium (Cr), silver (Ag), tungsten (W) and vanadium (V).

Additional Elements to Become Graded for Performance Assessment

PT results for select trace elements, including Co and Cr, are graded as part of this PT event, although the data are used for "Educational Purposes" only, to inform laboratory participants of where improved practices may be necessary. Laboratories that test and report these, and other, trace elements on patient specimens should continue to report results obtained for whole blood PT samples.

The next PT event for trace elements in whole blood is scheduled to be mailed Wednesday, September 16th, 2015. Please inform our laboratory staff at (518) 474-7161 if the test materials have not arrived within five days of the scheduled mail out date. **The deadline for reporting results is Wednesday, October 7th, 2015.**

Thank you for your participation.

Yours sincerely,

Patrick J. Parsons, Ph.D.
Chief, Laboratory of Inorganic and Nuclear Chemistry
Deputy Director, Division of Environmental Health

Mary Frances Verostek, Ph.D.
Assistant Section Head
PT Program for Blood Lead/Trace Elements
Biggs Laboratory - Wadsworth Center
New York State Department of Health
PO Box 509
Albany NY 12201-0509
www.wadsworth.org/testing/lead/index.htm

New York State Department of Health
Event #2, 2015

Whole Blood Arsenic

Test materials for arsenic were prepared from caprine (goat) whole blood preserved with K₂EDTA anticoagulant. A total of five pools were supplemented with arsenic as inorganic As³⁺.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for whole blood arsenic range from 6.1 µg/L (0.08 µmol/L) to 50.8 µg/L (0.68 µmol/L).

Acceptable range: The acceptable range for arsenic is set at ±6 µg/L or ±20%, whichever is greater. Thus, it is fixed at ±6 µg/L for concentrations below 30 µg/L.

Discussion: Based upon the above criteria, 95.8% of test results reported were judged as satisfactory, with none of the 19 laboratories reporting 2 or more of the 5 results outside the acceptable ranges.

New York State Department of Health
Blood Arsenic Test Results, 2015 Event #2
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ whole blood)

BE15-06 BE15-07 BE15-08 BE15-09 BE15-10

| | | | | | |
|-------------------------------|-------------|------------|-------------|-------------|------------|
| Robust Mean | 25.2 | 6.1 | 10.8 | 50.8 | 9.7 |
| Robust Standard Deviation | 3.3 | 2.1 | 2.1 | 7.0 | 1.9 |
| Standard Uncertainty | 0.9 | 0.7 | 0.7 | 2.0 | 0.6 |
| RSD (%) | 13.1 | 34.5 | 19.6 | 13.8 | 19.6 |
| Number of Sample Measurements | 19 | 15 | 16 | 19 | 16 |
| Acceptable Range: | | | | | |
| Upper Limit | 31.2 | 12.1 | 16.8 | 61.0 | 15.7 |
| Lower Limit | 19.2 | 0.1 | 4.8 | 40.6 | 3.7 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Blood Arsenic Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/L whole blood) | | | | | Info Only |
|----------------|---------------|----------------------------|---------|---------|---------|---------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | |
| Target Values: | | 25.2 | 6.1 | 10.8 | 50.8 | 9.7 | |
| 103 | DRC/CC-ICP-MS | 26.9 | 4.4 | 10.4 | 54.0 | 8.7 | Info |
| 110 | DRC/CC-ICP-MS | 23.3 | 4.6 | 9.2 | 51.2 | 7.9 | |
| 114 | ICP-MS | 28.0 | 9.0 | 13.0 | 50.0 | 13.0 | |
| 147 | ICP-MS | 24.4 | 4.4 | 9.4 | 52.7 | 8.2 | Info |
| 156 | DRC/CC-ICP-MS | 25.0 | <5.0 | 9.2 | 50.0 | 8.6 | |
| 164 | ICP-MS | 28.0 | 6.0 | 12.0 | 57.0 | 10.0 | |
| 179 | DRC/CC-ICP-MS | 26.0 | <12.0 | <12.0 | 56.0 | <12.0 | |
| 197 | DRC/CC-ICP-MS | 21.0 | <10.0 | <10.0 | 45.0 | <10.0 | |
| 200 | ICP-MS | 23.4 | 6.7 | 11.0 | 41.2 | 9.9 | Info |
| 206 | DRC/CC-ICP-MS | 24.4 | <10.0 | <10.0 | 52.4 | <10.0 | |
| 208 | ICP-MS | 21.6 | 7.5 | 12.0 | 40.1 ↓ | 10.8 | |
| 293 | DRC/CC-ICP-MS | 28.8 | 5.4 | 11.5 | 61.6 ↑ | 9.9 | Info |
| 305 | ICP-MS | 23.0 | 5.0 | 9.0 | 47.0 | 9.0 | |
| 312 | DRC/CC-ICP-MS | 22.0 | 6.0 | 9.0 | 42.0 | 8.0 | |
| 324 | ICP-MS | 31.3 ↑ | 10.9 | 15.0 | 50.5 | 15.4 | Info |
| 359 | ICP-MS | 27.1 | 9.1 | 13.4 | 55.9 | 12.3 | |
| 388 | ICP-MS | 30.3 | 5.9 | 12.3 | 61.2 ↑ | 10.9 | |
| 391 | DRC/CC-ICP-MS | 24.8 | 4.4 | 9.1 | 52.4 | 8.3 | Info |
| 484 | ICP-MS | 20.9 | 3.7 | 7.9 | 44.6 | 7.4 | |

Percent satisfactory results for all participants: 95.8 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
▼ Result unacceptable
▲ Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

New York State Department of Health
Blood Arsenic Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD

| Results ($\mu\text{g/L}$ whole blood) | | | | | |
|--|---------|---------|---------|---------|---------|
| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 9 | 5 | 6 | 9 | 6 |
| Mean: | 24.7 | 5.0 | 9.7 | 51.6 | 8.6 |
| Standard Deviation: | 2.4 | 0.7 | 1.0 | 5.8 | 0.7 |
| RSD (%): | 9.8 | 14.4 | 10.3 | 11.1 | 8.5 |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 10 | 10 | 10 | 10 | 10 |
| Mean: | 25.8 | 6.8 | 11.5 | 50.0 | 10.7 |
| Standard Deviation: | 3.6 | 2.3 | 2.2 | 6.9 | 2.4 |
| RSD (%): | 14.1 | 33.7 | 19.0 | 13.8 | 22.3 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 19 | 15 | 16 | 19 | 16 |
| Mean: | 25.3 | 6.2 | 10.8 | 50.8 | 9.9 |
| Standard Deviation: | 3.1 | 2.1 | 2.0 | 6.3 | 2.2 |
| RSD (%): | 12.2 | 33.7 | 18.4 | 12.3 | 21.9 |

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #2, 2015

Whole Blood Cadmium

Test materials for cadmium were prepared from caprine (goat) whole blood preserved with K₂EDTA anticoagulant. A total of five blood pools were supplemented with different amounts of cadmium (as Cd²⁺).

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for whole blood cadmium range from 2.0 µg/L (18 nmol/L) to 18.4 µg/L (164 nmol/L).

Acceptable ranges are based on the OSHA criteria of ±15%, or ±1 µg/L around the target value, whichever is greater. So, the range is fixed at ±1 µg/L for concentrations below 6.6 µg/L, where above 6.6 µg/L, it is ±15%.

Discussion: Based upon the above criteria, 96.2% of the results reported by all participants were satisfactory, with one of the 26 laboratories (3.8%) reporting 2 or more of the 5 results outside the acceptable ranges.

New York State Department of Health
Blood Cadmium Test Results, 2015 Event #2
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ whole blood)

| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
|-------------------------------|------------|------------|------------|-------------|------------|
| Robust Mean | 9.4 | 4.5 | 4.0 | 18.4 | 2.0 |
| Robust Standard Deviation | 0.7 | 0.4 | 0.3 | 1.4 | 0.1 |
| Standard Uncertainty | 0.2 | 0.1 | 0.1 | 0.3 | <0.1 |
| RSD (%) | 7.4 | 8.8 | 6.6 | 7.5 | 7.5 |
| Number of Sample Measurements | 26 | 26 | 26 | 26 | 26 |
| Acceptable Range: | | | | | |
| Upper Limit | 10.8 | 5.5 | 5.0 | 21.2 | 3.0 |
| Lower Limit | 8.0 | 3.5 | 3.0 | 15.6 | 1.0 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Blood Cadmium Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/L whole blood) | | | | | Info Only |
|----------------|---------------|----------------------------|---------|---------|---------|---------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | |
| Target Values: | | 9.4 | 4.5 | 4.0 | 18.4 | 2.0 | |
| 103 | DRC/CC-ICP-MS | 9.8 | 4.6 | 4.4 | 18.9 | 1.9 | Info |
| 106 | ICP-MS | 10.1 | 4.4 | 4.1 | 19.4 | 1.9 | Info |
| 107 | ICP-MS | 9.7 | 4.6 | 4.2 | 19 | 2.1 | Info |
| 109 | ICP-MS | 9.7 | 4.5 | 4.1 | 19.3 | 2.0 | Info |
| 110 | ICP-MS | 9.4 | 4.4 | 4.0 | 18.6 | 2.0 | |
| 114 | ICP-MS | 8.7 | 13.0 ↑ | 3.5 | 16.6 | 1.8 | |
| 116 | DRC/CC-ICP-MS | 9.6 | 4.8 | 4.2 | 20.5 | 2.0 | Info |
| 147 | ICP-MS | 9.8 | 4.4 | 4.2 | 18.9 | 2.0 | Info |
| 156 | DRC/CC-ICP-MS | 8.9 | 4.5 | 4.0 | 18.0 | 1.9 | |
| 164 | ICP-MS | 8.9 | 4.2 | 3.9 | 17.4 | 1.8 | |
| 179 | DRC/CC-ICP-MS | 9.5 | 4.3 | 4.0 | 19.3 | 2.0 | |
| 197 | DRC/CC-ICP-MS | 9.9 | 4.4 | 4.2 | 18.4 | 2.2 | |
| 200 | ICP-MS | 8.9 | 4.4 | 3.7 | 18.3 | 1.9 | Info |
| 206 | ICP-MS | 9.5 | 4.8 | 4.3 | 18.7 | 2.3 | |
| 208 | ICP-MS | 6.4 ↓ | 3.0 ↓ | 3.1 | 12.0 ↓ | 1.7 | |
| 293 | ICP-MS | 9.4 | 5.0 | 3.9 | 18.4 | 2.0 | Info |
| 305 | ICP-MS | 10.7 | 5.2 | 4.5 | 19.9 | 2.4 | |
| 312 | ICP-MS | 8.0 | 4.0 | 3.5 | 15.0 ↓ | 1.9 | |
| 324 | ICP-MS | 8.6 | 3.7 | 3.3 | 16.4 | 1.8 | Info |
| 325 | ETAAS-Z | 10.7 | 5.0 | 4.4 | 17.4 | 2.2 | Info |
| 359 | ICP-MS | 9.8 | 4.3 | 4.2 | 20.2 | 2.0 | |
| 366 | ETAAS-Z | 8.8 | 5.1 | 4.0 | 17.5 | 1.9 | Info |
| 388 | ICP-MS | 8.8 | 4.1 | 3.8 | 17.2 | 1.8 | |
| 391 | DRC/CC-ICP-MS | 9.0 | 4.1 | 3.8 | 17.8 | 1.8 | Info |
| 401 | DRC/CC-ICP-MS | 9.7 | 4.6 | 4.2 | 19.8 | 2.0 | Info |
| 410 | ICP-MS | 10.2 | 4.7 | 4.5 | 19.6 | 2.0 | Info |

Percent satisfactory results for all participants: 96.2 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
▼ Result unacceptable
▲ Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

New York State Department of Health
Blood Cadmium Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD

| Results (µg/L whole blood) | | | | | |
|--------------------------------|---------|---------|---------|---------|---------|
| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 7 | 7 | 7 | 7 | 7 |
| Mean: | 9.5 | 4.5 | 4.1 | 19.0 | 2.0 |
| Standard Deviation: | 0.4 | 0.2 | 0.2 | 1.0 | 0.1 |
| RSD (%): | 4.1 | 5.1 | 4.7 | 5.2 | 6.4 |
| ETAAS-Z | | | | | |
| Number of Sample Measurements: | 2 | 2 | 2 | 2 | 2 |
| Mean: | 9.8 | 5.1 | 4.2 | 17.5 | 2.1 |
| Standard Deviation: | 1.3 | 0.1 | 0.3 | 0.1 | 0.2 |
| RSD (%): | — | — | — | — | — |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 17 | 17 | 17 | 17 | 17 |
| Mean: | 9.2 | 4.9 | 3.9 | 17.9 | 2.0 |
| Standard Deviation: | 1.0 | 2.2 | 0.4 | 2.1 | 0.2 |
| RSD (%): | 10.7 | 44.3 | 10.2 | 11.5 | 9.2 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 26 | 26 | 26 | 26 | 26 |
| Mean: | 9.3 | 4.8 | 4.0 | 18.2 | 2.0 |
| Standard Deviation: | 0.9 | 1.7 | 0.4 | 1.8 | 0.2 |
| RSD (%): | 9.4 | 36.4 | 8.9 | 9.9 | 8.3 |

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #2, 2015

Whole Blood Mercury

Test materials for mercury were prepared from caprine (goat) whole blood preserved with K₂EDTA anticoagulant. A total of five pools were supplemented with different amounts of mercury as inorganic Hg²⁺.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for whole blood mercury range from 3.2 µg/L (16 nmol/L) to 57.5 µg/L (287 nmol/L).

Acceptable ranges were fixed at ±30%, or ±3 µg/L around the target value, whichever is greater. That is, the range is fixed at ±3 µg/L for concentrations below 10 µg/L, while above 10 µg/L, it is ±30%.

Discussion: Based on the above criteria, 96.9% of results reported by all participants were satisfactory, with one of the 26 laboratories (3.8%) reporting 2 or more of the 5 results outside the acceptable ranges.

New York State Department of Health
Blood Mercury Test Results, 2015 Event #2
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ whole blood)

| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
|--|---------|---------|---------|---------|---------|
|--|---------|---------|---------|---------|---------|

| | | | | | |
|-------------------------------|-------------|-------------|------------|-------------|------------|
| Robust Mean | 24.0 | 12.9 | 3.2 | 57.5 | 3.8 |
| Robust Standard Deviation | 2.3 | 1.0 | 0.3 | 5.9 | 0.3 |
| Standard Uncertainty | 0.6 | 0.2 | 0.1 | 1.5 | 0.1 |
| RSD (%) | 9.5 | 7.8 | 8.3 | 10.3 | 9.2 |
| Number of Sample Measurements | 26 | 26 | 22 | 25 | 23 |
| | | | | | |
| Acceptable Range: | | | | | |
| Upper Limit | 31.2 | 16.8 | 6.2 | 74.7 | 6.8 |
| Lower Limit | 16.8 | 9.0 | 0.2 | 40.3 | 0.8 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Blood Mercury Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/L whole blood) | | | | | Info Only |
|----------------|---------------------------|----------------------------|---------|---------|---------|---------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | |
| Target Values: | | 24.0 | 12.9 | 3.2 | 57.5 | 3.8 | |
| 103 | DRC/CC-ICP-MS | 23.8 | 13.1 | 3.2 | 57.8 | 3.8 | Info |
| 106 | DRC/CC-ICP-MS | 23.4 | 13.2 | 3.1 | 59.9 | 3.7 | Info |
| 107 | DRC/CC-ICP-MS | 21 | 13 | 3.0 | 58 | 3.6 | Info |
| 109 | ICP-MS | 24.7 | 12.7 | 3.1 | 59.3 | 3.8 | Info |
| 110 | ICP-MS | 23.8 | 12.6 | 3.1 | 56.9 | 3.7 | |
| 114 | ICP-MS | 25.8 | 4.1 ↓ | 3.9 | 61.0 | 4.6 | |
| 116 | DRC/CC-ICP-MS | 24.7 | 13.6 | 3.2 | 61.6 | 3.8 | Info |
| 147 | ICP-MS | 23.5 | 13.1 | 3.0 | 59.0 | 3.8 | Info |
| 156 | ICP-MS | 22.0 | 11.0 | <3.0 | 53.0 | <3.0 | |
| 164 | ICP-MS | 28.0 | 15.0 | 4.0 | 65.0 | 5.0 | |
| 179 | DRC/CC-ICP-MS | 25.0 | 13.0 | 3.0 | 61.0 | 3.0 | |
| 197 | DRC/CC-ICP-MS | 22.0 | 11.0 | <5.0 | 52.0 | <5.0 | |
| 200 | ICP-MS | 22.4 | 12.4 | 3.6 | 50.2 | 4.0 | Info |
| 206 | ICP-MS | 24.6 | 12.4 | <3.0 | 57.6 | 3.7 | |
| 208 | ICP-MS | 19.5 | 10.7 | 3.1 | 44.5 | 3.7 | |
| 293 | ICP-MS | 23.8 | 13.1 | 3.4 | 55.1 | 3.9 | Info |
| 305 | ICP-MS | 24.0 | 17.0 ↑ | 3.0 | 55.0 | 6.0 | |
| 312 | ICP-MS | 22.0 | 12.0 | 2.7 | 52.0 | 3.2 | |
| 359 | ICP-MS | 26.8 | 14.6 | 3.7 | 71.2 | 4.3 | |
| 366 | ICP-MS | 21.5 | 13.1 | 2.8 | 50.0 | 3.3 | Info |
| 388 | ICP-MS | 21.6 | 11.4 | 2.5 | 49.1 | 2.7 | |
| 391 | CV-AAS | 35.7 ↑ | 16.0 | 4.1 | 86.8 ↑ | 5.1 | Info |
| 401 | DRC/CC-ICP-MS | 25.1 | 12.8 | 3.8 | 62.2 | 3.8 | Info |
| 410 | ICP-MS | 25.9 | 13.8 | 3.3 | 63.3 | 3.9 | Info |
| 453 | Atomic Spectrometry Other | 24.5 | 12.7 | 3.0 | 56.2 | 3.4 | Info |
| 484 | ICP-MS | 27.3 | 13.7 | <5.0 | >60.0 | <5.0 | |

Percent satisfactory results for all participants: 96.9 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
▼ Result unacceptable
▲ Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

New York State Department of Health
Blood Mercury Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD

| | Results ($\mu\text{g/L}$ whole blood) | | | | |
|--------------------------------|--|---------|---------|---------|---------|
| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| Atomic Spectrometry | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 24.5 | 12.7 | 3.0 | 56.2 | 3.4 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| CV-AAS | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 35.7 | 16.0 | 4.1 | 86.8 | 5.1 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 7 | 7 | 6 | 7 | 6 |
| Mean: | 23.6 | 12.8 | 3.2 | 58.9 | 3.6 |
| Standard Deviation: | 1.6 | 0.8 | 0.3 | 3.5 | 0.3 |
| RSD (%): | 6.7 | 6.5 | 9.3 | 5.9 | 8.6 |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 17 | 17 | 14 | 16 | 15 |
| Mean: | 24.0 | 12.5 | 3.2 | 56.4 | 4.0 |
| Standard Deviation: | 2.3 | 2.7 | 0.4 | 6.8 | 0.8 |
| RSD (%): | 9.7 | 21.2 | 13.8 | 12.0 | 19.8 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 26 | 26 | 22 | 25 | 23 |
| Mean: | 24.3 | 12.7 | 3.3 | 58.3 | 3.9 |
| Standard Deviation: | 3.1 | 2.3 | 0.4 | 8.3 | 0.7 |
| RSD (%): | 12.6 | 17.8 | 13.1 | 14.2 | 18.4 |

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #2, 2015

Whole Blood Lead

Test materials for lead were prepared from caprine (goat) whole blood obtained from animals dosed with lead acetate to create physiologically-bound Pb. Whole blood was collected into collection bags containing K₂EDTA anticoagulant.

Target values were established as the mean of 19 measurements performed by 17 reference laboratories using ICP-MS and ETAAS methods. Values range from 5 µg/dL to 17 µg/dL. Among the reference group, imprecision (SD) varied from 0.4 - 0.8 µg/dL.

Acceptable ranges are based on the CLIA '88 criteria (Federal Register Volume 57, Number 40, §§ 493.2 and 493.937, February 28, 1992). The criteria are set at ±10% or ±4 µg/dL, whichever is greater.

Discussion Based on the CLIA '88 criteria, 99.8% of results reported by all participants were judged as satisfactory, with none of 82 participant laboratories reporting 2 or more of the 5 results outside the acceptable ranges.

New York State Department of Health
Blood Lead Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/dL whole blood) | | | | | Normalized Mean | Info Only |
|----------------|-----------------|-----------------------------|---------|---------|---------|---------|-----------------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | | |
| Target values: | | 12 | 17 | 10 | 5 | 5 | | |
| 103 | DRC/CC-ICP-MS | 12 | 17 | 10 | 5 | 5 | 1.00 | |
| 104 | ETAAS-Z | 12 | 18 | 10 | 5 | 5 | 1.03 | |
| 106 | ICP-MS | 12 | 17 | 10 | 6 | 5 | 1.00 | Info |
| 107 | ASV-LeadCare II | 12 | 17 | 9 | 6 | 5 | 1.00 | Info |
| 107 | ICP-MS | 12 | 17 | 9.7 | 5.5 | 4.8 | 1.00 | |
| 109 | ASV-LeadCare II | 11 | 15 | 8 | 5 | 4 | 0.90 | Info |
| 109 | ICP-MS | 12.6 | 16.5 | 9.6 | 5.4 | 4.6 | 1.04 | |
| 109 | ETAAS-Z | 12 | 18 | 10 | 5 | 4 | 1.03 | |
| 110 | ICP-MS | 12 | 16 | 10 | 5 | 4 | 0.97 | |
| 110 | ETAAS-Z | 12 | 16 | 9 | 5 | 4 | 0.97 | |
| 110 | ASV-LeadCare II | 16 | 21 | 11 | 6 | 5 | 1.22 | Info |
| 112 | ETAAS-Z | 13 | 18 | 10 | 5 | 5 | 1.07 | |
| 114 | ICP-MS | 12 | 17 | 10 | 5 | 5 | 1.00 | |
| 116 | DRC/CC-ICP-MS | 12 | 17 | 10 | 6 | 5 | 1.00 | Info |
| 121 | ETAAS-Z | 14 | 20 | 12 | 6 | 5 | 1.18 | Info |
| 123 | ETAAS-Z | 10 | 13 | 8 | 5 | 4 | 0.76 | |
| 126 | ICP-MS | 11 | 16 | 9 | 5 | 4 | 0.93 | |
| 131 | ETAAS-Z | 12 | 15 | 10 | 7 | 4 | 0.94 | |
| 143 | ETAAS-Z | 11 | 17 | 9 | 5 | 4 | 0.96 | |
| 144 | ETAAS-Z | 11 | 16 | 9 | 5 | 4 | 0.93 | |
| 147 | ICP-MS | 12 | 16 | 10 | 5 | 5 | 0.97 | |
| 150 | ETAAS-Z | 15 | 21 | 11 | 6 | 6 | 1.20 | |
| 156 | DRC/CC-ICP-MS | 12 | 15 | 10 | 6 | 5 | 0.94 | |

Notes: ↑ reported value outside upper limit
↓ reported value outside lower limit

Normalized Mean: The average of each reported result divided by the corresponding target value. It measures bias.

Info Only: results included for informational purposes only.

ND: non-detect

▼: Result unacceptable

New York State Department of Health
Blood Lead Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/dL whole blood) | | | | | Normalized Mean | Info Only |
|----------------|--------------------|-----------------------------|---------|---------|---------|---------|-----------------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | | |
| Target values: | | 12 | 17 | 10 | 5 | 5 | | |
| 158 | ICP-MS | 13 | 17 | 10 | 6 | 5 | 1.04 | |
| 160 | ICP-MS | 12 | 16 | 9 | 5 | 5 | 0.97 | |
| 164 | ICP-MS | 12 | 17 | 10 | 6 | 5 | 1.00 | |
| 166 | ETAAS-Z | 12 | 16 | 9 | 5 | 4 | 0.97 | |
| 168 | ETAAS-Z | 14 | 18 | 10 | 6 | 6 | 1.11 | |
| 179 | DRC/CC-ICP-MS | 12 | 17 | 10 | 6 | 5 | 1.00 | |
| 197 | ICP-MS | 11 | 15 | 9 | 5 | 4 | 0.90 | |
| 198 | ETAAS-Z | 12 | 17 | 10 | 5 | 5 | 1.00 | |
| 200 | ICP-MS | 11 | 16 | 10 | 6 | 5 | 0.93 | |
| 204 | ASV-LeadCare Ultra | 13 | 16 | 9 | 6 | 4 | 1.01 | |
| 206 | ICP-MS | 12 | 16 | 9 | 5 | 5 | 0.97 | |
| 208 | ETAAS-Z | 10 | 14 | 8 | 4 | 4 | 0.82 | |
| 237 | ETAAS-Z | 12 | 17 | 10 | 6 | 5 | 1.00 | |
| 254 | ETAAS-Z | 11 | 15 | 9 | 5 | 4 | 0.90 | |
| 255 | ETAAS-Z | 14 | 17 | 10 | 6 | 5 | 1.08 | |
| 269 | ETAAS-Z | 9 | 12 ↓ | 9 | 4 | 3 | 0.71 | |
| 272 | ETAAS-Z | 13 | 16 | 9 | 5 | 3 | 1.01 | |
| 279 | ETAAS-Z | 12 | 17 | 10 | 5 | 4 | 1.00 | |
| 290 | ICP-MS | 11 | 14 | 9 | 5 | 4 | 0.87 | |
| 291 | ASV-LeadCare Ultra | 12 | 17 | 10 | 6 | 5 | 1.00 | |
| 293 | ICP-MS | 12 | 16 | 9 | 5 | 4 | 0.97 | |
| 295 | ASV-3010 | 8 | 13 | 7 | 4 | 2 | 0.76 | |
| 301 | ETAAS Other | 11 | 15 | 9 | 5 | 4 | 0.90 | |

Notes: ↑ reported value outside upper limit
↓ reported value outside lower limit

Normalized Mean: The average of each reported result divided by the corresponding target value. It measures bias.

Info Only: results included for informational purposes only.

ND: non-detect

▼: Result unacceptable

New York State Department of Health
Blood Lead Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/dL whole blood) | | | | | Normalized Mean | Info Only |
|----------------|--------------------|-----------------------------|---------|---------|---------|---------|-----------------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | | |
| Target values: | | 12 | 17 | 10 | 5 | 5 | | |
| 305 | FAAS | 12 | 16 | 10 | 6 | 5 | 0.97 | |
| 312 | ICP-MS | 11 | 14 | 8 | 5 | 4 | 0.87 | |
| 317 | ETAAS-Z | 13 | 17 | 10 | 6 | 5 | 1.04 | |
| 324 | ETAAS-Z | 11.5 | 17.0 | 9.4 | 5.0 | 4.0 | 1.00 | |
| 325 | ETAAS-Z | 12 | 16 | 9 | 5 | 5 | 0.97 | |
| 333 | ETAAS-Z | 12 | 16 | 10 | 6 | 7 | 0.97 | |
| 337 | ASV-LeadCare II | 12 | 17 | 8 | <3 | 5 | 1.00 | |
| 340 | ETAAS-Z | 11 | 15 | 9 | 4 | 4 | 0.90 | |
| 343 | ASV-LeadCare Ultra | 13 | 17 | 9 | 5 | 3 | 1.04 | Info |
| 343 | ASV-LeadCare | 12 | 16 | 8 | 5 | 5 | 0.97 | Info |
| 343 | ASV-LeadCare II | 13 | 16 | 8 | 5 | 4 | 1.01 | Info |
| 345 | ASV-LeadCare II | 13 | 17 | 10 | 4 | 5 | 1.04 | |
| 348 | ETAAS-Z | 13 | 17 | 10 | 6 | 5 | 1.04 | |
| 349 | ETAAS-Z | 11 | 15 | 8 | 4 | 3 | 0.90 | |
| 350 | ASV-LeadCare Ultra | 11 | 15 | 9 | 5 | 4 | 0.90 | |
| 365 | ETAAS-Z | 12 | 17 | 10 | 6 | 5 | 1.00 | |
| 366 | ICP-MS | 12 | 17 | 10 | 5.2 | 4 | 1.00 | Info |
| 368 | ASV-3010 | 12 | 16 | 10 | 6 | 4 | 0.97 | |
| 369 | ASV-LeadCare Ultra | 10 | 16 | 9 | 3 | 4 | 0.94 | |
| 374 | ASV-LeadCare Ultra | 13 | 18 | 9 | 6 | 4 | 1.07 | |
| 384 | ASV-3010 | 11 | 14 | 7 | 3 | <1 | 0.87 | |
| 388 | ICP-MS | 11 | 16 | 8 | 4 | 3 | 0.93 | |
| 389 | ETAAS-Z | 12 | 16 | 9 | 5 | 4 | 0.97 | |

Notes: ↑ reported value outside upper limit
↓ reported value outside lower limit

Normalized Mean: The average of each reported result divided by the corresponding target value. It measures bias.

Info Only: results included for informational purposes only.

ND: non-detect

▼: Result unacceptable

**New York State Department of Health
Blood Lead Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES**

| Lab Code | Method | Results (µg/dL whole blood) | | | | | Normalized Mean | Info Only |
|----------------|--------------------|-----------------------------|---------|---------|---------|---------|-----------------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | | |
| Target values: | | 12 | 17 | 10 | 5 | 5 | | |
| 391 | ETAAS-Z | 10.4 | 14.7 | 8.2 | 4.8 | 4.2 | 0.86 | Info |
| 393 | ASV-LeadCare II | 12 | 15 | 8 | 4 | 4 | 0.94 | |
| 401 | DRC/CC-ICP-MS | 12.0 | 16.4 | 9.7 | 5.4 | 4.8 | 0.97 | Info |
| 410 | ICP-MS | 12 | 17 | 10 | 5 | 5 | 1.00 | Info |
| 461 | ASV-LeadCare Ultra | 12 | 16 | 9 | 6 | 4 | 0.97 | |
| 464 | ASV-LeadCare II | 14 | 21 | 10 | 5 | 5 | 1.20 | |
| 466 | ASV-LeadCare II | 12 | 16 | 8 | 4 | 5 | 0.97 | |
| 469 | ICP-MS | 11 | 15 | 9 | 5 | 4 | 0.90 | |
| 470 | ASV-LeadCare II | 11 | 16 | 8 | 5 | 4 | 0.93 | |
| 476 | ASV-LeadCare | 13 | 16 | 8 | 4 | 5 | 1.01 | |
| 477 | ASV-LeadCare II | 11 | 14 | 7 | 5 | 4 | 0.87 | |
| 482 | ASV-LeadCare II | 12 | 15 | 7 | 4 | 5 | 0.94 | |
| 484 | ICP-MS | 13 | 15 | 8 | 5 | 3 | 0.98 | |

Percent satisfactory results for all participants: 99.8 %

Notes: ↑ reported value outside upper limit
↓ reported value outside lower limit

Normalized Mean: The average of each reported result divided by the corresponding target value. It measures bias.

Info Only: results included for informational purposes only.

ND: non-detect

▼: Result unacceptable

**New York State Department of Health
Blood Lead Test Results, 2015 Event #2
STATISTICAL SUMMARY**

| Lab Code | Method | TARGET VALUE ASSIGNMENT AND STATISTICS | | | | |
|--------------------------------|---------------|--|---------|---------|---------|---------|
| | | Results (µg/dL whole blood) | | | | |
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 103 | DRC/CC-ICP-MS | 12 | 17 | 10 | 5 | 5 |
| 104 | ETAAS-Z | 12 | 18 | 10 | 5 | 5 |
| 107 | ICP-MS | 12 | 17 | 9.7 | 5.5 | 4.8 |
| 109 | ICP-MS | 12.6 | 16.5 | 9.6 | 5.4 | 4.6 |
| 109 | ETAAS-Z | 12 | 18 | 10 | 5 | 4 |
| 110 | ICP-MS | 12 | 16 | 10 | 5 | 4 |
| 110 | ETAAS-Z | 12 | 16 | 9 | 5 | 4 |
| 112 | ETAAS-Z | 13 | 18 | 10 | 5 | 5 |
| 147 | ICP-MS | 12 | 16 | 10 | 5 | 5 |
| 156 | DRC/CC-ICP-MS | 12 | 15 | 10 | 6 | 5 |
| 160 | ICP-MS | 12 | 16 | 9 | 5 | 5 |
| 164 | ICP-MS | 12 | 17 | 10 | 6 | 5 |
| 166 | ETAAS-Z | 12 | 16 | 9 | 5 | 4 |
| 179 | DRC/CC-ICP-MS | 12 | 17 | 10 | 6 | 5 |
| 198 | ETAAS-Z | 12 | 17 | 10 | 5 | 5 |
| 200 | ICP-MS | 11 | 16 | 10 | 6 | 5 |
| 293 | ICP-MS | 12 | 16 | 9 | 5 | 4 |
| 324 | ETAAS-Z | 11.5 | 17.0 | 9.4 | 5.0 | 4.0 |
| 325 | ETAAS-Z | 12 | 16 | 9 | 5 | 5 |
| Number of Sample Measurements: | | 19 | 19 | 19 | 19 | 19 |
| Mean (target value): | | 12 | 17 | 10 | 5 | 5 |
| Standard Deviation: | | 0.4 | 0.8 | 0.4 | 0.4 | 0.5 |
| RSD (%): | | 3.2 | 5.0 | 4.6 | 8.0 | 10.0 |
| Acceptable Range: | | | | | | |
| Upper Limit: | | 16 | 21 | 14 | 9 | 9 |
| Lower Limit: | | 8 | 13 | 6 | 1 | 1 |

notes: Results reported as less than the detection limits are treated as zero for statistical and grading purposes.

**New York State Department of Health
Blood Lead Test Results, 2015 Event #2
STATISTICAL SUMMARY BY CLASS**

| | Results (µg/dL whole blood) | | | | |
|--------------------------------|-----------------------------|---------|---------|---------|---------|
| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| Evaluated | | | | | |
| Number of Sample Measurements: | 50 | 50 | 50 | 49 | 49 |
| Mean: | 11.8 | 15.9 | 9.0 | 5.0 | 4.3 |
| Standard Deviation: | 1.3 | 1.7 | 1.0 | 0.9 | 0.9 |
| RSD (%): | 10.9 | 10.4 | 10.9 | 17.6 | 20.2 |
| Info | | | | | |
| Number of Sample Measurements: | 13 | 13 | 13 | 13 | 13 |
| Mean: | 12.4 | 17.0 | 9.5 | 5.4 | 4.5 |
| Standard Deviation: | 1.4 | 1.7 | 1.2 | 0.5 | 0.6 |
| RSD (%): | 11.2 | 10.3 | 13.1 | 9.2 | 14.1 |
| Reference | | | | | |
| Number of Sample Measurements: | 19 | 19 | 19 | 19 | 19 |
| Mean: | 12.0 | 16.6 | 9.7 | 5.3 | 4.7 |
| Standard Deviation: | 0.4 | 0.8 | 0.4 | 0.4 | 0.5 |
| RSD (%): | 3.2 | 5.0 | 4.6 | 8.0 | 10.0 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 82 | 82 | 82 | 81 | 81 |
| Mean: | 11.9 | 16.2 | 9.2 | 5.2 | 4.4 |
| Standard Deviation: | 1.2 | 1.6 | 1.0 | 0.8 | 0.8 |
| RSD (%): | 9.8 | 9.7 | 10.5 | 14.7 | 17.3 |

notes: ? Insufficient data for calculation.

**New York State Department of Health
Blood Lead Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD**

| | | Results (µg/dL whole blood) | | | | |
|--------------------------------|--|-----------------------------|---------|---------|---------|---------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| ASV-3010 | | | | | | |
| Number of Sample Measurements: | | 3 | 3 | 3 | 3 | 2 |
| Mean: | | 10.3 | 14.3 | 8.0 | 4.3 | 3.0 |
| Standard Deviation: | | 2.1 | 1.5 | 1.7 | 1.5 | 1.4 |
| RSD (%): | | — | — | — | — | — |
| ASV-LeadCare | | | | | | |
| Number of Sample Measurements: | | 2 | 2 | 2 | 2 | 2 |
| Mean: | | 12.5 | 16.0 | 8.0 | 4.5 | 5.0 |
| Standard Deviation: | | 0.7 | 0.0 | 0.0 | 0.7 | 0.0 |
| RSD (%): | | — | — | — | — | — |
| ASV-LeadCare II | | | | | | |
| Number of Sample Measurements: | | 12 | 12 | 12 | 11 | 12 |
| Mean: | | 12.4 | 16.7 | 8.5 | 4.8 | 4.6 |
| Standard Deviation: | | 1.4 | 2.2 | 1.2 | 0.8 | 0.5 |
| RSD (%): | | 11.6 | 13.4 | 14.6 | 15.6 | 11.2 |
| ASV-LeadCare Ultra | | | | | | |
| Number of Sample Measurements: | | 7 | 7 | 7 | 7 | 7 |
| Mean: | | 12.0 | 16.4 | 9.1 | 5.3 | 4.0 |
| Standard Deviation: | | 1.2 | 1.0 | 0.4 | 1.1 | 0.6 |
| RSD (%): | | 9.6 | 5.9 | 4.1 | 21.1 | 14.4 |
| DRC/CC-ICP-MS | | | | | | |
| Number of Sample Measurements: | | 5 | 5 | 5 | 5 | 5 |
| Mean: | | 12.0 | 16.5 | 9.9 | 5.7 | 5.0 |
| Standard Deviation: | | 0.0 | 0.9 | 0.1 | 0.5 | 0.1 |
| RSD (%): | | 0.0 | 5.3 | 1.3 | 8.1 | 1.8 |
| ETAAS Other | | | | | | |
| Number of Sample Measurements: | | 1 | 1 | 1 | 1 | 1 |
| Mean: | | 11.0 | 15.0 | 9.0 | 5.0 | 4.0 |
| Standard Deviation: | | ? | ? | ? | ? | ? |
| RSD (%): | | — | — | — | — | — |
| ETAAS-Z | | | | | | |
| Number of Sample Measurements: | | 30 | 30 | 30 | 30 | 30 |
| Mean: | | 12.0 | 16.4 | 9.5 | 5.2 | 4.5 |
| Standard Deviation: | | 1.3 | 1.8 | 0.9 | 0.7 | 0.9 |
| RSD (%): | | 11.0 | 11.1 | 9.3 | 14.0 | 20.0 |
| FAAS | | | | | | |
| Number of Sample Measurements: | | 1 | 1 | 1 | 1 | 1 |
| Mean: | | 12.0 | 16.0 | 10.0 | 6.0 | 5.0 |
| Standard Deviation: | | ? | ? | ? | ? | ? |
| RSD (%): | | — | — | — | — | — |

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1 (see DRC/CC-ICP-MS and HR-ICP-MS participants)

**New York State Department of Health
Blood Lead Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD**

| Results ($\mu\text{g/dL}$ whole blood) | | | | | |
|--|----------------|----------------|----------------|----------------|----------------|
| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 21 | 21 | 21 | 21 | 21 |
| Mean: | 11.8 | 16.0 | 9.3 | 5.2 | 4.4 |
| Standard Deviation: | 0.6 | 1.0 | 0.7 | 0.5 | 0.7 |
| RSD (%): | 5.5 | 6.0 | 7.7 | 9.3 | 14.9 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 82 | 82 | 82 | 81 | 81 |
| Mean: | 11.9 | 16.2 | 9.2 | 5.2 | 4.4 |
| Standard Deviation: | 1.2 | 1.6 | 1.0 | 0.8 | 0.8 |
| RSD (%): | 9.8 | 9.7 | 10.5 | 14.7 | 17.3 |

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1 (see DRC/CC-ICP-MS and HR-ICP-MS participants)

New York State Department of Health
Event #2, 2015

Additional Trace Elements Reported in Whole Blood

Participant laboratories reported their analytical results for any additional trace elements (other than As, Cd, Hg and Pb) that are routinely reported so that a more complete characterization can be recorded for these proficiency test materials. Results for the additional trace elements cobalt (Co) and chromium (Cr) are reported here. Although these data are provided solely for educational and informational purposes, target values and acceptable ranges are provided. The New York State grading criteria were established after discussions with the FDA and with other trace element PT scheme organizers. Departures from the acceptable ranges should trigger an internal Quality Assurance review.

Additional Elements

Co and Cr

New York State Department of Health
Event #2, 2015

Whole Blood Cobalt

Test materials for cobalt were prepared from caprine (goat) whole blood preserved with K₂EDTA anticoagulant. A total of five pools were supplemented with cobalt as inorganic Co²⁺.

The Target Values assigned for each PT material is the arithmetic mean of the results reported by all participants for the event. Values for whole blood cobalt range from 1.6 µg/L to 22.6 µg/L after outlier exclusion.

Acceptable range: The acceptable range for cobalt is set at ±1.5 µg/L or ±20%, whichever is greater. Thus, it is fixed at ±1.5 µg/L for concentrations below 7.5 µg/L. These NYS grading criteria were established after discussions with the FDA and with other trace element PT scheme organizers.

Discussion: Based upon the above criteria, 92.7% of test results reported were within the acceptable ranges, with one of the 11 laboratories (9.1%) reporting 2 or more of the 5 results outside the acceptable ranges. Upward and downward indicator arrows next to individual results should be used as part of a laboratory's on-going internal quality assessment (QA) program. Note that this grading scheme is intended for educational purposes. Departures from the acceptable ranges should trigger an internal QA review.

New York State Department of Health
Blood Cobalt Test Results, 2015 Event #2
STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ whole blood)

| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
|--------------------------------|-------------|------------|------------|-------------|------------|
| Arithmetic Mean* | 12.4 | 6.5 | 1.6 | 22.6 | 7.1 |
| Standard Deviation | 0.7 | 0.5 | 0.3 | 1.7 | 0.5 |
| RSD (%) | 5.8 | 6.9 | 21.1 | 7.6 | 6.6 |
| Number of Sample Measurements* | 10 | 10 | 10 | 11 | 10 |
| Acceptable Range: | | | | | |
| Upper Limit | 14.9 | 8.0 | 3.1 | 27.1 | 8.6 |
| Lower Limit | 9.9 | 5.0 | 0.1 | 18.1 | 5.6 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

* Outliers identified by Grubbs' test excluded

New York State Department of Health
Blood Cobalt Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/L whole blood) | | | | | Info Only |
|----------------|---------------|----------------------------|---------|---------|---------|---------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | |
| Target Values: | | 12.4 | 6.5 | 1.6 | 22.6 | 7.1 | |
| 103 | DRC/CC-ICP-MS | 13.1 | 6.84 | 1.66 | 23.3 | 7.36 | Info |
| 110 | ICP-MS | 13.4 | 7.1 | 2.6 | 24.2 | 7.7 | |
| 147 | ICP-MS | 12.7 | 6.60 | 1.59 | 23.5 | 7.25 | |
| 156 | DRC/CC-ICP-MS | 12 | 6.8 | 1.6 | 22 | 7.3 | |
| 164 | ICP-MS | 12.6 | 6.5 | 1.5 | 22.5 | 7.2 | |
| 197 | ICP-MS | 11.9 | 6.1 | 1.4 | 21.7 | 6.7 | |
| 206 | ICP-MS | 12.8 | 6.9 | 1.6 | 22.4 | 7.5 | |
| 305 | ICP-MS | 11.7 | 5.7 | 1.6 | 21.1 | 6.5 | |
| 312 | ICP-MS | 11 | 6.0 | 1.4 | 19 | 6.2 | |
| 324 | ICP-MS | 16.0 ↑ | 8.8 ↑ | 4.3 ↑ | 25.6 | 9.3 ↑ | Info |
| 391 | DRC/CC-ICP-MS | 12.6 | 6.3 | 1.5 | 22.8 | 7.1 | Info |

Percent satisfactory results for all participants: 92.7 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
▼ Result unacceptable
▲ Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

New York State Department of Health
Blood Cobalt Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD

| Results (µg/L whole blood) | | | | | |
|--------------------------------|---------|---------|---------|---------|---------|
| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 3 | 3 | 3 | 3 | 3 |
| Mean: | 12.6 | 6.6 | 1.6 | 22.7 | 7.3 |
| Standard Deviation: | 0.6 | 0.3 | 0.1 | 0.7 | 0.1 |
| RSD (%): | — | — | — | — | — |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 8 | 8 | 8 | 8 | 8 |
| Mean: | 12.8 | 6.7 | 2.0 | 22.5 | 7.3 |
| Standard Deviation: | 1.5 | 1.0 | 1.0 | 2.0 | 1.0 |
| RSD (%): | 11.8 | 14.4 | 50.4 | 8.9 | 13.1 |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 11 | 11 | 11 | 11 | 11 |
| Mean: | 12.7 | 6.7 | 1.9 | 22.6 | 7.3 |
| Standard Deviation: | 1.3 | 0.8 | 0.9 | 1.7 | 0.8 |
| RSD (%): | 10.1 | 12.2 | 45.9 | 7.6 | 11.0 |

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #2, 2015

Whole Blood Chromium

Test materials for chromium were prepared from caprine (goat) whole blood preserved with K₂EDTA anticoagulant. A total of five pools were supplemented with chromium as inorganic Cr³⁺.

The Target Values assigned for each PT material is the arithmetic mean of the results reported by all participants for the event. Values for whole blood chromium range from 2.7 µg/L (52 nmol/L) to 21.9 µg/L (421 nmol/L) after outlier exclusion.

Acceptable range: The acceptable range for chromium is set at ±2 µg/L or ±20%, whichever is greater. Thus, it is fixed at ±2 µg/L for concentrations below 10 µg/L. These NYS grading criteria were established after discussions with the FDA and with other trace element PT scheme organizers.

Discussion: Based upon the above criteria, 97.5% of test results reported were within the acceptable ranges, with none of the 8 laboratories reporting 2 or more of the 5 results outside the acceptable ranges. Upward and downward indicator arrows next to individual results should be used as part of a laboratory's on-going internal quality assessment (QA) program. Note that this grading scheme is intended for educational purposes. Departures from the acceptable ranges should trigger an internal QA review.

New York State Department of Health
Blood Chromium Test Results, 2015 Event #2
STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ whole blood)

BE15-06 BE15-07 BE15-08 BE15-09 BE15-10

| | | | | | |
|--------------------------------|-------------|------------|------------|-------------|------------|
| Arithmetic Mean* | 13.7 | 2.7 | 2.9 | 21.9 | 9.3 |
| Standard Deviation | 1.0 | 0.3 | 0.5 | 1.6 | 0.8 |
| RSD (%) | 7.5 | 12.5 | 17.5 | 7.2 | 9.1 |
| Number of Sample Measurements* | 8 | 8 | 7 | 8 | 8 |
| Acceptable Range: | | | | | |
| Upper Limit | 16.4 | 4.7 | 4.9 | 26.3 | 11.3 |
| Lower Limit | 11.0 | 0.7 | 0.9 | 17.5 | 7.3 |

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

* Outliers identified by Grubbs' test excluded

New York State Department of Health
Blood Chromium Test Results, 2015 Event #2
PERFORMANCE OF PARTICIPATING LABORATORIES

| Lab Code | Method | Results (µg/L whole blood) | | | | | Info Only |
|----------------|---------------|----------------------------|---------|---------|---------|---------|-----------|
| | | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 | |
| Target Values: | | 13.7 | 2.7 | 2.9 | 21.9 | 9.3 | |
| 103 | DRC/CC-ICP-MS | 14.8 | 2.83 | 3.38 | 23.1 | 9.61 | Info |
| 110 | DRC/CC-ICP-MS | 14.1 | 2.8 | 5.4 ↑ | 22.4 | 9.7 | |
| 147 | DRC/CC-ICP-MS | 15.0 | 2.85 | 3.34 | 24.9 | 10.5 | Info |
| 156 | DRC/CC-ICP-MS | 13 | 2.7 | 3 | 21 | 9.1 | |
| 164 | DRC/CC-ICP-MS | 12.8 | 2.4 | 2.9 | 20.6 | 9.1 | |
| 197 | DRC/CC-ICP-MS | 14.1 | 3.2 | 3.3 | 22.1 | 9.7 | |
| 305 | ICP-MS | 13.5 | 2.5 | 2.5 | 21.2 | 8.9 | |
| 312 | DRC/CC-ICP-MS | 12 | 2.1 | 2.0 | 20 | 7.6 | |

Percent satisfactory results for all participants: 97.5 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
▼: Result unacceptable
▲: Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

New York State Department of Health
Blood Chromium Test Results, 2015 Event #2
STATISTICAL SUMMARY BY METHOD

| Results ($\mu\text{g/L}$ whole blood) | | | | | |
|--|---------|---------|---------|---------|---------|
| | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| DRC/CC-ICP-MS | | | | | |
| Number of Sample Measurements: | 7 | 7 | 7 | 7 | 7 |
| Mean: | 13.7 | 2.7 | 3.3 | 22.0 | 9.3 |
| Standard Deviation: | 1.1 | 0.4 | 1.0 | 1.7 | 0.9 |
| RSD (%): | 8.1 | 13.1 | 30.9 | 7.6 | 9.6 |
| ICP-MS | | | | | |
| Number of Sample Measurements: | 1 | 1 | 1 | 1 | 1 |
| Mean: | 13.5 | 2.5 | 2.5 | 21.2 | 8.9 |
| Standard Deviation: | ? | ? | ? | ? | ? |
| RSD (%): | — | — | — | — | — |
| All Laboratories | | | | | |
| Number of Sample Measurements: | 8 | 8 | 8 | 8 | 8 |
| Mean: | 13.7 | 2.7 | 3.2 | 21.9 | 9.3 |
| Standard Deviation: | 1.0 | 0.3 | 1.0 | 1.6 | 0.8 |
| RSD (%): | 7.5 | 12.5 | 30.9 | 7.2 | 9.1 |

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #2, 2015

Additional Trace Elements Reported in Whole Blood

Participant laboratories reported their analytical results for any additional trace elements (other than As, Cd, Hg and Pb) that are routinely reported so that a more complete characterization can be recorded for these proficiency test materials. Results for additional trace elements are reported here, but no target value is implied nor are any acceptable ranges provided. These data are provided solely for educational and informational purposes.

In addition to As, Cd, Pb and Hg, the whole blood pools were supplemented with the following additional trace elements as indicated below

Additional Elements

Mn, Sn, Tl, Ti, V, W, Ni, Ag

**New York State Department of Health
Whole Blood Additional Elements, 2015 Event #2
Page 1**

| Blood Aluminum (µg/L) | | | | | | |
|------------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 147 | ICP-MS | <5.4 | <5.4 | <5.4 | <5.4 | <5.4 |
| 305 | ICP-MS | 11.4 | 10.3 | 11.7 | 16.9 | 22.6 |
| 359 | ICP-MS | 31.5 | 27.3 | 33.4 | 43.2 | 27.6 |

| Blood Antimony (µg/L) | | | | | | |
|------------------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 103 | DRC/CC-ICP-MS | <0.258 | <0.258 | <0.258 | <0.258 | <0.258 |
| 110 | ICP-MS | <0.10 | <0.10 | <0.10 | <0.10 | <0.10 |
| 147 | ICP-MS | <0.04 | <0.04 | <0.04 | 0.0535 | <0.04 |
| 206 | DRC/CC-ICP-MS | <1.0 | <1.0 | <1.0 | <1.0 | <1.0 |

| Blood Barium (µg/L) | | | | | | |
|----------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 147 | ICP-MS | 12.0 | 10.0 | 13.9 | 9.97 | 16.9 |
| 197 | ICP-MS | 12.8 | 10.6 | 14.4 | 10.1 | 17.3 |
| 312 | ICP-MS | 11.8 | 9.6 | 12.4 | 8.8 | 15.1 |
| Arithmetic Mean | | 12.2 | 10.1 | 13.6 | 9.6 | 16.4 |
| SD | | 0.5 | 0.5 | 1.0 | 0.7 | 1.2 |
| n | | 3 | 3 | 3 | 3 | 3 |

| Blood Beryllium (µg/L) | | | | | | |
|-------------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 110 | ICP-MS | <0.14 | 0.2 | 0.3 | 0.3 | 0.2 |
| 147 | ICP-MS | <1.8 | <1.8 | <1.8 | <1.8 | <1.8 |
| 197 | ICP-MS | 0.2 | 0.3 | 0.3 | 0.2 | 0.2 |

| Blood Bismuth (µg/L) | | | | | | |
|-----------------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 197 | ICP-MS | <1.0 | <1.0 | <1.0 | <1.0 | <1.0 |
| 206 | DRC/CC-ICP-MS | <1.0 | <1.0 | <1.0 | <1.0 | <1.0 |
| 305 | ICP-MS | <0.5 | <0.5 | <0.5 | <0.5 | <0.5 |

| Blood Cesium (µg/L) | | | | | | |
|----------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 110 | ICP-MS | 0.4 | 0.5 | 0.5 | 0.6 | 0.5 |

| Blood Copper (µg/L) | | | | | | |
|----------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 110 | ICP-MS | 1279 | 1447 | 1372 | 1218 | 1543 |
| 147 | ICP-MS | 1112 | 1290 | 1226 | 1067 | 1353 |
| 197 | ICP-MS | 1040 | 1260 | 1150 | 1050 | 1320 |
| 312 | ICP-MS | 960 | 1110 | 970 | 890 | 1130 |
| Arithmetic mean | | 1098 | 1277 | 1180 | 1056 | 1337 |
| SD | | 136 | 138 | 167 | 134 | 169 |
| n | | 4 | 4 | 4 | 4 | 4 |

New York State Department of Health
Whole Blood Additional Elements, 2015 Event #2
Page 2

| Blood Iodine (µg/L) | | | | | | |
|----------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 147 | ICP-MS | 53.8 | 91.6 | 52.7 | 73.3 | 57.6 |

| Blood Lithium (µg/L) | | | | | | |
|-----------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 147 | ICP-MS | 1.15 | 1.78 | 1.65 | 1.62 | 1.69 |

| Blood Manganese (µg/L) | | | | | | |
|-------------------------------|-----------------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 103 | DRC/CC-ICP-MS | 30.2 | 22.5 | 29.0 | 25.3 | 11.3 |
| 107 | DRC/CC-ICP-MS | 30 | 22 | 29 | 27 | 12 |
| 110 | ETAAS-Z | 29.0 | 21.7 | 28.7 | 25.9 | 11.9 |
| 114 | ICP-MS | 26.2 | 20.8 | 27.0 | 24.1 | 11.9 |
| 147 | ICP-MS | 29.2 | 20.9 | 28.3 | 25.5 | 11.8 |
| 156 | ICP-MS | 28 | 20 | 26 | 24 | 12 |
| 179 | DRC/CC-ICP-MS | 29.7 | 23.1 | 29.7 | 26.7 | 13.4 |
| 206 | DRC/CC-ICP-MS | 34.6 | 25.7 | 33.4 | 30.5 | 13.7 |
| 293 | ICP-MS | 27.0 | 21.3 | 24.5 | 22.8 | 9.8 |
| 305 | ICP-MS | 26.8 | 19.9 | 24.8 | 22.0 | 10.9 |
| 312 | DRC/CC-ICP-MS | 26 | 18 | 22 | 22 | 8 |
| 324 | ICP-MS | 22.5 | 16.0 | 16.9 | *14.8 | 5.7 |
| 391 | DRC/CC-ICP-MS | 29.4 | 20.4 | 26.5 | 25.0 | 10.2 |
| <i>*Outlier</i> | Arithmetic mean | 28.4 | 20.9 | 26.6 | 25.1 | 11.0 |
| | SD | 2.9 | 2.4 | 4.1 | 2.4 | 2.2 |
| | n | 13 | 13 | 13 | 12 | 13 |

| Blood Molybdenum (µg/L) | | | | | | |
|--------------------------------|-----------------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 103 | DRC/CC-ICP-MS | 128 | 22.2 | 89.1 | 26.8 | 46.4 |
| 147 | ICP-MS | 120 | 20.8 | 81.3 | 25.4 | 43.4 |
| 197 | ICP-MS | 132.5 | 23.4 | 91.4 | 28.4 | 49.1 |
| 305 | ICP-MS | 115.2 | 21.1 | 79.8 | 25.4 | 43.8 |
| 312 | ICP-MS | 110 | 20 | 71 | 23 | 40 |
| 324 | ICP-MS | 140.8 | 22.2 | 87.8 | 26.5 | 47.8 |
| | Arithmetic mean | 124 | 22 | 83 | 26 | 45 |
| | SD | 11 | 1 | 8 | 2 | 3 |
| | n | 6 | 6 | 6 | 6 | 6 |

| Blood Nickel (µg/L) | | | | | | |
|----------------------------|-----------------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 110 | DRC/CC-ICP-MS | 18.5 | 8.2 | *32.8 | 13.4 | 12.9 |
| 147 | ICP-MS | 18.9 | 8.10 | 3.48 | 13.5 | 12.6 |
| 197 | ICP-MS | 19.6 | 8.9 | 3.7 | 13.8 | 13.5 |
| 312 | ICP-MS | 17 | 7 | 4 | 12 | 11 |
| <i>*Outlier</i> | Arithmetic mean | 18.5 | 8.1 | 3.7 | 13.2 | 12.5 |
| | SD | 1.1 | 0.8 | 0.3 | 0.8 | 1.1 |
| | n | 4 | 4 | 3 | 4 | 4 |

**New York State Department of Health
Whole Blood Additional Elements, 2015 Event #2
Page 3**

| Blood Platinum (µg/L) | | | | | | |
|------------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 110 | ICP-MS | <0.10 | <0.10 | <0.10 | <0.10 | <0.10 |
| 312 | ICP-MS | <0.2 | <0.2 | <0.2 | <0.2 | <0.2 |

| Blood Selenium (µg/L) | | | | | | |
|------------------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 103 | DRC/CC-ICP-MS | 293 | 295 | 328 | 335 | 316 |
| 107 | DRC/CC-ICP-MS | 330 | 340 | 360 | 380 | 360 |
| 109 | ICP-MS | 298 | 297 | 322 | 339 | 323 |
| 114 | ICP-MS | 278 | 290 | 294 | 332 | 334 |
| 147 | ICP-MS | 280 | 293 | 330 | 333 | 314 |
| 305 | ICP-MS | 348 | 366 | 387 | 422 | 378 |
| 312 | ICP-MS | 278 | 278 | 287 | 309 | 282 |
| 359 | ICP-MS | 302.1 | 305.5 | 336.6 | 347.1 | 322.7 |
| Arithmetic Mean | | 301 | 308 | 331 | 350 | 329 |
| SD | | 26 | 30 | 33 | 35 | 29 |
| n | | 8 | 8 | 8 | 8 | 8 |

| Blood Silver (µg/L) | | | | | | |
|----------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 147 | ICP-MS | 8.76 | 1.76 | 5.75 | 4.07 | 4.65 |
| 197 | ICP-MS | 8.6 | 1.9 | 5.8 | 4.0 | 4.7 |

| Blood Strontium (µg/L) | | | | | | |
|-------------------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 103 | DRC/CC-ICP-MS | 32.1 | 23.1 | 36.7 | 25.3 | 28.9 |

| Blood Tellurium (µg/L) | | | | | | |
|-------------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 147 | ICP-MS | <0.08 | <0.08 | <0.08 | <0.08 | <0.08 |

| Blood Thorium (µg/L) | | | | | | |
|-----------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 147 | ICP-MS | <0.03 | <0.03 | <0.03 | <0.03 | <0.03 |

| Blood Thallium (µg/L) | | | | | | |
|------------------------------|---------------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 103 | DRC/CC-ICP-MS | 4.85 | 10.5 | 8.43 | 7.76 | 2.28 |
| 110 | ICP-MS | 4.9 | 10.8 | 8.6 | 8.0 | 2.4 |
| 147 | ICP-MS | 4.91 | 10.5 | 7.81 | 8.03 | 2.39 |
| 156 | DRC/CC-ICP-MS | 5 | 10 | 8.6 | 7.9 | 2.3 |
| 179 | ICP-MS | 5 | 11 | 9 | 8 | 2 |
| 197 | ICP-MS | 4.7 | 9.9 | 8.1 | 7.4 | 2.3 |
| 206 | ICP-MS | 5.0 | 11.2 | 8.5 | 7.9 | 2.4 |
| 305 | ICP-MS | 3.8 | 7.9 | 6.5 | 5.9 | 1.8 |
| 312 | ICP-MS | 4.2 | 9.1 | 6.7 | 6.4 | 1.9 |
| 324 | ICP-MS | 4.7 | 9.8 | 8.1 | 7.3 | 2.2 |
| Arithmetic mean | | 4.7 | 10.1 | 8.0 | 7.5 | 2.2 |
| SD | | 0.4 | 1.0 | 0.8 | 0.7 | 0.2 |
| n | | 10 | 10 | 10 | 10 | 10 |

**New York State Department of Health
Whole Blood Additional Elements, 2015 Event #2
Page 4**

| Blood Tin (µg/L) | | | | | | |
|-------------------------|------------------------|------------|-------------|------------|------------|-------------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 110 | ICP-MS | 3.3 | 13.8 | 4.7 | 6.9 | 11.6 |
| 147 | ICP-MS | 3.11 | 13.5 | 4.55 | 6.97 | 11.7 |
| 156 | DRC/CC-ICP-MS | 2.9 | 12 | 4.6 | 7.3 | 10 |
| 197 | ICP-MS | <5.0 | 14.5 | 5.1 | 7.4 | 12.5 |
| | Arithmetic Mean | 3.1 | 13.5 | 4.7 | 7.1 | 11.5 |
| | SD | 0.2 | 1.1 | 0.2 | 0.2 | 1.0 |
| | n | 3 | 4 | 4 | 4 | 4 |

| Blood Tungsten (µg/L) | | | | | | |
|------------------------------|--------|---------|---------|---------|---------|---------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 324 | ICP-MS | 4.1 | 12.8 | 1.7 | 10.6 | 7.6 |

| Blood Uranium (µg/L) | | | | | | |
|-----------------------------|---------------|----------|----------|----------|----------|----------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 103 | DRC/CC-ICP-MS | <0.00748 | <0.00748 | <0.00748 | <0.00748 | <0.00748 |
| 110 | ICP-MS | <0.02 | <0.02 | <0.02 | <0.02 | <0.02 |
| 147 | ICP-MS | <0.0071 | <0.0071 | <0.0071 | <0.0071 | <0.0071 |
| 312 | ICP-MS | <0.1 | <0.1 | <0.1 | <0.1 | <0.1 |

| Blood Vanadium (µg/L) | | | | | | |
|------------------------------|------------------------|------------|------------|------------|------------|-------------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 110 | DRC/CC-ICP-MS | 8.3 | 8.8 | 1.9 | 7.7 | 15.7 |
| 147 | DRC/CC-ICP-MS | 8.42 | 8.98 | 1.96 | 7.76 | 16.3 |
| 312 | DRC/CC-ICP-MS | 7.4 | 7.8 | 1.2 | 6.8 | 13 |
| 324 | ICP-MS | *19.8 | *18.0 | *9.3 | *16.6 | 31.5 |
| <i>*Outlier</i> | Arithmetic Mean | 8.0 | 8.5 | 1.7 | 7.4 | 19.1 |
| | SD | 0.6 | 0.6 | 0.4 | 0.5 | 8.4 |
| | n | 3 | 3 | 3 | 3 | 4 |

| Blood Zinc (µg/L) | | | | | | |
|--------------------------|------------------------|-------------|-------------|-------------|-------------|-------------|
| Lab Code | Method | BE15-06 | BE15-07 | BE15-08 | BE15-09 | BE15-10 |
| 110 | ICP-MS | 1883 | 2044 | 1968 | 2211 | 1863 |
| 114 | ICP-MS | 2010 | 2320 | 2030 | 2450 | 2060 |
| 147 | ICP-MS | 1830 | 2033 | 1941 | 2261 | 1856 |
| 197 | ICP-MS | 1570 | 1790 | 1630 | 1910 | 1690 |
| 206 | DRC/CC-ICP-MS | *197.8 | *219.5 | *196.7 | *239.6 | *200.4 |
| 312 | ICP-MS | 1690 | 1810 | 1540 | 1910 | 1590 |
| <i>*Outlier</i> | Arithmetic mean | 1797 | 1999 | 1822 | 2148 | 1812 |
| | SD | 171 | 215 | 221 | 235 | 180 |
| | n | 5 | 5 | 5 | 5 | 5 |

**New York State Department of Health
Trace Elements in Whole Blood
METHOD NOTES**

ATOMIC SPECTROMETRY METHODS

- A-1 ETAAS-Z (Electrothermal atomic absorption spectrometry with Zeeman background correction)
- A-2 ETAAS Other (i.e., D₂, S-H background correction)
- A-3 FAAS (Flame atomic absorption spectrometry)
- A-4 CV-AAS (Cold vapor atomic absorption spectrometry)
- A-5 HG-AAS (Hydride generation atomic absorption spectrometry)
- A-6 AFS (Atomic fluorescence spectrometry)

INDUCTIVELY COUPLED PLASMA

- P-1 ICP-MS (Inductively coupled plasma - mass spectrometry)
- P-2 DRC/CC-ICP-MS (ICP-MS used in the Dynamic Reaction Cell or Collision Cell mode)
- P-3 ICP-AES/OES (ICP atomic/optical emission spectrometry)
- P-4 HR-ICP-MS (High resolution ICP-MS)
- P-5 ETV-ICP-MS (Electrothermal vaporization ICP-MS)
- P-6 ID-ICP-MS (Isotope dilution ICP-MS)

ELECTROCHEMICAL METHODS

- E-1 ASV (Anodic stripping voltammetry without digestion)
- E-2 ASV-LeadCare® Blood Lead Testing System
- E-5 ASV-LeadCare® II Blood Lead Testing System
- E-6 ASV-LeadCare® Ultra™ Blood Lead Testing System
- E-3 Fluoride specific electrode

MOLECULAR FLUORIMETRY

- F-1 EtOAc (Ethyl acetate-acetic acid extraction method for determination of erythrocyte protoporphyrin)
- F-2 Aviv hematofluorometry (for determination of EP at hematocrit 35)
- F-3 Helena ZPP (for determination of zinc protoporphyrin in μmol ZPP/mol heme)

OTHER METHODS

If your method is not listed in the above list, please describe it briefly.
