
Wadsworth Center

NEW YORK STATE DEPARTMENT OF HEALTH

Trace Elements Laboratory

TRACE ELEMENTS IN URINE

Proficiency Test Report

Event #1, 2014

March 24th, 2014

March 24, 2014

Trace Elements in Urine

Dear Laboratory Director:

Results from the first proficiency test (PT) event for 2014 in the category Trace Elements in Urine have been tabulated and summarized. Target values for Arsenic, Cadmium, Mercury and Lead have been established along with acceptable ranges. Results are graded using element-specific criteria as indicated in each narrative section. A laboratory with an unacceptable significant analytical bias relative to the target value will be expected to investigate the source of the error. A confidential three-digit code number assigned by the PT program identifies participant laboratories.

PT Materials

The source of the test materials is human urine obtained from donor volunteers with informed consent. Urine was collected into polyethylene containers and then stored at 4°C. Following collection, urine from each donor was mixed and acidified to 1% v/v with nitric acid, and 1% (v/v) sulfamic acid solution containing 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100 was added to stabilize Hg. The urine was stored frozen at -80°C; after thawing at room temperature, precipitated salts were removed by centrifugation. The urine was separated into five pools and each was supplemented with different amounts of As, Cd, Hg and Pb as inorganic salts. Each pool was also spiked with additional trace elements that comprise the "NHANES suite" and include: Ba, Be, Co, Cs, Mo, Pt, Sb, Tl, U and W. Each pool was stirred for approximately 24 hours to ensure thorough mixing prior to aliquoting 10-mL samples into acid-leached polypropylene vials. Samples were stored at -80°C prior to circulating for proficiency testing.


The next PT event for trace elements in urine is scheduled to be mailed Wednesday, May 7th, 2014. Please inform our laboratory staff at (518) 474-7161 if the test materials have not arrived within five days of the scheduled mail out date. **The deadline for reporting results is Wednesday, June 4th, 2014.**

Thank you for your participation.

Sincerely,



Patrick J. Parsons, Ph.D.
Chief, Laboratory of Inorganic and Nuclear Chemistry
Deputy Director, Division of Environmental Health



Mary Frances Verostek, Ph.D.
Assistant Section Head
PT Program for Blood Lead /Trace Elements

New York State Department of Health
Event #1, 2014

Urine Arsenic

The source of the test materials is human urine obtained from donor volunteers with informed consent. Urine was collected into polyethylene containers and then stored at 4°C. Following collection, urine from each donor was mixed and acidified to 1% v/v with nitric acid, and 1% (v/v) sulfamic acid solution containing 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100 was added to stabilize Hg. The urine was stored frozen at -80°C; after thawing at room temperature, precipitated salts were removed by centrifugation. The urine was separated into five pools and each was supplemented with different amounts of inorganic As³⁺. Each pool was stirred for approximately 24 hours to ensure thorough mixing prior to aliquoting 10-mL samples into acid-leached polypropylene vials. Samples were stored at -80°C prior to circulating for proficiency testing.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for urine arsenic range from 24.2 µg/L (0.32 µmol/L) to 111.3 µg/L (1.49 µmol/L).

Acceptable ranges. The acceptable range is fixed at ±20% or ±6 µg/L for target values ≤30 µg/L. This provides a more realistic acceptability range at low concentrations of urine As, and the criteria are consistent with those in place for blood As.

Discussion. Based upon the above criteria, 96.7% of test results reported were judged as satisfactory, with one of the 24 participant laboratories (4.2%) reporting 2 or more of the 5 results outside the acceptable ranges.

**New York State Department of Health
Urine Arsenic Test Results, 2014 Event #1
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ urine)

	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
Robust Mean	62.4	79.8	24.2	47.4	111.3
Robust Standard Deviation	4.1	4.7	1.5	2.9	7.9
Standard Uncertainty	1.1	1.2	0.4	0.7	2.0
RSD (%)	6.7	5.9	6.1	6.2	7.1
Number of Sample Measurements	24	24	24	24	24
Acceptable Range:					
Upper Limit	74.9	95.8	30.2	56.9	133.6
Lower Limit	49.9	63.8	18.2	37.9	89.0

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Urine Arsenic Test Results, 2014 Event #1
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L urine)					Info Only
		UE14-01	UE14-02	UE14-03	UE14-04	UE14-05	
Target Values:		62.4	79.8	24.2	47.4	111.3	
103	DRC/CC-ICP-MS	63.1	79.1	24.0	47.1	112.1	Info
106	DRC/CC-ICP-MS	61.1	87.7	25.8	47.3	126	Info
107	DRC/CC-ICP-MS	63.1	78.8	23.9	47.0	112.7	Info
110	DRC/CC-ICP-MS	65.3	85.5	23.5	50.6	118.0	
116	DRC/CC-ICP-MS	62.1	79.7	24.0	46.5	112.2	Info
147	ICP-MS	59.3	74.8	23.8	44.2	98.1	Info
156	DRC/CC-ICP-MS	57.0	73.0	22.0	44.0	100.0	
164	ICP-MS	71.0	87.0	29.0	56.0	122.0	
179	ICP-MS	62.0	80.0	24.0	47.0	112.0	
197	DRC/CC-ICP-MS	60.0	79.0	24.0	45.0	110.0	
200	ICP-MS	67.6	99.3 ↑	28.8	60.9 ↑	144.0 ↑	Info
206	DRC/CC-ICP-MS	66.6	82.0	26.2	49.6	112.4	
208	ICP-MS	71.9	79.0	28.5	53.3	106.0	
293	DRC/CC-ICP-MS	59.8	76.4	23.1	45.6	108.6	Info
305	ICP-MS	60.0	76.8	23.1	44.3	118.5	
312	DRC/CC-ICP-MS	57.4	70.1	22.3	42.8	100.0	
324	HR-ICP-MS	63.3	82.4	24.9	49.0	116.2	Info
339	HR-ICP-MS	57.9	79.8	23.0	47.3	53.4 ↓	Info
359	ICP-MS	58.9	76.6	24.3	45.5	108.6	
366	ICP-MS	63.0	80.0	23.0	47.0	108.0	Info
367	DRC/CC-ICP-MS	67.4	84.2	25.3	49.1	115.7	Info
391	DRC/CC-ICP-MS	57.0	73.2	22.4	44.6	103.6	Info
401	DRC/CC-ICP-MS	64.0	82.8	24.9	49.8	116.8	Info
472	DRC/CC-ICP-MS	64.0	80.0	25.6	48.4	107.0	Info

Percent satisfactory results for all participants: 96.7 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
▼: Result unacceptable
▲: Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

**New York State Department of Health
Urine Arsenic Test Results, 2014 Event #1
STATISTICAL SUMMARY BY METHOD**

	Results ($\mu\text{g/L}$ urine)				
	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
DRC/CC-ICP-MS					
Number of Sample Measurements:	14	14	14	14	14
Mean:	62.0	79.4	24.1	47.0	111.1
Standard Deviation:	3.4	5.0	1.3	2.4	7.1
RSD (%):	5.5	6.3	5.6	5.0	6.4
HR-ICP-MS					
Number of Sample Measurements:	2	2	2	2	2
Mean:	60.6	81.1	24.0	48.2	84.8
Standard Deviation:	3.8	1.8	1.3	1.2	44.4
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	8	8	8	8	8
Mean:	64.2	81.7	25.6	49.8	114.7
Standard Deviation:	5.3	8.0	2.7	6.2	14.0
RSD (%):	8.2	9.8	10.5	12.5	12.2
All Laboratories					
Number of Sample Measurements:	24	24	24	24	24
Mean:	62.6	80.3	24.6	48.0	110.1
Standard Deviation:	4.1	5.9	2.0	4.1	15.4
RSD (%):	6.6	7.4	8.0	8.5	14.0

notes: ? Insufficient data for calculation.
A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #1, 2014

Urine Cadmium

The source of the test materials is human urine obtained from donor volunteers with informed consent. Urine was collected into polyethylene containers and then stored at 4°C. Following collection, urine from each donor was mixed and acidified to 1% v/v with nitric acid, and 1% (v/v) sulfamic acid solution containing 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100 was added to stabilize Hg. The urine was stored frozen at -80°C; after thawing at room temperature, precipitated salts were removed by centrifugation. The urine was separated into five pools and each was supplemented with different amounts of inorganic Cd²⁺. Each pool was stirred for approximately 24 hours to ensure thorough mixing prior to aliquoting 10-mL samples into acid-leached polypropylene vials. Samples were stored at -80°C prior to circulating for proficiency testing.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for urine cadmium range from 1.8 µg/L (16 nmol/L) to 13.3 µg/L (118 nmol/L).

Acceptable ranges. The acceptable range is fixed at ±15% or ±1 µg/L (9 nmol/L) around the target value whichever is greater. These criteria are used by the U.S. Occupational Safety and Health Administration (OSHA) to assess performance for occupational medicine.

Discussion. Based upon the above criteria, 98.3% of test results reported were judged as satisfactory, with none of the 24 participant laboratories reporting 2 or more of the 5 results outside the acceptable ranges.

**New York State Department of Health
Urine Cadmium Test Results, 2014 Event #1
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ urine)

	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
Robust Mean	13.3	3.3	4.9	9.9	1.8
Robust Standard Deviation	0.8	0.2	0.3	0.6	0.2
Standard Uncertainty	0.2	<0.1	0.1	0.2	<0.1
RSD (%)	5.8	5.5	6.3	6.1	9.0
Number of Sample Measurements	24	24	24	24	24
Acceptable Range:					
Upper Limit	15.3	4.3	5.9	11.4	2.8
Lower Limit	11.3	2.3	3.9	8.4	0.8

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Urine Cadmium Test Results, 2014 Event #1
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L urine)					Info Only
		UE14-01	UE14-02	UE14-03	UE14-04	UE14-05	
Target Values:		13.3	3.3	4.9	9.9	1.8	
103	DRC/CC-ICP-MS	13.5	3.3	4.9	10.1	1.8	Info
106	ICP-MS	12.9	3.1	4.9	9.6	1.7	Info
107	DRC/CC-ICP-MS	13.6	3.6	5.6	10.6	1.9	Info
110	ICP-MS	12.8	3.2	4.9	9.7	1.7	
116	ICP-MS	14.0	3.5	5.3	10.4	1.9	Info
147	ICP-MS	13.3	3.3	4.9	10.0	1.7	Info
156	ICP-MS	15.0	3.4	4.9	9.6	1.6	
164	ICP-MS	12.7	3.4	4.9	9.5	1.9	
179	ICP-MS	13.9	3.6	5.2	10.5	2.0	
197	DRC/CC-ICP-MS	12.1	3.2	4.7	9.0	1.7	
200	ICP-MS	13.2	3.4	4.8	9.7	1.8	Info
206	ICP-MS	13.6	3.3	5.0	9.8	1.8	
208	ICP-MS	13.1	3.2	5.0	9.7	1.6	
293	ICP-MS	13.7	3.4	5.2	10.3	1.8	Info
305	ICP-MS	13.8	3.4	5.2	10.6	1.8	
312	ICP-MS	13.7	3.3	4.9	10.1	1.7	
324	HR-ICP-MS	13.7	3.5	5.1	10.3	1.8	Info
339	HR-ICP-MS	11.8	2.9	4.3	9.2	3.0 ↑	Info
359	ICP-MS	14.2	3.7	5.4	10.5	2.2	
366	ICP-MS	13.0	3.0	4.6	9.5	1.5	Info
367	DRC/CC-ICP-MS	15.4 ↑	3.8	5.7	11.3	2.0	Info
391	DRC/CC-ICP-MS	12.7	3.3	4.7	9.5	1.8	Info
401	DRC/CC-ICP-MS	12.0	2.9	4.5	9.2	1.6	Info
472	ICP-MS	11.8	3.0	4.5	8.8	1.6	Info

Percent satisfactory results for all participants: 98.3 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
⚠ Result unacceptable
▲ Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

New York State Department of Health
Urine Cadmium Test Results, 2014 Event #1
STATISTICAL SUMMARY BY METHOD

	Results ($\mu\text{g/L}$ urine)				
	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
DRC/CC-ICP-MS					
Number of Sample Measurements:	6	6	6	6	6
Mean:	13.2	3.4	5.0	10.0	1.8
Standard Deviation:	1.3	0.3	0.5	0.9	0.1
RSD (%):	9.6	9.4	10.1	8.9	7.9
HR-ICP-MS					
Number of Sample Measurements:	2	2	2	2	2
Mean:	12.8	3.2	4.7	9.8	2.4
Standard Deviation:	1.3	0.4	0.6	0.8	0.8
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	16	16	16	16	16
Mean:	13.4	3.3	5.0	9.9	1.8
Standard Deviation:	0.7	0.2	0.2	0.5	0.2
RSD (%):	5.5	5.9	4.8	4.9	9.8
All Laboratories					
Number of Sample Measurements:	24	24	24	24	24
Mean:	13.3	3.3	5.0	9.9	1.8
Standard Deviation:	0.9	0.2	0.3	0.6	0.3
RSD (%):	6.8	7.1	6.8	6.0	16.1

notes: ? Insufficient data for calculation.
A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #1, 2014

Urine Mercury

The source of the test materials is human urine obtained from donor volunteers with informed consent. Urine was collected into polyethylene containers and then stored at 4°C. Following collection, urine from each donor was mixed and acidified to 1% v/v with nitric acid, and 1% (v/v) sulfamic acid solution containing 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100 was added to stabilize Hg. The urine was stored frozen at -80°C; after thawing at room temperature, precipitated salts were removed by centrifugation. The urine was separated into five pools and each was supplemented with different amounts of inorganic Hg. Each pool was stirred for approximately 24 hours to ensure thorough mixing prior to aliquoting 10-mL samples into acid-leached polypropylene vials. Samples were stored at -80°C prior to circulating for proficiency testing.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for urine mercury range from 18.5 µg/L (92 nmol/L) to 83.0 µg/L (414 nmol/L).

Acceptable ranges. The acceptable range is fixed at ±30% or ±3 µg/L (15 nmol/L) for target values ≤10 µg/L. The criteria are consistent with those in place for blood Hg.

Discussion. Based upon the above criteria, 93.9% of test results reported were judged as satisfactory, with one of the 23 participant laboratories (4.3%) reporting 2 or more of the 5 results outside the acceptable ranges.

**New York State Department of Health
Urine Mercury Test Results, 2014 Event #1
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ urine)

	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
Robust Mean	18.5	25.3	61.5	39.8	83.0
Robust Standard Deviation	2.2	2.5	6.2	5.0	8.5
Standard Uncertainty	0.6	0.7	1.6	1.3	2.2
RSD (%)	11.8	9.9	10.0	12.7	10.3
Number of Sample Measurements	23	23	23	23	23
Acceptable Range:					
Upper Limit	24.1	32.9	80.0	51.7	107.9
Lower Limit	12.9	17.7	43.0	27.9	58.1

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Urine Mercury Test Results, 2014 Event #1
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L urine)					Info Only
		UE14-01	UE14-02	UE14-03	UE14-04	UE14-05	
Target Values:		18.5	25.3	61.5	39.8	83.0	
103	DRC/CC-ICP-MS	17.7	24.3	59.9	38.5	81.3	Info
107	DRC/CC-ICP-MS	22.2	29.9	68.4	48.8	95.6	Info
109	AFS	21.0	30.2	71.4	44.0	92.7	Info
110	ICP-MS	19.5	27.1	65.4	42.3	88.0	
147	CV-AAS	17.4	24.5	59.2	38.0	78.0	Info
156	ICP-MS	20.0	26.0	64.0	43.0	85.0	
164	ICP-MS	22.0	28.0	67.0	45.0	90.0	
179	ICP-MS	17.0	22.0	55.0	35.0	72.0	
197	DRC/CC-ICP-MS	20.0	28.0	66.0	44.0	89.0	
199	ICP-MS	18.0	24.6	61.3	38.7	81.3	Info
200	ICP-MS	15.0	20.6	52.6	34.0	67.0	Info
206	ICP-MS	19.0	27.0	64.0	42.0	85.0	
208	CV-AAS	16.1	23.9	57.3	36.3	75.5	
293	ICP-MS	18.3	24.2	60.3	38.9	80.6	Info
305	ICP-MS	25.2 ↑	27.9	67.4	45.5	88.3	
312	ICP-MS	18.2	24.0	61.2	39.7	78.0	
324	AFS	18.6	26.2	61.2	38.2	80.6	Info
339	HR-ICP-MS	13.8	24.4	55.0	31.2	93.1	Info
359	ICP-MS	12.3 ↓	17.9	49.1	34.1	76.3	
366	ICP-MS	19.0	25.0	59.0	39.0	87.0	Info
367	CV-AAS	17.9	23.4	55.0	34.1	65.0	Info
391	DRC/CC-ICP-MS	38.0 ↑	52.2 ↑	133.2 ↑	84.1 ↑	177.9 ↑	Info
401	DRC/CC-ICP-MS	17.5	23.5	61.1	38.1	78.4	Info

Percent satisfactory results for all participants: 93.9 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
▴ Result unacceptable
▾ Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

New York State Department of Health
Urine Mercury Test Results, 2014 Event #1
STATISTICAL SUMMARY BY METHOD

	Results ($\mu\text{g/L}$ urine)				
	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
AFS					
Number of Sample Measurements:	2	2	2	2	2
Mean:	19.8	28.2	66.3	41.1	86.7
Standard Deviation:	1.7	2.8	7.2	4.1	8.6
RSD (%):	—	—	—	—	—
CV-AAS					
Number of Sample Measurements:	3	3	3	3	3
Mean:	17.1	23.9	57.2	36.1	72.8
Standard Deviation:	0.9	0.6	2.1	2.0	6.9
RSD (%):	—	—	—	—	—
DRC/CC-ICP-MS					
Number of Sample Measurements:	4	4	4	4	4
Mean:	19.4	26.4	63.9	42.4	86.1
Standard Deviation:	2.2	3.0	4.0	5.1	7.8
RSD (%):	11.4	11.5	6.3	12.0	9.0
HR-ICP-MS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	13.8	24.4	55.0	31.2	93.1
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	12	12	12	12	12
Mean:	18.6	24.5	60.5	39.8	81.5
Standard Deviation:	3.2	3.1	5.8	4.0	7.1
RSD (%):	17.3	12.6	9.5	10.0	8.7
All Laboratories					
Number of Sample Measurements:	22	22	22	22	22
Mean:	18.4	25.1	60.9	39.5	82.2
Standard Deviation:	2.8	2.9	5.5	4.4	8.1
RSD (%):	15.3	11.5	9.1	11.3	9.9

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #1, 2014

Urine Lead

The source of the test materials is human urine obtained from donor volunteers with informed consent. Urine was collected into polyethylene containers and then stored at 4°C. Following collection, urine from each donor was mixed and acidified to 1% v/v with nitric acid, and 1% (v/v) sulfamic acid solution containing 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100 was added to stabilize Hg. The urine was stored frozen at -80°C; after thawing at room temperature, precipitated salts were removed by centrifugation. The urine was separated into five pools and each was supplemented with different amounts of inorganic Pb²⁺. Each pool was stirred for approximately 24 hours to ensure thorough mixing prior to aliquoting 10-mL samples into acid-leached polypropylene vials. Samples were stored at -80°C prior to circulating for proficiency testing.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for urine lead range from 23.9 µg/L (0.12 µmol/L) to 188.4 µg/L (0.91 µmol/L).

Acceptable ranges. The acceptable range is fixed at ±10% or ±40 µg/L (0.19 µmol/L) around the target value, whichever is greater. These criteria are consistent with those established under CLIA '88 (Federal Register Volume 57, Number 40, §§ 493.2 and 493.937, February 28, 1992) for blood lead.

Discussion. Based upon the above criteria, 97.3% of test results were judged as satisfactory, with one of the 22 participant laboratories (4.5%) reporting 2 or more of the 5 results outside the acceptable ranges.

**New York State Department of Health
Urine Lead Test Results, 2014 Event #1
ROBUST STATISTICAL SUMMARY**

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ urine)

	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
Robust Mean	188.4	46.7	70.2	140.9	23.9
Robust Standard Deviation	14.4	2.9	4.2	9.6	2.1
Standard Uncertainty	3.8	0.8	1.1	2.6	0.6
RSD (%)	7.7	6.3	6.0	6.8	8.9
Number of Sample Measurements	22	22	22	22	22
Acceptable Range:					
Upper Limit	228.4	86.7	110.2	180.9	63.9
Lower Limit	148.4	6.7	30.2	100.9	0.0

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Urine Lead Test Results, 2014 Event #1
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L urine)					Info Only
		UE14-01	UE14-02	UE14-03	UE14-04	UE14-05	
Target Values:		188.4	46.7	70.2	140.9	23.9	
103	DRC/CC-ICP-MS	199.0	48.9	74.1	149.0	25.2	Info
106	ICP-MS	190	46.7	71.1	143	24.1	Info
107	DRC/CC-ICP-MS	201	50.6	75.8	151	25.9	Info
110	ICP-MS	187.0	45.8	69.1	140.0	23.1	
116	ICP-MS	193.4	47.6	72.2	144.7	24.4	Info
147	ICP-MS	193.3	48.1	71.1	143.6	24.2	Info
156	DRC/CC-ICP-MS	180.0	45.0	67.0	140.0	22.0	
164	ICP-MS	211.0	53.0	74.0	149.0	28.0	
179	ICP-MS	195.0	48.0	74.0	146.0	25.0	
197	DRC/CC-ICP-MS	193.0	49.3	71.4	148.7	23.8	
200	ICP-MS	176.0	43.1	67.8	131.0	22.1	Info
206	ICP-MS	172.0	45.0	69.0	136.0	21.0	
208	ICP-MS	194.0	46.5	70.8	141.1	24.1	
293	ICP-MS	193.1	48.3	71.7	144.0	24.2	Info
305	ICP-MS	185.8	44.8	66.8	132.6	23.0	
312	ICP-MS	212.0	49.7	78.2	157.0	26.4	
324	HR-ICP-MS	178.2	42.7	64.6	131.3	21.7	Info
339	HR-ICP-MS	142.2 ↓	45.6	67.3	116.2	45.6	Info
359	ICP-MS	172.9	42.9	64.3	129.7	21.9	
366	ICP-MS	205.0	50.0	74.0	152.0	26.0	Info
391	ETAAS-Z	105.9 ↓	30.9	45.4	85.7 ↓	15.4	Info
472	ICP-MS	179.9	45.4	67.7	137.2	23.4	Info

Percent satisfactory results for all participants: 97.3 %

notes: ↑ Reported outside upper limit
↓ Reported outside lower limit
▼ Result unacceptable
▲ Result not reported

notes: Results reported as less than the method detection limit are excluded from statistical calculations.
Info only: results included for informational purposes only.

**New York State Department of Health
Urine Lead Test Results, 2014 Event #1
STATISTICAL SUMMARY BY METHOD**

	Results ($\mu\text{g/L}$ urine)				
	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
DRC/CC-ICP-MS					
Number of Sample Measurements:	4	4	4	4	4
Mean:	193.3	48.5	72.1	147.2	24.2
Standard Deviation:	9.5	2.4	3.8	4.9	1.7
RSD (%):	4.9	5.0	5.3	3.3	7.1
ETAAS-Z					
Number of Sample Measurements:	1	1	1	1	1
Mean:	105.9	30.9	45.4	85.7	15.4
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
HR-ICP-MS					
Number of Sample Measurements:	2	2	2	2	1
Mean:	160.2	44.2	66.0	123.8	21.7
Standard Deviation:	25.5	2.1	1.9	10.7	?
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	15	15	15	15	15
Mean:	190.7	47.0	70.8	141.8	24.1
Standard Deviation:	12.4	2.7	3.5	7.7	1.8
RSD (%):	6.5	5.7	4.9	5.4	7.5
All Laboratories					
Number of Sample Measurements:	22	22	22	22	21
Mean:	184.5	46.3	69.4	138.6	23.6
Standard Deviation:	23.2	4.3	6.5	15.0	2.6
RSD (%):	12.6	9.4	9.3	10.8	10.8

notes: ? Insufficient data for calculation.

A Standard Deviation displayed as 0.0 should be interpreted as <0.1

New York State Department of Health
Event #1, 2014

Additional Trace Elements Reported in Urine

Participating laboratories reported analytical results for any other elements that are routinely reported in order to characterize these materials more completely. Results and descriptive statistics are provided for reference purposes. No target value or acceptable range is implied. As, Cd, and Pb were spiked using a stock standard containing all elements in the National Health and Nutrition Examination Survey (NHANES) conducted by the Centers for Disease Control and Prevention. Refer to www.cdc.gov/exposurereport for more information on recent NHANES data for these elements in urine. In addition, these samples were spiked with leading elements present in other proficiency testing programs. The following table shows the additional elements spiked in the samples.

<u>NHANES Elements</u>	<u>Additional Elements</u>
Ba	Al
Be	Cr
Co	Cu
Cs	Mn
Mo	Ni
Pt	Se
Sb	Sn
Tl	Te
U	V
W	Zn

**New York State Department of Health
Urine Additional Elements, 2014 Event #1
Page 1**

Urine Aluminum (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
147	DRC/CC-ICP-MS	105	25.1	43.4	77.4	20.8
164	ICP-MS	93	24	36	70	13
179	DRC/CC-ICP-MS	95	24	37	72	12
197	ICP-MS	88	<20	29	65	<20
206	ICP-MS	>20.0	*19.0	>20.0	>20.0	12.0
305	ICP-MS	103	26	43	78	18
312	ICP-MS	97.6	25.4	40.2	76.6	17.4
324	HR-ICP-MS	88.1	23.6	33.8	66.4	13.7
359	ICP-MS	86.8	26.7	36.9	65	15.6
*Outlier	Arithmetic Mean	95	25	37	71	15
	SD	7	1	5	6	3
	n	8	7	8	8	8

Urine Antimony (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	12.1	2.94	4.42	9.04	1.56
110	ICP-MS	12.6	3.11	4.67	9.43	1.56
147	ICP-MS	12.2	2.95	4.43	9.11	1.49
197	ICP-MS	13.7	3.4	4.8	9.9	1.7
206	ICP-MS	11.0	3.0	4.0	9.0	*1.0
312	ICP-MS	13.1	3	4.7	9.3	1.6
359	ICP-MS	11.1	2.7	4	8.3	1.4
472	ICP-MS	11.9	2.9	4.4	8.8	1.6
*Outlier	Arithmetic Mean	12.2	3.0	4.4	9.1	1.6
	SD	0.9	0.2	0.3	0.5	0.1
	n	8	8	8	8	7

Urine Barium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
106	ICP-MS	16.2	4.1	6.2	12.2	2.2
107	DRC/CC-ICP-MS	16.2	4.11	6.14	12.3	2.27
110	ICP-MS	15.8	4.02	6.08	11.9	2.59
116	ICP-MS	15.9	3.94	5.94	11.7	2.00
147	ICP-MS	15.1	3.86	5.56	11.3	1.99
197	ICP-MS	16.9	4.2	6.5	12.9	2.2
312	ICP-MS	17.6	4.2	6.6	13.2	2.4
359	ICP-MS	12.9	*3.1	4.7	9.8	1.6
472	ICP-MS	15.1	3.9	5.8	11.4	2.2
*Outlier	Arithmetic Mean	16	4.0	5.9	12	2.2
	SD	1	0.1	0.6	1	0.3
	n	9	8	9	9	9

Urine Beryllium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
106	ICP-MS	15.5	4.0	6.1	12.4	2.1
107	DRC/CC-ICP-MS	16.3	4.07	6.49	12.9	2.18
110	ICP-MS	15.9	4.00	5.78	11.9	2.22
116	ICP-MS	16.1	3.91	6.32	12.3	1.97
147	ICP-MS	16.5	3.94	6.33	12.1	2.00
197	ICP-MS	14.2	3.4	5.2	10.7	1.7
312	ICP-MS	*20.5	4.4	6.9	*16.3	2.5
472	ICP-MS	15.4	3.8	5.8	11.9	1.9
*Outlier	Arithmetic Mean	16	3.9	6.1	12	2.1
	SD	1	0.3	0.5	1	0.2
	n	7	8	8	7	8

**New York State Department of Health
Urine Additional Elements, 2014 Event #1
Page 2**

Urine Bismuth (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
147	ICP-MS	<0.104	<0.104	<0.104	<0.104	<0.104
197	ICP-MS	<1.0	<1.0	<1.0	<1.0	<1.0
206	ICP-MS	0.0	0.0	0.0	0.0	0.0
305	ICP-MS	<0.5	<0.5	<0.5	<0.5	<0.5
312	ICP-MS	<0.08	<0.08	<0.08	<0.08	<0.08

Urine Cesium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	66.0	16.9	25.3	50.2	9.49
110	ICP-MS	65.7	16.9	25.0	49.3	9.20
147	ICP-MS	61.8	16.3	23.5	46.8	8.71
312	ICP-MS	70.8	17.4	26.5	52.7	9.9
359	ICP-MS	47.4	*13.6	19.9	39.9	7.5
472	ICP-MS	62.5	16.1	23.9	47.1	8.9
<i>*Outlier</i>	Arithmetic Mean	62	16.7	24	48	9.0
	SD	8	0.5	2	4	0.8
	n	6	5	6	6	6

Urine Chromium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	32.3	8.0	11.7	24.1	4.3
110	DRC/CC-ICP-MS	32.4	8.25	12.3	24.3	4.41
147	DRC/CC-ICP-MS	30.0	7.64	11.3	22.5	3.93
156	DRC/CC-ICP-MS	28	7.2	10	21	3.7
164	DRC/CC-ICP-MS	30.9	7.9	11.6	22.9	4.0
179	DRC/CC-ICP-MS	30.2	7.5	11.3	22.8	3.9
197	DRC/CC-ICP-MS	31.5	8.1	12.1	24.5	4.3
206	DRC/CC-ICP-MS	29.3	7.8	11.6	22.8	4.0
305	ICP-MS	31.9	8.0	12.2	24.2	4.0
312	DRC/CC-ICP-MS	*41.9	7.3	11.1	23.1	3.8
359	ICP-MS	28.2	8.5	11.8	21.8	*5.4
366	DRC/CC-ICP-MS	30.5	7.4	10.5	21.5	3.6
391	DRC/CC-ICP-MS	26.89	7.0	10.4	21.2	3.8
401	DRC/CC-ICP-MS	28.6	7.1	10.7	21.9	3.6
<i>*Outlier</i>	Arithmetic Mean	30	7.7	11.3	23	3.9
	SD	2	0.5	0.7	1	0.3
	n	13	14	14	14	13

Urine Cobalt (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	6.64	6.88	3.75	5.16	1.49
110	ICP-MS	6.67	6.88	3.60	4.96	1.42
147	ICP-MS	6.54	7.01	3.51	4.96	1.36
156	DRC/CC-ICP-MS	5.6	6.1	3.2	4.4	1.2
164	ICP-MS	6.2	6.7	3.6	5.0	1.4
179	DRC/CC-ICP-MS	6.4	6.9	3.4	4.9	1.4
197	ICP-MS	6.7	7.0	3.7	5.2	1.4
206	ICP-MS	5.6	6.2	3.1	4.6	1.2
305	ICP-MS	6.8	7.3	3.8	5.4	1.4
312	ICP-MS	6.9	7.0	3.8	5.2	1.5
324	HR-ICP-MS	6.22	6.49	3.38	4.71	1.31
359	ICP-MS	5.6	5.7	3	4.2	1.2
366	ICP-MS	6.1	6.6	3.4	4.9	1.3
391	DRC/CC-ICP-MS	5.9	6.2	3.3	4.7	1.3
401	DRC/CC-ICP-MS	6.1	6.3	3.3	4.7	1.3
472	ICP-MS	6.6	6.8	3.6	4.9	1.5
	Arithmetic Mean	6.3	6.6	3.5	4.9	1.4
	SD	0.4	0.4	0.2	0.3	0.1
	n	16	16	16	16	16

**New York State Department of Health
Urine Additional Elements, 2014 Event #1
Page 3**

Urine Copper (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
110	ICP-MS	654	165	245	488	86.1
147	ICP-MS	661	166	244	487	85.1
164	ICP-MS	593	147	220	445	76
179	DRC/CC-ICP-MS	650	163	241	485	84
197	ICP-MS	742.9	195.9	286.6	579.1	104.4
206	ICP-MS	604.0	151.0	248.0	502.0	81.0
305	ICP-MS	605	158	241	479	82
312	ICP-MS	666	163	243	488	86.9
324	HR-ICP-MS	648	160	243	485	83.5
359	ICP-MS	501.8	138.3	185.9	366.8	71.6
391	DRC/CC-ICP-MS	*343.0	*90.5	*135.3	*265.7	62.1
401	DRC/CC-ICP-MS	613.1	154.5	230.9	461.1	80.8
<i>*Outlier</i>	Arithmetic Mean	631	160	239	479	82
	SD	60	15	24	50	10
	n	11	11	11	11	12

Urine Iodine (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	45.8	45.8	46.3	45.3	46.0
164	DRC/CC-ICP-MS	43	43	43	42	43
179	ICP-MS	43	43	43	43	43
197	ICP-MS	*16	*19	*19	*19	*21
206	ICP-MS	38.1	40.2	44.9	40.7	41.0
312	ICP-MS	46.1	46.4	45.7	46.3	48.7
<i>*Outlier</i>	Arithmetic Mean	43	44	45	43	44
	SD	3	2	2	2	3
	n	5	5	5	5	5

Urine Iron (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
391	DRC/CC-ICP-MS	6.4	11.7	12.3	7.3	11.8

Urine Lithium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
147	ICP-MS	4.69	4.91	4.71	4.77	4.73

Urine Manganese (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
103	DRC/CC-ICP-MS	15.6	3.9	5.9	11.7	2.1
110	DRC/CC-ICP-MS	15.8	4.19	6.00	12.1	2.01
179	DRC/CC-ICP-MS	17.0	4.1	6.2	12.3	2.2
206	ICP-MS	13.9	3.9	5.5	11.4	2.1
305	ICP-MS	16.0	4.0	6.2	12.6	2.3
312	DRC/CC-ICP-MS	15.6	4.1	6.0	11.4	2.3
324	HR-ICP-MS	15.4	3.75	5.72	11.5	1.93
359	ICP-MS	11.4	3.2	4.6	*8.7	1.9
391	DRC/CC-ICP-MS	13.1	3.4	5.1	10.5	1.8
<i>*Outlier</i>	Arithmetic Mean	15	3.8	5.7	11.7	2.1
	SD	2	0.3	0.5	0.7	0.2
	n	9	9	9	8	9

**New York State Department of Health
Urine Additional Elements, 2014 Event #1
Page 4**

Urine Molybdenum (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	180	63.2	83.3	140	45.1
110	ICP-MS	180	62.9	82.7	141	44.1
147	ICP-MS	174	66.0	80.0	134	43.5
179	ICP-MS	173	61	79	136	43
197	ICP-MS	*233.3	*83.5	*101.8	*172.3	*58.6
312	ICP-MS	180	64	84.5	144	45.3
324	HR-ICP-MS	192	66.7	88.1	150	47.4
359	ICP-MS	154.7	55.2	72.5	124.5	38.6
472	ICP-MS	179.0	62.9	82.6	141.0	45.1
*Outlier	Arithmetic Mean	177	63	82	139	44
	SD	11	4	5	8	3
	n	8	8	8	8	8

Urine Nickel (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	16.8	5.9	7.3	12.8	2.7
110	ICP-MS	19.2	*8.92	*10.1	16.0	*5.77
147	DRC/CC-ICP-MS	16.9	6.69	7.52	13.3	2.67
164	ICP-MS	16.0	5.0	6.1	11.7	2.5
179	DRC/CC-ICP-MS	17.3	5.5	6.8	12.5	2.8
197	ICP-MS	17.3	5.7	7.1	13.6	2.7
206	ICP-MS	17.2	6.5	6.3	13.0	3.2
312	ICP-MS	21.0	6.2	6.9	12.0	2.9
324	HR-ICP-MS	16.4	5.42	6.89	12.3	2.51
359	ICP-MS	10.2	3.7	4.7	7.8	2.1
391	DRC/CC-ICP-MS	13.9	4.9	6.0	11.1	2.6
401	DRC/CC-ICP-MS	14.3	5.1	6.1	11.2	2.6
*Outlier	Arithmetic Mean	16	5.5	6.5	12	2.7
	SD	3	0.8	0.8	2	0.3
	n	12	11	11	12	11

Urine Platinum (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	6.33	1.59	2.36	4.71	0.81
110	ICP-MS	6.35	1.51	2.31	4.71	0.729
147	ICP-MS	6.11	1.48	2.24	4.62	0.747
312	ICP-MS	6.5	1.6	2.4	4.9	0.8
359	ICP-MS	5.5	<2.7	<2.7	4	<2.7
472	ICP-MS	6.1	1.5	2.2	4.5	0.8
	Arithmetic Mean	6.1	1.54	2.30	4.6	0.78
	SD	0.4	0.06	0.08	0.3	0.04
	n	6	5	5	6	5

Urine Selenium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
110	DRC/CC-ICP-MS	163	53.9	66.5	129	35.0
147	ICP-MS	154	46.8	63.6	114	29.4
179	DRC/CC-ICP-MS	170	52	72	133	33
206	ICP-MS	149.0	53.0	67.0	126.0	36.0
305	ICP-MS	174	49	65	132	35
312	ICP-MS	159	49.1	68.4	119	32.1
324	HR-ICP-MS	*213	*65.8	*88.5	*160	40.4
359	ICP-MS	152.8	53.0	67.4	117.8	33.2
*Outlier	Arithmetic Mean	160	51	67	124	34
	SD	9	3	3	7	3
	n	7	7	7	7	8

**New York State Department of Health
Urine Additional Elements, 2014 Event #1
Page 5**

Urine Silver (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
147	ICP-MS	2.23	2.20	2.28	1.93	2.29

Urine Strontium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	39.4	39.4	40.2	39.8	40.6
472	ICP-MS	38.3	38.3	38.8	38.8	39.6

Urine Tellurium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
110	ICP-MS	12.9	3.33	4.93	9.65	1.85
197	ICP-MS	12.8	3.1	4.7	9.2	1.5
206	ICP-MS	>10.0	3.0	4.0	9.0	1.0
312	ICP-MS	13.1	3.2	4.7	9.3	1.6
	Arithmetic Mean	12.9	3.2	4.6	9.3	1.5
	SD	0.2	0.1	0.4	0.3	0.4
	n	3	4	4	4	4

Urine Thallium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
106	ICP-MS	31.9	7.8	12.0	23.9	4.0
107	DRC/CC-ICP-MS	32.9	8.27	12.4	24.6	4.25
110	ICP-MS	31.3	7.76	11.6	23.5	3.96
116	ICP-MS	31.3	7.67	11.7	23.6	3.94
147	ICP-MS	31.3	7.85	11.6	23.5	3.95
179	ICP-MS	31	8	12	24	4
197	ICP-MS	30.9	7.7	11.7	23.4	3.9
206	ICP-MS	>10.0	7.2	>10.0	>10.0	3.7
305	ICP-MS	30.3	7.4	11.5	22.3	3.9
312	ICP-MS	32.9	7.8	12.2	24.3	4.0
324	HR-ICP-MS	31.7	7.58	11.5	23.4	3.81
359	ICP-MS	29	7.2	*10.4	*21	3.7
472	ICP-MS	29.7	7.6	11.3	22.8	3.9
<i>*Outlier</i>	Arithmetic Mean	31	7.7	11.8	23.6	3.9
	SD	1	0.3	0.3	0.6	0.1
	n	12	13	11	11	13

Urine Thorium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
147	ICP-MS	<0.0046	<0.0046	<0.0046	<0.0046	<0.0046

Urine Tin (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
107	DRC/CC-ICP-MS	33.5	8.29	12.5	25.5	4.45
110	ICP-MS	31.8	7.90	11.7	23.9	4.08
147	ICP-MS	30.2	7.90	11.7	22.9	4.01
312	ICP-MS	31.3	8.1	12.1	23.3	4.2
359	ICP-MS	26.7	6.8	9.8	20.6	3.6
472	ICP-MS	30.3	7.6	11.3	22.9	4.0
	Arithmetic Mean	31	7.8	11.5	23	4.1
	SD	2	0.5	0.9	2	0.3
	n	6	6	6	6	6

**New York State Department of Health
Urine Additional Elements, 2014 Event #1
Page 6**

Urine Tungsten (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
103	DRC/CC-ICP-MS	12.3	3.0	4.6	9.1	1.5
107	DRC/CC-ICP-MS	12.2	3.07	4.57	9.11	1.58
110	ICP-MS	12.7	3.13	4.72	9.56	1.62
147	ICP-MS	12.0	3.27	4.85	9.05	1.79
312	ICP-MS	14.1	3.4	5.3	10.2	1.7
324	HR-ICP-MS	10.4	2.63	3.92	7.67	1.41
359	ICP-MS	10.4	2.7	4.1	8.2	1.3
472	ICP-MS	11.6	2.9	4.5	8.9	1.6
	Arithmetic Mean	12	3.0	4.6	9.0	1.6
	SD	1	0.3	0.4	0.8	0.2
	n	8	8	8	8	8

Urine Uranium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
106	ICP-MS	1.6	0.4	0.6	1.2	0.2
107	DRC/CC-ICP-MS	1.70	0.43	0.65	1.29	0.22
110	ICP-MS	1.51	0.371	0.555	1.13	0.188
116	ICP-MS	1.48	0.367	0.544	1.10	0.184
147	ICP-MS	1.57	0.393	0.600	1.20	0.202
197	ICP-MS	1.4	<1.0	<1.0	1.0	<1.0
312	ICP-MS	1.7	0.4	0.6	1.3	0.2
324	HR-ICP-MS	1.37	0.335	0.507	1.02	0.173
359	ICP-MS	1.5	0.3	0.6	1.1	0.2
472	ICP-MS	1.4	0.4	0.6	1.1	0.2
	Arithmetic Mean	1.5	0.38	0.58	1.1	0.20
	SD	0.1	0.04	0.04	0.1	0.01
	n	10	9	9	10	9

Urine Vanadium (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
147	DRC/CC-ICP-MS	12.1	3.17	4.58	9.29	1.55
179	DRC/CC-ICP-MS	11.6	2.8	4.2	8.4	1.5
197	ICP-MS	12.50	3.00	4.70	9.40	1.60
312	DRC/CC-ICP-MS	9.4	2.1	3.7	7.3	1.1
324	HR-ICP-MS	*34.2	8.24	*12.6	*25.9	4.43
359	ICP-MS	10.4	2	3.3	7.6	1.4
391	DRC/CC-ICP-MS	13.0	6.1	7.9	12.2	5.6
<i>*Outlier</i>	Arithmetic Mean	12	4	5	9	2
	SD	1	2	2	2	2
	n	6	7	6	6	7

Urine Zinc (µg/L)						
Lab Code	Method	UE14-01	UE14-02	UE14-03	UE14-04	UE14-05
110	ICP-MS	710	238	318	560	163
147	ICP-MS	804	265	352	625	174
164	ICP-MS	679	241	301	531	151
179	DRC/CC-ICP-MS	767	256	324	597	165
197	ICP-MS	698	234	313	558	<200
206	ICP-MS	680.0	220.0	330.0	560.0	150.0
305	ICP-MS	677	225	307	565	142
312	ICP-MS	740	231	305	568	167
324	HR-ICP-MS	736	237	319	586	157
359	ICP-MS	669.2	216.8	284.2	507.8	141.9
391	DRC/CC-ICP-MS	*429.5	*143.4	*189.4	*347.4	*94.6
401	DRC/CC-ICP-MS	693.2	228.9	313.9	542.8	150.4
<i>*Outlier</i>	Arithmetic Mean	714	236	315	564	156
	SD	43	14	17	32	11
	n	11	11	11	11	10

**New York State Department of Health
Trace Elements in Urine
METHOD NOTES**

ATOMIC SPECTROMETRY METHODS

- A-1 ETAAS-Z (Electrothermal atomic absorption spectrometry with Zeeman background correction)
- A-2 ETAAS other (i.e., D₂, S-H background correction)
- A-3 FAAS (Flame atomic absorption spectrometry)
- A-4 CV-AAS (Cold vapor atomic absorption spectrometry)
- A-5 HG-AAS (Hydride generation atomic absorption spectrometry)
- A-6 AFS (Atomic fluorescence spectrometry)
- A-7 Other

INDUCTIVELY COUPLED PLASMA

- P-1 ICP-MS (Inductively coupled plasma - mass spectrometry)
- P-2 DRC/CC-ICP-MS (ICP-MS used in the Dynamic Reaction Cell or Collision Cell mode)
- P-3 ICP-AES/OES (ICP atomic/optical emission spectrometry)
- P-4 HR-ICP-MS (High resolution ICP-MS)
- P-5 ETV-ICP-MS (Electrothermal vaporization ICP-MS)
- P-6 ID-ICP-MS (Isotope dilution ICP-MS)
- P-7 Other

ELECTROCHEMICAL METHODS

- E-1 ASV (Anodic stripping voltammetry without digestion)
- E-2 ASV-LeadCare[®] (Anodic stripping voltammetry using the ESA LeadCare[®] system)
- E-3 Fluoride specific electrode
- E-4 Other

MOLECULAR FLUORIMETRY

- F-1 EtOAc (Ethyl acetate-acetic acid extraction method for determination of erythrocyte protoporphyrin)
- F-2 Aviv hematofluorometry (for determination of EP at hematocrit 35)
- F-3 Helena ZPP (for determination of zinc protoporphyrin in $\mu\text{mol ZPP/mol heme}$)
- F-4 Other

OTHER METHODS

If your method is not listed in the above list, please describe it briefly.⁵⁰
