Toxicology
Clinical Toxicology

Standard	Guidance
The following specialty sustaining standards of practices shall be incorporated into the laboratory's quality management system, where applicable to the scope of services provided.	
Clinical Toxicology Standard 1 (CT S1) Comprehensive Toxicology laboratories shall maintain, on-site, acceptable methods for the confirmation of presumptive positive drug screens.	The analytical methods used to confirm presumptive positives are based on chemical and physical principles that differ from those used in the screening procedure.
Clinical Toxicology Standard 2 (CT S2) For substance of abuse testing, reports of unconfirmed drug screen findings shall clearly state that the positive findings are unconfirmed.	
Clinical Toxicology Standard 3 (CT S3) For substance of abuse testing, reports shall state the assay cutoff levels used to discern positive from negative results.	