



**Department
of Health**

**Wadsworth
Center**

TRACE ELEMENTS IN SERUM

Proficiency Test Report

Event #3, 2015

November 12th, 2015



Department of Health

ANDREW M. CUOMO
Governor

HOWARD A. ZUCKER, M.D., J.D.
Commissioner

SALLY DRESLIN, M.S., R.N.
Executive Deputy Commissioner

November 12, 2015

Trace Elements in Serum Event #3, 2015

Dear Laboratory Director:

Results from the third proficiency test (PT) event for 2015 in the category Trace Elements in Serum have been tabulated and are summarized. Target values for aluminum, copper, selenium and zinc have been established along with acceptable ranges. Results are graded using element-specific criteria as indicated in each narrative section. A laboratory with an unacceptable significant analytical bias relative to the target value will be expected to investigate the source of the error. A confidential three-digit code number assigned by the PT program identifies participant laboratories.

PT Materials

Test materials were prepared from human serum obtained from Tennessee Blood Services, Inc. Serum units were spiked with a suite of additional trace elements as described in each narrative. In addition to aluminum (Al), copper (Cu), selenium (Se) and zinc (Zn), serum pools were supplemented with the trace elements chromium (Cr), cobalt, (Co), manganese (Mn), thallium (Tl), and vanadium (V).

Thank you for your participation.

Sincerely,

Patrick J. Parsons, Ph.D.
Chief, Laboratory of Inorganic and Nuclear Chemistry
Deputy Director, Division of Environmental Health

Mary Frances Verostek, Ph.D.
Assistant Section Head
PT Program for Blood Lead /Trace Elements
Biggs Laboratory - Wadsworth Center
New York State Department of Health
PO Box 509
Albany NY 12201-0509
www.wadsworth.org/testing/lead/index.htm

New York State Department of Health
Event #3, 2015

Serum Aluminum

The test materials for serum Al were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be negative for HIV 1/2 and HIV-1 RNA, and non-reactive to HBsAg, HCV3 and STS. Serum was dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including aluminum as Al³⁺ at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum aluminum range from 31 µg/L (1.15 µmol/L) to 143 µg/L (5.30 µmol/L).

Acceptable ranges for serum aluminum are based on fixed criteria of ±20%, or ±5 µg/L below 25 µg/L. These criteria are based on consensus recommendations from several EQAS organizers (1).

Discussion. Based on the above criteria, 97.1% of test results reported were judged as satisfactory, with one of the 21 participant laboratories (4.8%) reporting 2 or more of the 5 results outside the acceptable ranges.

1. Taylor, A., Angerer, J., Claeys, F., Kristiansen, J., Mazarrasa, O., Menditto, A., Patriarca, M., Pineau, A., Schoeters, I., Sykes, C., Valkonen, S. and Weykamp, C. Comparison of procedures for evaluating laboratory performance in external quality assessment schemes for lead in blood and aluminum in serum demonstrates the need for common quality specifications. Clinical Chemistry 2002 **48** 2000-2007.

New York State Department of Health
Serum Aluminum Test Results, 2015 Event #3
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

	Results ($\mu\text{g/L}$ serum)				
	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
Robust Mean	61	83	31	143	105
Robust Standard Deviation	4	6	3	8	7
Standard Uncertainty	1.2	1.5	0.8	2.3	1.9
RSD (%)	7.2	6.7	9.2	5.8	6.4
Number of Sample Measurements	21	21	21	20	20
Acceptable Range:					
Upper Limit:	73	100	37	172	126
Lower Limit:	49	66	25	114	84

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Aluminum Test Results, 2015 Event #3
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L serum)					Info Only
		SE15-11	SE15-12	SE15-13	SE15-14	SE15-15	
	Target Values:	61	83	31	143	105	
114	ICP-MS	69	92	36	160	128 ↑	
147	FAAS	65	85	33	148	109	Info
156	ICP-MS	57	78	30	140	98	
160	ICP-MS	60	78	29	137	98	
164	ICP-MS	64	87	32	154	114	
179	DRC/CC-ICP-MS	65	85	32	151	109	
200	DRC/CC-ICP-MS	56	80	27	132	95	Info
206	DRC/CC-ICP-MS	66	83	29	>100	>100	
287	ETAAS-Z	59	78	33	144	102	
293	ICP-MS	61	83	37	148	115	Info
305	ICP-MS	65	88	30	148	108	
325	ETAAS-Z	63	82	31	133	101	Info
355	ICP-MS	64	91	33	150	108	
357	ICP-MS	58	80	28	140	101	
358	ICP-MS	60	79	30	145	105	
363	ICP-MS	66	87	34	150	112	
366	ETAAS-Z	54	74	29	117	104	Info
388	ICP-MS	58	87	27	145	101	
401	DRC/CC-ICP-MS	62	84	32	127	105	Info
458	ETAAS other	48 ↓	65 ↓	28	116	84	
485	HR-ICP-MS	63.4	79.4	32.3	144.8	109.4	Info

Percent satisfactory results for all participants: 97.1 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

▲: Result not reported

New York State Department of Health
Serum Aluminum Test Results, 2015 Event #3
STATISTICAL SUMMARY BY METHOD

	Results ($\mu\text{g/L}$ serum)				
	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
DRC/CC-ICP-MS					
Number of Sample Measurements:	4	4	4	3	3
Mean:	62	83	30	137	103
Standard Deviation:	5	2	2	13	7
RSD (%):	7.2	2.6	8.2	—	—
ETAAS other					
Number of Sample Measurements:	1	1	1	1	1
Mean:	48	65	28	116	84
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ETAAS-Z					
Number of Sample Measurements:	3	3	3	3	3
Mean:	59	78	31	131	102
Standard Deviation:	5	4	2	14	2
RSD (%):	—	—	—	—	—
FAAS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	65	85	33	148	109
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
HR-ICP-MS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	63	79	32	145	109
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	11	11	11	11	11
Mean:	62	85	31	147	108
Standard Deviation:	4	5	3	7	9
RSD (%):	6.2	6.1	10.3	4.5	8.3
All Laboratories					
Number of Sample Measurements:	21	21	21	20	20
Mean:	61	82	31	141	105
Standard Deviation:	5	6	3	12	9
RSD (%):	8.0	7.4	8.9	8.2	8.5

notes: ? Insufficient data for calculation.

Serum Copper

The test materials for serum Cu were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to be STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including copper as Cu^{2+} at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum copper range from 950 $\mu\text{g/L}$ (14.95 $\mu\text{mol/L}$) to 2472 $\mu\text{g/L}$ (38.90 $\mu\text{mol/L}$).

Acceptable ranges for serum copper are based on fixed criteria of $\pm 15\%$, or $\pm 95 \mu\text{g/L}$ below 635 $\mu\text{g/L}$. These criteria are consistent with those proposed by the OELM Network of EQAS organizers (1, 2) for trace elements in serum, and are slightly less stringent than those previously suggested for NYS ($\pm 10\%$).

Discussion. Based on the above criteria, 99.0% of test results reported were judged as satisfactory, with none of the 21 participant laboratories reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 11 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 54 1892-1899.

New York State Department of Health
Serum Copper Test Results, 2015 Event #3
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ serum)

	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
Robust Mean	989	1549	2472	1531	950
Robust Standard Deviation	43	86	168	78	37
Standard Uncertainty	12	23	46	21	10
RSD (%)	4.4	5.5	6.8	5.1	3.9
Number of Sample Measurements	21	21	21	21	21
Acceptable Range:					
Upper Limit:	1137	1781	2843	1761	1093
Lower Limit:	841	1317	2101	1301	807

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Copper Test Results, 2015 Event #3
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L serum)					Info Only
		SE15-11	SE15-12	SE15-13	SE15-14	SE15-15	
	Target Values:	989	1549	2472	1531	950	
107	ICP-MS	1000	1600	2500	1500	930	Info
110	ICP-MS	1008	1587	2674	1532	960	
114	ICP-MS	970	1500	2360	1480	940	
147	ICP-MS	985	1544	2459	1499	959	Info
156	ICP-AES/OES	1000	1700	2600	1700	1000	
160	ICP-MS	879	1422	2281	1422	882	
164	ICP-MS	919	1446	2316	1418	898	
179	DRC/CC-ICP-MS	1030	1620	2650	1620	1030	
197	ICP-MS	990	1640	2490	1550	970	
200	ICP-MS	1041	1683	2750	1695	1048	Info
206	ICP-MS	920	1430	2222	1500	900	
293	ICP-MS	998	1519	2479	1551	941	Info
305	ICP-MS	960	1500	2350	1480	920	
325	ICP-MS	963	1557	2498	1558	1228 ↑	Info
359	ICP-MS	1038	1563	2521	1567	942	
366	ETAAS-Z	977	1547	2348	1517	940	Info
388	ICP-MS	1073	1606	2644	1636	1017	
401	DRC/CC-ICP-MS	972	1487	2383	1474	934	Info
457	ICP-AES/OES	1071	1609	2676	1603	1040	Info
483	DRC/CC-ICP-MS	951	1441	2278	1436	928	Info
484	ICP-MS	1002	1552	2460	1529	954	

Percent satisfactory results for all participants: 99.0 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

▲: Result not reported

New York State Department of Health
Serum Copper Test Results, 2015 Event #3
STATISTICAL SUMMARY BY METHOD

	Results ($\mu\text{g/L}$ serum)				
	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
DRC/CC-ICP-MS					
Number of Sample Measurements:	3	3	3	3	3
Mean:	984	1516	2437	1510	964
Standard Deviation:	41	93	192	97	57
RSD (%):	—	—	—	—	—
ETAAS-Z					
Number of Sample Measurements:	1	1	1	1	1
Mean:	977	1547	2348	1517	940
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ICP-AES/OES					
Number of Sample Measurements:	2	2	2	2	2
Mean:	1036	1655	2638	1652	1020
Standard Deviation:	50	64	54	69	28
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	15	15	15	15	15
Mean:	983	1543	2467	1528	966
Standard Deviation:	51	76	147	72	84
RSD (%):	5.2	4.9	6.0	4.7	8.7
All Laboratories					
Number of Sample Measurements:	21	21	21	21	21
Mean:	988	1550	2473	1537	970
Standard Deviation:	48	80	151	80	75
RSD (%):	4.9	5.1	6.1	5.2	7.8

notes: ? Insufficient data for calculation.

Serum Selenium

The test materials for serum Se were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to be STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including selenium as Se^{4+} at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum selenium range from 128 µg/L (1.62 µmol/L) to 363 µg/L (4.60 µmol/L).

Acceptable ranges for serum selenium are based on fixed criteria of $\pm 20\%$, or ± 2 µg/L below 10 µg/L. These criteria are a little less stringent than those proposed by the OELM Network of EQAS organizers ($\pm 15\%$ or ± 8 µg/L below 55 µg/L) (1, 2) for trace elements in serum. As performance for serum Se improves among NYS-permit laboratories, consideration will be given to adopting the OELM criteria.

Discussion. Based on the above criteria, 97.9% of test results reported were judged as satisfactory, with one of the 19 participant laboratories (5.3%) reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 **11** 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 **54** 1892-1899.

New York State Department of Health
Serum Selenium Test Results, 2015 Event #3
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

Results ($\mu\text{g/L}$ serum)

	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
--	---------	---------	---------	---------	---------

Robust Mean	132	229	128	161	363
--------------------	------------	------------	------------	------------	------------

Robust Standard Deviation	6	11	6	10	16
---------------------------	---	----	---	----	----

Standard Uncertainty	2	3	2	3	5
----------------------	---	---	---	---	---

RSD (%)	4.9	4.9	4.6	6.1	4.5
---------	-----	-----	-----	-----	-----

Number of Sample Measurements	19	19	19	19	19
-------------------------------	----	----	----	----	----

Acceptable Range:

Upper Limit:	158	275	154	193	436
--------------	-----	-----	-----	-----	-----

Lower Limit:	106	183	102	129	290
--------------	-----	-----	-----	-----	-----

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Selenium Test Results, 2015 Event #3
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L serum)					Info Only
		SE15-11	SE15-12	SE15-13	SE15-14	SE15-15	
Target Values:		132	229	128	161	363	
103	DRC/CC-ICP-MS	132	229	126	160	369	Info
107	ICP-MS	130	220	120	150	330	Info
110	DRC/CC-ICP-MS	131	228	129	159	362	
114	ICP-MS	140	236	136	169	350	
147	ICP-MS	119	205	114	148	325	Info
156	DRC/CC-ICP-MS	120	220	120	150	330	
160	ICP-MS	134	226	127	167	368	
164	DRC/CC-ICP-MS	128	239	124	161	364	
179	DRC/CC-ICP-MS	130	227	127	154	366	
200	DRC/CC-ICP-MS	133	231	129	165	379	Info
206	DRC/CC-ICP-MS	145	233	133	172	371	
293	DRC/CC-ICP-MS	140	239	134	164	375	Info
305	ICP-MS	136	223	139	170	360	
325	ETAAS-Z	123	200	128	141	385	Info
366	ETAAS-Z	126	248	128	153	374	Info
367	DRC/CC-ICP-MS	126	245	132	342 ↑	174 ↓	Info
388	ICP-MS	134	223	123	166	358	
401	DRC/CC-ICP-MS	137	236	131	167	385	Info
483	DRC/CC-ICP-MS	132	220	126	160	356	Info

Percent satisfactory results for all participants: 97.9 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

▲: Result not reported

New York State Department of Health
Serum Selenium Test Results, 2015 Event #3
STATISTICAL SUMMARY BY METHOD

Results ($\mu\text{g/L}$ serum)					
	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
DRC/CC-ICP-MS					
Number of Sample Measurements:	11	11	11	10	11
Mean:	132	232	128	161	348
Standard Deviation:	7	8	4	6	60
RSD (%):	5.1	3.4	3.3	3.9	17.1
ETAAS-Z					
Number of Sample Measurements:	2	2	2	2	2
Mean:	125	224	128	147	380
Standard Deviation:	2	34	0	8	8
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	6	6	6	6	6
Mean:	132	222	127	162	349
Standard Deviation:	7	10	10	10	17
RSD (%):	5.5	4.5	7.6	6.1	5.0
All Laboratories					
Number of Sample Measurements:	19	19	19	18	19
Mean:	131	228	128	160	352
Standard Deviation:	7	12	6	9	46
RSD (%):	5.2	5.3	4.7	5.5	13.2

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #3, 2015

Serum Zinc

The test materials for serum Zn were prepared from human serum obtained from Tennessee Blood Services, Inc. The units were tested by FDA approved methods and found to be Non-reactive for Anti-HIV-1/2, Anti-HCV 3.0 and HBsAg. The serum has also been found to be STS (RPR) Non-reactive and Negative for HIV-1 and HCV by PCR. Serum units were dispensed into acid-washed 500-mL polypropylene containers to make up five (5) serum pools. Each pool was spiked with a suite of additional trace elements including zinc as Zn^{2+} at various concentrations.

The Target Value assigned for each PT material is the robust mean of the results reported by all participants in this event. The robust statistics were obtained utilizing algorithms based on those presented in **ISO 13528:2005E** Statistical methods for use in proficiency testing by interlaboratory comparisons. Values for serum zinc range from 617 µg/L (9.44 µmol/L) to 2578 µg/L (39.42 µmol/L).

Acceptable ranges for serum zinc are based on fixed criteria of $\pm 15\%$, or ± 15 µg/L below 100 µg/L. These criteria are consistent with those proposed by the OELM network of EQAS organizers (1) for trace elements in serum.

Discussion. Based on the above criteria, 94.6% of test results reported were judged as satisfactory, with two of the 26 participant laboratories (7.7%) reporting 2 or more of the 5 results outside the acceptable ranges.

1. A. Taylor, J. Angerer, J. Arnaud, F. Claeys, R.L. Jones, O. Mazarrasa, E. Mairiaux, A. Menditto, P.J. Parsons, M. Patriarca, A. Pineau, S. Valkonen, J.-P. Weber and C. Weykamp Accreditation and Quality Assurance 2006 **11** 440-445.

2. J. Arnaud, J.-P. Weber, C.W. Weykamp, P.J. Parsons, J. Angerer, E. Mairiaux, O. Mazarrasa, S. Valkonen, A. Menditto, M. Patriarca, and A. Taylor Clinical Chemistry 2008 **54** 1892-1899.

New York State Department of Health
Serum Zinc Test Results, 2015 Event #3
ROBUST STATISTICAL SUMMARY

TARGET VALUE ASSIGNMENT AND STATISTICS

	Results ($\mu\text{g/L}$ serum)				
	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
Robust Mean	2578	617	712	1078	1943
Robust Standard Deviation	140	35	44	60	118
Standard Uncertainty	34	9	11	15	29
RSD (%)	5.4	5.7	6.3	5.6	6.0
Number of Sample Measurements	26	26	26	26	26
Acceptable Range:					
Upper Limit:	2965	710	819	1240	2235
Lower Limit:	2191	525	605	916	1651

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

New York State Department of Health
Serum Zinc Test Results, 2015 Event #3
PERFORMANCE OF PARTICIPATING LABORATORIES

Lab Code	Method	Results (µg/L serum)					Info Only
		SE15-11	SE15-12	SE15-13	SE15-14	SE15-15	
	Target Values:	2578	617	712	1078	1943	
107	ICP-MS	2700	560	570 ↓	900 ↓	1900	Info
110	ICP-MS	2717	635	764	1113	2038	
114	ICP-MS	2470	600	680	1030	1840	
147	ICP-MS	2333	620	673	1013	1752	Info
156	ICP-AES/OES	2600	620	700	1100	2000	
160	ICP-MS	2406	570	658	1025	1849	
164	ICP-MS	2422	582	679	1030	1929	
179	DRC/CC-ICP-MS	2680	580	720	1080	2040	
197	ICP-MS	2590	640	710	1080	1910	
200	ICP-MS	2668	667	778	1158	2112	Info
206	ICP-MS	2410	580	700	1080	1860	
287	FAAS	2550	600	710	1010	1810	
293	ICP-MS	2614	652	758	1105	1915	Info
305	ICP-MS	2290	550	640	990	1740	
325	ICP-MS	2561	625	738	1082	2425 ↑	Info
355	ICP-MS	2482	620	683	1043	1883	
357	ICP-MS	2653	620	715	1092	1998	
358	ICP-MS	2870	720 ↑	850 ↑	1360 ↑	2400 ↑	
359	ICP-MS	2843	678	769	1162	2023	
366	FAAS	2525	624	683	1017	1827	Info
388	ICP-MS	2571	579	704	1095	1886	
401	DRC/CC-ICP-MS	2805	641	752	1144	2125	Info
457	ICP-AES/OES	2568	619	702	1090	1982	Info
458	FAAS	2657	661	774	1141	2008	
483	ICP-MS	2542	628	726	1096	1955	Info
484	ICP-MS	2570	600	680	1090	1920	

Percent satisfactory results for all participants: 94.6 %

notes: ↑ reported outside upper limit
↓ reported outside lower limit

Info only: results included for informational purposes only.

notes: Results reported as less than the method detection limit are excluded from statistical calculations.

▲: Result not reported

New York State Department of Health
Serum Zinc Test Results, 2015 Event #3
STATISTICAL SUMMARY BY METHOD

Results ($\mu\text{g/L}$ serum)					
	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
DRC/CC-ICP-MS					
Number of Sample Measurements:	2	2	2	2	2
Mean:	2743	611	736	1112	2083
Standard Deviation:	88	43	23	45	60
RSD (%):	—	—	—	—	—
FAAS					
Number of Sample Measurements:	3	3	3	3	3
Mean:	2577	628	722	1056	1882
Standard Deviation:	70	31	47	74	110
RSD (%):	—	—	—	—	—
ICP-AES/OES					
Number of Sample Measurements:	2	2	2	2	2
Mean:	2584	620	701	1095	1991
Standard Deviation:	23	1	1	7	13
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	19	19	19	19	19
Mean:	2564	617	709	1081	1965
Standard Deviation:	158	43	61	91	182
RSD (%):	6.2	7.0	8.5	8.4	9.2
All Laboratories					
Number of Sample Measurements:	26	26	26	26	26
Mean:	2581	618	712	1082	1966
Standard Deviation:	145	39	54	81	164
RSD (%):	5.6	6.3	7.6	7.5	8.3

notes: ? Insufficient data for calculation.

New York State Department of Health
Event #3, 2015

Additional Trace Elements Reported in Serum

Participant laboratories reported their analytical results for any additional trace elements (other than Al, Cu, Se and Zn) that are routinely reported so that a more complete characterization can be recorded for these PT materials. Results for additional trace elements are reported here, but no target value is implied nor are any acceptable ranges provided. These data are provided solely for educational and informational purposes.

In addition to Al, Cu, Se and Zn, the serum pools were supplemented with additional trace elements as indicated below.

Additional Elements

Cr, Co, Mn, Tl, V

New York State Department of Health
Serum Additional Elements, 2015 Event #3

Page 1

Serum Antimony (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	0.0729	<0.00973	<0.00973	<0.00973	<0.00973
110	ICP-MS	0.035	<0.02	<0.02	<0.02	<0.02
147	ICP-MS	0.0421	<0.030	<0.030	<0.030	<0.030

Serum Arsenic (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	0.0911	0.117	0.119	0.670	0.102
110	DRC/CC-ICP-MS	0.27	0.24	0.22	0.88	0.28
197	DRC/CC-ICP-MS	<10	<10	<10	<10	<10

Serum Barium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
110	ICP-MS	1.77	1.52	1.27	1.41	1.41
147	ICP-MS	1.52	1.36	1.07	1.32	1.17
197	ICP-MS	<2.0	<2.0	<2.0	<2.0	<2.0

Serum Beryllium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
110	ICP-MS	<0.04	<0.04	<0.04	<0.04	<0.04
147	ICP-MS	<0.36	<0.36	<0.36	<0.36	<0.36
197	ICP-MS	<0.2	<0.2	<0.2	<0.2	<0.2

Serum Bismuth (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
147	ICP-MS	<0.042	<0.042	<0.042	<0.042	<0.042
197	ICP-MS	<1.0	<1.0	<1.0	<1.0	<1.0

Serum Cadmium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	0.0133	0.0131	0.0136	0.0268	0.00872
110	ICP-MS	<0.03	<0.03	<0.03	<0.03	<0.03
147	ICP-MS	0.0232	<0.022	0.0247	0.0360	<0.022
197	DRC/CC-ICP-MS	<0.5	<0.5	<0.5	<0.5	<0.5

Serum Chromium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	1.19	3.68	2.01	0.670	1.71
110	DRC/CC-ICP-MS	1.54	3.81	2.29	0.57	2.24
147	DRC/CC-ICP-MS	1.18	3.63	1.97	0.707	1.60
156	DRC/CC-ICP-MS	1	3.5	2	<1	1.7
160	ICP-MS	1.0	3	2	1	2
164	DRC/CC-ICP-MS	1	3.5	1.9	0.6	1.6
179	DRC/CC-ICP-MS	1.3	3.6	2.0	0.7	1.7
197	DRC/CC-ICP-MS	1.3	3.5	1.9	<1.0	1.7
206	DRC/CC-ICP-MS	1.1	3.7	*3.0	<1.0	1.9
305	ICP-MS	1.2	3.7	2.0	0.3	1.6
366	ICP-MS	1.0	3.2	1.4	0.5	1.5
485	HR-ICP-MS	1.17	3.49	2.15	0.67	1.72
*Outlier	Arithmetic mean	1.2	3.5	2.0	0.6	1.7
	SD	0.2	0.2	0.2	0.2	0.2
	n	12	12	11	9	12

New York State Department of Health
Serum Additional Elements, 2015 Event #3
Page 2

Serum Cobalt (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	0.260	4.49	2.70	0.975	0.559
110	ICP-MS	0.37	4.74	2.97	1.14	0.71
147	ICP-MS	0.320	4.27	2.50	0.949	0.552
156	DRC/CC-ICP-MS	<0.5	4.2	2.5	0.9	0.53
164	ICP-MS	0.3	4.4	2.4	1.0	1.0
179	DRC/CC-ICP-MS	0.3	4.5	2.7	1.0	0.6
197	ICP-MS	<1.0	3.9	2.4	<1.0	<1.0
206	ICP-MS	<1.0	4.6	2.7	1.2	<1.0
366	ICP-MS	0.5	5.3	2.3	*1.5	0.9
485	HR-ICP-MS	0.25	4.65	2.85	1.02	0.58
*Outlier	Arithmetic mean	0.3	4.5	2.6	1.0	0.7
	SD	0.1	0.4	0.2	0.1	0.2
	n	7	10	10	8	8

Serum Iodine (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
147	ICP-MS	54.4	57.5	49.4	62.5	46.2
156	DRC/CC-ICP-MS	62	69	51	78	56
160	ICP-MS	56	62	53	69	52
164	ICP-MS	49	56	51	60	44
179	ICP-MS	53	59	50	65	48
197	ICP-MS	54.8	62.6	52.4	68.9	49.6
206	ICP-MS	47.0	52.2	45.0	58.6	42.8
Arithmetic mean		54	60	50	66	48
SD		5	5	3	7	5
n		7	7	7	7	7

Serum Iron (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
457	ICP-AES/OES	980	882	690	873	2433

Serum Lead (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	0.0343	0.0196	<.00862	0.0504	0.0189
110	ICP-MS	<0.6	<0.6	<0.6	<0.6	<0.6
147	ICP-MS	<0.062	<0.062	<0.062	<0.062	<0.062
197	DRC/CC-ICP-MS	<0.4	<0.4	<0.4	<0.4	<0.4

Serum Lithium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
147	ICP-MS	0.840	1.17	0.777	2.71	0.589

Serum Manganese (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	7.00	0.848	0.784	3.71	1.89
147	ICP-MS	6.69	0.984	0.890	3.46	2.02
179	DRC/CC-ICP-MS	7.8	1.0	0.9	3.9	2.3
197	DRC/CC-ICP-MS	7.3	1.4	1.0	4.6	2.3
206	ICP-MS	6.8	1.0	1.1	3.4	2.0
305	ICP-MS	7.4	1.4	1.2	4.0	2.5
Arithmetic mean		7.2	1.1	1.0	3.8	2.2
SD		0.4	0.2	0.2	0.4	0.2
n		6	6	6	6	6

New York State Department of Health
Serum Additional Elements, 2015 Event #3

Page 3

Serum Mercury (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	0.135	0.141	0.119	0.373	0.182
110	ICP-MS	0.18	0.23	0.21	0.60	0.26
147	ICP-MS	<0.16	<0.16	<0.16	0.365	<0.16
197	ICP-MS	<5	<5	<5	<5	<5

Serum Molybdenum (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	1.37	1.57	1.42	0.975	1.69
110	ICP-MS	1.53	1.64	1.65	0.99	1.73
147	ICP-MS	1.23	1.46	1.35	0.880	1.56
179	ICP-MS	1.5	1.6	1.4	0.9	1.6
197	ICP-MS	<2.0	<2.0	<2.0	<2.0	<2.0
485	HR-ICP-MS	1.23	1.59	1.44	0.90	1.65
Arithmetic mean		1.4	1.57	1.5	0.93	1.65
SD		0.1	0.07	0.1	0.05	0.07
n		5	5	5	5	5

Serum Nickel (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
110	DRC/CC-ICP-MS	<1	<1	<1	<1	<1
147	ICP-MS	<0.29	<0.29	<0.29	<0.29	<0.29
179	DRC/CC-ICP-MS	0.2	0.2	0.2	<0.2	0.3
197	ICP-MS	<2.0	<2.0	<2.0	<2.0	<2.0
206	ICP-MS	<10.0	<10.0	<10.0	<10.0	<10.0

Serum Platinum (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
110	ICP-MS	<0.04	<0.04	<0.04	<0.04	<0.04

Serum Silver (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
147	ICP-MS	0.143	0.099	0.0791	0.247	0.150
197	ICP-MS	<1.0	<1.0	<1.0	<1.0	<1.0

Serum Strontium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	31.5	34.5	30.2	27.6	24.4
200	ICP-MS	32.4	35.9	31.2	29.2	25.7

Serum Tellurium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
110	ICP-MS	<0.2	<0.2	<0.2	<0.2	<0.2
147	ICP-MS	<0.077	<0.077	<0.077	<0.077	<0.077

Serum Thallium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	0.552	2.16	3.34	0.919	0.724
110	ICP-MS	0.60	2.32	3.70	1.02	0.78
147	ICP-MS	0.509	2.00	3.07	0.844	0.673
197	ICP-MS	<1.0	2.1	3.2	<1.0	<1.0
Arithmetic mean		0.55	2.1	3.3	0.93	0.73
SD		0.05	0.1	0.3	0.09	0.05
n		3	4	4	3	3

New York State Department of Health
Serum Additional Elements, 2015 Event #3
Page 4

Serum Thorium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
147	ICP-MS	<0.0070	<0.0070	<0.0070	<0.0070	<0.0070

Serum Tin (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
110	ICP-MS	0.15	0.12	0.21	0.18	0.13
197	ICP-MS	<5.0	<5.0	<5.0	<5.0	<5.0

Serum Tungsten (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
110	ICP-MS	<0.04	<0.04	<0.04	<0.04	<0.04
147	ICP-MS	<0.239	<0.239	<0.239	<0.239	<0.239

Serum Uranium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
103	DRC/CC-ICP-MS	0.00198	0.00141	0.00131	0.00215	0.00206
110	ICP-MS	0.003	<0.003	<0.003	<0.003	<0.003
147	ICP-MS	<0.0071	<0.0071	<0.0071	<0.0071	<0.0071

Serum Vanadium (µg/L)						
Lab Code	Method	SE15-11	SE15-12	SE15-13	SE15-14	SE15-15
110	ICP-MS	5.13	*0.89	2.27	*1.19	2.98
147	DRC/CC-ICP-MS	3.97	0.587	1.61	0.913	2.49
179	DRC/CC-ICP-MS	4.5	0.6	1.7	0.9	2.8
485	HR-ICP-MS	4.19	0.63	1.77	0.94	2.96
*Outlier	Arithmetic mean	4.4	0.61	1.8	0.92	2.8
	SD	0.5	0.02	0.3	0.02	0.2
	n	4	3	4	3	4

**New York State Department of Health
Trace Elements in Serum
METHOD NOTES**

ATOMIC SPECTROMETRY METHODS

- A-1 ETAAS-Z (Electrothermal atomic absorption spectrometry with Zeeman background correction)
- A-2 ETAAS Other (i.e., D₂, S-H background correction)
- A-3 FAAS (Flame atomic absorption spectrometry)
- A-4 CV-AAS (Cold vapor atomic absorption spectrometry)
- A-5 HG-AAS (Hydride generation atomic absorption spectrometry)
- A-6 AFS (Atomic fluorescence spectrometry)

INDUCTIVELY COUPLED PLASMA

- P-1 ICP-MS (Inductively coupled plasma - mass spectrometry)
- P-2 DRC/CC-ICP-MS (ICP-MS used in the Dynamic Reaction Cell or Collision Cell mode)
- P-3 ICP-AES/OES (ICP atomic/optical emission spectrometry)
- P-4 HR-ICP-MS (High resolution ICP-MS)
- P-5 ETV-ICP-MS (Electrothermal vaporization ICP-MS)
- P-6 ID-ICP-MS (Isotope dilution ICP-MS)

ELECTROCHEMICAL METHODS

- E-1 ASV (Anodic stripping voltammetry without digestion)
- E-2 ASV-LeadCare® Blood Lead Testing System
- E-5 ASV-LeadCare® II Blood Lead Testing System
- E-6 ASV-LeadCare® Ultra™ Blood Lead Testing System
- E-3 Fluoride specific electrode

MOLECULAR FLUORIMETRY

- F-1 EtOAc (Ethyl acetate-acetic acid extraction method for determination of erythrocyte protoporphyrin)
- F-2 Aviv hematofluorometry (for determination of EP at hematocrit 35)
- F-3 Helena ZPP (for determination of zinc protoporphyrin in μmol ZPP/mol heme)

OTHER METHODS

If your method is not listed in the above list, please describe it briefly.
