

Statistical Report
Urinalysis/Urine Pregnancy Testing
Mail out date: April 9, 2012

This report summarizes data from the urinalysis/urine pregnancy proficiency testing of April 9, 2012.

If you have any questions regarding this report, please contact Regina Saulsbery at 518-474-6156. You may also contact us by e-mail: clinchem@wadsworth.org

Results are shown for each analyte arranged by instrument and reagent system. Results for specific gravity and pH are listed as mean \pm 1SD for each of the instruments and reagents. All other results are shown as the number of participant results for that instrument or reagent.

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

PLEASE NOTE: The evaluation report and summary of the educational quantitative urine clinical chemistry proficiency test will be mailed separately if your laboratory participated in that proficiency test.

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Specific Gravity

Specimen: U01	Specimen: U02	Specimen: U03	Specimen: U04	Specimen: U05	Number	Instrument or Reagent System
1.019 ± 0.005	1.016 ± 0.003	1.016 ± 0.005	1.020 ± 0.006	1.017 ± 0.004	n = 373	All Methods & Instruments
1.017	1.015	1.015	1.020	1.016	n = 373	Median value, All Methods & Instruments
<Instruments>						
1.021 ± 0.003	1.013 ± 0.004	1.015 ± 0.003	1.023 ± 0.004	1.015 ± 0.003	n = 27	Visual examination
1.016 ± 0.002	1.016 ± 0.002	1.015 ± 0.002	1.017 ± 0.002	1.016 ± 0.002	n = 96	Siemens Atlas
1.025 ± 0.003	1.019 ± 0.001	1.023 ± 0.003	1.025 ± 0.003	1.023 ± 0.003	n = 38	Siemens Clinitek Status
1.024 ± 0.004	1.014 ± 0.002	1.021 ± 0.002	1.029 ± 0.002	1.020 ± 0.003	n = 7	Siemens Clinitek 50
1.024 ± 0.002	1.020 ± 0.000	1.024 ± 0.002	1.025 ± 0.000	1.023 ± 0.003	n = 6	Siemens Clinitek 100
1.025 ± 0.002	1.017 ± 0.003	1.021 ± 0.002	1.026 ± 0.003	1.021 ± 0.002	n = 48	Siemens Clinitek 500
1.025 ± 0.003	1.017 ± 0.003	1.021 ± 0.002	1.026 ± 0.003	1.021 ± 0.001	n = 38	Siemens Clinitek Advantus
1.017 ± 0.002	1.011 ± 0.002	1.013 ± 0.003	1.021 ± 0.002	1.014 ± 0.002	n = 13	Roche Chemstrip Criterion
1.017 ± 0.003	1.014 ± 0.002	1.013 ± 0.002	1.019 ± 0.003	1.015 ± 0.001	n = 36	Roche Urisys
1.013 ± 0.000	1.012 ± 0.001	1.010 ± 0.000	1.013 ± 0.000	1.012 ± 0.001	n = 38	IRIS AUTION MAX
1.015 ± 0.002	1.015 ± 0.001	1.013 ± 0.002	1.015 ± 0.001	1.014 ± 0.001	n = 17	Refractometer
1.017 ± 0.005	1.010 ± 0.003	1.011 ± 0.004	1.017 ± 0.006	1.011 ± 0.004	n = 6	Other
<Reagents>						
1.024 ± 0.004	1.017 ± 0.004	1.020 ± 0.004	1.025 ± 0.004	1.020 ± 0.004	n = 164	Siemens Reagent Strips
1.016 ± 0.002	1.016 ± 0.002	1.015 ± 0.002	1.017 ± 0.002	1.016 ± 0.002	n = 88	Siemens Atlas Reagent
1.017 ± 0.003	1.012 ± 0.003	1.013 ± 0.003	1.021 ± 0.002	1.014 ± 0.002	n = 43	Roche Chemstrips
1.016 ± 0.001	1.016 ± 0.001	1.014 ± 0.001	1.016 ± 0.001	1.015 ± 0.000	n = 14	Roche Urisys Cassette
1.013 ± 0.000	1.012 ± 0.001	1.010 ± 0.000	1.013 ± 0.001	1.012 ± 0.001	n = 22	ARKRAY Aution Sticks
1.013 ± 0.001	1.012 ± 0.001	1.010 ± 0.001	1.013 ± 0.001	1.012 ± 0.000	n = 11	IRIS Reagent Strips
1.014 ± 0.001	1.014 ± 0.002	1.012 ± 0.002	1.014 ± 0.001	1.014 ± 0.002	n = 23	None, no reagent or strip used
1.019 ± 0.005	1.014 ± 0.005	1.015 ± 0.007	1.019 ± 0.006	1.015 ± 0.007	n = 8	Other

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

pH

Specimen: U01	Specimen: U02	Specimen: U03	Specimen: U04	Specimen: U05	Number	Instrument or Reagent System
6.00 ± 0.18	6.99 ± 0.10	7.06 ± 0.18	5.97 ± 0.19	7.02 ± 0.13	n = 379	All Methods & Instruments
6.00	7.00	7.00	6.00	7.00	n = 379	Median value, All Methods & Instruments
<Instruments>						
5.94 ± 0.31	7.01 ± 0.08	7.21 ± 0.30	5.83 ± 0.36	7.08 ± 0.22	n = 36	Visual examination
6.04 ± 0.13	7.00 ± 0.00	7.01 ± 0.09	6.02 ± 0.12	7.01 ± 0.05	n = 95	Siemens Atlas
5.96 ± 0.20	6.99 ± 0.08	7.00 ± 0.00	5.96 ± 0.13	7.01 ± 0.08	n = 42	Siemens Clinitek Status
6.00 ± 0.00	7.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	n = 7	Siemens Clinitek 50
5.94 ± 0.18	7.13 ± 0.23	7.31 ± 0.26	5.94 ± 0.18	7.38 ± 0.23	n = 8	Siemens Clinitek 100
5.98 ± 0.10	6.95 ± 0.15	7.11 ± 0.21	5.98 ± 0.10	7.03 ± 0.18	n = 51	Siemens Clinitek 500
5.95 ± 0.19	6.94 ± 0.20	7.11 ± 0.21	6.00 ± 0.00	7.01 ± 0.08	n = 40	Siemens Clinitek Advantus
6.00 ± 0.00	7.00 ± 0.00	7.00 ± 0.00	5.93 ± 0.27	7.00 ± 0.00	n = 14	Roche Chemstrip Criterion
6.06 ± 0.26	7.00 ± 0.00	7.00 ± 0.00	5.94 ± 0.33	7.00 ± 0.00	n = 36	Roche Urisys
6.00 ± 0.00	7.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	n = 39	IRIS AUTION MAX
6.07 ± 0.13	7.03 ± 0.05	7.21 ± 0.27	5.96 ± 0.08	7.10 ± 0.17	n = 7	Other
<Reagents>						
5.97 ± 0.18	6.98 ± 0.15	7.11 ± 0.22	5.97 ± 0.14	7.04 ± 0.17	n = 179	Siemens Reagent Strips
6.04 ± 0.14	7.00 ± 0.00	7.01 ± 0.09	6.02 ± 0.13	7.01 ± 0.05	n = 89	Siemens Atlas Reagent
5.99 ± 0.24	7.00 ± 0.00	7.01 ± 0.07	5.88 ± 0.38	7.00 ± 0.00	n = 47	Roche Chemstrips
6.04 ± 0.37	7.00 ± 0.00	7.00 ± 0.00	5.93 ± 0.27	7.00 ± 0.00	n = 14	Roche Urisys Cassette
6.00 ± 0.00	7.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	n = 30	ARKRAY Aution Sticks
6.00 ± 0.00	7.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	n = 11	IRIS Reagent Strips
6.04 ± 0.15	6.99 ± 0.12	7.15 ± 0.46	5.85 ± 0.20	7.06 ± 0.37	n = 8	Other

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Bilirubin	Specimen: U01	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		69	0	0	0	0
Siemens Atlas		94	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0
Siemens Clinitek 500		49	0	0	0	0
Siemens Clinitek Advantus		40	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		35	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0
Other		6	0	0	0	0
<Reagents>						
Siemens Reagent Strips		172	0	0	0	0
Siemens Atlas Reagent		87	0	0	0	0
Roche Chemstrips		45	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0
Siemens Ictotest		40	0	0	0	0
Other		6	0	0	0	0
All Instruments and Reagents		404	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Bilirubin	Specimen: U02	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
	Visual examination	69	0	0	0	0
	Siemens Atlas	93	0	1	0	0
	Siemens Clinitek Status	41	0	0	0	0
	Siemens Clinitek 50	6	0	1	0	0
	Siemens Clinitek 100	9	0	0	0	0
	Siemens Clinitek 500	49	0	0	0	0
	Siemens Clinitek Advantus	40	0	0	0	0
	Roche UA Reader/Super UA Read	2	0	0	0	0
	Roche Chemstrip Criterion	13	0	0	0	0
	Roche Urisys	35	0	0	0	0
	IRIS AUTION MAX	39	0	0	0	0
	Other	6	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	171	0	1	0	0
	Siemens Atlas Reagent	86	0	1	0	0
	Roche Chemstrips	45	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0
	ARKRAY Aution Sticks	30	0	0	0	0
	IRIS Reagent Strips	11	0	0	0	0
	Siemens Ictotest	40	0	0	0	0
	Other	6	0	0	0	0
	All Instruments and Reagents	402	0	2	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Bilirubin	Specimen: U03	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		69	0	0	0	0
Siemens Atlas		94	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0
Siemens Clinitek 500		49	0	0	0	0
Siemens Clinitek Advantus		40	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		35	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0
Other		6	0	0	0	0
<Reagents>						
Siemens Reagent Strips		172	0	0	0	0
Siemens Atlas Reagent		87	0	0	0	0
Roche Chemstrips		45	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0
Siemens Ictotest		40	0	0	0	0
Other		6	0	0	0	0
All Instruments and Reagents		404	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Bilirubin	Specimen: U04	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		3	0	25	25	145
Siemens Atlas		2	20	69	0	3
Siemens Clinitek Status		0	3	38	0	0
Siemens Clinitek 50		0	0	7	0	0
Siemens Clinitek 100		0	0	2	6	1
Siemens Clinitek 500		0	1	47	0	3
Siemens Clinitek Advantus		1	0	39	0	0
Roche UA Reader/Super UA Read		0	2	0	0	0
Roche Chemstrip Criterion		6	6	1	0	0
Roche Urisys		2	32	1	0	0
IRIS AUTION MAX		1	2	35	0	1
Other		1	2	1	1	1
<Reagents>						
Siemens Reagent Strips		1	4	142	23	2
Siemens Atlas Reagent		2	19	63	0	3
Roche Chemstrips		7	30	6	3	0
Roche Urisys Cassette		1	12	0	0	0
ARKRAY Aution Sticks		0	1	28	0	1
IRIS Reagent Strips		2	1	8	0	0
Siemens Ictotest		3	0	16	4	147
Other		0	1	2	2	1
All Instruments and Reagents		16	68	265	32	154

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Bilirubin	Specimen: U05	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		69	0	0	0	0
Siemens Atlas		94	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0
Siemens Clinitek 500		49	0	0	0	0
Siemens Clinitek Advantus		40	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		35	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0
Other		6	0	0	0	0
<Reagents>						
Siemens Reagent Strips		172	0	0	0	0
Siemens Atlas Reagent		87	0	0	0	0
Roche Chemstrips		45	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0
Siemens Ictotest		40	0	0	0	0
Other		6	0	0	0	0
All Instruments and Reagents		404	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Glucose	Specimen: U01	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		88	0	0	0	0	0	0
Siemens Atlas		95	1	0	0	0	0	0
Siemens Clinitek Status		44	0	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0	0
Siemens Clinitek 500		50	0	0	0	0	0	0
Siemens Clinitek Advantus		40	0	0	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		37	0	0	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
Other		6	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		181	0	0	0	0	0	0
Siemens Atlas Reagent		85	0	0	0	0	0	0
Roche Chemstrips		48	0	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0	0
Bayer Clinitest		50	1	0	0	0	0	0
Siemens Ictotest		3	0	0	0	0	0	0
Hitachi/Roche Reagents		3	0	0	0	0	0	0
Other		8	0	0	0	0	0	0
All Instruments and Reagents		433	1	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Glucose	Specimen: U02	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		9	34	38	33	0	2	0
Siemens Atlas		1	0	93	2	0	0	0
Siemens Clinitek Status		1	42	1	0	0	0	0
Siemens Clinitek 50		0	7	0	0	0	0	0
Siemens Clinitek 100		4	3	1	1	0	0	0
Siemens Clinitek 500		9	42	1	1	0	0	0
Siemens Clinitek Advantus		3	37	1	0	0	0	0
Roche UA Reader/Super UA Read		0	0	2	0	0	0	0
Roche Chemstrip Criterion		0	1	12	1	0	0	0
Roche Urisys		0	8	29	0	0	0	0
IRIS AUTION MAX		0	0	35	2	0	2	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other Chemistry Analyzer		0	0	0	1	0	0	0
Other		1	0	3	2	0	0	0
<Reagents>								
Siemens Reagent Strips		26	142	12	2	0	0	0
Siemens Atlas Reagent		0	0	83	2	0	0	0
Roche Chemstrips		0	12	39	1	0	0	0
Roche Urisys Cassette		0	0	14	0	0	0	0
ARKRAY Aution Sticks		0	0	28	1	0	1	0
IRIS Reagent Strips		0	0	8	2	0	1	0
Bayer Clinitest		1	10	31	34	0	2	0
Siemens Ictotest		0	1	0	1	0	0	0
Bayer Diastix		0	1	1	0	0	0	0
Hitachi/Roche Reagents		0	1	2	0	0	0	0
Other		1	7	0	0	0	0	0
All Instruments and Reagents		28	174	218	43	0	4	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Glucose	Specimen: U03	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		24	58	34	1	0	0	0
Siemens Atlas		0	93	2	1	0	0	0
Siemens Clinitek Status		0	44	0	0	0	0	0
Siemens Clinitek 50		0	7	0	0	0	0	0
Siemens Clinitek 100		5	4	0	0	0	0	0
Siemens Clinitek 500		4	48	1	0	0	0	0
Siemens Clinitek Advantus		2	39	0	0	0	0	0
Roche UA Reader/Super UA Read		0	2	0	0	0	0	0
Roche Chemstrip Criterion		0	13	1	0	0	0	0
Roche Urisys		0	37	0	0	0	0	0
IRIS AUTION MAX		0	36	2	1	0	0	0
Hitachi/Roche Analyzer		0	2	0	0	0	0	0
Other Chemistry Analyzer		1	0	0	0	0	0	0
Other		1	3	2	0	0	0	0
<Reagents>								
Siemens Reagent Strips		16	164	2	0	0	0	0
Siemens Atlas Reagent		0	83	2	0	0	0	0
Roche Chemstrips		0	50	1	0	0	0	0
Roche Urisys Cassette		0	14	0	0	0	0	0
ARKRAY Aution Sticks		0	29	1	0	0	0	0
IRIS Reagent Strips		0	8	2	1	0	0	0
Bayer Clinitest		19	26	33	2	0	0	0
Siemens Ictotest		0	1	1	0	0	0	0
Bayer Diastix		0	2	0	0	0	0	0
Hitachi/Roche Reagents		0	3	0	0	0	0	0
Other		2	6	0	0	0	0	0
All Instruments and Reagents		37	386	42	3	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Glucose	Specimen: U04	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		3	28	72	15	0	0	0
Siemens Atlas		0	33	60	2	0	1	0
Siemens Clinitek Status		0	30	14	0	0	0	0
Siemens Clinitek 50		0	0	7	0	0	0	0
Siemens Clinitek 100		0	6	3	0	0	0	0
Siemens Clinitek 500		0	14	39	0	0	0	0
Siemens Clinitek Advantus		1	14	26	0	0	0	0
Roche UA Reader/Super UA Read		0	2	0	0	0	0	0
Roche Chemstrip Criterion		0	12	2	0	0	0	0
Roche Urisys		0	26	11	0	0	0	0
IRIS AUTION MAX		0	9	28	1	1	0	0
Hitachi/Roche Analyzer		0	1	1	0	0	0	0
Other Chemistry Analyzer		0	0	1	0	0	0	0
Other		0	2	2	2	0	0	0
<Reagents>								
Siemens Reagent Strips		0	73	107	2	0	0	0
Siemens Atlas Reagent		0	30	53	2	0	0	0
Roche Chemstrips		0	40	11	0	0	0	0
Roche Urisys Cassette		0	5	9	0	0	0	0
ARKRAY Aution Sticks		0	7	22	1	0	0	0
IRIS Reagent Strips		0	2	8	0	1	0	0
Bayer Clinitest		4	15	47	13	0	1	0
Siemens Ictotest		0	1	1	0	0	0	0
Bayer Diastix		0	0	2	1	0	0	0
Hitachi/Roche Reagents		0	1	2	0	0	0	0
Other		0	3	4	1	0	0	0
All Instruments and Reagents		4	177	266	20	1	1	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Glucose	Specimen: U05	Negative	Trace	50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>									
Visual examination		88	0	0	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0	0	0
Siemens Clinitek Status		44	0	0	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0	0	0
Siemens Clinitek 500		50	0	0	0	0	0	0	0
Siemens Clinitek Advantus		40	0	0	0	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0	0
Roche Urisys		37	0	0	0	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0	0
Other		6	0	0	0	0	0	0	0
<Reagents>									
Siemens Reagent Strips		181	0	0	0	0	0	0	0
Siemens Atlas Reagent		85	0	0	0	0	0	0	0
Roche Chemstrips		48	0	0	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0	0	0
Bayer Clinitest		51	0	0	0	0	0	0	0
Siemens Ictotest		3	0	0	0	0	0	0	0
Hitachi/Roche Reagents		3	0	0	0	0	0	0	0
Other		8	0	0	0	0	0	0	0
All Instruments and Reagents		434	0	0	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Hemoglobin	Specimen: U01	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		39	0	0	0	0	0
Siemens Atlas		95	0	0	0	0	0
Siemens Clinitek Status		44	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		50	0	0	0	0	0
Siemens Clinitek Advantus		39	0	1	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
Other		5	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	1	0	0	0
Siemens Atlas Reagent		88	0	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		6	0	0	0	0	0
All Instruments and Reagents		379	0	1	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Hemoglobin	Specimen: U02	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		39	0	0	0	0	0
Siemens Atlas		95	0	0	0	0	0
Siemens Clinitek Status		44	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		7	1	0	0	0	0
Siemens Clinitek 500		49	0	0	0	0	1
Siemens Clinitek Advantus		39	1	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
Other		5	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		179	2	0	0	0	1
Siemens Atlas Reagent		88	0	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		6	0	0	0	0	0
All Instruments and Reagents		377	2	0	0	0	1

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Hemoglobin	Specimen: U03	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		0	0	0	1	0	41
Siemens Atlas		0	0	0	2	0	93
Siemens Clinitek Status		0	0	0	5	0	39
Siemens Clinitek 50		0	0	0	1	0	6
Siemens Clinitek 100		0	0	0	1	0	7
Siemens Clinitek 500		0	0	0	2	0	49
Siemens Clinitek Advantus		0	0	0	0	0	40
Roche UA Reader/Super UA Read		0	0	0	0	0	2
Roche Chemstrip Criterion		0	0	0	0	0	14
Roche Urisys		0	0	0	3	0	33
IRIS AUTION MAX		0	0	1	24	0	14
Hitachi/Roche Analyzer		0	0	0	0	0	1
Other		0	0	0	3	0	2
<Reagents>							
Siemens Reagent Strips		0	0	0	9	0	176
Siemens Atlas Reagent		0	0	0	2	0	86
Roche Chemstrips		0	0	0	4	0	44
Roche Urisys Cassette		0	0	0	0	0	14
ARKRAY Aution Sticks		0	0	0	18	0	12
IRIS Reagent Strips		0	0	1	8	0	2
Siemens Hemastix		0	0	0	0	0	1
None, no reagent or strip use		0	0	0	0	0	1
Other		0	0	0	1	0	5
All Instruments and Reagents		0	0	1	42	0	341

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Hemoglobin	Specimen: U04	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		38	1	0	0	0	0
Siemens Atlas		95	0	0	0	0	0
Siemens Clinitek Status		39	5	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		7	1	0	0	0	0
Siemens Clinitek 500		48	2	0	0	0	0
Siemens Clinitek Advantus		37	3	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0
Hitachi/Roche Analyzer		0	1	0	0	0	0
Other		5	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		170	12	0	0	0	0
Siemens Atlas Reagent		88	0	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		5	1	0	0	0	0
All Instruments and Reagents		367	13	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Hemoglobin	Specimen: U05	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		0	0	3	36	0	3
Siemens Atlas		1	0	1	93	0	0
Siemens Clinitek Status		1	0	0	43	0	0
Siemens Clinitek 50		0	1	1	5	0	0
Siemens Clinitek 100		0	0	0	8	0	0
Siemens Clinitek 500		0	0	2	47	0	2
Siemens Clinitek Advantus		0	0	1	39	0	0
Roche UA Reader/Super UA Read		0	0	0	0	1	1
Roche Chemstrip Criterion		0	0	1	4	8	1
Roche Urisys		0	0	6	23	6	1
IRIS AUTION MAX		0	0	32	7	0	0
Hitachi/Roche Analyzer		0	0	0	0	0	1
Other		0	0	1	3	1	0
<Reagents>							
Siemens Reagent Strips		1	1	4	174	0	5
Siemens Atlas Reagent		1	0	1	86	0	0
Roche Chemstrips		0	0	7	22	16	3
Roche Urisys Cassette		0	0	2	12	0	0
ARKRAY Aution Sticks		0	0	26	4	0	0
IRIS Reagent Strips		0	0	7	4	0	0
Siemens Hemastix		0	0	0	1	0	0
None, no reagent or strip use		0	0	1	0	0	0
Other		0	0	0	5	0	1
All Instruments and Reagents		2	1	48	308	16	9

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Ketones	Specimen: U01	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		45	0	0	0	0	0
Siemens Atlas		95	1	0	0	0	0
Siemens Clinitek Status		43	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		50	0	0	0	0	0
Siemens Clinitek Advantus		41	0	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0
Other		5	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	0	0	0	0
Siemens Atlas Reagent		87	1	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0
Siemens Acetest		7	0	0	0	0	0
Siemens Ictotest		1	0	0	0	0	0
Bayer Diastix		1	0	0	0	0	0
Other		6	0	0	0	0	0
All Instruments and Reagents		385	1	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Ketones	Specimen: U02	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		0	0	6	41	14	2
Siemens Atlas		2	0	1	93	0	0
Siemens Clinitek Status		0	0	1	23	19	0
Siemens Clinitek 50		0	0	0	7	0	0
Siemens Clinitek 100		0	0	0	8	0	0
Siemens Clinitek 500		0	0	5	43	3	0
Siemens Clinitek Advantus		0	0	4	37	0	0
Roche UA Reader/Super UA Read		0	0	0	2	0	0
Roche Chemstrip Criterion		0	0	0	14	0	0
Roche Urisys		0	0	0	30	6	0
IRIS AUTION MAX		0	0	1	38	0	0
Other		0	0	1	3	1	0
<Reagents>							
Siemens Reagent Strips		0	0	9	141	32	0
Siemens Atlas Reagent		2	0	1	85	0	0
Roche Chemstrips		0	0	1	40	7	0
Roche Urisys Cassette		0	0	0	14	0	0
ARKRAY Aution Sticks		0	0	0	30	0	0
IRIS Reagent Strips		0	0	2	9	0	0
Siemens Acetest		0	0	5	13	3	2
Siemens Ictotest		0	0	0	1	0	0
Bayer Diastix		0	0	0	1	0	0
Other		0	0	1	5	1	0
All Instruments and Reagents		2	0	19	339	43	2

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Ketones	Specimen: U03	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		45	0	0	0	0	0
Siemens Atlas		95	0	1	0	0	0
Siemens Clinitek Status		43	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		50	0	0	0	0	0
Siemens Clinitek Advantus		41	0	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0
Other		5	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	0	0	0	0
Siemens Atlas Reagent		87	0	1	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0
Siemens Acetest		7	0	0	0	0	0
Siemens Ictotest		1	0	0	0	0	0
Bayer Diastix		1	0	0	0	0	0
Other		6	0	0	0	0	0
All Instruments and Reagents		385	0	1	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Ketones	Specimen: U04	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		45	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		43	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		50	0	0	0	0	0
Siemens Clinitek Advantus		41	0	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		35	0	1	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0
Other		5	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	0	0	0	0
Siemens Atlas Reagent		88	0	0	0	0	0
Roche Chemstrips		46	0	1	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0
Siemens Acetest		7	0	0	0	0	0
Siemens Ictotest		1	0	0	0	0	0
Bayer Diastix		1	0	0	0	0	0
Other		6	0	0	0	0	0
All Instruments and Reagents		385	0	1	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Ketones	Specimen: U05	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		45	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		43	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		50	0	0	0	0	0
Siemens Clinitek Advantus		41	0	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		39	0	0	0	0	0
Other		5	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	0	0	0	0
Siemens Atlas Reagent		88	0	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		30	0	0	0	0	0
IRIS Reagent Strips		11	0	0	0	0	0
Siemens Acetest		7	0	0	0	0	0
Siemens Ictotest		1	0	0	0	0	0
Bayer Diastix		1	0	0	0	0	0
Other		6	0	0	0	0	0
All Instruments and Reagents		386	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Protein	Specimen: U01	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		2	11	64	5	0	0	20
Siemens Atlas		0	0	95	0	1	0	0
Siemens Clinitek Status		0	0	24	19	0	0	0
Siemens Clinitek 50		0	0	4	3	0	0	0
Siemens Clinitek 100		0	0	7	1	0	0	1
Siemens Clinitek 500		0	11	40	0	0	0	0
Siemens Clinitek Advantus		0	9	32	0	0	0	0
Roche UA Reader/Super UA Read		0	0	1	1	0	0	0
Roche Chemstrip Criterion		0	1	13	0	0	0	0
Roche Urisys		3	3	18	12	0	0	0
IRIS AUTION MAX		0	1	38	0	0	0	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other		0	0	5	0	0	0	0
<Reagents>								
Siemens Reagent Strips		0	23	137	22	0	0	0
Siemens Atlas Reagent		0	0	88	0	1	0	0
Sulfosalicylic Acid		2	8	27	3	0	0	21
Roche Chemstrips		1	3	34	9	0	0	0
Roche Urisys Cassette		2	1	6	5	0	0	0
ARKRAY Aution Sticks		0	1	29	0	0	0	0
IRIS Reagent Strips		0	0	11	0	0	0	0
Bayer Diastix		0	0	1	0	0	0	0
Hitachi/Roche Reagents		0	0	1	0	0	0	0
Other		0	0	9	2	0	0	0
All Instruments and Reagents		5	36	343	41	1	0	21

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Protein	Specimen: U02	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		1	2	47	31	0	0	21
Siemens Atlas		1	0	40	55	0	0	0
Siemens Clinitek Status		0	0	0	43	0	0	0
Siemens Clinitek 50		0	0	0	7	0	0	0
Siemens Clinitek 100		0	0	3	5	0	0	1
Siemens Clinitek 500		0	0	47	4	0	0	0
Siemens Clinitek Advantus		0	0	36	5	0	0	0
Roche UA Reader/Super UA Read		0	0	0	2	0	0	0
Roche Chemstrip Criterion		0	0	14	0	0	0	0
Roche Urisys		0	4	6	26	0	0	0
IRIS AUTION MAX		0	1	36	2	0	0	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other		0	0	2	3	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	106	76	0	0	0
Siemens Atlas Reagent		1	0	36	52	0	0	0
Sulfosalicylic Acid		1	2	22	14	0	0	22
Roche Chemstrips		0	1	25	21	0	0	0
Roche Urisys Cassette		0	3	1	10	0	0	0
ARKRAY Aution Sticks		0	1	28	1	0	0	0
IRIS Reagent Strips		0	0	9	2	0	0	0
Bayer Diastix		0	0	1	0	0	0	0
Hitachi/Roche Reagents		0	0	1	0	0	0	0
Other		0	0	4	7	0	0	0
All Instruments and Reagents		2	7	233	183	0	0	22

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Protein	Specimen: U03	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	0	3	54	23	21
Siemens Atlas		0	0	0	1	77	19	0
Siemens Clinitek Status		0	0	0	0	42	1	0
Siemens Clinitek 50		0	0	0	0	7	0	0
Siemens Clinitek 100		0	0	0	1	7	0	1
Siemens Clinitek 500		0	0	0	4	47	0	0
Siemens Clinitek Advantus		0	0	0	1	40	0	0
Roche UA Reader/Super UA Read		0	0	0	0	2	0	0
Roche Chemstrip Criterion		0	0	0	0	13	1	0
Roche Urisys		0	0	0	1	35	0	0
IRIS AUTION MAX		0	0	0	1	37	1	0
Hitachi/Roche Analyzer		0	0	0	0	2	0	0
Other		0	0	0	0	5	0	0
<Reagents>								
Siemens Reagent Strips		0	0	0	6	165	11	0
Siemens Atlas Reagent		0	0	0	1	69	19	0
Sulfosalicylic Acid		0	0	0	3	24	11	22
Roche Chemstrips		0	0	0	1	45	1	0
Roche Urisys Cassette		0	0	0	0	14	0	0
ARKRAY Aution Sticks		0	0	0	0	29	1	0
IRIS Reagent Strips		0	0	0	1	10	0	0
Bayer Diastix		0	0	0	0	1	0	0
Hitachi/Roche Reagents		0	0	0	0	1	0	0
Other		0	0	0	0	10	2	0
All Instruments and Reagents		0	0	0	12	368	45	22

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Protein	Specimen: U04	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	15	63	3	0	21
Siemens Atlas		0	0	1	95	0	0	0
Siemens Clinitek Status		0	0	1	41	1	0	0
Siemens Clinitek 50		0	0	0	5	2	0	0
Siemens Clinitek 100		0	0	0	8	0	0	1
Siemens Clinitek 500		0	0	2	49	0	0	0
Siemens Clinitek Advantus		0	0	3	38	0	0	0
Roche UA Reader/Super UA Read		0	0	0	2	0	0	0
Roche Chemstrip Criterion		0	0	3	10	1	0	0
Roche Urisys		1	0	4	31	0	0	0
IRIS AUTION MAX		0	0	24	15	0	0	0
Hitachi/Roche Analyzer		0	0	0	2	0	0	0
Other		0	0	0	5	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	8	169	5	0	0
Siemens Atlas Reagent		0	0	1	88	0	0	0
Sulfosalicylic Acid		0	0	12	26	1	0	22
Roche Chemstrips		1	0	6	39	1	0	0
Roche Urisys Cassette		0	0	1	13	0	0	0
ARKRAY Aution Sticks		0	0	19	11	0	0	0
IRIS Reagent Strips		0	0	5	6	0	0	0
Bayer Diastix		0	0	0	1	0	0	0
Hitachi/Roche Reagents		0	0	0	1	0	0	0
Other		0	0	1	10	0	0	0
All Instruments and Reagents		1	0	53	364	7	0	22

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Protein	Specimen: U05	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		1	0	1	4	60	15	21
Siemens Atlas		1	0	0	1	90	5	0
Siemens Clinitek Status		0	0	0	0	41	2	0
Siemens Clinitek 50		0	0	0	0	7	0	0
Siemens Clinitek 100		0	0	0	0	8	0	1
Siemens Clinitek 500		0	0	0	9	42	0	0
Siemens Clinitek Advantus		0	0	0	5	36	0	0
Roche UA Reader/Super UA Read		0	0	0	0	2	0	0
Roche Chemstrip Criterion		0	0	0	0	13	1	0
Roche Urisys		0	0	1	1	34	0	0
IRIS AUTION MAX		0	0	1	3	35	0	0
Hitachi/Roche Analyzer		0	0	0	0	2	0	0
Other		0	0	0	0	5	0	0
<Reagents>								
Siemens Reagent Strips		0	0	0	15	159	8	0
Siemens Atlas Reagent		1	0	0	1	82	5	0
Sulfosalicylic Acid		0	0	1	3	26	9	22
Roche Chemstrips		1	0	1	1	43	1	0
Roche Urisys Cassette		0	0	0	0	14	0	0
ARKRAY Aution Sticks		0	0	1	1	28	0	0
IRIS Reagent Strips		0	0	0	2	9	0	0
Bayer Diastix		0	0	0	0	1	0	0
Hitachi/Roche Reagents		0	0	0	0	1	0	0
Other		0	0	0	0	12	0	0
All Instruments and Reagents		2	0	3	23	375	23	22

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Qualitative hCG report:

Bayer Clinitek Status	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	2	2	0	2	0
Indeterminant	0	0	0	0	0
Positive	0	0	2	0	2

Beckman Coulter ICON II	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	2	2	0	2	0
Indeterminant	0	0	0	0	0
Positive	0	0	2	0	2

Beckman Coulter ICON 25	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	25	25	0	25	0
Indeterminant	0	0	0	0	0
Positive	0	0	25	0	25

Beckman Coulter ICON 20	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	23	23	0	23	0
Indeterminant	0	0	0	0	0
Positive	0	0	23	0	23

Cardinal Health hCG Combo	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	17	17	0	17	0
Indeterminant	0	0	0	0	0
Positive	0	0	17	0	17

Wampole Clearview hCG Combo	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	2	2	0	2	0
Indeterminant	0	0	0	0	0
Positive	0	0	2	0	2

Cen Med Elite	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	19	19	0	19	0
Indeterminant	0	0	0	0	0
Positive	0	0	19	0	19

Wampole Clearview hCG II	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	5	5	0	5	0
Indeterminant	0	0	0	0	0
Positive	0	0	5	0	5

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Genzyme OSOM Combo	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	17	17	0	17	0
Indeterminant	0	0	0	0	0
Positive	0	0	17	0	17
Inverness Medical BioStar Acceava	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	3	3	0	3	0
Indeterminant	0	0	0	0	0
Positive	0	0	3	0	3
Lifesign Status (serum/urine)	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	2	2	0	2	0
Indeterminant	0	0	0	0	0
Positive	0	0	2	0	2
Polymedco Poly stat	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	6	6	0	6	0
Indeterminant	0	0	0	0	0
Positive	0	0	6	0	6
Quidel QuickVue	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	78	78	0	78	0
Indeterminant	0	0	0	0	0
Positive	0	0	78	0	78
SA Scientific	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	2	2	0	2	0
Indeterminant	0	0	0	0	0
Positive	0	0	2	0	2
Stanbio QuPID	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	18	17	1	17	1
Indeterminant	0	0	0	0	0
Positive	0	1	17	1	17
Fisher Sure-Vue serum/urine	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	42	42	0	42	0
Indeterminant	0	0	0	0	0
Positive	0	0	42	0	42

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

Fisher Sure-Vue STAT serum/urine	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	23	23	0	23	0
Indeterminant	0	0	0	0	0
Positive	0	0	23	0	23
Signify hCG Urine/Serum	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	5	5	0	5	0
Indeterminant	0	0	0	0	0
Positive	0	0	5	0	5
Sacks Medical RefuAH	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	2	2	0	2	0
Indeterminant	0	0	0	0	0
Positive	0	0	2	0	2
Fisher Sure-Vue (urine only)	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	11	11	0	11	0
Indeterminant	0	0	0	0	0
Positive	0	0	11	0	11
True Medix	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	4	4	0	4	0
Indeterminant	0	0	0	0	0
Positive	0	0	4	0	4
Method not listed	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	25	25	0	25	0
Indeterminant	0	0	0	0	0
Positive	0	0	25	0	25
All Methods & Instruments	Specimen:U01	Specimen:U02	Specimen:U03	Specimen:U04	Specimen:U05
Negative	333	332	1	332	1
Indeterminant	0	0	0	0	0
Positive	0	1	332	1	332

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 9 April 2012

hCG Summary:

[Code], hCG Kit name	U01			U02			U03			U04			U05			
	N	I	P	N	I	P	N	I	P	N	I	P	N	I	P	
[BA] Bayer Clinitek Status	2	0	0	2	0	0	0	0	2	2	0	0	0	0	0	2
[BC] Beckman Coulter ICON II	2	0	0	2	0	0	0	0	2	2	0	0	0	0	0	2
[BN] Beckman Coulter ICON 25	25	0	0	25	0	0	0	0	25	25	0	0	0	0	0	25
[BT] Beckman Coulter ICON 20	23	0	0	23	0	0	0	0	23	23	0	0	0	0	0	23
[CA] Cardinal Health hCG Combo	17	0	0	17	0	0	0	0	17	17	0	0	0	0	0	17
[CC] Wampole Clearview hCG Combo	2	0	0	2	0	0	0	0	2	2	0	0	0	0	0	2
[CM] Cen Med Elite	19	0	0	19	0	0	0	0	19	19	0	0	0	0	0	19
[CV] Wampole Clearview hCG II	5	0	0	5	0	0	0	0	5	5	0	0	0	0	0	5
[GO] Genzyme OSOM Combo	17	0	0	17	0	0	0	0	17	17	0	0	0	0	0	17
[IM] Inverness Medical BioStar Acceava	3	0	0	3	0	0	0	0	3	3	0	0	0	0	0	3
[LS] Lifesign Status (serum/urine)	2	0	0	2	0	0	0	0	2	2	0	0	0	0	0	2
[PM] Polymedco Poly stat	6	0	0	6	0	0	0	0	6	6	0	0	0	0	0	6
[QQ] Quidel QuickVue	78	0	0	78	0	0	0	0	78	78	0	0	0	0	0	78
[SA] SA Scientific	2	0	0	2	0	0	0	0	2	2	0	0	0	0	0	2
[SB] Stanbio QuPID	18	0	0	17	0	1	1	0	17	17	0	1	1	0	0	17
[SE] Fisher Sure-Vue serum/urine	42	0	0	42	0	0	0	0	42	42	0	0	0	0	0	42
[SF] Fisher Sure-Vue STAT serum/urine	23	0	0	23	0	0	0	0	23	23	0	0	0	0	0	23
[SG] Signify hCG Urine/Serum	5	0	0	5	0	0	0	0	5	5	0	0	0	0	0	5
[SM] Sacks Medical RefuAH	2	0	0	2	0	0	0	0	2	2	0	0	0	0	0	2
[SU] Fisher Sure-Vue (urine only)	11	0	0	11	0	0	0	0	11	11	0	0	0	0	0	11
[TX] True Medix	4	0	0	4	0	0	0	0	4	4	0	0	0	0	0	4
[90] Method not listed	25	0	0	25	0	0	0	0	25	25	0	0	0	0	0	25
[99] All hCG Methods	333	0	0	332	0	1	1	0	332	332	0	1	1	0	0	332