

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

6-Acetylmorphine

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	4	2	2	0	0	0
GCMS	5	1	1	0	0	0
GCMS	10	22	22	0	0	0
GCMS	15	1	1	0	0	0
GCMS	300	1	0	1	0	0
LCMM	2	1	1	0	0	0
LCMM	5	1	1	0	0	0
LCMM	10	8	7	0	0	0
LCMM	20	2	1	1	0	0
LCMM	50	1	1	0	0	0
TOXI	200	2	0	2	0	0
TOXI	1000	1	0	1	0	0

Alprazolam

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	50	1	0	1	0	0
GCMS	100	3	0	3	0	0
GCMS	150	1	0	1	0	0
GCMS	200	3	0	3	0	0
GCMS	300	1	0	1	0	0
LCDA	200	1	0	1	0	0
LCDA	400	1	0	1	0	0
LCMM	5	1	0	1	0	0
LCMM	15	1	0	1	0	0
LCMM	20	1	0	1	0	0
LCMM	25	1	0	1	0	0
LCMM	50	4	0	4	0	0
LCMM	60	1	0	1	0	0
LCMM	75	2	0	2	0	0
LCMM	100	2	0	2	0	0
LCMM	200	2	0	2	0	0
LCMM	300	3	0	3	0	0

Amitriptyline

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	50	1	1	0	0	0
GCMS	100	1	1	0	0	0
GCMS	250	1	1	0	0	0
GCMS	500	1	1	0	0	0
GCNP	100	1	1	0	0	0
LCFW	25	1	1	0	0	0
LCMM	10	2	1	0	1	0
LCMM	25	1	1	0	0	0
LCMM	50	4	4	0	0	0
LCMM	100	4	4	0	0	0
LCMM	150	1	1	0	0	0
LCMM	200	2	2	0	0	0

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

TOXI	500	2	2	0	0	0
------	-----	---	---	---	---	---

Amobarbital

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCFL	1000	1	0	1	0	0
GCMS	20	1	0	1	0	0
GCMS	50	2	0	2	0	0
GCMS	100	5	0	5	0	0
GCMS	150	1	0	1	0	0
GCMS	200	18	0	18	0	0
GCMS	300	5	0	5	0	0
GCNP	500	1	0	0	0	0
LCDA	200	1	0	0	0	0
LCMM	40	1	0	1	0	0
LCMM	100	1	0	1	0	0
TLCC	300	1	0	0	0	0
TLCC	1000	1	0	1	0	0

Amphetamine

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCMS	50	3	3	0	0	0
GCMS	100	1	1	0	0	0
GCMS	200	3	3	0	0	0
GCMS	250	8	8	0	0	0
GCMS	300	1	1	0	0	0
GCMS	500	17	17	0	0	0
GCMS	1000	1	1	0	0	0
LCDA	800	1	0	1	0	0
LCMM	20	1	1	0	0	0
LCMM	25	1	1	0	0	0
LCMM	50	4	4	0	0	0
LCMM	100	3	3	0	0	0
LCMM	125	2	2	0	0	0
LCMM	200	2	2	0	0	0
LCMM	500	1	1	0	0	0
LCMS	500	1	1	0	0	0
TLCC	250	1	1	0	0	0
TLCC	1000	1	1	0	0	0
TOXI	500	3	1	0	0	0

Barbiturates

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
TOXI	175	1	1	0	0	0
TOXI	500	1	1	0	0	0
TOXI	1000	1	1	0	0	0

Benzodiazepines

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
TLCC	1000	1	1	0	0	0
TOXI	100	2	2	0	0	0

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

TOXI	1000	1	1	0	0	0
------	------	---	---	---	---	---

Benzoylcegonine

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCMS	40	1	1	0	0	0
GCMS	50	2	2	0	0	0
GCMS	100	8	8	0	0	0
GCMS	150	9	9	0	0	0
LCMM	5	1	1	0	0	0
LCMM	25	1	1	0	0	0
LCMM	30	2	2	0	0	0
LCMM	50	2	2	0	0	0
LCMM	100	6	6	0	0	0
TOXI	1000	1	0	1	0	0

Butalbital

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCFL	500	1	1	0	0	0
GCFL	1500	1	1	0	0	0
GCMS	40	1	1	0	0	0
GCMS	50	4	4	0	0	0
GCMS	100	4	4	0	0	0
GCMS	150	1	1	0	0	0
GCMS	200	18	18	0	0	0
GCMS	300	6	6	0	0	0
GCNP	500	1	1	0	0	0
LCDA	200	1	1	0	0	0
LCMM	20	1	1	0	0	0
LCMM	50	1	1	0	0	0
LCMM	100	5	5	0	0	0
LCMM	200	1	1	0	0	0
TLCC	300	1	1	0	0	0
TLCC	1000	1	1	0	0	0

Codeine

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCMS	50	2	0	2	0	0
GCMS	100	4	0	4	0	0
GCMS	150	1	0	1	0	0
GCMS	200	1	0	1	0	0
GCMS	300	15	0	15	0	0
GCMS	2000	5	0	5	0	0
GCNP	500	1	0	1	0	0
LCMM	20	1	0	1	0	0
LCMM	25	1	0	1	0	0
LCMM	30	1	0	1	0	0
LCMM	50	7	0	7	0	0
LCMM	100	6	0	6	0	0
LCMM	300	3	0	3	0	0
THSC	2000	1	0	1	0	0

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

TLCC	300	1	0	1	0	0
TLCC	500	1	0	1	0	0
TOXI	200	2	0	2	0	0
TOXI	500	1	0	1	0	0

Doxepin

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCMS	50	1	0	1	0	0
GCMS	100	1	0	1	0	0
GCMS	250	1	0	1	0	0
LCFW	25	1	0	1	0	0
LCMM	50	3	0	3	0	0
LCMM	100	1	0	1	0	0
LCMM	150	1	0	1	0	0
LCMM	200	2	0	2	0	0
TOXI	500	1	0	1	0	0

EDDP

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCMS	100	2	2	0	0	0
GCMS	150	2	2	0	0	0
GCMS	200	2	2	0	0	0
GCMS	250	1	1	0	0	0
GCMS	300	3	3	0	0	0
LCMM	5	1	1	0	0	0
LCMM	10	1	1	0	0	0
LCMM	25	1	1	0	0	0
LCMM	50	2	2	0	0	0
LCMM	100	10	10	0	0	0
LCMM	200	2	2	0	0	0
LCMM	300	2	2	0	0	0
THSD	150	1	1	0	0	0
TLCC	500	1	1	0	0	0
TOXI	1000	1	0	1	0	0

Ethanol

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
BECO	10	2	2	0	0	0
BECO	20	2	2	0	0	0
GCED	10	2	2	0	0	0
GCFL	10	24	24	0	0	0
GCFL	20	29	29	0	0	0
GCFL	40	4	4	0	0	0
GCFL	50	2	2	0	0	0
GCMS	20	2	2	0	0	0
ROON	20	2	2	0	0	0
THSD	50	2	2	0	0	0
ZZZ	20	2	2	0	0	0

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

Methadone

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCMS	50	1	1	0	0	0
GCMS	100	5	5	0	0	0
GCMS	150	4	4	0	0	0
GCMS	200	4	4	0	0	0
GCMS	250	1	1	0	0	0
GCMS	300	13	13	0	0	0
GCNP	400	1	1	0	0	0
LCMM	3	1	1	0	0	0
LCMM	10	1	1	0	0	0
LCMM	25	1	1	0	0	0
LCMM	50	2	2	0	0	0
LCMM	100	10	10	0	0	0
LCMM	200	2	2	0	0	0
LCMM	300	2	2	0	0	0
TLCC	300	1	1	0	0	0
TLCC	500	1	1	0	0	0
TOXI	250	1	1	0	0	0
TOXI	500	2	2	0	0	0

Methamphetamine

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCMM	125	1	1	0	0	0
GCMS	50	3	3	0	0	0
GCMS	100	1	1	0	0	0
GCMS	200	3	3	0	0	0
GCMS	250	8	8	0	0	0
GCMS	300	1	1	0	0	0
GCMS	500	17	17	0	0	0
GCMS	1000	1	1	0	0	0
LCDA	800	1	1	0	0	0
LCMM	50	5	5	0	0	0
LCMM	100	3	3	0	0	0
LCMM	125	1	1	0	0	0
LCMM	200	2	2	0	0	0
LCMM	250	1	1	0	0	0
LCMM	500	1	1	0	0	0
LCMS	500	1	1	0	0	0
TLCC	250	1	1	0	0	0
TLCC	1000	1	0	0	0	0
TOXI	500	3	3	0	0	0

Morphine

<i>Method Code</i>	<i>Cutoff</i>	<i>No. Labs</i>	<i>True Positive</i>	<i>True Negative</i>	<i>False Positive</i>	<i>False Negative</i>
GCMS	50	2	2	0	0	0
GCMS	100	4	4	0	0	0
GCMS	150	1	1	0	0	0
GCMS	200	1	1	0	0	0

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

GCMS	300	15	15	0	0	0
GCMS	2000	6	6	0	0	0
GCNP	400	1	1	0	0	0
LCMM	20	1	1	0	0	0
LCMM	25	2	2	0	0	0
LCMM	30	1	1	0	0	0
LCMM	50	7	7	0	0	0
LCMM	100	6	6	0	0	0
LCMM	300	3	3	0	0	0
TLCC	300	1	1	0	0	0
TLCC	500	1	1	0	0	0
TOXI	200	2	2	0	0	0
TOXI	1000	1	1	0	0	0

Nordiazepam

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	50	1	0	0	1	0
GCMS	100	4	0	4	0	0
GCMS	150	1	0	1	0	0
GCMS	200	8	0	8	0	0
GCMS	300	4	0	4	0	0
LCDA	200	1	0	1	0	0
LCDA	400	1	0	1	0	0
LCMM	10	1	0	1	0	0
LCMM	20	2	0	2	0	0
LCMM	25	1	0	1	0	0
LCMM	40	2	0	2	0	0
LCMM	50	5	0	5	0	0
LCMM	75	2	0	2	0	0
LCMM	100	5	0	5	0	0
LCMM	200	2	0	2	0	0
LCMM	300	4	0	4	0	0
LCMS	100	1	0	1	0	0

Nordoxepin

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	50	1	0	1	0	0
GCMS	100	1	0	1	0	0
GCMS	250	1	0	1	0	0
LCFW	25	1	0	1	0	0
LCMM	50	1	0	1	0	0
LCMM	150	1	0	1	0	0
TOXI	500	1	0	1	0	0

Norpropoxyphene

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCFL	2000	1	1	0	0	0
GCMS	50	1	1	0	0	0
GCMS	100	3	3	0	0	0
GCMS	150	2	2	0	0	0

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

GCMS	200	8	8	0	0	0
GCMS	300	10	10	0	0	0
LCMM	10	1	1	0	0	0
LCMM	50	2	2	0	0	0
LCMM	100	6	6	0	0	0
LCMM	200	1	1	0	0	0
LCMM	300	1	1	0	0	0
TOXI	2000	1	0	1	0	0

Nortriptyline

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	50	1	1	0	0	0
GCMS	100	1	1	0	0	0
GCMS	250	1	1	0	0	0
GCMS	500	1	1	0	0	0
GCNP	100	1	1	0	0	0
LCFW	25	1	1	0	0	0
LCMM	10	2	1	0	1	0
LCMM	25	1	1	0	0	0
LCMM	50	4	4	0	0	0
LCMM	100	4	4	0	0	0
LCMM	150	1	1	0	0	0
LCMM	200	2	2	0	0	0
TOXI	500	2	2	0	0	0

Oxazepam

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	50	1	1	0	0	0
GCMS	100	5	5	0	0	0
GCMS	150	1	1	0	0	0
GCMS	200	8	8	0	0	0
GCMS	300	5	5	0	0	0
LCDA	200	1	1	0	0	0
LCDA	400	1	1	0	0	0
LCMM	5	1	1	0	0	0
LCMM	20	2	2	0	0	0
LCMM	25	1	1	0	0	0
LCMM	40	2	2	0	0	0
LCMM	50	5	5	0	0	0
LCMM	75	2	2	0	0	0
LCMM	100	5	5	0	0	0
LCMM	200	2	2	0	0	0
LCMM	300	4	4	0	0	0
LCMS	100	1	1	0	0	0
TLCC	300	1	1	0	0	0

Pentobarbital

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCFL	500	1	0	1	0	0
GCMS	50	4	0	4	0	0

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

GCMS	60	1	0	1	0	0
GCMS	100	4	0	4	0	0
GCMS	150	1	0	1	0	0
GCMS	200	18	0	18	0	0
GCMS	300	5	0	5	0	0
GCNP	500	1	0	1	0	0
LCDA	200	1	0	1	0	0
LCMM	40	1	0	1	0	0
LCMM	100	3	0	3	0	0
TLCC	300	1	0	1	0	0
TLCC	1000	1	0	1	0	0

Phencyclidine

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	5	2	2	0	0	0
GCMS	10	3	3	0	0	0
GCMS	13	1	1	0	0	0
GCMS	15	1	1	0	0	0
GCMS	20	3	3	0	0	0
GCMS	25	27	27	0	0	0
LCMM	5	2	2	0	0	0
LCMM	10	2	2	0	0	0
LCMM	25	6	6	0	0	0
LCMS	50	1	1	0	0	0
TLCC	100	1	0	1	0	0
TLCC	500	1	0	1	0	0
TOXI	500	1	0	1	0	0

Propoxyphene

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCFL	500	1	0	0	0	0
GCMS	50	1	1	0	0	0
GCMS	100	1	1	0	0	0
GCMS	150	3	3	0	0	0
GCMS	200	2	2	0	0	0
GCMS	300	6	6	0	0	0
GCNP	150	1	1	0	0	0
LCMM	3	1	1	0	0	0
LCMM	10	1	1	0	0	0
LCMM	25	1	1	0	0	0
LCMM	50	1	1	0	0	0
LCMM	100	5	5	0	0	0
LCMM	150	1	1	0	0	0
LCMM	200	1	1	0	0	0
LCMM	300	1	1	0	0	0
TOXI	500	2	2	0	0	0

Secobarbital

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCFL	500	1	0	1	0	0

Participant Performance Evaluation - Confirmation Test

August 2012 Test Event

GCMS	20	1	0	1	0	0
GCMS	50	4	0	4	0	0
GCMS	100	4	0	4	0	0
GCMS	150	1	0	1	0	0
GCMS	200	18	0	18	0	0
GCMS	300	6	0	6	0	0
GCNP	500	1	0	1	0	0
LCDA	200	1	0	0	0	0
LCMM	50	1	0	1	0	0
LCMM	60	1	0	1	0	0
LCMM	100	4	0	4	0	0
LCMM	200	1	0	1	0	0
TLCC	300	1	0	0	0	0
TLCC	1000	1	0	1	0	0

Temazepam

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	50	1	1	0	0	0
GCMS	100	4	4	0	0	0
GCMS	150	1	1	0	0	0
GCMS	200	4	4	0	0	0
GCMS	300	3	3	0	0	0
LCDA	200	1	1	0	0	0
LCDA	400	1	1	0	0	0
LCMM	5	1	1	0	0	0
LCMM	20	2	2	0	0	0
LCMM	25	1	1	0	0	0
LCMM	50	7	7	0	0	0
LCMM	75	2	2	0	0	0
LCMM	100	5	5	0	0	0
LCMM	200	2	2	0	0	0
LCMM	300	4	4	0	0	0
LCMS	100	1	1	0	0	0

THCA

Method Code	Cutoff	No. Labs	True Positive	True Negative	False Positive	False Negative
GCMS	3	2	2	0	0	0
GCMS	5	3	3	0	0	0
GCMS	8	1	1	0	0	0
GCMS	10	6	6	0	0	0
GCMS	15	27	27	0	0	0
LCMM	5	4	4	0	0	0
LCMM	10	3	3	0	0	0
LCMM	15	4	4	0	0	0
LCMS	50	1	1	0	0	0
TLCC	25	1	1	0	0	0
TLCC	50	1	1	0	0	0
TLCC	100	1	1	0	0	0
TOXI	15	1	1	0	0	0
TOXI	25	1	1	0	0	0