

Nirav R. Shah, M.D., M.P.H.
Commissioner

NEW YORK
state department of
HEALTH

Sue Kelly
Executive Deputy Commissioner

Statistical Report
Urinalysis/Urine Pregnancy Testing
Mail out date: April 8, 2013

This report summarizes data from the urinalysis/urine pregnancy proficiency testing of April 8, 2013.

If you have any questions regarding this report, please contact Regina Saulsbery at 518-474-6156. You may also contact us by e-mail: clinchem@wadsworth.org

Results are shown for each analyte arranged by instrument and reagent system. Results for specific gravity and pH are listed as mean \pm 1SD for each of the instruments and reagents. All other results are shown as the number of participant results for that instrument or reagent.

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

PLEASE NOTE: The evaluation report and summary of the educational quantitative urine clinical chemistry proficiency test will be mailed separately if your laboratory participated in that proficiency test.

HEALTH.NY.GOV
[facebook.com/NYSDOH](https://www.facebook.com/NYSDOH)
twitter.com/HealthNYGov

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Specific Gravity

Specimen: U11	Specimen: U12	Specimen: U13	Specimen: U14	Specimen: U15	Number	Instrument or Reagent System
1.011 ± 0.003	1.017 ± 0.005	1.019 ± 0.005	1.016 ± 0.004	1.015 ± 0.003	n = 371	All Methods & Instruments
1.010	1.015	1.019	1.015	1.015	n = 371	Median value, All Methods & Instruments <Instruments>
1.012 ± 0.003	1.015 ± 0.004	1.021 ± 0.004	1.014 ± 0.004	1.010 ± 0.003	n = 27	Visual examination
1.010 ± 0.002	1.016 ± 0.002	1.016 ± 0.002	1.015 ± 0.002	1.018 ± 0.002	n = 96	Siemens Atlas
1.013 ± 0.003	1.023 ± 0.002	1.024 ± 0.002	1.021 ± 0.002	1.017 ± 0.003	n = 36	Siemens Clinitek Status
1.011 ± 0.002	1.022 ± 0.004	1.026 ± 0.003	1.017 ± 0.004	1.012 ± 0.004	n = 7	Siemens Clinitek 50
1.014 ± 0.004	1.023 ± 0.003	1.025 ± 0.000	1.022 ± 0.003	1.019 ± 0.004	n = 5	Siemens Clinitek 100
1.012 ± 0.003	1.021 ± 0.002	1.025 ± 0.002	1.019 ± 0.002	1.014 ± 0.003	n = 36	Siemens Clinitek 500
1.011 ± 0.004	1.020 ± 0.001	1.024 ± 0.002	1.019 ± 0.002	1.013 ± 0.003	n = 50	Siemens Clinitek Advantus
1.017 ± 0.003	1.015 ± 0.000	1.018 ± 0.003	1.013 ± 0.003	1.013 ± 0.003	n = 3	Roche UA Reader/Super UA Reader
1.018 ± 0.002	1.015 ± 0.001	1.017 ± 0.003	1.013 ± 0.002	1.012 ± 0.002	n = 13	Roche Chemstrip Criterion
1.014 ± 0.005	1.015 ± 0.001	1.017 ± 0.002	1.013 ± 0.003	1.015 ± 0.003	n = 34	Roche Urisys
1.008 ± 0.001	1.010 ± 0.001	1.013 ± 0.001	1.011 ± 0.001	1.010 ± 0.001	n = 37	IRIS AUTION MAX
1.009 ± 0.001	1.011 ± 0.001	1.014 ± 0.001	1.013 ± 0.001	1.012 ± 0.001	n = 7	IRIS iChem Velocity
1.009 ± 0.001	1.014 ± 0.001	1.015 ± 0.001	1.013 ± 0.001	1.016 ± 0.002	n = 15	Refractometer
						<Reagents>
1.012 ± 0.003	1.020 ± 0.004	1.024 ± 0.003	1.019 ± 0.004	1.014 ± 0.004	n = 161	Siemens Reagent Strips
1.010 ± 0.002	1.016 ± 0.002	1.016 ± 0.002	1.015 ± 0.002	1.018 ± 0.002	n = 89	Siemens Atlas Reagent
1.017 ± 0.003	1.015 ± 0.002	1.018 ± 0.003	1.013 ± 0.003	1.012 ± 0.003	n = 43	Roche Chemstrips
1.009 ± 0.001	1.015 ± 0.001	1.015 ± 0.001	1.013 ± 0.004	1.017 ± 0.001	n = 14	Roche Urisys Cassette
1.008 ± 0.001	1.010 ± 0.000	1.013 ± 0.001	1.011 ± 0.001	1.010 ± 0.001	n = 22	ARKRAY Aution Sticks
1.008 ± 0.001	1.010 ± 0.001	1.013 ± 0.001	1.011 ± 0.001	1.012 ± 0.001	n = 15	IRIS Reagent Strips
1.009 ± 0.001	1.014 ± 0.002	1.015 ± 0.002	1.013 ± 0.002	1.015 ± 0.003	n = 21	None, no reagent or strip used
1.008 ± 0.004	1.012 ± 0.007	1.016 ± 0.005	1.012 ± 0.005	1.010 ± 0.006	n = 6	Other

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

pH

Specimen: U11	Specimen: U12	Specimen: U13	Specimen: U14	Specimen: U15	Number	Instrument or Reagent System
5.98 ± 0.24	7.03 ± 0.13	5.99 ± 0.21	7.14 ± 0.24	7.36 ± 0.27	n = 377	All Methods & Instruments
6.00	7.00	6.00	7.00	7.50	n = 377	Median value, All Methods & Instruments
						<Instruments>
5.94 ± 0.34	7.11 ± 0.21	5.99 ± 0.33	7.26 ± 0.25	7.47 ± 0.24	n = 35	Visual examination
6.03 ± 0.11	7.01 ± 0.05	6.05 ± 0.15	7.22 ± 0.25	7.50 ± 0.00	n = 96	Siemens Atlas
6.06 ± 0.20	7.00 ± 0.00	6.03 ± 0.16	7.06 ± 0.17	7.39 ± 0.21	n = 40	Siemens Clinitek Status
6.00 ± 0.00	7.00 ± 0.00	5.93 ± 0.19	7.14 ± 0.38	7.43 ± 0.19	n = 7	Siemens Clinitek 50
6.00 ± 0.00	7.33 ± 0.26	5.92 ± 0.49	7.42 ± 0.20	7.50 ± 0.00	n = 6	Siemens Clinitek 100
5.97 ± 0.11	7.00 ± 0.00	6.00 ± 0.00	7.09 ± 0.26	7.36 ± 0.23	n = 38	Siemens Clinitek 500
6.00 ± 0.00	7.05 ± 0.15	6.00 ± 0.00	7.16 ± 0.26	7.42 ± 0.19	n = 53	Siemens Clinitek Advantus
6.17 ± 0.29	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	7.67 ± 0.58	n = 3	Roche UA Reader/Super UA Reader
6.04 ± 0.13	7.00 ± 0.00	6.04 ± 0.13	7.14 ± 0.36	7.21 ± 0.43	n = 14	Roche Chemstrip Criterion
5.77 ± 0.60	7.00 ± 0.00	5.83 ± 0.46	7.00 ± 0.00	7.15 ± 0.36	n = 33	Roche Urisys
5.97 ± 0.11	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	7.04 ± 0.18	n = 37	IRIS AUTION MAX
6.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	7.38 ± 0.52	n = 8	IRIS iChem Velocity
						<Reagents>
6.02 ± 0.13	7.05 ± 0.15	6.02 ± 0.15	7.16 ± 0.26	7.41 ± 0.19	n = 176	Siemens Reagent Strips
6.03 ± 0.12	7.01 ± 0.05	6.05 ± 0.15	7.22 ± 0.25	7.50 ± 0.00	n = 89	Siemens Atlas Reagent
5.86 ± 0.46	7.00 ± 0.00	5.85 ± 0.37	7.05 ± 0.21	7.29 ± 0.44	n = 47	Roche Chemstrips
5.75 ± 0.64	7.00 ± 0.00	5.89 ± 0.45	7.00 ± 0.00	7.07 ± 0.27	n = 14	Roche Urisys Cassette
5.97 ± 0.13	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	7.00 ± 0.00	n = 29	ARKRAY Aution Sticks
6.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	7.28 ± 0.45	n = 16	IRIS Reagent Strips
5.84 ± 0.27	7.12 ± 0.69	5.82 ± 0.29	7.26 ± 0.50	7.36 ± 0.44	n = 5	Other

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Bilirubin	Specimen: U11	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		45	0	0	0	0
Siemens Atlas		95	0	0	0	0
Siemens Clinitek Status		36	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0
Siemens Clinitek 500		37	0	0	0	0
Siemens Clinitek Advantus		53	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		33	0	0	0	0
Roche Cobas U411		2	0	0	0	0
IRIS AUTION MAX		37	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0
Other		1	0	0	0	0
<Reagents>						
Siemens Reagent Strips		170	0	0	0	0
Siemens Atlas Reagent		86	0	0	0	0
Roche Chemstrips		43	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0
IRIS Reagent Strips		16	0	0	0	0
Siemens Ictotest		16	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Hitachi/Roche Reagents		1	0	0	0	0
Other		2	0	0	0	0
All Instruments and Reagents		375	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Bilirubin	Specimen: U12	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
	Visual examination	45	0	0	0	0
	Siemens Atlas	95	0	0	0	0
	Siemens Clinitek Status	36	0	0	0	0
	Siemens Clinitek 50	7	0	0	0	0
	Siemens Clinitek 100	6	0	0	0	0
	Siemens Clinitek 500	37	0	0	0	0
	Siemens Clinitek Advantus	53	0	0	0	0
	Roche UA Reader/Super UA Read	3	0	0	0	0
	Roche Chemstrip Criterion	13	0	0	0	0
	Roche Urisys	33	0	0	0	0
	Roche Cobas U411	2	0	0	0	0
	IRIS AUTION MAX	37	0	0	0	0
	IRIS iChem Velocity	7	0	0	0	0
	Other	1	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	170	0	0	0	0
	Siemens Atlas Reagent	86	0	0	0	0
	Roche Chemstrips	43	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0
	ARKRAY Aution Sticks	27	0	0	0	0
	IRIS Reagent Strips	16	0	0	0	0
	Siemens Ictotest	16	0	0	0	0
	Siemens Hemastix	1	0	0	0	0
	Hitachi/Roche Reagents	1	0	0	0	0
	Other	2	0	0	0	0
	All Instruments and Reagents	375	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Bilirubin	Specimen: U13	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		5	2	18	22	45
Siemens Atlas		0	6	86	0	3
Siemens Clinitek Status		0	1	35	0	1
Siemens Clinitek 50		0	0	7	0	0
Siemens Clinitek 100		0	0	1	5	0
Siemens Clinitek 500		0	0	38	0	0
Siemens Clinitek Advantus		0	0	53	0	0
Roche UA Reader/Super UA Read		0	2	0	0	1
Roche Chemstrip Criterion		7	4	1	0	1
Roche Urisys		1	32	0	0	0
Roche Cobas U411		0	2	0	0	0
IRIS AUTION MAX		1	1	34	0	1
IRIS iChem Velocity		7	0	0	0	0
Other		1	0	0	0	0
<Reagents>						
Siemens Reagent Strips		0	1	147	22	0
Siemens Atlas Reagent		0	7	76	0	3
Roche Chemstrips		7	28	3	4	3
Roche Urisys Cassette		1	12	0	0	0
ARKRAY Aution Sticks		0	1	25	0	1
IRIS Reagent Strips		9	1	8	0	1
Siemens Ictotest		2	0	12	1	43
Siemens Hemastix		0	0	1	0	0
Hitachi/Roche Reagents		1	0	0	0	0
Other		2	0	1	0	1
All Instruments and Reagents		22	50	273	27	52

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Bilirubin	Specimen: U14	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
	Visual examination	46	0	0	0	0
	Siemens Atlas	95	0	0	0	0
	Siemens Clinitek Status	36	0	0	0	0
	Siemens Clinitek 50	7	0	0	0	0
	Siemens Clinitek 100	6	0	0	0	0
	Siemens Clinitek 500	37	0	0	0	0
	Siemens Clinitek Advantus	53	0	0	0	0
	Roche UA Reader/Super UA Read	3	0	0	0	0
	Roche Chemstrip Criterion	13	0	0	0	0
	Roche Urisys	33	0	0	0	0
	Roche Cobas U411	2	0	0	0	0
	IRIS AUTION MAX	37	0	0	0	0
	IRIS iChem Velocity	7	0	0	0	0
	Other	1	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	170	0	0	0	0
	Siemens Atlas Reagent	86	0	0	0	0
	Roche Chemstrips	43	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0
	ARKRAY Aution Sticks	27	0	0	0	0
	IRIS Reagent Strips	16	0	0	0	0
	Siemens Ictotest	17	0	0	0	0
	Siemens Hemastix	1	0	0	0	0
	Hitachi/Roche Reagents	1	0	0	0	0
	Other	2	0	0	0	0
	All Instruments and Reagents	376	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Bilirubin	Specimen: U15	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
	Visual examination	46	0	0	0	0
	Siemens Atlas	95	0	0	0	0
	Siemens Clinitek Status	36	0	0	0	0
	Siemens Clinitek 50	7	0	0	0	0
	Siemens Clinitek 100	6	0	0	0	0
	Siemens Clinitek 500	37	0	0	0	0
	Siemens Clinitek Advantus	52	0	0	1	0
	Roche UA Reader/Super UA Read	3	0	0	0	0
	Roche Chemstrip Criterion	13	0	0	0	0
	Roche Urisys	33	0	0	0	0
	Roche Cobas U411	2	0	0	0	0
	IRIS AUTION MAX	37	0	0	0	0
	IRIS iChem Velocity	7	0	0	0	0
	Other	1	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	169	0	0	1	0
	Siemens Atlas Reagent	86	0	0	0	0
	Roche Chemstrips	43	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0
	ARKRAY Aution Sticks	27	0	0	0	0
	IRIS Reagent Strips	16	0	0	0	0
	Siemens Ictotest	17	0	0	0	0
	Siemens Hemastix	1	0	0	0	0
	Hitachi/Roche Reagents	1	0	0	0	0
	Other	2	0	0	0	0
	All Instruments and Reagents	375	0	0	1	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Glucose	Specimen: U11	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		6	47	69	0	0	0	0
Siemens Atlas		0	96	1	0	0	0	0
Siemens Clinitek Status		1	36	4	0	0	0	0
Siemens Clinitek 50		0	4	3	0	0	0	0
Siemens Clinitek 100		0	6	0	0	0	0	0
Siemens Clinitek 500		0	25	15	0	0	0	0
Siemens Clinitek Advantus		1	37	15	0	0	0	0
Roche UA Reader/Super UA Read		1	3	0	0	0	0	0
Roche Chemstrip Criterion		0	13	1	0	0	0	0
Roche Urisys		0	33	1	0	0	0	0
Roche Cobas U411		0	2	0	0	0	0	0
IRIS AUTION MAX		0	34	2	1	0	0	0
IRIS iChem Velocity		0	0	7	0	0	0	0
Hitachi/Roche Analyzer		0	2	0	0	0	0	0
Other Chemistry Analyzer		1	0	0	0	0	0	0
Other		0	0	1	0	0	0	0
<Reagents>								
Siemens Reagent Strips		0	134	47	0	0	0	0
Siemens Atlas Reagent		0	86	1	0	0	0	0
Roche Chemstrips		1	44	6	0	0	0	0
Roche Urisys Cassette		0	12	1	0	0	0	0
ARKRAY Aution Sticks		0	27	1	1	0	0	0
IRIS Reagent Strips		0	7	8	0	0	0	0
Bayer Clinitest		9	21	53	0	0	0	0
Bayer Diastix		0	2	0	0	0	0	0
Siemens Uristix		0	2	1	0	0	0	0
Hitachi/Roche Reagents		0	2	0	0	0	0	0
Other		0	1	1	0	0	0	0
All Instruments and Reagents		10	338	119	1	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Glucose	Specimen: U12	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		0	5	62	56	0	0	0
Siemens Atlas		0	1	94	2	0	0	0
Siemens Clinitek Status		1	1	32	5	0	2	0
Siemens Clinitek 50		0	0	4	2	0	1	0
Siemens Clinitek 100		0	0	5	1	0	0	0
Siemens Clinitek 500		0	0	35	5	0	0	0
Siemens Clinitek Advantus		0	2	44	7	0	0	0
Roche UA Reader/Super UA Read		0	0	4	0	0	0	0
Roche Chemstrip Criterion		0	0	13	1	0	0	0
Roche Urisys		0	0	32	1	0	1	0
Roche Cobas U411		0	0	2	0	0	0	0
IRIS AUTION MAX		0	0	30	5	0	2	0
IRIS iChem Velocity		0	0	0	7	0	0	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other Chemistry Analyzer		0	0	0	1	0	0	0
Other		0	0	1	0	0	0	0
<Reagents>								
Siemens Reagent Strips		1	3	147	27	0	3	0
Siemens Atlas Reagent		0	1	84	2	0	0	0
Roche Chemstrips		0	0	46	5	0	0	0
Roche Urisys Cassette		0	0	12	0	0	1	0
ARKRAY Aution Sticks		0	0	25	3	0	1	0
IRIS Reagent Strips		0	0	5	9	0	1	0
Bayer Clinitest		0	5	33	46	0	0	0
Bayer Diastix		0	0	2	0	0	0	0
Siemens Uristix		0	0	2	1	0	0	0
Hitachi/Roche Reagents		0	0	2	0	0	0	0
Other		0	0	2	0	0	0	0
All Instruments and Reagents		1	9	360	93	0	6	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Glucose	Specimen: U13	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		89	1	0	0	0	0	0
Siemens Atlas		97	0	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0	0
Siemens Clinitek 500		39	0	0	0	0	0	0
Siemens Clinitek Advantus		53	0	0	0	0	0	0
Roche UA Reader/Super UA Read		4	0	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0	0
Roche Cobas U411		2	0	0	0	0	0	0
IRIS AUTION MAX		37	0	0	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
Other		1	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		180	0	0	0	0	0	0
Siemens Atlas Reagent		87	0	0	0	0	0	0
Roche Chemstrips		49	0	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0	0
IRIS Reagent Strips		15	0	0	0	0	0	0
Bayer Clinitest		53	1	0	0	0	0	0
Siemens Uristix		2	0	0	0	0	0	0
Hitachi/Roche Reagents		2	0	0	0	0	0	0
Other		2	0	0	0	0	0	0
All Instruments and Reagents		432	1	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Glucose	Specimen: U14	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		94	1	0	0	0	0	0
Siemens Atlas		97	0	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0	0
Siemens Clinitek 500		39	0	0	0	0	0	0
Siemens Clinitek Advantus		53	0	0	0	0	0	0
Roche UA Reader/Super UA Read		4	0	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0	0
Roche Cobas U411		2	0	0	0	0	0	0
IRIS AUTION MAX		37	0	0	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
Other		1	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		180	0	0	0	0	0	0
Siemens Atlas Reagent		87	0	0	0	0	0	0
Roche Chemstrips		51	0	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0	0
IRIS Reagent Strips		15	0	0	0	0	0	0
Bayer Clinitest		56	1	0	0	0	0	0
Siemens Uristix		2	0	0	0	0	0	0
Hitachi/Roche Reagents		2	0	0	0	0	0	0
Other		2	0	0	0	0	0	0
All Instruments and Reagents		437	1	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Glucose	Specimen: U15	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		0	0	1	11	17	83	15
Siemens Atlas		0	0	0	81	0	16	0
Siemens Clinitek Status		0	0	0	29	1	9	2
Siemens Clinitek 50		0	0	0	3	0	4	0
Siemens Clinitek 100		0	0	0	0	0	6	0
Siemens Clinitek 500		0	0	0	11	0	29	0
Siemens Clinitek Advantus		1	0	1	6	2	43	0
Roche UA Reader/Super UA Read		0	0	0	0	1	3	0
Roche Chemstrip Criterion		0	0	1	3	0	10	0
Roche Urisys		0	0	1	1	0	32	0
Roche Cobas U411		0	0	0	0	0	2	0
IRIS AUTION MAX		0	1	0	2	0	34	0
IRIS iChem Velocity		0	0	0	7	0	0	0
Hitachi/Roche Analyzer		0	0	0	0	2	0	0
Other Chemistry Analyzer		0	0	0	0	1	0	0
Other		0	0	0	0	0	1	0
<Reagents>								
Siemens Reagent Strips		1	0	1	60	3	109	7
Siemens Atlas Reagent		0	0	0	72	0	15	0
Roche Chemstrips		0	0	1	3	1	45	1
Roche Urisys Cassette		0	0	1	1	0	11	0
ARKRAY Aution Sticks		0	1	0	2	0	26	0
IRIS Reagent Strips		0	0	0	7	0	8	0
Bayer Clinitest		0	0	1	8	18	51	9
Bayer Diastix		0	0	0	0	0	2	0
Siemens Uristix		0	0	0	0	0	3	0
Hitachi/Roche Reagents		0	0	0	0	2	0	0
Other		0	0	0	1	0	2	0
All Instruments and Reagents		1	1	4	154	24	272	17

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Hemoglobin	Specimen: U11	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		38	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		38	0	0	0	0	0
Siemens Clinitek Advantus		52	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		2	0	0	0	0	0
IRIS AUTION MAX		37	0	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		178	0	0	0	0	0
Siemens Atlas Reagent		88	0	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		15	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		3	0	0	0	0	0
All Instruments and Reagents		376	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Hemoglobin	Specimen: U12	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		0	0	0	1	1	39
Siemens Atlas		0	0	0	6	0	90
Siemens Clinitek Status		0	0	0	14	0	26
Siemens Clinitek 50		0	0	0	2	0	5
Siemens Clinitek 100		0	0	0	1	0	5
Siemens Clinitek 500		0	0	0	8	0	31
Siemens Clinitek Advantus		0	0	0	4	0	48
Roche UA Reader/Super UA Read		0	0	0	0	1	2
Roche Chemstrip Criterion		0	0	0	0	1	13
Roche Urisys		0	0	0	3	5	26
Roche Cobas U411		0	0	0	0	0	2
IRIS AUTION MAX		0	0	1	34	0	2
IRIS iChem Velocity		0	0	1	6	0	0
Hitachi/Roche Analyzer		0	0	0	1	0	0
Other		0	0	0	0	0	1
<Reagents>							
Siemens Reagent Strips		0	0	0	29	0	152
Siemens Atlas Reagent		0	0	0	6	0	82
Roche Chemstrips		0	0	0	2	4	42
Roche Urisys Cassette		0	0	0	1	3	10
ARKRAY Aution Sticks		0	0	1	27	0	1
IRIS Reagent Strips		0	0	1	13	0	1
Siemens Hemastix		0	0	0	0	0	1
None, no reagent or strip use		0	0	0	0	1	0
Other		0	0	0	2	0	1
All Instruments and Reagents		0	0	2	80	8	290

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Hemoglobin	Specimen: U13	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
	Visual examination	37	1	0	0	0	0
	Siemens Atlas	96	0	0	0	0	0
	Siemens Clinitek Status	39	1	0	0	0	0
	Siemens Clinitek 50	7	0	0	0	0	0
	Siemens Clinitek 100	6	0	0	0	0	0
	Siemens Clinitek 500	38	0	0	0	0	0
	Siemens Clinitek Advantus	52	0	0	0	0	0
	Roche UA Reader/Super UA Read	3	0	0	0	0	0
	Roche Chemstrip Criterion	14	0	0	0	0	0
	Roche Urisys	34	0	0	0	0	0
	Roche Cobas U411	2	0	0	0	0	0
	IRIS AUTION MAX	37	0	0	0	0	0
	IRIS iChem Velocity	7	0	0	0	0	0
	Hitachi/Roche Analyzer	1	0	0	0	0	0
	Other	1	0	0	0	0	0
<Reagents>							
	Siemens Reagent Strips	176	2	0	0	0	0
	Siemens Atlas Reagent	88	0	0	0	0	0
	Roche Chemstrips	47	0	0	0	0	0
	Roche Urisys Cassette	14	0	0	0	0	0
	ARKRAY Aution Sticks	29	0	0	0	0	0
	IRIS Reagent Strips	15	0	0	0	0	0
	Siemens Hemastix	1	0	0	0	0	0
	None, no reagent or strip use	1	0	0	0	0	0
	Other	3	0	0	0	0	0
	All Instruments and Reagents	374	2	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Hemoglobin	Specimen: U14	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		0	0	0	2	1	38
Siemens Atlas		0	0	0	3	0	93
Siemens Clinitek Status		0	0	0	8	0	32
Siemens Clinitek 50		0	0	0	2	0	5
Siemens Clinitek 100		0	0	0	0	0	6
Siemens Clinitek 500		0	0	0	4	0	35
Siemens Clinitek Advantus		0	0	0	1	0	51
Roche UA Reader/Super UA Read		0	0	0	0	0	3
Roche Chemstrip Criterion		0	0	0	0	0	14
Roche Urisys		0	0	0	1	1	32
Roche Cobas U411		0	0	0	0	0	2
IRIS AUTION MAX		0	0	1	32	0	4
IRIS iChem Velocity		0	0	1	6	0	0
Hitachi/Roche Analyzer		0	0	0	1	0	0
Other		0	0	0	0	0	1
<Reagents>							
Siemens Reagent Strips		0	0	0	16	0	165
Siemens Atlas Reagent		0	0	0	3	0	85
Roche Chemstrips		0	0	0	1	1	46
Roche Urisys Cassette		0	0	0	0	0	14
ARKRAY Aution Sticks		0	0	0	26	0	3
IRIS Reagent Strips		0	0	2	12	0	1
Siemens Hemastix		0	0	0	0	0	1
None, no reagent or strip use		0	0	0	0	1	0
Other		0	0	0	2	0	1
All Instruments and Reagents		0	0	2	60	2	316

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Hemoglobin	Specimen: U15	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		0	0	1	0	0	40
Siemens Atlas		0	0	0	4	0	92
Siemens Clinitek Status		0	0	0	18	0	22
Siemens Clinitek 50		0	0	0	2	0	5
Siemens Clinitek 100		0	0	0	1	0	5
Siemens Clinitek 500		0	0	0	3	0	36
Siemens Clinitek Advantus		0	0	0	1	0	51
Roche UA Reader/Super UA Read		0	0	0	0	0	3
Roche Chemstrip Criterion		0	0	0	0	0	14
Roche Urisys		0	0	0	1	0	33
Roche Cobas U411		0	0	0	0	0	2
IRIS AUTION MAX		0	0	1	34	0	2
IRIS iChem Velocity		0	0	1	6	0	0
Hitachi/Roche Analyzer		0	0	0	1	0	0
Other		0	0	0	0	0	1
<Reagents>							
Siemens Reagent Strips		0	0	1	24	0	156
Siemens Atlas Reagent		0	0	0	4	0	84
Roche Chemstrips		0	0	0	1	0	47
Roche Urisys Cassette		0	0	0	0	0	14
ARKRAY Aution Sticks		0	0	1	27	0	1
IRIS Reagent Strips		0	0	1	13	0	1
Siemens Hemastix		0	0	0	0	0	1
None, no reagent or strip use		0	0	0	0	0	1
Other		0	0	0	2	0	1
All Instruments and Reagents		0	0	3	71	0	306

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Ketones	Specimen: U11	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		47	0	0	0	0	0
Siemens Atlas		97	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		38	0	0	0	0	0
Siemens Clinitek Advantus		52	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		2	0	0	0	0	0
IRIS AUTION MAX		37	0	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	0	0	0	0
Siemens Atlas Reagent		87	0	0	0	0	0
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0	0
IRIS Reagent Strips		16	0	0	0	0	0
Siemens Acetest		9	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		2	0	0	0	0	0
All Instruments and Reagents		385	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Ketones	Specimen: U12	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		47	0	0	0	0	0
Siemens Atlas		97	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		38	0	0	0	0	0
Siemens Clinitek Advantus		52	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		2	0	0	0	0	0
IRIS AUTION MAX		37	0	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	0	0	0	0
Siemens Atlas Reagent		87	0	0	0	0	0
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0	0
IRIS Reagent Strips		16	0	0	0	0	0
Siemens Acetest		9	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		2	0	0	0	0	0
All Instruments and Reagents		385	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Ketones	Specimen: U13	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
	Visual examination	46	0	0	0	1	0
	Siemens Atlas	97	0	0	0	0	0
	Siemens Clinitek Status	40	0	0	0	0	0
	Siemens Clinitek 50	7	0	0	0	0	0
	Siemens Clinitek 100	6	0	0	0	0	0
	Siemens Clinitek 500	38	0	0	0	0	0
	Siemens Clinitek Advantus	52	0	0	0	0	0
	Roche UA Reader/Super UA Read	3	0	0	0	0	0
	Roche Chemstrip Criterion	14	0	0	0	0	0
	Roche Urisys	34	0	0	0	0	0
	Roche Cobas U411	2	0	0	0	0	0
	IRIS AUTION MAX	37	0	0	0	0	0
	IRIS iChem Velocity	7	0	0	0	0	0
	Other	1	0	0	0	0	0
<Reagents>							
	Siemens Reagent Strips	180	0	0	0	1	0
	Siemens Atlas Reagent	87	0	0	0	0	0
	Roche Chemstrips	48	0	0	0	0	0
	Roche Urisys Cassette	14	0	0	0	0	0
	ARKRAY Aution Sticks	27	0	0	0	0	0
	IRIS Reagent Strips	16	0	0	0	0	0
	Siemens Acetest	9	0	0	0	0	0
	Siemens Ketostix	1	0	0	0	0	0
	Other	2	0	0	0	0	0
	All Instruments and Reagents	384	0	0	0	1	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Ketones	Specimen: U14	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		0	0	4	24	28	7
Siemens Atlas		0	0	1	96	0	0
Siemens Clinitek Status		0	0	0	14	26	0
Siemens Clinitek 50		0	0	0	5	2	0
Siemens Clinitek 100		0	0	0	4	2	0
Siemens Clinitek 500		0	0	5	31	3	0
Siemens Clinitek Advantus		0	0	3	47	2	0
Roche UA Reader/Super UA Read		0	0	0	1	2	0
Roche Chemstrip Criterion		0	0	0	10	4	0
Roche Urisys		0	0	0	12	22	0
Roche Cobas U411		0	0	0	1	1	0
IRIS AUTION MAX		0	0	1	36	0	0
IRIS iChem Velocity		0	0	6	1	0	0
Other		0	0	0	1	0	0
<Reagents>							
Siemens Reagent Strips		0	0	9	119	54	0
Siemens Atlas Reagent		0	0	1	86	0	0
Roche Chemstrips		0	0	0	23	26	0
Roche Urisys Cassette		0	0	0	7	7	0
ARKRAY Aution Sticks		0	0	0	27	0	0
IRIS Reagent Strips		0	0	7	9	0	0
Siemens Acetest		0	0	3	8	5	6
Bayer Diastix		0	0	0	1	0	0
Siemens Ketostix		0	0	0	1	0	1
Other		0	0	0	2	0	0
All Instruments and Reagents		0	0	20	283	92	7

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Ketones	Specimen: U15	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		47	0	0	0	0	0
Siemens Atlas		96	0	0	0	1	0
Siemens Clinitek Status		40	0	0	0	0	0
Siemens Clinitek 50		7	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		38	0	0	0	0	0
Siemens Clinitek Advantus		51	0	0	1	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		2	0	0	0	0	0
IRIS AUTION MAX		37	0	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		180	0	0	1	0	0
Siemens Atlas Reagent		86	0	0	0	1	0
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0	0
IRIS Reagent Strips		16	0	0	0	0	0
Siemens Acetest		9	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		2	0	0	0	0	0
All Instruments and Reagents		383	0	0	1	1	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Protein	Specimen: U11	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		4	21	49	1	0	0	15
Siemens Atlas		0	0	97	0	0	0	0
Siemens Clinitek Status		0	1	36	3	0	0	0
Siemens Clinitek 50		0	0	6	1	0	0	0
Siemens Clinitek 100		0	0	5	1	0	0	0
Siemens Clinitek 500		0	30	9	0	0	0	0
Siemens Clinitek Advantus		0	35	18	0	0	0	0
Roche UA Reader/Super UA Read		0	1	1	1	0	0	0
Roche Chemstrip Criterion		1	7	6	0	0	0	0
Roche Urisys		0	4	27	3	0	0	0
Roche Cobas U411		0	0	1	1	0	0	0
IRIS AUTION MAX		0	0	36	1	0	0	0
IRIS iChem Velocity		0	0	7	0	0	0	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other		0	0	1	0	0	0	0
<Reagents>								
Siemens Reagent Strips		1	75	99	5	0	0	0
Siemens Atlas Reagent		0	0	90	0	0	0	0
Sulfosalicylic Acid		3	7	24	1	0	0	15
Roche Chemstrips		1	14	28	4	0	0	0
Roche Urisys Cassette		0	2	11	1	0	0	0
ARKRAY Aution Sticks		0	0	28	1	0	0	0
IRIS Reagent Strips		0	0	15	0	0	0	0
Siemens Uristix		0	1	2	0	0	0	0
Hitachi/Roche Reagents		0	0	1	0	0	0	0
Other		0	0	3	0	0	0	0
All Instruments and Reagents		5	99	301	12	0	0	15

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Protein	Specimen: U12	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	0	2	66	11	17
Siemens Atlas		0	0	0	0	88	9	0
Siemens Clinitek Status		0	0	0	0	39	1	0
Siemens Clinitek 50		0	0	0	0	7	0	0
Siemens Clinitek 100		0	0	0	0	6	0	0
Siemens Clinitek 500		0	0	0	4	35	0	0
Siemens Clinitek Advantus		0	0	0	0	52	1	0
Roche UA Reader/Super UA Read		0	0	0	0	3	0	0
Roche Chemstrip Criterion		0	0	0	0	14	0	0
Roche Urisys		0	0	0	0	34	0	0
Roche Cobas U411		0	0	0	0	2	0	0
IRIS AUTION MAX		0	0	0	0	36	1	0
IRIS iChem Velocity		0	0	0	1	6	0	0
Hitachi/Roche Analyzer		0	0	0	0	2	0	0
Other		0	0	0	0	0	1	0
<Reagents>								
Siemens Reagent Strips		0	0	0	5	167	8	0
Siemens Atlas Reagent		0	0	0	0	81	9	0
Sulfosalicylic Acid		0	0	0	1	33	5	17
Roche Chemstrips		0	0	0	0	47	0	0
Roche Urisys Cassette		0	0	0	0	14	0	0
ARKRAY Aution Sticks		0	0	0	0	29	0	0
IRIS Reagent Strips		0	0	0	1	13	1	0
Siemens Uristix		0	0	0	0	3	0	0
Hitachi/Roche Reagents		0	0	0	0	1	0	0
Other		0	0	0	0	2	1	0
All Instruments and Reagents		0	0	0	7	390	24	17

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Protein	Specimen: U13	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	5	47	23	1	0	16
Siemens Atlas		0	0	74	23	0	0	0
Siemens Clinitek Status		0	0	2	38	0	0	0
Siemens Clinitek 50		0	0	0	7	0	0	0
Siemens Clinitek 100		0	0	2	4	0	0	0
Siemens Clinitek 500		0	0	38	1	0	0	0
Siemens Clinitek Advantus		0	0	52	1	0	0	0
Roche UA Reader/Super UA Read		0	0	1	2	0	0	0
Roche Chemstrip Criterion		0	1	13	0	0	0	0
Roche Urisys		0	0	7	27	0	0	0
Roche Cobas U411		0	0	0	2	0	0	0
IRIS AUTION MAX		0	0	37	0	0	0	0
IRIS iChem Velocity		0	0	7	0	0	0	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other		0	0	1	0	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	116	64	0	0	0
Siemens Atlas Reagent		0	0	69	21	0	0	0
Sulfosalicylic Acid		0	5	23	8	0	0	16
Roche Chemstrips		0	1	23	22	1	0	0
Roche Urisys Cassette		0	0	2	12	0	0	0
ARKRAY Aution Sticks		0	0	29	0	0	0	0
IRIS Reagent Strips		0	0	15	0	0	0	0
Siemens Uristix		0	0	3	0	0	0	0
Hitachi/Roche Reagents		0	0	1	0	0	0	0
Other		0	0	2	1	0	0	0
All Instruments and Reagents		0	6	283	128	1	0	16

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Protein	Specimen: U14	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	1	26	49	1	18
Siemens Atlas		0	0	0	15	82	0	0
Siemens Clinitek Status		0	0	0	0	38	2	0
Siemens Clinitek 50		0	0	0	0	7	0	0
Siemens Clinitek 100		0	0	0	1	5	0	0
Siemens Clinitek 500		0	0	0	38	1	0	0
Siemens Clinitek Advantus		0	0	0	52	1	0	0
Roche UA Reader/Super UA Read		0	0	0	2	1	0	0
Roche Chemstrip Criterion		0	0	0	8	6	0	0
Roche Urisys		0	0	0	24	10	0	0
Roche Cobas U411		0	0	0	1	1	0	0
IRIS AUTION MAX		0	0	0	35	2	0	0
IRIS iChem Velocity		0	0	0	7	0	0	0
Hitachi/Roche Analyzer		0	0	0	2	0	0	0
Other		0	0	0	1	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	0	98	80	2	0
Siemens Atlas Reagent		0	0	0	14	76	0	0
Sulfosalicylic Acid		0	0	1	18	17	1	18
Roche Chemstrips		0	0	0	27	20	0	0
Roche Urisys Cassette		0	0	0	9	5	0	0
ARKRAY Aution Sticks		0	0	0	27	2	0	0
IRIS Reagent Strips		0	0	0	15	0	0	0
Siemens Uristix		0	0	0	1	2	0	0
Hitachi/Roche Reagents		0	0	0	1	0	0	0
Other		0	0	0	2	1	0	0
All Instruments and Reagents		0	0	1	212	203	3	18

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Protein	Specimen: U15	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	1	5	60	11	0	18
Siemens Atlas		0	0	0	96	1	0	0
Siemens Clinitek Status		0	0	0	11	28	1	0
Siemens Clinitek 50		0	0	0	1	6	0	0
Siemens Clinitek 100		0	0	0	3	3	0	0
Siemens Clinitek 500		0	0	0	39	0	0	0
Siemens Clinitek Advantus		0	0	0	53	0	0	0
Roche UA Reader/Super UA Read		0	0	0	3	0	0	0
Roche Chemstrip Criterion		0	0	0	14	0	0	0
Roche Urisys		0	0	0	34	0	0	0
Roche Cobas U411		0	0	0	2	0	0	0
IRIS AUTION MAX		0	0	0	37	0	0	0
IRIS iChem Velocity		0	0	0	7	0	0	0
Hitachi/Roche Analyzer		0	0	0	2	0	0	0
Other		0	0	0	1	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	0	138	41	1	0
Siemens Atlas Reagent		0	0	0	89	1	0	0
Sulfosalicylic Acid		0	1	5	26	5	0	18
Roche Chemstrips		0	0	0	46	1	0	0
Roche Urisys Cassette		0	0	0	14	0	0	0
ARKRAY Aution Sticks		0	0	0	29	0	0	0
IRIS Reagent Strips		0	0	0	15	0	0	0
Siemens Uristix		0	0	0	3	0	0	0
Hitachi/Roche Reagents		0	0	0	1	0	0	0
Other		0	0	0	2	1	0	0
All Instruments and Reagents		0	1	5	363	49	1	18

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Qualitative hCG report:

Beckman Coulter ICON II	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	2	0	2	0	0
Indeterminant	0	0	0	0	0
Positive	0	2	0	2	2

Beckman Coulter ICON 25	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	22	0	22	0	0
Indeterminant	0	0	0	0	0
Positive	0	22	0	22	22

Beckman Coulter ICON 20	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	22	0	22	0	1
Indeterminant	0	0	0	0	0
Positive	0	22	0	22	21

Cardinal Health hCG Combo	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	23	0	23	0	0
Indeterminant	0	0	0	0	0
Positive	0	23	0	23	23

Wampole Clearview hCG Combo	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	2	0	2	0	0
Indeterminant	0	0	0	0	0
Positive	0	2	0	2	2

Cen Med Elite	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	21	0	21	0	0
Indeterminant	0	0	0	0	0
Positive	0	21	0	21	21

Wampole Clearview hCG II	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	5	0	5	0	0
Indeterminant	0	0	0	0	0
Positive	0	5	0	5	5

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Sekisui OSOM Combo (Genzyme)	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	21	0	21	0	0
Indeterminant	0	0	0	0	0
Positive	0	21	0	21	21
Henry Schein One Step	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	2	0	2	0	0
Indeterminant	0	0	0	0	0
Positive	0	2	0	2	2
Inverness Medical BioStar Acceava	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	1	0	1	0	0
Indeterminant	0	0	0	0	0
Positive	0	1	0	1	1
Lifesign Status (serum/urine)	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	1	0	1	0	0
Indeterminant	0	0	0	0	0
Positive	0	1	0	1	1
Meridian Diagnostics	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	1	0	1	0	0
Indeterminant	0	0	0	0	0
Positive	0	1	0	1	1
McKesson hCG Combo	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	1	0	1	0	0
Indeterminant	0	0	0	0	0
Positive	0	1	0	1	1
Polymedco Poly stat	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	6	0	6	0	0
Indeterminant	0	0	0	0	0
Positive	0	6	0	6	6
Quidel QuickVue	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	78	0	78	0	0
Indeterminant	0	0	0	0	0
Positive	0	78	0	78	78

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

R. Medical One Step Strip	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	3	0	3	0	0
Indeterminant	0	0	0	0	0
Positive	0	3	0	3	3
SA Scientific	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	1	0	1	0	0
Indeterminant	0	0	0	0	0
Positive	0	1	0	1	1
Stanbio QuPID	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	16	0	16	0	0
Indeterminant	0	0	0	0	0
Positive	0	16	0	16	16
Fisher Sure-Vue serum/urine	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	37	1	37	0	0
Indeterminant	0	0	0	0	0
Positive	0	36	0	37	37
Fisher Sure-Vue STAT serum/urine	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	25	0	25	0	1
Indeterminant	0	0	0	0	0
Positive	0	25	0	25	24
Signify hCG Urine/Serum	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	4	0	4	0	0
Indeterminant	0	0	0	0	0
Positive	0	4	0	4	4
Sacks Medical RefuAH	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	2	0	2	1	0
Indeterminant	0	0	0	0	0
Positive	0	2	0	1	2

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

Fisher Sure-Vue (urine only)	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	11	0	11	0	0
Indeterminant	0	0	0	0	0
Positive	0	11	0	11	11
True Medix	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	5	0	5	1	0
Indeterminant	0	0	0	0	0
Positive	0	5	0	4	5
Method not listed	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	7	0	7	0	0
Indeterminant	0	0	0	0	0
Positive	0	7	0	7	7
All Methods & Instruments	Specimen:U11	Specimen:U12	Specimen:U13	Specimen:U14	Specimen:U15
Negative	319	1	319	2	2
Indeterminant	0	0	0	0	0
Positive	0	318	0	317	317

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 8 April 2013

hCG Summary:

[Code], hCG Kit name	U11			U12			U13			U14			U15		
	N	I	P	N	I	P	N	I	P	N	I	P	N	I	P
[BC] Beckman Coulter ICON II	2	0	0	0	0	2	2	0	0	0	0	2	0	0	2
[BN] Beckman Coulter ICON 25	22	0	0	0	0	22	22	0	0	0	0	22	0	0	22
[BT] Beckman Coulter ICON 20	22	0	0	0	0	22	22	0	0	0	0	22	1	0	21
[CA] Cardinal Health hCG Combo	23	0	0	0	0	23	23	0	0	0	0	23	0	0	23
[CC] Wampole Clearview hCG Combo	2	0	0	0	0	2	2	0	0	0	0	2	0	0	2
[CM] Cen Med Elite	21	0	0	0	0	21	21	0	0	0	0	21	0	0	21
[CV] Wampole Clearview hCG II	5	0	0	0	0	5	5	0	0	0	0	5	0	0	5
[GO] Sekisui OSOM Combo (Genzyme)	21	0	0	0	0	21	21	0	0	0	0	21	0	0	21
[HS] Henry Schein One Step	2	0	0	0	0	2	2	0	0	0	0	2	0	0	2
[IM] Inverness Medical BioStar Acceava	1	0	0	0	0	1	1	0	0	0	0	1	0	0	1
[LS] Lifesign Status (serum/urine)	1	0	0	0	0	1	1	0	0	0	0	1	0	0	1
[MD] Meridian Diagnostics	1	0	0	0	0	1	1	0	0	0	0	1	0	0	1
[MK] McKesson hCG Combo	1	0	0	0	0	1	1	0	0	0	0	1	0	0	1
[PM] Polymedco Poly stat	6	0	0	0	0	6	6	0	0	0	0	6	0	0	6
[QQ] Quidel QuickVue	78	0	0	0	0	78	78	0	0	0	0	78	0	0	78
[RM] R. Medical One Step Strip	3	0	0	0	0	3	3	0	0	0	0	3	0	0	3
[SA] SA Scientific	1	0	0	0	0	1	1	0	0	0	0	1	0	0	1
[SB] Stanbio QuPID	16	0	0	0	0	16	16	0	0	0	0	16	0	0	16
[SE] Fisher Sure-Vue serum/urine	37	0	0	1	0	36	37	0	0	0	0	37	0	0	37
[SF] Fisher Sure-Vue STAT serum/urine	25	0	0	0	0	25	25	0	0	0	0	25	1	0	24
[SG] Signify hCG Urine/Serum	4	0	0	0	0	4	4	0	0	0	0	4	0	0	4
[SM] Sacks Medical RefuAH	2	0	0	0	0	2	2	0	0	1	0	1	0	0	2
[SU] Fisher Sure-Vue (urine only)	11	0	0	0	0	11	11	0	0	0	0	11	0	0	11
[TX] True Medix	5	0	0	0	0	5	5	0	0	1	0	4	0	0	5
[90] Method not listed	7	0	0	0	0	7	7	0	0	0	0	7	0	0	7
[99] All hCG Methods	319	0	0	1	0	318	319	0	0	2	0	317	2	0	317