

---

**Wadsworth Center**

New York State Department of Health

**BLOOD LEAD**

**Proficiency Test Report**

**Event #1, 2012**

**March 5, 2012**

---

Nirav R. Shah, M.D., M.P.H.  
Commissioner

**NEW YORK**  
state department of  
**HEALTH**

Sue Kelly  
Executive Deputy Commissioner

March 5, 2012

**TOXICOLOGY – Blood Lead**  
**Event #1, 2012**

Dear Laboratory Director:

A statistical summary report for all proficiency test (PT) results evaluated in the first blood lead event of 2012 is enclosed. A confidential three-digit code number assigned by the PT program identifies participating laboratories. Each laboratory will receive an individual performance summary for the last three PT test events under separate cover. To pass the PT for Blood Lead, a laboratory must achieve a minimum score of 80% (4 out of 5 correct) on two consecutive testing events, or two out of three consecutive testing events. Summary reports for Trace Elements in Whole Blood (other than Blood Lead) will be distributed shortly.

**PT Materials**

The blood-based test materials were obtained from lead-dosed goats prior to the test. On Tuesday, January 24th, 2012, 400-500 mL of blood were drawn from each animal into a blood bag containing 750 mg K<sub>2</sub>EDTA. The animals provided pools with lead concentrations ranging from 5 µg/dL to 45 µg/dL. Aliquots of whole blood were transferred into cryovials, and shipped to participating laboratories Wednesday, January 25th, 2012. PT samples for laboratories using the LeadCare® system were shipped by overnight express for delivery Wednesday, January 25th, 2012. Target values were established by a ≥90% consensus of 22 measurements performed by 20 reference laboratories using ICP-MS, ETAAS and ASV methods.

**Certification for CLIA '88 and OSHA Purposes**

Laboratories outside of New York State can have their PT results from this program evaluated for federal regulatory purposes under CLIA '88. The laboratory director should notify the regional CMS office, and should provide our program with the address to which PT results should be sent. Participation in this program may also be used to obtain approval for blood lead testing from the Occupational Safety and Health Administration (OSHA), U.S. Dept. of Labor. For further information on OSHA approval, contact: James S. Pike at 801-233-4927.


**Scheduled PT Events**

Samples are scheduled to be mailed Wednesday, May 9th, 2012. The laboratory is required to notify the section (trel@wadsworth.org) within five days of shipment that samples have not arrived or are unacceptable for testing. For laboratories using the LeadCare® system samples are scheduled to be mailed Tuesday, May 8th, 2012, for delivery Wednesday, May 9th, 2012. These laboratories are required to notify the section (trel@wadsworth.org) by 12:00 pm (EST) Wednesday, May 9th, 2012, that samples have not arrived or are unacceptable for testing. Failure to notify the section will result in a score of zero. The deadline for reporting results is Wednesday, May 30th, 2012.

Sincerely,



Patrick J. Parsons, Ph.D.  
Chief, Laboratory of Inorganic and Nuclear Chemistry  
Deputy Director, Division of Environmental Health Sciences



Mary Frances Verostek, Ph.D.  
Assistant Section Head  
PT Program for Blood Lead/Trace Elements.

**New York State Department of Health  
Blood Lead Test Results, 2012 Event #1  
PERFORMANCE OF PARTICIPATING LABORATORIES**

| Lab Code       | Method        | Results (µg/dL whole blood) |         |         |         |         | Normalized Mean | Info Only |
|----------------|---------------|-----------------------------|---------|---------|---------|---------|-----------------|-----------|
|                |               | BE12-01                     | BE12-02 | BE12-03 | BE12-04 | BE12-05 |                 |           |
| Target values: |               | 5                           | 22      | 10      | 20      | 45      |                 |           |
| 103            | DRC/CC-ICP-MS | 5                           | 23      | 9       | 20      | 47      | 1.03            |           |
| 104            | ETAAS-Z       | 8                           | 22      | 9       | 19      | 41      | 0.95            |           |
| 107            | DRC/CC-ICP-MS | 5                           | 23      | 9       | 20      | 48      | 1.04            |           |
| 107            | ASV-LeadCare  | 5                           | 24      | 9       | 20      | 46      | 1.04            | Info      |
| 109            | ETAAS-Z       | 4                           | 22      | 9       | 20      | 45      | 1.00            |           |
| 109            | ICP-MS        | 5                           | 22      | 9       | 19      | 46      | 0.99            |           |
| 109            | ASV-LeadCare  | 5                           | 23      | 8       | 19      | 41      | 0.97            | Info      |
| 109            | ASV-LeadCare  | 5                           | 23      | 9       | 21      | 43      | 1.02            | Info      |
| 110            | ETAAS-Z       | 6                           | 24      | 10      | 21      | 47      | 1.06            |           |
| 110            | ASV-LeadCare  | 5                           | 24      | 9       | 21      | 46      | 1.05            | Info      |
| 110            | ASV-LeadCare  | 5                           | 27 ↑    | 12      | 19      | 50      | 1.12            | Info      |
| 110            | ICP-MS        | 5                           | 22      | 9       | 20      | 46      | 1.01            |           |
| 112            | ETAAS-Z       | 3                           | 18      | 8       | 15 ↓    | 33 ↓    | 0.77            |           |
| 114            | ETAAS-Z       | 5                           | 21      | 9       | 19      | 44      | 0.96            |           |
| 116            | ICP-MS        | 5                           | 22      | 9       | 19      | 46      | 0.99            | Info      |
| 121            | ETAAS-Z       | 4                           | 17 ↓    | 7       | 15 ↓    | 32 ↓    | 0.74            | Info      |
| 123            | ETAAS-Z       | 6                           | 21      | 9       | 17      | 42      | 0.91            |           |
| 126            | ETAAS-Z       | 6                           | 25      | 10      | 23      | 44      | 1.09            |           |
| 131            | ETAAS-Z       | 6                           | 23      | 9       | 17      | 43      | 0.95            |           |
| 143            | ETAAS-Z       | 4                           | 19      | 8       | 17      | 40 ↓    | 0.87            |           |
| 144            | ETAAS-Z       | 4                           | 23      | 9       | 20      | 46      | 1.02            |           |
| 146            | ETAAS-Z       | 5                           | 22      | 8       | 18      | 42      | 0.94            |           |
| 147            | ICP-MS        | 5                           | 21      | 9       | 19      | 44      | 0.96            |           |
| 150            | ASV-LeadCare  | 4                           | >8      | 7       | >8      | >8      | ??              |           |
| 156            | ICP-MS        | 4                           | 20      | 8       | 18      | 43      | 0.92            |           |
| 158            | ICP-MS        | 6                           | 23      | 10      | 20      | 46      | 1.02            |           |
| 159            | ICP-MS        | 5                           | 22      | 9       | 20      | 44      | 0.99            |           |
| 160            | ETAAS-Z       | 5                           | 22      | 10      | 22      | 43      | 1.02            |           |
| 164            | ICP-MS        | 5                           | 22      | 9       | 19      | 45      | 0.98            |           |
| 166            | ASV-3010      | 6                           | 22      | 10      | 20      | 45      | 1.00            |           |
| 168            | ETAAS-Z       | 6                           | 23      | 10      | 21      | 46      | 1.04            |           |

**notes:** ↑ reported value outside upper limit  
↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.  
Info only: results included for informational purposes only.

**New York State Department of Health  
Blood Lead Test Results, 2012 Event #1  
PERFORMANCE OF PARTICIPATING LABORATORIES**

| Lab Code       | Method      | Results (µg/dL whole blood) |         |         |         |         | Normalized Mean | Info Only |
|----------------|-------------|-----------------------------|---------|---------|---------|---------|-----------------|-----------|
|                |             | BE12-01                     | BE12-02 | BE12-03 | BE12-04 | BE12-05 |                 |           |
| Target values: |             | 5                           | 22      | 10      | 20      | 45      |                 |           |
| 179            | ICP-MS      | 5                           | 22      | 9       | 19      | 45      | 0.98            |           |
| 197            | ICP-MS      | 5                           | 22      | 8       | 19      | 44      | 0.98            |           |
| 198            | ETAAS-Z     | 6                           | 23      | 10      | 21      | 42      | 1.01            |           |
| 199            | ETAAS-Z     | 5                           | 24      | 11      | 21      | 48      | 1.08            |           |
| 200            | ETAAS-Z     | 5                           | 23      | 9       | 20      | 43      | 1.00            |           |
| 204            | ASV-3010    | 4                           | 20      | 6       | 16      | 48      | 0.93            |           |
| 206            | ICP-MS      | 5                           | 22      | 10      | 19      | 46      | 0.99            |           |
| 208            | ETAAS-Z     | 7                           | 24      | 9       | 20      | 44      | 1.02            |           |
| 221            | ETAAS-Z     | 5                           | 23      | 10      | 20      | 44      | 1.01            |           |
| 232            | ASV-3010    | 6                           | 23      | 11      | 19      | 46      | 1.03            |           |
| 237            | ETAAS-Z     | 5                           | 22      | 9       | 19      | 46      | 0.99            |           |
| 243            | ASV-3010    | 5                           | 22      | 9       | 20      | 47      | 1.01            |           |
| 254            | ETAAS-Z     | 5                           | 24      | 10      | 21      | 47      | 1.06            |           |
| 255            | ETAAS-Z     | 5                           | 22      | 9       | 19      | 45      | 0.98            |           |
| 261            | ETAAS-Z     | 5                           | 22      | 9       | 19      | 45      | 0.98            |           |
| 269            | ETAAS-Z     | 4                           | 19      | 7       | 17      | 39 ↓    | 0.86            |           |
| 271            | ASV-3010    | 4                           | 20      | 7       | 18      | 39 ↓    | 0.89            |           |
| 272            | ETAAS-Z     | 5                           | 22      | 10      | 20      | 45      | 1.00            |           |
| 279            | ETAAS-Z     | 4                           | 20      | 8       | 18      | 43      | 0.92            |           |
| 282            | ASV-3010    | 6                           | 24      | 11      | 21      | 39 ↓    | 1.03            |           |
| 290            | ICP-MS      | 5                           | 21      | 9       | 17      | 41      | 0.91            |           |
| 291            | ASV-3010    | 6                           | 24      | 10      | 20      | 52 ↑    | 1.08            |           |
| 293            | ICP-MS      | 5                           | 22      | 12      | 18      | 44      | 1.02            |           |
| 295            | ASV-3010    | 8                           | 24      | 11      | 20      | 48      | 1.06            |           |
| 301            | ETAAS Other | 4                           | 20      | 8       | 15 ↓    | 44      | 0.88            |           |
| 305            | ETAAS-Z     | 5                           | 20      | 9       | 18      | 42      | 0.91            |           |
| 312            | ICP-MS      | 5                           | 24      | 9       | 20      | 48      | 1.05            |           |
| 317            | ETAAS-Z     | 6                           | 21      | 9       | 19      | 43      | 0.95            |           |
| 324            | HR-ICP-MS   | 5                           | 22      | 9       | 19      | 42      | 0.96            |           |
| 325            | ETAAS-Z     | 5                           | 21      | 9       | 19      | 50      | 1.01            |           |
| 333            | ETAAS-Z     | 5                           | 22      | 9       | 20      | 44      | 0.99            |           |

**notes:** ↑ reported value outside upper limit  
↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.  
Info only: results included for informational purposes only.

**New York State Department of Health  
Blood Lead Test Results, 2012 Event #1  
PERFORMANCE OF PARTICIPATING LABORATORIES**

| Lab Code       | Method        | Results (µg/dL whole blood) |         |         |         |         | Normalized Mean | Info Only |
|----------------|---------------|-----------------------------|---------|---------|---------|---------|-----------------|-----------|
|                |               | BE12-01                     | BE12-02 | BE12-03 | BE12-04 | BE12-05 |                 |           |
| Target values: |               | 5                           | 22      | 10      | 20      | 45      |                 |           |
| 337            | ASV-LeadCare  | 5                           | 22      | 9       | 21      | 47      | 1.03            | Info      |
| 339            | HR-ICP-MS     | 5                           | 22      | 9       | 18      | 44      | 0.96            |           |
| 340            | ETAAS-Z       | 4                           | 22      | 8       | 19      | 45      | 0.98            |           |
| 343            | ASV-LeadCare  | 5                           | 23      | 8       | 20      | 42      | 0.99            | Info      |
| 348            | ETAAS-Z       | 5                           | 24      | 10      | 22      | 47      | 1.08            |           |
| 349            | ETAAS-Z       | 5                           | 23      | 9       | 19      | 45      | 1.00            |           |
| 350            | ASV-3010      | 9                           | 26      | 13      | 22      | 48      | 1.16            |           |
| 352            | ASV-3010      | 4                           | 23      | 10      | 20      | 45      | 1.02            |           |
| 353            | ETAAS-Z       | 4                           | 25      | 9       | 20      | 45      | 1.05            |           |
| 365            | ETAAS-Z       | 5                           | 22      | 9       | 20      | 43      | 0.99            |           |
| 366            | ETAAS-Z       | 6                           | 22      | 8       | 23      | 43      | 1.04            | Info      |
| 367            | DRC/CC-ICP-MS | 5                           | 23      | 9       | 20      | 47      | 1.03            | Info      |
| 368            | ASV-3010      | 4                           | 20      | 8       | 19      | 42      | 0.93            |           |
| 369            | ASV-3010      | 4                           | 22      | 9       | 18      | 45      | 0.97            |           |
| 374            | ASV-3010      | 4                           | 20      | 7       | 17      | 44      | 0.91            |           |
| 383            | ETAAS-Z       | 5                           | 22      | 9       | 19      | 44      | 0.98            |           |
| 384            | ASV-3010      | 4                           | 20      | 9       | 19      | 39 ↓    | 0.91            |           |
| 388            | ASV-3010      | 3                           | 21      | 9       | 17      | 44      | 0.93            |           |
| 389            | ETAAS-Z       | 5                           | 22      | 9       | 20      | 45      | 1.00            |           |
| 391            | ETAAS-Z       | 6                           | 25      | 11      | 21      | 50      | 1.10            | Info      |
| 393            | ASV-LeadCare  | 4                           | 24      | 7       | 21      | >37     | 1.07            |           |
| 401            | ETAAS-Z       | 6                           | 25      | 10      | 22      | 50      | 1.12            | Info      |
| 410            | ICP-MS        | 5                           | 23      | 9       | 20      | 47      | 1.03            | Info      |
| 461            | ASV-3010      | 4                           | 21      | 8       | 18      | 44      | 0.94            |           |
| 463            | ASV-LeadCare  | 4                           | 25      | 9       | 21      | 53 ↑    | 1.12            |           |
| 464            | ASV-LeadCare  | 5                           | 26      | 9       | 23      | 48      | 1.13            |           |
| 469            | ICP-MS        | 6                           | 42 ↑    | 16 ↑    | 40 ↑    | 105 ↑   | 1.96            |           |
| 470            | ASV-LeadCare  | 4                           | 24      | 8       | 21      | 47      | 1.06            |           |

Percent satisfactory results for all participants: 96.0 %

**notes:** ↑ reported value outside upper limit  
↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.  
Info only: results included for informational purposes only.

**New York State Department of Health  
Blood Lead Test Results, 2012 Event #1  
STATISTICAL SUMMARY**

| TARGET VALUE ASSIGNMENT AND STATISTICS |               |                             |         |         |         |         |
|----------------------------------------|---------------|-----------------------------|---------|---------|---------|---------|
| Lab Code                               | Method        | Results (µg/dL whole blood) |         |         |         |         |
|                                        |               | BE12-01                     | BE12-02 | BE12-03 | BE12-04 | BE12-05 |
| 103                                    | DRC/CC-ICP-MS | 5                           | 23      | 9       | 20      | 47      |
| 104                                    | ETAAS-Z       | 8                           | 22      | 9       | 19      | 41      |
| 107                                    | DRC/CC-ICP-MS | 5                           | 23      | 9       | 20      | 48      |
| 109                                    | ETAAS-Z       | 4                           | 22      | 9       | 20      | 45      |
| 109                                    | ICP-MS        | 5                           | 22      | 9       | 19      | 46      |
| 110                                    | ETAAS-Z       | 6                           | 24      | 10      | 21      | 47      |
| 110                                    | ICP-MS        | 5                           | 22      | 9       | 20      | 46      |
| 147                                    | ICP-MS        | 5                           | 21      | 9       | 19      | 44      |
| 156                                    | ICP-MS        | 4                           | 20      | 8       | 18      | 43      |
| 159                                    | ICP-MS        | 5                           | 22      | 9       | 20      | 44      |
| 160                                    | ETAAS-Z       | 5                           | 22      | 10      | 22      | 43      |
| 164                                    | ICP-MS        | 5                           | 22      | 9       | 19      | 45      |
| 166                                    | ASV-3010      | 6                           | 22      | 10      | 20      | 45      |
| 179                                    | ICP-MS        | 5                           | 22      | 9       | 19      | 45      |
| 198                                    | ETAAS-Z       | 6                           | 23      | 10      | 21      | 42      |
| 199                                    | ETAAS-Z       | 5                           | 24      | 11      | 21      | 48      |
| 200                                    | ETAAS-Z       | 5                           | 23      | 9       | 20      | 43      |
| 243                                    | ASV-3010      | 5                           | 22      | 9       | 20      | 47      |
| 293                                    | ICP-MS        | 5                           | 22      | 12      | 18      | 44      |
| 324                                    | HR-ICP-MS     | 5                           | 22      | 9       | 19      | 42      |
| 325                                    | ETAAS-Z       | 5                           | 21      | 9       | 19      | 50      |
| 350                                    | ASV-3010      | 9                           | 26      | 13      | 22      | 48      |
| <hr/>                                  |               |                             |         |         |         |         |
| Number of Sample Measurements:         |               | 22                          | 22      | 22      | 22      | 22      |
| Mean (target value):                   |               | 5                           | 22      | 10      | 20      | 45      |
| Standard Deviation:                    |               | 1.1                         | 1.2     | 1.1     | 1.1     | 2.3     |
| RSD (%):                               |               | 21.2                        | 5.4     | 12.0    | 5.5     | 5.2     |
| <hr/>                                  |               |                             |         |         |         |         |
| Acceptable Range:                      |               |                             |         |         |         |         |
| Upper Limit:                           |               | 9                           | 26      | 14      | 24      | 50      |
| Lower Limit:                           |               | 1                           | 18      | 6       | 16      | 41      |

**notes:** Results reported as less than the detection limits are treated as zero for statistical and grading purposes.

**New York State Department of Health  
Blood Lead Test Results, 2012 Event #1  
STATISTICAL SUMMARY BY METHOD**

|                                | Results (µg/dL whole blood) |         |         |         |         |
|--------------------------------|-----------------------------|---------|---------|---------|---------|
|                                | BE12-01                     | BE12-02 | BE12-03 | BE12-04 | BE12-05 |
| <b>ASV-3010</b>                |                             |         |         |         |         |
| Number of Sample Measurements: | 15                          | 16      | 16      | 16      | 16      |
| Mean:                          | 4.8                         | 22.0    | 9.3     | 19.0    | 44.7    |
| Standard Deviation:            | 1.3                         | 1.9     | 1.8     | 1.6     | 3.6     |
| RSD (%):                       | 27.5                        | 8.5     | 19.5    | 8.4     | 8.2     |
| <b>ASV-LeadCare</b>            |                             |         |         |         |         |
| Number of Sample Measurements: | 12                          | 11      | 12      | 11      | 10      |
| Mean:                          | 4.7                         | 24.1    | 8.7     | 20.6    | 46.3    |
| Standard Deviation:            | 0.5                         | 1.4     | 1.3     | 1.1     | 3.7     |
| RSD (%):                       | 10.6                        | 6.0     | 15.0    | 5.4     | 7.9     |
| <b>DRC/CC-ICP-MS</b>           |                             |         |         |         |         |
| Number of Sample Measurements: | 3                           | 3       | 3       | 3       | 3       |
| Mean:                          | 5.0                         | 23.0    | 9.0     | 20.0    | 47.3    |
| Standard Deviation:            | 0.0                         | 0.0     | 0.0     | 0.0     | 0.6     |
| RSD (%):                       | —                           | —       | —       | —       | —       |
| <b>ETAAS Other</b>             |                             |         |         |         |         |
| Number of Sample Measurements: | 1                           | 1       | 1       | 1       | 1       |
| Mean:                          | 4.0                         | 20.0    | 8.0     | 15.0    | 44.0    |
| Standard Deviation:            | ?                           | ?       | ?       | ?       | ?       |
| RSD (%):                       | —                           | —       | —       | —       | —       |
| <b>ETAAS-Z</b>                 |                             |         |         |         |         |
| Number of Sample Measurements: | 40                          | 40      | 40      | 40      | 40      |
| Mean:                          | 5.1                         | 22.2    | 9.1     | 19.5    | 43.9    |
| Standard Deviation:            | 0.9                         | 1.8     | 0.9     | 1.9     | 3.6     |
| RSD (%):                       | 18.3                        | 8.3     | 9.9     | 9.5     | 8.2     |
| <b>HR-ICP-MS</b>               |                             |         |         |         |         |
| Number of Sample Measurements: | 2                           | 2       | 2       | 2       | 2       |
| Mean:                          | 5.0                         | 22.0    | 9.0     | 18.5    | 43.0    |
| Standard Deviation:            | 0.0                         | 0.0     | 0.0     | 0.7     | 1.4     |
| RSD (%):                       | —                           | —       | —       | —       | —       |
| <b>ICP-MS</b>                  |                             |         |         |         |         |
| Number of Sample Measurements: | 16                          | 15      | 16      | 15      | 15      |
| Mean:                          | 5.1                         | 22.0    | 9.6     | 19.1    | 45.0    |
| Standard Deviation:            | 0.4                         | 0.9     | 1.9     | 0.9     | 1.7     |
| RSD (%):                       | 8.7                         | 4.2     | 20.0    | 4.6     | 3.8     |
| <b>All Laboratories</b>        |                             |         |         |         |         |
| Number of Sample Measurements: | 89                          | 88      | 90      | 88      | 87      |
| Mean:                          | 5.0                         | 22.3    | 9.1     | 19.4    | 44.6    |
| Standard Deviation:            | 0.9                         | 1.7     | 1.4     | 1.7     | 3.3     |
| RSD (%):                       | 17.7                        | 7.8     | 14.8    | 8.5     | 7.5     |

**notes:** ? Insufficient data for calculation.

**New York State Department of Health**  
**Blood Lead Test Results, 2012 Event #1**  
**STATISTICAL SUMMARY BY CLASS**

|                                | Results ( $\mu\text{g/dL}$ whole blood) |         |         |         |         |
|--------------------------------|-----------------------------------------|---------|---------|---------|---------|
|                                | BE12-01                                 | BE12-02 | BE12-03 | BE12-04 | BE12-05 |
| <b>Evaluated</b>               |                                         |         |         |         |         |
| Number of Sample Measurements: | 54                                      | 52      | 54      | 52      | 51      |
| Mean:                          | 4.9                                     | 22.1    | 9.0     | 19.1    | 44.3    |
| Standard Deviation:            | 1.0                                     | 1.8     | 1.4     | 1.7     | 3.3     |
| RSD (%):                       | 19.8                                    | 8.0     | 16.0    | 9.1     | 7.4     |
| <b>Info</b>                    |                                         |         |         |         |         |
| Number of Sample Measurements: | 14                                      | 14      | 14      | 14      | 14      |
| Mean:                          | 5.1                                     | 23.1    | 9.1     | 19.9    | 44.8    |
| Standard Deviation:            | 0.5                                     | 2.2     | 1.3     | 1.9     | 4.7     |
| RSD (%):                       | 10.4                                    | 9.7     | 14.0    | 9.6     | 10.5    |
| <b>Reference</b>               |                                         |         |         |         |         |
| Number of Sample Measurements: | 21                                      | 22      | 22      | 22      | 22      |
| Mean:                          | 5.2                                     | 22.4    | 9.5     | 19.8    | 45.1    |
| Standard Deviation:            | 0.8                                     | 1.2     | 1.1     | 1.1     | 2.3     |
| RSD (%):                       | 15.7                                    | 5.4     | 12.0    | 5.5     | 5.2     |
| <b>All Laboratories</b>        |                                         |         |         |         |         |
| Number of Sample Measurements: | 89                                      | 88      | 90      | 88      | 87      |
| Mean:                          | 5.0                                     | 22.3    | 9.1     | 19.4    | 44.6    |
| Standard Deviation:            | 0.9                                     | 1.7     | 1.4     | 1.7     | 3.3     |
| RSD (%):                       | 17.7                                    | 7.8     | 14.8    | 8.5     | 7.5     |

**notes:** ? Insufficient data for calculation.



**New York State Department of Health  
Blood Lead Test Results, 2012 Event #1**

**METHOD NOTES**

---

**ATOMIC SPECTROMETRY METHODS**

- A-1     ETAAS-Z (Electrothermal atomic absorption spectrometry with Zeeman background correction)
- A-2     ETAAS other (i.e., D<sub>2</sub>, S-H background correction)
- A-3     FAAS (Flame atomic absorption spectrometry)
- A-4     CV-AAS (Cold vapor atomic absorption spectrometry)
- A-5     HG-AAS (Hydride generation atomic absorption spectrometry)
- A-6     AFS (Atomic fluorescence spectrometry)
- A-7     Other

**INDUCTIVELY COUPLED PLASMA**

- P-1     ICP-MS (Inductively coupled plasma - mass spectrometry)
- P-2     DRC/CC-ICP-MS (ICP-MS used in the Dynamic Reaction Cell or Collision Cell mode)
- P-3     ICP-AES/OES (ICP atomic/optical emission spectrometry)
- P-4     HR-ICP-MS (High resolution ICP-MS)
- P-5     ETV-ICP-MS (Electrothermal vaporization ICP-MS)
- P-6     ID-ICP-MS (Isotope dilution ICP-MS)
- P-7     Other

**ELECTROCHEMICAL METHODS**

- E-1     ASV (Anodic stripping voltammetry without digestion)
- E-2     ASV-LeadCare® (Anodic stripping voltammetry using the ESA LeadCare® system)
- E-3     Fluoride specific electrode
- E-4     Other

**MOLECULAR FLUORIMETRY**

- F-1     EtOAc (Ethyl acetate-acetic acid extraction method for determination of erythrocyte protoporphyrin)
- F-2     Aviv hematofluorometry (for determination of EP at hematocrit 35)
- F-3     Helena ZPP (for determination of zinc protoporphyrin in  $\mu\text{mol ZPP/mol heme}$ )
- F-4     Other

**OTHER METHODS**

If your method is not listed in the above list, please describe it briefly.

---