
Wadsworth Center

New York State Department of Health

BLOOD LEAD

Proficiency Test Report

Event #1, 2010

March 2, 2010

STATE OF NEW YORK DEPARTMENT OF HEALTH

Wadsworth Center The Governor Nelson A. Rockefeller Empire State Plaza P.O. Box 509 Albany, New York 12201-0509

Richard F. Daines, M.D.
Commissioner

James W. Clyne, Jr.,
Executive Deputy Commissioner

March 2, 2010

TOXICOLOGY – Blood Lead Event #1, 2010

Dear Laboratory Director:

A statistical summary report for all proficiency test (PT) results evaluated in the first blood lead event of 2010 is enclosed. A confidential three-digit code number assigned by the PT program identifies participating laboratories. Each laboratory will receive an individual performance summary for the last three PT test events under separate cover. To pass the PT for Blood Lead, a laboratory must achieve a minimum score of 80% (4 out of 5 correct) on two consecutive testing events, or two out of three consecutive testing events. Summary reports for Trace Elements in Whole Blood (other than Blood Lead) will be distributed shortly.

PT Materials

The blood-based test materials were obtained from lead-dosed goats prior to the test. On Tuesday, January 12th, 2010, 400-500 mL of blood were drawn from each animal into a blood bag containing 750 mg K₂EDTA. The animals provided pools with lead concentrations ranging from 3 µg/dL to 43 µg/dL. Aliquots of whole blood were transferred into cryovials, and shipped to participating laboratories Wednesday, January 13th, 2010. PT samples for laboratories using the LeadCare[®] system were shipped by overnight express for delivery Wednesday, January 13th, 2010. Target values were established by a ≥90% consensus of 20 measurements performed by 19 reference laboratories using ICP-MS, ETAAS and ASV methods.


Certification for CLIA '88 and OSHA Purposes

Laboratories outside of New York State can have their PT results from this program evaluated for federal regulatory purposes under CLIA '88. The laboratory director should notify the regional CMS office, and should provide our program with the address to which PT results should be sent. Participation in this program may also be used to obtain approval for blood lead testing from the Occupational Safety and Health Administration (OSHA), U.S. Dept. of Labor. For further information on OSHA approval, contact: James S. Pike at 801-233-4927.

The next PT event for blood lead is scheduled to be mailed Wednesday, April 28th, 2010. PT samples for laboratories using the LeadCare[®] system are scheduled to be shipped by overnight express for delivery Wednesday, April 28th, 2010. Please inform our laboratory staff at (518) 474-4484 if the test materials have not arrived within five days of the scheduled mail out date. The deadline for reporting results is **Wednesday, May 19th, 2010.**

Thank you for your participation.

Sincerely,



Patrick J. Parsons, Ph.D.
Section Head, Blood Lead Proficiency Testing Program

**New York State Department of Health
Blood Lead Test Results, 2010 Event #1
PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ($\mu\text{g/dL}$ whole blood)					Normalized Mean	Info Only
		BE10-01	BE10-02	BE10-03	BE10-04	BE10-05		
Target Values:		17	21	43	38	3		
103	ASV-LeadCare	18	21	46	40	3	1.05	Info
103	DRC/CC-ICP-MS	17	20	41	36	3	0.96	
104	ETAAS-Z	18	22	44	39	3	1.04	
106	ICP-MS	17	21	43	38	3	1.00	Info
107	ICP-MS	18	22	44	39	3	1.04	
107	ASV-LeadCare	14	17	37 ↓	30 ↓	<3	0.82	Info
108	ETAAS-Z	16	21	39	36	5	0.95	
109	ETAAS-Z	19	23	46	41	4	1.09	
109	ASV-LeadCare	18	20	42	36	3	0.98	Info
109	ASV-LeadCare	16	19	39	34	<3	0.91	Info
110	ETAAS-Z	18	22	44	38	4	1.03	
110	ICP-MS	18	21	43	38	3	1.01	
110	ASV-LeadCare	18	21	44	38	3	1.02	Info
110	ASV-LeadCare	18	20	46	36	<3	1.01	Info
112	ASV-3010	15	19	39	34	<2	0.90	
114	ETAAS-Z	19	24	47	40	6	1.10	
115	ETAAS-Z	17	20	43	30 ↓	2	0.94	
116	ICP-MS	18	22	45	39	3	1.04	Info
121	ETAAS-Z	20	24	52 ↑	60 ↑	3	1.28	Info
123	ETAAS-Z	16	20	40	35	3	0.94	
126	ETAAS-Z	17	22	43	39	3	1.02	
131	ETAAS-Z	19	22	43	32 ↓	7	1.00	
132	ETAAS-Z	17	20	41	36	3	0.96	
143	ETAAS-Z	16	20	41	36	3	0.95	
144	ETAAS-Z	17	21	43	37	3	0.99	
146	ETAAS-Z	16	20	41	35	2	0.94	
147	ICP-MS	17	20	41	36	3	0.96	
150	ASV-LeadCare	17	21	43	38	3	1.00	
156	ICP-MS	18	25	56 ↑	44 ↑	3	1.18	
158	ETAAS-Z	18	21	44	38	4	1.02	

notes: ↑ reported value outside upper limit
↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.
Info only: results included for informational purposes only.

**New York State Department of Health
Blood Lead Test Results, 2010 Event #1
PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ($\mu\text{g/dL}$ whole blood)					Normalized Mean	Info Only
		BE10-01	BE10-02	BE10-03	BE10-04	BE10-05		
Target Values:		17	21	43	38	3		
159	ICP-MS	17	20	42	36	3	0.97	
160	ETAAS-Z	19	25	56 \uparrow	51 \uparrow	4	1.24	
164	ICP-MS	18	22	45	39	3	1.04	
166	ASV-3010	17	20	41	36	2	0.96	
168	ETAAS-Z	18	23	45	39	4	1.06	
170	ETAAS Other	17	21	43	36	3	0.99	
179	ICP-MS	17	22	44	37	3	1.01	
197	ICP-MS	15	19	38 \downarrow	33 \downarrow	3	0.88	
198	ETAAS-Z	17	22	43	38	3	1.01	
199	ETAAS-Z	18	22	43	39	3	1.03	
200	ETAAS-Z	17	21	45	38	2	1.01	
204	ASV-3010	17	18	39	37	2	0.93	
206	ICP-MS	17	21	43	36	3	0.99	
208	ETAAS-Z	21	25	52 \uparrow	43 \uparrow	<3	1.19	
215	ETAAS-Z	20	23	46	41	6	1.11	
221	ETAAS-Z	19	21	41	36	4	1.00	
232	ASV-3010	18	22	46	40	3	1.06	
237	ETAAS-Z	18	22	45	40	3	1.05	
243	ASV-3010	17	20	42	37	2	0.98	
249	ASV-3010	17	20	40	33 \downarrow	3	0.94	
254	ETAAS-Z	18	21	42	38	4	1.01	
255	ETAAS-Z	17	22	42	37	3	1.00	
261	ETAAS-Z	16	20	39	34	3	0.92	
269	ETAAS-Z	18	20	42	36	3	0.98	
271	ASV-3010	15	18	40	35	1	0.90	
272	ETAAS-Z	17	21	44	37	3	1.00	
279	ETAAS-Z	15	18	33 \downarrow	28 \downarrow	3	0.81	
282	ASV-3010	13	15 \downarrow	34 \downarrow	30 \downarrow	1	0.76	
286	ASV-LeadCare	16	18	39	31 \downarrow	<3	0.88	
290	ICP-MS	18	21	43	38	3	1.01	

notes: \uparrow reported value outside upper limit
 \downarrow reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.
Info only: results included for informational purposes only.

**New York State Department of Health
Blood Lead Test Results, 2010 Event #1
PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ($\mu\text{g/dL}$ whole blood)					Normalized Mean	Info Only
		BE10-01	BE10-02	BE10-03	BE10-04	BE10-05		
Target Values:		17	21	43	38	3		
291	ASV-3010	14	16 ↓	28 ↓	26 ↓	17 ↑	1.72	
293	ICP-MS	17	22	44	38	3	1.02	
295	ASV-3010	15	19	40	35	3	0.91	
300	ASV-3010	18	21	44	38	<3	1.02	
301	ETAAS-Z	17	21	42	36	3	0.98	
305	ETAAS-Z	17	21	41	36	3	0.98	
312	ICP-MS	17	21	44	38	3	1.01	
317	ETAAS-Z	19	23	47	40	4	1.09	
324	ICP-MS	18	22	42	39	3	1.03	
325	ETAAS-Z	15	18	40	35	3	0.90	
333	ETAAS-Z	18	22	43	38	4	1.03	
339	HR-ICP-MS	16	19	40	35	3	0.92	Info
340	ETAAS-Z	17	20	40	35	3	0.95	
343	ASV-LeadCare	17	18	42	35	3	0.94	Info
348	ETAAS-Z	17	21	41	36	3	0.98	
349	ETAAS-Z	17	22	43	39	3	1.02	
350	ASV-3010	18	23	46	41	4	1.08	
352	ASV-3010	18	23	46	41	4	1.08	
353	ETAAS-Z	16	20	43	37	<2	0.97	
359	ICP-MS	15	18	38 ↓	31 ↓	3	0.86	
365	ETAAS-Z	19	24	49 ↑	42	3	1.13	
366	ETAAS-Z	17	23	46	41	3	1.06	Info
367	ETAAS-Z	17	21	42	39	3	1.00	Info
368	ASV-3010	15	16 ↓	40	31 ↓	2	0.85	
369	ASV-3010	17	19	40	33 ↓	3	0.93	
374	ASV-3010	15	18	40	34	2	0.89	
376	ASV-LeadCare	14	20	40	33 ↓	3	0.89	
383	ETAAS-Z	17	20	42	35	3	0.96	
384	ASV-3010	17	20	42	38	4	0.98	
385	ICP-MS	17	21	41	35	4	0.97	Info

notes: ↑ reported value outside upper limit
↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.
Info only: results included for informational purposes only.

**New York State Department of Health
Blood Lead Test Results, 2010 Event #1
PERFORMANCE OF PARTICIPATING LABORATORIES**

Lab Code	Method	Results ($\mu\text{g/dL}$ whole blood)					Normalized Mean	Info Only
		BE10-01	BE10-02	BE10-03	BE10-04	BE10-05		
Target Values:		17	21	43	38	3		
388	ASV-LeadCare	16	22	46	34	2	0.99	
389	ETAAS-Z	18	21	43	38	3	1.01	
391	ETAAS Other	16	20	40	32 ↓	6	0.92	Info
395	ICP-MS	17	21	43	38	3	1.00	
401	ETAAS-Z	19	23	46	39	4	1.08	Info
408	ICP-MS	16	20	40	35	3	0.94	Info
410	ICP-MS	18	22	45	39	3	1.04	Info
449	ASV-LeadCare	17	16 ↓	47	37	3	0.96	
455	ASV-LeadCare	14	15 ↓	35 ↓	33 ↓	<3	0.81	
456	ASV-LeadCare	13	16 ↓	37 ↓	32 ↓	3	0.81	
Percent satisfactory results for all participants:							92.0 %	

notes: ↑ reported value outside upper limit
↓ reported value outside lower limit

Normalized mean: The average of each reported result divided by the corresponding target value. It measures bias.
Info only: results included for informational purposes only.

**New York State Department of Health
Blood Lead Test Results, 2010 Event #1
STATISTICAL SUMMARY**

		TARGET VALUE ASSIGNMENT AND STATISTICS				
Lab Code	Method	Results ($\mu\text{g/dL}$ whole blood)				
		BE10-01	BE10-02	BE10-03	BE10-04	BE10-05
103	DRC/CC-ICP-MS	17	20	41	36	3
104	ETAAS-Z	18	22	44	39	3
107	ICP-MS	18	22	44	39	3
109	ETAAS-Z	19	23	46	41	4
110	ETAAS-Z	18	22	44	38	4
110	ICP-MS	18	21	43	38	3
112	ASV-3010	15	19	39	34	<2
147	ICP-MS	17	20	41	36	3
159	ICP-MS	17	20	42	36	3
164	ICP-MS	18	22	45	39	3
166	ASV-3010	17	20	41	36	2
179	ICP-MS	17	22	44	37	3
198	ETAAS-Z	17	22	43	38	3
199	ETAAS-Z	18	22	43	39	3
200	ETAAS-Z	17	21	45	38	2
243	ASV-3010	17	20	42	37	2
293	ICP-MS	17	22	44	38	3
324	ICP-MS	18	22	42	39	3
325	ETAAS-Z	15	18	40	35	3
350	ASV-3010	18	23	46	41	4
Number of Sample Measurements:		20	20	20	20	19
Mean (target value):		17	21	43	38	3
Standard Deviation:		1.0	1.3	1.9	1.8	0.6
RSD (%):		5.7	6.4	4.5	4.9	19.2
Acceptable Range:						
Upper Limit:		21	25	47	42	7
Lower Limit:		13	17	39	34	0

notes: Results reported as less than the detection limits are treated as zero for statistical and grading purposes.

**New York State Department of Health
Blood Lead Test Results, 2010 Event #1
STATISTICAL SUMMARY BY METHOD**

	Results ($\mu\text{g/dL}$ whole blood)				
	BE10-01	BE10-02	BE10-03	BE10-04	BE10-05
ASV-3010					
Number of Sample Measurements:	17	17	17	17	14
Mean:	16.2	19.2	40.4	35.2	2.6
Standard Deviation:	1.6	2.3	4.4	4.0	1.0
RSD (%):	9.6	12.0	10.9	11.3	39.5
ASV-LeadCare					
Number of Sample Measurements:	14	14	14	14	9
Mean:	16.1	18.9	41.6	34.8	2.9
Standard Deviation:	1.7	2.2	3.9	2.9	0.3
RSD (%):	10.8	11.7	9.4	8.3	11.5
DRC/CC-ICP-MS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	17.0	20.0	41.0	36.0	3.0
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ETAAS Other					
Number of Sample Measurements:	2	2	2	2	1
Mean:	16.5	20.5	41.5	34.0	3.0
Standard Deviation:	0.7	0.7	2.1	2.8	?
RSD (%):	—	—	—	—	—
ETAAS-Z					
Number of Sample Measurements:	45	45	45	45	40
Mean:	17.6	21.5	43.5	38.1	3.2
Standard Deviation:	1.3	1.6	3.8	4.9	0.6
RSD (%):	7.4	7.3	8.7	12.8	19.2
HR-ICP-MS					
Number of Sample Measurements:	1	1	1	1	1
Mean:	16.0	19.0	40.0	35.0	3.0
Standard Deviation:	?	?	?	?	?
RSD (%):	—	—	—	—	—
ICP-MS					
Number of Sample Measurements:	20	20	20	20	20
Mean:	17.2	21.2	43.2	37.3	3.1
Standard Deviation:	0.9	1.4	3.7	2.7	0.2
RSD (%):	5.4	6.7	8.5	7.2	7.3
All Laboratories					
Number of Sample Measurements:	100	100	100	100	86
Mean:	17.0	20.6	42.6	36.8	3.0
Standard Deviation:	1.4	2.1	4.0	4.2	0.6
RSD (%):	8.5	10.0	9.3	11.4	21.1

notes: ? Insufficient data for calculation.

**New York State Department of Health
Blood Lead Test Results, 2010 Event #1
STATISTICAL SUMMARY BY CLASS**

	Results ($\mu\text{g/dL}$ whole blood)				
	BE10-01	BE10-02	BE10-03	BE10-04	BE10-05
Evaluated					
Number of Sample Measurements:	62	62	62	62	53
Mean:	16.9	20.4	42.3	36.3	3.0
Standard Deviation:	1.6	2.3	4.5	4.0	0.7
RSD (%):	9.4	11.2	10.7	10.9	23.9
Info					
Number of Sample Measurements:	18	18	18	18	14
Mean:	17.2	20.7	43.1	37.8	3.1
Standard Deviation:	1.4	1.8	3.5	6.2	0.4
RSD (%):	7.9	8.6	8.2	16.5	11.6
Reference					
Number of Sample Measurements:	20	20	20	20	19
Mean:	17.3	21.2	43.0	37.7	3.0
Standard Deviation:	1.0	1.3	1.9	1.8	0.6
RSD (%):	5.7	6.4	4.5	4.9	19.2
All Laboratories					
Number of Sample Measurements:	100	100	100	100	86
Mean:	17.0	20.6	42.6	36.8	3.0
Standard Deviation:	1.4	2.1	4.0	4.2	0.6
RSD (%):	8.5	10.0	9.3	11.4	21.1

notes: ? Insufficient data for calculation.

**New York State Department of Health
Blood Lead Test Results, 2010 Event #1**

METHOD NOTES

ATOMIC SPECTROMETRY METHODS

- A-1 ETAAS-Z (Electrothermal atomic absorption spectrometry with Zeeman background correction)
- A-2 ETAAS other (i.e., D₂, S-H background correction)
- A-3 FAAS (Flame atomic absorption spectrometry)
- A-4 CV-AAS (Cold vapor atomic absorption spectrometry)
- A-5 HG-AAS (Hydride generation atomic absorption spectrometry)
- A-6 AFS (Atomic fluorescence spectrometry)
- A-7 Other

INDUCTIVELY COUPLED PLASMA

- P-1 ICP-MS (Inductively coupled plasma - mass spectrometry)
- P-2 DRC/CC-ICP-MS (ICP-MS used in the Dynamic Reaction Cell or Collision Cell mode)
- P-3 ICP-AES/OES (ICP atomic/optical emission spectrometry)
- P-4 HR-ICP-MS (High resolution ICP-MS)
- P-5 ETV-ICP-MS (Electrothermal vaporization ICP-MS)
- P-6 ID-ICP-MS (Isotope dilution ICP-MS)
- P-7 Other

ELECTROCHEMICAL METHODS

- E-1 ASV (Anodic stripping voltammetry without digestion)
- E-2 ASV-LeadCare® (Anodic stripping voltammetry using the ESA LeadCare® system)
- E-3 Fluoride specific electrode
- E-4 Other

MOLECULAR FLUORIMETRY

- F-1 EtOAc (Ethyl acetate-acetic acid extraction method for determination of erythrocyte protoporphyrin)
- F-2 Aviv hematofluorometry (for determination of EP at hematocrit 35)
- F-3 Helena ZPP (for determination of zinc protoporphyrin in $\mu\text{mol ZPP/mol heme}$)
- F-4 Other

OTHER METHODS

If your method is not listed in the above list, please describe it briefly.
