

**Immunohematology Proficiency Testing Program
New York State Department of Health
Wadsworth Center
Empire State Plaza, PO Box 509
Albany, New York 12201-0509**

Performance Summary for the June 16, 2014 Proficiency Test Event

Number of Participating Laboratories 261

Satisfactory* 255 Unsatisfactory 6**

Specimen #	Number (%) of acceptable responses									
	ABO Group	Subgroups of A	Direct Antiglobulin Test			Rh ₀ (D) Group	Rh Phenotype			
			Polyspecific	IgG	Complement		C (rh')	E (rh'')	c (hr')	e (hr'')
NYS-14201	257(100%)	191(98%)	193(99%)	215(100%)	155(99%)	259(100%)	194(99%)	195(100%)	195(100%)	191(100%)
NYS-14202	257(100%)	193(99%)	194(100%)	215(100%)	156(100%)	259(100%)	195(100%)	195(100%)	194(99%)	191(100%)
NYS-14203	257(100%)	192(99%)	194(100%)	215(100%)	156(100%)	259(100%)	195(100%)	195(100%)	195(100%)	191(100%)
NYS-14204	257(100%)	194(100%)	194(100%)	215(100%)	156(100%)	259(100%)	195(100%)	195(100%)	195(100%)	191(100%)
NYS-14205	257(100%)	194(100%)	194(100%)	215(100%)	155(99%)	258(99%)	194(99%)	194(99%)	195(100%)	191(100%)
Total labs	257	194	194	215	156	259	195	195	195	191

Specimen #	Number (%) of acceptable responses		
	Unexpected Antibody Detection	Antibody Identification	Compatibility Testing
NYS-14201	241(99%)	195(100%)	205(100%)
NYS-14202	240(99%)	195(100%)	204(99%)
NYS-14203	242(100%)	195(100%)	205(100%)
NYS-14204	242(100%)	195(100%)	205(100%)
NYS-14205	241(99%)	195(100%)	205(100%)
Total labs	242	195	205

* Satisfactory – No or minor error(s) in interpretation

** Unsatisfactory – Significant error(s) in interpretation or failure to participate

REMINDER: The next scheduled shipment date is October 27, 2014. If specimens are not received within five business days of the mailout date, the specimens are received in unsatisfactory condition, or the shipment is incomplete, please contact the Immunohematology PT Program at (518) 485-5341.