

Nirav R. Shah, M.D., M.P.H.
Commissioner

NEW YORK
state department of
HEALTH

Sue Kelly
Executive Deputy Commissioner

Statistical Report
Urinalysis/Urine Pregnancy Testing
Mail out date: November 7, 2011

This report summarizes data from the urinalysis/urine pregnancy proficiency testing of November 7, 2011.

If you have any questions regarding this report, please contact Regina Saulsbery at 518-474-6156. You may also contact us by e-mail: clinchem@wadsworth.org

Results are shown for each analyte arranged by instrument and reagent system. Results for specific gravity and pH are listed as mean \pm 1SD for each of the instruments and reagents. All other results are shown as the number of participant results for that instrument or reagent.

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

PLEASE NOTE: The evaluation report and summary of the educational quantitative urine clinical chemistry proficiency test will be mailed separately if your laboratory participated in that proficiency test.

HEALTH.NY.GOV
[facebook.com/NYSDOH](https://www.facebook.com/NYSDOH)
twitter.com/HealthNYGov

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Specific Gravity

Specimen: U91	Specimen: U92	Specimen: U93	Specimen: U94	Specimen: U95	Number	Instrument or Reagent System
1.016 ± 0.003	1.014 ± 0.003	1.023 ± 0.006	1.018 ± 0.004	1.016 ± 0.005	n = 374	All Methods & Instruments
1.015	1.015	1.021	1.015	1.017	n = 374	Median value, All Methods & Instruments
<Instruments>						
1.016 ± 0.003	1.010 ± 0.003	1.026 ± 0.005	1.013 ± 0.003	1.010 ± 0.003	n = 30	Visual examination
1.015 ± 0.002	1.016 ± 0.002	1.020 ± 0.002	1.022 ± 0.002	1.021 ± 0.001	n = 92	Siemens Atlas
1.018 ± 0.003	1.017 ± 0.003	1.029 ± 0.002	1.018 ± 0.003	1.018 ± 0.003	n = 36	Siemens Clinitek Status
1.016 ± 0.005	1.013 ± 0.003	1.030 ± 0.000	1.013 ± 0.003	1.012 ± 0.003	n = 8	Siemens Clinitek 50
1.019 ± 0.004	1.017 ± 0.003	1.027 ± 0.003	1.019 ± 0.002	1.018 ± 0.004	n = 6	Siemens Clinitek 100
1.018 ± 0.003	1.013 ± 0.003	1.029 ± 0.003	1.016 ± 0.004	1.014 ± 0.003	n = 54	Siemens Clinitek 500
1.018 ± 0.004	1.013 ± 0.005	1.028 ± 0.004	1.015 ± 0.003	1.014 ± 0.003	n = 37	Siemens Clinitek Advantus
1.017 ± 0.003	1.012 ± 0.003	1.022 ± 0.003	1.015 ± 0.005	1.008 ± 0.003	n = 3	Roche UA Reader/Super UA Reader
1.016 ± 0.002	1.011 ± 0.003	1.020 ± 0.001	1.013 ± 0.002	1.010 ± 0.002	n = 13	Roche Chemstrip Criterion
1.015 ± 0.002	1.012 ± 0.003	1.020 ± 0.001	1.018 ± 0.003	1.014 ± 0.007	n = 36	Roche Urisys
1.011 ± 0.001	1.012 ± 0.001	1.014 ± 0.001	1.015 ± 0.002	1.018 ± 0.001	n = 37	IRIS AUTION MAX
1.013 ± 0.001	1.014 ± 0.001	1.019 ± 0.003	1.021 ± 0.002	1.020 ± 0.001	n = 18	Refractometer
1.015 ± 0.000	1.012 ± 0.003	1.020 ± 0.000	1.015 ± 0.000	1.013 ± 0.003	n = 3	Other
<Reagents>						
1.018 ± 0.004	1.013 ± 0.004	1.028 ± 0.003	1.016 ± 0.004	1.014 ± 0.004	n = 164	Siemens Reagent Strips
1.015 ± 0.002	1.016 ± 0.002	1.020 ± 0.002	1.022 ± 0.002	1.021 ± 0.002	n = 83	Siemens Atlas Reagent
1.016 ± 0.002	1.011 ± 0.003	1.020 ± 0.001	1.015 ± 0.002	1.010 ± 0.003	n = 44	Roche Chemstrips
1.014 ± 0.001	1.015 ± 0.000	1.019 ± 0.001	1.022 ± 0.001	1.021 ± 0.000	n = 15	Roche Urisys Cassette
1.011 ± 0.001	1.012 ± 0.001	1.015 ± 0.001	1.015 ± 0.001	1.018 ± 0.001	n = 22	ARKRAY Aution Sticks
1.011 ± 0.001	1.012 ± 0.001	1.014 ± 0.001	1.015 ± 0.002	1.018 ± 0.001	n = 9	IRIS Reagent Strips
1.017 ± 0.004	1.017 ± 0.007	1.025 ± 0.006	1.018 ± 0.004	1.018 ± 0.006	n = 8	Bayer Clinitest
1.013 ± 0.002	1.014 ± 0.002	1.018 ± 0.004	1.019 ± 0.004	1.019 ± 0.002	n = 24	None, no reagent or strip used
1.015 ± 0.001	1.015 ± 0.004	1.025 ± 0.006	1.018 ± 0.003	1.016 ± 0.004	n = 5	Other

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

pH

Specimen: U91	Specimen: U92	Specimen: U93	Specimen: U94	Specimen: U95	Number	Instrument or Reagent System
6.16 ± 0.28	6.96 ± 0.17	6.18 ± 0.28	6.85 ± 0.26	7.36 ± 0.27	n = 380	All Methods & Instruments
6.00	7.00	6.00	7.00	7.50	n = 380	Median value, All Methods & Instruments
						<Instruments>
6.08 ± 0.31	7.03 ± 0.26	6.09 ± 0.38	6.77 ± 0.32	7.51 ± 0.18	n = 39	Visual examination
6.31 ± 0.28	6.99 ± 0.05	6.45 ± 0.15	6.98 ± 0.13	7.48 ± 0.12	n = 91	Siemens Atlas
6.09 ± 0.22	7.00 ± 0.00	6.18 ± 0.24	6.99 ± 0.08	7.38 ± 0.22	n = 40	Siemens Clinitek Status
6.00 ± 0.00	6.94 ± 0.18	6.00 ± 0.00	6.81 ± 0.26	7.50 ± 0.00	n = 8	Siemens Clinitek 50
6.00 ± 0.00	7.00 ± 0.00	6.06 ± 0.18	7.00 ± 0.00	7.31 ± 0.37	n = 8	Siemens Clinitek 100
6.06 ± 0.17	6.90 ± 0.20	6.01 ± 0.07	6.77 ± 0.31	7.37 ± 0.24	n = 57	Siemens Clinitek 500
6.05 ± 0.15	6.86 ± 0.28	6.04 ± 0.18	6.78 ± 0.30	7.42 ± 0.18	n = 39	Siemens Clinitek Advantus
6.33 ± 0.29	7.00 ± 0.00	6.33 ± 0.29	6.83 ± 0.29	7.00 ± 0.00	n = 3	Roche UA Reader/Super UA Reader
6.11 ± 0.21	7.00 ± 0.00	6.04 ± 0.13	6.93 ± 0.18	7.21 ± 0.43	n = 14	Roche Chemstrip Criterion
6.10 ± 0.35	6.97 ± 0.17	6.07 ± 0.34	6.96 ± 0.14	7.19 ± 0.40	n = 36	Roche Urisys
6.42 ± 0.22	6.93 ± 0.17	6.21 ± 0.25	6.51 ± 0.08	7.03 ± 0.16	n = 38	IRIS AUTION MAX
6.08 ± 0.18	6.86 ± 0.22	6.16 ± 0.36	6.66 ± 0.23	7.22 ± 0.44	n = 5	Other
						<Reagents>
6.08 ± 0.19	6.95 ± 0.19	6.09 ± 0.19	6.84 ± 0.27	7.41 ± 0.22	n = 179	Siemens Reagent Strips
6.31 ± 0.28	6.99 ± 0.05	6.45 ± 0.16	6.99 ± 0.08	7.48 ± 0.12	n = 84	Siemens Atlas Reagent
6.02 ± 0.31	6.98 ± 0.18	5.96 ± 0.31	6.83 ± 0.28	7.21 ± 0.37	n = 48	Roche Chemstrips
6.20 ± 0.41	6.93 ± 0.26	6.20 ± 0.41	7.00 ± 0.00	7.33 ± 0.49	n = 15	Roche Urisys Cassette
6.38 ± 0.26	6.91 ± 0.19	6.22 ± 0.25	6.52 ± 0.09	7.03 ± 0.19	n = 29	ARKRAY Aution Sticks
6.50 ± 0.00	7.00 ± 0.00	6.15 ± 0.24	6.50 ± 0.00	7.00 ± 0.00	n = 10	IRIS Reagent Strips
6.19 ± 0.26	6.88 ± 0.35	6.44 ± 0.32	6.88 ± 0.35	7.44 ± 0.18	n = 8	Bayer Clinitest
6.10 ± 0.30	6.87 ± 0.20	6.13 ± 0.33	6.80 ± 0.24	7.23 ± 0.23	n = 6	Other

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Bilirubin	Specimen: U91	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
	Visual examination	67	0	0	0	0
	Siemens Atlas	91	0	0	0	0
	Siemens Clinitek Status	38	0	0	0	0
	Siemens Clinitek 50	8	0	0	0	0
	Siemens Clinitek 100	9	0	0	0	0
	Siemens Clinitek 500	55	0	0	0	0
	Siemens Clinitek Advantus	39	0	0	0	0
	Roche UA Reader/Super UA Read	3	0	0	0	0
	Roche Chemstrip Criterion	13	0	0	0	0
	Roche Urisys	34	0	0	0	0
	IRIS AUTION MAX	38	0	0	0	0
	Other	5	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	172	0	0	0	0
	Siemens Atlas Reagent	82	0	0	0	0
	Roche Chemstrips	45	0	0	0	0
	Roche Urisys Cassette	14	0	0	0	0
	ARKRAY Aution Sticks	29	0	0	0	0
	IRIS Reagent Strips	10	0	0	0	0
	Bayer Clinitest	8	0	0	0	0
	Siemens Ictotest	35	0	0	0	0
	Other	5	0	0	0	0
	All Instruments and Reagents	400	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Bilirubin	Specimen: U92	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		1	2	18	47	136
Siemens Atlas		1	0	31	55	4
Siemens Clinitek Status		0	0	16	22	0
Siemens Clinitek 50		0	0	1	7	0
Siemens Clinitek 100		0	0	0	8	1
Siemens Clinitek 500		1	1	51	3	1
Siemens Clinitek Advantus		1	0	38	0	0
Roche UA Reader/Super UA Read		0	2	0	0	1
Roche Chemstrip Criterion		4	8	1	0	0
Roche Urisys		0	34	0	0	0
IRIS AUTION MAX		1	1	35	0	1
Other		0	3	1	0	1
<Reagents>						
Siemens Reagent Strips		1	1	102	66	3
Siemens Atlas Reagent		1	0	29	49	3
Roche Chemstrips		4	33	2	6	1
Roche Urisys Cassette		0	14	0	0	0
ARKRAY Aution Sticks		1	1	26	0	1
IRIS Reagent Strips		0	0	10	0	0
Bayer Clinitest		1	0	6	1	1
Siemens Ictotest		1	1	16	18	135
Other		0	1	1	2	1
All Instruments and Reagents		9	51	192	142	145

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Bilirubin	Specimen: U93	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		67	0	0	0	0
Siemens Atlas		91	0	0	0	0
Siemens Clinitek Status		38	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0
Siemens Clinitek 500		55	0	0	0	0
Siemens Clinitek Advantus		38	0	1	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		34	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0
Other		5	0	0	0	0
<Reagents>						
Siemens Reagent Strips		172	0	0	0	0
Siemens Atlas Reagent		82	0	0	0	0
Roche Chemstrips		45	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0
Bayer Clinitest		7	0	1	0	0
Siemens Ictotest		35	0	0	0	0
Other		5	0	0	0	0
All Instruments and Reagents		399	0	1	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Bilirubin	Specimen: U94	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		67	0	0	0	0
Siemens Atlas		90	0	1	0	0
Siemens Clinitek Status		38	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0
Siemens Clinitek 500		55	0	0	0	0
Siemens Clinitek Advantus		39	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		34	0	0	0	0
IRIS AUTION MAX		37	0	1	0	0
Other		5	0	0	0	0
<Reagents>						
Siemens Reagent Strips		172	0	0	0	0
Siemens Atlas Reagent		81	0	1	0	0
Roche Chemstrips		45	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0
ARKRAY Aution Sticks		28	0	1	0	0
IRIS Reagent Strips		10	0	0	0	0
Bayer Clinitest		8	0	0	0	0
Siemens Ictotest		35	0	0	0	0
Other		5	0	0	0	0
All Instruments and Reagents		398	0	2	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Bilirubin	Specimen: U95	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		67	0	0	0	0
Siemens Atlas		91	0	0	0	0
Siemens Clinitek Status		38	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0
Siemens Clinitek 500		55	0	0	0	0
Siemens Clinitek Advantus		39	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		34	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0
Other		5	0	0	0	0
<Reagents>						
Siemens Reagent Strips		172	0	0	0	0
Siemens Atlas Reagent		82	0	0	0	0
Roche Chemstrips		45	0	0	0	0
Roche Urisys Cassette		14	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0
Bayer Clinitest		8	0	0	0	0
Siemens Ictotest		35	0	0	0	0
Other		5	0	0	0	0
All Instruments and Reagents		400	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Glucose	Specimen: U91	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		47	48	24	1	0	0	0
Siemens Atlas		19	73	1	0	0	0	0
Siemens Clinitek Status		4	37	0	0	0	0	0
Siemens Clinitek 50		0	8	0	0	0	0	0
Siemens Clinitek 100		7	2	0	0	0	0	0
Siemens Clinitek 500		9	49	0	0	0	0	0
Siemens Clinitek Advantus		5	34	0	0	0	0	0
Roche UA Reader/Super UA Read		0	3	0	0	0	0	0
Roche Chemstrip Criterion		1	13	0	0	0	0	0
Roche Urisys		0	37	0	0	0	0	0
IRIS AUTION MAX		1	38	0	0	0	0	0
Hitachi/Roche Analyzer		0	2	0	0	0	0	0
Other Chemistry Analyzer		1	0	0	0	0	0	0
Other		2	3	1	0	0	0	0
<Reagents>								
Siemens Reagent Strips		41	142	0	0	0	0	0
Siemens Atlas Reagent		17	65	1	0	0	0	0
Sulfosalicylic Acid		0	1	0	0	0	0	0
Roche Chemstrips		2	50	0	0	0	0	0
Roche Urisys Cassette		0	15	0	0	0	0	0
ARKRAY Aution Sticks		1	29	0	0	0	0	0
IRIS Reagent Strips		0	10	0	0	0	0	0
Bayer Clinitest		33	23	25	1	0	0	0
Bayer Diastix		0	4	0	0	0	0	0
Hitachi/Roche Reagents		0	2	0	0	0	0	0
Other		2	6	0	0	0	0	0
All Instruments and Reagents		96	347	26	1	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Glucose	Specimen: U92	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		86	0	0	0	0	0	0
Siemens Atlas		92	0	0	1	0	0	0
Siemens Clinitek Status		41	0	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0	0
Siemens Clinitek Advantus		38	1	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		37	0	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	1	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
Other Chemistry Analyzer		1	0	0	0	0	0	0
Other		6	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		183	0	0	0	0	0	0
Siemens Atlas Reagent		82	0	0	1	0	0	0
Sulfosalicylic Acid		1	0	0	0	0	0	0
Roche Chemstrips		49	0	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	1	0
IRIS Reagent Strips		10	0	0	0	0	0	0
Bayer Clinitest		52	1	0	0	0	0	0
Hitachi/Roche Reagents		2	0	0	0	0	0	0
Other		8	0	0	0	0	0	0
All Instruments and Reagents		431	1	0	1	0	1	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Glucose	Specimen: U93	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		0	23	40	46	12	9	0
Siemens Atlas		0	1	10	82	0	0	0
Siemens Clinitek Status		0	25	15	1	0	0	0
Siemens Clinitek 50		0	4	4	0	0	0	0
Siemens Clinitek 100		0	6	1	1	0	1	0
Siemens Clinitek 500		3	37	19	0	0	0	0
Siemens Clinitek Advantus		2	26	12	0	0	0	0
Roche UA Reader/Super UA Read		0	0	1	1	0	1	0
Roche Chemstrip Criterion		0	0	7	2	1	4	0
Roche Urisys		0	0	15	3	0	19	0
IRIS AUTION MAX		0	0	19	16	1	3	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other Chemistry Analyzer		0	0	1	0	0	0	0
Other		2	0	2	1	0	1	0
<Reagents>								
Siemens Reagent Strips		4	108	60	11	0	0	0
Siemens Atlas Reagent		0	2	9	72	0	0	0
Sulfosalicylic Acid		0	0	1	0	0	0	0
Roche Chemstrips		0	0	23	11	1	18	0
Roche Urisys Cassette		0	0	3	1	0	11	0
ARKRAY Aution Sticks		0	0	17	10	0	3	0
IRIS Reagent Strips		0	0	3	6	1	0	0
Bayer Clinitest		1	8	24	42	12	6	0
Bayer Diastix		0	3	1	0	0	0	0
Hitachi/Roche Reagents		0	0	2	0	0	0	0
Other		2	1	5	0	0	0	0
All Instruments and Reagents		7	122	148	153	14	38	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Glucose	Specimen: U94	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		0	0	0	9	20	82	18
Siemens Atlas		1	0	0	72	0	20	0
Siemens Clinitek Status		0	1	2	24	1	12	1
Siemens Clinitek 50		0	0	0	0	0	8	0
Siemens Clinitek 100		0	0	0	1	0	7	1
Siemens Clinitek 500		2	0	0	10	1	45	0
Siemens Clinitek Advantus		0	0	1	5	1	33	0
Roche UA Reader/Super UA Read		0	0	0	0	0	3	0
Roche Chemstrip Criterion		0	0	1	1	0	12	0
Roche Urisys		0	0	1	3	0	33	0
IRIS AUTION MAX		1	0	5	5	2	25	1
Hitachi/Roche Analyzer		0	0	0	0	2	0	0
Other Chemistry Analyzer		0	0	0	0	1	0	0
Other		0	0	1	1	1	3	0
<Reagents>								
Siemens Reagent Strips		2	1	2	44	2	123	9
Siemens Atlas Reagent		1	0	0	64	0	18	0
Sulfosalicylic Acid		0	0	0	1	0	0	0
Roche Chemstrips		0	0	1	5	1	45	0
Roche Urisys Cassette		0	0	1	1	0	13	0
ARKRAY Aution Sticks		1	0	5	4	1	18	1
IRIS Reagent Strips		0	0	0	2	1	7	0
Bayer Clinitest		0	0	1	9	22	51	10
Bayer Diastix		0	0	0	0	0	3	0
Hitachi/Roche Reagents		0	0	0	0	2	0	0
Other		0	0	1	1	0	5	1
All Instruments and Reagents		4	1	11	131	29	283	21

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Glucose	Specimen: U95	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		85	0	0	0	0	0	0
Siemens Atlas		92	1	0	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0	0
Siemens Clinitek Advantus		38	1	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		37	0	0	0	0	0	0
IRIS AUTION MAX		38	1	0	0	0	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
Other Chemistry Analyzer		1	0	0	0	0	0	0
Other		6	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		182	1	0	0	0	0	0
Siemens Atlas Reagent		82	1	0	0	0	0	0
Sulfosalicylic Acid		1	0	0	0	0	0	0
Roche Chemstrips		49	0	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0	0
ARKRAY Aution Sticks		29	1	0	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0	0
Bayer Clinitest		52	0	0	0	0	0	0
Hitachi/Roche Reagents		2	0	0	0	0	0	0
Other		8	0	0	0	0	0	0
All Instruments and Reagents		430	3	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Hemoglobin	Specimen: U91	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		43	0	0	0	0	0
Siemens Atlas		92	0	0	0	0	0
Siemens Clinitek Status		39	2	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0
Siemens Clinitek Advantus		37	2	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		35	1	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
Hitachi/Roche Analyzer		0	0	1	0	0	0
Other		4	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		178	4	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0	0
Roche Chemstrips		47	1	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0
Bayer Clinitest		8	0	0	0	0	0
Siemens Hemastix		2	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		4	0	1	0	0	0
All Instruments and Reagents		377	5	1	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Hemoglobin	Specimen: U92	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		42	0	0	0	0	0
Siemens Atlas		92	0	0	0	0	0
Siemens Clinitek Status		40	1	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0
Siemens Clinitek Advantus		38	0	0	0	0	1
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
Other		4	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		180	1	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0	0
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0
Bayer Clinitest		7	0	0	0	0	1
Siemens Hemastix		2	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		5	0	0	0	0	0
All Instruments and Reagents		380	1	0	0	0	1

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Hemoglobin	Specimen: U93	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		1	0	0	3	0	42
Siemens Atlas		0	0	0	13	0	79
Siemens Clinitek Status		0	0	0	15	0	26
Siemens Clinitek 50		0	0	0	1	0	7
Siemens Clinitek 100		1	0	0	0	0	7
Siemens Clinitek 500		1	0	0	6	0	49
Siemens Clinitek Advantus		1	0	0	0	0	38
Roche UA Reader/Super UA Read		0	0	0	1	1	1
Roche Chemstrip Criterion		0	0	0	0	0	14
Roche Urisys		0	0	0	9	13	14
IRIS AUTION MAX		0	0	1	18	0	19
Hitachi/Roche Analyzer		0	0	0	0	0	1
Other		0	0	0	0	1	3
<Reagents>							
Siemens Reagent Strips		2	0	0	22	0	158
Siemens Atlas Reagent		0	0	0	10	0	74
Roche Chemstrips		0	0	0	9	10	31
Roche Urisys Cassette		0	0	0	3	5	7
ARKRAY Aution Sticks		0	0	1	13	0	15
IRIS Reagent Strips		0	0	0	5	0	5
Bayer Clinitest		1	0	0	2	0	5
Siemens Hemastix		0	0	0	0	0	2
None, no reagent or strip use		1	0	0	0	0	0
Other		0	0	0	2	0	3
All Instruments and Reagents		4	0	1	66	15	300

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Hemoglobin	Specimen: U94	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		41	1	0	0	0	0
Siemens Atlas		91	0	0	0	0	1
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0
Siemens Clinitek Advantus		39	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
Other		4	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	0	0	0	0
Siemens Atlas Reagent		82	0	0	0	0	1
Roche Chemstrips		47	1	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0
Bayer Clinitest		8	0	0	0	0	0
Siemens Hemastix		2	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		5	0	0	0	0	0
All Instruments and Reagents		380	1	0	0	0	1

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Hemoglobin	Specimen: U95	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		42	0	0	0	0	0
Siemens Atlas		91	0	0	0	0	1
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0
Siemens Clinitek 100		8	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0
Siemens Clinitek Advantus		39	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
Other		4	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	0	0	0	0	0
Siemens Atlas Reagent		82	0	0	0	0	1
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0
Bayer Clinitest		8	0	0	0	0	0
Siemens Hemastix		2	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		5	0	0	0	0	0
All Instruments and Reagents		381	0	0	0	0	1

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Ketones	Specimen: U91	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg	Large 3+ >=80mg/dL	Positive
<Instruments>							
	Visual examination	48	0	0	0	0	0
	Siemens Atlas	93	0	0	0	0	0
	Siemens Clinitek Status	40	0	0	0	0	0
	Siemens Clinitek 50	8	0	0	0	0	0
	Siemens Clinitek 100	9	0	0	0	0	0
	Siemens Clinitek 500	56	0	0	0	0	0
	Siemens Clinitek Advantus	40	0	0	0	0	0
	Roche UA Reader/Super UA Read	3	0	0	0	0	0
	Roche Chemstrip Criterion	14	0	0	0	0	0
	Roche Urisys	35	0	0	0	0	0
	IRIS AUTION MAX	38	0	0	0	0	0
	Other	4	0	0	0	0	0
<Reagents>							
	Siemens Reagent Strips	180	0	0	0	0	0
	Siemens Atlas Reagent	84	0	0	0	0	0
	Roche Chemstrips	47	0	0	0	0	0
	Roche Urisys Cassette	15	0	0	0	0	0
	ARKRAY Aution Sticks	29	0	0	0	0	0
	IRIS Reagent Strips	10	0	0	0	0	0
	Siemens Acetest	9	0	0	0	0	0
	Bayer Clinitest	8	0	0	0	0	0
	Siemens Ictotest	1	0	0	0	0	0
	Other	5	0	0	0	0	0
	All Instruments and Reagents	388	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Ketones	Specimen: U92	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		48	0	0	0	0	0
Siemens Atlas		93	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0
Siemens Clinitek Advantus		39	0	0	1	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		35	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
Other		4	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		180	0	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0
Siemens Acetest		9	0	0	0	0	0
Bayer Clinitest		7	0	0	1	0	0
Siemens Ictotest		1	0	0	0	0	0
Other		5	0	0	0	0	0
All Instruments and Reagents		387	0	0	1	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Ketones	Specimen: U93	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		0	0	7	29	33	3
Siemens Atlas		0	0	2	90	1	0
Siemens Clinitek Status		0	0	1	25	14	0
Siemens Clinitek 50		0	0	0	6	2	0
Siemens Clinitek 100		0	0	0	8	0	1
Siemens Clinitek 500		0	0	4	52	1	0
Siemens Clinitek Advantus		1	0	1	37	1	0
Roche UA Reader/Super UA Read		0	0	0	3	0	0
Roche Chemstrip Criterion		0	0	0	12	2	0
Roche Urisys		0	0	1	20	14	0
IRIS AUTION MAX		0	0	0	38	0	0
Other		0	0	1	2	1	0
<Reagents>							
Siemens Reagent Strips		0	0	5	142	34	0
Siemens Atlas Reagent		0	0	2	81	1	0
Roche Chemstrips		0	0	1	24	23	0
Roche Urisys Cassette		0	0	0	13	2	0
ARKRAY Aution Sticks		0	0	0	29	0	0
IRIS Reagent Strips		0	0	0	10	0	0
Siemens Acetest		0	0	7	13	6	4
Bayer Clinitest		1	0	1	5	2	0
Siemens Ictotest		0	0	0	1	0	0
Other		0	0	1	4	1	0
All Instruments and Reagents		1	0	17	322	69	4

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Ketones	Specimen: U94	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		48	0	0	0	0	0
Siemens Atlas		93	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0
Siemens Clinitek Advantus		40	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		35	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
Other		4	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		180	0	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0
Siemens Acetest		9	0	0	0	0	0
Bayer Clinitest		8	0	0	0	0	0
Siemens Ictotest		1	0	0	0	0	0
Other		5	0	0	0	0	0
All Instruments and Reagents		388	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Ketones	Specimen: U95	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		48	0	0	0	0	0
Siemens Atlas		93	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0
Siemens Clinitek Advantus		40	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		35	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
Other		4	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		180	0	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0
Siemens Acetest		9	0	0	0	0	0
Bayer Clinitest		8	0	0	0	0	0
Siemens Ictotest		1	0	0	0	0	0
Other		5	0	0	0	0	0
All Instruments and Reagents		388	0	0	0	0	0

**New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011**

Protein	Specimen: U91	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	0	9	68	15	19
Siemens Atlas		1	0	0	0	90	3	0
Siemens Clinitek Status		0	0	0	1	38	1	0
Siemens Clinitek 50		0	0	0	0	8	0	0
Siemens Clinitek 100		1	0	0	1	6	0	1
Siemens Clinitek 500		0	0	0	25	32	0	0
Siemens Clinitek Advantus		0	0	0	6	34	0	0
Roche UA Reader/Super UA Read		0	0	0	0	3	0	0
Roche Chemstrip Criterion		0	0	0	0	14	0	0
Roche Urisys		0	0	0	1	35	0	0
IRIS AUTION MAX		1	0	0	3	34	0	0
Hitachi/Roche Analyzer		0	0	0	0	2	0	0
Other		0	0	0	1	3	0	0
<Reagents>								
Siemens Reagent Strips		1	0	0	34	144	5	0
Siemens Atlas Reagent		1	0	0	1	80	3	0
Sulfosalicylic Acid		0	0	0	6	31	9	20
Roche Chemstrips		0	0	0	1	46	1	0
Roche Urisys Cassette		0	0	0	0	15	0	0
ARKRAY Aution Sticks		1	0	0	2	26	0	0
IRIS Reagent Strips		0	0	0	1	9	0	0
Bayer Clinitest		0	0	0	1	7	0	0
Bayer Diastix		0	0	0	0	1	0	0
Other		0	0	0	1	8	1	0
All Instruments and Reagents		3	0	0	47	367	19	20

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Protein	Specimen: U92	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		58	1	0	0	0	0	0
Siemens Atlas		92	0	1	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0	0
Siemens Clinitek Advantus		39	0	0	1	0	0	0
Roche UA Reader/Super UA Read		2	1	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0	0
IRIS AUTION MAX		37	0	1	0	0	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
Other		4	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		183	0	0	0	0	0	0
Siemens Atlas Reagent		84	0	1	0	0	0	0
Sulfosalicylic Acid		15	0	0	0	0	0	0
Roche Chemstrips		45	2	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0	0
ARKRAY Aution Sticks		28	0	1	0	0	0	0
IRIS Reagent Strips		10	0	0	0	0	0	0
Bayer Clinitest		7	0	0	1	0	0	0
Other		10	0	0	0	0	0	0
All Instruments and Reagents		397	2	2	1	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Protein	Specimen: U93	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		1	0	2	53	32	2	19
Siemens Atlas		0	0	0	56	37	0	0
Siemens Clinitek Status		0	0	0	5	34	1	0
Siemens Clinitek 50		0	0	0	0	8	0	0
Siemens Clinitek 100		0	0	0	3	5	0	1
Siemens Clinitek 500		0	0	0	55	2	0	0
Siemens Clinitek Advantus		1	0	0	39	0	0	0
Roche UA Reader/Super UA Read		0	0	0	2	1	0	0
Roche Chemstrip Criterion		0	0	1	10	3	0	0
Roche Urisys		0	0	0	35	1	0	0
IRIS AUTION MAX		0	0	0	37	1	0	0
Hitachi/Roche Analyzer		0	0	0	2	0	0	0
Other		0	0	0	4	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	0	119	63	1	0
Siemens Atlas Reagent		0	0	0	52	33	0	0
Sulfosalicylic Acid		1	0	2	27	12	2	20
Roche Chemstrips		0	0	1	39	8	0	0
Roche Urisys Cassette		0	0	0	14	1	0	0
ARKRAY Aution Sticks		0	0	0	28	1	0	0
IRIS Reagent Strips		0	0	0	10	0	0	0
Bayer Clinitest		1	0	0	4	3	0	0
Bayer Diastix		0	0	0	1	0	0	0
Other		0	0	0	7	3	0	0
All Instruments and Reagents		2	0	3	301	124	3	20

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Protein	Specimen: U94	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	9	59	22	1	0	20
Siemens Atlas		1	0	86	6	0	0	0
Siemens Clinitek Status		0	0	4	36	0	0	0
Siemens Clinitek 50		0	0	1	7	0	0	0
Siemens Clinitek 100		0	0	4	4	0	0	1
Siemens Clinitek 500		0	2	53	2	0	0	0
Siemens Clinitek Advantus		0	2	36	2	0	0	0
Roche UA Reader/Super UA Read		0	0	1	2	0	0	0
Roche Chemstrip Criterion		0	0	13	1	0	0	0
Roche Urisys		0	2	8	26	0	0	0
IRIS AUTION MAX		1	1	36	0	0	0	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other		0	0	4	0	0	0	0
<Reagents>								
Siemens Reagent Strips		0	6	122	54	1	0	0
Siemens Atlas Reagent		1	0	78	6	0	0	0
Sulfosalicylic Acid		0	7	29	9	0	0	21
Roche Chemstrips		0	2	26	20	0	0	0
Roche Urisys Cassette		0	0	3	12	0	0	0
ARKRAY Aution Sticks		1	1	27	0	0	0	0
IRIS Reagent Strips		0	0	10	0	0	0	0
Bayer Clinitest		0	0	3	5	0	0	0
Bayer Diastix		0	0	1	0	0	0	0
Other		0	0	8	2	0	0	0
All Instruments and Reagents		2	16	307	108	1	0	21

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Protein	Specimen: U95	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		59	0	0	0	0	0	1
Siemens Atlas		91	0	0	1	1	0	0
Siemens Clinitek Status		40	0	0	0	0	0	0
Siemens Clinitek 50		8	0	0	0	0	0	0
Siemens Clinitek 100		9	0	0	0	0	0	0
Siemens Clinitek 500		56	0	0	0	0	0	0
Siemens Clinitek Advantus		40	0	0	0	0	0	0
Roche UA Reader/Super UA Read		2	1	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		36	0	0	0	0	0	0
IRIS AUTION MAX		37	0	0	0	1	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
Other		4	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		183	0	0	0	0	0	0
Siemens Atlas Reagent		83	0	0	1	1	0	0
Sulfosalicylic Acid		15	0	0	0	0	0	1
Roche Chemstrips		46	1	0	0	0	0	0
Roche Urisys Cassette		15	0	0	0	0	0	0
ARKRAY Aution Sticks		28	0	0	0	1	0	0
IRIS Reagent Strips		10	0	0	0	0	0	0
Bayer Clinitest		8	0	0	0	0	0	0
Other		10	0	0	0	0	0	0
All Instruments and Reagents		398	1	0	1	2	0	1

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Qualitative hCG report:

Bayer Clinitek Status	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	2	9	0	9	9
Indeterminant	0	0	0	0	0
Positive	7	0	9	0	0

Beckman Coulter ICON II	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	0	2	0	2	2
Indeterminant	0	0	0	0	0
Positive	2	0	2	0	0

Beckman Coulter ICON 25	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	1	25	0	25	25
Indeterminant	0	0	0	0	0
Positive	24	0	25	0	0

Beckman Coulter ICON 20	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	3	23	0	23	21
Indeterminant	1	0	0	0	0
Positive	19	0	23	0	2

Cardinal Health hCG Combo	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	0	14	0	14	14
Indeterminant	1	0	0	0	0
Positive	13	0	14	0	0

Wampole Clearview hCG Combo	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	1	3	0	3	3
Indeterminant	0	0	0	0	0
Positive	2	0	3	0	0

Cen Med Elite	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	4	19	0	19	19
Indeterminant	1	0	0	0	0
Positive	14	0	19	0	0

Genzyme Contrast hCG	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	0	1	0	1	1
Indeterminant	0	0	0	0	0
Positive	1	0	1	0	0

**New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011**

Wampole Clearview hCG II	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	1	4	0	4	4
Indeterminant	0	0	0	0	0
Positive	3	0	4	0	0
Genzyme OSOM Combo	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	5	17	0	17	17
Indeterminant	0	0	0	0	0
Positive	12	0	17	0	0
Inverness Medical BioStar Acceava	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	0	3	0	3	3
Indeterminant	0	0	0	0	0
Positive	3	0	3	0	0
Lifesign Status (serum/urine)	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	0	1	0	1	1
Indeterminant	0	0	0	0	0
Positive	1	0	1	0	0
Polymedco Poly stat	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	0	6	0	6	6
Indeterminant	0	0	0	0	0
Positive	6	0	6	0	0
Quidel QuickVue	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	26	72	0	72	72
Indeterminant	2	0	0	0	0
Positive	44	0	72	0	0
SA Scientific	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	2	2	0	2	2
Indeterminant	0	0	0	0	0
Positive	0	0	2	0	0
Stanbio QuPID	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	4	20	0	20	20
Indeterminant	0	0	0	0	0
Positive	16	0	20	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011

Fisher Sure-Vue serum/urine	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	3	43	0	42	42
Indeterminant	2	0	0	0	0
Positive	38	0	43	1	1
Fisher Sure-Vue STAT serum/urine	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	4	27	0	28	28
Indeterminant	5	0	0	0	0
Positive	19	1	28	0	0
Signify hCG Urine/Serum	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	0	5	0	5	5
Indeterminant	0	0	0	0	0
Positive	5	0	5	0	0
Sacks Medical RefuAH	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	0	2	0	2	2
Indeterminant	0	0	0	0	0
Positive	2	0	2	0	0
Fisher Sure-Vue (urine only)	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	3	7	0	7	7
Indeterminant	0	0	0	0	0
Positive	4	0	7	0	0
True Medix	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	1	4	0	4	4
Indeterminant	0	0	0	0	0
Positive	3	0	4	0	0
Method not listed	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	2	28	1	28	28
Indeterminant	0	0	0	0	0
Positive	26	0	27	0	0
All Methods & Instruments	Specimen:U91	Specimen:U92	Specimen:U93	Specimen:U94	Specimen:U95
Negative	62	337	1	337	335
Indeterminant	12	0	0	0	0
Positive	264	1	337	1	3

**New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 7 November 2011**

hCG Summary:

[Code], hCG Kit name	U91			U92			U93			U94			U95		
	N	I	P	N	I	P	N	I	P	N	I	P	N	I	P
[BA] Bayer Clinitek Status	2	0	7	9	0	0	0	0	9	9	0	0	9	0	0
[BC] Beckman Coulter ICON II	0	0	2	2	0	0	0	0	2	2	0	0	2	0	0
[BN] Beckman Coulter ICON 25	1	0	24	25	0	0	0	0	25	25	0	0	25	0	0
[BT] Beckman Coulter ICON 20	3	1	19	23	0	0	0	0	23	23	0	0	21	0	2
[CA] Cardinal Health hCG Combo	0	1	13	14	0	0	0	0	14	14	0	0	14	0	0
[CC] Wampole Clearview hCG Combo	1	0	2	3	0	0	0	0	3	3	0	0	3	0	0
[CM] Cen Med Elite	4	1	14	19	0	0	0	0	19	19	0	0	19	0	0
[CT] Genzyme Contrast hCG	0	0	1	1	0	0	0	0	1	1	0	0	1	0	0
[CV] Wampole Clearview hCG II	1	0	3	4	0	0	0	0	4	4	0	0	4	0	0
[GO] Genzyme OSOM Combo	5	0	12	17	0	0	0	0	17	17	0	0	17	0	0
[IM] Inverness Medical BioStar Acceava	0	0	3	3	0	0	0	0	3	3	0	0	3	0	0
[LS] Lifesign Status (serum/urine)	0	0	1	1	0	0	0	0	1	1	0	0	1	0	0
[PM] Polymedco Poly stat	0	0	6	6	0	0	0	0	6	6	0	0	6	0	0
[QQ] Quidel QuickVue	26	2	44	72	0	0	0	0	72	72	0	0	72	0	0
[SA] SA Scientific	2	0	0	2	0	0	0	0	2	2	0	0	2	0	0
[SB] Stanbio QuPID	4	0	16	20	0	0	0	0	20	20	0	0	20	0	0
[SE] Fisher Sure-Vue serum/urine	3	2	38	43	0	0	0	0	43	42	0	1	42	0	1
[SF] Fisher Sure-Vue STAT serum/urine	4	5	19	27	0	1	0	0	28	28	0	0	28	0	0
[SG] Signify hCG Urine/Serum	0	0	5	5	0	0	0	0	5	5	0	0	5	0	0
[SM] Sacks Medical RefuAH	0	0	2	2	0	0	0	0	2	2	0	0	2	0	0
[SU] Fisher Sure-Vue (urine only)	3	0	4	7	0	0	0	0	7	7	0	0	7	0	0
[TX] True Medix	1	0	3	4	0	0	0	0	4	4	0	0	4	0	0
[90] Method not listed	2	0	26	28	0	0	1	0	27	28	0	0	28	0	0
[99] All hCG Methods	62	12	264	337	0	1	1	0	337	337	0	1	335	0	3