

**Participant Performance Summary - Initial (Screen) Test**  
**Clinical Toxicology - August 2016 Test Event**

Amphetamines Screen (ng/mL)		UTD 1			UTD 2			UTD 3			UTD 4		
Method Code	Cutoff	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%	Response	No. labs	No. Resp.	%
LCMM - LC MS/MS	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Siemens Syva Emit / II Plus	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
ABAR - Abbott Architect	1000	Positive	8	8	100	Negative	8	8	100	Negative	8	8	100
ALTM - Alere (Biosite) Triage Meter/PRO	1000	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100
Alere (BioSite) Triage Plus TCA	1000	Positive	11	11	100	Negative	11	11	100	Negative	11	11	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	1000	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	1000	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	1000	Positive	19	19	100	Negative	19	19	100	Negative	19	19	100
BMC QuickTox	1000	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BRTS - BioRad Tox/See	1000	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	1000	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100
Diatron Pictus	1000	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
ESCR - E-Screen	1000	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
ORVI - Ortho Vitros	1000	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100
ROCO - Roche Cobas	1000	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100
ROKI - Roche KIMS	1000	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100
SIAD - Siemens ADVIA	1000	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100
SIDI - Siemens Dimension	1000	Positive	23	23	100	Negative	23	23	100	Negative	23	23	100
Siemens Syva Emit / II Plus	1000	Positive	8	8	100	Negative	8	8	100	Negative	8	8	100
THSD - Thermo Scientific / Microgenics DRI	1000	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100
Other	1000	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	371.5	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
ABAR - Abbott Architect	500	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	500	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100
Immunalysis HEIA	500	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	500	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100
ROCO - Roche Cobas	500	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100
ROKI - Roche KIMS	500	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100
SIDI - Siemens Dimension	500	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100
Siemens Syva Emit / II Plus	500	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	500	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100
All Labs	All Labs	Negative	142	1	0.7	Negative	142	142	100	Negative	142	142	100
		Positive	142	141	99								

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Barbiturates Screen (ng/mL)</b>		<b>UTD 1</b>			<b>UTD 2</b>			<b>UTD 3</b>			<b>UTD 4</b>		
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>
Siemens Syva Emit / II Plus	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
ABAR - Abbott Architect	200	Negative	9	9	100	Negative	9	9	100	Negative	9	9	100
BECO - Beckman Coulter	200	Negative	22	22	100	Negative	22	22	100	Negative	22	22	100
CALC - Carolina Liquid Chemistries	200	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
Diatron Pictus	200	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	200	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	200	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100
ORVI - Ortho Vitros	200	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100
ROCO - Roche Cobas	200	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100
ROKI - Roche KIMS	200	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100
SIAD - Siemens ADVIA	200	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
SIDI - Siemens Dimension	200	Negative	30	30	100	Negative	30	30	100	Negative	30	30	100
Siemens Syva Emit / II Plus	200	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100
THSD - Thermo Scientific / Microgenics DRI	200	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100
ALTM - Alere (Biosite) Triage Meter/PRO	300	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
Alere (BioSite) Triage Plus TCA	300	Negative	11	11	100	Negative	11	11	100	Negative	11	11	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	300	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	300	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
BMC QuickTox	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BRTS - BioRad Tox/See	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
SIAD - Siemens ADVIA	300	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
Siemens Syva Emit / II Plus	300	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
THSC - Thermo Scientific / Microgenics CEDIA	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	391.5	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
All Labs	All Labs	Negative	142	142	100	Negative	142	142	100	Negative	142	142	100

**Participant Performance Summary - Initial (Screen) Test**  
**Clinical Toxicology - August 2016 Test Event**

<b>Benzodiazepines Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
LCMM - LC MS/MS	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ROCO - Roche Cobas	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ROKI - Roche KIMS	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Siemens Syva Emit / II Plus	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
THSC - Thermo Scientific / Microgenics CEDIA	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	150	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100	Negative	6	6	100
ABAR - Abbott Architect	200	Negative	8	8	100	Negative	8	8	100	Positive	8	8	100	Negative	8	8	100
BECO - Beckman Coulter	200	Negative	19	19	100	Negative	19	19	100	Positive	19	19	100	Negative	19	19	100
CALC - Carolina Liquid Chemistries	200	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
Diatron Pictus	200	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Immunalysis HEIA	200	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ORVI - Ortho Vitros	200	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
ROCO - Roche Cobas	200	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100
ROKI - Roche KIMS	200	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
SIAD - Siemens ADVIA	200	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
SIDI - Siemens Dimension	200	Negative	30	30	100	Negative	30	30	100	Positive	30	30	100	Negative	30	30	100
Siemens Syva Emit / II Plus	200	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100	Negative	7	7	100
THSD - Thermo Scientific / Microgenics DRI	200	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
Other	200	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ABAR - Abbott Architect	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ALTM - Alere (Biosite) Triage Meter/PRO	300	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
Alere (BioSite) Triage Plus TCA	300	Negative	11	11	100	Negative	11	11	100	Positive	11	11	100	Negative	11	11	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	300	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100	Negative	7	7	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	300	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100	Negative	7	7	100
BMC QuickTox	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
BRTS - BioRad Tox/See	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ORVI - Ortho Vitros	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ROCO - Roche Cobas	300	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
ROKI - Roche KIMS	300	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
SIAD - Siemens ADVIA	300	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
Siemens Syva Emit / II Plus	300	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
THSC - Thermo Scientific / Microgenics CEDIA	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	348.9	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
All Labs	All Labs	Negative	143	143	100	Negative	143	143	100	Positive	143	143	100	Negative	143	143	100

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Cocaine/metab Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
LCMM - LC MS/MS	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Siemens Syva Emit / II Plus	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
ABAR - Abbott Architect	150	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100
BECO - Beckman Coulter	150	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
Immunalysis HEIA	150	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	150	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100
ROCO - Roche Cobas	150	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	150	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
THSD - Thermo Scientific / Microgenics DRI	150	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
ABAR - Abbott Architect	300	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
ALTM - Alere (Biosite) Triage Meter/PRO	300	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
Alere (BioSite) Triage Plus TCA	300	Positive	10	10	100	Negative	10	10	100	Negative	10	10	100	Negative	10	10	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	300	Positive	8	8	100	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	300	Positive	22	22	100	Negative	22	22	100	Negative	22	22	100	Negative	22	22	100
BMC QuickTox	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BRTS - BioRad Tox/See	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	300	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
Diatron Pictus	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
ESCR - E-Screen	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
ORVI - Ortho Vitros	300	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100
ROCO - Roche Cobas	300	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100
ROKI - Roche KIMS	300	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100
SIAD - Siemens ADVIA	300	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
SIDI - Siemens Dimension	300	Positive	26	26	100	Negative	26	26	100	Negative	26	26	100	Negative	26	26	100
Siemens Syva Emit / II Plus	300	Positive	8	8	100	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100
THSD - Thermo Scientific / Microgenics DRI	300	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100
Other	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	400.9	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
All Labs	All Labs	Positive	144	144	100	Negative	144	144	100	Negative	144	144	100	Negative	144	144	100

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Ethyl glucuronide Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
Immunalysis HEIA	250	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	250	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	500	Negative	4	4	100	Negative	4	1	25	Negative	4	4	100	Negative	4	4	100
	500					Positive	4	3	75								
CALC - Carolina Liquid Chemistries	500	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	500	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100
SIAD - Siemens ADVIA	500	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	500	Negative	4	4	100	Negative	4	1	25	Negative	4	4	100	Negative	4	4	100
	500					Positive	4	3	75								
All Labs	All Labs	Negative	14	14	100	Negative	14	2	14	Negative	14	14	100	Negative	14	14	100
						Positive	14	12	86								

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Ethanol Screen (mg/dL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
ABAR - Abbott Architect	10	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
Alcohol Dehydrogenases/UV or Visible Spectrom	10	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	10	Negative	7	7	100	Positive	7	7	100	Positive	7	7	100	Negative	7	7	100
Diatron Pictus	10	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
ORVI - Ortho Vitros	10	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
ROCO - Roche Cobas	10	Negative	2	2	100	Positive	2	2	100	Positive	2	2	100	Negative	2	2	100
SIAD - Siemens ADVIA	10	Negative	2	2	100	Positive	2	2	100	Positive	2	2	100	Negative	2	2	100
SIDI - Siemens Dimension	10	Negative	2	2	100	Positive	2	2	100	Positive	2	2	100	Negative	2	2	100
Siemens Syva Emit / II Plus	10	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
Siemens Syva Emit / II Plus	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
Alcohol Dehydrogenases/UV or Visible Spectrom	20	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	20	Negative	4	4	100	Positive	4	4	100	Positive	4	4	100	Negative	4	4	100
SIAD - Siemens ADVIA	20	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	20	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
Siemens Syva Emit / II Plus	20	Negative	2	2	100	Positive	2	2	100	Positive	2	2	100	Negative	2	2	100
THSD - Thermo Scientific / Microgenics DRI	20	Negative	5	5	100	Positive	5	5	100	Positive	5	5	100	Negative	5	5	100
BECO - Beckman Coulter	40	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
SIAD - Siemens ADVIA	40	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	5	Negative	1	1	100	Positive	1	1	100	Positive	1	1	100	Negative	1	1	100
Alcohol Dehydrogenases/UV or Visible Spectrom	50	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	50	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ROCO - Roche Cobas	50	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	50	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100
All Labs	All Labs	Negative	42	42	100	Negative	42	7	17	Positive	42	42	100	Negative	42	42	100
						Positive	42	35	83								

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Fentanyl Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
IMEL - Immunalysis ELISA	1	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
LCMM - LC MS/MS	10	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	2	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	2	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	2	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100
Other	2	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	2000	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
LCMM - LC MS/MS	5	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
All Labs	All Labs	Negative	12	12	100	Negative	12	12	100	Negative	12	12	100	Negative	12	12	100

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>K2/Spice Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
LCMM - LC MS/MS	1	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
Immunalysis HEIA	10	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	20	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	20	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100
LCMM - LC MS/MS	20	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	5	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
All Labs	All Labs	Negative	9	9	100	Negative	9	9	100	Negative	9	9	100	Negative	9	9	100



**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>MDMA Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
LCMM - LC MS/MS	100	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
ALTM - Alere (Biosite) Triage Meter/PRO	1000	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	300	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	500	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100
BRTS - BioRad Tox/See	500	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	500	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	500	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
SIAD - Siemens ADVIA	500	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	500	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Siemens Syva Emit / II Plus	500	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
THSD - Thermo Scientific / Microgenics DRI	500	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100
All Labs	All Labs	Positive	20	20	100	Negative	20	20	100	Negative	20	20	100	Negative	20	20	100

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Methodadone Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
LCMM - LC MS/MS	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Siemens Syva Emit / II Plus	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	20	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	200	Negative	5	5	100	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100
ABAR - Abbott Architect	300	Negative	6	6	100	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100
ALTM - Alere (Biosite) Triage Meter/PRO	300	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	300	Negative	5	5	100	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	300	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	300	Negative	20	20	100	Positive	20	20	100	Negative	20	20	100	Negative	20	20	100
CALC - Carolina Liquid Chemistries	300	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Diatron Pictus	300	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	300	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
ORVI - Ortho Vitros	300	Negative	5	5	100	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100
ROCO - Roche Cobas	300	Negative	6	6	100	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100
ROKI - Roche KIMS	300	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100
SIAD - Siemens ADVIA	300	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100	Negative	2	2	100
SIDI - Siemens Dimension	300	Negative	22	22	100	Positive	22	22	100	Negative	22	22	100	Negative	22	22	100
Siemens Syva Emit / II Plus	300	Negative	7	7	100	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100
THSD - Thermo Scientific / Microgenics DRI	300	Negative	6	6	100	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100
Other	300	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	434.3	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
All Labs	All Labs	Negative	99	99	100	Positive	99	99	100	Negative	99	99	100	Negative	99	99	100

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Opiates Screen (ng/mL)</b>		<b>UTD 1</b>			<b>UTD 2</b>			<b>UTD 3</b>			<b>UTD 4</b>						
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>				
LCMM - LC MS/MS	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
MEPR - MedTox Profile	100	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
Siemens Syva Emit / II Plus	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	2000	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
BECO - Beckman Coulter	2000	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100
ESCR - E-Screen	2000	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Immunalysis HEIA	2000	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
SIAD - Siemens ADVIA	2000	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
SIDI - Siemens Dimension	2000	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Siemens Syva Emit / II Plus	2000	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
ABAR - Abbott Architect	300	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100	Positive	8	8	100
ALTM - Alere (Biosite) Triage Meter/PRO	300	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100
Alere (BioSite) Triage Plus TCA	300	Negative	11	11	100	Negative	11	11	100	Negative	11	11	100	Positive	11	11	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	300	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
BECO - Beckman Coulter	300	Negative	23	23	100	Negative	23	23	100	Negative	23	23	100	Positive	23	23	100
BMC QuickTox	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BRTS - BioRad Tox/See	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
CALC - Carolina Liquid Chemistries	300	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
Diatron Pictus	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
ORVI - Ortho Vitros	300	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Positive	5	5	100
ROCO - Roche Cobas	300	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100	Positive	8	8	100
ROKI - Roche KIMS	300	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100
SIAD - Siemens ADVIA	300	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100
SIDI - Siemens Dimension	300	Negative	28	28	100	Negative	28	28	100	Negative	28	28	100	Positive	28	28	100
Siemens Syva Emit / II Plus	300	Negative	7	7	100	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100
THSD - Thermo Scientific / Microgenics DRI	300	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
Other	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
BECO - Beckman Coulter	387.5	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
All Labs	All Labs	Negative	144	144	100	Negative	144	144	100	Negative	144	144	100	Negative	144	1	0.69
		Positive												Positive	144	143	99.3

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Oxycodone Screen (ng/mL)</b>		<b>UTD 1</b>			<b>UTD 2</b>			<b>UTD 3</b>			<b>UTD 4</b>		
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>
ABAR - Abbott Architect	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	100	Negative	4	4	100	Negative	4	4	100	Negative	4	4	100
BRTS - BioRad Tox/See	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100
CALC - Carolina Liquid Chemistries	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Innovacon Oxycodone Single Drug Test	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
LCMM - LC MS/MS	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	100	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100
ORVI - Ortho Vitros	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
ROCO - Roche Cobas	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100
ROKI - Roche KIMS	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
SIAD - Siemens ADVIA	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	100	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100
Other	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	300	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
CALC - Carolina Liquid Chemistries	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	300	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
All Labs	All Labs	Negative	44	44	100	Negative	44	44	100	Negative	44	44	100

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Phencyclidine (PCP) Screen (ng/mL)</b>		<b>UTD 1</b>			<b>UTD 2</b>			<b>UTD 3</b>			<b>UTD 4</b>		
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
Siemens Syva Emit / II Plus	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
ABAR - Abbott Architect	25	Negative	8	8	100	Negative	8	8	100	Positive	8	8	100
ALTM - Alere (Biosite) Triage Meter/PRO	25	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100
Alere (BioSite) Triage Plus TCA	25	Negative	11	11	100	Negative	11	11	100	Positive	11	11	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	25	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	25	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
BECO - Beckman Coulter	25	Negative	24	24	100	Negative	24	24	100	Positive	24	24	100
BMC QuickTox	25	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
BRTS - BioRad Tox/See	25	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
CALC - Carolina Liquid Chemistries	25	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Diatron Pictus	25	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
ESCR - E-Screen	25	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
Immunalysis HEIA	25	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100
LCMM - LC MS/MS	25	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
MEPR - MedTox Profile	25	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
ORVI - Ortho Vitros	25	Negative	5	5	100	Negative	5	5	100	Positive	5	5	100
ROCO - Roche Cobas	25	Negative	8	8	100	Negative	8	8	100	Positive	8	8	100
ROKI - Roche KIMS	25	Negative	7	7	100	Negative	7	7	100	Positive	7	7	100
SIAD - Siemens ADVIA	25	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100
SIDI - Siemens Dimension	25	Negative	27	27	100	Negative	27	27	100	Positive	27	27	100
Siemens Syva Emit / II Plus	25	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
THSD - Thermo Scientific / Microgenics DRI	25	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100
BECO - Beckman Coulter	479.2	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100
All Labs	All Labs	Negative	131	131	100	Negative	131	131	100	Positive	131	131	100

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Propoxyphene Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
Siemens Syva Emit / II Plus	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
ABAR - Abbott Architect	300	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
BECO - Beckman Coulter	300	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100	Negative	8	8	100
Diatron Pictus	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	300	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100	Negative	5	5	100
ORVI - Ortho Vitros	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
ROCO - Roche Cobas	300	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
ROKI - Roche KIMS	300	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
SIAD - Siemens ADVIA	300	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100	Negative	2	2	100
Siemens Syva Emit / II Plus	300	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100	Negative	3	3	100
THSD - Thermo Scientific / Microgenics DRI	300	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100	Negative	6	6	100
All Labs	All Labs	Negative	32	32	100	Negative	32	32	100	Negative	32	32	100	Negative	32	32	100

**Participant Performance Summary - Initial (Screen) Test**  
**Clinical Toxicology - August 2016 Test Event**

<b>Cannabinoids (THCA) Screen (ng/mL)</b>		<b>UTD 1</b>			<b>UTD 2</b>			<b>UTD 3</b>			<b>UTD 4</b>		
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp. %</b>
Siemens Syva Emit / II Plus	100	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	20	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Siemens Syva Emit / II Plus	20	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	20	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
SIDI - Siemens Dimension	25	Positive	3	3	100	Negative	3	3	100	Negative	3	3	100
THSC - Thermo Scientific / Microgenics CEDIA	25	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Other	25	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	379.7	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
ABAR - Abbott Architect	50	Positive	9	9	100	Negative	9	9	100	Negative	9	9	100
ALTM - Alere (Biosite) Triage Meter/PRO	50	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100
Alere (BioSite) Triage Plus TCA	50	Positive	11	11	100	Negative	11	11	100	Negative	11	11	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	50	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	50	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	50	Positive	25	25	100	Negative	25	25	100	Negative	25	25	100
BMC QuickTox	50	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
BRTS - BioRad Tox/See	50	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	50	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Diatron Pictus	50	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
ESCR - E-Screen	50	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
Immunalysis HEIA	50	Positive	1	1	100	Negative	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	50	Positive	6	6	100	Negative	6	6	100	Negative	6	6	100
ORVI - Ortho Vitros	50	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100
ROCO - Roche Cobas	50	Positive	8	8	100	Negative	8	8	100	Negative	8	8	100
ROKI - Roche KIMS	50	Positive	7	7	100	Negative	7	7	100	Negative	7	7	100
SIAD - Siemens ADVIA	50	Positive	4	4	100	Negative	4	4	100	Negative	4	4	100
SIDI - Siemens Dimension	50	Positive	27	27	100	Negative	27	27	100	Negative	27	27	100
Siemens Syva Emit / II Plus	50	Positive	8	8	100	Negative	8	8	100	Negative	8	8	100
THSD - Thermo Scientific / Microgenics DRI	50	Positive	5	5	100	Negative	5	5	100	Negative	5	5	100
All Labs	All Labs	Positive	143	143	100	Negative	143	143	100	Negative	143	143	100

**Participant Performance Summary - Initial (Screen) Test  
Clinical Toxicology - August 2016 Test Event**

<b>Tricyclic Antidepressants Screen (ng/mL)</b>		<b>UTD 1</b>				<b>UTD 2</b>				<b>UTD 3</b>				<b>UTD 4</b>			
<b>Method Code</b>	<b>Cutoff</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>	<b>Response</b>	<b>No. labs</b>	<b>No. Resp.</b>	<b>%</b>
IMEL - Immunalysis ELISA	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100	Negative	1	1	100
LCMM - LC MS/MS	100	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ALTM - Alere (Biosite) Triage Meter/PRO	1000	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
Alere (BioSite) Triage Plus TCA	1000	Negative	10	10	100	Negative	10	10	100	Positive	10	10	100	Negative	10	10	100
ALTT - Alere (Biosite) Triage TOX Drug Screen	1000	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100	Negative	6	6	100
AMIM - Ameditech ImmuTest Multi-Drug Screen	1000	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
BRTS - BioRad Tox/See	1000	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
Siemens Syva Rapid Test	1000	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100
DIPR - Diagnostic Products	110	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
ABAR - Abbott Architect	300	Negative	3	3	100	Negative	3	3	100	Positive	3	3	100	Negative	3	3	100
Alere (BioSite) Triage Plus TCA	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
BECO - Beckman Coulter	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
CALC - Carolina Liquid Chemistries	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
MEPR - MedTox Profile	300	Negative	6	6	100	Negative	6	6	100	Positive	6	6	100	Negative	6	6	100
ORVI - Ortho Vitros	300	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
THSD - Thermo Scientific / Microgenics DRI	300	Negative	4	4	100	Negative	4	4	100	Positive	4	4	100	Negative	4	4	100
BECO - Beckman Coulter	500	Negative	2	2	100	Negative	2	2	100	Positive	2	2	100	Negative	2	2	100
ROCO - Roche Cobas	500	Negative	1	1	100	Negative	1	1	100	Positive	1	1	100	Negative	1	1	100
All Labs	All Labs	Negative	49	49	100	Negative	49	49	100	Negative	49	1	2	Negative	49	49	100
										Positive	49	48	98				