

Bacteriology – Restricted
Group A *Streptococcus* Antigen Detection
September 2015

Number of participating laboratories: 82

Grade distribution

100%	77
80-89%	5

Correct responses:

A – Negative
B – Negative
C – Positive
D – Negative
E – Positive

Method	# Labs Correct Samples B, C, E	# Labs Correct (Incorrect) Sample:	
		A	D
Acceava Strep A	21	19 (2)	21
Sekisui (Genzyme) OSOM Ultra Strep A	13	13	12 (1)
Quidel QuickVue + Strep A	10	10	9 (1)
Sekisui (Genzyme) OSOM Strep A Test	6	6	6
Fisher Sure-Vue Signature Strep A Test	5	5	5
Meridian Bioscience ImmunoCard STAT Strep A	4	4	3 (1)
Cardinal Health SP Brand Strep A Dipstick	4	4	4
BD Chek Group A Strep	4	4	4
Henry Schein One Step + Strep A Dipstick	3	3	3
Abbott Signify Strep A Dipstick	3	3	3
Beckman Coulter Icon DS Strep A	2	2	2
Quidel QuickVue Dipstick Strep A	2	2	2
Clearview Exact Strep A Dipstick	1	1	1
CONSULT Strep A Test	1	1	1
Quidel QuickVue Inline Strep A	1	1	1
Fisher Sure-Vue Strep A Lateral Flow Test	1	1	1
ProAdvantage by NDC	1	1	1