

**Bacteriology - Restricted
Throat Culture Results
September 2015**

Number of participating laboratories: 17

Grade distribution

100%	14
80-89%	3

Correct Responses:

- A – Positive (*Streptococcus pyogenes* & *Streptococcus mitis*)
- B – Positive (*Streptococcus pyogenes* & *Staphylococcus epidermidis*)
- C – Negative (*Staphylococcus aureus* & *Neisseria mucosa*)
- D – Negative (*Corynebacterium pseudiptheriticum* & *Streptococcus bovis*)
- E – Negative (*Streptococcus agalactiae* & *Citrobacter freundii*)

Methods used by participating laboratories				
		# Labs Correct (Incorrect) Sample:		
Method	# Labs Correct Samples A, D	B	C	E
Bacitracin disk only	7	7	7	7
Selective Strep Agar	5	4 (1)	4 (1)	5
Remel PathoDx	2	2	2	1 (1)
Bacitracin disk and Remel PathoDx	2	2	2	2
BD BBL Streptocard	1	1	1	1