



## Department of Health

ANDREW M. CUOMO  
Governor

HOWARD A. ZUCKER, M.D., J.D.  
Commissioner

SALLY DRESLIN, M.S., R.N.  
Executive Deputy Commissioner

### **Statistical Report Urinalysis/Urine Pregnancy Testing Mail out date: April 13, 2015**

This report summarizes data from the urinalysis/urine pregnancy proficiency testing of April 13, 2015.

If you have any questions regarding this report, please contact Regina Saulsbery at 518-474-6156.  
You may also contact us by e-mail: [clinchem@wadsworth.org](mailto:clinchem@wadsworth.org)

Results are shown for each analyte arranged by instrument and reagent system. Results for specific gravity and pH are listed as the mean  $\pm$  1SD for each of the instruments and reagents. All other results are shown as the number of participant results for that instrument or reagent.

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

**PLEASE NOTE:** The evaluation report and summary of the educational quantitative urine clinical chemistry proficiency test will be separate if your laboratory participated in that proficiency test.

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Specific Gravity

Specimen: U31	Specimen: U32	Specimen: U33	Specimen: U34	Specimen: U35	Number	Instrument or Reagent System
1.017 ± 0.006	1.017 ± 0.005	1.012 ± 0.004	1.021 ± 0.008	1.013 ± 0.003	n = 358	All Methods & Instruments
1.015	1.016	1.011	1.020	1.013	n = 358	Median value, All Methods & Instruments
						<Instruments>
1.019 ± 0.004	1.014 ± 0.002	1.008 ± 0.002	1.026 ± 0.005	1.012 ± 0.003	n = 14	Visual examination
1.013 ± 0.002	1.016 ± 0.002	1.011 ± 0.001	1.015 ± 0.002	1.013 ± 0.002	n = 96	Siemens Atlas
1.024 ± 0.003	1.025 ± 0.002	1.018 ± 0.003	1.029 ± 0.003	1.016 ± 0.002	n = 51	Siemens Clinitek Status
1.021 ± 0.002	1.020 ± 0.000	1.012 ± 0.003	1.030 ± 0.000	1.011 ± 0.002	n = 5	Siemens Clinitek 50
1.023 ± 0.002	1.020 ± 0.001	1.015 ± 0.003	1.029 ± 0.002	1.013 ± 0.003	n = 25	Siemens Clinitek 500
1.022 ± 0.002	1.021 ± 0.002	1.013 ± 0.003	1.029 ± 0.003	1.013 ± 0.003	n = 66	Siemens Clinitek Advantus
1.010 ± 0.000	1.010 ± 0.000	1.012 ± 0.003	1.020 ± 0.000	1.012 ± 0.003	n = 3	Roche Chemstrip Criterion
1.014 ± 0.002	1.013 ± 0.003	1.011 ± 0.002	1.018 ± 0.003	1.013 ± 0.002	n = 26	Roche Urisys
1.015 ± 0.003	1.013 ± 0.003	1.011 ± 0.004	1.021 ± 0.002	1.013 ± 0.003	n = 6	Roche Cobas U411
1.011 ± 0.001	1.011 ± 0.001	1.007 ± 0.001	1.013 ± 0.001	1.008 ± 0.002	n = 23	IRIS AUTION MAX
1.012 ± 0.001	1.011 ± 0.001	1.007 ± 0.000	1.014 ± 0.001	1.008 ± 0.001	n = 27	IRIS iChem Velocity
1.013 ± 0.001	1.015 ± 0.001	1.010 ± 0.001	1.014 ± 0.001	1.011 ± 0.001	n = 13	Refractometer
						<Reagents>
1.022 ± 0.004	1.021 ± 0.004	1.014 ± 0.004	1.028 ± 0.005	1.014 ± 0.003	n = 169	Siemens Reagent Strips
1.013 ± 0.002	1.016 ± 0.002	1.011 ± 0.001	1.015 ± 0.002	1.013 ± 0.002	n = 87	Siemens Atlas Reagent
1.015 ± 0.002	1.012 ± 0.002	1.011 ± 0.003	1.021 ± 0.001	1.013 ± 0.002	n = 27	Roche Chemstrips
1.013 ± 0.001	1.015 ± 0.002	1.010 ± 0.001	1.015 ± 0.002	1.012 ± 0.001	n = 10	Roche Urisys Cassette
1.011 ± 0.001	1.011 ± 0.002	1.007 ± 0.001	1.013 ± 0.001	1.009 ± 0.002	n = 16	ARKRAY Aution Sticks
1.012 ± 0.001	1.011 ± 0.002	1.007 ± 0.001	1.014 ± 0.002	1.008 ± 0.001	n = 30	IRIS Reagent Strips
1.013 ± 0.002	1.014 ± 0.002	1.009 ± 0.001	1.014 ± 0.003	1.010 ± 0.002	n = 18	None, no reagent or strip used

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

pH

Specimen: U31	Specimen: U32	Specimen: U33	Specimen: U34	Specimen: U35	Number	Instrument or Reagent System
6.37 ± 0.25 6.50	7.21 ± 0.25 7.00	7.68 ± 0.28 7.50	5.99 ± 0.15 6.00	6.78 ± 0.27 7.00	n = 361 n = 361	All Methods & Instruments Median value, All Methods & Instruments
<Instruments>						
6.37 ± 0.23	7.32 ± 0.30	7.68 ± 0.30	5.97 ± 0.31	6.82 ± 0.25	n = 19	Visual examination
6.49 ± 0.05	7.25 ± 0.25	7.53 ± 0.12	6.00 ± 0.00	6.80 ± 0.25	n = 96	Siemens Atlas
6.51 ± 0.22	7.18 ± 0.24	7.71 ± 0.28	5.99 ± 0.12	6.96 ± 0.17	n = 56	Siemens Clinitek Status
6.40 ± 0.22	7.10 ± 0.22	7.50 ± 0.00	6.00 ± 0.00	6.40 ± 0.22	n = 5	Siemens Clinitek 50
6.27 ± 0.25	7.35 ± 0.24	7.71 ± 0.25	6.00 ± 0.00	6.71 ± 0.29	n = 26	Siemens Clinitek 500
6.34 ± 0.24	7.37 ± 0.22	7.67 ± 0.32	5.99 ± 0.06	6.69 ± 0.30	n = 68	Siemens Clinitek Advantus
6.50 ± 0.41	7.00 ± 0.00	8.00 ± 0.00	6.00 ± 0.00	6.75 ± 0.29	n = 4	Roche Chemstrip Criterion
6.23 ± 0.25	7.00 ± 0.00	7.92 ± 0.27	5.85 ± 0.37	6.75 ± 0.25	n = 26	Roche Urisys
6.25 ± 0.27	7.00 ± 0.00	7.50 ± 0.55	6.00 ± 0.00	6.50 ± 0.00	n = 6	Roche Cobas U411
6.26 ± 0.26	7.00 ± 0.00	7.54 ± 0.14	6.00 ± 0.00	6.50 ± 0.21	n = 23	IRIS AUTION MAX
6.00 ± 0.00	7.00 ± 0.00	8.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	n = 27	IRIS iChem Velocity
<Reagents>						
6.40 ± 0.24	7.29 ± 0.25	7.67 ± 0.29	5.99 ± 0.12	6.78 ± 0.28	n = 183	Siemens Reagent Strips
6.49 ± 0.05	7.26 ± 0.25	7.54 ± 0.13	6.00 ± 0.00	6.81 ± 0.24	n = 85	Siemens Atlas Reagent
6.18 ± 0.28	7.00 ± 0.00	7.80 ± 0.39	5.88 ± 0.32	6.63 ± 0.22	n = 28	Roche Chemstrips
6.45 ± 0.16	7.00 ± 0.00	8.00 ± 0.00	5.90 ± 0.32	6.95 ± 0.16	n = 10	Roche Urisys Cassette
6.24 ± 0.26	7.00 ± 0.00	7.55 ± 0.16	6.00 ± 0.00	6.53 ± 0.20	n = 19	ARKRAY Aution Sticks
6.05 ± 0.15	7.03 ± 0.18	7.94 ± 0.17	6.00 ± 0.00	6.92 ± 0.22	n = 32	IRIS Reagent Strips

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

<b>Bilirubin</b>	<b>Specimen: U31</b>	Negative	Small   1+   1mg/dL	Moderate   2+   3mg/dL	Large   3+   6mg/dL	Positive
<Instruments>						
Visual examination		26	0	0	0	0
Siemens Atlas		96	0	0	0	0
Siemens Clinitek Status		55	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0
Siemens Clinitek 500		26	0	0	0	0
Siemens Clinitek Advantus		67	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0
Roche Urisys		25	0	0	0	0
Roche Cobas U411		6	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0
Other Chemistry Analyzer		0	0	0	0	0
Other		1	0	0	0	0
<Reagents>						
Siemens Reagent Strips		178	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0
Roche Chemstrips		28	0	0	0	0
Roche Urisys Cassette		9	0	0	0	0
ARKRAY Aution Sticks		18	0	0	0	0
IRIS Reagent Strips		34	0	0	0	0
Siemens Ictotest		9	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Siemens Uristix		2	0	0	0	0
Biorex Ictero-Check		0	0	0	0	0
Other		1	0	0	0	0
All Instruments and Reagents		363	0	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Bilirubin	Specimen: U32	Negative	Small   1+   1mg/dL	Moderate   2+   3mg/dL	Large   3+   6mg/dL	Positive
<Instruments>						
Visual examination		26	0	0	0	0
Siemens Atlas		96	0	0	0	0
Siemens Clinitek Status		55	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0
Siemens Clinitek 500		26	0	0	0	0
Siemens Clinitek Advantus		67	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0
Roche Urisys		25	0	0	0	0
Roche Cobas U411		6	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0
Other Chemistry Analyzer		0	0	0	0	0
Other		1	0	0	0	0
<Reagents>						
Siemens Reagent Strips		178	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0
Roche Chemstrips		28	0	0	0	0
Roche Urisys Cassette		9	0	0	0	0
ARKRAY Aution Sticks		18	0	0	0	0
IRIS Reagent Strips		34	0	0	0	0
Siemens Ictotest		9	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Siemens Uristix		2	0	0	0	0
Biorex Ictero-Check		0	0	0	0	0
Other		1	0	0	0	0
All Instruments and Reagents		363	0	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

<b>Bilirubin</b>	<b>Specimen: U33</b>	Negative	Small   1+   1mg/dL	Moderate   2+   3mg/dL	Large   3+   6mg/dL	Positive
<Instruments>						
Visual examination		26	0	0	0	0
Siemens Atlas		96	0	0	0	0
Siemens Clinitek Status		55	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0
Siemens Clinitek 500		26	0	0	0	0
Siemens Clinitek Advantus		67	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0
Roche Urisys		25	0	0	0	0
Roche Cobas U411		6	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0
Other Chemistry Analyzer		0	0	0	0	0
Other		1	0	0	0	0
<Reagents>						
Siemens Reagent Strips		178	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0
Roche Chemstrips		28	0	0	0	0
Roche Urisys Cassette		9	0	0	0	0
ARKRAY Aution Sticks		18	0	0	0	0
IRIS Reagent Strips		34	0	0	0	0
Siemens Ictotest		9	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Siemens Uristix		2	0	0	0	0
Biorex Ictero-Check		0	0	0	0	0
Other		1	0	0	0	0
All Instruments and Reagents		363	0	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

<b>Bilirubin</b>	<b>Specimen: U34</b>	Negative	Small   1+   1mg/dL	Moderate   2+   3mg/dL	Large   3+   6mg/dL	Positive
<Instruments>						
Visual examination		26	0	0	0	0
Siemens Atlas		96	0	0	0	0
Siemens Clinitek Status		54	1	0	0	0
Siemens Clinitek 50		5	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0
Siemens Clinitek 500		26	0	0	0	0
Siemens Clinitek Advantus		67	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0
Roche Urisys		25	0	0	0	0
Roche Cobas U411		6	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0
Other Chemistry Analyzer		0	0	0	0	0
Other		1	0	0	0	0
<Reagents>						
Siemens Reagent Strips		177	1	0	0	0
Siemens Atlas Reagent		83	0	0	0	0
Roche Chemstrips		28	0	0	0	0
Roche Urisys Cassette		9	0	0	0	0
ARKRAY Aution Sticks		18	0	0	0	0
IRIS Reagent Strips		34	0	0	0	0
Siemens Ictotest		9	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Siemens Uristix		2	0	0	0	0
Biorex Ictero-Check		0	0	0	0	0
Other		1	0	0	0	0
All Instruments and Reagents		362	1	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

<b>Bilirubin</b>	<b>Specimen: U35</b>	Negative	Small   1+   1mg/dL	Moderate   2+   3mg/dL	Large   3+   6mg/dL	Positive
<Instruments>						
Visual examination		2	0	1	24	58
Siemens Atlas		0	0	5	86	5
Siemens Clinitek Status		0	0	1	53	1
Siemens Clinitek 50		0	0	0	5	0
Siemens Clinitek 100		0	0	0	1	1
Siemens Clinitek 500		1	0	17	6	2
Siemens Clinitek Advantus		0	1	45	21	0
Roche Chemstrip Criterion		0	2	2	0	0
Roche Urisys		0	1	22	2	0
Roche Cobas U411		0	0	6	0	0
IRIS AUTION MAX		2	1	18	2	0
IRIS iChem Velocity		27	0	0	0	0
Other Chemistry Analyzer		0	0	0	0	1
Other		1	0	0	0	0
<Reagents>						
Siemens Reagent Strips		0	1	63	108	7
Siemens Atlas Reagent		0	0	5	74	4
Roche Chemstrips		0	3	22	4	0
Roche Urisys Cassette		0	0	8	1	0
ARKRAY Aution Sticks		2	1	14	1	0
IRIS Reagent Strips		29	0	4	1	2
Siemens Ictotest		1	0	1	7	52
Siemens Hemastix		0	0	0	1	0
Siemens Uristix		1	0	0	1	0
Biorex Ictero-Check		0	0	0	1	3
Other		0	0	0	1	0
All Instruments and Reagents		33	5	117	200	68



New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Glucose Specimen: U31	Negative	Trace   50-100mg/dL	1/4%   150-300mg/dL	1/2%   500mg/dL	3/4%   750mg/dL	1%   >=1000mg/dL	2%   >=2000mg/dL
<Instruments>							
Visual examination	72	1	0	0	0	0	0
Siemens Atlas	96	0	0	0	0	0	0
Siemens Clinitek Status	57	0	0	0	0	0	0
Siemens Clinitek 50	5	0	0	0	0	0	0
Siemens Clinitek 100	2	0	0	0	0	0	0
Siemens Clinitek 500	27	0	0	0	0	0	0
Siemens Clinitek Advantus	67	0	0	0	0	0	0
Roche Chemstrip Criterion	4	0	0	0	0	0	0
Roche Urisys	26	0	0	0	0	0	0
Roche Cobas U411	6	0	0	0	0	0	0
IRIS AUTION MAX	23	0	0	0	0	0	0
IRIS iChem Velocity	27	0	0	0	0	0	0
Other Chemistry Analyzer	0	0	0	0	0	0	0
Other	1	0	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips	184	0	0	0	0	0	0
Siemens Atlas Reagent	84	0	0	0	0	0	0
Roche Chemstrips	29	0	0	0	0	0	0
Roche Urisys Cassette	10	0	0	0	0	0	0
ARKRAY Aution Sticks	19	0	0	0	0	0	0
IRIS Reagent Strips	32	0	0	0	0	0	0
Bayer Clinitest	49	1	0	0	0	0	0
Bayer Diastix	0	0	0	0	0	0	0
Siemens Hemastix	1	0	0	0	0	0	0
Siemens Uristix	4	0	0	0	0	0	0
Other	1	0	0	0	0	0	0
All Instruments and Reagents	413	1	0	0	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Glucose Specimen: U32	Negative	Trace	50-100mg/dL	1/4%   150-300mg/dL	1/2%   500mg/dL	3/4%   750mg/dL	1%   >=1000mg/dL	2%   >=2000mg/dL
<Instruments>								
Visual examination	0	4	20	55	11	5	0	0
Siemens Atlas	0	0	28	66	2	0	0	0
Siemens Clinitek Status	0	1	16	40	0	0	0	0
Siemens Clinitek 50	0	0	1	3	0	1	0	0
Siemens Clinitek 100	0	0	0	2	0	0	0	0
Siemens Clinitek 500	0	1	6	21	0	0	0	0
Siemens Clinitek Advantus	0	1	14	51	0	1	0	0
Roche Chemstrip Criterion	0	0	2	1	0	1	0	0
Roche Urisys	0	0	11	0	1	14	0	0
Roche Cobas U411	0	0	6	0	0	0	0	0
IRIS AUTION MAX	0	0	2	20	1	0	0	0
IRIS iChem Velocity	1	0	0	25	0	0	0	1
Other Chemistry Analyzer	0	0	0	1	0	0	0	0
Other	0	0	1	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips	0	3	47	132	0	3	0	0
Siemens Atlas Reagent	0	0	22	60	2	0	0	0
Roche Chemstrips	0	0	20	3	1	8	0	0
Roche Urisys Cassette	0	0	1	0	0	9	0	0
ARKRAY Aution Sticks	1	0	1	16	1	0	0	0
IRIS Reagent Strips	0	0	2	29	0	0	0	1
Bayer Clinitest	0	4	14	38	11	2	0	0
Bayer Diastix	0	0	0	1	0	0	0	0
Siemens Hemastix	0	0	0	1	0	0	0	0
Siemens Uristix	0	0	0	4	0	0	0	0
Other	0	0	0	1	0	0	0	0
All Instruments and Reagents	1	7	107	285	15	22	0	1

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Glucose Specimen: U33	Negative	Trace	50-100mg/dL	1/4%   150-300mg/dL	1/2%   500mg/dL	3/4%   750mg/dL	1%   >=1000mg/dL	2%   >=2000mg/dL
<Instruments>								
Visual examination	1	4	47	41	0	2	0	0
Siemens Atlas	0	3	91	2	0	0	0	0
Siemens Clinitek Status	0	3	28	26	0	0	0	0
Siemens Clinitek 50	0	0	1	4	0	0	0	0
Siemens Clinitek 100	0	0	1	1	0	0	0	0
Siemens Clinitek 500	0	0	10	18	0	0	0	0
Siemens Clinitek Advantus	1	0	35	30	0	1	0	0
Roche Chemstrip Criterion	0	0	4	0	0	0	0	0
Roche Urisys	0	0	18	0	0	8	0	0
Roche Cobas U411	0	0	6	0	0	0	0	0
IRIS AUTION MAX	0	0	19	4	0	0	0	0
IRIS iChem Velocity	0	0	0	26	0	0	0	1
Other Chemistry Analyzer	0	0	1	0	0	0	0	0
Other	0	0	1	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips	2	6	94	82	0	1	0	0
Siemens Atlas Reagent	0	0	82	2	0	0	0	0
Roche Chemstrips	0	0	28	4	0	0	0	0
Roche Urisys Cassette	0	0	2	0	0	8	0	0
ARKRAY Aution Sticks	0	0	15	4	0	0	0	0
IRIS Reagent Strips	0	0	5	26	0	0	0	1
Bayer Clinitest	0	4	34	29	0	2	0	0
Bayer Diastix	0	0	1	0	0	0	0	0
Siemens Hemastix	0	0	0	1	0	0	0	0
Siemens Uristix	0	0	0	4	0	0	0	0
Other	0	0	1	0	0	0	0	0
All Instruments and Reagents	2	10	262	152	0	11	0	1

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Glucose	Specimen: U34	Negative	Trace	50-100mg/dL	1/4%   150-300mg/dL	1/2%   500mg/dL	3/4%   750mg/dL	1%   >=1000mg/dL	2%   >=2000mg/dL
<Instruments>									
Visual examination		71	0	0	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0	0	0
Siemens Clinitek Status		56	1	0	0	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0	0	0	0
Siemens Clinitek 500		27	0	0	0	0	0	0	0
Siemens Clinitek Advantus		67	0	0	0	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0	0	0	0
Roche Urisys		26	0	0	0	0	0	0	0
Roche Cobas U411		6	0	0	0	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0	0	0	0
Other Chemistry Analyzer		0	0	0	0	0	0	0	0
Other		1	0	0	0	0	0	0	0
<Reagents>									
Siemens Reagent Strips		183	1	0	0	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	0	0	0	0
Roche Chemstrips		28	0	0	0	0	0	0	0
Roche Urisys Cassette		10	0	0	0	0	0	0	0
ARKRAY Aution Sticks		19	0	0	0	0	0	0	0
IRIS Reagent Strips		32	0	0	0	0	0	0	0
Bayer Clinitest		49	0	0	0	0	0	0	0
Bayer Diastix		0	0	0	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0	0	0
Siemens Uristix		4	0	0	0	0	0	0	0
Other		1	0	0	0	0	0	0	0
All Instruments and Reagents		411	1	0	0	0	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Glucose	Specimen: U35	Negative	Trace	50-100mg/dL	1/4%   150-300mg/dL	1/2%   500mg/dL	3/4%   750mg/dL	1%   >=1000mg/dL	2%   >=2000mg/dL
<Instruments>									
Visual examination		1	1	22	61	5	6	0	
Siemens Atlas		0	0	17	77	2	0	0	
Siemens Clinitek Status		0	1	19	34	0	3	0	
Siemens Clinitek 50		0	0	0	4	0	1	0	
Siemens Clinitek 100		0	0	0	2	0	0	0	
Siemens Clinitek 500		0	1	7	19	0	1	0	
Siemens Clinitek Advantus		0	0	16	46	0	5	0	
Roche Chemstrip Criterion		0	0	3	0	0	1	0	
Roche Urisys		0	0	12	0	0	14	0	
Roche Cobas U411		0	0	5	0	0	1	0	
IRIS AUTION MAX		0	0	3	19	1	0	0	
IRIS iChem Velocity		0	0	1	25	0	0	1	
Other Chemistry Analyzer		0	0	1	0	0	0	0	
Other		0	0	0	1	0	0	0	
<Reagents>									
Siemens Reagent Strips		1	2	57	114	0	11	0	
Siemens Atlas Reagent		0	0	10	72	2	0	0	
Roche Chemstrips		0	0	20	6	0	6	0	
Roche Urisys Cassette		0	0	0	0	0	10	0	
ARKRAY Aution Sticks		0	0	2	16	1	0	0	
IRIS Reagent Strips		0	0	2	29	0	0	1	
Bayer Clinitest		0	1	15	44	5	5	0	
Bayer Diastix		0	0	0	1	0	0	0	
Siemens Hemastix		0	0	0	1	0	0	0	
Siemens Uristix		0	0	0	4	0	0	0	
Other		0	0	0	1	0	0	0	
All Instruments and Reagents		1	3	106	288	8	32	1	

New York State Department of Health – Wadsworth Center  
 Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Hemoglobin	Specimen: U31	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		21	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		51	7	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0	0
Siemens Clinitek 500		25	2	0	0	0	0
Siemens Clinitek Advantus		62	4	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0	0
Roche Urisys		26	0	0	0	0	0
Roche Cobas U411		6	0	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0	0
IRIS iChem Velocity		26	0	1	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		173	13	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0	0
Roche Chemstrips		28	0	0	0	0	0
Roche Urisys Cassette		10	0	0	0	0	0
ARKRAY Aution Sticks		19	0	0	0	0	0
IRIS Reagent Strips		31	0	1	0	0	0
Siemens Uristix		1	0	0	0	0	0
None, no reagent or strip use		2	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		348	13	1	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Hemoglobin	Specimen: U32	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		1	0	0	1	0	23
Siemens Atlas		0	0	0	1	0	95
Siemens Clinitek Status		0	0	0	19	0	39
Siemens Clinitek 50		0	0	0	0	0	5
Siemens Clinitek 100		0	0	0	0	0	2
Siemens Clinitek 500		2	0	0	1	0	24
Siemens Clinitek Advantus		0	0	0	2	0	64
Roche Chemstrip Criterion		0	0	0	0	0	4
Roche Urisys		0	0	0	1	0	25
Roche Cobas U411		0	0	0	0	0	6
IRIS AUTION MAX		0	0	0	14	0	9
IRIS iChem Velocity		0	0	0	21	0	6
Other		0	0	0	0	0	1
<Reagents>							
Siemens Reagent Strips		2	0	0	23	0	164
Siemens Atlas Reagent		0	0	0	1	0	82
Roche Chemstrips		0	0	0	1	0	28
Roche Urisys Cassette		0	0	0	0	0	10
ARKRAY Aution Sticks		0	0	0	11	0	8
IRIS Reagent Strips		0	0	0	24	0	8
Siemens Uristix		0	0	0	0	0	1
None, no reagent or strip use		1	0	0	0	0	1
Other		0	0	0	0	0	1
All Instruments and Reagents		3	0	0	60	0	303

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Hemoglobin	Specimen: U33	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		1	0	0	0	0	24
Siemens Atlas		0	0	0	0	0	96
Siemens Clinitek Status		1	0	0	8	0	49
Siemens Clinitek 50		0	0	0	0	0	5
Siemens Clinitek 100		0	0	0	0	0	2
Siemens Clinitek 500		1	0	0	1	0	25
Siemens Clinitek Advantus		0	0	0	0	0	66
Roche Chemstrip Criterion		0	0	0	0	0	4
Roche Urisys		0	0	0	1	0	25
Roche Cobas U411		0	0	0	0	0	6
IRIS AUTION MAX		0	0	0	7	0	16
IRIS iChem Velocity		0	0	0	4	0	23
Other		0	0	0	1	0	0
<Reagents>							
Siemens Reagent Strips		2	0	0	9	0	178
Siemens Atlas Reagent		0	0	0	0	0	83
Roche Chemstrips		0	0	0	1	0	28
Roche Urisys Cassette		0	0	0	0	0	10
ARKRAY Aution Sticks		0	0	0	6	0	13
IRIS Reagent Strips		0	0	0	6	0	26
Siemens Uristix		0	0	0	0	0	1
None, no reagent or strip use		1	0	0	0	0	1
Other		0	0	0	0	0	1
All Instruments and Reagents		3	0	0	22	0	341



New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Hemoglobin	Specimen: U34	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		21	0	0	0	0	0
Siemens Atlas		95	1	0	0	0	0
Siemens Clinitek Status		46	12	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0	0
Siemens Clinitek 500		12	15	0	0	0	0
Siemens Clinitek Advantus		38	28	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0	0
Roche Urisys		26	0	0	0	0	0
Roche Cobas U411		6	0	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0	0
IRIS iChem Velocity		24	0	3	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		131	55	0	0	0	0
Siemens Atlas Reagent		82	1	0	0	0	0
Roche Chemstrips		28	0	0	0	0	0
Roche Urisys Cassette		10	0	0	0	0	0
ARKRAY Aution Sticks		19	0	0	0	0	0
IRIS Reagent Strips		29	0	3	0	0	0
Siemens Uristix		1	0	0	0	0	0
None, no reagent or strip use		2	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		303	56	3	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Hemoglobin	Specimen: U35	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		21	0	0	0	0	0
Siemens Atlas		95	0	0	0	0	1
Siemens Clinitek Status		54	4	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0	0
Siemens Clinitek 500		27	0	0	0	0	0
Siemens Clinitek Advantus		65	1	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0	0
Roche Urisys		26	0	0	0	0	0
Roche Cobas U411		6	0	0	0	0	0
IRIS AUTION MAX		22	0	1	0	0	0
IRIS iChem Velocity		26	0	1	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		181	5	0	0	0	0
Siemens Atlas Reagent		82	0	0	0	0	1
Roche Chemstrips		28	0	0	0	0	0
Roche Urisys Cassette		10	0	0	0	0	0
ARKRAY Aution Sticks		19	0	0	0	0	0
IRIS Reagent Strips		30	0	2	0	0	0
Siemens Uristix		1	0	0	0	0	0
None, no reagent or strip use		2	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		354	5	2	0	0	1

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Ketones	Specimen: U31	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<Instruments>							
Visual examination		0	0	0	4	36	1
Siemens Atlas		0	0	0	1	95	0
Siemens Clinitek Status		0	0	0	1	56	1
Siemens Clinitek 50		0	0	0	0	5	0
Siemens Clinitek 100		0	0	0	0	2	0
Siemens Clinitek 500		0	0	0	5	22	0
Siemens Clinitek Advantus		0	0	0	6	60	0
Roche Chemstrip Criterion		0	0	0	0	4	0
Roche Urisys		0	0	0	3	23	0
Roche Cobas U411		0	0	0	0	6	0
IRIS AUTION MAX		0	0	0	1	22	0
IRIS iChem Velocity		0	0	0	13	14	0
Other		0	0	0	1	0	0
<Reagents>							
Siemens Reagent Strips		0	0	0	13	173	1
Siemens Atlas Reagent		0	0	0	1	82	0
Roche Chemstrips		0	0	0	5	25	0
Roche Urisys Cassette		0	0	0	0	10	0
ARKRAY Aution Sticks		0	0	0	1	17	0
IRIS Reagent Strips		0	0	0	14	19	0
Siemens Acetest		0	0	0	1	16	1
Bayer Diastix		0	0	0	0	0	0
Siemens Uristix		0	0	0	0	1	0
Siemens Ketostix		0	0	0	0	1	0
Other		0	0	0	0	1	0
All Instruments and Reagents		0	0	0	35	345	2

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

<b>Ketones</b>	<b>Specimen: U32</b>	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<b>&lt;Instruments&gt;</b>							
Visual examination		23	1	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		58	0	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0	0
Siemens Clinitek 500		27	0	0	0	0	0
Siemens Clinitek Advantus		66	0	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0	0
Roche Urisys		26	0	0	0	0	0
Roche Cobas U411		6	0	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0	0
Other		1	0	0	0	0	0
<b>&lt;Reagents&gt;</b>							
Siemens Reagent Strips		185	1	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0	0
Roche Chemstrips		29	0	0	0	0	0
Roche Urisys Cassette		10	0	0	0	0	0
ARKRAY Aution Sticks		18	0	0	0	0	0
IRIS Reagent Strips		33	0	0	0	0	0
Siemens Acetest		4	0	0	0	0	0
Bayer Diastix		0	0	0	0	0	0
Siemens Uristix		1	0	0	0	0	0
Siemens Ketostix		0	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		364	1	0	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

<b>Ketones</b>	<b>Specimen: U33</b>	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<Instruments>							
Visual examination		24	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		58	0	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0	0
Siemens Clinitek 500		26	0	0	0	1	0
Siemens Clinitek Advantus		66	0	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0	0
Roche Urisys		26	0	0	0	0	0
Roche Cobas U411		6	0	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		185	0	0	0	1	0
Siemens Atlas Reagent		83	0	0	0	0	0
Roche Chemstrips		29	0	0	0	0	0
Roche Urisys Cassette		10	0	0	0	0	0
ARKRAY Aution Sticks		18	0	0	0	0	0
IRIS Reagent Strips		33	0	0	0	0	0
Siemens Acetest		4	0	0	0	0	0
Bayer Diastix		0	0	0	0	0	0
Siemens Uristix		1	0	0	0	0	0
Siemens Ketostix		0	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		364	0	0	0	1	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

<b>Ketones</b>	<b>Specimen: U34</b>	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<b>&lt;Instruments&gt;</b>							
Visual examination		24	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		58	0	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0	0
Siemens Clinitek 500		27	0	0	0	0	0
Siemens Clinitek Advantus		66	0	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0	0
Roche Urisys		26	0	0	0	0	0
Roche Cobas U411		6	0	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0	0
Other		1	0	0	0	0	0
<b>&lt;Reagents&gt;</b>							
Siemens Reagent Strips		186	0	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0	0
Roche Chemstrips		29	0	0	0	0	0
Roche Urisys Cassette		10	0	0	0	0	0
ARKRAY Aution Sticks		18	0	0	0	0	0
IRIS Reagent Strips		33	0	0	0	0	0
Siemens Acetest		4	0	0	0	0	0
Bayer Diastix		0	0	0	0	0	0
Siemens Uristix		1	0	0	0	0	0
Siemens Ketostix		0	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		365	0	0	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

<b>Ketones</b>	<b>Specimen: U35</b>	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<Instruments>							
Visual examination		24	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		58	0	0	0	0	0
Siemens Clinitek 50		5	0	0	0	0	0
Siemens Clinitek 100		2	0	0	0	0	0
Siemens Clinitek 500		27	0	0	0	0	0
Siemens Clinitek Advantus		66	0	0	0	0	0
Roche Chemstrip Criterion		4	0	0	0	0	0
Roche Urisys		26	0	0	0	0	0
Roche Cobas U411		6	0	0	0	0	0
IRIS AUTION MAX		23	0	0	0	0	0
IRIS iChem Velocity		27	0	0	0	0	0
Other		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		186	0	0	0	0	0
Siemens Atlas Reagent		83	0	0	0	0	0
Roche Chemstrips		29	0	0	0	0	0
Roche Urisys Cassette		10	0	0	0	0	0
ARKRAY Aution Sticks		18	0	0	0	0	0
IRIS Reagent Strips		33	0	0	0	0	0
Siemens Acetest		4	0	0	0	0	0
Bayer Diastix		0	0	0	0	0	0
Siemens Uristix		1	0	0	0	0	0
Siemens Ketostix		0	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		365	0	0	0	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Protein	Specimen: U31	Negative	Trace   10-15mg/dL	1+   25-30mg/dL	2+   75-150mg/dL	3+   300-600mg/dL	4+   >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	16	33	0	0	8
Siemens Atlas		0	0	4	92	0	0	0
Siemens Clinitek Status		0	0	0	57	0	0	0
Siemens Clinitek 50		0	0	0	5	0	0	0
Siemens Clinitek 100		0	0	0	2	0	0	0
Siemens Clinitek 500		0	0	14	13	0	0	0
Siemens Clinitek Advantus		0	0	25	42	0	0	0
Roche Chemstrip Criterion		0	0	3	1	0	0	0
Roche Urisys		0	0	5	21	0	0	0
Roche Cobas U411		0	0	0	6	0	0	0
IRIS AUTION MAX		0	0	18	5	0	0	0
IRIS iChem Velocity		0	0	0	27	0	0	0
Hitachi/Roche Analyzer		0	0	0	0	0	0	0
Other		0	0	0	1	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	45	138	0	0	0
Siemens Atlas Reagent		0	0	3	82	0	0	0
Sulfosalicylic Acid		0	0	8	18	0	0	8
Roche Chemstrips		0	0	9	20	0	0	0
Roche Urisys Cassette		0	0	1	9	0	0	0
ARKRAY Aution Sticks		0	0	15	4	0	0	0
IRIS Reagent Strips		0	0	3	29	0	0	0
Siemens Uristix		0	0	0	5	0	0	0
Hitachi/Roche Reagents		0	0	0	0	0	0	0
Other		0	0	1	0	0	0	0
All Instruments and Reagents		0	0	85	305	0	0	8



New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Protein	Specimen: U32	Negative	Trace   10-15mg/dL	1+   25-30mg/dL	2+   75-150mg/dL	3+   300-600mg/dL	4+   >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	0	5	40	6	8
Siemens Atlas		0	0	0	0	88	8	0
Siemens Clinitek Status		0	0	0	1	56	0	0
Siemens Clinitek 50		0	0	0	0	5	0	0
Siemens Clinitek 100		0	0	0	0	2	0	0
Siemens Clinitek 500		0	0	0	8	19	0	0
Siemens Clinitek Advantus		0	0	0	7	60	0	0
Roche Chemstrip Criterion		0	0	0	0	4	0	0
Roche Urisys		0	0	0	1	25	0	0
Roche Cobas U411		0	0	0	0	6	0	0
IRIS AUTION MAX		0	0	0	2	21	0	0
IRIS iChem Velocity		0	0	0	1	26	0	0
Hitachi/Roche Analyzer		0	0	0	0	0	0	1
Other		0	0	0	0	1	0	0
<Reagents>								
Siemens Reagent Strips		0	0	0	17	165	2	0
Siemens Atlas Reagent		0	0	0	0	77	8	0
Sulfosalicylic Acid		0	0	0	3	20	4	8
Roche Chemstrips		0	0	0	1	28	0	0
Roche Urisys Cassette		0	0	0	1	9	0	0
ARKRAY Aution Sticks		0	0	0	2	17	0	0
IRIS Reagent Strips		0	0	0	1	31	0	0
Siemens Uristix		0	0	0	0	5	0	0
Hitachi/Roche Reagents		0	0	0	0	0	0	1
Other		0	0	0	0	1	0	0
All Instruments and Reagents		0	0	0	25	353	14	9

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Protein	Specimen: U33	Negative	Trace   10-15mg/dL	1+   25-30mg/dL	2+   75-150mg/dL	3+   300-600mg/dL	4+   >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	2	42	5	1	8
Siemens Atlas		0	0	0	76	20	0	0
Siemens Clinitek Status		0	0	0	6	51	0	0
Siemens Clinitek 50		0	0	0	0	5	0	0
Siemens Clinitek 100		0	0	0	0	2	0	0
Siemens Clinitek 500		0	0	0	27	0	0	0
Siemens Clinitek Advantus		0	0	0	65	2	0	0
Roche Chemstrip Criterion		0	0	0	4	0	0	0
Roche Urisys		0	0	0	26	0	0	0
Roche Cobas U411		0	0	0	6	0	0	0
IRIS AUTION MAX		0	0	0	23	0	0	0
IRIS iChem Velocity		0	0	0	26	1	0	0
Hitachi/Roche Analyzer		0	0	0	0	0	0	1
Other		0	0	0	1	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	1	118	64	0	0
Siemens Atlas Reagent		0	0	0	66	19	0	0
Sulfosalicylic Acid		0	0	1	24	1	1	8
Roche Chemstrips		0	0	0	29	0	0	0
Roche Urisys Cassette		0	0	0	10	0	0	0
ARKRAY Aution Sticks		0	0	0	19	0	0	0
IRIS Reagent Strips		0	0	0	31	1	0	0
Siemens Uristix		0	0	0	4	1	0	0
Hitachi/Roche Reagents		0	0	0	0	0	0	1
Other		0	0	0	1	0	0	0
All Instruments and Reagents		0	0	2	302	86	1	9

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Protein	Specimen: U34	Negative	Trace   10-15mg/dL	1+   25-30mg/dL	2+   75-150mg/dL	3+   300-600mg/dL	4+   >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	6	42	1	0	8
Siemens Atlas		0	0	0	96	0	0	0
Siemens Clinitek Status		0	0	0	57	0	0	0
Siemens Clinitek 50		0	0	0	4	1	0	0
Siemens Clinitek 100		0	0	0	2	0	0	0
Siemens Clinitek 500		0	0	4	23	0	0	0
Siemens Clinitek Advantus		0	0	4	63	0	0	0
Roche Chemstrip Criterion		0	0	2	2	0	0	0
Roche Urisys		0	0	5	21	0	0	0
Roche Cobas U411		0	0	0	6	0	0	0
IRIS AUTION MAX		0	0	17	6	0	0	0
IRIS iChem Velocity		0	0	0	27	0	0	0
Hitachi/Roche Analyzer		0	0	0	0	0	0	0
Other		0	0	0	1	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	9	173	1	0	0
Siemens Atlas Reagent		0	0	0	85	0	0	0
Sulfosalicylic Acid		0	0	5	20	1	0	8
Roche Chemstrips		0	0	6	23	0	0	0
Roche Urisys Cassette		0	0	1	9	0	0	0
ARKRAY Aution Sticks		0	0	14	5	0	0	0
IRIS Reagent Strips		0	0	3	29	0	0	0
Siemens Uristix		0	0	0	5	0	0	0
Hitachi/Roche Reagents		0	0	0	0	0	0	0
Other		0	0	0	1	0	0	0
All Instruments and Reagents		0	0	38	350	2	0	8

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Protein	Specimen: U35	Negative	Trace   10-15mg/dL	1+   25-30mg/dL	2+   75-150mg/dL	3+   300-600mg/dL	4+   >=1000mg/dL	Positive
<Instruments>								
Visual examination		1	2	29	17	0	0	8
Siemens Atlas		0	0	77	19	0	0	0
Siemens Clinitek Status		0	0	3	54	0	0	0
Siemens Clinitek 50		0	0	0	5	0	0	0
Siemens Clinitek 100		0	0	0	2	0	0	0
Siemens Clinitek 500		0	0	25	2	0	0	0
Siemens Clinitek Advantus		1	0	64	2	0	0	0
Roche Chemstrip Criterion		0	0	4	0	0	0	0
Roche Urisys		0	0	6	20	0	0	0
Roche Cobas U411		0	0	0	6	0	0	0
IRIS AUTION MAX		0	1	20	2	0	0	0
IRIS iChem Velocity		0	0	17	10	0	0	0
Hitachi/Roche Analyzer		0	0	0	0	0	0	1
Other		0	0	0	1	0	0	0
<Reagents>								
Siemens Reagent Strips		2	1	108	72	0	0	0
Siemens Atlas Reagent		0	0	67	18	0	0	0
Sulfosalicylic Acid		0	1	17	8	0	0	8
Roche Chemstrips		0	0	11	18	0	0	0
Roche Urisys Cassette		0	0	2	8	0	0	0
ARKRAY Aution Sticks		0	1	17	1	0	0	0
IRIS Reagent Strips		0	0	20	12	0	0	0
Siemens Uristix		0	0	3	2	0	0	0
Hitachi/Roche Reagents		0	0	0	0	0	0	1
Other		0	0	0	1	0	0	0
All Instruments and Reagents		2	3	245	140	0	0	9

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Qualitative hCG report:

Other	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	1	0	0	1	1
Indeterminant	0	0	0	0	0
Positive	0	1	1	0	0

Alere hCG Cassette	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	13	0	0	13	13
Indeterminant	0	0	0	0	0
Positive	0	13	13	0	0

Siemens Clinitek Status	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	13	0	0	12	13
Indeterminant	0	0	0	0	0
Positive	0	13	13	1	0

Beckman Coulter ICON II	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	2	0	0	2	2
Indeterminant	0	0	0	0	0
Positive	0	2	2	0	0

Beckman Coulter ICON 25	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	17	0	0	17	17
Indeterminant	0	0	0	0	0
Positive	0	17	17	0	0

BTNX Rapid Response	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	2	0	0	2	2
Indeterminant	0	0	0	0	0
Positive	0	2	2	0	0

Beckman Coulter ICON 20	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	27	0	0	27	27
Indeterminant	0	0	0	0	0
Positive	0	27	27	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Cardinal Health hCG Combo	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	31	0	0	31	31
Indeterminant	0	0	0	0	0
Positive	0	31	31	0	0

Wampole Clearview hCG Combo	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	1	0	0	1	1
Indeterminant	0	0	0	0	0
Positive	0	1	1	0	0

Cardinal Health hCG Cassette	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	3	0	0	3	3
Indeterminant	0	0	0	0	0
Positive	0	3	3	0	0

Clarity hCG Cassette	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	1	0	0	1	1
Indeterminant	0	0	0	0	0
Positive	0	1	1	0	0

Cen Med Elite	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	13	0	0	13	13
Indeterminant	0	0	0	0	0
Positive	0	13	13	0	0

Sekisui OSOM Combo (Genzyme)	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	13	0	0	13	13
Indeterminant	0	0	0	0	0
Positive	0	13	13	0	0

Henry Schein One Step	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	4	0	0	4	4
Indeterminant	0	0	0	0	0
Positive	0	4	4	0	0

Immunostics hCG Detector	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	1	0	0	1	1
Indeterminant	0	0	0	0	0
Positive	0	1	1	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Inverness Medical BioStar Acceava	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	1	0	0	1	1
Indeterminant	0	0	0	0	0
Positive	0	1	1	0	0

Jant Accutest Value+	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	2	0	0	2	2
Indeterminant	0	0	0	0	0
Positive	0	2	2	0	0

McKesson hCG Cassette	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	1	0	0	1	1
Indeterminant	0	0	0	0	0
Positive	0	1	1	0	0

Meridian Diagnostics	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	1	0	0	1	1
Indeterminant	0	0	0	0	0
Positive	0	1	1	0	0

Polymedco Poly stat	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	5	0	0	5	5
Indeterminant	0	0	0	0	0
Positive	0	5	5	0	0

Quidel QuickVue	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	65	0	0	65	64
Indeterminant	0	0	0	0	0
Positive	0	65	65	0	1

R. Medical One Step Strip	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	2	0	0	2	2
Indeterminant	0	0	0	0	0
Positive	0	2	2	0	0

Stanbio QuPID	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	15	0	0	15	15
Indeterminant	0	0	0	0	0
Positive	0	15	15	0	0

New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

Fisher Sure-Vue serum/urine	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	39	0	0	39	39
Indeterminant	0	0	0	0	0
Positive	0	39	39	0	0

Fisher Sure-Vue STAT serum/urine	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	26	0	0	26	26
Indeterminant	0	0	0	0	0
Positive	0	26	26	0	0

Signify hCG Urine/Serum	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	3	0	0	3	3
Indeterminant	0	0	0	0	0
Positive	0	3	3	0	0

Siemens Immulite	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	5	0	0	5	5
Indeterminant	0	0	0	0	0
Positive	0	5	5	0	0

Fisher Sure-Vue (urine only)	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	8	0	0	8	8
Indeterminant	0	0	0	0	0
Positive	0	8	8	0	0

True Medix	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	5	0	0	5	5
Indeterminant	0	0	0	0	0
Positive	0	5	5	0	0

All Methods & Instruments	Specimen:U31	Specimen:U32	Specimen:U33	Specimen:U34	Specimen:U35
Negative	320	0	0	319	319
Indeterminant	0	0	0	0	0
Positive	0	320	320	1	1



New York State Department of Health – Wadsworth Center  
Urinalysis/ Urine Pregnancy Proficiency Testing – 13 April 2015

hCG Summary:

[Code], hCG Kit name	<u>U31</u>			<u>U32</u>			<u>U33</u>			<u>U34</u>			<u>U35</u>		
	N	I	P	N	I	P	N	I	P	N	I	P	N	I	P
[ZZZ] Other	1	0	0	0	0	1	0	0	1	1	0	0	1	0	0
[AC ] Alere hCG Cassette	13	0	0	0	0	13	0	0	13	13	0	0	13	0	0
[BA ] Siemens Clinitek Status	13	0	0	0	0	13	0	0	13	12	0	1	13	0	0
[BC ] Beckman Coulter ICON II	2	0	0	0	0	2	0	0	2	2	0	0	2	0	0
[BN ] Beckman Coulter ICON 25	17	0	0	0	0	17	0	0	17	17	0	0	17	0	0
[BR ] BTNX Rapid Response	2	0	0	0	0	2	0	0	2	2	0	0	2	0	0
[BT ] Beckman Coulter ICON 20	27	0	0	0	0	27	0	0	27	27	0	0	27	0	0
[CA ] Cardinal Health hCG Combo	31	0	0	0	0	31	0	0	31	31	0	0	31	0	0
[CC ] Wampole Clearview hCG Combo	1	0	0	0	0	1	0	0	1	1	0	0	1	0	0
[CH ] Cardinal Health hCG Cassette	3	0	0	0	0	3	0	0	3	3	0	0	3	0	0
[CL ] Clarity hCG Cassette	1	0	0	0	0	1	0	0	1	1	0	0	1	0	0
[CM ] Cen Med Elite	13	0	0	0	0	13	0	0	13	13	0	0	13	0	0
[GO ] Sekisui OSOM Combo (Genzyme)	13	0	0	0	0	13	0	0	13	13	0	0	13	0	0
[HS ] Henry Schein One Step	4	0	0	0	0	4	0	0	4	4	0	0	4	0	0
[ID ] Immunostics hCG Detector	1	0	0	0	0	1	0	0	1	1	0	0	1	0	0
[IM ] Inverness Medical BioStar Aceava	1	0	0	0	0	1	0	0	1	1	0	0	1	0	0
[JA ] Jant Accutest Value+	2	0	0	0	0	2	0	0	2	2	0	0	2	0	0
[MC ] McKesson hCG Cassette	1	0	0	0	0	1	0	0	1	1	0	0	1	0	0
[MD ] Meridian Diagnostics	1	0	0	0	0	1	0	0	1	1	0	0	1	0	0
[PM ] Polymedco Poly stat	5	0	0	0	0	5	0	0	5	5	0	0	5	0	0
[QQ ] Quidel QuickVue	65	0	0	0	0	65	0	0	65	65	0	0	64	0	1
[RM ] R. Medical One Step Strip	2	0	0	0	0	2	0	0	2	2	0	0	2	0	0
[SB ] Stanbio QuPID	15	0	0	0	0	15	0	0	15	15	0	0	15	0	0
[SE ] Fisher Sure-Vue serum/urine	39	0	0	0	0	39	0	0	39	39	0	0	39	0	0
[SF ] Fisher Sure-Vue STAT serum/urine	26	0	0	0	0	26	0	0	26	26	0	0	26	0	0
[SG ] Signify hCG Urine/Serum	3	0	0	0	0	3	0	0	3	3	0	0	3	0	0
[SI ] Siemens Immulite	5	0	0	0	0	5	0	0	5	5	0	0	5	0	0
[SU ] Fisher Sure-Vue (urine only)	8	0	0	0	0	8	0	0	8	8	0	0	8	0	0
[TX ] True Medix	5	0	0	0	0	5	0	0	5	5	0	0	5	0	0
[99 ] All hCG Methods	320	0	0	0	0	320	0	0	320	319	0	1	319	0	1