

**Statistical Report**  
**Urinalysis/Urine Pregnancy Testing**  
**Mail out date: October 28, 2013**

This report summarizes data from the urinalysis/urine pregnancy proficiency testing of October 28, 2013.

If you have any questions regarding this report, please contact Regina Saulsbery at 518-474-6156. You may also contact us by e-mail: [clinchem@wadsworth.org](mailto:clinchem@wadsworth.org)

Results are shown for each analyte arranged by instrument and reagent system. Results for specific gravity and pH are listed as mean  $\pm$  1SD for each of the instruments and reagents. All other results are shown as the number of participant results for that instrument or reagent.

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

PLEASE NOTE: The evaluation report and summary of the educational quantitative urine clinical chemistry proficiency test will be mailed separately if your laboratory participated in that proficiency test.

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

**Specific Gravity**

Specimen: U16 -----	Specimen: U17 -----	Specimen: U18 -----	Specimen: U19 -----	Specimen: U20 -----	Number	Instrument or Reagent System -----
1.012 ± 0.004	1.016 ± 0.006	1.019 ± 0.006	1.018 ± 0.005	1.018 ± 0.004	n = 361	All Methods & Instruments
1.010	1.015	1.019	1.020	1.019	n = 361	Median value, All Methods & Instruments
<Instruments>						
1.010 ± 0.004	1.018 ± 0.004	1.017 ± 0.005	1.021 ± 0.003	1.013 ± 0.003	n = 24	Visual examination
1.009 ± 0.001	1.011 ± 0.001	1.019 ± 0.001	1.015 ± 0.001	1.020 ± 0.001	n = 95	Siemens Atlas
1.018 ± 0.002	1.024 ± 0.002	1.025 ± 0.001	1.022 ± 0.002	1.020 ± 0.001	n = 35	Siemens Clinitek Status
1.015 ± 0.003	1.022 ± 0.003	1.026 ± 0.004	1.023 ± 0.003	1.018 ± 0.006	n = 6	Siemens Clinitek 50
1.019 ± 0.003	1.024 ± 0.002	1.025 ± 0.000	1.023 ± 0.003	1.021 ± 0.002	n = 4	Siemens Clinitek 100
1.015 ± 0.003	1.022 ± 0.002	1.024 ± 0.002	1.023 ± 0.002	1.016 ± 0.003	n = 32	Siemens Clinitek 500
1.014 ± 0.002	1.022 ± 0.002	1.024 ± 0.001	1.022 ± 0.002	1.016 ± 0.003	n = 56	Siemens Clinitek Advantus
1.012 ± 0.002	1.015 ± 0.002	1.009 ± 0.002	1.019 ± 0.002	1.014 ± 0.002	n = 12	Roche Chemstrip Criterion
1.011 ± 0.003	1.014 ± 0.003	1.014 ± 0.004	1.018 ± 0.003	1.017 ± 0.003	n = 33	Roche Urisys
1.012 ± 0.003	1.015 ± 0.000	1.010 ± 0.000	1.020 ± 0.000	1.015 ± 0.000	n = 3	Roche Cobas U411
1.007 ± 0.001	1.008 ± 0.001	1.014 ± 0.001	1.009 ± 0.001	1.017 ± 0.001	n = 36	IRIS AUTION MAX
1.008 ± 0.001	1.009 ± 0.000	1.013 ± 0.000	1.009 ± 0.000	1.018 ± 0.000	n = 8	IRIS iChem Velocity
1.009 ± 0.001	1.009 ± 0.001	1.016 ± 0.002	1.012 ± 0.002	1.019 ± 0.001	n = 16	Refractometer
<Reagents>						
1.015 ± 0.004	1.021 ± 0.004	1.024 ± 0.004	1.022 ± 0.003	1.017 ± 0.004	n = 158	Siemens Reagent Strips
1.009 ± 0.001	1.011 ± 0.001	1.019 ± 0.001	1.015 ± 0.001	1.020 ± 0.001	n = 88	Siemens Atlas Reagent
1.012 ± 0.003	1.015 ± 0.002	1.011 ± 0.002	1.019 ± 0.002	1.015 ± 0.002	n = 40	Roche Chemstrips
1.009 ± 0.001	1.010 ± 0.000	1.018 ± 0.001	1.014 ± 0.000	1.020 ± 0.001	n = 13	Roche Urisys Cassette
1.006 ± 0.001	1.008 ± 0.001	1.014 ± 0.001	1.009 ± 0.001	1.017 ± 0.001	n = 22	ARKRAY Aution Sticks
1.007 ± 0.001	1.009 ± 0.001	1.013 ± 0.000	1.009 ± 0.000	1.017 ± 0.001	n = 15	IRIS Reagent Strips
1.009 ± 0.002	1.010 ± 0.003	1.017 ± 0.003	1.012 ± 0.003	1.019 ± 0.001	n = 22	None, no reagent or strip used
1.012 ± 0.007	1.015 ± 0.009	1.018 ± 0.006	1.015 ± 0.008	1.019 ± 0.001	n = 3	Other

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

**pH**

Specimen: U16	Specimen: U17	Specimen: U18	Specimen: U19	Specimen: U20	Number	Instrument or Reagent System
7.58 ± 0.24	6.60 ± 0.25	7.07 ± 0.19	5.91 ± 0.26	6.94 ± 0.17	n = 367	All Methods & Instruments
7.50	6.50	7.00	6.00	7.00	n = 367	Median value, All Methods & Instruments
<Instruments>						
7.61 ± 0.21	6.58 ± 0.31	7.20 ± 0.31	5.75 ± 0.44	6.95 ± 0.20	n = 32	Visual examination
7.51 ± 0.05	6.52 ± 0.12	7.08 ± 0.18	6.01 ± 0.05	6.99 ± 0.05	n = 95	Siemens Atlas
7.50 ± 0.16	6.95 ± 0.19	7.05 ± 0.15	5.97 ± 0.16	7.00 ± 0.00	n = 40	Siemens Clinitek Status
7.50 ± 0.00	6.50 ± 0.32	7.00 ± 0.00	5.75 ± 0.27	7.00 ± 0.00	n = 6	Siemens Clinitek 50
7.80 ± 0.27	6.80 ± 0.27	7.40 ± 0.22	6.00 ± 0.00	6.80 ± 0.45	n = 5	Siemens Clinitek 100
7.50 ± 0.21	6.49 ± 0.19	7.06 ± 0.16	5.96 ± 0.14	6.88 ± 0.25	n = 34	Siemens Clinitek 500
7.53 ± 0.17	6.53 ± 0.22	7.10 ± 0.20	5.99 ± 0.07	6.97 ± 0.11	n = 59	Siemens Clinitek Advantus
8.00 ± 0.00	6.65 ± 0.24	7.00 ± 0.00	6.00 ± 0.00	6.96 ± 0.14	n = 13	Roche Chemstrip Criterion
7.81 ± 0.40	6.63 ± 0.28	7.03 ± 0.18	5.59 ± 0.50	6.98 ± 0.09	n = 32	Roche Urisys
7.33 ± 0.58	6.50 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	n = 3	Roche Cobas U411
7.51 ± 0.08	6.51 ± 0.08	7.00 ± 0.00	5.76 ± 0.25	6.72 ± 0.25	n = 36	IRIS AUTION MAX
8.00 ± 0.00	7.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	n = 9	IRIS iChem Velocity
<Reagents>						
7.52 ± 0.17	6.64 ± 0.29	7.11 ± 0.22	5.97 ± 0.12	6.96 ± 0.17	n = 174	Siemens Reagent Strips
7.51 ± 0.05	6.52 ± 0.13	7.09 ± 0.19	6.01 ± 0.05	6.99 ± 0.05	n = 88	Siemens Atlas Reagent
7.83 ± 0.37	6.53 ± 0.25	7.02 ± 0.15	5.66 ± 0.48	6.95 ± 0.15	n = 44	Roche Chemstrips
7.92 ± 0.28	6.85 ± 0.24	7.00 ± 0.00	5.54 ± 0.52	7.00 ± 0.00	n = 13	Roche Urisys Cassette
7.50 ± 0.00	6.50 ± 0.00	7.00 ± 0.00	5.77 ± 0.25	6.66 ± 0.24	n = 28	ARKRAY Aution Sticks
7.79 ± 0.25	6.79 ± 0.25	7.00 ± 0.00	5.88 ± 0.22	6.97 ± 0.12	n = 17	IRIS Reagent Strips

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Bilirubin	Specimen: U16	Negative	Small  1+  1mg/dL	Moderate  2+  3mg/dL	Large  3+  6mg/dL	Positive
<Instruments>						
	Visual examination	37	0	0	0	0
	Siemens Atlas	95	0	0	0	0
	Siemens Clinitek Status	39	0	0	0	0
	Siemens Clinitek 50	6	0	0	0	0
	Siemens Clinitek 100	5	0	0	0	0
	Siemens Clinitek 500	34	0	0	0	0
	Siemens Clinitek Advantus	58	0	0	0	0
	Roche UA Reader/Super UA Read	1	0	0	0	0
	Roche Chemstrip Criterion	12	0	0	0	0
	Roche Urisys	33	0	0	0	0
	Roche Cobas U411	3	0	0	0	0
	IRIS AUTION MAX	36	0	0	0	0
	IRIS iChem Velocity	9	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	170	0	0	0	0
	Siemens Atlas Reagent	86	0	0	0	0
	Roche Chemstrips	42	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0
	ARKRAY Aution Sticks	26	0	0	0	0
	IRIS Reagent Strips	19	0	0	0	0
	Siemens Ictotest	8	0	0	0	0
	Siemens Hemastix	1	0	0	0	0
	Siemens Uristix	1	0	0	0	0
	Hitachi/Roche Reagents	1	0	0	0	0
	Other	1	0	0	0	0
	All Instruments and Reagents	368	0	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Bilirubin	Specimen: U17	Negative	Small  1+  1mg/dL	Moderate  2+  3mg/dL	Large  3+  6mg/dL	Positive
<b>&lt;Instruments&gt;</b>						
Visual examination		3	3	9	25	17
Siemens Atlas		0	0	79	12	4
Siemens Clinitek Status		0	0	32	7	0
Siemens Clinitek 50		0	0	4	2	0
Siemens Clinitek 100		0	0	1	4	0
Siemens Clinitek 500		0	0	34	0	0
Siemens Clinitek Advantus		1	0	57	0	0
Roche UA Reader/Super UA Read		0	0	0	0	1
Roche Chemstrip Criterion		0	11	0	0	1
Roche Urisys		1	14	17	1	0
Roche Cobas U411		0	2	1	0	0
IRIS AUTION MAX		1	1	34	0	0
IRIS iChem Velocity		9	0	0	0	0
<b>&lt;Reagents&gt;</b>						
Siemens Reagent Strips		1	0	138	31	1
Siemens Atlas Reagent		0	1	71	11	3
Roche Chemstrips		1	16	16	8	2
Roche Urisys Cassette		0	10	3	0	0
ARKRAY Aution Sticks		0	1	25	0	0
IRIS Reagent Strips		11	2	8	0	3
Siemens Ictotest		0	0	4	0	14
Siemens Hemastix		0	0	0	1	0
Siemens Uristix		0	0	1	0	0
Hitachi/Roche Reagents		0	1	0	0	0
None, no reagent or strip use		1	0	0	0	0
Other		1	0	2	0	0
All Instruments and Reagents		15	31	268	51	23

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Bilirubin	Specimen: U18	Negative	Small  1+  1mg/dL	Moderate  2+  3mg/dL	Large  3+  6mg/dL	Positive
<Instruments>						
	Visual examination	37	0	0	0	0
	Siemens Atlas	95	0	0	0	0
	Siemens Clinitek Status	39	0	0	0	0
	Siemens Clinitek 50	6	0	0	0	0
	Siemens Clinitek 100	5	0	0	0	0
	Siemens Clinitek 500	34	0	0	0	0
	Siemens Clinitek Advantus	58	0	0	0	0
	Roche UA Reader/Super UA Read	1	0	0	0	0
	Roche Chemstrip Criterion	12	0	0	0	0
	Roche Urisys	33	0	0	0	0
	Roche Cobas U411	3	0	0	0	0
	IRIS AUTION MAX	36	0	0	0	0
	IRIS iChem Velocity	9	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	170	0	0	0	0
	Siemens Atlas Reagent	86	0	0	0	0
	Roche Chemstrips	42	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0
	ARKRAY Aution Sticks	26	0	0	0	0
	IRIS Reagent Strips	19	0	0	0	0
	Siemens Ictotest	8	0	0	0	0
	Siemens Hemastix	1	0	0	0	0
	Siemens Uristix	1	0	0	0	0
	Hitachi/Roche Reagents	1	0	0	0	0
	Other	1	0	0	0	0
	All Instruments and Reagents	368	0	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Bilirubin	Specimen: U19	Negative	Small  1+  1mg/dL	Moderate  2+  3mg/dL	Large  3+  6mg/dL	Positive
<Instruments>						
	Visual examination	37	0	0	0	0
	Siemens Atlas	95	0	0	0	0
	Siemens Clinitek Status	39	0	0	0	0
	Siemens Clinitek 50	6	0	0	0	0
	Siemens Clinitek 100	5	0	0	0	0
	Siemens Clinitek 500	34	0	0	0	0
	Siemens Clinitek Advantus	58	0	0	0	0
	Roche UA Reader/Super UA Read	1	0	0	0	0
	Roche Chemstrip Criterion	12	0	0	0	0
	Roche Urisys	33	0	0	0	0
	Roche Cobas U411	3	0	0	0	0
	IRIS AUTION MAX	36	0	0	0	0
	IRIS iChem Velocity	9	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	170	0	0	0	0
	Siemens Atlas Reagent	86	0	0	0	0
	Roche Chemstrips	42	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0
	ARKRAY Aution Sticks	26	0	0	0	0
	IRIS Reagent Strips	19	0	0	0	0
	Siemens Ictotest	8	0	0	0	0
	Siemens Hemastix	1	0	0	0	0
	Siemens Uristix	1	0	0	0	0
	Hitachi/Roche Reagents	1	0	0	0	0
	Other	1	0	0	0	0
	All Instruments and Reagents	368	0	0	0	0

**New York State Department of Health - Wadsworth Center  
Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Bilirubin	Specimen: U20	Negative	Small  1+  1mg/dL	Moderate  2+  3mg/dL	Large  3+  6mg/dL	Positive
<Instruments>						
	Visual examination	37	0	0	0	0
	Siemens Atlas	95	0	0	0	0
	Siemens Clinitek Status	39	0	0	0	0
	Siemens Clinitek 50	6	0	0	0	0
	Siemens Clinitek 100	5	0	0	0	0
	Siemens Clinitek 500	34	0	0	0	0
	Siemens Clinitek Advantus	58	0	0	0	0
	Roche UA Reader/Super UA Read	1	0	0	0	0
	Roche Chemstrip Criterion	12	0	0	0	0
	Roche Urisys	33	0	0	0	0
	Roche Cobas U411	3	0	0	0	0
	IRIS AUTION MAX	36	0	0	0	0
	IRIS iChem Velocity	9	0	0	0	0
<Reagents>						
	Siemens Reagent Strips	170	0	0	0	0
	Siemens Atlas Reagent	86	0	0	0	0
	Roche Chemstrips	42	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0
	ARKRAY Aution Sticks	26	0	0	0	0
	IRIS Reagent Strips	19	0	0	0	0
	Siemens Ictotest	8	0	0	0	0
	Siemens Hemastix	1	0	0	0	0
	Siemens Uristix	1	0	0	0	0
	Hitachi/Roche Reagents	1	0	0	0	0
	Other	1	0	0	0	0
	All Instruments and Reagents	368	0	0	0	0



**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Glucose	Specimen: U16	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		69	14	9	0	0	0	0
Siemens Atlas		90	5	0	0	0	0	0
Siemens Clinitek Status		35	5	0	0	0	0	0
Siemens Clinitek 50		3	3	0	0	0	0	0
Siemens Clinitek 100		5	0	0	0	0	0	0
Siemens Clinitek 500		34	1	0	0	0	0	0
Siemens Clinitek Advantus		52	6	0	0	0	0	0
Roche UA Reader/Super UA Read		1	1	0	0	0	0	0
Roche Chemstrip Criterion		2	11	0	0	0	0	0
Roche Urisys		20	12	0	0	0	0	0
Roche Cobas U411		2	1	0	0	0	0	0
IRIS AUTION MAX		5	31	0	0	0	0	0
IRIS iChem Velocity		0	6	3	0	0	0	0
Hitachi/Roche Analyzer		0	1	0	0	0	0	0
Other Chemistry Analyzer		1	1	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		162	16	0	0	0	0	0
Siemens Atlas Reagent		80	5	0	0	0	0	0
Roche Chemstrips		19	27	0	0	0	0	0
Roche Urisys Cassette		7	5	0	0	0	0	0
ARKRAY Aution Sticks		4	24	0	0	0	0	0
IRIS Reagent Strips		1	13	3	0	0	0	0
Bayer Clinitest		43	4	9	0	0	0	0
Bayer Diastix		1	0	0	0	0	0	0
Siemens Uristix		1	2	0	0	0	0	0
Hitachi/Roche Reagents		0	1	0	0	0	0	0
Other		1	1	0	0	0	0	0
All Instruments and Reagents		319	98	12	0	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Glucose	Specimen: U17	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<b>&lt;Instruments&gt;</b>								
Visual examination		87	1	0	1	0	0	0
Siemens Atlas		95	0	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0	0
Siemens Clinitek 50		6	0	0	0	0	0	0
Siemens Clinitek 100		5	0	0	0	0	0	0
Siemens Clinitek 500		35	0	0	0	0	0	0
Siemens Clinitek Advantus		58	0	0	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0	0	0
Roche Urisys		32	0	0	0	0	0	0
Roche Cobas U411		3	0	0	0	0	0	0
IRIS AUTION MAX		36	0	0	0	0	0	0
IRIS iChem Velocity		9	0	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0	0
Other Chemistry Analyzer		2	0	0	0	0	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		178	0	0	0	0	0	0
Siemens Atlas Reagent		85	0	0	0	0	0	0
Roche Chemstrips		45	0	0	0	0	0	0
Roche Urisys Cassette		12	0	0	0	0	0	0
ARKRAY Aution Sticks		28	0	0	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0	0
Bayer Clinitest		52	1	0	1	0	0	0
Bayer Diastix		1	0	0	0	0	0	0
Siemens Uristix		3	0	0	0	0	0	0
Hitachi/Roche Reagents		1	0	0	0	0	0	0
Other		2	0	0	0	0	0	0
All Instruments and Reagents		424	1	0	1	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Glucose	Specimen: U18	Negative	Trace 50-100mg/dL	1/4 150-300mg/dL	1/2 500mg/dL	3/4 750mg/dL	1 >=1000mg/dL	2 >=2000mg/dL
<b>&lt;Instruments&gt;</b>								
Visual examination		8	25	51	32	1	1	0
Siemens Atlas		0	0	96	0	0	0	0
Siemens Clinitek Status		0	39	1	0	0	0	0
Siemens Clinitek 50		0	6	0	0	0	0	0
Siemens Clinitek 100		2	3	0	0	0	0	0
Siemens Clinitek 500		7	25	3	0	0	0	0
Siemens Clinitek Advantus		5	52	0	1	0	0	0
Roche UA Reader/Super UA Read		0	0	2	0	0	0	0
Roche Chemstrip Criterion		0	1	11	1	0	0	0
Roche Urisys		0	2	30	0	0	0	0
Roche Cobas U411		0	0	3	0	0	0	0
IRIS AUTION MAX		0	0	29	6	1	0	0
IRIS iChem Velocity		0	0	0	9	0	0	0
Hitachi/Roche Analyzer		0	0	1	0	0	0	0
Other Chemistry Analyzer		0	0	2	0	0	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		22	140	15	2	0	0	0
Siemens Atlas Reagent		0	0	86	0	0	0	0
Roche Chemstrips		0	5	41	2	0	0	0
Roche Urisys Cassette		0	0	12	0	0	0	0
ARKRAY Aution Sticks		0	0	23	4	1	0	0
IRIS Reagent Strips		0	0	6	11	0	0	0
Bayer Clinitest		0	4	42	30	1	1	0
Bayer Diastix		0	1	1	0	0	0	0
Siemens Uristix		0	2	1	0	0	0	0
Hitachi/Roche Reagents		0	0	1	0	0	0	0
Other		0	1	1	0	0	0	0
All Instruments and Reagents		22	153	229	49	2	1	0

**New York State Department of Health - Wadsworth Center  
Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Glucose	Specimen: U19	Negative	Trace   50-100mg/dL	1/4%   150-300mg/dL	1/2%   500mg/dL	3/4%   750mg/dL	1%   >=1000mg/dL	2%   >=2000mg/dL
<b>&lt;Instruments&gt;</b>								
Visual examination		2	0	1	23	24	58	10
Siemens Atlas		0	0	0	85	1	10	0
Siemens Clinitek Status		0	0	1	29	1	8	1
Siemens Clinitek 50		0	0	1	1	0	4	0
Siemens Clinitek 100		0	0	0	1	0	4	0
Siemens Clinitek 500		0	0	0	14	0	21	1
Siemens Clinitek Advantus		0	0	1	15	1	40	2
Roche UA Reader/Super UA Read		0	0	0	0	0	2	0
Roche Chemstrip Criterion		0	0	1	2	0	10	0
Roche Urisys		0	0	0	2	0	30	0
Roche Cobas U411		0	0	0	0	0	3	0
IRIS AUTION MAX		0	0	7	28	1	0	0
IRIS iChem Velocity		0	0	0	9	0	0	0
Hitachi/Roche Analyzer		0	0	0	1	0	0	0
Other Chemistry Analyzer		0	0	0	2	0	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		0	0	4	73	2	91	10
Siemens Atlas Reagent		0	0	0	78	1	7	0
Roche Chemstrips		0	0	1	5	0	42	0
Roche Urisys Cassette		0	0	0	1	0	11	0
ARKRAY Aution Sticks		0	0	5	22	1	0	0
IRIS Reagent Strips		0	0	2	15	0	0	0
Bayer Clinitest		2	0	0	15	24	34	4
Bayer Diastix		0	0	0	0	0	2	0
Siemens Uristix		0	0	0	0	0	3	0
Hitachi/Roche Reagents		0	0	0	1	0	0	0
Other		0	0	0	2	0	0	0
All Instruments and Reagents		2	0	12	212	28	190	14

**New York State Department of Health - Wadsworth Center  
Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Glucose	Specimen: U20	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<b>&lt;Instruments&gt;</b>								
Visual examination		88	0	0	0	0	1	0
Siemens Atlas		95	0	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0	0
Siemens Clinitek 50		6	0	0	0	0	0	0
Siemens Clinitek 100		4	0	0	0	0	1	0
Siemens Clinitek 500		35	0	0	0	0	0	0
Siemens Clinitek Advantus		58	0	0	0	0	0	0
Roche UA Reader/Super UA Read		2	0	0	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0	0	0
Roche Urisys		32	0	0	0	0	0	0
Roche Cobas U411		3	0	0	0	0	0	0
IRIS AUTION MAX		36	0	0	0	0	0	0
IRIS iChem Velocity		9	0	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0	0
Other Chemistry Analyzer		2	0	0	0	0	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		177	0	0	0	0	1	0
Siemens Atlas Reagent		85	0	0	0	0	0	0
Roche Chemstrips		45	0	0	0	0	0	0
Roche Urisys Cassette		12	0	0	0	0	0	0
ARKRAY Aution Sticks		28	0	0	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0	0
Bayer Clinitest		53	0	0	0	0	1	0
Bayer Diastix		1	0	0	0	0	0	0
Siemens Uristix		3	0	0	0	0	0	0
Hitachi/Roche Reagents		1	0	0	0	0	0	0
Other		2	0	0	0	0	0	0
All Instruments and Reagents		424	0	0	0	0	2	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Hemoglobin	Specimen: U16	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<b>&lt;Instruments&gt;</b>							
Visual examination		0	0	0	3	1	33
Siemens Atlas		0	0	0	9	0	86
Siemens Clinitek Status		0	0	0	23	0	18
Siemens Clinitek 50		0	0	0	4	0	2
Siemens Clinitek 100		0	0	0	1	0	4
Siemens Clinitek 500		0	0	0	6	0	29
Siemens Clinitek Advantus		0	0	0	9	0	48
Roche UA Reader/Super UA Read		0	0	0	0	0	1
Roche Chemstrip Criterion		0	0	0	0	1	12
Roche Urisys		0	0	0	9	15	9
Roche Cobas U411		0	0	0	1	1	1
IRIS AUTION MAX		0	0	0	36	0	0
IRIS iChem Velocity		0	0	1	8	0	0
Hitachi/Roche Analyzer		0	0	0	1	0	0
<b>&lt;Reagents&gt;</b>							
Siemens Reagent Strips		0	0	0	45	0	133
Siemens Atlas Reagent		0	0	0	8	0	79
Roche Chemstrips		0	0	0	6	10	29
Roche Urisys Cassette		0	0	0	5	7	1
ARKRAY Aution Sticks		0	0	0	28	0	0
IRIS Reagent Strips		0	0	1	16	0	0
Siemens Hemastix		0	0	0	0	0	1
None, no reagent or strip use		0	0	0	0	1	0
Other		0	0	0	2	0	0
All Instruments and Reagents		0	0	1	110	18	243

**New York State Department of Health - Wadsworth Center  
Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Hemoglobin	Specimen: U17	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
	Visual examination	35	1	0	0	0	0
	Siemens Atlas	94	1	0	0	0	0
	Siemens Clinitek Status	34	7	0	0	0	0
	Siemens Clinitek 50	5	1	0	0	0	0
	Siemens Clinitek 100	5	0	0	0	0	0
	Siemens Clinitek 500	26	8	0	1	0	0
	Siemens Clinitek Advantus	36	21	0	0	0	0
	Roche UA Reader/Super UA Read	1	0	0	0	0	0
	Roche Chemstrip Criterion	13	0	0	0	0	0
	Roche Urisys	33	0	0	0	0	0
	Roche Cobas U411	3	0	0	0	0	0
	IRIS AUTION MAX	36	0	0	0	0	0
	IRIS iChem Velocity	9	0	0	0	0	0
	Hitachi/Roche Analyzer	1	0	0	0	0	0
<Reagents>							
	Siemens Reagent Strips	138	38	0	1	0	0
	Siemens Atlas Reagent	86	1	0	0	0	0
	Roche Chemstrips	45	0	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0	0
	ARKRAY Aution Sticks	28	0	0	0	0	0
	IRIS Reagent Strips	17	0	0	0	0	0
	Siemens Hemastix	1	0	0	0	0	0
	None, no reagent or strip use	1	0	0	0	0	0
	Other	2	0	0	0	0	0
	All Instruments and Reagents	331	39	0	1	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Hemoglobin	Specimen: U18	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<b>&lt;Instruments&gt;</b>							
Visual examination		0	0	0	1	0	36
Siemens Atlas		0	0	0	2	0	93
Siemens Clinitek Status		0	0	0	10	0	31
Siemens Clinitek 50		0	0	0	1	0	5
Siemens Clinitek 100		0	0	0	1	0	4
Siemens Clinitek 500		0	0	0	1	0	34
Siemens Clinitek Advantus		0	0	0	4	0	53
Roche UA Reader/Super UA Read		0	0	0	0	0	1
Roche Chemstrip Criterion		0	0	0	0	0	13
Roche Urisys		0	0	0	0	0	33
Roche Cobas U411		0	0	0	0	0	3
IRIS AUTION MAX		0	0	0	29	0	7
IRIS iChem Velocity		0	0	1	8	0	0
Hitachi/Roche Analyzer		0	0	0	0	0	1
<b>&lt;Reagents&gt;</b>							
Siemens Reagent Strips		0	0	0	17	0	161
Siemens Atlas Reagent		0	0	0	1	0	86
Roche Chemstrips		0	0	0	1	0	44
Roche Urisys Cassette		0	0	0	0	0	13
ARKRAY Aution Sticks		0	0	0	21	0	7
IRIS Reagent Strips		0	0	1	16	0	0
Siemens Hemastix		0	0	0	0	0	1
None, no reagent or strip use		0	0	0	0	0	1
Other		0	0	0	1	0	1
All Instruments and Reagents		0	0	1	57	0	314



**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Hemoglobin	Specimen: U19	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		36	0	0	0	0	0
Siemens Atlas		95	0	0	0	0	0
Siemens Clinitek Status		37	4	0	0	0	0
Siemens Clinitek 50		6	0	0	0	0	0
Siemens Clinitek 100		5	0	0	0	0	0
Siemens Clinitek 500		33	1	1	0	0	0
Siemens Clinitek Advantus		57	0	0	0	0	0
Roche UA Reader/Super UA Read		1	0	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0	0
Roche Urisys		33	0	0	0	0	0
Roche Cobas U411		2	1	0	0	0	0
IRIS AUTION MAX		36	0	0	0	0	0
IRIS iChem Velocity		9	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		171	5	1	0	0	0
Siemens Atlas Reagent		87	0	0	0	0	0
Roche Chemstrips		44	1	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		28	0	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		2	0	0	0	0	0
All Instruments and Reagents		364	6	1	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Hemoglobin	Specimen: U20	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<b>&lt;Instruments&gt;</b>							
Visual examination		36	0	0	0	0	0
Siemens Atlas		95	0	0	0	0	0
Siemens Clinitek Status		40	1	0	0	0	0
Siemens Clinitek 50		6	0	0	0	0	0
Siemens Clinitek 100		5	0	0	0	0	0
Siemens Clinitek 500		35	0	0	0	0	0
Siemens Clinitek Advantus		57	0	0	0	0	0
Roche UA Reader/Super UA Read		1	0	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0	0
Roche Urisys		33	0	0	0	0	0
Roche Cobas U411		3	0	0	0	0	0
IRIS AUTION MAX		35	1	0	0	0	0
IRIS iChem Velocity		9	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
<b>&lt;Reagents&gt;</b>							
Siemens Reagent Strips		176	1	0	0	0	0
Siemens Atlas Reagent		87	0	0	0	0	0
Roche Chemstrips		45	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		27	1	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		2	0	0	0	0	0
All Instruments and Reagents		369	2	0	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Ketones	Specimen: U16	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
	Visual examination	43	0	0	0	0	0
	Siemens Atlas	95	0	0	0	0	0
	Siemens Clinitek Status	41	0	0	0	0	0
	Siemens Clinitek 50	6	0	0	0	0	0
	Siemens Clinitek 100	5	0	0	0	0	0
	Siemens Clinitek 500	35	0	0	0	0	0
	Siemens Clinitek Advantus	57	0	0	0	0	0
	Roche UA Reader/Super UA Read	1	0	0	0	0	0
	Roche Chemstrip Criterion	13	0	0	0	0	0
	Roche Urisys	33	0	0	0	0	0
	Roche Cobas U411	3	0	0	0	0	0
	IRIS AUTION MAX	36	0	0	0	0	0
	IRIS iChem Velocity	9	0	0	0	0	0
<Reagents>							
	Siemens Reagent Strips	179	0	0	0	0	0
	Siemens Atlas Reagent	85	0	0	0	0	0
	Roche Chemstrips	46	0	0	0	0	0
	Roche Urisys Cassette	13	0	0	0	0	0
	ARKRAY Aution Sticks	26	0	0	0	0	0
	IRIS Reagent Strips	18	0	0	0	0	0
	Siemens Acetest	6	0	0	0	0	0
	Bayer Diastix	2	0	0	0	0	0
	Siemens Ketostix	1	0	0	0	0	0
	Other	1	0	0	0	0	0
	All Instruments and Reagents	377	0	0	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Ketones	Specimen: U17	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<Instruments>							
Visual examination		43	0	0	0	0	0
Siemens Atlas		96	0	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		6	0	0	0	0	0
Siemens Clinitek 100		5	0	0	0	0	0
Siemens Clinitek 500		35	0	0	0	0	0
Siemens Clinitek Advantus		57	0	0	0	0	0
Roche UA Reader/Super UA Read		1	0	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0	0
Roche Urisys		33	0	0	0	0	0
Roche Cobas U411		3	0	0	0	0	0
IRIS AUTION MAX		36	0	0	0	0	0
IRIS iChem Velocity		9	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		179	0	0	0	0	0
Siemens Atlas Reagent		86	0	0	0	0	0
Roche Chemstrips		46	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		26	0	0	0	0	0
IRIS Reagent Strips		18	0	0	0	0	0
Siemens Acetest		6	0	0	0	0	0
Bayer Diastix		2	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		378	0	0	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Ketones	Specimen: U18	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<b>&lt;Instruments&gt;</b>							
Visual examination		0	0	1	24	30	4
Siemens Atlas		0	0	0	96	0	0
Siemens Clinitek Status		0	0	0	14	27	0
Siemens Clinitek 50		0	0	0	4	2	0
Siemens Clinitek 100		0	0	0	2	3	0
Siemens Clinitek 500		0	0	0	32	3	0
Siemens Clinitek Advantus		0	0	3	48	6	0
Roche UA Reader/Super UA Read		0	0	0	0	1	0
Roche Chemstrip Criterion		0	0	0	6	7	0
Roche Urisys		0	0	0	10	23	0
Roche Cobas U411		0	0	0	1	2	0
IRIS AUTION MAX		0	0	0	34	2	0
IRIS iChem Velocity		0	0	4	2	3	0
<b>&lt;Reagents&gt;</b>							
Siemens Reagent Strips		0	0	3	114	62	0
Siemens Atlas Reagent		0	0	0	86	0	0
Roche Chemstrips		0	0	0	18	28	0
Roche Urisys Cassette		0	0	0	5	8	0
ARKRAY Aution Sticks		0	0	0	25	1	0
IRIS Reagent Strips		0	0	4	10	4	0
Siemens Acetest		0	0	1	10	6	4
Bayer Diastix		0	0	0	2	0	0
Siemens Ketostix		0	0	0	2	0	0
Other		0	0	0	1	0	0
All Instruments and Reagents		0	0	8	273	109	4

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Ketones	Specimen: U19	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<b>&lt;Instruments&gt;</b>							
Visual examination		44	0	0	0	1	0
Siemens Atlas		95	0	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		6	0	0	0	0	0
Siemens Clinitek 100		5	0	0	0	0	0
Siemens Clinitek 500		35	0	0	0	0	0
Siemens Clinitek Advantus		57	0	0	0	0	0
Roche UA Reader/Super UA Read		1	0	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0	0
Roche Urisys		33	0	0	0	0	0
Roche Cobas U411		3	0	0	0	0	0
IRIS AUTION MAX		36	0	0	0	0	0
IRIS iChem Velocity		9	0	0	0	0	0
<b>&lt;Reagents&gt;</b>							
Siemens Reagent Strips		179	0	0	0	0	0
Siemens Atlas Reagent		85	0	0	0	0	0
Roche Chemstrips		46	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		26	0	0	0	0	0
IRIS Reagent Strips		18	0	0	0	0	0
Siemens Acetest		7	0	0	0	1	0
Bayer Diastix		2	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		378	0	0	0	1	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Ketones	Specimen: U20	Negative	Trace   5mg/dL	Small   1+   15mg/dL	Moderate   2+   40-50mg/dL	Large   3+   >=80mg/dL	Positive
<Instruments>							
Visual examination		44	0	0	0	0	0
Siemens Atlas		95	0	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		6	0	0	0	0	0
Siemens Clinitek 100		5	0	0	0	0	0
Siemens Clinitek 500		35	0	0	0	0	0
Siemens Clinitek Advantus		57	0	0	0	0	0
Roche UA Reader/Super UA Read		1	0	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0	0
Roche Urisys		33	0	0	0	0	0
Roche Cobas U411		3	0	0	0	0	0
IRIS AUTION MAX		36	0	0	0	0	0
IRIS iChem Velocity		9	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		179	0	0	0	0	0
Siemens Atlas Reagent		85	0	0	0	0	0
Roche Chemstrips		46	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		26	0	0	0	0	0
IRIS Reagent Strips		18	0	0	0	0	0
Siemens Acetest		7	0	0	0	0	0
Bayer Diastix		2	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		1	0	0	0	0	0
All Instruments and Reagents		378	0	0	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Protein	Specimen: U16	Negative	Trace   10-15mg/dL	1+   25-30mg/dL	2+   75-150mg/dL	3+   300-600mg/dL	4+   >=1000mg/dL	Positive
<b>&lt;Instruments&gt;</b>								
Visual examination		0	0	0	0	61	11	14
Siemens Atlas		0	0	0	0	86	10	0
Siemens Clinitek Status		0	0	0	0	39	1	0
Siemens Clinitek 50		0	0	0	0	6	0	0
Siemens Clinitek 100		0	0	0	0	5	0	0
Siemens Clinitek 500		0	0	0	16	19	0	0
Siemens Clinitek Advantus		0	0	0	13	45	0	0
Roche UA Reader/Super UA Read		0	0	0	0	1	0	0
Roche Chemstrip Criterion		0	0	0	0	13	0	0
Roche Urisys		0	0	0	1	32	0	0
Roche Cobas U411		0	0	0	0	3	0	0
IRIS AUTION MAX		0	0	0	0	36	0	0
IRIS iChem Velocity		0	0	0	1	7	1	0
Hitachi/Roche Analyzer		0	0	0	0	1	0	0
Other Chemistry Analyzer		0	0	0	0	1	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		0	0	0	28	144	5	0
Siemens Atlas Reagent		0	0	0	0	79	10	0
Sulfosalicylic Acid		0	0	0	0	28	7	13
Roche Chemstrips		0	0	0	1	43	0	0
Roche Urisys Cassette		0	0	0	0	13	0	0
ARKRAY Aution Sticks		0	0	0	0	28	0	0
IRIS Reagent Strips		0	0	0	1	15	1	0
Siemens Uristix		0	0	0	1	2	0	0
Hitachi/Roche Reagents		0	0	0	0	1	0	0
Other		0	0	0	0	2	0	1
All Instruments and Reagents		0	0	0	31	355	23	14



**New York State Department of Health - Wadsworth Center  
Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Protein	Specimen: U17	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<b>&lt;Instruments&gt;</b>								
Visual examination		1	0	2	51	16	0	13
Siemens Atlas		0	0	0	96	0	0	0
Siemens Clinitek Status		0	0	0	36	4	0	0
Siemens Clinitek 50		0	0	0	3	3	0	0
Siemens Clinitek 100		0	0	0	3	2	0	0
Siemens Clinitek 500		1	0	0	34	0	0	0
Siemens Clinitek Advantus		0	0	0	58	0	0	0
Roche UA Reader/Super UA Read		0	0	0	1	0	0	0
Roche Chemstrip Criterion		0	0	0	13	0	0	0
Roche Urisys		0	0	1	32	0	0	0
Roche Cobas U411		0	0	0	3	0	0	0
IRIS AUTION MAX		0	0	3	33	0	0	0
IRIS iChem Velocity		0	0	0	9	0	0	0
Hitachi/Roche Analyzer		0	0	0	1	0	0	0
Other Chemistry Analyzer		0	0	0	1	0	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		2	0	1	158	16	0	0
Siemens Atlas Reagent		0	0	0	89	0	0	0
Sulfosalicylic Acid		0	0	1	27	5	0	13
Roche Chemstrips		0	0	1	39	4	0	0
Roche Urisys Cassette		0	0	0	13	0	0	0
ARKRAY Aution Sticks		0	0	3	25	0	0	0
IRIS Reagent Strips		0	0	0	17	0	0	0
Siemens Uristix		0	0	0	3	0	0	0
Hitachi/Roche Reagents		0	0	0	1	0	0	0
Other		0	0	0	2	0	0	0
All Instruments and Reagents		2	0	6	374	25	0	13

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Protein	Specimen: U18	Negative	Trace   10-15mg/dL	1+   25-30mg/dL	2+   75-150mg/dL	3+   300-600mg/dL	4+   >=1000mg/dL	Positive
<b>&lt;Instruments&gt;</b>								
Visual examination		0	0	1	48	20	1	14
Siemens Atlas		0	0	0	74	22	0	0
Siemens Clinitek Status		0	0	0	11	29	0	0
Siemens Clinitek 50		0	0	0	1	5	0	0
Siemens Clinitek 100		0	0	0	3	2	0	0
Siemens Clinitek 500		0	0	0	35	0	0	0
Siemens Clinitek Advantus		0	0	0	58	0	0	0
Roche UA Reader/Super UA Read		0	0	0	0	1	0	0
Roche Chemstrip Criterion		0	0	0	13	0	0	0
Roche Urisys		0	0	0	33	0	0	0
Roche Cobas U411		0	0	0	3	0	0	0
IRIS AUTION MAX		0	0	0	36	0	0	0
IRIS iChem Velocity		0	0	0	9	0	0	0
Hitachi/Roche Analyzer		0	0	0	1	0	0	0
Other Chemistry Analyzer		0	0	0	1	0	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		0	0	1	131	45	0	0
Siemens Atlas Reagent		0	0	0	67	22	0	0
Sulfosalicylic Acid		0	0	0	25	7	1	13
Roche Chemstrips		0	0	0	40	4	0	0
Roche Urisys Cassette		0	0	0	13	0	0	0
ARKRAY Aution Sticks		0	0	0	28	0	0	0
IRIS Reagent Strips		0	0	0	17	0	0	0
Siemens Uristix		0	0	0	3	0	0	0
Hitachi/Roche Reagents		0	0	0	1	0	0	0
Other		0	0	0	1	1	0	1
All Instruments and Reagents		0	0	1	326	79	1	14

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Protein	Specimen: U19	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<b>&lt;Instruments&gt;</b>								
Visual examination		1	3	45	22	1	0	12
Siemens Atlas		0	0	46	50	0	0	0
Siemens Clinitek Status		0	0	3	37	0	0	0
Siemens Clinitek 50		0	0	3	3	0	0	0
Siemens Clinitek 100		0	0	2	3	0	0	0
Siemens Clinitek 500		0	2	33	0	0	0	0
Siemens Clinitek Advantus		1	0	56	1	0	0	0
Roche UA Reader/Super UA Read		0	0	0	1	0	0	0
Roche Chemstrip Criterion		0	0	11	2	0	0	0
Roche Urisys		0	1	5	27	0	0	0
Roche Cobas U411		0	0	0	3	0	0	0
IRIS AUTION MAX		1	1	34	0	0	0	0
IRIS iChem Velocity		0	0	8	1	0	0	0
Hitachi/Roche Analyzer		0	0	1	0	0	0	0
Other Chemistry Analyzer		0	0	1	0	0	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		1	3	117	55	1	0	0
Siemens Atlas Reagent		0	0	42	47	0	0	0
Sulfosalicylic Acid		1	2	21	11	0	0	12
Roche Chemstrips		0	1	19	24	0	0	0
Roche Urisys Cassette		0	0	1	12	0	0	0
ARKRAY Aution Sticks		1	1	26	0	0	0	0
IRIS Reagent Strips		0	0	16	1	0	0	0
Siemens Uristix		0	0	3	0	0	0	0
Hitachi/Roche Reagents		0	0	1	0	0	0	0
Other		0	0	2	0	0	0	0
All Instruments and Reagents		3	7	248	150	1	0	12

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Protein	Specimen: U20	Negative	Trace   10-15mg/dL	1+   25-30mg/dL	2+   75-150mg/dL	3+   300-600mg/dL	4+   >=1000mg/dL	Positive
<b>&lt;Instruments&gt;</b>								
Visual examination		59	0	0	1	0	0	0
Siemens Atlas		95	0	0	0	0	0	0
Siemens Clinitek Status		40	0	0	0	0	0	0
Siemens Clinitek 50		6	0	0	0	0	0	0
Siemens Clinitek 100		4	0	0	1	0	0	0
Siemens Clinitek 500		35	0	0	0	0	0	0
Siemens Clinitek Advantus		58	0	0	0	0	0	0
Roche UA Reader/Super UA Read		1	0	0	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0	0	0
Roche Urisys		33	0	0	0	0	0	0
Roche Cobas U411		3	0	0	0	0	0	0
IRIS AUTION MAX		36	0	0	0	0	0	0
IRIS iChem Velocity		9	0	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0	0
Other Chemistry Analyzer		1	0	0	0	0	0	0
<b>&lt;Reagents&gt;</b>								
Siemens Reagent Strips		176	0	0	1	0	0	0
Siemens Atlas Reagent		88	0	0	0	0	0	0
Sulfosalicylic Acid		22	0	0	1	0	0	0
Roche Chemstrips		44	0	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0	0
ARKRAY Aution Sticks		28	0	0	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0	0
Siemens Uristix		3	0	0	0	0	0	0
Hitachi/Roche Reagents		1	0	0	0	0	0	0
Other		2	0	0	0	0	0	0
All Instruments and Reagents		394	0	0	2	0	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

**Qualitative hCG report:**

Other	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	10	0	10	10
Indeterminant	0	0	0	0	0
Positive	10	0	10	0	0
Beckman Coulter ICON II	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	2	0	2	2
Indeterminant	0	0	0	0	0
Positive	2	0	2	0	0
Beckman Coulter ICON 25	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	23	0	23	23
Indeterminant	1	0	0	0	0
Positive	22	0	23	0	0
Beckman Coulter ICON 20	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	22	0	22	22
Indeterminant	0	0	0	0	0
Positive	22	0	22	0	0
Cardinal Health hCG Combo	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	26	0	26	26
Indeterminant	0	0	0	0	0
Positive	26	0	26	0	0
Wampole Clearview hCG Combo	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	2	0	2	2
Indeterminant	0	0	0	0	0
Positive	2	0	2	0	0
Cen Med Elite	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	21	0	21	21
Indeterminant	0	0	0	0	0
Positive	21	0	21	0	0

**New York State Department of Health - Wadsworth Center  
Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

Sekisui OSOM Combo (Genzyme)	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	18	1	18	18
Indeterminant	0	0	0	0	0
Positive	18	0	17	0	0
Henry Schein One Step	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	2	0	2	2
Indeterminant	0	0	0	0	0
Positive	2	0	2	0	0
Inverness Medical BioStar Acceava	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	1	0	1	1
Indeterminant	0	0	0	0	0
Positive	1	0	1	0	0
Lifesign Status (serum/urine)	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	1	0	1	1
Indeterminant	0	0	0	0	0
Positive	1	0	1	0	0
McKesson hCG Combo	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	1	0	1	1
Indeterminant	0	0	0	0	0
Positive	1	0	1	0	0
Polymedco Poly stat	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	5	0	5	5
Indeterminant	0	0	0	0	0
Positive	5	0	5	0	0
Quidel QuickVue	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	74	0	74	74
Indeterminant	0	0	0	0	0
Positive	74	0	74	0	0
R. Medical One Step Strip	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	3	0	3	3
Indeterminant	0	0	0	0	0
Positive	3	0	3	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

SA Scientific	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	1	0	1	1
Indeterminant	0	0	0	0	0
Positive	1	0	1	0	0

Stanbio QuPID	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	14	0	14	14
Indeterminant	0	0	0	0	0
Positive	14	0	14	0	0

Fisher Sure-Vue serum/urine	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	38	0	38	38
Indeterminant	0	0	0	0	0
Positive	38	0	38	0	0

Fisher Sure-Vue STAT serum/urine	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	23	0	23	23
Indeterminant	0	0	0	0	0
Positive	23	0	23	0	0

Signify hCG Urine/Serum	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	3	0	3	3
Indeterminant	0	0	0	0	0
Positive	3	0	3	0	0

Fisher Sure-Vue (urine only)	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	11	0	11	11
Indeterminant	0	0	0	0	0
Positive	11	0	11	0	0

True Medix	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	6	0	6	6
Indeterminant	0	0	0	0	0
Positive	6	0	6	0	0

All Methods & Instruments	Specimen:U16	Specimen:U17	Specimen:U18	Specimen:U19	Specimen:U20
Negative	0	307	1	307	307
Indeterminant	1	0	0	0	0
Positive	306	0	306	0	0

**New York State Department of Health - Wadsworth Center**  
**Urinalysis/Urine Pregnancy Proficiency Testing - 28 October 2013**

**hCG Summary:**

[Code], hCG Kit name	U16			U17			U18			U19			U20		
	N	I	P	N	I	P	N	I	P	N	I	P	N	I	P
[ZZZ] Other	0	0	10	10	0	0	0	0	10	10	0	0	10	0	0
[BC ] Beckman Coulter ICON II	0	0	2	2	0	0	0	0	2	2	0	0	2	0	0
[BN ] Beckman Coulter ICON 25	0	1	22	23	0	0	0	0	23	23	0	0	23	0	0
[BT ] Beckman Coulter ICON 20	0	0	22	22	0	0	0	0	22	22	0	0	22	0	0
[CA ] Cardinal Health hCG Combo	0	0	26	26	0	0	0	0	26	26	0	0	26	0	0
[CC ] Wampole Clearview hCG Combo	0	0	2	2	0	0	0	0	2	2	0	0	2	0	0
[CM ] Cen Med Elite	0	0	21	21	0	0	0	0	21	21	0	0	21	0	0
[GO ] Sekisui OSOM Combo (Genzyme)	0	0	18	18	0	0	1	0	17	18	0	0	18	0	0
[HS ] Henry Schein One Step	0	0	2	2	0	0	0	0	2	2	0	0	2	0	0
[IM ] Inverness Medical BioStar Acceava	0	0	1	1	0	0	0	0	1	1	0	0	1	0	0
[LS ] Lifesign Status (serum/urine)	0	0	1	1	0	0	0	0	1	1	0	0	1	0	0
[MK ] McKesson hCG Combo	0	0	1	1	0	0	0	0	1	1	0	0	1	0	0
[PM ] Polymedco Poly stat	0	0	5	5	0	0	0	0	5	5	0	0	5	0	0
[QQ ] Quidel QuickVue	0	0	74	74	0	0	0	0	74	74	0	0	74	0	0
[RM ] R. Medical One Step Strip	0	0	3	3	0	0	0	0	3	3	0	0	3	0	0
[SA ] SA Scientific	0	0	1	1	0	0	0	0	1	1	0	0	1	0	0
[SB ] Stanbio QuPID	0	0	14	14	0	0	0	0	14	14	0	0	14	0	0
[SE ] Fisher Sure-Vue serum/urine	0	0	38	38	0	0	0	0	38	38	0	0	38	0	0
[SF ] Fisher Sure-Vue STAT serum/urine	0	0	23	23	0	0	0	0	23	23	0	0	23	0	0
[SG ] Signify hCG Urine/Serum	0	0	3	3	0	0	0	0	3	3	0	0	3	0	0
[SU ] Fisher Sure-Vue (urine only)	0	0	11	11	0	0	0	0	11	11	0	0	11	0	0
[TX ] True Medix	0	0	6	6	0	0	0	0	6	6	0	0	6	0	0
[99 ] All hCG Methods	0	1	306	307	0	0	1	0	306	307	0	0	307	0	0