



**Department  
of Health**

**Wadsworth  
Center**

# New York State Biomonitoring Program for Trace Elements

**Event #3, 2017**

**Trace Elements in Whole Blood,  
Urine, and Serum**

**December, 2017**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*

**Event #3, 2017:  
Trace Elements in Whole Blood, Urine, and Serum**

12/14/2017

Dear Laboratory Director,

This report summarizes performance for the third biomonitoring proficiency test (PT) event of 2017 for **Trace Elements in Whole Blood, Urine, and Serum**. One of the key goals of this PT program is to achieve harmonization of biomonitoring data for trace elements.

**Target Value Assignment and Performance Evaluation**

For these PT materials, target values have been assigned for a limited number of trace elements that are gradable under criteria set by the NYS DOH Biomonitoring PT program. See assay-specific narratives for details. Data for additional trace elements are reported and are included here in order to characterize the PT materials more completely. Participant data and descriptive statistics are provided for educational purposes. No target value or acceptable range is implied.

Where the data permit, robust statistics were used to assign target values based on Algorithm A as defined by ISO 13528:2005E “*Statistical methods for use in proficiency testing by inter-laboratory comparisons*” [1]. Acceptable ranges for the “graded elements” are based on consensus criteria and/or those set by the NYS DOH’s PT program. For example, some are fixed based on US regulatory guidelines (Pb, Cd) while for other elements the criteria are based on a consensus of the Network of PT scheme organizers for trace elements in occupational and environmental laboratory medicine [2]. Quality specifications are element and matrix specific; full details are provided under each element specific narrative.

A confidential, three-digit code number assigned by PT program staff identifies all laboratory participants.

Samples for the next PT event (Event #1 of 2018) will be shipped February 28, 2018. Comments about this report may be directed to [trel@health.ny.gov](mailto:trel@health.ny.gov). If you have not yet enrolled for 2018 please contact Dr. Galusha at [Aubrey.Galusha@health.ny.gov](mailto:Aubrey.Galusha@health.ny.gov).

Sincerely,



Patrick J. Parsons, PhD  
Chief, Inorganic and Nuclear Chemistry  
Division of Environmental Sciences  
Wadsworth Center



Aubrey L. Galusha, PhD  
Coordinator, Biomonitoring PT Program,  
Division of Environmental Sciences  
Wadsworth Center



**Department  
of Health**

Wadsworth  
Center

## **Event #3, 2017**

# **Trace Elements in Whole Blood**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*

## **Event #3, 2017: Trace Elements in Whole Blood**

### **PT Materials**

Human whole blood was purchased from ZenBio, Inc. and preserved with K<sub>2</sub>EDTA. The company certifies that this material was "non-reactive" for HBsAg, HBV DNA, HIV-1,2 Ab, HIV-1 RNA, HCV Ab, HCV RNA, and STS. Units of whole blood were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with arsenic (As), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), barium (Ba), beryllium (Be), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), titanium (Ti), thallium (Tl), uranium (U), vanadium (V), tungsten (W), and zinc (Zn). Whole blood samples were placed on a rocker overnight prior to aliquoting 2-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### **Graded Elements**

Seven elements in whole blood are formally graded: As, Cd, Co, Cr, Hg, Mn and Pb. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) where a robust mean is not possible, the arithmetic mean after outlier deletion.

### **Additional Elements**

An additional 27 elements (beyond the seven graded) were reported by at least one participant: Ag, Al, Ba, Be, Bi, Ca, Cs, Cu, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, Tl, U, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.

## Results for Event #3, 2017: Summary Statistics

Whole Blood As ( $\mu\text{g/L}$ )					
	BE17-11	BE17-12	BE17-13	BE17-14	BE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	6.8	11.2	25.1	44.4	13.1
<b>Upper Limit</b>	12.8	17.2	31.1	53.28	19.1
<b>Lower Limit</b>	0.8	5.2	19.1	35.52	7.1
<b>Arithmetic SD (s)</b>	1.4	1.6	4.7	7.6	2.3
<b>Arithmetic RSD (%)</b>	20.6	14.3	18.7	17.1	17.6
<b>Number of Sample Measurements (N)</b>	7	8	8	8	8

The acceptable range is based on quality specifications:

$\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Whole Blood As (<math>\mu\text{g/L}</math>)</b>				
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
	<b>Target</b>	<b>6.8</b>	<b>11.2</b>	<b>25.1</b>	<b>44.4</b>	<b>13.1</b>
103	DRC/CC-ICP-MS	6.20	10.9	24.2	44.0	12.9
110	DRC/CC-ICP-MS	6.1	10.3	23.2	44.0	12.6
147	ICP-MS	6.46	10.5	22.3	39.6	11.7
200	ICP-MS	9.0	12.8	29.3	46.5	15.8
293	ICP-MS	NR	11.6	25.55	47.05	13.54
391	DRC/CC-ICP-MS	4.7	8.2	16.375 ↓	28.985 ↓	8.375
597	DRC/CC-ICP-MS	7.26	12.2	28.7	52.8	14.2
598	DRC/CC-ICP-MS	7.57	13.4	31.3 ↑	52.4	15.3

Based on the grading criteria for As in Whole Blood, 93% of results were satisfactory, with 1 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Whole Blood As



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 6 \mu\text{g}/\text{L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $30 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Whole Blood Cd (<math>\mu\text{g}/\text{L}</math>)</b>				
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Target (Robust Mean (<math>x^*</math>))</b>	1.1	2.7	6.2	6.4	9.7
<b>Upper Limit</b>	2.1	3.7	7.2	7.4	11.2
<b>Lower Limit</b>	0.1	1.7	5.2	5.4	8.2
<b>Robust SD (<math>s^*</math>)</b>	0.1	0.2	0.4	0.4	0.7
<b>Robust RSD (%)</b>	9.1	7.4	6.5	6.3	7.2
<b>Number of Sample Measurements (N)</b>	15	16	16	16	16
<b>Standard Uncertainty (<math>u</math>)</b>	0.046	0.057	0.123	0.136	0.226

The acceptable range is based on quality specifications:

$\pm 1 \mu\text{g}/\text{L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $6.7 \mu\text{g}/\text{L}$ . These quality specifications are based on those used by US OSHA for occupational exposure.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Whole Blood Cd (<math>\mu\text{g/L}</math>)</b>				
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
		<b>Target</b>	<b>1.1</b>	<b>2.7</b>	<b>6.2</b>	<b>6.4</b>
103	DRC/CC-ICP-MS	1.11	2.67	6.34	6.49	10.1
107	ICP-MS	1.2	2.7	6.1	6.7	9.7
110	ICP-MS	1.04	2.55	5.98	6.24	9.63
116	DRC/CC-ICP-MS	<1.06	2.73	6.08	6.67	9.32
147	ICP-MS	1.04	2.45	5.55	5.95	8.95
200	ICP-MS	2.1 ↑	4.0 ↑	8.8 ↑	8.2 ↑	12.0 ↑
293	ICP-MS	1.14	2.55	5.97	6.25	9.55
391	DRC/CC-ICP-MS	0.681	1.577 ↓	3.746 ↓	3.833 ↓	5.826 ↓
399	DRC/CC-ICP-MS	1.08	2.59	5.98	6.25	9.49
401	DRC/CC-ICP-MS	1.1	2.7	6.5	6.7	10.6
597	DRC/CC-ICP-MS	1.39	3.27	7.60 ↑	7.96 ↑	11.69 ↑
598	DRC/CC-ICP-MS	1.01	2.64	6.36	6.17	9.23
599	DRC/CC-ICP-MS	1.39	2.72	7.69 ↑	5.62	5.96 ↓
605	ICP-MS	1.23	2.78	6.03	6.25	9.80
606	DRC/CC-ICP-MS	1.34	2.87	6.48	6.57	10.3
686	ICP-MS	1.05	2.49	6.07	6.28	9.69

Based on the grading criteria for Cd in Whole Blood, 83% of results were satisfactory, with 4 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Whole Blood Cd



### Legend:

○ O/C/HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 1 \mu\text{g}/\text{L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $6.7 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Summary Statistics

	Whole Blood Co ( $\mu\text{g/L}$ )				
	BE17-11	BE17-12	BE17-13	BE17-14	BE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	5.8	0.8	10.9	1.4	7.6
<b>Upper Limit</b>	7.3	2.3	13.08	2.9	9.12
<b>Lower Limit</b>	4.3	0	8.72	0	6.08
<b>Arithmetic SD (s)</b>	0.5	0.1	2.0	0.2	1.2
<b>Arithmetic RSD (%)</b>	8.6	12.5	18.3	14.3	15.8
<b>Number of Sample Measurements (N)</b>	8	9	9	8	8

The acceptable range is based on quality specifications:

$\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Whole Blood Co (<math>\mu\text{g/L}</math>)</b>				
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
		<b>Target</b>	<b>5.8</b>	<b>0.8</b>	<b>10.9</b>	<b>1.4</b>
103	DRC/CC-ICP-MS	5.43	0.753	10.7	1.34	7.58
110	ICP-MS	5.5	0.8	11.1	1.3	7.9
147	ICP-MS	5.56	0.790	10.7	1.30	7.42
293	ICP-MS	5.42	0.71	10.75	1.35	7.56
391	DRC/CC-ICP-MS	*0.352 ↓	0.482	6.605 ↓	0.955	4.962 ↓
401	DRC/CC-ICP-MS	5.5	0.7	10.3	1.4	7.9
597	DRC/CC-ICP-MS	6.83	0.959	13.8 ↑	1.76	9.20 ↑
598	ICP-MS	5.64	0.97	11.9	1.57	8.02
599	DRC/CC-ICP-MS	6.12	0.872	12.6	*13.7 ↑	*1.91 ↓






Based on the grading criteria for Co in Whole Blood, 84% of results were satisfactory, with 3 of the 9 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier



## Results for Event #3, 2017: Summary Figures

### Whole Blood Co



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
 $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

Whole Blood Cr ( $\mu\text{g/L}$ )					
	BE17-11	BE17-12	BE17-13	BE17-14	BE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	2.7	2	11.4	5.2	7.1
<b>Upper Limit</b>	4.7	4	13.68	7.2	9.1
<b>Lower Limit</b>	0.7	0	9.12	3.2	5.1
<b>Arithmetic SD (s)</b>	1.0	0.9	0.6	0.5	0.8
<b>Arithmetic RSD (%)</b>	37.0	45.0	5.3	9.6	11.3
<b>Number of Sample Measurements (N)</b>	8	7	7	7	8

The acceptable range is based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Whole Blood Cr (<math>\mu\text{g/L}</math>)</b>				
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
	<b>Target</b>	<b>2.7</b>	<b>2</b>	<b>11.4</b>	<b>5.2</b>	<b>7.1</b>
103	DRC/CC-ICP-MS	2.34	1.31	11.1	4.99	6.91
110	DRC/CC-ICP-MS	2.2	1.2	10.9	4.6	6.3
147	DRC/CC-ICP-MS	2.51	1.43	11.4	5.07	6.92
293	ICP-MS	2.75	1.77	11.6	5.55	7.3
391	DRC/CC-ICP-MS	4.66	3.36	*8.8 ↓	*9.089 ↑	8.5
401	DRC/CC-ICP-MS	3.4	3.1	10.8	6.0	7.9
597	DRC/CC-ICP-MS	1.26	*6.04 ↑	12.2	4.68	6.32
598	DRC/CC-ICP-MS	2.33	1.73	12.1	5.22	6.85

Based on the grading criteria for Cr in Whole Blood, 93% of results were satisfactory, with 1 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Whole Blood Cr



#### Legend:

○ C/HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±2 µg/L at concentrations less than or equal to 10 µg/L.

## Results for Event #3, 2017: Summary Statistics

	Whole Blood Hg ( $\mu\text{g}/\text{L}$ )				
	BE17-11	BE17-12	BE17-13	BE17-14	BE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	5.9	2.6	22.2	13.6	10.7
<b>Upper Limit</b>	8.9	5.6	28.9	17.7	13.9
<b>Lower Limit</b>	2.9	0	15.5	9.5	7.5
<b>Robust SD (<math>s^*</math>)</b>	0.6	0.2	1.1	0.8	0.6
<b>Robust RSD (%)</b>	10.2	7.7	5.0	5.9	5.6
<b>Number of Sample Measurements (N)</b>	15	15	15	15	15
<b>Standard Uncertainty (<math>u</math>)</b>	0.190	0.071	0.339	0.273	0.200

The acceptable range is based on quality specifications:

$\pm 3 \mu\text{g}/\text{L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $10 \mu\text{g}/\text{L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Whole Blood Hg (µg/L)</b>				
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
	<b>Target</b>	<b>5.9</b>	<b>2.6</b>	<b>22.2</b>	<b>13.6</b>	<b>10.7</b>
103	DRC/CC-ICP-MS	5.90	2.55	22.4	13.5	10.6
107	DRC/CC-ICP-MS	6.1	2.6	23	14	11
110	ICP-MS	5.41	2.42	21.9	12.3	10.2
116	DRC/CC-ICP-MS	6.27	2.80	23.9	13.9	10.9
147	ICP-MS	5.28	2.35	19.9	11.9	9.51
200	ICP-MS	6.8	3.0	27.0	15.8	13.0
293	ICP-MS	5.51	2.77	22.1	12.87	11.16
391	CV-AAS	7.02	2.67	16.4	23.1 ↑	10.91
399	DRC/CC-ICP-MS	5.74	2.57	22.4	13.5	10.5
401	DRC/CC-ICP-MS	5.8	2.4	20.9	12.8	9.8
598	ICP-MS	5.51	2.90	22.7	13.2	10.3
599	DRC/CC-ICP-MS	4.65	1.28	20.8	21.7 ↑	12.6
605	ICP-MS	6.28	2.67	21.7	13.4	10.7
606	DRC/CC-ICP-MS	6.29	2.57	22.7	13.6	10.7
686	ICP-MS	6.22	2.67	23.0	14.0	11.1

Based on the grading criteria for Hg in Whole Blood, 97% of results were satisfactory, with 0 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Whole Blood Hg



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

	Whole Blood Mn ( $\mu\text{g/L}$ )				
	BE17-11	BE17-12	BE17-13	BE17-14	BE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	7.1	25.9	18.9	39.5	9.2
<b>Upper Limit</b>	10.1	30.3	22.1	46	12.2
<b>Lower Limit</b>	4.1	21.5	15.7	33	6.2
<b>Robust SD (<math>s^*</math>)</b>	1.0	1.8	1.4	3.1	0.7
<b>Robust RSD (%)</b>	14.1	6.9	7.4	7.8	7.6
<b>Number of Sample Measurements (N)</b>	12	12	12	12	11
<b>Standard Uncertainty (<math>u</math>)</b>	0.374	0.655	0.495	1.110	0.281

The acceptable range is based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ . These quality specifications were recently proposed by a network of Trace Element PT program organizers (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry and Laboratory Medicine. 2016 In press.)

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Whole Blood Mn (<math>\mu\text{g/L}</math>)</b>				
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
	<b>Target</b>	<b>7.1</b>	<b>25.9</b>	<b>18.9</b>	<b>39.5</b>	<b>9.2</b>
103	DRC/CC-ICP-MS	7.14	26.5	19.2	41.3	9.30
107	DRC/CC-ICP-MS	7.7	27	19	42	9.7
110	ICP-MS	5.3	23.3	18.0	37.8	9.7
147	ICP-MS	7.61	24.5	18.3	37.4	9.04
293	ICP-MS	6.12	24.25	17.7	38.43	7.76
391	DRC/CC-ICP-MS	15.254 ↑	25.309	16.1	36.379	11.943
399	DRC/CC-ICP-MS	7.05	25.7	18.6	40.5	8.86
401	DRC/CC-ICP-MS	6.5	24.6	18.0	38.7	8.4
597	DRC/CC-ICP-MS	6.51	27.6	20.8	45.3	7.69
598	ICP-MS	6.20	25.4	19.6	40.1	9.29
599	DRC/CC-ICP-MS	10.3 ↑	29.1	23.0 ↑	25.9 ↓	*57.1 ↑
606	DRC/CC-ICP-MS	7.73	28.1	20.7	43.1	9.86

Based on the grading criteria for Mn in Whole Blood, 92% of results were satisfactory, with 1 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Whole Blood Mn



#### Legend:

○ C/HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 3 \text{ } \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \text{ } \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \text{ } \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

	Whole Blood Pb ( $\mu\text{g/dL}$ )				
	BE17-11	BE17-12	BE17-13	BE17-14	BE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	0.97	51.9	5.3	9.54	15.8
<b>Upper Limit</b>	2.97	57.1	7.3	11.5	17.8
<b>Lower Limit</b>	0	46.7	3.3	7.5	13.8
<b>Robust SD (<math>s^*</math>)</b>	0.10	2.9	0.28	0.55	0.7
<b>Robust RSD (%)</b>	10.3	5.6	5.3	5.8	4.4
<b>Number of Sample Measurements (N)</b>	14	17	17	17	17
<b>Standard Uncertainty (<math>u</math>)</b>	0.033	0.877	0.085	0.165	0.212

The acceptable range is based on quality specifications:

$\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/dL}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ . These quality specifications are recommended by the Clinical Laboratory Standards Institute (CLSI, C40-A2) and have been proposed for use in proficiency testing programs approved under CLIA by the Centers for Medicare and Medicaid Services (CMS) in the USA. (<http://shop.clsi.org/C40.html>)

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Whole Blood Pb (<math>\mu\text{g/dL}</math>)</b>				
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
		<b>Target</b>	<b>0.97</b>	<b>51.9</b>	<b>5.3</b>	<b>9.54</b>
103	DRC/CC-ICP-MS	1.00	53.9	5.53	10.2	16.8
107	ICP-MS	0.92	52	5.4	9.9	16
110	ICP-MS	0.94	51.1	5.34	9.75	16.1
116	DRC/CC-ICP-MS	<0.948	51.6	5.34	9.80	16.3
147	ICP-MS	0.951	49.7	5.14	9.18	15.3
200	ICP-MS	0.8	48.4	5.2	9.1	14.9
293	ICP-MS	1.03	52.94	5.58	9.93	16.34
343	ASV-LeadCare	<1.9	54.7	3.8	9.0	15.6
391	ETAAS-Z	0.92	41.45 ↓	3.795	6.98 ↓	12.817 ↓
399	DRC/CC-ICP-MS	0.964	52.4	5.41	10.0	16.3
401	DRC/CC-ICP-MS	0.8	49.3	5.0	9.3	15.1
597	DRC/CC-ICP-MS	1.30	61.1 ↑	6.68	12.2 ↑	18.4 ↑
598	ICP-MS	0.877	48.8	5.25	9.08	14.7
599	DRC/CC-ICP-MS	1.14	57.5 ↑	5.95	6.35 ↓	11.1 ↓
605	ICP-MS	1.00	53.4	5.09	9.68	16.2
606	DRC/CC-ICP-MS	1.09	52.0	5.23	9.43	15.8
686	ICP-MS	<1.0	52.1	5.32	9.82	16.2






Based on the grading criteria for Pb in Whole Blood, 89% of results were satisfactory, with 3 of the 17 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier



## Results for Event #3, 2017: Summary Figures

### Whole Blood Pb



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±2 μg/dL or ±10% around the target value, whichever is greater; thus, it is fixed at ±2 μg/dL at concentrations less than or equal to 20 μg/dL.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Ba (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	ICP-MS	3.6	2.1	10.9	6.3	0.5
147	ICP-MS	3.71	1.95	10.6	5.56	0.720
597	DRC/CC-ICP-MS	4.36	2.41	13.2	7.53	0.75
598	ICP-MS	4.08	3.49	12.20	6.31	1.40
599	DRC/CC-ICP-MS	3.84	1.95	11.27	9.03	*6.83






  

<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	3.9	2.4	11.6	6.9	0.8
<b>Arithmetic SD (s)</b>	0.3	0.6	1.1	1.4	0.4
<b>Arithmetic RSD (%)</b>	7.7	25	9.5	20.3	50
<b>Number of Sample Measurements (N)</b>	5	5	5	5	4

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Whole Blood Ba



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Cu (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	ICP-MS	901	2220	1710	740	2970
147	ICP-MS	947	2224	1677	731	2891
597	DRC/CC-ICP-MS	1170	2780	2150	961	3590
598	ICP-MS	868	2128	1693	719	2740
599	DRC/CC-ICP-MS	992	2393	1938	*2076	*987






  

<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	976	2349	1834	788	3048
<b>Arithmetic SD (s)</b>	118	259	207	116	374
<b>Arithmetic RSD (%)</b>	12.1	11	11.3	14.7	12.3
<b>Number of Sample Measurements (N)</b>	5	5	5	4	4

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Whole Blood Cu



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Sb (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
103	DRC/CC-ICP-MS	<0.0750	<0.258	2.68	5.65	1.15
110	ICP-MS	0.01	0.27	2.80	5.97	1.21
147	ICP-MS	< 0.0645	0.201	2.46	5.13	1.02
293	ICP-MS	0.04	0.21	2.56	5.13	1.14
597	DRC/CC-ICP-MS	0.076	0.326	3.24	6.60	1.29
598	ICP-MS	0.07	0.32	3.27	6.33	1.30






<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	00	0.27	2.84	5.8	1.19
<b>Arithmetic SD (s)</b>	00	0.06	0.34	0.61	0.11
<b>Arithmetic RSD (%)</b>	?	22.2	12	10.5	9.2
<b>Number of Sample Measurements (N)</b>	NA	5	6	6	6

\*Denotes a statistical Outlier.

Statistics for BE17-11 were not calculated based on a lack of consensus among participating laboratories.

## Results for Event #3, 2017: Summary Figures

### Whole Blood Sb



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Se (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
103	DRC/CC-ICP-MS	207	180	329	381	280
107	DRC/CC-ICP-MS	190	180	320	370	280
110	DRC/CC-ICP-MS	178	158	292	347	253
147	ICP-MS	188	164	290	325	246
399	DRC/CC-ICP-MS	185	165	304	355	259
401	DRC/CC-ICP-MS	176	159	274	328	242
597	DRC/CC-ICP-MS	238	217	398	471	320
598	DRC/CC-ICP-MS	216	193	375	428	313






  

<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	197	177	323	376	274
<b>Arithmetic SD (s)</b>	21	20	43	51	30
<b>Arithmetic RSD (%)</b>	10.7	11.3	13.3	13.6	10.9
<b>Number of Sample Measurements (N)</b>	8	8	8	8	8

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Whole Blood Se



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Tl (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
103	DRC/CC-ICP-MS	0.666	0.0600	2.17	0.793	0.140
110	ICP-MS	0.70	0.03	2.12	0.80	0.13
147	ICP-MS	0.658	< 0.0348	2.02	0.754	0.130
293	ICP-MS	0.65	0.02	2.07	0.78	0.14
597	DRC/CC-ICP-MS	0.749	0.074	2.44	0.924	0.175
598	ICP-MS	0.70	*0.38	2.38	0.91	0.23






<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	0.69	00	2.2	0.83	0.16
<b>Arithmetic SD (s)</b>	0.04	00	0.17	0.07	0.04
<b>Arithmetic RSD (%)</b>	5.8	?	7.7	8.4	25
<b>Number of Sample Measurements (N)</b>	6	NA	6	6	6

\*Denotes a statistical Outlier.

Statistics for BE17-12 were not calculated based on a lack of consensus among participating laboratories.

## Results for Event #3, 2017: Summary Figures

### Whole Blood Tl



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood V (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	DRC/CC-ICP-MS	2.1	1.4	5.3	0.6	0.5
147	DRC/CC-ICP-MS	2.40	1.36	5.41	0.607	0.332
293	ICP-MS	2.4	1.35	5.56	0.68	0.45
597	DRC/CC-ICP-MS	2.49	1.36	6.27	0.690	0.379
598	DRC/CC-ICP-MS	2.26	1.40	6.04	0.73	0.34






<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	2.33	1.37	5.72	0.66	0.4
<b>Arithmetic SD (s)</b>	0.15	0.02	0.42	0.06	0.07
<b>Arithmetic RSD (%)</b>	6.4	1.5	7.3	9.1	17.5
<b>Number of Sample Measurements (N)</b>	5	5	5	5	5

\*Denotes a statistical Outlier.



## Results for Event #3, 2017: Summary Figures

### Whole Blood V



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Zn (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	ICP-MS	5090	5940	8030	6130	6490
147	ICP-MS	4954	5667	7516	5621	6020
597	DRC/CC-ICP-MS	6520	7340	10100	7800	7820
598	ICP-MS	5018	5679	7761	6155	5851
599	DRC/CC-ICP-MS	5816	6519	9264	9965	8475






  

<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	5480	6229	8534	7134	6931
<b>Arithmetic SD (s)</b>	678	711	1104	1783	1158
<b>Arithmetic RSD (%)</b>	12.4	11.4	12.9	25	16.7
<b>Number of Sample Measurements (N)</b>	5	5	5	5	5

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Whole Blood Zn



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Be (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	ICP-MS	1.08	0.77	4.37	3.18	2.25
147	ICP-MS	<1.17	<1.17	4.38	3.02	1.87
293	ICP-MS	1.1	0.7	4.2	2.9	1.9
598	ICP-MS	1.33	1.21	4.46	3.35	2.47
<b>Summary Statistics</b>						
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		1.2	0.9	4.4	3.1	2.1
<b>Arithmetic SD (s)</b>		0.1	0.3	0.1	0.2	0.3
<b>Arithmetic RSD (%)</b>		8.3	33.3	2.3	6.5	14.3
<b>Number of Sample Measurements (N)</b>		3	3	4	4	4

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Cs (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	ICP-MS	1.8	2.3	1.6	2.1	1.8
597	DRC/CC-ICP-MS	2.12	2.41	1.90	2.52	1.95
598	ICP-MS	1.79	2.26	1.78	2.17	1.81

<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	1.9	2.3	1.8	2.3	1.9
<b>Arithmetic SD (s)</b>	0.2	0.1	0.2	0.2	0.1
<b>Arithmetic RSD (%)</b>	10.5	4.3	11.1	8.7	5.3
<b>Number of Sample Measurements (N)</b>	3	3	3	3	3

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Mo (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
103	DRC/CC-ICP-MS	2.87	7.76	13.0	0.224	3.86
147	ICP-MS	2.92	7.24	12.2	0.368	3.66
597	DRC/CC-ICP-MS	3.30	8.90	15.8	0.523	4.50
598	DRC/CC-ICP-MS	3.84	9.14	16.29	0.87	5.11
<b>Summary Statistics</b>						
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		3.23	8.26	14.3	NA	4.28
<b>Arithmetic SD (s)</b>		0.45	0.91	2.0	NA	0.66
<b>Arithmetic RSD (%)</b>		13.9	11.0	14.0	NA	15.4
<b>Number of Sample Measurements (N)</b>		4	4	4	NA	4

\*Denotes a statistical Outlier.

Statistics for BE17-14 were not calculated based on a lack of consensus among participating laboratories.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Ni (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	DRC/CC-ICP-MS	5.5	4.1	10.2	2.1	6.4
147	ICP-MS	5.93	4.13	10.5	1.66	6.40
598	ICP-MS	4.70	*8.23	11.7	1.33	6.84
599	DRC/CC-ICP-MS	6.57	3.29	11.6	*13.5	*2.70
<b>Summary Statistics</b>						
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		5.7	3.8	11	1.7	6.6
<b>Arithmetic SD (s)</b>		0.8	0.5	0.8	0.4	0.3
<b>Arithmetic RSD (%)</b>		14	13.2	7.3	23.5	4.5
<b>Number of Sample Measurements (N)</b>		4	3	4	3	3

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Pt (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	ICP-MS	0.5	7.2	2.2	1.3	0.3
598	ICP-MS	0.78	7.02	2.54	1.35	0.51
<b>Summary Statistics</b>						
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.64	7.1	2.4	1.3	0.41
<b>Arithmetic SD (s)</b>		0.20	0.1	0.2	<0.1	0.15
<b>Arithmetic RSD (%)</b>		31.3	1.4	8.3	<0.1	36.6
<b>Number of Sample Measurements (N)</b>		2	2	2	2	2

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood Sn (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	ICP-MS	1.5	3.2	5.8	2.4	0.8
147	ICP-MS	1.25	2.87	5.19	2.02	0.591
597	DRC/CC-ICP-MS	1.71	3.79	6.83	2.91	0.837
598	ICP-MS	1.73	3.96	7.06	2.91	1.15
<b>Summary Statistics</b>						
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		1.5	3.5	6.2	2.6	0.8
<b>Arithmetic SD (s)</b>		0.2	0.5	0.9	0.4	0.2
<b>Arithmetic RSD (%)</b>		13.3	14.3	14.5	15.4	25.0
<b>Number of Sample Measurements (N)</b>		4	4	4	4	4

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood U (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
103	DRC/CC-ICP-MS	0.0587	0.0193	0.107	0.0714	0.0823
110	ICP-MS	0.060	0.024	0.125	0.079	0.086
147	ICP-MS	0.0540	0.0219	0.112	0.0645	0.0786
598	ICP-MS	*0.10	*0.14	*0.14	*0.09	*0.11

<b>Summary Statistics</b>					
	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	0.058	0.022	0.110	0.072	0.082
<b>Arithmetic SD (s)</b>	0.003	0.002	0.009	0.007	0.004
<b>Arithmetic RSD (%)</b>	5.5	10.8	8.1	10.1	4.5
<b>Number of Sample Measurements (N)</b>	3	3	3	3	3

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Whole Blood W (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
110	ICP-MS	0.30	0.05	1.17	0.20	0.41
200	ICP-MS	0.39	0.08	1.25	0.23	0.48
598	ICP-MS	*1.00	*0.70	1.90	*0.60	*0.90
<b>Summary Statistics</b>						
		<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.35	0.07	1.4	0.22	0.45
<b>Arithmetic SD (s)</b>		0.06	0.02	0.4	0.02	0.05
<b>Arithmetic RSD (%)</b>		17.1	28.6	27.8	9.1	11.1
<b>Number of Sample Measurements (N)</b>		2	2	3	2	2

\*Denotes a statistical Outlier.



## Results for Event #3, 2017: Additional Elements in Whole Blood

<b>Whole Blood Ag (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
147	ICP-MS	< 0.205	< 0.205	< 0.205	< 0.205	< 0.205
<b>Whole Blood Al (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
147	ICP-MS	<9.17	<9.17	<9.17	<9.17	<9.17
597	DRC/CC-ICP-MS	16.4	16.7	14.2	19.0	20.6
<b>Whole Blood Bi (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
147	ICP-MS	< 0.0836	< 0.0836	< 0.0836	< 0.0836	0.196
<b>Whole Blood I (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
147	ICP-MS	31.8	28.2	29.1	20.9	30.8
<b>Whole Blood Li (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
147	ICP-MS	0.840	2.53	0.434	0.701	0.355
<b>Whole Blood Mg (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
597	DRC/CC-ICP-MS	39100	35400	35900	39600	39400
<b>Whole Blood Sr (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
103	DRC/CC-ICP-MS	15.3	37.6	17.2	21.7	16.8
<b>Whole Blood Te (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
147	ICP-MS	< 0.128	< 0.128	< 0.128	< 0.128	< 0.128
<b>Whole Blood Th (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
147	ICP-MS	< 0.0148	< 0.0148	< 0.0148	< 0.0148	< 0.0148
<b>Whole Blood Ti (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>BE17-11</b>	<b>BE17-12</b>	<b>BE17-13</b>	<b>BE17-14</b>	<b>BE17-15</b>
200	DRC/CC-ICP-MS	5.5	1.7	10.3	3.5	2.2



**Department  
of Health**

Wadsworth  
Center

## **Event #3, 2017**

# **Trace Elements in Urine**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*

### Event #3, 2017: Trace Elements in Urine

#### PT Materials

Urine was collected from volunteer donors into polyethylene containers and stored at 4°C. Following collection, urine was acidified to 1% (v/v) with nitric acid and mixed with a sulfamic acid solution (stock solution contained 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100) to a final concentration of 1% (v/v) to stabilize Hg. Urine was stored frozen at -80°C pending further preparation. The urine was thawed at room temperature and precipitated salts removed by centrifugation. Urine supernatants were combined into five separate pools. Each urine pool was supplemented with arsenic (As), barium (Ba), beryllium (Be), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), thallium (Tl), uranium (U), aluminum (Al), cesium (Cs), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), strontium (Sr), tellurium (Te), vanadium (V), tungsten (W), and zinc (Zn) and stirred overnight to ensure thorough mixing prior to aliquoting 10-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

#### Graded Elements

Eleven elements in urine are formally graded: As, Ba, Be, Cd, Co, Cr, Hg, Mn, Pb, Tl, and U. Target values for the graded elements are assigned to these pools based on the robust mean calculated from data reported by all laboratories.

#### Additional Elements

An additional 24 elements (beyond the eleven graded) were reported by at least one participant: Ag, Al, B, Bi, Cs, Cu, Fe, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, V, W, Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.

## Results for Event #3, 2017: Summary Statistics

	<b>Urine As (<math>\mu\text{g/L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	85.5	225	20.8	12	68.1
<b>Upper Limit</b>	103	270	26.8	18	82
<b>Lower Limit</b>	68	180	14.8	6	54
<b>Robust SD (<math>s^*</math>)</b>	4.9	14	1.6	1.1	4.0
<b>Robust RSD (%)</b>	5.7	6.2	7.7	9.2	5.9
<b>Number of Sample Measurements (N)</b>	21	21	21	21	21
<b>Standard Uncertainty (<math>u</math>)</b>	1.350	3.730	0.442	0.288	1.100

The acceptable range is based on quality specifications:

$\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine As (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
		<b>Target</b>	<b>85.5</b>	<b>225</b>	<b>20.8</b>	<b>12</b>
103	DRC/CC-ICP-MS	88.0	224	20.7	12.1	68.0
107	DRC/CC-ICP-MS	83	220	20	11	65
110	DRC/CC-ICP-MS	92.9	235	22.1	12.7	72.2
116	DRC/CC-ICP-MS	88.2	229	20.8	11.8	67.7
147	ICP-MS	83.1	210	19.9	11.2	65.5
200	ICP-MS	84.0	230.0	25.5	18.8 ↑	80.3
293	ICP-MS	83.5	219.2	19.8	11.6	68.6
324	ICP-MS	88.48	226.89	20.82	12.00	69.12
391	DRC/CC-ICP-MS	75.2	194.2	19.0	10.5	57.9
399	DRC/CC-ICP-MS	86.8	218	21.0	11.9	67.1
401	DRC/CC-ICP-MS	92.3	230	21.0	12.7	69.8
597	DRC/CC-ICP-MS	82.1	264	21.2	12.5	63.2
598	DRC/CC-ICP-MS	86.3	242.1	22.8	12.4	73.4
599	DRC/CC-ICP-MS	52.6 ↓	160.4 ↓	10.0 ↓	4.19 ↓	54.8
604	DRC/CC-ICP-MS	77.9	209	19.0	11.1	65.9
605	ICP-MS	81.3	211	19.6	11.2	65.2
606	DRC/CC-ICP-MS	91.9	251	22.2	12.7	73.1
630	ICP-MS	82.1	222.0	20.2	12.1	68.5
632	DRC/CC-ICP-MS	88.2	226	22.5	14.5	70.5
684	DRC/CC-ICP-MS	85.6	230.0	19.6	11.6	66.8
686	DRC/CC-ICP-MS	92.7	245	23.3	14.7	72.7






Based on the grading criteria for As in Urine, 95% of results were satisfactory, with 1 of the 21 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier



## Results for Event #3, 2017: Summary Figures

### Urine As



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Urine Ba (<math>\mu\text{g}/\text{L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	10.7	3.3	7.7	1.6	5.9
<b>Upper Limit</b>	12.8	4.3	9.2	2.6	7.1
<b>Lower Limit</b>	8.6	2.3	6.2	0.6	4.7
<b>Robust SD (<math>s^*</math>)</b>	0.6	0.2	0.3	0.1	0.4
<b>Robust RSD (%)</b>	5.6	6.1	3.9	6.3	6.8
<b>Number of Sample Measurements (N)</b>	13	13	13	13	13
<b>Standard Uncertainty (<math>u</math>)</b>	0.209	0.074	0.120	0.041	0.124

The acceptable range is based on quality specifications:

$\pm 1 \mu\text{g}/\text{L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $5 \mu\text{g}/\text{L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine Ba (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>10.7</b>	<b>3.3</b>	<b>7.7</b>	<b>1.6</b>	<b>5.9</b>
107	ICP-MS	11	3.3	7.8	1.6	5.9
110	ICP-MS	10.9	3.2	7.5	1.6	5.8
116	DRC/CC-ICP-MS	10.4	3.12	7.47	1.49	5.79
147	ICP-MS	10.60	3.20	7.72	1.57	5.88
399	ICP-MS	10.8	3.29	7.82	1.64	6.11
597	DRC/CC-ICP-MS	9.64	3.57	7.48	1.47	5.16
598	ICP-MS	10.4	3.63	8.32	1.75	6.46
599	DRC/CC-ICP-MS	8.79	3.10	6.32	2.54	5.46
605	ICP-MS	10.30	3.13	7.48	1.52	5.74
606	DRC/CC-ICP-MS	10.9	3.64	7.36	1.21	5.70
607	ICP-MS	11.4	3.32	7.87	1.63	6.13
632	ICP-MS	11.5	3.45	7.94	1.59	6.29
686	ICP-MS	11.4	3.37	8.13	1.69	6.29

Based on the grading criteria for Ba in Urine, 100% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Urine Ba



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 1 \mu\text{g}/\text{L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $5 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Urine Be (<math>\mu\text{g/L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	1.8	2.8	4.7	0.5	1.3
<b>Upper Limit</b>	2.8	3.8	5.7	1.5	2.3
<b>Lower Limit</b>	0.8	1.8	3.7	0	0.3
<b>Robust SD (<math>s^*</math>)</b>	0.1	0.2	0.3	0.02	0.1
<b>Robust RSD (%)</b>	5.6	7.1	6.4	4.0	7.7
<b>Number of Sample Measurements (N)</b>	12	12	12	11	12
<b>Standard Uncertainty (<math>u</math>)</b>	0.0480	0.0720	0.1210	0.0069	0.0310

The acceptable range is based on quality specifications:

$\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Urine Be (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>1.8</b>	<b>2.8</b>	<b>4.7</b>	<b>0.5</b>	<b>1.3</b>
107	ICP-MS	2	3	4.9	0.51	1.3
110	ICP-MS	2.0	3.2	5.0	0.5	1.3
116	ICP-MS	1.74	2.66	4.62	0.436	1.17
147	ICP-MS	1.74	2.68	4.25	< 0.450	1.33
293	ICP-MS	1.84	2.85	4.59	0.50	1.33
399	ICP-MS	1.76	2.73	4.46	0.504	1.27
598	ICP-MS	1.74	2.89	5.15	0.49	1.45
599	DRC/CC-ICP-MS	1.18	2.20	3.27 ↓	1.62 ↑	0.829
605	ICP-MS	1.73	2.63	4.51	0.474	1.24
607	ICP-MS	1.96	2.97	4.93	0.559	1.42
632	ICP-MS	1.95	2.98	4.65	0.500	1.31
686	ICP-MS	1.88	2.88	4.78	0.536	1.38

Based on the grading criteria for Be in Urine, 97% of results were satisfactory, with 1 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Urine Be



#### Legend:

○ C/HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 5 µg/L.

## Results for Event #3, 2017: Summary Statistics

	<b>Urine Cd (<math>\mu\text{g}/\text{L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	2.63	1.65	4.47	0.71	1.21
<b>Upper Limit</b>	3.63	2.65	5.47	1.71	2.21
<b>Lower Limit</b>	1.63	0.65	3.47	0	0.21
<b>Robust SD (<math>s^*</math>)</b>	0.09	0.13	0.17	0.07	0.23
<b>Robust RSD (%)</b>	3.4	7.9	3.8	9.8	19
<b>Number of Sample Measurements (N)</b>	20	20	20	19	20
<b>Standard Uncertainty (<math>u</math>)</b>	0.026	0.036	0.047	0.020	0.064

The acceptable range is based on quality specifications:

$\pm 1 \mu\text{g}/\text{L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $6.6 \mu\text{g}/\text{L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine Cd (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>2.63</b>	<b>1.65</b>	<b>4.47</b>	<b>0.71</b>	<b>1.21</b>
103	DRC/CC-ICP-MS	2.98	1.72	4.69	0.760	1.20
107	DRC/CC-ICP-MS	2.7	1.5	4.5	0.69	1
110	ICP-MS	2.56	1.39	4.25	0.61	0.84
116	DRC/CC-ICP-MS	2.63	1.59	4.53	0.623	1.18
147	ICP-MS	2.55	1.47	4.43	0.574	0.883
200	ICP-MS	2.80	1.80	4.40	0.70	1.30
293	ICP-MS	2.77	1.62	4.23	0.71	1.18
324	ICP-MS	2.56	1.61	4.34	<1	1.35
391	DRC/CC-ICP-MS	2.59	1.64	4.31	0.74	1.33
399	DRC/CC-ICP-MS	2.55	1.63	4.43	0.738	1.18
401	DRC/CC-ICP-MS	2.59	1.69	4.50	0.67	1.24
597	DRC/CC-ICP-MS	2.56	1.93	4.51	0.67	1.02
598	DRC/CC-ICP-MS	2.58	1.57	4.56	0.68	1.12
599	DRC/CC-ICP-MS	2.09	1.51	3.45 ↓	1.77 ↑	0.912
605	ICP-MS	2.67	1.65	4.51	0.761	1.30
606	DRC/CC-ICP-MS	2.56	1.61	4.38	0.672	1.13
607	ICP-MS	2.65	1.80	4.50	0.807	1.54
630	ICP-MS	2.69	1.71	4.72	0.79	1.37
632	DRC/CC-ICP-MS	2.71	1.74	4.63	0.727	1.50
686	ICP-MS	2.79	1.88	4.77	0.862	1.62

Based on the grading criteria for Cd in Urine, 98% of results were satisfactory, with 1 of the 20 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Urine Cd



### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Urine Co (<math>\mu\text{g}/\text{L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	6.2	1.5	2.8	0.91	2.4
<b>Upper Limit</b>	7.7	3	4.3	2.41	3.9
<b>Lower Limit</b>	4.7	0	1.3	0	0.9
<b>Robust SD (<math>s^*</math>)</b>	0.4	0.2	0.2	0.11	0.1
<b>Robust RSD (%)</b>	6.5	13.3	7.1	12.1	4.2
<b>Number of Sample Measurements (N)</b>	15	15	15	14	15
<b>Standard Uncertainty (<math>u</math>)</b>	0.126	0.051	0.057	0.036	0.045

The acceptable range is based on quality specifications:

$\pm 1.5 \mu\text{g}/\text{L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $10 \mu\text{g}/\text{L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine Co (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>6.2</b>	<b>1.5</b>	<b>2.8</b>	<b>0.91</b>	<b>2.4</b>
103	DRC/CC-ICP-MS	6.39	1.42	2.84	0.869	2.42
107	ICP-MS	6.7	1.8	2.9	1	2.5
110	ICP-MS	6.75	1.82	2.95	1.19	2.76
147	ICP-MS	6.07	1.40	2.67	0.831	2.40
324	ICP-MS	6.18	1.51	2.74	<1	2.40
391	DRC/CC-ICP-MS	5.15	1.19	2.3	0.76	1.90
399	DRC/CC-ICP-MS	6.28	1.45	2.74	0.843	2.34
401	DRC/CC-ICP-MS	6.79	1.42	2.86	0.91	2.47
485	HR-ICP-MS	6.14	1.38	2.81	0.880	2.45
597	DRC/CC-ICP-MS	5.51	1.60	2.54	0.79	1.93
598	ICP-MS	5.98	1.70	3.06	1.01	2.62
599	DRC/CC-ICP-MS	4.91	1.17	2.2	2.18	2.14
605	ICP-MS	6.00	1.36	2.69	0.834	2.34
606	DRC/CC-ICP-MS	6.25	1.45	2.81	0.875	2.47
632	ICP-MS	6.81	1.73	2.95	1.04	2.61

Based on the grading criteria for Co in Urine, 100% of results were satisfactory, with 0 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Urine Co



#### Legend:

○ C/HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Urine Cr (<math>\mu\text{g/L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	17	4.8	0.53	2.3	9.1
<b>Upper Limit</b>	20.4	7.8	3.53	5.3	12.1
<b>Lower Limit</b>	13.6	1.8	0	0	6.1
<b>Robust SD (<math>s^*</math>)</b>	1.2	0.5	0.09	0.2	0.6
<b>Robust RSD (%)</b>	7.1	10.4	17.0	8.7	6.6
<b>Number of Sample Measurements (N)</b>	12	12	10	12	12
<b>Standard Uncertainty (<math>u</math>)</b>	0.438	0.166	0.036	0.057	0.203

The acceptable range is based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $15 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine Cr (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>17</b>	<b>4.8</b>	<b>0.53</b>	<b>2.3</b>	<b>9.1</b>
103	DRC/CC-ICP-MS	17.3	4.69	0.381	2.18	9.14
107	DRC/CC-ICP-MS	18	5.1	0.57	2.4	9.6
110	DRC/CC-ICP-MS	17.3	5.3	<0.8	2.9	9.5
116	DRC/CC-ICP-MS	17.2	4.48	0.578	2.03	9.10
147	DRC/CC-ICP-MS	16.7	4.47	0.422	2.23	8.58
324	ICP-MS	15.52	3.76	<1	2.34	8.26
391	DRC/CC-ICP-MS	14.45	4.33	0.96	2.26	7.80
401	DRC/CC-ICP-MS	18.3	4.800	0.50	2.36	9.37
485	HR-ICP-MS	17.3	4.82	0.475	2.33	9.43
597	DRC/CC-ICP-MS	15.9	5.66	1.08	2.60	7.99
598	DRC/CC-ICP-MS	16.39	4.70	0.51	2.22	9.26
632	DRC/CC-ICP-MS	19.0	5.11	0.545	2.43	9.76

Based on the grading criteria for Cr in Urine, 100% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Urine Cr



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g}/\text{L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $15 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Urine Hg (<math>\mu\text{g/L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	34.3	6.1	12.8	3.3	24.1
<b>Upper Limit</b>	45	9.1	16.6	6.3	31.3
<b>Lower Limit</b>	24	3.1	9	0.3	16.9
<b>Robust SD (<math>s^*</math>)</b>	4.3	0.6	1.2	0.4	1.6
<b>Robust RSD (%)</b>	12.5	9.8	9.4	12.1	6.6
<b>Number of Sample Measurements (N)</b>	15	15	15	15	15
<b>Standard Uncertainty (<math>u</math>)</b>	1.380	0.189	0.373	0.134	0.522

The acceptable range is based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine Hg (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
		<b>Target</b>	<b>34.3</b>	<b>6.1</b>	<b>12.8</b>	<b>3.3</b>
103	DRC/CC-ICP-MS	37.3	6.54	13.9	3.47	26.2
107	DRC/CC-ICP-MS	35	5.7	13	3.4	25
110	ICP-MS	35.5	5.94	13.0	3.19	24.4
147	CV-AAS	35.2	6.03	13.4	3.13	25.0
200	ICP-MS	40.9	6.7	14.8	3.9	29.2
293	ICP-MS	31.44	6.43	12.6	3.23	23.55
391	DRC/CC-ICP-MS	30.94	12.77 ↑	11.6	3.74	21.45
401	DRC/CC-ICP-MS	44.5	8.18	16.1	3.73	31.5 ↑
598	ICP-MS	29.9	5.79	12.4	2.60	21.46
599	DRC/CC-ICP-MS	21.3 ↓	3.48	9.5	3.83	16.6 ↓
604	DRC/CC-ICP-MS	31.9	5.46	11.5	2.87	22.9
605	ICP-MS	34.9	6.09	12.5	3.31	24.5
606	DRC/CC-ICP-MS	37.5	6.56	13.8	3.04	24.1
632	ICP-MS	30.7	5.99	12.1	2.32	23.2
684	DRC/CC-ICP-MS	34.60	5.74	12.6	3.23	24.10






Based on the grading criteria for Hg in Urine, 95% of results were satisfactory, with 1 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier



## Results for Event #3, 2017: Summary Figures

### Urine Hg



#### Legend:

○ C/HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±3 µg/L or ±30% around the target value, whichever is greater; thus, it is fixed at ±3 µg/L at concentrations less than or equal to 10 µg/L.

## Results for Event #3, 2017: Summary Statistics

	<b>Urine Mn (<math>\mu\text{g/L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	2.8	7.9	1.4	1.1	2.1
<b>Upper Limit</b>	3.5	9.9	1.95	1.65	2.65
<b>Lower Limit</b>	2.1	5.9	0.85	0.55	1.55
<b>Robust SD (<math>s^*</math>)</b>	0.2	0.8	0.2	0.2	0.3
<b>Robust RSD (%)</b>	7.1	10.1	14.3	18.2	14.3
<b>Number of Sample Measurements (N)</b>	16	16	16	16	16
<b>Standard Uncertainty (<math>u</math>)</b>	0.075	0.257	0.065	0.067	0.095

The acceptable range is based on quality specifications:

$\pm 0.55 \mu\text{g/L}$  or  $\pm 25\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.55 \mu\text{g/L}$  at concentrations less than or equal to  $2.2 \mu\text{g/L}$ . Quality specifications for Mn are consistent with those used by other External Quality Assessment Schemes for trace elements. (Praamsma M, et al. Clinical Chemistry and Laboratory Medicine.2016; 54(12): 1921-1928)

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine Mn (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>2.8</b>	<b>7.9</b>	<b>1.4</b>	<b>1.1</b>	<b>2.1</b>
103	DRC/CC-ICP-MS	2.76	7.86	1.36	1.00	2.13
107	DRC/CC-ICP-MS	2.9	8.4	1.5	1.3	2.3
110	ICP-MS	2.65	7.63	1.31	0.94	2.05
116	DRC/CC-ICP-MS	2.74	7.42	1.35	0.964	1.94
147	DRC/CC-ICP-MS	2.57	7.64	1.21	0.962	1.93
324	ICP-MS	2.83	7.85	1.40	1.09	2.14
391	DRC/CC-ICP-MS	2.55	6.62	1.13	1.07	1.75
399	DRC/CC-ICP-MS	2.66	7.68	1.28	0.996	2.02
597	DRC/CC-ICP-MS	2.75	8.78	1.69	1.31	2.12
598	ICP-MS	2.99	8.35	2.17 ↑	1.61	2.81 ↑
599	DRC/CC-ICP-MS	2.10	6.36	1.24	2.88 ↑	1.55
605	ICP-MS	2.58	7.10	1.18	0.816	1.87
606	DRC/CC-ICP-MS	2.72	8.00	1.39	0.994	2.16
630	ICP-MS	3.49	9.43	2.67 ↑	2.25 ↑	3.39 ↑
632	DRC/CC-ICP-MS	3.20	8.62	1.50	1.09	2.34
684	ICP-MS	3.22	8.86	1.89	1.44	2.63






Based on the grading criteria for Mn in Urine, 93% of results were satisfactory, with 2 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier



## Results for Event #3, 2017: Summary Figures

### Urine Mn



#### Legend:

○ C/HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 0.55 \mu\text{g}/\text{L}$  or  $\pm 25\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.55 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $2.2 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Urine Pb (<math>\mu\text{g}/\text{L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	7	12.4	2.4	1.4	5.1
<b>Upper Limit</b>	8.4	14.9	3.4	2.4	6.12
<b>Lower Limit</b>	5.6	9.9	1.4	0.4	4.08
<b>Robust SD (<math>s^*</math>)</b>	0.5	1.0	0.1	0.1	0.3
<b>Robust RSD (%)</b>	7.1	8.1	4.2	7.1	5.9
<b>Number of Sample Measurements (N)</b>	18	18	18	17	18
<b>Standard Uncertainty (<math>u</math>)</b>	0.146	0.291	0.044	0.029	0.089

The acceptable range is based on quality specifications:

$\pm 1 \mu\text{g}/\text{L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $5 \mu\text{g}/\text{L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine Pb (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>7</b>	<b>12.4</b>	<b>2.4</b>	<b>1.4</b>	<b>5.1</b>
103	DRC/CC-ICP-MS	7.36	13.2	2.53	1.41	5.39
107	ICP-MS	7.5	13	2.5	1.4	5.3
110	ICP-MS	7.2	12.5	2.5	1.3	5.2
116	ICP-MS	6.85	12.2	2.41	1.33	5.12
147	ICP-MS	6.76	12.2	2.32	*11.9 <span style="color: red;">↑</span>	5.10
200	ICP-MS	7.7	13.7	2.9	1.2	5.6
293	ICP-MS	7.4	13.5	2.3	1.5	4.9
324	ICP-MS	6.57	9.90 ↓	2.27	1.26	4.73
391	DRC/CC-ICP-MS	6.4	11.3	2.3	1.3	4.9
399	ICP-MS	7.15	12.5	2.46	1.34	5.23
597	DRC/CC-ICP-MS	6.88	14.7	2.71	1.51	4.61
598	ICP-MS	6.18	11.1	2.22	1.23	4.34
599	DRC/CC-ICP-MS	5.65	10.6	2.04	2.16	4.72
605	ICP-MS	6.7	12.2	2.36	1.34	5.04
606	ICP-MS	6.75	12.0	2.34	1.33	4.99
607	ICP-MS	7.50	13.3	2.53	1.48	5.39
632	ICP-MS	7.15	12.5	2.39	1.34	5.16
686	ICP-MS	7.21	12.6	2.48	1.39	5.29






Based on the grading criteria for Pb in Urine, 98% of results were satisfactory, with 0 of the 18 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier



## Results for Event #3, 2017: Summary Figures

### Urine Pb



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 1 \mu\text{g}/\text{L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $5 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Urine TI (<math>\mu\text{g}/\text{L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	1.08	1.85	5.3	0.31	0.75
<b>Upper Limit</b>	1.3	2.22	6.36	0.51	0.95
<b>Lower Limit</b>	0.86	1.48	4.24	0.11	0.55
<b>Robust SD (<math>s^*</math>)</b>	0.03	0.05	0.17	0.005	0.02
<b>Robust RSD (%)</b>	2.8	2.7	3.2	1.5	2.7
<b>Number of Sample Measurements (N)</b>	14	14	14	14	14
<b>Standard Uncertainty (<math>u</math>)</b>	0.0110	0.0170	0.0560	0.0015	0.0053

The acceptable range is based on quality specifications:

$\pm 0.2 \mu\text{g}/\text{L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.2 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $1 \mu\text{g}/\text{L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Urine TI (<math>\mu\text{g/L}</math>)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>1.08</b>	<b>1.85</b>	<b>5.3</b>	<b>0.31</b>	<b>0.75</b>
103	DRC/CC-ICP-MS	1.12	1.92	5.46	0.307	0.755
107	ICP-MS	1.1	1.9	5.5	0.31	0.74
110	ICP-MS	1.09	1.86	5.34	0.31	0.72
116	ICP-MS	1.08	1.82	5.26	0.300	0.731
147	ICP-MS	1.08	1.83	5.36	0.307	0.767
293	ICP-MS	1.10	1.84	5.28	0.31	0.76
399	ICP-MS	1.07	1.84	5.39	0.308	0.759
597	DRC/CC-ICP-MS	0.958	2.01	4.73	0.274	0.605
598	ICP-MS	0.94	1.68	4.90	0.27	0.62
605	ICP-MS	1.05	1.85	5.33	0.307	0.752
606	ICP-MS	1.01	1.76	4.93	0.291	0.699
607	ICP-MS	1.12	1.88	5.49	0.332	0.754
632	ICP-MS	1.09	1.84	5.11	0.324	0.76
686	ICP-MS	1.11	1.92	5.42	0.309	0.760






Based on the grading criteria for TI in Urine, 100% of results were satisfactory, with 0 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier



## Results for Event #3, 2017: Summary Figures

### Urine Tl



#### Legend:

○ C/HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 0.2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.2 \mu\text{g/L}$  at concentrations less than or equal to  $1 \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

	<b>Urine U (<math>\mu\text{g/L}</math>)</b>				
	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	0.057	0.011	0.19	0.027	0.063
<b>Upper Limit</b>	0.087	0.041	0.228	0.057	0.093
<b>Lower Limit</b>	0.027	0	0.152	0	0.033
<b>Robust SD (<math>s^*</math>)</b>	0.003	0.002	0.009	0.003	0.003
<b>Robust RSD (%)</b>	5.3	18.2	4.7	11.1	4.8
<b>Number of Sample Measurements (N)</b>	16	12	16	16	16
<b>Standard Uncertainty (<math>u</math>)</b>	0.00110	0.00068	0.00270	0.00083	0.00100

The acceptable range is based on quality specifications:

$\pm 0.03 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.03 \mu\text{g/L}$  at concentrations less than or equal to  $0.15 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Urine U (µg/L)</b>				
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
	<b>Target</b>	<b>0.057</b>	<b>0.011</b>	<b>0.19</b>	<b>0.027</b>	<b>0.063</b>
103	DRC/CC-ICP-MS	0.0533	0.00784	0.185	0.0256	0.0620
107	ICP-MS	0.058	0.013	0.2	0.028	0.066
110	ICP-MS	0.058	0.012	0.193	0.028	0.061
116	ICP-MS	0.0526	<0.0108	0.197	0.0248	0.0607
147	ICP-MS	0.0455	< 0.0105	0.167	0.0112	0.0502
324	ICP-MS	0.054	0.011	0.175	0.026	0.060
399	ICP-MS	0.055	0.01	0.175	0.023	0.062
598	ICP-MS	0.076	0.026	0.188	0.035	0.068
604	ICP-MS	0.056	0.012	0.182	0.026	0.060
605	ICP-MS	0.061	0.012	0.200	0.028	0.067
606	ICP-MS	0.057	0.010	0.186	0.025	0.063
607	ICP-MS	0.054	0.010	0.198	0.029	0.064
630	ICP-MS	0.056	0.010	0.191	0.026	0.064
632	ICP-MS	0.0594	<0.015	0.194	0.0309	0.0673
684	ICP-MS	0.058	0.013	0.196	0.030	0.064
686	ICP-MS	0.0594	<0.015	0.193	0.0265	0.0646

Based on the grading criteria for U in Urine, 100% of results were satisfactory, with 0 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Urine U



#### Legend:

○ C/HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 0.03 \mu\text{g}/\text{L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.03 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $0.15 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Cs (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
107	ICP-MS	5.6	10	2.6	7.6	15
110	ICP-MS	5.7	10.6	2.7	7.4	15.1
147	ICP-MS	5.41	10.0	2.83	7.19	14.5
399	ICP-MS	5.56	10.30	2.68	7.52	15.30
597	DRC/CC-ICP-MS	4.77	10.9	2.35	6.58	12.1
598	ICP-MS	5.32	10.6	2.77	7.52	15.7
599	DRC/CC-ICP-MS	4.98	9.78	2.35	8.52	14.9
605	ICP-MS	5.27	10.0	2.52	7.18	14.6
606	DRC/CC-ICP-MS	5.63	10.4	2.71	7.51	15.4
632	ICP-MS	5.95	10.7	2.69	7.83	15.5






  

<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Robust Mean (<math>x^*</math>)</b>	5.5	10.3	2.6	7.5	15
<b>Robust SD (<math>s^*</math>)</b>	0.3	0.4	0.1	0.3	0.5
<b>Robust RSD (%)</b>	5.5	3.9	3.8	4	3.3
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (<math>u</math>)</b>	0.124	0.167	0.052	0.13	0.213

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Urine Cs



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = robust mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Cu (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
110	ICP-MS	28.7	74.6	42.4	20.4	112
116	DRC/CC-ICP-MS	24.5	68.4	37.1	15.5	98.3
147	ICP-MS	28.3	71.8	41.9	19.6	107
293	ICP-MS	24.8	67	37.5	17.2	103.0
324	ICP-MS	25.81	69.70	39.86	17.72	104.92
391	DRC/CC-ICP-MS	19.5	48	31.0	15.1	74.2
401	DRC/CC-ICP-MS	24.8	67.4	37.5	15.9	101
597	DRC/CC-ICP-MS	24.1	80.1	37.5	17.1	87.9
598	ICP-MS	30.5	69.3	40.69	18.33	105
632	ICP-MS	32.4	88.1	45.9	22.8	116






<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Robust Mean (<math>x^*</math>)</b>	26.2	70.2	39.2	17.8	103
<b>Robust SD (<math>s^*</math>)</b>	3.3	4.2	3.1	2.4	8
<b>Robust RSD (%)</b>	12.6	6	7.9	13.5	7.8
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (<math>u</math>)</b>	1.29	1.66	1.21	0.946	2.97

\*Denotes a statistical Outlier.



## Results for Event #3, 2017: Summary Figures

### Urine Cu



#### Legend:

○ C/HHEAR Labs   ● Other Labs  
Horizontal purple line = robust mean of all laboratories.  
Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Mo (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
103	DRC/CC-ICP-MS	51.3	128	122	98.7	288
107	ICP-MS	52	130	120	100	280
110	ICP-MS	51.5	129	128	103	298
147	ICP-MS	49.2	126	122	96.0	280
293	ICP-MS	50.3	125.4	119.9	99.5	279.7
324	ICP-MS	51.01	124.52	120.36	96.98	275.96
399	ICP-MS	50.2	128	121	97.7	289
485	HR-ICP-MS	51.6	128	122	98.3	291
597	DRC/CC-ICP-MS	43.5	140	110	87.2	233
598	DRC/CC-ICP-MS	47.3	128.4	129.6	96.5	289.1
599	DRC/CC-ICP-MS	34.9	96.2	85.5	77.8	234.9
605	ICP-MS	48.4	123	118	93.3	279
606	DRC/CC-ICP-MS	50.2	129	122	97.0	290
632	ICP-MS	56.6	142	134	109	313






<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Robust Mean (<math>x^*</math>)</b>	50.1	128	121	97	284
<b>Robust SD (<math>s^*</math>)</b>	2.1	3	3	3	9
<b>Robust RSD (%)</b>	4.2	2.3	2.5	3.1	3.2
<b>Number of Sample Measurements (N)</b>	14	14	14	14	14
<b>Standard Uncertainty (<math>u</math>)</b>	0.71	1.09	0.929	1.01	3.15

\*Denotes a statistical Outlier.



## Results for Event #3, 2017: Summary Figures

### Urine Mo



#### Legend:

○ C/HHEAR Labs   ● Other Labs  
Horizontal purple line = robust mean of all laboratories.  
Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.






## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Ni (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
107	DRC/CC-ICP-MS	4.9	2.4	4.5	8.1	14
110	ICP-MS	5.6	3.2	5.1	9.5	14.3
293	ICP-MS	4.38	1.57	3.58	7.12	11.81
324	ICP-MS	5.12	2.85	4.65	7.75	13.20
391	DRC/CC-ICP-MS	*8.0	5.5	*7.5	9.9	14.7
401	DRC/CC-ICP-MS	5.400	2.88	4.58	7.63	13.8
485	HR-ICP-MS	4.59	1.66	3.98	7.74	13.7
597	DRC/CC-ICP-MS	4.59	3.26	4.98	7.98	12.9
598	ICP-MS	5.70	4.31	5.52	8.70	14.6
<b>Summary Statistics</b>						
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		5	3.1	4.6	8.3	13.7
<b>Arithmetic SD (s)</b>		0.5	1.2	0.6	0.9	0.9
<b>Arithmetic RSD (%)</b>		10	38.7	13	10.8	6.6
<b>Number of Sample Measurements (N)</b>		8	9	8	9	9

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Urine Ni



#### Legend:

○ C/HHEAR Labs   ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Pt (µg/L)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
107	ICP-MS	2.2	0.65	5.6	0.3	1.2
110	ICP-MS	2.23	0.60	5.39	0.20	1.22
147	ICP-MS	2.07	0.554	4.94	0.222	1.11
399	ICP-MS	2.16	0.597	5.21	0.234	1.19
598	ICP-MS	2.01	0.63	5.24	0.21	1.12
632	ICP-MS	2.17	0.658	4.89	0.244	1.24






<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	2.14	0.61	5.21	0.23	1.18
<b>Arithmetic SD (s)</b>	0.08	0.04	0.27	0.04	0.05
<b>Arithmetic RSD (%)</b>	3.7	6.6	5.2	17.4	4.2
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

\*Denotes a statistical Outlier.



## Results for Event #3, 2017: Summary Figures

### Urine Pt

**Legend:**

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.






## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Sb (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
103	DRC/CC-ICP-MS	0.324	0.825	1.64	0.589	1.27
107	ICP-MS	0.52	0.83	1.6	0.84	1.7
110	ICP-MS	0.3	0.8	1.5	0.6	1.2
147	ICP-MS	0.379	0.800	1.62	0.631	1.24
293	ICP-MS	0.4	0.8	1.5	0.6	1.2
399	ICP-MS	0.365	0.858	1.65	0.647	1.33
597	DRC/CC-ICP-MS	0.524	1.13	1.73	0.731	1.22
598	ICP-MS	0.38	0.92	1.82	0.66	1.36
605	ICP-MS	0.361	0.834	1.59	0.642	1.23
606	DRC/CC-ICP-MS	0.371	0.853	1.57	0.605	1.28
632	ICP-MS	0.384	0.905	1.63	0.649	1.28
<b>Summary Statistics</b>						
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Robust Mean (<math>x^*</math>)</b>		0.38	0.85	1.62	0.64	1.27
<b>Robust SD (<math>s^*</math>)</b>		0.03	0.05	0.05	0.05	0.07
<b>Robust RSD (%)</b>		7.9	5.9	3.1	7.8	5.5
<b>Number of Sample Measurements (N)</b>		11	11	11	11	11
<b>Standard Uncertainty (<math>u</math>)</b>		0.012	0.019	0.02	0.019	0.026

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Urine Sb



#### Legend:

○ C/HHEAR Labs    ● Other Labs

Horizontal purple line = robust mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Se (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
103	DRC/CC-ICP-MS	77.3	168	211	48.3	151
110	DRC/CC-ICP-MS	70.7	162	208	50.4	152
147	ICP-MS	83.7	183	234	54.6	164
597	DRC/CC-ICP-MS	67.4	191	200	48.0	131
598	DRC/CC-ICP-MS	65.4	167.1	217.2	48.1	153.2
632	DRC/CC-ICP-MS	66.8	151	189	45.7	136
684	DRC/CC-ICP-MS	82.0	188	223.0	55.6	166.0

<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	73.3	173	212	50.1	150
<b>Arithmetic SD (s)</b>	7.6	15	15	3.7	13
<b>Arithmetic RSD (%)</b>	10.4	8.7	7.1	7.4	8.7
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

\*Denotes a statistical Outlier.








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Urine Se



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Sn (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
107	ICP-MS	1.9	3.3	1.2	2.8	13
110	ICP-MS	1.6	2.8	1.0	2.5	11.6
147	ICP-MS	1.69	2.92	1.10	2.49	12.5
399	ICP-MS	1.63	2.92	1.01	2.52	12.5
598	ICP-MS	1.93	3.3	1.22	2.76	13.70
599	DRC/CC-ICP-MS	1.11	2.28	1.01	2.09	10.22
605	ICP-MS	1.43	2.57	0.775	2.17	11.6

<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	1.6	2.9	1	2.5	12.2
<b>Arithmetic SD (s)</b>	0.3	0.4	0.2	0.3	1.1
<b>Arithmetic RSD (%)</b>	18.8	13.8	20	12	9
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

\*Denotes a statistical Outlier.








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Urine Sn



#### Legend:

○C/HHEAR Labs   ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.  
Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Sr (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
103	DRC/CC-ICP-MS	136	251	60.1	119	279
107	ICP-MS	140	270	62	120	300
200	ICP-MS	123.5	246.2	57.8	120.0	270.7
399	DRC/CC-ICP-MS	138	251	60.8	120	282
605	ICP-MS	134	242	58.6	116	275

<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	134	252	59.9	119	281
<b>Arithmetic SD (s)</b>	6	11	1.7	2	11
<b>Arithmetic RSD (%)</b>	4.5	4.4	2.8	1.7	3.9
<b>Number of Sample Measurements (N)</b>	5	5	5	5	5

\*Denotes a statistical Outlier.








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Urine Sr



#### Legend:

○ C/HHEAR Labs   ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.  
Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine V (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
116	DRC/CC-ICP-MS	0.821	0.399	1.50	1.23	6.06
147	DRC/CC-ICP-MS	0.776	0.360	1.30	1.00	5.31
293	ICP-MS	0.97	0.6	1.67	1.41	6.22
485	HR-ICP-MS	0.856	0.394	1.57	1.25	6.41
597	DRC/CC-ICP-MS	0.835	0.519	1.42	1.17	5.08
598	DRC/CC-ICP-MS	0.97	0.56	1.85	1.43	6.93






<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	0.87	0.47	1.55	1.25	6
<b>Arithmetic SD (s)</b>	0.08	0.1	0.19	0.16	0.69
<b>Arithmetic RSD (%)</b>	9.2	21.3	12.3	12.8	11.5
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

\*Denotes a statistical Outlier.



## Results for Event #3, 2017: Summary Figures

### Urine V

**Legend:**

○ C/HHEAR Labs   ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.  
Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.






## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine W (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
107	ICP-MS	0.36	1.5	0.87	0.55	1.2
110	ICP-MS	0.39	1.60	0.96	0.63	1.28
147	ICP-MS	0.638	1.58	0.896	0.539	1.24
200	ICP-MS	0.54	1.75	0.98	0.64	1.38
324	ICP-MS	<1	1.51	<1	<1	1.22
399	ICP-MS	0.313	1.530	0.880	0.552	1.23
598	ICP-MS	0.40	1.70	*1.21	0.59	1.30
605	ICP-MS	0.311	1.50	0.860	0.542	1.24
606	DRC/CC-ICP-MS	0.306	1.52	0.875	0.544	1.26
<b>Summary Statistics</b>						
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.41	1.58	0.9	0.57	1.26
<b>Arithmetic SD (s)</b>		0.12	0.09	0.05	0.04	0.05
<b>Arithmetic RSD (%)</b>		29.3	5.7	5.6	7	4
<b>Number of Sample Measurements (N)</b>		8	9	7	8	9

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Urine W



#### Legend:

○ C/HHEAR Labs   ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.






## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine Zn (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
110	ICP-MS	354	874	435	883	1335
147	ICP-MS	356	876	449	902	1399
293	ICP-MS	342.5	907.8	460.1	902.6	1333.3
324	ICP-MS	335.53	844.73	417.25	837.74	1272.46
401	DRC/CC-ICP-MS	366	915	451	883	1373
597	DRC/CC-ICP-MS	299	976	384	753	*1061
598	ICP-MS	315	837	436	834	1333
599	DRC/CC-ICP-MS	300	867	383	866	1364
632	ICP-MS	377	985	465	951	1459
<b>Summary Statistics</b>						
		<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		338	898	431	868	1359
<b>Arithmetic SD (s)</b>		28	53	31	56	55
<b>Arithmetic RSD (%)</b>		8.3	5.9	7.2	6.5	4
<b>Number of Sample Measurements (N)</b>		9	9	9	9	8

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Urine Zn



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Urine AI (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
147	DRC/CC-ICP-MS	<13.8	18.0	16.7	<13.8	26.4
324	ICP-MS	9.63	17.46	18.83	11.90	25.61
485	HR-ICP-MS	5.56	12.3	16.7	8.93	28.7
597	DRC/CC-ICP-MS	1.60	12.0	15.4	5.50	21.7

<b>Summary Statistics</b>					
	<b>UE17-11</b>	<b>UE17-12</b>	<b>UE17-13</b>	<b>UE17-14</b>	<b>UE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	NA	NA	16.9	NA	26
<b>Arithmetic SD (s)</b>	NA	NA	1.4	NA	3
<b>Arithmetic RSD (%)</b>	NA	NA	8.3	NA	11.5
<b>Number of Sample Measurements (N)</b>	NA	NA	4	NA	4

\*Denotes a statistical Outlier.

Statistics for UE17-11, UE17-12, and UE17-14 were not calculated based on a lack of consensus among participating laboratories.



## Results for Event #3, 2017: Additional Elements in Urine

Urine Ag ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
147	ICP-MS	< 0.302	< 0.302	< 0.302	< 0.302	< 0.302
Urine B ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
200	ICP-MS	1253	1501	1307	799	994
Urine Bi ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
147	ICP-MS	< 0.230	< 0.230	< 0.230	< 0.230	< 0.230
Urine Fe ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
324	ICP-MS	7.93	10.61	9.30	8.85	7.78
Urine I ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
107	ICP-MS	43	100	43	130	100
Urine Li ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
147	ICP-MS	35.9	24.0	36.6	15.0	16.6
Urine Mg ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
597	DRC/CC-ICP-MS	31800	78200	34500	13800	28500
Urine Te ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
110	ICP-MS	1.1	1.0	0.3	0.7	2.4
Urine Th ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
147	ICP-MS	< 0.00557	< 0.00557	< 0.00557	< 0.00557	< 0.00557
Urine Ti ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	UE17-11	UE17-12	UE17-13	UE17-14	UE17-15
485	HR-ICP-MS	<0.5	<0.5	<0.5	5.96	24.6



**Department  
of Health**

Wadsworth  
Center

## **Event #3, 2017**

# **Trace Elements in Serum**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*

### **Event #3, 2017: Trace Elements in Serum**

#### **PT Materials**

Test materials were prepared from human serum obtained from ZenBio, Inc. The company certifies that these materials were tested by FDA approved methods and found to be negative for HIV 1 $\ddot{Z}2$  and HIV-1 RNA, and non-reactive to HBsAg, HCV3 and STS. Units of serum were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with aluminum (Al), cobalt (Co), chromium (Cr), copper (Cu), selenium (Se), zinc (Zn), arsenic (As), beryllium (Be), cadmium (Cd), mercury (Hg), manganese (Mn), molybdenum (Mo), nickel (Ni), lead (Pb), platinum (Pt), antimony (Sb), tin (Sn), strontium (Sr), titanium (Ti), thallium (Tl), uranium (U), vanadium (V) and tungsten (W). Serum units were homogenized overnight prior to aliquoting 2-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

#### **Graded Elements**

Six elements in serum are formally graded: Al, Co, Cr, Cu, Se, and Zn. Target values for the graded elements are assigned to these pools based on the arithmetic mean after outlier deletion.

#### **Additional Elements**

An additional 29 elements (beyond the six graded) were reported by at least one participant: Ag, As, B, Ba, Be, Bi, Cd, Cs, Fe, Hg, I, Li, Mg, Mn, Mo, Ni, Pb, Pt, Sb, Sn, Sr, Te, Th, Ti, Tl, U, V, and W. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.

## Results for Event #3, 2017: Summary Statistics

Serum AI ( $\mu\text{g/L}$ )					
	SE17-11	SE17-12	SE17-13	SE17-14	SE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	56.2	93	147	196	29.4
<b>Upper Limit</b>	67.44	111.6	176.4	235.2	35.28
<b>Lower Limit</b>	44.96	74.4	117.6	156.8	23.52
<b>Arithmetic SD (s)</b>	4.5	5.4	14	22	3.0
<b>Arithmetic RSD (%)</b>	8.0	5.8	9.5	11.2	10.2
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

The acceptable range is based on quality specifications:

$\pm 5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 5 \mu\text{g/L}$  at concentrations less than or equal to 25  $\mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Serum AI (<math>\mu\text{g/L}</math>)</b>				
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
		<b>Target</b>	<b>56.2</b>	<b>93</b>	<b>147</b>	<b>196</b>
147	ETAAS-Z	54.0	87.8	134	173	28
200	DRC/CC-ICP-MS	48.60	89.6	130	169.3	25.7
293	ICP-MS	58.87	92.74	145.7	197.9	32.26
391	ETAAS-Z	72.14 ↑	88.15	171.5	214	50.65 ↑
485	HR-ICP-MS	55.80	90.7	154	202	28.80
597	DRC/CC-ICP-MS	59.1	103	167	226	33.7
598	ICP-MS	60.8	94.3	150.9	209.5	27.9

Based on the grading criteria for AI in Serum, 94% of results were satisfactory, with 1 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier








**Department  
of Health**

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Serum AI



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±5  $\mu\text{g}/\text{L}$  or ±20% around the target value, whichever is greater; thus, it is fixed at ±5  $\mu\text{g}/\text{L}$  at concentrations less than or equal to 25  $\mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Summary Statistics

	Serum Co ( $\mu\text{g/L}$ )				
	SE17-11	SE17-12	SE17-13	SE17-14	SE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	6.6	3.5	4.1	1.1	10.7
<b>Upper Limit</b>	8.1	5	5.6	2.6	12.305
<b>Lower Limit</b>	5.1	2	2.6	0	9.095
<b>Arithmetic SD (s)</b>	0.3	0.2	0.3	0.2	0.7
<b>Arithmetic RSD (%)</b>	4.5	5.7	7.3	18.2	6.5
<b>Number of Sample Measurements (N)</b>	7	7	6	7	7

The acceptable range is based on quality specifications:

$\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Serum Co (<math>\mu\text{g/L}</math>)</b>				
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
		<b>Target</b>	<b>6.6</b>	<b>3.5</b>	<b>4.1</b>	<b>1.1</b>
103	DRC/CC-ICP-MS	6.27	3.25	3.85	1.00	10.3
110	ICP-MS	6.58	3.48	3.99	1.14	10.9
147	ICP-MS	6.36	3.31	3.77	0.984	10.3
293	ICP-MS	6.41	3.43	*22.37 ↑	1.08	10.13
485	HR-ICP-MS	6.60	3.48	4.28	1.03	11.1
597	DRC/CC-ICP-MS	7.02	3.84	4.61	1.27	12.0
598	ICP-MS	6.72	3.79	4.35	1.46	10.38
599	DRC/CC-ICP-MS	*11.7 ↑	*6.62 ↑	*7.94 ↑	*2.40	*15.7 ↑

Based on the grading criteria for Co in Serum, 88% of results were satisfactory, with 1 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Serum Co



#### Legend:

○ C/HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:

$\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

Serum Cr ( $\mu\text{g/L}$ )					
	SE17-11	SE17-12	SE17-13	SE17-14	SE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	0.91	2.8	4.1	1.1	10.1
<b>Upper Limit</b>	2.91	4.8	6.1	3.1	12.12
<b>Lower Limit</b>	0	0.8	2.1	0	8.08
<b>Arithmetic SD (s)</b>	0.09	0.2	0.2	0.1	0.5
<b>Arithmetic RSD (%)</b>	9.9	7.1	4.9	9.1	5.0
<b>Number of Sample Measurements (N)</b>	6	6	5	6	6

The acceptable range is based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Serum Cr (µg/L)</b>				
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
		<b>Target</b>	<b>0.91</b>	<b>2.8</b>	<b>4.1</b>	<b>1.1</b>
103	DRC/CC-ICP-MS	0.792	2.66	4.04	0.982	10.0
110	DRC/CC-ICP-MS	0.9	2.8	4.2	1.1	10.7
147	DRC/CC-ICP-MS	0.832	2.49	3.75	1.00	9.31
293	ICP-MS	0.97	3.04	*37.89 ↑	1.27	9.88
485	HR-ICP-MS	0.899	2.93	4.25	1.00	10.4
597	DRC/CC-ICP-MS	*3.12 ↑	*5.57 ↑	*7.13 ↑	*4.07 ↑	*14.3 ↑
598	DRC/CC-ICP-MS	1.05	2.87	4.13	0.99	10.2

Based on the grading criteria for Cr in Serum, % of results were satisfactory, with of the laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Serum Cr



#### Legend:

○ O/C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:

±2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±2 µg/L at concentrations less than or equal to 10 µg/L.

## Results for Event #3, 2017: Summary Statistics

Serum Cu ( $\mu\text{g/L}$ )					
	SE17-11	SE17-12	SE17-13	SE17-14	SE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	2614	1657	1037	966	775
<b>Upper Limit</b>	3006.1	1905.55	1192.55	1110.9	891.25
<b>Lower Limit</b>	2221.9	1408.45	881.45	821.1	658.75
<b>Arithmetic SD (s)</b>	161	70	66	70	45
<b>Arithmetic RSD (%)</b>	6.2	4.2	6.4	7.2	5.8
<b>Number of Sample Measurements (N)</b>	9	9	9	9	9

The acceptable range is based on quality specifications:

$\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Serum Cu (µg/L)</b>				
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
		<b>Target</b>	<b>2614</b>	<b>1657</b>	<b>1037</b>	<b>966</b>
107	DRC/CC-ICP-MS	2800	1700	1100	990	810
110	ICP-MS	2730	1700	1040	985	813
147	ICP-MS	2598	1607	997	934	769
200	ICP-MS	2261	1581	921	838	756
293	ICP-MS	2614.1	1587.4	1022.3	938.3	733.6
457	ICP-AES/OES	2667	1712	1049	1010	789
483	DRC/CC-ICP-MS	2660	1680	1010	953	766
597	DRC/CC-ICP-MS	2720	1770	1160	1100	841
598	ICP-MS	2479	1577	1031	949	693

Based on the grading criteria for Cu in Serum, 100% of results were satisfactory, with 0 of the 9 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Serum Cu



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:

$\pm 95 \mu\text{g}/\text{L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $635 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Summary Statistics

Serum Se ( $\mu\text{g/L}$ )					
	SE17-11	SE17-12	SE17-13	SE17-14	SE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	175	128	262	212	148
<b>Upper Limit</b>	210	153.6	314.4	254.4	177.6
<b>Lower Limit</b>	140	102.4	209.6	169.6	118.4
<b>Arithmetic SD (s)</b>	15	9	25	23	11
<b>Arithmetic RSD (%)</b>	8.6	7.0	9.5	10.8	7.4
<b>Number of Sample Measurements (N)</b>	9	9	9	9	9

The acceptable range is based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.

## Results for Event #3, 2017: Performance of Participating Laboratories

<b>Lab Code</b>	<b>Method</b>	<b>Serum Se (<math>\mu\text{g/L}</math>)</b>				
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
		<b>Target</b>	<b>175</b>	<b>128</b>	<b>262</b>	<b>212</b>
103	DRC/CC-ICP-MS	183	132	278	226	155
107	DRC/CC-ICP-MS	160	120	250	200	140
110	DRC/CC-ICP-MS	175	122	251	208	145
147	ICP-MS	171	119	236	193	137
200	DRC/CC-ICP-MS	146.2	120.1	226.7	178.5	136.7
293	ICP-MS	187.9	128.7	280.9	223.4	146.8
483	DRC/CC-ICP-MS	179	130	254	211	145
597	DRC/CC-ICP-MS	194	146	306	261 ↑	172
598	DRC/CC-ICP-MS	183	134	271	211	152

Based on the grading criteria for Se in Serum, 98% of results were satisfactory, with 0 of the 9 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Serum Se



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
 $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .

## Results for Event #3, 2017: Summary Statistics

Serum Zn ( $\mu\text{g/L}$ )					
	SE17-11	SE17-12	SE17-13	SE17-14	SE17-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	795	2019	1600	1337	785
<b>Upper Limit</b>	914.25	2321.85	1840	1537.55	902.75
<b>Lower Limit</b>	675.75	1716.15	1360	1136.45	667.25
<b>Arithmetic SD (s)</b>	61	123	119	104	56
<b>Arithmetic RSD (%)</b>	7.7	6.1	7.4	7.8	7.1
<b>Number of Sample Measurements (N)</b>	9	9	9	9	9

The acceptable range is based on quality specifications:

$\pm 15 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 15 \mu\text{g/L}$  at concentrations less than or equal to  $100 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.

## Results for Event #3, 2017: Performance of Participating Laboratories






<b>Lab Code</b>	<b>Method</b>	<b>Serum Zn (µg/L)</b>				
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
		<b>Target</b>	<b>795</b>	<b>2019</b>	<b>1600</b>	<b>1337</b>
107	DRC/CC-ICP-MS	900	2200	1700	1400	830
110	ICP-MS	808	2020	1620	1340	820
147	ICP-MS	778	1954	1536	1268	745
200	ICP-MS	693	2001	1465	1184	778
293	ICP-MS	781.7	1926.1	1634.6	1322.2	751
457	ICP-AES/OES	775	1932	1551	1349	818
483	DRC/CC-ICP-MS	843	2090	1590	1340	786
597	DRC/CC-ICP-MS	841	2200	1840 ↑	1560 ↑	860
598	ICP-MS	739	1849	1460	1267	674

Based on the grading criteria for Zn in Serum, 98% of results were satisfactory, with 0 of the 9 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.

\* Denotes a statistical Outlier

## Results for Event #3, 2017: Summary Figures

### Serum Zn



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:

$\pm 15 \mu\text{g}/\text{L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 15 \mu\text{g}/\text{L}$  at concentrations less than or equal to  $100 \mu\text{g}/\text{L}$ .

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum As (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	12.2	17.9	2.18	9.62	6.37
110	DRC/CC-ICP-MS	11.7	16.8	2.1	8.7	6.4
147	ICP-MS	11.3	15.7	1.56	7.85	5.36
597	DRC/CC-ICP-MS	13.3	19.3	3.57	11.9	8.16
598	DRC/CC-ICP-MS	12.7	18.3	2.12	9.58	6.66






<b>Summary Statistics</b>					
	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	12.2	17.6	2.3	9.5	6.6
<b>Arithmetic SD (s)</b>	0.8	1.4	0.7	1.5	1
<b>Arithmetic RSD (%)</b>	6.6	8	30.4	15.8	15.2
<b>Number of Sample Measurements (N)</b>	5	5	5	5	5

\*Denotes a statistical Outlier.



## Results for Event #3, 2017: Summary Figures

### Serum As



### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Cd (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	4.59	3.76	1.38	0.509	2.65
110	ICP-MS	4.7	3.6	1.2	0.5	2.7
147	ICP-MS	4.42	3.55	1.29	0.483	2.61
597	DRC/CC-ICP-MS	4.64	3.93	1.53	0.499	2.96
598	DRC/CC-ICP-MS	5.02	3.70	1.32	0.51	2.69
599	DRC/CC-ICP-MS	5.69	*4.61	1.60	0.478	*3.58






  

<b>Summary Statistics</b>					
	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	4.8	3.7	1.4	0.5	2.7
<b>Arithmetic SD (s)</b>	0.5	0.1	0.2	0.01	0.1
<b>Arithmetic RSD (%)</b>	10.4	2.7	14.3	2	3.7
<b>Number of Sample Measurements (N)</b>	6	5	6	6	5

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Serum Cd



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.






## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Mn (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	10.2	1.63	2.01	4.15	16.8
110	ICP-MS	10.1	1.7	2.0	4.3	16.9
147	ICP-MS	10.0	1.70	2.13	4.31	16.3
293	ICP-MS	10.19	1.72	2.32	4.26	16.06
597	DRC/CC-ICP-MS	10.4	*2.14	2.63	5.02	17.4
598	ICP-MS	10.0	1.76	2.18	4.55	16.2
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		10.1	1.7	2.2	4.4	16.6
<b>Arithmetic SD (s)</b>		0.2	0.1	0.2	0.3	0.5
<b>Arithmetic RSD (%)</b>		2	5.9	9.1	6.8	3
<b>Number of Sample Measurements (N)</b>		6	5	6	6	6

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Serum Mn



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Mo (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	5.07	2.86	6.64	1.19	3.41
110	ICP-MS	5.0	3.0	6.4	1.4	3.7
147	ICP-MS	5.11	3.00	6.33	1.19	3.44
293	ICP-MS	5.28	3.1	6.54	1.38	3.4
485	HR-ICP-MS	5.40	3.28	6.64	1.27	3.82
597	DRC/CC-ICP-MS	5.33	3.00	7.20	1.39	3.58
598	DRC/CC-ICP-MS	5.77	3.52	7.08	1.48	3.92

<b>Summary Statistics</b>					
	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	5.28	3.11	6.69	1.33	3.61
<b>Arithmetic SD (s)</b>	0.26	0.22	0.33	0.11	0.21
<b>Arithmetic RSD (%)</b>	4.9	7.1	4.9	8.3	5.8
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

\*Denotes a statistical Outlier.








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Serum Mo



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Ni (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
110	DRC/CC-ICP-MS	5.4	8.5	2.5	11.8	5.6
147	ICP-MS	5.55	8.69	2.60	10.7	4.97
293	ICP-MS	4.86	7.34	2.04	10.2	4.34
485	HR-ICP-MS	5.12	8.91	2.17	10.9	4.96
597	DRC/CC-ICP-MS	4.88	8.42	1.71	11.8	4.48
598	ICP-MS	*7.87	*11.9	*5.92	*14.9	*7.54
599	DRC/CC-ICP-MS	*10.0	*16.2	*4.37	*21.7	*7.85


  

<b>Summary Statistics</b>					
	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	5.2	8.4	2.2	11.1	4.9
<b>Arithmetic SD (s)</b>	0.3	0.6	0.4	0.7	0.5
<b>Arithmetic RSD (%)</b>	5.8	7.1	18.2	6.3	10.2
<b>Number of Sample Measurements (N)</b>	5	5	5	5	5

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Serum Ni



#### Legend:

○ O/C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Pb (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	4.50	7.89	2.56	6.19	7.21
110	ICP-MS	4.5	8.0	2.6	6.1	7.5
147	ICP-MS	4.87	8.18	2.65	6.44	7.07
597	DRC/CC-ICP-MS	5.44	9.45	3.61	8.38	8.99
598	ICP-MS	4.07	7.17	2.26	5.52	6.00






  

<b>Summary Statistics</b>					
	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	4.7	8.1	2.7	6.5	7.4
<b>Arithmetic SD (s)</b>	0.5	0.8	0.5	1.1	1.1
<b>Arithmetic RSD (%)</b>	10.6	9.9	18.5	16.9	14.9
<b>Number of Sample Measurements (N)</b>	5	5	5	5	5

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Serum Pb



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Sb (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	4.45	2.87	0.462	0.847	4.67
110	ICP-MS	4.30	2.65	0.45	0.83	4.36
147	ICP-MS	4.12	2.62	0.416	0.735	4.04
597	DRC/CC-ICP-MS	4.36	2.84	0.478	0.869	4.59
598	ICP-MS	*5.35	3.33	0.52	0.98	5.00






  

<b>Summary Statistics</b>					
	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	4.3	2.9	0.47	0.85	4.5
<b>Arithmetic SD (s)</b>	0.1	0.3	0.04	0.09	0.4
<b>Arithmetic RSD (%)</b>	2.3	10.3	8.5	10.6	8.9
<b>Number of Sample Measurements (N)</b>	4	5	5	5	5

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Serum Sb



#### Legend:

○ C/HHEAR Labs   ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Tl (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	5.15	1.60	3.77	2.79	2.09
110	ICP-MS	5.03	1.55	3.79	2.75	2.20
147	ICP-MS	5.05	1.48	3.68	2.64	2.01
597	DRC/CC-ICP-MS	5.07	1.65	4.08	3.07	2.25
598	ICP-MS	4.86	1.53	3.54	2.61	1.81

<b>Summary Statistics</b>					
	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	5	1.6	3.8	2.8	2.1
<b>Arithmetic SD (s)</b>	0.1	0.1	0.2	0.2	0.2
<b>Arithmetic RSD (%)</b>	2	6.3	5.3	7.1	9.5
<b>Number of Sample Measurements (N)</b>	5	5	5	5	5

\*Denotes a statistical Outlier.








Department  
of Health

Wadsworth  
Center

## Results for Event #3, 2017: Summary Figures

### Serum Tl



#### Legend:

○ C/HHEAR Labs   ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum V (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
110	DRC/CC-ICP-MS	3.4	7.9	2.8	4.7	1.3
147	DRC/CC-ICP-MS	3.25	7.50	2.32	4.84	0.969
293	ICP-MS	3.56	7.79	2.56	5.26	1.1
485	HR-ICP-MS	3.37	8.00	2.36	4.99	1.00
597	DRC/CC-ICP-MS	3.31	7.86	2.65	5.50	1.15
598	DRC/CC-ICP-MS	3.68	8.17	2.55	5.47	1.16






  

<b>Summary Statistics</b>					
	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	3.4	7.9	2.5	5.1	1.1
<b>Arithmetic SD (s)</b>	0.2	0.2	0.2	0.3	0.1
<b>Arithmetic RSD (%)</b>	5.9	2.5	8	5.9	9.1
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Summary Figures

### Serum V



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area =  $\pm 2\text{SD}$  of the mean.

The mean and  $\pm 2\text{SD}$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Ba (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
110	ICP-MS	2.7	2.7	0.7	0.7	0.6
147	ICP-MS	2.55	2.57	0.633	0.652	0.652
597	DRC/CC-ICP-MS	2.85	3.86	0.899	0.915	0.790
598	ICP-MS	2.53	2.39	0.46	0.63	0.69
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		2.7	2.9	0.67	0.72	0.68
<b>Arithmetic SD (s)</b>		0.1	0.7	0.18	0.13	0.08
<b>Arithmetic RSD (%)</b>		3.7	24.1	26.9	18.1	11.8
<b>Number of Sample Measurements (N)</b>		4	4	4	4	4

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Be (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
110	ICP-MS	3.11	3.95	5.24	0.12	0.65
147	ICP-MS	3.23	3.83	4.96	< 0.441	0.665
293	ICP-MS	3.4	4.18	5.45	0.15	0.65
598	ICP-MS	3.54	4.38	5.99	0.24	0.71
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		3.3	4.1	5.4	NA	0.67
<b>Arithmetic SD (s)</b>		0.2	0.2	0.4	NA	0.03
<b>Arithmetic RSD (%)</b>		6.1	4.9	7.4	NA	4.2
<b>Number of Sample Measurements (N)</b>		4	4	4	NA	4

\*Denotes a statistical Outlier.

Statistics for SE17-14 were not calculated based on a lack of consensus among participating laboratories.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Cs (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
110	ICP-MS	0.7	0.7	0.5	0.4	0.4
597	DRC/CC-ICP-MS	0.786	0.822	0.640	0.502	0.547
598	ICP-MS	0.72	0.73	0.51	0.42	0.38
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.7	0.8	0.6	0.4	0.4
<b>Arithmetic SD (s)</b>		<0.1	0.1	0.1	0.1	0.1
<b>Arithmetic RSD (%)</b>		<0.1	12.5	16.7	25.0	25.0
<b>Number of Sample Measurements (N)</b>		3	3	3	3	3

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Hg (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	1.96	4.13	0.524	7.93	4.92
110	ICP-MS	2.0	4.0	0.5	7.5	4.7
147	ICP-MS	1.58	3.95	0.680	6.66	4.23
598	ICP-MS	2.12	4.67	0.85	8.85	5.43
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		1.9	4.2	0.6	7.7	4.8
<b>Arithmetic SD (s)</b>		0.2	0.3	0.2	0.9	0.5
<b>Arithmetic RSD (%)</b>		10.5	7.1	33.3	11.7	10.4
<b>Number of Sample Measurements (N)</b>		4	4	4	4	4

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Pt (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
110	ICP-MS	0.5	2.1	0.2	1.4	0.1
598	ICP-MS	0.46	2.26	0.23	1.50	0.11
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.48	2.2	0.22	1.5	0.11
<b>Arithmetic SD (s)</b>		0.03	0.1	0.02	0.1	0.01
<b>Arithmetic RSD (%)</b>		6.3	4.5	9.1	6.7	9.1
<b>Number of Sample Measurements (N)</b>		2	2	2	2	2

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Sn (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
110	ICP-MS	5.6	3.1	8.1	1.0	5.7
147	ICP-MS	5.33	3.03	7.89	0.887	5.40
597	DRC/CC-ICP-MS	5.86	3.45	9.00	1.22	6.27
598	ICP-MS	6.26	3.52	8.66	1.11	6.41
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		5.8	3.3	8.4	1.1	5.9
<b>Arithmetic SD (s)</b>		0.4	0.2	0.5	0.1	0.5
<b>Arithmetic RSD (%)</b>		6.9	6.1	6.0	9.1	8.5
<b>Number of Sample Measurements (N)</b>		4	4	4	4	4

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Sr (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	133	68.1	25.0	56.8	87.5
200	ICP-MS	131.4	70.1	24.5	51.7	90.2
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		132	69.1	24.8	54.3	88.9
<b>Arithmetic SD (s)</b>		1	1.4	0.4	3.6	1.9
<b>Arithmetic RSD (%)</b>		0.8	2.0	1.6	6.6	2.1
<b>Number of Sample Measurements (N)</b>		2	2	2	2	2

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum Ti (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
200	DRC/CC-ICP-MS	8.4	5.5	4.3	1.1	10.8
485	HR-ICP-MS	12.4	5.28	5.70	1.09	14.9
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		10	5.4	5	1.1	13
<b>Arithmetic SD (s)</b>		3	0.2	1	<0.1	3
<b>Arithmetic RSD (%)</b>		30.0	3.7	20.0	<0.1	23.1
<b>Number of Sample Measurements (N)</b>		2	2	2	2	2

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum U (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
103	DRC/CC-ICP-MS	0.0967	0.0673	0.161	0.0259	0.00876
110	ICP-MS	0.106	0.074	0.165	0.030	0.009
147	ICP-MS	0.114	0.0702	0.151	0.0286	< 0.0145
598	ICP-MS	0.15	*0.12	0.20	*0.07	*0.04
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.12	0.071	0.17	0.028	0.0089
<b>Arithmetic SD (s)</b>		0.02	0.003	0.02	0.002	0.0002
<b>Arithmetic RSD (%)</b>		16.7	4.2	11.8	7.1	2.2
<b>Number of Sample Measurements (N)</b>		4	3	4	3	2

\*Denotes a statistical Outlier.

## Results for Event #3, 2017: Laboratory Data and Summary Statistics

<b>Serum W (<math>\mu\text{g/L}</math>)</b>						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
110	ICP-MS	3.05	4.36	2.43	1.32	0.5
147	ICP-MS	2.94	4.34	2.37	1.26	0.476
200	ICP-MS	2.78	4.63	2.43	1.16	0.51
598	ICP-MS	3.68	5.33	2.85	1.55	0.60
<b>Summary Statistics</b>						
		<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		3.1	4.7	2.5	1.3	0.5
<b>Arithmetic SD (s)</b>		0.4	0.5	0.2	0.2	0.1
<b>Arithmetic RSD (%)</b>		12.9	10.6	8.0	15.4	20.0
<b>Number of Sample Measurements (N)</b>		4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2017: Additional Elements in Serum

Serum Ag ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
147	ICP-MS	< 0.248	< 0.248	< 0.248	< 0.248	< 0.248
Serum B ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
200	ICP-MS	41.0	45.4	33.5	34.6	34.6
Serum Bi ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
147	ICP-MS	0.028	< 0.201	< 0.201	< 0.201	< 0.0201
Serum Fe ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
457	ICP-AES/OES	633.0	648	1119	704	734
Serum I ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
147	ICP-MS	55.4	55.8	61.6	54.1	53.7
Serum Li ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
147	ICP-MS	1.20	1.12	0.439	0.694	0.675
Serum Mg ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
457	ICP-AES/OES	1520	1552	1590	1716	101
597	DRC/CC-ICP-MS	20200	21000	23300	25200	23600
Serum Te ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
147	ICP-MS	< 0.0880	< 0.0880	< 0.0880	< 0.0880	< 0.0880
Serum Th ( $\mu\text{g/L}$ )						
<b>Lab Code</b>	<b>Method</b>	<b>SE17-11</b>	<b>SE17-12</b>	<b>SE17-13</b>	<b>SE17-14</b>	<b>SE17-15</b>
147	ICP-MS	< 0.00789	< 0.00789	< 0.00789	< 0.00789	< 0.00789



## References

1. ISO/FDIS-13528 (2005) Statistical methods for use in proficiency testing by interlaboratory comparisons. International Organization for Standardization, Geneva.
2. Taylor A, Angerer J, Arnaud J, Claeys F, Jones RL, Mazarrasa O, Mairiaux E, Menditto A, Parsons PJ, Patriarca M, Pineau A, Valkonen S, Weber J-P, Weykamp C. Occupational and environmental laboratory medicine: A network of EQAS organisers. Accreditation and Quality Assurance. 2006;11(8-9):435-9. PubMed PMID: 086NJ-0011.