



**Department  
of Health**

**Wadsworth  
Center**

# **New York State Biomonitoring Program for Trace Elements**

## **Event #1, 2022**

### **Trace Elements in Whole Blood, Urine, and Serum**

### **April, 2022**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



**Event #1, 2022:  
Trace Elements in Whole Blood, Urine, and Serum**

4/18/2022

Dear Laboratory Director,

This report summarizes performance for the first biomonitoring proficiency test (PT) event of 2022 for Trace Elements in Whole Blood, Urine, and Serum. One of the key goals of this PT program is to achieve harmonization of biomonitoring data for trace elements.

**Target Value Assignment and Performance Evaluation:**

For these PT materials, target values have been assigned for a limited number of trace elements that are gradable under criteria set by the NYS DOH Biomonitoring PT program. See assay-specific narratives for details. Data for additional trace elements are reported and are included here in order to characterize the PT materials more completely. Participant data and descriptive statistics are provided for educational purposes. No target value or acceptable range is implied.

Where the data permit, robust statistics were used to assign target values based on Algorithm A as defined by ISO 13528:2005E *Statistical methods for use in proficiency testing by inter-laboratory comparisons* [1]. Acceptable ranges for the graded elements are based on consensus criteria and/or those set by the NYS DOH's PT program. For example, some are fixed based on US regulatory guidelines (Pb, Cd) while for other elements the criteria are based on a consensus of the Network of PT scheme organizers for trace elements in occupational and environmental laboratory medicine [2]. Quality specifications are element and matrix specific; full details are provided under each element specific narrative.

A confidential, three-digit code number assigned by PT program staff identifies all laboratory participants.

Samples for the next PT event (Event #2, 2022) will be shipped May 18, 2022. Comments about this report may be directed to [trel@health.ny.gov](mailto:trel@health.ny.gov).

Sincerely,

Patrick J. Parsons, PhD  
Chief, Inorganic and Nuclear Chemistry,  
Division of Environmental Sciences  
Wadsworth Center

Kayla Mehigan  
Coordinator, Biomonitoring PT Program,  
Division of Environmental Sciences  
Wadsworth Center



**Department  
of Health**

**Wadsworth  
Center**

**Event #1, 2022**

**Trace Elements in  
Whole Blood**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



**Event #1, 2022:  
Trace Elements in Whole Blood**

**PT Materials**

Human whole blood was purchased from Zen-Bio, Inc. and preserved with K<sub>2</sub>EDTA. The company certifies that this material was "non-reactive" for HBsAg, HBV DNA, HIV-1,2 Ab, HIV-1 RNA, HCV Ab, HCV RNA, and STS. Units of whole blood were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with arsenic (As), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), barium (Ba), beryllium (Be), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), titanium (Ti), thallium (Tl), uranium (U), vanadium (V), tungsten (W), and zinc (Zn). Whole blood samples were homogenized overnight prior to aliquoting 2-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories

**Graded Elements**

Seven elements in whole blood are formally graded: As, Cd, Co, Cr, Hg, Mn, and Pb. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

**Additional Elements**

An additional 25 elements were reported by at least one participant: Ag, Al, Ba, Be, Bi, Cs, Cu, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, Tl, U, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #1, 2022: Summary Statistics

Whole Blood As (µg/L)					
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	12.1	37.9	16.3	1.96	8.99
<b>Upper Limit</b>	18.1	45.5	22.3	7.96	14.99
<b>Lower Limit</b>	6.1	30.3	10.3	0.00	2.99
<b>Arithmetic SD (s)</b>	0.7	2.3	1.0	0.23	0.41
<b>Arithmetic RSD (%)</b>	6.0	6.2	6.1	12	4.5
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2022: Performance of Participating Laboratories






Whole Blood As (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
	<b>Target</b>	<b>12.1</b>	<b>37.9</b>	<b>16.3</b>	<b>1.96</b>	<b>8.99</b>
103	ICP-MS/MS	12.5	38.7	16.8	1.84	8.99
110	DRC/CC-ICP-MS	12.4	39.3	16.9	2.12	9.49
147	ICP-MS	11.90	37.1	16.4	1.55	8.35
264	ICP-MS	11.22	36.34	15.61	2.13	8.93
293	DRC/CC-ICP-MS	13.18	41.7	17.68	1.94	9.5
391	DRC/CC-ICP-MS	*7.28	*11.0 ↓	*6.15 ↓	*4.14	*4.73
597	ICP-MS/MS	12.4	37.5	16.4	2.24	8.93
598	DRC/CC-ICP-MS	11.20	34.4	14.60	1.88	8.7

Based on the grading criteria for As in Whole Blood, 95% of results were satisfactory, with 1 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Whole Blood As



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

Whole Blood Cd (µg/L)					
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Target (Robust Mean (x*))	9.5	12.0	4.0	1.87	0.54
Upper Limit	10.9	13.8	5.0	2.87	1.54
Lower Limit	8.1	10.2	3.0	0.87	0.00
Robust SD (s*)	0.3	0.6	0.3	0.16	0.05
Robust RSD (%)	3.4	5.0	7.5	8.6	9.3
Number of Sample Measurements (N)	13	13	13	13	11
Standard Uncertainty (u)	0.1	0.2	0.1	0.06	0.02

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.7 \mu\text{g/L}$ . These quality specifications are based on those used by US OSHA for occupational exposure.



### Results for Event #1, 2022: Performance of Participating Laboratories






Whole Blood Cd (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
	Target	9.5	12.0	4.0	1.87	0.54
103	ICP-MS/MS	9.76	12.2	4.11	1.80	0.527
107	ICP-MS/MS	9.614	12.442	3.983	1.898	0.541
110	ICP-MS	9.69	12.1	4.12	1.83	0.55
116	ICP-MS/MS	9.49	11.8	4.08	1.96	<1.50
147	ICP-MS	9.80	12.5	4.21	1.82	0.577
264	ICP-MS	8.26	10.26	3.50	1.72	0.57
293	DRC/CC-ICP-MS	9.61	11.82	3.880	1.8	0.53
391	DRC/CC-ICP-MS	8.39	12.3	2.74 ↓	0.911	0.012
597	ICP-MS/MS	9.31	11.4	3.80	1.69	0.510
598	DRC/CC-ICP-MS	9.40	9.60 ↓	3.720	2.1	0.35
605	ICP-MS	11.3 ↑	13.1	4.37	2.05	<0.5
606	ICP-MS/MS	9.49	12.0	4.12	1.91	0.715
686	ICP-MS	9.26	12.6	4.47	2.07	0.660

Based on the grading criteria for Cd in Whole Blood, 95% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Whole Blood Cd



### Legend:

○ C/HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 6.7 µg/L.



## Results for Event #1, 2022: Summary Statistics

	Whole Blood Co (µg/L)				
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	213	1.18	79	6.1	5.05
<b>Upper Limit</b>	256	2.68	95	7.6	6.55
<b>Lower Limit</b>	170	0.00	63	4.6	3.55
<b>Arithmetic SD (s)</b>	9	0.10	4	0.4	0.21
<b>Arithmetic RSD (%)</b>	4.2	8.5	5.1	6.6	4.2
<b>Number of Sample Measurements (N)</b>	8	8	8	8	8

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #1, 2022: Performance of Participating Laboratories






Whole Blood Co (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
	<b>Target</b>	<b>213</b>	<b>1.18</b>	<b>79</b>	<b>6.1</b>	<b>5.05</b>
103	ICP-MS/MS	216	1.13	80.7	5.96	4.94
110	ICP-MS	223	1.24	82.3	6.39	5.24
147	ICP-MS	223	1.30	84.9	6.31	5.16
255	ICP-MS	210	1.1	79	6.1	4.9
264	ICP-MS	195.81	0.99	73.03	5.72	4.79
293	DRC/CC-ICP-MS	213.65	1.29	80.25	7.00	4.99
391	DRC/CC-ICP-MS	*137 ↓	*2.61	*56.3 ↓	*8.34 ↑	5.42
597	ICP-MS/MS	211	1.21	78.4	5.97	4.99
598	ICP-MS	209.00	1.18	72.50	5.67	*4.19

Based on the grading criteria for Co in Whole Blood, 93% of results were satisfactory, with 1 of the 9 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Whole Blood Co



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

Whole Blood Cr ( $\mu\text{g/L}$ )					
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	5.7	3.7	1.0	15.3	2.00
<b>Upper Limit</b>	7.7	5.7	3.0	18.4	4.00
<b>Lower Limit</b>	3.7	1.7	0.0	12.2	0.00
<b>Arithmetic SD (s)</b>	0.3	0.3	0.3	1.3	0.23
<b>Arithmetic RSD (%)</b>	5.8	7.8	28	8.5	12
<b>Number of Sample Measurements (N)</b>	8	8	5	9	8

The acceptable range is based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #1, 2022: Performance of Participating Laboratories






Whole Blood Cr (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
	Target	5.7	3.7	1.0	15.3	2.00
103	ICP-MS/MS	5.19	3.54	<1.50	14.1	1.62
110	DRC/CC-ICP-MS	5.96	3.64	0.59	17.0	2.06
147	DRC/CC-ICP-MS	6.26	4.26	1.19	16.8	2.34
255	ICP-MS	5.7	3.6	<1.0	15	1.9
264	ICP-MS	5.75	3.49	0.91	15.26	1.87
293	DRC/CC-ICP-MS	5.78	3.87	1.24	14.95	2.27
391	DRC/CC-ICP-MS	*10.5 ↑	*11.4 ↑	*8.93 ↑	16.4	*9.68 ↑
597	ICP-MS/MS	5.35	3.91	<0.957	14.6	1.95
598	DRC/CC-ICP-MS	5.74	3.36	1.27	13.30	1.99

Based on the grading criteria for Cr in Whole Blood, 91% of results were satisfactory, with 1 of the 9 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Whole Blood Cr



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

Whole Blood Hg (µg/L)					
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Target (Robust Mean (x*))	18.2	7.2	25.0	5.1	0.95
Upper Limit	23.7	10.2	32.5	8.1	3.95
Lower Limit	12.7	4.2	17.5	2.1	0.00
Robust SD (s*)	1.6	0.5	1.3	0.4	0.16
Robust RSD (%)	8.8	6.9	5.2	6.8	17
Number of Sample Measurements (N)	13	13	13	13	12
Standard Uncertainty (u)	0.6	0.2	0.5	0.1	0.06

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2022: Performance of Participating Laboratories






Whole Blood Hg (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
	Target	18.2	7.2	25.0	5.1	0.95
103	ICP-MS/MS	18.7	7.62	26.1	5.40	1.12
107	ICP-MS/MS	18.29	7.52	25.53	5.38	0.99
110	ICP-MS	18.0	7.26	25.0	5.12	0.92
116	ICP-MS/MS	18.9	6.82	25.4	5.16	<1.50
147	ICP-MS	18.4	7.16	26.3	5.06	0.554
264	ICP-MS	17.25	7.61	24.18	5.37	1.40
293	DRC/CC-ICP-MS	14.6	5.87	19.7	4.31	0.80
391	CV-AAS	20.8	8.55	23.9	7.29	2.30
597	ICP-MS/MS	17.1	6.68	23.4	4.92	0.976
598	ICP-MS	16.0	5.32	20.2	3.82	0.80
605	ICP-MS	20.2	7.5	26.8	5.31	0.948
606	ICP-MS/MS	17.8	7.26	25.6	4.89	0.900
686	ICP-MS	19.4	7.11	25.6	4.99	0.867

Based on the grading criteria for Hg in Whole Blood, 100% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Whole Blood Hg



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

Whole Blood Mn (µg/L)					
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
<b>Target (Robust Mean (x*))</b>	10.1	13.2	33.7	23.0	15.0
<b>Upper Limit</b>	13.1	16.2	39.4	26.9	18.0
<b>Lower Limit</b>	7.1	10.2	28.0	19.1	12.0
<b>Robust SD (s*)</b>	0.8	1.0	1.7	1.3	1.0
<b>Robust RSD (%)</b>	7.9	7.6	5.0	5.7	6.7
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (u)</b>	0.3	0.4	0.7	0.5	0.4

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ . These quality specifications were recently proposed by a network of Trace Element PT program organizers (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry Laboratory Medicine 2016; 54(12): 1921-1928).



### Results for Event #1, 2022: Performance of Participating Laboratories






Whole Blood Mn (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
	<b>Target</b>	<b>10.1</b>	<b>13.2</b>	<b>33.7</b>	<b>23.0</b>	<b>15.0</b>
103	ICP-MS/MS	10.5	13.4	34.1	22.5	15.3
107	ICP-MS/MS	10.39	13.61	33.41	23.90	15.61
110	ICP-MS	10.1	12.7	34.9	22.8	14.7
147	ICP-MS	12.0	14.8	35.6	24.2	17.1
264	ICP-MS	15.39 ↑	18.58 ↑	39.74 ↑	27.58 ↑	20.80 ↑
293	DRC/CC-ICP-MS	9.1	11.15	33.1	21.5	13.45
391	DRC/CC-ICP-MS	8.92	13.6	32.6	24.1	14.9
597	ICP-MS/MS	9.94	12.5	32.2	21.9	14.4
598	ICP-MS	9.6	12.50	30.7	22.4	13.70
606	ICP-MS/MS	9.91	12.9	33.6	22.3	14.9

Based on the grading criteria for Mn in Whole Blood, 90% of results were satisfactory, with 1 of the 10 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Whole Blood Mn



**Legend:**

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

Whole Blood Pb (µg/dL)					
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
<b>Target (Robust Mean (x*))</b>	18.3	29.1	12.2	4.46	0.94
<b>Upper Limit</b>	20.3	32.0	14.2	6.46	2.94
<b>Lower Limit</b>	16.3	26.2	10.2	2.46	0.00
<b>Robust SD (s*)</b>	0.7	1.4	0.5	0.25	0.11
<b>Robust RSD (%)</b>	3.8	4.8	4.1	5.6	12
<b>Number of Sample Measurements (N)</b>	14	14	14	13	10
<b>Standard Uncertainty (u)</b>	0.2	0.5	0.2	0.09	0.04

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/dL}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ . These quality specifications are recommended by the Clinical Laboratory Standards Institute (CLSI, C40-A2) and have been proposed for use in proficiency testing programs approved under CLIA by the Centers for Medicare and Medicaid Services (CMS) in the USA. (<https://clsi.org/standards/products/clinical-chemistry-and-toxicology/documents/c40/>)



### Results for Event #1, 2022: Performance of Participating Laboratories






Whole Blood Pb (µg/dL)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
	Target	18.3	29.1	12.2	4.46	0.94
103	ICP-MS/MS	18.7	29.9	12.4	4.45	0.944
107	ICP-MS/MS	18.395	30.369	11.948	4.608	0.963
110	ICP-MS	18.1	29.0	12.0	4.33	0.90
116	ICP-MS/MS	19.0	30.9	12.7	4.78	<3.00
147	ICP-MS	18.7	29.4	12.6	4.50	0.945
264	ICP-MS	15.60 ↓	24.81 ↓	10.32	3.88	0.77
293	DRC/CC-ICP-MS	18.20	29.36	12.20	4.3	1.03
343	ASV-LeadCare	17.1	27.2	11.3	<3.3	<3.3
391	ETAAS-Z	19.3	28.3	12.4	5.96	2.93
597	ICP-MS/MS	18.1	28.5	11.9	4.30	0.930
598	ICP-MS	17.00	24.60 ↓	10.50	4.0	0.80
605	ICP-MS	21.5 ↑	32.4 ↑	13.5	4.88	1.04
606	ICP-MS/MS	18.2	29.2	12.3	4.43	<1.00
686	ICP-MS	18.6	29.8	12.3	4.53	<1.00

Based on the grading criteria for Pb in Whole Blood, 93% of results were satisfactory, with 2 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Whole Blood Pb



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/dL}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ .



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood Mo (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
103	ICP-MS/MS	<1.50	10.0	6.23	2.32	4.66
147	ICP-MS	0.453	9.59	6.11	2.35	4.27
264	ICP-MS	*8.20	*20.90	*17.30	*10.70	*13.10
442	ICP-MS	0.463	9.32	6.12	2.37	4.18
597	ICP-MS/MS	0.400	9.53	5.83	2.29	4.17
598	DRC/CC-ICP-MS	0.611	8.49	5.48	2.63	4.19

### Summary Statistics






	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	0.48	9.4	5.9	2.39	4.29
Arithmetic SD (s)	0.09	0.6	0.3	0.14	0.21
Arithmetic RSD (%)	19	6.4	5.0	5.9	4.9
Number of Sample Measurements (N)	4	5	5	5	5

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Whole Blood Mo



**Legend:**

○ C/HHEAR Labs

● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood Sb (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
103	ICP-MS/MS	2.07	4.98	7.17	*3.90	1.63
110	ICP-MS	2.70	5.52	8.27	4.89	1.89
147	ICP-MS	2.92	6.05	8.86	5.33	2.19
264	ICP-MS	2.94	5.97	8.91	5.67	2.17
293	DRC/CC-ICP-MS	2.740	5.73	8.43	5.00	1.97
442	ICP-MS	2.96	6.19	8.89	5.51	2.38
597	ICP-MS/MS	2.72	5.27	7.92	5.01	2.01
598	ICP-MS	3.280	5.26	8.52	5.28	2.16

### Summary Statistics






	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	2.8	5.6	8.4	5.2	2.05
Arithmetic SD (s)	0.4	0.4	0.6	0.3	0.23
Arithmetic RSD (%)	13	7.1	7.1	5.5	11
Number of Sample Measurements (N)	8	8	8	7	8

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Whole Blood Sb



**Legend:**

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood Se (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
103	ICP-MS/MS	169	335	153	411	197
107	ICP-MS/MS	182.2	378.1	160.2	460.5	211.5
110	DRC/CC-ICP-MS	180	336	161	415	202
147	ICP-MS	186	356	167	442	206
293	DRC/CC-ICP-MS	171	335	155	424	197
597	ICP-MS/MS	174	332	154	402	196
598	DRC/CC-ICP-MS	165	306	*133	400	184

### Summary Statistics






	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	175	340	158	422	199
Arithmetic SD (s)	8	22	5	22	9
Arithmetic RSD (%)	4.6	6.5	3.2	5.2	4.5
Number of Sample Measurements (N)	7	7	6	7	7

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Summary Figures

### Whole Blood Se



**Legend:**

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.  
Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics






Whole Blood TI (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
103	ICP-MS/MS	0.963	1.72	1.04	2.50	0.329
110	ICP-MS	0.980	1.78	1.03	2.58	0.339
147	ICP-MS	0.940	1.68	0.998	2.47	0.325
264	ICP-MS	0.90	1.65	0.95	2.45	0.28
293	DRC/CC-ICP-MS	0.95	1.64	0.95	2.35	0.31
597	ICP-MS/MS	0.964	1.62	1.00	2.39	0.329
598	ICP-MS	0.96	1.50	0.84	2.54	0.30
Summary Statistics						
		BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )		0.95	1.66	0.97	2.47	0.316
Arithmetic SD (s)		0.03	0.09	0.07	0.08	0.021
Arithmetic RSD (%)		2.7	5.4	7.2	3.2	6.6
Number of Sample Measurements (N)		7	7	7	7	7

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Whole Blood TI



### Legend:

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.  
Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood U (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
103	ICP-MS/MS	0.294	0.166	0.210	0.336	0.0694
110	ICP-MS	0.290	0.176	0.218	0.329	0.076
147	ICP-MS	0.298	0.176	0.228	0.331	0.0698
597	ICP-MS/MS	0.274	0.162	0.209	0.307	0.0611
598	ICP-MS	0.338	0.167	0.220	0.298	0.076

### Summary Statistics






	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	0.30	0.169	0.217	0.320	0.070
Arithmetic SD (s)	0.02	0.006	0.008	0.017	0.006
Arithmetic RSD (%)	8.0	3.6	3.7	5.3	8.6
Number of Sample Measurements (N)	5	5	5	5	5

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Whole Blood U



### Legend:

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood Ba (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	ICP-MS	4.3	13.8	8.5	2.9	8.6
147	ICP-MS	4.55	14.3	8.89	2.91	9.18
597	ICP-MS/MS	4.18	13.8	8.24	2.71	8.42
598	ICP-MS	4.3	12.1	7.6	3.4	7.5

### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	4.32	13.5	8.3	3.0	8.4
Arithmetic SD (s)	0.16	1.0	0.5	0.3	0.7
Arithmetic RSD (%)	3.7	7.4	6.0	10	8.3
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

#### Whole Blood Be (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	ICP-MS	1.53	5.87	5.28	3.72	0.610
147	ICP-MS	1.35	5.59	5.09	3.29	<1.17
597	ICP-MS/MS	1.18	5.37	4.83	3.14	0.60
598	ICP-MS	0.94	5.82	4.120	3.10	0.61

#### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	1.3	5.7	4.8	3.3	0.605
Arithmetic SD (s)	0.3	0.2	0.5	0.3	0.007
Arithmetic RSD (%)	20	4.1	10	8.5	1.2
Number of Sample Measurements (N)	4	4	4	4	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Whole Blood Cs (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	ICP-MS	1.41	1.32	1.45	1.36	1.33
147	ICP-MS	1.44	1.29	1.45	1.36	1.27
597	ICP-MS/MS	1.55	1.41	1.55	1.48	1.38
598	ICP-MS	1.32	1.20	1.23	1.27	1.21
Summary Statistics						
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05	
Arithmetic Mean ( $\bar{x}$ )	1.43	1.31	1.42	1.37	1.30	
Arithmetic SD (s)	0.09	0.09	0.14	0.09	0.07	
Arithmetic RSD (%)	6.3	6.9	9.9	6.6	5.4	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

#### Whole Blood Cu (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	ICP-MS	2863	1915	834	1044	1224
147	ICP-MS	2853	1887	864	1042	1201
597	ICP-MS/MS	2650	1780	775	959	1140
598	ICP-MS	2630	1650	726	941	1020

#### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	2750	1810	800	1000	1150
Arithmetic SD (s)	130	120	60	50	90
Arithmetic RSD (%)	4.7	6.6	7.5	5.0	7.8
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

#### Whole Blood Ni (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	DRC/CC-ICP-MS	72.3	5.21	23.2	4.58	12.2
147	ICP-MS	75.2	5.54	24.4	4.16	12.9
597	ICP-MS/MS	72.5	5.02	21.8	4.03	12.7
598	ICP-MS	68.0	5.64	20.50	5.78	13.4

#### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	72	5.4	22.5	4.6	12.8
Arithmetic SD (s)	3	0.3	1.7	0.8	0.5
Arithmetic RSD (%)	4.2	5.4	7.6	17	3.9
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood Pt (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	ICP-MS	5.57	0.48	0.42	0.69	4.77
293	DRC/CC-ICP-MS	4.700	0.460	0.37	0.58	3.99
598	ICP-MS	4.780	0.417	0.63	0.66	3.88

### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	5.0	0.45	0.47	0.64	4.2
Arithmetic SD (s)	0.5	0.03	0.14	0.06	0.5
Arithmetic RSD (%)	10	6.7	30	9.4	12
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood Sn (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	ICP-MS	2.57	4.19	0.42	2.89	0.91
147	ICP-MS	2.81	4.71	<0.368	3.21	0.765
597	ICP-MS/MS	2.58	4.20	0.383	2.87	0.756
598	ICP-MS	2.42	3.440	<0.5	2.45	0.51

### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	2.60	4.1	0.40	2.9	0.74
Arithmetic SD (s)	0.16	0.5	0.02	0.3	0.17
Arithmetic RSD (%)	6.2	12	5.0	10	23
Number of Sample Measurements (N)	4	4	2	4	4

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Whole Blood Sr ( $\mu\text{g/L}$ )						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
103	ICP-MS/MS	26.0	25.8	26.2	24.3	25.2
147	ICP-MS	25.6	24.6	26.5	24.4	23.8
597	ICP-MS/MS	26.2	25.8	26.5	24.7	24.8
Summary Statistics						
		BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )		25.9	25.4	26.42	24.5	24.6
Arithmetic SD (s)		0.3	0.7	0.16	0.2	0.7
Arithmetic RSD (%)		1.2	2.8	0.61	0.90	2.8
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood Ti (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
200	DRC/CC-ICP-MS	7.4	4.6	4.4	4.8	3.5
442	DRC/CC-ICP-MS	7.90	2.65	3.43	5.12	2.89
597	ICP-MS/MS	7.87	2.95	4.65	5.64	4.25

### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	7.7	3.4	4.2	5.2	3.5
Arithmetic SD (s)	0.3	1.1	0.6	0.4	0.7
Arithmetic RSD (%)	3.9	32	14	7.7	20
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Whole Blood V (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	DRC/CC-ICP-MS	0.387	0.603	0.211	0.808	0.134
147	DRC/CC-ICP-MS	0.354	0.592	0.0842	0.745	0.117
597	ICP-MS/MS	0.376	0.661	0.144	0.766	0.188
598	DRC/CC-ICP-MS	<0.5	<0.5	<0.5	0.71	<0.5

### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	0.37	0.62	NA	0.76	0.15
Arithmetic SD (s)	0.02	0.04	NA	0.04	0.04
Arithmetic RSD (%)	4.6	6.5	NA	5.3	27
Number of Sample Measurements (N)	3	3	NA	4	3

\*Denotes a statistical Outlier.

Statistical data was not calculated for BE22-03 based on a lack of consensus among participating labs.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Whole Blood W (µg/L)						
Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	ICP-MS	1.92	0.767	1.18	0.191	0.709
200	ICP-MS	2.10	0.77	1.310	0.22	0.77
597	ICP-MS/MS	1.80	0.696	1.13	0.189	0.647
598	ICP-MS	2.05	0.86	1.08	0.26	0.55
Summary Statistics						
	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05	
Arithmetic Mean ( $\bar{x}$ )	1.97	0.77	1.18	0.21	0.67	
Arithmetic SD (s)	0.14	0.07	0.10	0.03	0.09	
Arithmetic RSD (%)	7.1	9.1	8.5	14	13	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

#### Whole Blood Zn (µg/L)

Lab Code	Method	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
110	ICP-MS	6220	7790	5956	5172	7781
147	ICP-MS	6255	7647	6190	5196	7712
597	ICP-MS/MS	6240	7780	6010	5150	7820
598	ICP-MS	5810	6720	5300	4740	6310

#### Summary Statistics

	BE22-01	BE22-02	BE22-03	BE22-04	BE22-05
Arithmetic Mean ( $\bar{x}$ )	6130	7500	5900	5060	7400
Arithmetic SD (s)	210	500	400	220	700
Arithmetic RSD (%)	3.4	6.7	6.8	4.3	9.5
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



Results for Event #1, 2022:
Additional Elements in Whole Blood

Whole Blood Ag (µg/L)

Table with 7 columns: Lab Code, Method, BE22-01, BE22-02, BE22-03, BE22-04, BE22-05. Row 1: 147, ICP-MS, <0.302, <0.302, <0.302, <0.302, <0.302

Whole Blood Al (µg/L)

Table with 7 columns: Lab Code, Method, BE22-01, BE22-02, BE22-03, BE22-04, BE22-05. Row 1: 147, ICP-MS, <5.13, <5.13, <5.13, <5.13, <5.13. Row 2: 597, ICP-MS/MS, <4.34, 4.81, <4.34, 0.00, 7.55

Whole Blood Bi (µg/L)

Table with 7 columns: Lab Code, Method, BE22-01, BE22-02, BE22-03, BE22-04, BE22-05. Row 1: 147, ICP-MS, <0.0334, <0.0334, <0.0334, <0.0334, <0.0334. Row 2: 597, ICP-MS/MS, <0.0423, <0.0423, <0.0423, <0.0423, <0.0423

Whole Blood I (µg/L)

Table with 7 columns: Lab Code, Method, BE22-01, BE22-02, BE22-03, BE22-04, BE22-05. Row 1: 147, ICP-MS, 32.8, 22.9, 33.8, 47.3, 22.3

Whole Blood Li (µg/L)

Table with 7 columns: Lab Code, Method, BE22-01, BE22-02, BE22-03, BE22-04, BE22-05. Row 1: 147, ICP-MS, 0.660, 0.736, 0.620, 1.11, 0.644

Whole Blood Mg (µg/L)

Table with 7 columns: Lab Code, Method, BE22-01, BE22-02, BE22-03, BE22-04, BE22-05. Row 1: 597, ICP-MS/MS, 26800, 30000, 26900, 27900, 29800

Whole Blood Te (µg/L)

Table with 7 columns: Lab Code, Method, BE22-01, BE22-02, BE22-03, BE22-04, BE22-05. Row 1: 147, ICP-MS, <0.117, <0.117, <0.117, <0.117, <0.117

Whole Blood Th (µg/L)

Table with 7 columns: Lab Code, Method, BE22-01, BE22-02, BE22-03, BE22-04, BE22-05. Row 1: 147, ICP-MS, <0.0278, <0.0278, <0.0278, <0.0278, <0.0278. Row 2: 597, ICP-MS/MS, 0.00887, 0.0233, <0.00875, <0.00875, 0.00986



**Department  
of Health**

**Wadsworth  
Center**

**Event #1, 2022**

**Trace Elements in  
Urine**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #1, 2022: Trace Elements in Urine

### PT Materials

Urine was collected from volunteer donors into polyethylene containers and stored at 4°C. Following collection, urine was acidified to 1% (v/v) with nitric acid and mixed with a sulfamic acid solution (stock solution contained 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100) to a final concentration of 1% (v/v) to stabilize Hg. Urine was stored frozen at -80°C pending further preparation. The urine was thawed at room temperature and precipitated salts removed by centrifugation. Urine supernatants were combined into five separate pools. Each urine pool was supplemented with arsenic (As), barium (Ba), beryllium (Be), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), thallium (Tl), uranium (U), aluminum (Al), cesium (Cs), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), strontium (Sr), tellurium (Te), titanium (Ti), vanadium (V), tungsten (W), and zinc (Zn). PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Eleven elements in urine are formally graded: As, Ba, Be, Cd, Co, Cr, Hg, Mn, Pb, Tl, and U. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 23 elements were reported by at least one participant: Ag, Al, B, Bi, Cs, Cu, Fe, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #1, 2022: Summary Statistics

	Urine As (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	13.8	47.9	111	4.6	34.9
<b>Upper Limit</b>	19.8	57.5	133	10.6	41.9
<b>Lower Limit</b>	7.8	38.3	89	0.0	27.9
<b>Robust SD (s*)</b>	0.5	2.0	4	0.4	1.6
<b>Robust RSD (%)</b>	3.5	4.2	3.9	7.8	4.6
<b>Number of Sample Measurements (N)</b>	16	16	16	15	16
<b>Standard Uncertainty (u)</b>	0.1	0.6	1	0.1	0.5

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2022: Performance of Participating Laboratories






		Urine As (µg/L)				
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
Target		13.8	47.9	111	4.6	34.9
103	ICP-MS/MS	13.7	47.2	111	4.18	33.1
107	DRC/CC-ICP-MS	13.70	46.78	106.76	4.24	32.93
110	DRC/CC-ICP-MS	14.6	51.1	117	4.94	36.3
116	ICP-MS/MS	13.5	45.8	112	4.78	36.9
147	ICP-MS	13.3	47.5	109	<4.04	33.9
264	ICP-MS	14.29	50.19	114.38	4.59	36.37
293	DRC/CC-ICP-MS	13.38	45.84	109.61	5.14	33.95
324	ICP-MS	13.691	47.616	111.999	4.398	34.004
391	DRC/CC-ICP-MS	32.8 ↑	61.9 ↑	137 ↑	61.0 ↑	64.7 ↑
399	DRC/CC-ICP-MS	13.9	48.0	114	4.50	33.8
442	DRC/CC-ICP-MS	14.5	49.7	117	4.49	35.5
597	ICP-MS/MS	13.4	46.1	108	4.65	32.9
598	DRC/CC-ICP-MS	13.4	47	106	4.16	34.6
605	ICP-MS	13.7	48.9	111	4.53	35.1
606	ICP-MS/MS	13.5	48.5	112	4.53	35.7
686	DRC/CC-ICP-MS	16.8	45.2	104	6.46	36.0

Based on the grading criteria for As in Urine, 94% of results were satisfactory, with 1 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine As



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

	Urine Ba (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	13.1	1.04	5.8	2.08	8.9
<b>Upper Limit</b>	15.7	2.04	7.0	3.08	10.7
<b>Lower Limit</b>	10.5	0.04	4.6	1.08	7.1
<b>Robust SD (s*)</b>	0.5	0.06	0.3	0.09	0.3
<b>Robust RSD (%)</b>	3.8	5.8	4.5	4.3	3.6
<b>Number of Sample Measurements (N)</b>	11	10	11	11	11
<b>Standard Uncertainty (u)</b>	0.2	0.02	0.1	0.03	0.1

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2022: Performance of Participating Laboratories






		Urine Ba (µg/L)				
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>13.1</b>	<b>1.04</b>	<b>5.8</b>	<b>2.08</b>	<b>8.9</b>
107	ICP-MS	13.314	1.049	5.983	2.056	8.759
110	ICP-MS	12.8	1.04	5.61	2.07	8.69
116	ICP-MS/MS	12.4	0.951	5.56	1.94	8.54
147	ICP-MS	13.0	1.02	5.70	2.06	8.83
264	ICP-MS	11.26	0.89	4.93	1.71	7.32
399	ICP-MS/MS	13.7	1.12	6.15	2.20	9.12
597	ICP-MS/MS	12.7	<1.56	5.90	2.08	8.98
598	ICP-MS	13.2	1	5.98	2.12	9.21
605	ICP-MS	13.6	1.11	6.08	2.21	9.33
606	ICP-MS/MS	13.4	1.05	5.66	2.43	9.09
686	ICP-MS	13.5	1.05	5.95	2.03	9.16

Based on the grading criteria for Ba in Urine, 100% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine Ba



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

	Urine Be (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	0.82	12.8	8.6	1.50	2.34
<b>Upper Limit</b>	1.82	15.4	10.3	2.50	3.34
<b>Lower Limit</b>	0.00	10.2	6.9	0.50	1.34
<b>Robust SD (s*)</b>	0.05	0.6	0.5	0.07	0.10
<b>Robust RSD (%)</b>	6.1	4.7	5.8	4.7	4.3
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.02	0.2	0.2	0.03	0.04

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2022: Performance of Participating Laboratories






		Urine Be (µg/L)				
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>0.82</b>	<b>12.8</b>	<b>8.6</b>	<b>1.50</b>	<b>2.34</b>
107	ICP-MS	0.793	12.769	8.531	1.572	2.291
110	ICP-MS	0.911	14.4	9.80	1.74	2.65
116	ICP-MS/MS	0.842	13	8.39	1.53	2.28
147	ICP-MS	0.79	12.5	8.74	1.46	2.36
264	ICP-MS	0.69	10.87	7.28	1.28	1.92
293	ICP-MS	0.83	12.65	8.55	1.51	2.35
399	ICP-MS/MS	0.807	12.8	8.24	1.50	2.32
597	ICP-MS/MS	0.787	12.1	7.98	1.44	2.23
598	ICP-MS	0.824	13.7	9.04	1.46	2.51
605	ICP-MS	0.878	13.7	9.52	1.55	2.39
686	ICP-MS	0.865	12.8	8.76	1.52	2.38

Based on the grading criteria for Be in Urine, 100% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine Be



**Legend:**  
○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
±1 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 5 µg/L.



### Results for Event #1, 2022: Summary Statistics

	Urine Cd (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	1.78	0.59	6.00	4.13	3.35
<b>Upper Limit</b>	2.78	1.59	7.00	5.13	4.35
<b>Lower Limit</b>	0.78	0.00	5.00	3.13	2.35
<b>Robust SD (s*)</b>	0.07	0.04	0.24	0.19	0.18
<b>Robust RSD (%)</b>	3.9	6.3	4.0	4.6	5.4
<b>Number of Sample Measurements (N)</b>	16	15	16	16	16
<b>Standard Uncertainty (u)</b>	0.02	0.01	0.07	0.06	0.06

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2022: Performance of Participating Laboratories






Lab Code	Method	Urine Cd (µg/L)				
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>1.78</b>	<b>0.59</b>	<b>6.00</b>	<b>4.13</b>	<b>3.35</b>
103	ICP-MS/MS	1.82	0.561	5.92	4.06	3.38
107	DRC/CC-ICP-MS	1.848	0.577	6.240	4.175	3.358
110	ICP-MS	1.83	0.572	5.71	3.87	3.27
116	ICP-MS/MS	1.81	0.557	5.92	3.87	3.18
147	ICP-MS	1.94	0.605	6.61	4.34	3.68
264	ICP-MS	1.73	0.59	6.00	4.04	3.30
293	DRC/CC-ICP-MS	1.76	0.62	6.01	4.29	3.47
324	ICP-MS	1.734	<1	5.899	4.019	3.495
391	DRC/CC-ICP-MS	1.66	0.549	6.23	4.41	3.09
399	DRC/CC-ICP-MS	1.81	0.591	6.11	4.24	3.43
442	DRC/CC-ICP-MS	1.78	0.602	5.89	4.22	3.22
597	ICP-MS/MS	1.75	0.572	5.82	4.17	3.21
598	DRC/CC-ICP-MS	1.66	0.644	5.41	3.71	3.18
605	ICP-MS	1.85	0.617	6.35	4.22	3.48
606	ICP-MS/MS	1.71	0.537	5.85	3.95	3.23
686	ICP-MS	1.81	0.662	6.25	4.20	3.69

Based on the grading criteria for Cd in Urine, 100% of results were satisfactory, with 0 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine Cd



**Legend:**  
 ○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ .



### Results for Event #1, 2022: Summary Statistics

	Urine Co (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	8.9	0.83	4.47	4.2	4.05
<b>Upper Limit</b>	10.4	2.33	5.97	5.7	5.55
<b>Lower Limit</b>	7.4	0.00	2.97	2.7	2.55
<b>Robust SD (s*)</b>	0.6	0.06	0.19	0.6	0.18
<b>Robust RSD (%)</b>	6.7	7.2	4.3	14	4.4
<b>Number of Sample Measurements (N)</b>	13	12	13	13	13
<b>Standard Uncertainty (u)</b>	0.2	0.02	0.07	0.2	0.06

The acceptable range is based on quality specifications: ±1.5 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 10 µg/L. These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #1, 2022: Performance of Participating Laboratories






Lab Code	Method	Urine Co (µg/L)				
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>8.9</b>	<b>0.83</b>	<b>4.47</b>	<b>4.2</b>	<b>4.05</b>
103	ICP-MS/MS	9.16	0.803	4.47	4.20	4.00
107	ICP-MS	8.424	0.784	4.277	3.586	3.907
110	ICP-MS	9.90	0.912	4.95	4.75	4.54
147	ICP-MS	8.31	0.760	4.16	4.67	3.89
264	ICP-MS	8.27	0.76	4.16	4.56	3.83
293	DRC/CC-ICP-MS	9.06	0.85	4.47	4.83	4.06
324	ICP-MS	8.559	<1	4.364	3.310	3.953
391	DRC/CC-ICP-MS	10.5 ↑	2.71 ↑	6.73 ↑	4.95	5.71 ↑
399	DRC/CC-ICP-MS	9.12	0.817	4.58	4.31	4.18
597	ICP-MS/MS	8.55	0.797	4.44	3.73	3.87
598	ICP-MS	8.55	0.857	4.42	4.44	4.02
605	ICP-MS	9.18	0.835	4.66	3.60	4.12
606	ICP-MS/MS	9.05	0.879	4.56	3.63	4.19

Based on the grading criteria for Co in Urine, 94% of results were satisfactory, with 1 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine Co



**Legend:**  
 ○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



### Results for Event #1, 2022: Summary Statistics

	Urine Cr (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	5.2	11.5	1.00	1.16	20.4
<b>Upper Limit</b>	8.2	14.5	4.00	4.16	24.5
<b>Lower Limit</b>	2.2	8.5	0.00	0.00	16.3
<b>Robust SD (s*)</b>	0.5	1.0	0.15	0.23	1.4
<b>Robust RSD (%)</b>	9.6	8.7	15	20	6.9
<b>Number of Sample Measurements (N)</b>	12	12	12	12	12
<b>Standard Uncertainty (u)</b>	0.2	0.3	0.05	0.08	0.5

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $15 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #1, 2022: Performance of Participating Laboratories






		Urine Cr (µg/L)				
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
Target		5.2	11.5	1.00	1.16	20.4
103	ICP-MS/MS	4.51	10.5	0.859	0.766	19.0
107	DRC/CC-ICP-MS	5.08	11.35	0.87	0.99	19.48
110	DRC/CC-ICP-MS	5.70	12.6	1.08	1.37	22.1
116	ICP-MS/MS	5.58	12.6	1.15	1.30	23.0
147	DRC/CC-ICP-MS	4.81	10.8	0.842	1.12	19.9
264	ICP-MS	4.67	10.61	0.93	1.20	19.19
293	DRC/CC-ICP-MS	5.29	11.24	1.02	1.01	20.07
324	ICP-MS	5.410	11.777	1.152	1.325	21.042
391	DRC/CC-ICP-MS	14.3 ↑	18.9 ↑	6.89 ↑	24.0 ↑	34.6 ↑
597	ICP-MS/MS	4.77	10.9	0.922	1.00	19.1
598	DRC/CC-ICP-MS	4.97	10.8	0.895	0.994	19.6
605	ICP-MS	5.30	12.5	1.05	1.29	21.7

Based on the grading criteria for Cr in Urine, 92% of results were satisfactory, with 1 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine Cr



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 3 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $15 \mu\text{g/L}$ .



### Results for Event #1, 2022: Summary Statistics

	Urine Hg (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	45	18.6	0.38	5.9	24.1
<b>Upper Limit</b>	59	24.2	3.38	8.9	31.3
<b>Lower Limit</b>	32	13.0	0.00	2.9	16.9
<b>Robust SD (s*)</b>	4	1.4	0.16	0.4	2.2
<b>Robust RSD (%)</b>	8.0	7.5	42	7.3	9.1
<b>Number of Sample Measurements (N)</b>	14	14	10	14	14
<b>Standard Uncertainty (u)</b>	1	0.5	0.06	0.1	0.7

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2022: Performance of Participating Laboratories






Lab Code	Method	Urine Hg (µg/L)				
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>45</b>	<b>18.6</b>	<b>0.38</b>	<b>5.9</b>	<b>24.1</b>
103	ICP-MS/MS	45.1	18.3	0.303	6.06	25.6
107	DRC/CC-ICP-MS	44.85	18.87	0.22	5.63	24.05
110	ICP-MS	44.1	18.7	0.546	5.95	23.4
116	ICP-MS/MS	38.2	15.4	<0.500	4.78	20.6
147	ICP-MS	46.1	18.4	0.358	5.90	24.9
264	ICP-MS	48.04	19.68	0.31	6.30	26.07
293	DRC/CC-ICP-MS	46.15	18.69	0.43	6.04	25.66
391	DRC/CC-ICP-MS	61.9 ↑	25.1 ↑	1.11	7.50	18.1
442	DRC/CC-ICP-MS	48.1	19.3	0.214	6.18	26.2
597	ICP-MS/MS	42.9	17.8	0.264	5.86	24.0
598	ICP-MS	39.5	16.3	0.555	4.94	22
605	ICP-MS	48.0	19.7	<1.0	5.86	24.6
606	ICP-MS/MS	41.6	17.2	<1.00	5.38	22.4
686	ICP-MS	47.2	20.0	<1.00	6.29	26.9

Based on the grading criteria for Hg in Urine, 97% of results were satisfactory, with 1 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine Hg



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



### Results for Event #1, 2022: Summary Statistics

	Urine Mn (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	0.42	8.0	1.75	4.4	3.04
<b>Upper Limit</b>	0.97	10.0	2.30	5.5	3.80
<b>Lower Limit</b>	0.00	6.0	1.20	3.3	2.28
<b>Robust SD (s*)</b>	0.05	0.5	0.13	0.4	0.19
<b>Robust RSD (%)</b>	12	6.3	7.4	8.3	6.3
<b>Number of Sample Measurements (N)</b>	13	14	14	14	14
<b>Standard Uncertainty (u)</b>	0.02	0.2	0.04	0.1	0.06

The acceptable range is based on quality specifications:  $\pm 0.55 \mu\text{g/L}$  or  $\pm 25\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.55 \mu\text{g/L}$  at concentrations less than or equal to  $2.2 \mu\text{g/L}$ . Quality specifications for Mn are consistent with those used by other External Quality Assessment Schemes for trace elements. (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry and Laboratory Medicine.2016; 54(12): 1921-1928).



### Results for Event #1, 2022: Performance of Participating Laboratories






Lab Code	Method	Urine Mn (µg/L)				
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>0.42</b>	<b>8.0</b>	<b>1.75</b>	<b>4.4</b>	<b>3.04</b>
103	ICP-MS/MS	0.398	7.93	1.70	4.23	2.83
107	DRC/CC-ICP-MS	0.407	7.999	1.822	4.336	2.975
110	DRC/CC-ICP-MS	0.485	8.46	1.92	4.64	3.25
116	ICP-MS/MS	0.420	8.98	1.88	4.90	3.47
147	DRC/CC-ICP-MS	0.414	7.47	1.65	4.23	2.92
264	ICP-MS	0.39	7.47	1.62	4.10	2.89
293	DRC/CC-ICP-MS	0.5	7.59	1.74	4.24	2.92
324	ICP-MS	<1	8.052	1.767	4.776	3.225
391	DRC/CC-ICP-MS	0.904	5.14 ↓	0.904 ↓	3.28 ↓	1.99 ↓
399	DRC/CC-ICP-MS	0.382	8.32	1.76	4.76	3.18
597	ICP-MS/MS	0.322	7.57	1.59	4.36	2.99
598	ICP-MS	0.436	7.93	1.76	4.62	3.15
605	ICP-MS	0.468	8.63	2.08	4.19	3.08
606	ICP-MS/MS	0.384	8.04	1.74	5.82 ↑	3.05

Based on the grading criteria for Mn in Urine, 93% of results were satisfactory, with 1 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine Mn



**Legend:**  
 ○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 0.55 \mu\text{g/L}$  or  $\pm 25\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.55 \mu\text{g/L}$  at concentrations less than or equal to  $2.2 \mu\text{g/L}$ .



### Results for Event #1, 2022: Summary Statistics

	Urine Pb (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	10.8	2.74	6.52	14.6	1.01
<b>Upper Limit</b>	13.0	3.74	7.82	17.5	2.01
<b>Lower Limit</b>	8.6	1.74	5.22	11.7	0.01
<b>Robust SD (s*)</b>	0.4	0.09	0.21	0.5	0.05
<b>Robust RSD (%)</b>	3.3	3.3	3.2	3.2	4.6
<b>Number of Sample Measurements (N)</b>	16	16	16	16	16
<b>Standard Uncertainty (u)</b>	0.1	0.03	0.06	0.1	0.01

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2022: Performance of Participating Laboratories






		Urine Pb (µg/L)				
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>10.8</b>	<b>2.74</b>	<b>6.52</b>	<b>14.6</b>	<b>1.01</b>
103	ICP-MS/MS	11.0	2.77	6.77	15.0	1.03
107	ICP-MS	11.055	2.779	6.577	14.755	1.020
110	ICP-MS	10.8	2.82	6.69	14.4	1.04
116	ICP-MS/MS	10.8	2.72	6.48	14.4	0.95
147	ICP-MS	10.9	2.74	6.67	14.8	1.10
264	ICP-MS	10.42	2.62	6.30	13.67	0.92
293	DRC/CC-ICP-MS	10.63	2.77	6.55	14.27	1.01
324	ICP-MS	10.462	2.708	6.203	14.446	1.011
391	DRC/CC-ICP-MS	11.1	2.33	6.69	16.4	0.986
399	ICP-MS/MS	10.4	2.66	6.32	14.3	1.01
442	DRC/CC-ICP-MS	11.1	2.85	6.6	15.4	0.981
597	ICP-MS/MS	10.5	2.74	6.38	14.3	0.991
598	ICP-MS	10.1	2.66	6.39	13.6	0.94
605	ICP-MS	11.5	2.95	7.10	15.5	1.07
606	ICP-MS/MS	10.7	2.73	6.35	14.6	0.984
686	ICP-MS	11.0	2.82	6.57	15.2	1.06

Based on the grading criteria for Pb in Urine, 100% of results were satisfactory, with 0 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine Pb



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



### Results for Event #1, 2022: Summary Statistics

	Urine TI (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	3.64	4.05	1.73	0.69	2.73
<b>Upper Limit</b>	4.37	4.86	2.08	0.89	3.28
<b>Lower Limit</b>	2.91	3.24	1.38	0.49	2.18
<b>Robust SD (s*)</b>	0.12	0.19	0.08	0.03	0.13
<b>Robust RSD (%)</b>	3.3	4.7	4.6	4.5	4.8
<b>Number of Sample Measurements (N)</b>	14	14	14	14	14
<b>Standard Uncertainty (u)</b>	0.04	0.06	0.03	0.01	0.04

The acceptable range is based on quality specifications:  $\pm 0.2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.2 \mu\text{g/L}$  at concentrations less than or equal to  $1 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2022: Performance of Participating Laboratories






Lab Code	Method	Urine TI (µg/L)				
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>3.64</b>	<b>4.05</b>	<b>1.73</b>	<b>0.69</b>	<b>2.73</b>
103	ICP-MS/MS	3.72	4.18	1.79	0.713	2.86
107	ICP-MS	3.687	4.186	1.751	0.691	2.685
110	ICP-MS	4.45	↑ 4.30	2.08	0.786	2.94
116	ICP-MS/MS	3.64	4.12	1.73	0.687	2.73
147	ICP-MS	3.43	3.84	1.62	0.711	2.72
264	ICP-MS	3.47	3.81	1.64	0.63	2.53
293	DRC/CC-ICP-MS	3.66	4.04	1.76	0.69	2.7
391	DRC/CC-ICP-MS	3.72	3.70	1.80	0.698	2.79
399	ICP-MS/MS	3.61	4.05	1.71	0.683	2.70
597	ICP-MS/MS	3.62	3.98	1.66	0.662	2.62
598	ICP-MS	3.52	3.93	1.68	0.627	2.55
605	ICP-MS	3.91	4.38	1.83	0.73	2.92
606	ICP-MS/MS	3.57	4.01	1.70	0.671	2.70
686	ICP-MS	3.63	4.13	1.68	0.687	2.81

Based on the grading criteria for TI in Urine, 99% of results were satisfactory, with 0 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine TI



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 0.2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.2 \mu\text{g/L}$  at concentrations less than or equal to  $1 \mu\text{g/L}$ .



### Results for Event #1, 2022: Summary Statistics

	Urine U (µg/L)				
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Target (Robust Mean (x*))</b>	0.039	0.241	0.314	0.151	0.193
<b>Upper Limit</b>	0.069	0.289	0.377	0.181	0.232
<b>Lower Limit</b>	0.009	0.193	0.251	0.121	0.154
<b>Robust SD (s*)</b>	0.003	0.018	0.026	0.010	0.010
<b>Robust RSD (%)</b>	8.5	7.5	8.3	6.6	5.2
<b>Number of Sample Measurements (N)</b>	13	14	14	14	14
<b>Standard Uncertainty (u)</b>	0.001	0.006	0.009	0.003	0.003

The acceptable range is based on quality specifications:  $\pm 0.03 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.03 \mu\text{g/L}$  at concentrations less than or equal to  $0.15 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #1, 2022: Performance of Participating Laboratories






Lab Code	Method	Urine U (µg/L)				
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
	<b>Target</b>	<b>0.039</b>	<b>0.241</b>	<b>0.314</b>	<b>0.151</b>	<b>0.193</b>
103	ICP-MS/MS	0.0391	0.256	0.336	0.157	0.201
107	ICP-MS	0.0414	0.2596	0.3446	0.1625	0.2033
110	ICP-MS	0.042	0.255	0.330	0.166	0.205
116	ICP-MS/MS	0.0365	0.241	0.29	0.151	0.200
147	ICP-MS	0.0381	0.235	0.295	0.143	0.190
264	ICP-MS	0.04	0.18 ↓	0.25 ↓	0.14	0.17
324	ICP-MS	0.037	0.228	0.300	0.143	0.192
391	DRC/CC-ICP-MS	0.025	0.209	0.312	0.141	0.180
399	ICP-MS/MS	0.0320	0.218	0.297	0.143	0.188
597	ICP-MS/MS	<0.00695 ↓	0.249	0.346	0.158	0.194
598	ICP-MS	0.043	0.258	0.293	0.143	0.178
605	ICP-MS	0.040	0.256	0.341	0.160	0.203
606	ICP-MS/MS	0.041	0.236	0.314	0.154	0.198
686	ICP-MS	0.0389	0.247	0.320	0.155	0.194

Based on the grading criteria for U in Urine, 96% of results were satisfactory, with 1 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Urine U



**Legend:**  
○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
±0.03 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.03 µg/L at concentrations less than or equal to 0.15 µg/L.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics






Urine AI (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
147	ICP-MS	<9.44	12.8	20.9	12.5	14.0
264	ICP-MS	6.54	14.26	20.33	13.52	8.83
293	DRC/CC-ICP-MS	7.8	15.86	24.46	13.71	9.41
324	ICP-MS	8.090	14.802	19.377	16.555	11.949
597	ICP-MS/MS	<4.81	12.7	18.9	8.75	10.4
Summary Statistics						
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		7.5	14.1	21	13	10.9
<b>Arithmetic SD (s)</b>		0.8	1.3	2	3	2.1
<b>Arithmetic RSD (%)</b>		11	9.2	11	22	19
<b>Number of Sample Measurements (N)</b>		3	5	5	5	5

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Summary Figures

### Urine AI



#### Legend:

- C/HHEAR Labs
- Other Labs
- Horizontal purple line = arithmetic mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics






Urine Cs (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
107	ICP-MS	1.74	4.25	6.32	13.55	1.48
110	ICP-MS	1.63	4.12	6.01	13.6	1.42
147	ICP-MS	1.67	4.11	6.27	13.4	1.45
264	ICP-MS	*1.24	*3.05	*4.50	*9.86	*1.05
399	ICP-MS/MS	1.72	4.29	6.40	14.2	1.50
597	ICP-MS/MS	1.76	4.26	6.30	13.2	1.56
598	ICP-MS	1.7	4.2	6.31	13.6	1.46
605	ICP-MS	1.75	4.40	6.50	14.5	1.52
606	ICP-MS/MS	1.74	4.28	6.28	14.2	1.49
Summary Statistics						
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05	
Arithmetic Mean ( $\bar{x}$ )	1.71	4.24	6.30	13.8	1.48	
Arithmetic SD (s)	0.05	0.10	0.14	0.5	0.04	
Arithmetic RSD (%)	2.9	2.4	2.2	3.6	2.7	
Number of Sample Measurements (N)	8	8	8	8	8	

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Urine Cs



### Legend:

- C/HHEAR Labs
- Other Labs
- Horizontal purple line = arithmetic mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics






Urine Cu (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
110	ICP-MS	80.0	17.1	115	43.8	89.0
116	ICP-MS/MS	66.0	12.0	102	38.7	82.8
147	ICP-MS	71.8	16.5	106	46.6	91.5
264	ICP-MS	62.60	11.80	99.27	36.85	78.77
293	DRC/CC-ICP-MS	61.03	11.44	98.54	35.6	76.92
324	ICP-MS	63.364	11.743	98.422	38.709	78.713
391	DRC/CC-ICP-MS	67.5	16.3	102	46.5	80.8
597	ICP-MS/MS	61.5	12.0	97.1	36.2	76.0
598	ICP-MS	58	11.7	93.1	34.3	72.3
Summary Statistics						
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05	
Arithmetic Mean ( $\bar{x}$ )	66	13.4	101	40	81	
Arithmetic SD (s)	7	2.4	6	5	6	
Arithmetic RSD (%)	11	18	5.9	13	7.4	
Number of Sample Measurements (N)	9	9	9	9	9	

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Urine Cu



### Legend:

- C/HHEAR Labs
- Other Labs
- Horizontal purple line = arithmetic mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

#### Urine Mo (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
103	ICP-MS/MS	10.9	30.3	77.2	33.2	133
107	ICP-MS	10.98	31.90	76.78	34.14	134.46
110	ICP-MS	11.8	33.8	80.5	36.4	141
147	ICP-MS	10.4	29.2	69.8	32.1	127
264	ICP-MS	6.34	19.79	49.89	22.41	90.24
293	DRC/CC-ICP-MS	11.33	32.24	81.61	34.3	136.29
324	ICP-MS	10.865	30.913	77.842	32.803	132.157
399	ICP-MS/MS	11.3	32.1	79.2	34.7	138
597	ICP-MS/MS	10.2	29.1	71.9	30.9	123
598	DRC/CC-ICP-MS	11	31.5	73.2	34.3	142
605	ICP-MS	11.3	32.2	77.7	35.2	136
606	ICP-MS/MS	11.2	31.4	78.1	34.3	136

#### Summary Statistics






	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Robust Mean (x*)</b>	11.0	31.2	77	33.7	135
<b>Robust SD (s*)</b>	0.5	1.3	4	1.5	5
<b>Robust RSD (%)</b>	4.5	4.2	5.0	4.5	3.7
<b>Number of Sample Measurements (N)</b>	12	12	12	12	12
<b>Standard Uncertainty (u)</b>	0.2	0.5	1	0.5	2

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Urine Mo



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = robust mean of all laboratories.  
 Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

#### Urine Ni (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
107	DRC/CC-ICP-MS	10.57	4.69	2.27	8.79	4.80
110	ICP-MS	12.0	5.75	3.00	10.1	6.02
147	ICP-MS	11.5	4.98	2.58	8.75	8.22
264	ICP-MS	10.12	4.50	2.32	8.21	4.54
293	DRC/CC-ICP-MS	11.14	4.7	2.41	8.4	4.6
324	ICP-MS	10.566	4.814	2.583	8.413	5.010
391	DRC/CC-ICP-MS	13.9	11.4	10.3	13.7	12.0
442	DRC/CC-ICP-MS	11.0	4.83	2.43	8.57	5.97
597	ICP-MS/MS	9.70	4.31	2.09	7.91	4.08
598	ICP-MS	10.5	5.09	2.6	9.07	5.16
605	ICP-MS	11.2	4.93	2.54	8.46	4.58

#### Summary Statistics






	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Robust Mean (x*)</b>	11.0	4.9	2.51	8.6	5.2
<b>Robust SD (s*)</b>	0.8	0.3	0.23	0.4	0.8
<b>Robust RSD (%)</b>	7.3	5.6	9.2	4.5	15
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.3	0.1	0.09	0.1	0.3

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Urine Ni



### Legend:

- C/HHEAR Labs
- Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics






Urine Pt (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
107	ICP-MS	0.6889	2.1022	0.3756	1.1195	3.2464
110	ICP-MS	0.703	2.16	0.394	1.15	3.34
264	ICP-MS	0.62	1.89	0.35	1.03	2.93
293	DRC/CC-ICP-MS	0.68	2	0.36	1.05	3.08
399	ICP-MS/MS	0.676	2.08	0.359	1.11	3.16
598	ICP-MS	0.747	2.01	0.422	1.13	3.17
605	ICP-MS	0.665	2.05	0.369	1.09	3.19
Summary Statistics						
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
Arithmetic Mean ( $\bar{x}$ )		0.68	2.04	0.376	1.10	3.16
Arithmetic SD (s)		0.04	0.09	0.025	0.04	0.13
Arithmetic RSD (%)		5.6	4.4	6.6	3.6	4.1
Number of Sample Measurements (N)		7	7	7	7	7

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Urine Pt



### Legend:

- C/HHEAR Labs
- Other Labs
- Horizontal purple line = arithmetic mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

Urine Sb (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
103	ICP-MS/MS	0.213	2.23	0.991	0.800	0.526
107	ICP-MS	0.245	2.323	1.026	0.814	0.517
110	ICP-MS	0.218	2.33	0.977	0.800	0.577
147	ICP-MS	0.246	2.54	1.12	0.928	0.603
264	ICP-MS	0.11	1.32	0.59	0.49	0.31
293	DRC/CC-ICP-MS	0.27	2.45	1.1	0.8	0.57
399	ICP-MS/MS	0.226	2.46	1.08	0.884	0.520
597	ICP-MS/MS	0.286	2.36	1.08	0.836	0.582
598	ICP-MS	0.221	2.3	1.06	0.843	0.526
605	ICP-MS	<0.8	2.51	1.10	0.883	<0.8
606	ICP-MS/MS	0.230	2.26	1.04	0.837	0.559






Summary Statistics					
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Robust Mean (x*)</b>	0.23	2.36	1.05	0.83	0.55
<b>Robust SD (s*)</b>	0.03	0.14	0.06	0.05	0.04
<b>Robust RSD (%)</b>	11	5.9	5.7	6.0	7.3
<b>Number of Sample Measurements (N)</b>	10	11	11	11	10
<b>Standard Uncertainty (u)</b>	0.01	0.05	0.02	0.02	0.02

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Summary Figures

### Urine Sb



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = robust mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics






Urine Se (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
103	ICP-MS/MS	261	30.4	174	107	24.7
110	DRC/CC-ICP-MS	293	37.4	184	132	30.9
147	ICP-MS	258	32.5	175	115	27.2
293	DRC/CC-ICP-MS	223.36	29.99	146.01	105.76	23.68
597	ICP-MS/MS	231	31.1	157	107	25.5
598	DRC/CC-ICP-MS	252	30	153	110	23.8
Summary Statistics						
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
Arithmetic Mean ( $\bar{x}$ )		253	32	165	113	26
Arithmetic SD (s)		25	3	15	10	3
Arithmetic RSD (%)		9.9	9.1	9.1	8.8	10
Number of Sample Measurements (N)		6	6	6	6	6

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Summary Figures

### Urine Se



#### Legend:

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Urine Sn (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
107	ICP-MS	1.47	0.67	0.23	2.30	5.04
110	ICP-MS	1.30	0.68	<0.21	2.44	5.03
147	ICP-MS	1.07	0.588	<0.238	1.85	4.01
264	ICP-MS	*0.37	*0.14	*0.01	*0.59	*1.51
399	ICP-MS/MS	1.41	0.638	<0.300	2.44	5.55
597	ICP-MS/MS	1.14	0.617	0.273	1.88	3.90
598	ICP-MS	1.08	0.52	0.186	1.87	4.22
605	ICP-MS	1.06	<0.9	<0.9	1.89	4.34






Summary Statistics						
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05	
Arithmetic Mean ( $\bar{x}$ )	1.22	0.62	0.23	2.1	4.6	
Arithmetic SD (s)	0.17	0.06	0.04	0.3	0.6	
Arithmetic RSD (%)	14	9.7	17	13	13	
Number of Sample Measurements (N)	7	6	3	7	7	

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Urine Sn



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = arithmetic mean of all laboratories.  
 Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics






Urine Sr (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
103	ICP-MS/MS	241	63.0	160	118	60.7
107	ICP-MS	251.6	65.2	163.8	124.4	63.6
200	ICP-MS	258	67	168	139	67
264	ICP-MS	207.53	*43.58	*110.06	82.16	*42.11
399	DRC/CC-ICP-MS	255	65.9	165	129	64.4
597	ICP-MS/MS	227	59.5	155	113	59.4
605	ICP-MS	251	67.0	167	129	65.3
Summary Statistics						
		UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		242	65	163	119	63
<b>Arithmetic SD (s)</b>		18	3	5	18	3
<b>Arithmetic RSD (%)</b>		7.4	4.5	3.1	15	4.6
<b>Number of Sample Measurements (N)</b>		7	6	6	7	6

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Summary Figures

### Urine Sr



**Legend:**

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.  
Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Urine V (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
116	ICP-MS/MS	0.399	<0.100	0.290	0.636	0.156
147	DRC/CC-ICP-MS	0.402	<0.0612	0.267	0.602	0.170
293	DRC/CC-ICP-MS	0.52	0.23	0.34	0.77	0.3
597	ICP-MS/MS	0.370	<0.0844	0.160	0.407	<0.0844
598	DRC/CC-ICP-MS	0.434	<0.2	0.247	0.726	0.207
605	ICP-MS	0.381	<0.3	<0.3	0.799	<0.3

Summary Statistics						
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05	
Arithmetic Mean ( $\bar{x}$ )	0.42	NA	0.26	0.66	0.21	
Arithmetic SD (s)	0.05	NA	0.07	0.14	0.06	
Arithmetic RSD (%)	12	NA	27	21	29	
Number of Sample Measurements (N)	6	NA	5	6	4	






\*Denotes a statistical Outlier.

Statistical data was not calculated for UE22-02 based on a lack of consensus among participating labs.



# Results for Event #1, 2022: Summary Figures

## Urine V



### Legend:

- C/HHEAR Labs
- Other Labs
- Horizontal purple line = arithmetic mean of all laboratories.
- Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

Urine W (µg/L)						
Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
107	ICP-MS	1.651	0.554	1.288	0.766	0.396
110	ICP-MS	1.76	0.577	1.35	0.797	0.423
147	ICP-MS	1.64	0.541	1.26	0.765	0.458
200	ICP-MS	1.60	0.55	1.18	0.86	0.46
264	ICP-MS	1.46	0.46	1.14	0.64	0.32
324	ICP-MS	1.643	<1	1.303	<1	<1
399	ICP-MS/MS	1.71	0.538	1.32	0.740	0.399
597	ICP-MS/MS	1.59	0.528	1.27	0.797	0.380
598	ICP-MS	1.79	0.597	1.4	0.779	0.468
605	ICP-MS	1.72	0.583	1.36	0.795	0.420
606	ICP-MS/MS	1.68	0.568	1.30	0.774	0.430






Summary Statistics					
	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Robust Mean (x*)</b>	1.67	0.55	1.30	0.78	0.42
<b>Robust SD (s*)</b>	0.08	0.03	0.07	0.03	0.04
<b>Robust RSD (%)</b>	4.8	5.6	5.4	3.9	9.5
<b>Number of Sample Measurements (N)</b>	11	10	11	10	10
<b>Standard Uncertainty (u)</b>	0.03	0.01	0.02	0.01	0.02

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Urine W



### Legend:

- C/HHEAR Labs
- Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Urine Zn (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
110	ICP-MS	250	209	128	354	233
147	ICP-MS	255	214	108	382	*284
264	ICP-MS	226.08	199.71	97.36	363.96	232.51
293	DRC/CC-ICP-MS	205.88	179.08	72.55	339.22	211.76
324	ICP-MS	219.049	192.318	89.811	370.480	229.353
391	DRC/CC-ICP-MS	199	187	100	356	221
597	ICP-MS/MS	210	188	86.9	353	218
598	ICP-MS	209	187	89.5	344	220

### Summary Statistics






	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
Arithmetic Mean ( $\bar{x}$ )	222	194	97	358	224
Arithmetic SD (s)	21	12	16	14	8
Arithmetic RSD (%)	9.5	6.2	16	3.9	3.6
Number of Sample Measurements (N)	8	8	8	8	7

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Urine Zn



### Legend:

○ C/HHEAR Labs

● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Urine Te (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
110	ICP-MS	2.12	1.25	0.626	1.44	0.300
147	ICP-MS	2.63	1.36	0.775	1.59	0.385

### Summary Statistics

	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
Arithmetic Mean ( $\bar{x}$ )	2.4	1.30	0.70	1.51	0.34
Arithmetic SD (s)	0.4	0.08	0.11	0.11	0.06
Arithmetic RSD (%)	17	6.2	16	7.3	18
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Urine Ti (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
442	DRC/CC-ICP-MS	2.11	6.66	7.43	1.74	2.72
597	ICP-MS/MS	<2.05	7.95	8.12	3.04	2.83

### Summary Statistics

	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	NA	7.3	7.8	NA	2.78
<b>Arithmetic SD (s)</b>	NA	0.9	0.5	NA	0.08
<b>Arithmetic RSD (%)</b>	NA	12	6.4	NA	2.9
<b>Number of Sample Measurements (N)</b>	NA	2	2	NA	2

\*Denotes a statistical Outlier.

Statistical data was not calculated for UE22-01 and UE22-04 based on a lack of consensus among participating labs.



## Results for Event #1, 2022: Additional Elements in Urine

### Urine Ag (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
147	ICP-MS	<0.183	<0.183	<0.183	<0.183	<0.183

### Urine B (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
200	ICP-MS	338	374	188	806	609

### Urine Bi (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
147	ICP-MS	<0.0815	<0.0815	<0.0815	<0.0815	<0.0815
264	ICP-MS	0.01	0.01	0.01	0.01	0.01
597	ICP-MS/MS	<0.0510	<0.0510	<0.0510	<0.0510	<0.0510

### Urine Fe (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
324	ICP-MS	12.176	6.407	4.558	18.058	13.029

### Urine I (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
147	ICP-MS	99.4	143	60.0	273	163

### Urine Li (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
147	ICP-MS	6.37	12.7	3.37	26.4	10.5

### Urine Mg (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
597	ICP-MS/MS	18400	22200	9710	43300	30000

### Urine Th (µg/L)

Lab Code	Method	UE22-01	UE22-02	UE22-03	UE22-04	UE22-05
147	ICP-MS	<0.0673	<0.0673	<0.0673	<0.0673	<0.0673
597	ICP-MS/MS	<0.0298	0.0380	<0.0298	0.0482	<0.0298



**Department  
of Health**

**Wadsworth  
Center**

**Event #1, 2022**

**Trace Elements in  
Serum**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #1, 2022: Trace Elements in Serum

### PT Materials

Test materials were prepared from human serum obtained from Zen-Bio, Inc. The company certifies that these materials were tested by FDA approved methods and found to be negative for HIV 1Z2 and HIV-1 RNA, and non-reactive to HBsAg, HCV3 and STS. Units of serum were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with aluminum (Al), cobalt (Co), chromium (Cr), copper (Cu), selenium (Se), zinc (Zn), arsenic (As), beryllium (Be), cadmium (Cd), mercury (Hg), manganese (Mn), molybdenum (Mo), nickel (Ni), lead (Pb), platinum (Pt), antimony (Sb), tin (Sn), strontium (Sr), titanium (Ti), thallium (Tl), uranium (U), vanadium (V) and tungsten (W). PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Six elements in serum are formally graded: Al, Co, Cr, Cu, Se, and Zn. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 26 were reported by at least one participant: As, B, Ba, Be, Bi, Cd, Cs, Fe, Hg, I, Li, Mg, Mn, Mo, Ni, Pb, Pt, Sb, Sn, Sr, Th, Ti, Tl, U, V, and W. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #1, 2022: Summary Statistics

	Serum AI (µg/L)				
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	93	30.5	68	37	54.6
<b>Upper Limit</b>	112	36.6	82	44	65.5
<b>Lower Limit</b>	74	24.4	54	30	43.7
<b>Arithmetic SD (s)</b>	7	1.9	4	2	1.6
<b>Arithmetic RSD (%)</b>	7.5	6.2	5.9	6.5	2.9
<b>Number of Sample Measurements (N)</b>	6	6	6	6	5

The acceptable range is based on quality specifications:  $\pm 5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 5 \mu\text{g/L}$  at concentrations less than or equal to  $25 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2022: Performance of Participating Laboratories






Lab Code	Method	Serum AI (µg/L)				
		SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
	<b>Target</b>	<b>93</b>	<b>30.5</b>	<b>68</b>	<b>37</b>	<b>54.6</b>
147	ETAAS-Z	82.7	33.2	63.7	38.7	52.8
264	ICP-MS	99.27	30.92	74.01	38.46	*63.79
293	DRC/CC-ICP-MS	99.46	32.26	69.89	37.63	56.45
391	ETAAS-Z	86.4	29.0	64.8	37.8	55.4
485	HR-ICP-MS	97.00	28.2	68.3	35.7	55.4
597	ICP-MS/MS	90.4	29.6	65.3	32.4	53.1

Based on the grading criteria for AI in Serum, 100% of results were satisfactory, with 0 of the 6 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Serum AI



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 5 \mu\text{g/L}$  at concentrations less than or equal to  $25 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

	Serum Co (µg/L)				
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	61.1	9.9	5.3	4.02	1.60
<b>Upper Limit</b>	70.3	11.4	6.8	5.52	3.10
<b>Lower Limit</b>	51.9	8.4	3.8	2.52	0.10
<b>Arithmetic SD (s)</b>	1.1	0.6	0.3	0.20	0.07
<b>Arithmetic RSD (%)</b>	1.8	6.1	6.6	5.0	4.4
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #1, 2022: Performance of Participating Laboratories






		Serum Co (µg/L)				
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
	<b>Target</b>	<b>61.1</b>	<b>9.9</b>	<b>5.3</b>	<b>4.02</b>	<b>1.60</b>
103	ICP-MS/MS	61.4	9.71	5.06	3.92	1.56
110	ICP-MS	63.1	10.3	5.82	4.32	1.65
147	DRC/CC-ICP-MS	61.2	8.72	4.77	3.67	1.47
264	ICP-MS	61.15	9.95	5.51	4.10	1.65
293	DRC/CC-ICP-MS	61.0	10.66	5.54	4.09	1.67
485	HR-ICP-MS	60.4	10.2	5.46	4.1	1.61
597	ICP-MS/MS	59.3	9.59	5.20	3.94	1.61

Based on the grading criteria for Co in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #1, 2022: Summary Figures

### Serum Co



**Legend:**

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
±1.5 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 10 µg/L.



## Results for Event #1, 2022: Summary Statistics

	Serum Cr (µg/L)				
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	11.0	8.3	3.2	6.3	0.62
<b>Upper Limit</b>	13.2	10.3	5.2	8.3	2.62
<b>Lower Limit</b>	8.8	6.3	1.2	4.3	0.00
<b>Arithmetic SD (s)</b>	0.8	0.5	0.4	0.5	0.10
<b>Arithmetic RSD (%)</b>	7.3	6.0	12	7.9	16
<b>Number of Sample Measurements (N)</b>	7	7	7	7	6

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



## Results for Event #1, 2022: Performance of Participating Laboratories






		Serum Cr (µg/L)				
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
	<b>Target</b>	<b>11.0</b>	<b>8.3</b>	<b>3.2</b>	<b>6.3</b>	<b>0.62</b>
103	ICP-MS/MS	11.4	8.55	3.20	6.45	0.504
110	DRC/CC-ICP-MS	12.5	8.86	3.73	7.00	0.67
147	DRC/CC-ICP-MS	10.70	8.01	3.22	6.19	0.546
264	ICP-MS	10.55	8.24	3.29	5.98	0.69
293	DRC/CC-ICP-MS	11.03	8.49	3.54	6.50	0.77
485	HR-ICP-MS	10.6	8.49	3.27	6.38	0.55
597	ICP-MS/MS	10.1	7.41	2.46	5.54	<1.01

Based on the grading criteria for Cr in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #1, 2022: Summary Figures

### Serum Cr



**Legend:**

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
 $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

	Serum Cu (µg/L)				
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	860	2280	1860	1080	940
<b>Upper Limit</b>	990	2620	2140	1240	1080
<b>Lower Limit</b>	730	1940	1580	920	800
<b>Arithmetic SD (s)</b>	50	120	90	40	60
<b>Arithmetic RSD (%)</b>	5.8	5.3	4.8	3.7	6.4
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2022: Performance of Participating Laboratories






		Serum Cu (µg/L)				
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
	<b>Target</b>	<b>860</b>	<b>2280</b>	<b>1860</b>	<b>1080</b>	<b>940</b>
107	DRC/CC-ICP-MS	858	2327	1932	1116	989
110	ICP-MS	893	2434	1988	1144	978
147	DRC/CC-ICP-MS	845	2300	1849	1086	959
264	ICP-MS	812.3	2286.9	1874.3	1088.8	976.8
293	DRC/CC-ICP-MS	826	2238	1805	1055	909
436	FAAS	953	2033	1715.7	1016.73	826.09
597	ICP-MS/MS	849	2340	1870	1080	948

Based on the grading criteria for Cu in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Serum Cu



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ .



### Results for Event #1, 2022: Summary Statistics

	Serum Se (µg/L)				
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	208	123	174	136	287
<b>Upper Limit</b>	250	148	209	163	344
<b>Lower Limit</b>	166	98	139	109	230
<b>Arithmetic SD (s)</b>	9	5	7	5	11
<b>Arithmetic RSD (%)</b>	4.3	4.1	4.0	3.7	3.8
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 2$  µg/L or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2$  µg/L at concentrations less than or equal to 10 µg/L. These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2022: Performance of Participating Laboratories






		Serum Se (µg/L)				
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
	<b>Target</b>	<b>208</b>	<b>123</b>	<b>174</b>	<b>136</b>	<b>287</b>
103	ICP-MS/MS	217	127	176	139	303
107	DRC/CC-ICP-MS	207.8	124.4	178.1	137.5	280.8
110	DRC/CC-ICP-MS	211	126	180	139	276
147	DRC/CC-ICP-MS	195	113	162	126	275
264	ICP-MS	196.91	120.53	169.61	132.64	283.62
293	DRC/CC-ICP-MS	220	128	180	140	297
597	ICP-MS/MS	210	125	175	135	292

Based on the grading criteria for Se in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #1, 2022: Summary Figures

## Serum Se



### Legend:

○ C/HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #1, 2022: Summary Statistics

	Serum Zn (µg/L)				
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	680	668	709	798	737
<b>Upper Limit</b>	782	768	815	918	848
<b>Lower Limit</b>	578	568	603	678	626
<b>Arithmetic SD (s)</b>	31	34	36	35	26
<b>Arithmetic RSD (%)</b>	4.6	5.1	5.1	4.4	3.5
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

The acceptable range is based on quality specifications:  $\pm 15 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 15 \mu\text{g/L}$  at concentrations less than or equal to  $100 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #1, 2022: Performance of Participating Laboratories






		Serum Zn (µg/L)				
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
	<b>Target</b>	<b>680</b>	<b>668</b>	<b>709</b>	<b>798</b>	<b>737</b>
107	DRC/CC-ICP-MS	681	664	722	830	752
110	ICP-MS	729	724	772	845	773
147	DRC/CC-ICP-MS	673	660	693	791	732
264	ICP-MS	637.5	639.2	681.4	762.8	718.2
293	DRC/CC-ICP-MS	667	634	673	758	699
597	ICP-MS/MS	694	689	714	801	749

Based on the grading criteria for Zn in Serum, 100% of results were satisfactory, with 0 of the 6 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #1, 2022: Summary Figures

### Serum Zn



**Legend:**

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
±15 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±15 µg/L at concentrations less than or equal to 100 µg/L.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics






Serum Mn (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	3.02	15.4	5.94	8.27	4.11
110	ICP-MS	3.29	16.1	6.56	8.49	4.10
147	DRC/CC-ICP-MS	2.84	15.2	5.99	8.13	4.08
264	ICP-MS	3.50	16.32	6.56	8.93	*4.68
293	DRC/CC-ICP-MS	3.06	15.73	6.22	8.44	4.2
597	ICP-MS/MS	2.77	14.8	5.98	7.99	3.92
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	3.1	15.6	6.2	8.4	4.09	
Arithmetic SD (s)	0.3	0.6	0.3	0.3	0.11	
Arithmetic RSD (%)	8.8	3.8	4.7	3.9	2.7	
Number of Sample Measurements (N)	6	6	6	6	5	

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Serum Mn



### Legend:

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics






Serum Mo (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	7.19	3.94	0.740	2.78	5.04
110	ICP-MS	7.39	4.36	0.97	3.06	5.39
147	DRC/CC-ICP-MS	6.62	3.71	0.746	2.63	4.90
293	DRC/CC-ICP-MS	7.24	4.13	0.83	2.85	5.34
485	HR-ICP-MS	7.08	3.99	0.79	2.81	5.19
597	ICP-MS/MS	7.05	4.17	0.816	2.78	5.19
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	7.1	4.05	0.81	2.82	5.17	
Arithmetic SD (s)	0.3	0.22	0.08	0.14	0.18	
Arithmetic RSD (%)	3.7	5.4	9.9	5.0	3.5	
Number of Sample Measurements (N)	6	6	6	6	6	

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Serum Mo



### Legend:

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Ni (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
110	DRC/CC-ICP-MS	24.0	8.95	2.96	2.01	4.98
147	DRC/CC-ICP-MS	23.1	6.58	2.72	1.78	4.64
293	DRC/CC-ICP-MS	24.29	7.55	3.27	2.08	4.82
442	DRC/CC-ICP-MS	23.9	7.48	3.30	2.07	5.13
485	HR-ICP-MS	23.3	7.62	3.08	1.97	4.95
597	ICP-MS/MS	22.4	6.35	2.30	*0.887	*3.87






Summary Statistics					
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
Arithmetic Mean ( $\bar{x}$ )	23.5	7.4	2.9	1.98	4.90
Arithmetic SD (s)	0.7	0.9	0.4	0.12	0.18
Arithmetic RSD (%)	3.0	12	14	6.1	3.7
Number of Sample Measurements (N)	6	6	6	5	5

\*Denotes a statistical Outlier.



# Results for Event #1, 2022: Summary Figures

## Serum Ni



### Legend:

- C/HHEAR Labs
- Other Labs
- Horizontal purple line = arithmetic mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum V (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
110	DRC/CC-ICP-MS	0.3	0.3	0.6	0.9	1.1
147	DRC/CC-ICP-MS	0.154	0.143	0.247	0.628	0.862
293	DRC/CC-ICP-MS	0.6	0.6	0.7	1.2	1.3
485	HR-ICP-MS	0.168	0.152	0.295	0.69	0.944
597	ICP-MS/MS	<0.157	<0.157	<0.157	0.264	0.53
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	NA	NA	NA	NA	NA	NA
Arithmetic SD (s)	NA	NA	NA	NA	NA	NA
Arithmetic RSD (%)	NA	NA	NA	NA	NA	NA
Number of Sample Measurements (N)	NA	NA	NA	NA	NA	NA






\*Denotes a statistical Outlier.

Statistical data was not calculated for SE22-01, SE22-02, SE22-03, SE22-04 and SE22-05 based on a lack of consensus among participating labs.



# Results for Event #1, 2022: Summary Figures

## Serum V



### Legend:

○ C/HHEAR Labs    ● Other Labs  
Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum As (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	7.15	1.20	21.2	2.65	13.1
110	DRC/CC-ICP-MS	7.47	1.24	22.1	2.65	12.8
147	DRC/CC-ICP-MS	6.57	1.22	19.9	2.54	12.7
597	ICP-MS/MS	7.11	1.57	21.1	2.99	12.8
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	7.1	1.31	21.1	2.7	12.84	
Arithmetic SD (s)	0.4	0.17	0.9	0.2	0.15	
Arithmetic RSD (%)	5.6	13	4.3	7.0	1.2	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Ba (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
110	ICP-MS	1.68	0.73	1.03	0.74	1.36
147	ICP-MS	1.35	0.658	0.981	0.750	0.898
597	ICP-MS/MS	1.46	0.749	1.16	0.724	0.915
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	1.5	0.71	1.06	0.738	1.1	
Arithmetic SD (s)	0.2	0.05	0.09	0.013	0.3	
Arithmetic RSD (%)	11	7.0	8.5	1.8	27	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Be (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
110	ICP-MS	0.595	6.94	3.61	2.81	4.38
147	ICP-MS	0.495	5.77	2.89	2.47	3.94
293	ICP-MS	0.580	6.55	3.26	2.60	3.930
597	ICP-MS/MS	0.550	6.65	3.44	2.61	4.09
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	0.55	6.5	3.3	2.62	4.1	
Arithmetic SD (s)	0.04	0.5	0.3	0.14	0.2	
Arithmetic RSD (%)	7.3	7.7	9.1	5.3	5.1	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Cd (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	1.02	2.69	6.25	3.39	0.304
110	ICP-MS	0.989	2.89	6.91	3.50	0.353
147	ICP-MS	0.940	2.53	6.09	3.28	0.294
597	ICP-MS/MS	0.970	2.77	6.63	3.39	0.298
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	0.98	2.72	6.5	3.39	0.31	
Arithmetic SD (s)	0.03	0.15	0.4	0.09	0.03	
Arithmetic RSD (%)	3.1	5.5	6.2	2.7	9.0	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Serum Cs (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
110	ICP-MS	0.740	0.886	0.649	0.738	0.856
597	ICP-MS/MS	0.649	0.812	0.653	0.640	0.800

### Summary Statistics

	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
Arithmetic Mean ( $\bar{x}$ )	0.69	0.85	0.651	0.69	0.83
Arithmetic SD (s)	0.06	0.05	0.003	0.07	0.04
Arithmetic RSD (%)	8.7	5.9	0.46	10	4.8
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Hg ( $\mu\text{g/L}$ )						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	6.61	3.74	0.793	2.15	5.04
110	ICP-MS	6.85	3.71	0.93	2.28	4.93
597	ICP-MS/MS	6.68	3.83	1.02	2.31	5.09

Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	6.71	3.76	0.91	2.25	5.02	
Arithmetic SD (s)	0.12	0.07	0.11	0.08	0.08	
Arithmetic RSD (%)	1.8	1.9	12	3.6	1.6	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum I ( $\mu\text{g/L}$ )						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
147	ICP-MS	34.9	48.7	34.6	72.5	51.0
442	ICP-MS	40.7	55.9	40.6	83.3	57.0

Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	38	52	38	78	54	
Arithmetic SD (s)	4	5	4	8	4	
Arithmetic RSD (%)	11	9.6	11	10	7.4	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Mg ( $\mu\text{g/L}$ )						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
264	ICP-MS	21636	24311	22003	19564	25154
597	ICP-MS/MS	20100	22900	20500	18200	22800

Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	20900	23600	21300	18900	24000	
Arithmetic SD (s)	1100	1000	1100	1000	1700	
Arithmetic RSD (%)	5.3	4.2	5.2	5.3	7.1	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Pb (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	4.08	1.45	7.92	4.45	3.20
110	ICP-MS	4.20	1.49	8.49	4.71	3.34
597	ICP-MS/MS	4.08	1.58	7.98	4.42	3.26
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	4.12	1.50	8.1	4.53	3.27	
Arithmetic SD (s)	0.07	0.07	0.3	0.16	0.07	
Arithmetic RSD (%)	1.7	4.7	3.7	3.5	2.1	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Pt (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
110	ICP-MS	0.41	1.62	1.85	1.23	0.58
264	ICP-MS	*0.10	*1.09	*1.24	*0.75	*0.24
293	DRC/CC-ICP-MS	0.35	1.46	1.64	1.08	0.51
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	0.38	1.5	1.7	1.2	0.54	
Arithmetic SD (s)	0.04	0.1	0.1	0.1	0.05	
Arithmetic RSD (%)	11	7.4	8.6	9.4	8.7	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Sb (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	6.01	4.11	2.06	1.06	0.705
110	ICP-MS	6.81	4.38	2.48	1.16	0.739
147	ICP-MS	6.21	3.93	2.26	1.06	0.712
597	ICP-MS/MS	6.47	4.21	2.42	1.20	0.82
Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	6.4	4.16	2.30	1.12	0.75	
Arithmetic SD (s)	0.3	0.19	0.19	0.07	0.06	
Arithmetic RSD (%)	4.7	4.6	8.3	6.3	8.0	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Serum Sn (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
110	ICP-MS	2.74	5.20	11.3	9.70	0.99
597	ICP-MS/MS	2.38	4.71	10.0	8.90	1.08

### Summary Statistics

	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
Arithmetic Mean ( $\bar{x}$ )	2.6	5.0	10.6	9.3	1.04
Arithmetic SD (s)	0.3	0.3	0.9	0.6	0.06
Arithmetic RSD (%)	12	6.0	8.5	6.5	5.8
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Sr ( $\mu\text{g/L}$ )						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	101	63.7	57.6	98.6	60.9
200	ICP-MS	115.6	69.2	68.3	119.1	73.6
597	ICP-MS/MS	99.6	64.5	58.4	97.9	61.3

Summary Statistics						
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05	
Arithmetic Mean ( $\bar{x}$ )	105	66	61	105	65	
Arithmetic SD (s)	9	3	6	12	7	
Arithmetic RSD (%)	8.6	4.5	9.8	11	11	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum Ti (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
200	DRC/CC-ICP-MS	4.80	4.60	3.60	4.50	*4.30
442	DRC/CC-ICP-MS	6.72	3.04	3.78	5.23	10.6
485	HR-ICP-MS	7.03	2.6	3.62	5.09	11.1
597	ICP-MS/MS	8.26	4.57	6.13	6.90	12.0

Summary Statistics					
	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
Arithmetic Mean ( $\bar{x}$ )	6.7	3.7	4.3	5.4	11.2
Arithmetic SD (s)	1.4	1.0	1.2	1.0	0.7
Arithmetic RSD (%)	21	27	28	19	6.5
Number of Sample Measurements (N)	4	4	4	4	3

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

#### Serum TI (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	4.47	0.899	2.71	1.18	0.214
110	ICP-MS	4.53	0.901	2.85	1.24	0.218
147	ICP-MS	4.29	0.838	2.60	1.14	0.203
597	ICP-MS/MS	4.53	0.937	2.79	1.22	0.203

#### Summary Statistics

	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
Arithmetic Mean ( $\bar{x}$ )	4.45	0.89	2.74	1.20	0.210
Arithmetic SD (s)	0.11	0.04	0.11	0.05	0.008
Arithmetic RSD (%)	2.5	4.5	4.0	4.2	3.8
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



### Results for Event #1, 2022: Laboratory Data and Summary Statistics

Serum U (µg/L)						
Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
103	ICP-MS/MS	0.0466	0.224	0.0961	0.135	0.209
110	ICP-MS	0.049	0.230	0.097	0.142	0.220
597	ICP-MS/MS	0.0456	0.212	0.0933	0.131	0.192
Summary Statistics						
		SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
Arithmetic Mean ( $\bar{x}$ )		0.0469	0.222	0.095	0.136	0.207
Arithmetic SD (s)		0.0015	0.009	0.002	0.005	0.014
Arithmetic RSD (%)		3.2	4.1	1.8	3.7	6.8
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Laboratory Data and Summary Statistics

### Serum W (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
110	ICP-MS	4.68	0.94	3.52	2.26	0.42
200	ICP-MS	4.34	0.92	3.49	2.11	0.40
597	ICP-MS/MS	4.45	0.87	3.28	2.08	0.395

### Summary Statistics

	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
Arithmetic Mean ( $\bar{x}$ )	4.5	0.91	3.43	2.15	0.404
Arithmetic SD (s)	0.2	0.04	0.13	0.10	0.011
Arithmetic RSD (%)	3.8	4.4	3.8	4.7	2.7
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #1, 2022: Additional Elements in Serum

### Serum B (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
200	ICP-MS	43.00	45.00	52.00	43.00	91.00

### Serum Bi (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
147	ICP-MS	<0.0397	<0.0397	<0.0397	<0.0397	<0.0397
597	ICP-MS/MS	<0.0357	<0.0357	<0.0357	<0.0357	<0.0357

### Serum Fe (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
264	ICP-MS	1230.7	1175.3	1224.7	655.6	1223.7

### Serum Li (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
147	ICP-MS	1.12	0.916	1.03	0.346	0.972

### Serum Th (µg/L)

Lab Code	Method	SE22-01	SE22-02	SE22-03	SE22-04	SE22-05
597	ICP-MS/MS	<0.0142	<0.0142	<0.0142	<0.0142	<0.0142



## References

1. ISO/FDIS-13528 (2005) Statistical methods for use in proficiency testing by interlaboratory comparisons. International Organization for Standardization, Geneva.
2. Taylor A, Angerer J, Arnaud J, Claeys F, Jones RL, Mazarrasa O, Mairiaux E, Menditto A, Parsons PJ, Patriarca M, Pineau A, Valkonen S, Weber J-P, Weykamp C. Occupational and environmental laboratory medicine: A network of EQAS organisers. Accreditation and Quality Assurance. 2006;11(8-9):435-9. PubMed PMID: 086NJ-0011.