



**Department  
of Health**

**Wadsworth  
Center**

# **New York State Biomonitoring Program for Trace Elements**

## **Event #3, 2023**

### **Trace Elements in Whole Blood, Urine, and Serum**

## **December, 2023**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



**Event #3, 2023:  
Trace Elements in Whole Blood, Urine, and Serum**

12/7/2023

Dear Laboratory Director,

This report summarizes performance for the third biomonitoring proficiency test (PT) event of 2023 for Trace Elements in Whole Blood, Urine, and Serum. One of the key goals of this PT program is to achieve harmonization of biomonitoring data for trace elements. In this report, we summarize the responses to our recent survey request. Please refer to the attachment at the end of the report for more details.

**Target Value Assignment and Performance Evaluation:**

For these PT materials, target values have been assigned for a limited number of trace elements that are gradable under criteria set by the NYS DOH Biomonitoring PT program. See assay-specific narratives for details. Data for additional trace elements are reported and are included here in order to characterize the PT materials more completely. Participant data and descriptive statistics are provided for educational purposes. No target value or acceptable range is implied.

Where the data permit, robust statistics were used to assign target values based on Algorithm A as defined by ISO 13528:2005E *Statistical methods for use in proficiency testing by inter-laboratory comparisons* [1]. Acceptable ranges for the graded elements are based on consensus criteria and/or those set by the NYS DOH's PT program. For example, some are fixed based on US regulatory guidelines (Pb, Cd) while for other elements the criteria are based on a consensus of the Network of PT scheme organizers for trace elements in occupational and environmental laboratory medicine [2]. Quality specifications are element and matrix specific; full details are provided under each element specific narrative.

A confidential, three-digit code number assigned by PT program staff identifies all laboratory participants.

Samples for the next PT event (Event #1, 2024) will be shipped February 7, 2024. Comments about this report may be directed to [trel@health.ny.gov](mailto:trel@health.ny.gov). If you have not yet enrolled for next year, please contact PT program staff at [trel@health.ny.gov](mailto:trel@health.ny.gov).

Sincerely,

Patrick J. Parsons, PhD  
Chief, Inorganic and Nuclear Chemistry,  
Division of Environmental Sciences  
Wadsworth Center

Kayla Mehigan  
Coordinator, Biomonitoring PT Program,  
Division of Environmental Sciences  
Wadsworth Center



**Department  
of Health**

**Wadsworth  
Center**

**Event #3, 2023**

**Trace Elements in  
Whole Blood**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



**Event #3, 2023:  
Trace Elements in Whole Blood**

**PT Materials**

Human whole blood was purchased from Zen-Bio, Inc. and preserved with K<sub>2</sub>EDTA. The company certifies that this material was "non-reactive" for HBsAg, HBV DNA, HIV-1,2 Ab, HIV-1 RNA, HCV Ab, HCV RNA, and STS. Units of whole blood were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with arsenic (As), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), barium (Ba), beryllium (Be), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), titanium (Ti), thallium (Tl), uranium (U), vanadium (V), tungsten (W), and zinc (Zn). Whole blood samples were homogenized overnight prior to aliquoting 2-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories

**Graded Elements**

Seven elements in whole blood are formally graded: As, Cd, Co, Cr, Hg, Mn, and Pb. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

**Additional Elements**

An additional 25 elements were reported by at least one participant: Ag, Al, Ba, Be, Bi, Cs, Cu, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, Tl, U, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #3, 2023: Summary Statistics

Whole Blood As (µg/L)					
	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	1.89	12.3	36.5	6.7	3.75
<b>Upper Limit</b>	7.89	18.3	43.8	12.7	9.75
<b>Lower Limit</b>	0.00	6.3	29.2	0.7	0.00
<b>Arithmetic SD (s)</b>	0.09	0.5	1.4	0.3	0.23
<b>Arithmetic RSD (%)</b>	4.8	4.1	3.8	3.8	6.1
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #3, 2023: Performance of Participating Laboratories






Whole Blood As (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
	<b>Target</b>	<b>1.89</b>	<b>12.3</b>	<b>36.5</b>	<b>6.7</b>	<b>3.75</b>
103	ICP-MS/MS	1.90	13.1	38.0	6.95	3.94
110	ICP-MS/MS	1.97	12.6	38.0	6.99	4.15
147	ICP-MS	1.99	12.1	37.1	6.53	3.65
247	ICP-MS/MS	1.84	11.5	34.2	6.42	3.48
264	ICP-MS	1.96	11.87	35.98	6.36	3.66
293	DRC/CC-ICP-MS	1.75	12.1	35.65	6.58	3.6
597	ICP-MS/MS	1.85	12.5	36.5	6.82	3.77

Based on the grading criteria for As in Whole Blood, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #3, 2023: Summary Figures

### Whole Blood As



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ .



## Results for Event #3, 2023: Summary Statistics

Whole Blood Cd (µg/L)					
	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Target (Robust Mean (x*))</b>	1.24	2.81	4.95	12.0	0.569
<b>Upper Limit</b>	2.24	3.81	5.95	13.8	1.569
<b>Lower Limit</b>	0.24	1.81	3.95	10.2	0.000
<b>Robust SD (s*)</b>	0.08	0.14	0.26	0.6	0.020
<b>Robust RSD (%)</b>	6.5	4.9	5.3	4.7	3.5
<b>Number of Sample Measurements (N)</b>	11	12	12	12	11
<b>Standard Uncertainty (u)</b>	0.03	0.05	0.09	0.2	0.007

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.7 \mu\text{g/L}$ . These quality specifications are based on those used by US OSHA for occupational exposure.



### Results for Event #3, 2023: Performance of Participating Laboratories






Whole Blood Cd (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
	Target	1.24	2.81	4.95	12.0	0.569
103	ICP-MS/MS	1.26	2.87	5.15	12.2	0.578
107	ICP-MS/MS	1.22	2.73	4.96	11.70	0.55
110	ICP-MS	1.21	2.93	5.26	12.3	0.43
116	ICP-MS/MS	<1.50	2.83	4.60	11.7	<1.50
147	ICP-MS	1.18	2.76	4.95	11.6	0.570
247	ICP-MS/MS	1.14	2.59	4.47	11.4	0.565
264	ICP-MS	1.14	2.50	4.56	10.72	0.42
293	DRC/CC-ICP-MS	1.21	2.78	5.010	12.1	0.57
597	ICP-MS/MS	1.30	2.94	5.09	12.3	0.581
605	ICP-MS	1.33	2.90	5.06	12.5	0.578
606	ICP-MS/MS	1.32	2.73	4.83	12.5	0.600
686	ICP-MS	1.31	2.98	5.23	12.9	0.589

Based on the grading criteria for Cd in Whole Blood, 100% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Whole Blood Cd



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 6.7 µg/L.



## Results for Event #3, 2023: Summary Statistics

Whole Blood Co (µg/L)					
	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	1.99	0.50	11.7	6.3	29.3
<b>Upper Limit</b>	3.49	2.00	14.0	7.8	35.2
<b>Lower Limit</b>	0.49	0.00	9.4	4.8	23.4
<b>Arithmetic SD (s)</b>	0.11	0.05	0.5	0.3	1.1
<b>Arithmetic RSD (%)</b>	5.5	10	4.3	4.3	3.8
<b>Number of Sample Measurements (N)</b>	8	7	8	8	8

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #3, 2023: Performance of Participating Laboratories






Whole Blood Co (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Target		1.99	0.50	11.7	6.3	29.3
103	ICP-MS/MS	2.05	0.460	11.7	6.23	29.3
110	ICP-MS/MS	2.04	0.51	12.1	6.54	30.5
147	ICP-MS	2.10	0.575	11.8	6.42	29.0
247	ICP-MS/MS	1.90	0.476	11.1	6.01	27.7
255	ICP-MS	1.9	<0.5	11	6.2	29
264	ICP-MS	1.79	0.44	11.09	5.83	28.12
293	DRC/CC-ICP-MS	2.08	0.54	12.43	6.50	30.28
597	ICP-MS/MS	2.05	0.511	12.1	6.61	30.6

Based on the grading criteria for Co in Whole Blood, 100% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Whole Blood Co



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ .



## Results for Event #3, 2023: Summary Statistics

Whole Blood Cr ( $\mu\text{g/L}$ )					
	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Target (Arithmetic Mean ( $\bar{x}$ ))	4.14	1.19	10.2	0.49	3.04
Upper Limit	6.14	3.19	12.2	2.49	5.04
Lower Limit	2.14	0.00	8.2	0.00	1.04
Arithmetic SD (s)	0.24	0.19	0.8	0.05	0.11
Arithmetic RSD (%)	5.8	16	7.8	10	3.6
Number of Sample Measurements (N)	7	6	8	6	7

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



## Results for Event #3, 2023: Performance of Participating Laboratories






Whole Blood Cr (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
	<b>Target</b>	<b>4.14</b>	<b>1.19</b>	<b>10.2</b>	<b>0.49</b>	<b>3.04</b>
103	ICP-MS/MS	4.27	<1.50	10.9	<1.50	2.95
110	ICP-MS/MS	4.27	1.04	10.4	0.44	3.22
147	DRC/CC-ICP-MS	*5.36	1.27	11.7	0.551	*3.53
247	ICP-MS/MS	3.93	1.05	9.36	0.495	2.89
255	ICP-MS	4.2	<1.0	10	<1.0	3.1
264	ICP-MS	3.69	1.53	9.06	0.43	3.06
293	DRC/CC-ICP-MS	4.33	1.13	10.34	0.53	3.04
597	ICP-MS/MS	4.28	1.10	10.0	0.510	3.01

Based on the grading criteria for Cr in Whole Blood, 100% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Whole Blood Cr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±2 µg/L at concentrations less than or equal to 10 µg/L.



### Results for Event #3, 2023: Summary Statistics

Whole Blood Hg (µg/L)					
	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Target (Robust Mean (x*))</b>	5.1	0.77	7.7	6.5	17.3
<b>Upper Limit</b>	8.1	3.77	10.7	9.5	22.5
<b>Lower Limit</b>	2.1	0.00	4.7	3.5	12.1
<b>Robust SD (s*)</b>	0.4	0.09	0.8	2.0	1.8
<b>Robust RSD (%)</b>	7.8	12	10	31	10
<b>Number of Sample Measurements (N)</b>	13	11	13	13	13
<b>Standard Uncertainty (u)</b>	0.1	0.03	0.3	0.7	0.6

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #3, 2023: Performance of Participating Laboratories






		Whole Blood Hg (µg/L)				
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Target		5.1	0.77	7.7	6.5	17.3
103	ICP-MS/MS	5.26	0.828	7.78	6.52	17.9
107	ICP-MS/MS	5.35	0.71	8.12	7.37	17.68
110	ICP-MS	5.11	0.81	7.29	4.89	16.4
116	ICP-MS/MS	5.13	<1.50	7.99	7.40	17.5
147	ICP-MS	5.22	0.658	7.66	7.54	18.1
247	ICP-MS/MS	4.93	0.804	7.54	6.54	17.1
264	ICP-MS	1.82 ↓	<0.50	3.06 ↓	2.36 ↓	8.21 ↓
293	DRC/CC-ICP-MS	4.2	0.72	5.8	4.19	13.87
391	CV-AAS	3.8	0.48	6.8	4.69	14.76
597	ICP-MS/MS	5.12	0.749	7.37	5.94	17.2
605	ICP-MS	5.78	0.885	8.93	8.41	19.3
606	ICP-MS/MS	6.50	0.880	10.4	9.26	22.5
686	ICP-MS	5.57	0.774	8.63	7.87	18.7

Based on the grading criteria for Hg in Whole Blood, 94% of results were satisfactory, with 1 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Whole Blood Hg



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #3, 2023: Summary Statistics

Whole Blood Mn ( $\mu\text{g/L}$ )					
	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	7.7	26.0	9.6	20.3	27.6
<b>Upper Limit</b>	10.7	30.4	12.6	23.8	32.3
<b>Lower Limit</b>	4.7	21.6	6.6	16.8	22.9
<b>Arithmetic SD (s)</b>	1.3	2.2	1.2	1.6	1.7
<b>Arithmetic RSD (%)</b>	17	8.5	13	7.6	6.2
<b>Number of Sample Measurements (N)</b>	8	9	8	8	8

The acceptable range is based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ . These quality specifications were recently proposed by a network of Trace Element PT program organizers (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry Laboratory Medicine 2016; 54(12): 1921-1928).



### Results for Event #3, 2023: Performance of Participating Laboratories






Whole Blood Mn (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
	Target	7.7	26.0	9.6	20.3	27.6
103	ICP-MS/MS	7.85	26.1	9.92	19.6	27.3
107	ICP-MS/MS	7.59	26.36	9.38	20.67	28.06
110	ICP-MS	7.7	25.6	9.7	19.7	28.3
147	ICP-MS	10.50	27.5	12.2	22.9	30.3
247	ICP-MS/MS	6.65	24.0	8.41	19.0	26.1
264	ICP-MS	*12.78 ↑	30.59 ↑	*14.72 ↑	*26.52 ↑	*34.66 ↑
293	DRC/CC-ICP-MS	6.2	22.95	8.6	18.1	24.66
597	ICP-MS/MS	7.00	25.0	9.10	20.4	27.1
606	ICP-MS/MS	8.26	26.3	9.47	21.9	28.8

Based on the grading criteria for Mn in Whole Blood, 89% of results were satisfactory, with 1 of the 9 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Whole Blood Mn



### Legend:

○ HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories. Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ .



### Results for Event #3, 2023: Summary Statistics

Whole Blood Pb (µg/dL)					
	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Target (Robust Mean (x*))</b>	0.92	24.5	0.57	1.12	6.0
<b>Upper Limit</b>	2.92	27.0	2.57	3.12	8.0
<b>Lower Limit</b>	0.00	22.1	0.00	0.00	4.0
<b>Robust SD (s*)</b>	0.10	0.6	0.10	0.10	0.3
<b>Robust RSD (%)</b>	11	2.4	18	8.9	5.8
<b>Number of Sample Measurements (N)</b>	10	14	9	12	14
<b>Standard Uncertainty (u)</b>	0.04	0.2	NA	0.04	0.1

The acceptable range is based on quality specifications: ±2 µg/dL or ±10% around the target value, whichever is greater; thus, it is fixed at ±2 µg/dL at concentrations less than or equal to 20 µg/dL. These quality specifications are recommended by the Clinical Laboratory Standards Institute (CLSI, C40-A2) and have been proposed for use in proficiency testing programs approved under CLIA by the Centers for Medicare and Medicaid Services (CMS) in the USA. (<https://clsi.org/standards/products/clinical-chemistry-and-toxicology/documents/c40/>)

An arithmetic mean, SD, RSD and n are provided for sample BE23-13.



### Results for Event #3, 2023: Performance of Participating Laboratories






Whole Blood Pb (µg/dL)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
	Target	0.92	24.5	0.57	1.12	6.0
103	ICP-MS/MS	0.957	24.8	0.670	1.23	6.17
107	ICP-MS/MS	0.941	24.795	0.565	1.147	6.114
110	ICP-MS	0.91	24.3	0.47	0.95	6.20
116	ICP-MS/MS	<3.00	22.9	<3.00	<3.00	5.72
147	ICP-MS	0.937	25.3	0.562	1.20	6.30
247	ICP-MS/MS	0.798	22.8	0.485	1.02	5.66
264	ICP-MS	0.73	21.17 ↓	0.41	0.93	5.37
293	DRC/CC-ICP-MS	1.03	24.61	0.62	1.0	6.00
343	ASV-LeadCare	<3.3	22.4	<3.3	<3.3	4.1
391	ETAAS-Z	0.85	24.59	0.61	1.10	5.83
597	ICP-MS/MS	1.02	24.9	0.728	1.18	6.36
605	ICP-MS	1.01	26.3	<1.00	1.24	6.54
606	ICP-MS/MS	<1.00	24.2	<1.00	1.15	6.06
686	ICP-MS	<1.00	24.7	<1.00	1.16	6.15

Based on the grading criteria for Pb in Whole Blood, 99% of results were satisfactory, with 0 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Whole Blood Pb



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/dL}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ .



### Results for Event #3, 2023: Laboratory Data and Summary Statistics






Whole Blood Mo (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
103	ICP-MS/MS	6.13	<1.50	3.41	<1.50	<1.50
110	ICP-MS/MS	6.40	0.62	3.44	1.30	0.52
147	ICP-MS	6.37	0.605	3.35	1.24	0.545
442	DRC/CC-ICP-MS	6.22	0.488	3.31	1.19	0.443
597	ICP-MS/MS	6.39	0.603	3.47	1.36	0.519
Summary Statistics						
		BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		6.30	0.58	3.40	1.27	0.51
<b>Arithmetic SD (s)</b>		0.12	0.06	0.07	0.07	0.04
<b>Arithmetic RSD (%)</b>		1.9	10	2.1	5.5	7.8
<b>Number of Sample Measurements (N)</b>		5	4	5	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Summary Figures

### Whole Blood Mo



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics






Whole Blood Sb (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
103	ICP-MS/MS	4.18	2.00	3.07	1.03	0.506
110	ICP-MS/MS	5.51	2.69	3.55	1.31	0.65
147	ICP-MS	5.25	2.56	3.35	1.27	0.628
264	ICP-MS	4.20	2.15	2.87	0.98	0.45
293	DRC/CC-ICP-MS	5.12	2.43	2.800	1.30	0.61
442	DRC/CC-ICP-MS	6.11	2.97	3.85	1.48	0.605
597	ICP-MS/MS	4.93	2.62	3.28	1.34	0.634
Summary Statistics						
		BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		5.0	2.5	3.3	1.24	0.58
<b>Arithmetic SD (s)</b>		0.7	0.3	0.4	0.18	0.08
<b>Arithmetic RSD (%)</b>		14	13	11	15	14
<b>Number of Sample Measurements (N)</b>		7	7	7	7	7

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Summary Figures

### Whole Blood Sb



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood Se (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
103	ICP-MS/MS	366	135	149	139	287
107	ICP-MS/MS	375.2	132.7	147.7	134.7	282.0
110	ICP-MS/MS	359	129	148	131	286
147	ICP-MS	352	129	145	137	279
247	ICP-MS/MS	346	120	134	122	268
264	ICP-MS	330.8	131.0	140.3	127.8	262.7
293	DRC/CC-ICP-MS	*200.63	*281.99	*113.74	116.90	*102.69
597	ICP-MS/MS	337	134	146	138	274

### Summary Statistics






	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	352	130	144	131	277
Arithmetic SD (s)	16	5	5	8	9
Arithmetic RSD (%)	4.5	3.8	3.5	6.1	3.2
Number of Sample Measurements (N)	7	7	7	8	7

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Whole Blood Se



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood TI (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
103	ICP-MS/MS	3.62	0.263	0.610	1.46	0.119
110	ICP-MS/MS	3.72	0.30	0.64	1.63	0.12
147	ICP-MS	3.72	0.311	0.638	1.66	0.126
264	ICP-MS	3.37	0.26	0.56	1.42	*0.10
293	DRC/CC-ICP-MS	3.54	0.28	0.60	1.49	0.12
597	ICP-MS/MS	3.61	0.279	0.627	1.58	0.120

### Summary Statistics






	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	3.60	0.282	0.61	1.54	0.121
<b>Arithmetic SD (s)</b>	0.13	0.020	0.03	0.10	0.003
<b>Arithmetic RSD (%)</b>	3.6	7.1	4.9	6.5	2.3
<b>Number of Sample Measurements (N)</b>	6	6	6	6	5

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Summary Figures

### Whole Blood TI



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood Ba (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
110	ICP-MS/MS	2.97	4.23	10.3	3.77	1.76
147	ICP-MS	2.47	4.01	9.52	3.46	1.69
597	ICP-MS/MS	2.70	4.49	9.92	3.91	1.95

### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	2.7	4.2	9.9	3.7	1.80
Arithmetic SD (s)	0.3	0.2	0.4	0.2	0.13
Arithmetic RSD (%)	9.2	5.7	3.9	6.2	7.2
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Whole Blood Be (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
110	ICP-MS/MS	0.94	3.57	1.29	0.48	6.14
147	ICP-MS	<0.991	3.21	1.28	<0.991	6.22
293	ICP-MS	0.9	3.3	1.3	0.5	6.3
597	ICP-MS/MS	1.03	3.53	1.43	0.419	6.55
Summary Statistics						
		BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )		0.96	3.40	1.33	0.47	6.30
Arithmetic SD (s)		0.07	0.17	0.07	0.04	0.18
Arithmetic RSD (%)		7.3	5.1	5.3	8.5	2.9
Number of Sample Measurements (N)		3	4	4	3	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Whole Blood Bi (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
147	ICP-MS	<0.0334	0.0420	<0.0334	<0.0334	<0.0334
597	ICP-MS/MS	<0.0282	0.0402	<0.0282	<0.0282	<0.0282

Summary Statistics						
	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15	
Arithmetic Mean ( $\bar{x}$ )	NA	0.0411	NA	NA	NA	
Arithmetic SD (s)	NA	0.0013	NA	NA	NA	
Arithmetic RSD (%)	NA	3.2	NA	NA	NA	
Number of Sample Measurements (N)	NA	2	NA	NA	NA	

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Whole Blood Cs (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
110	ICP-MS/MS	3.34	1.27	2.01	2.15	4.71
147	ICP-MS	3.35	1.30	1.97	2.11	4.47
597	ICP-MS/MS	3.36	1.26	1.96	2.19	4.60
Summary Statistics						
		BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )		3.35	1.28	1.98	2.15	4.59
Arithmetic SD (s)		0.01	0.02	0.03	0.04	0.12
Arithmetic RSD (%)		0.30	1.6	1.5	1.9	2.6
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood Cu (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
110	ICP-MS/MS	1400	818	685	2560	1070
147	ICP-MS	1455	832	712	2560	1074
247	ICP-MS/MS	1366	813	667	2503	1036
597	ICP-MS/MS	1490	862	716	2720	1110

### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	1430	831	695	2590	1070
Arithmetic SD (s)	60	22	23	90	30
Arithmetic RSD (%)	4.2	2.6	3.3	3.5	2.8
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood Ni (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
103	ICP-MS/MS	3.28	<1.50	7.68	<1.50	9.48
110	ICP-MS/MS	3.67	0.88	7.44	1.09	10.1
147	ICP-MS	3.44	0.746	6.81	0.752	9.45
597	ICP-MS/MS	3.88	0.939	7.41	1.49	10.7

### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	3.6	0.85	7.3	1.1	9.9
Arithmetic SD (s)	0.3	0.10	0.4	0.4	0.6
Arithmetic RSD (%)	7.3	12	5.5	36	6.1
Number of Sample Measurements (N)	4	3	4	3	4

\*Denotes a statistical Outlier.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics

#### Whole Blood Sn (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
110	ICP-MS/MS	3.95	0.39	7.33	0.89	2.25
147	ICP-MS	3.71	0.462	6.73	0.867	1.98
293	DRC/CC-ICP-MS	3.95	0.36	6.91	0.84	2.20
597	ICP-MS/MS	3.92	0.374	7.09	0.864	2.16

#### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	3.88	0.40	7.0	0.87	2.15
Arithmetic SD (s)	0.12	0.05	0.3	0.02	0.12
Arithmetic RSD (%)	3.1	13	3.7	2.3	5.6
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood Sr (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
103	ICP-MS/MS	38.7	47.8	24.9	34.9	25.3
110	ICP-MS/MS	39.5	48.7	25.1	36.6	25.9
147	ICP-MS	37.4	47.2	23.8	36.2	24.6
597	ICP-MS/MS	39.6	48.1	25.2	36.8	26.1

### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	38.8	48.0	24.8	36.1	25.5
Arithmetic SD (s)	1.0	0.6	0.6	0.9	0.7
Arithmetic RSD (%)	2.6	1.3	2.4	2.5	2.7
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood Ti (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
200	DRC/CC-ICP-MS	5.9	3.3	9.2	2.8	3.7
597	ICP-MS/MS	6.43	3.98	9.97	3.06	4.40

### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	6.2	3.6	9.6	2.9	4.1
Arithmetic SD (s)	0.4	0.5	0.5	0.2	0.5
Arithmetic RSD (%)	6.5	14	5.2	6.1	12
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Whole Blood U (µg/L)						
Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
103	ICP-MS/MS	0.126	0.184	<0.0500	0.0622	0.0919
110	ICP-MS/MS	0.139	0.186	0.0457	0.0786	0.0936
147	ICP-MS	0.116	0.164	0.0386	0.0629	0.0829
597	ICP-MS/MS	0.133	0.167	0.0394	0.0698	0.0803
Summary Statistics						
		BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )		0.129	0.175	0.041	0.068	0.087
Arithmetic SD (s)		0.010	0.011	0.004	0.008	0.007
Arithmetic RSD (%)		7.8	6.3	9.8	12	7.5
Number of Sample Measurements (N)		4	4	3	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood V (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
110	ICP-MS/MS	1.04	3.74	0.73	1.93	0.48
147	DRC/CC-ICP-MS	1.11	4.32	0.806	2.24	0.531
597	ICP-MS/MS	0.975	3.49	0.704	1.94	0.427

### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	1.04	3.9	0.75	2.0	0.48
Arithmetic SD (s)	0.07	0.4	0.05	0.2	0.05
Arithmetic RSD (%)	6.7	10	6.7	8.8	10
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Whole Blood W (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
110	ICP-MS/MS	0.99	3.75	0.78	1.91	0.44
200	ICP-MS	1.0	4.4	0.9	2.1	0.5
597	ICP-MS/MS	1.00	3.68	0.744	1.84	0.429

### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	0.997	3.9	0.81	1.95	0.46
Arithmetic SD (s)	0.006	0.4	0.08	0.13	0.04
Arithmetic RSD (%)	0.58	10	9.9	6.7	8.7
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics

#### Whole Blood Zn (µg/L)

Lab Code	Method	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
110	ICP-MS/MS	5060	6070	5370	5370	6320
147	ICP-MS	4627	5379	4837	4824	5595
247	ICP-MS/MS	4255	5013	4536	4542	5288
597	ICP-MS/MS	5440	6400	5710	5800	6670

#### Summary Statistics

	BE23-11	BE23-12	BE23-13	BE23-14	BE23-15
Arithmetic Mean ( $\bar{x}$ )	4800	5700	5100	5100	6000
Arithmetic SD (s)	500	600	500	600	600
Arithmetic RSD (%)	10	11	9.8	12	10
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



Results for Event #3, 2023:
Additional Elements in Whole Blood

Whole Blood Ag (µg/L)

Table with 7 columns: Lab Code, Method, BE23-11, BE23-12, BE23-13, BE23-14, BE23-15. Row 1: 147, ICP-MS, <0.151, <0.151, <0.151, <0.151, <0.151

Whole Blood Al (µg/L)

Table with 7 columns: Lab Code, Method, BE23-11, BE23-12, BE23-13, BE23-14, BE23-15. Row 1: 147, ICP-MS, <4.86, <4.86, <4.86, <4.86, <4.86. Row 2: 597, ICP-MS/MS, 6.01, 9.18, 6.77, 4.91, 4.14

Whole Blood I (µg/L)

Table with 7 columns: Lab Code, Method, BE23-11, BE23-12, BE23-13, BE23-14, BE23-15. Row 1: 147, ICP-MS, 34.4, 31.0, 36.2, 25.6, 31.9

Whole Blood Li (µg/L)

Table with 7 columns: Lab Code, Method, BE23-11, BE23-12, BE23-13, BE23-14, BE23-15. Row 1: 147, ICP-MS, 0.749, 0.708, 0.909, 0.902, 1.39

Whole Blood Mg (µg/L)

Table with 7 columns: Lab Code, Method, BE23-11, BE23-12, BE23-13, BE23-14, BE23-15. Row 1: 597, ICP-MS/MS, 27700, 28200, 30000, 28200, 38100

Whole Blood Pt (µg/L)

Table with 7 columns: Lab Code, Method, BE23-11, BE23-12, BE23-13, BE23-14, BE23-15. Row 1: 293, DRC/CC-ICP-MS, 4.20, 0.58, 1.76, 2.59, 0.79

Whole Blood Te (µg/L)

Table with 7 columns: Lab Code, Method, BE23-11, BE23-12, BE23-13, BE23-14, BE23-15. Row 1: 110, ICP-MS/MS, 0.01, 0.02, 0.01, 0.01, 0.01. Row 2: 147, ICP-MS, <0.0561, <0.0561, <0.0561, <0.0561, <0.0561

Whole Blood Th (µg/L)

Table with 7 columns: Lab Code, Method, BE23-11, BE23-12, BE23-13, BE23-14, BE23-15. Row 1: 147, ICP-MS, <0.0255, <0.0255, <0.0255, <0.0255, <0.0255. Row 2: 597, ICP-MS/MS, 0.0197, 0.00824, 0.00654, 0.00715, 0.00658



**Department  
of Health**

**Wadsworth  
Center**

**Event #3, 2023**

**Trace Elements in  
Urine**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #3, 2023: Trace Elements in Urine

### PT Materials

Urine was collected from volunteer donors into polyethylene containers and stored at 4°C. Following collection, urine was acidified to 1% (v/v) with nitric acid and mixed with a sulfamic acid solution (stock solution contained 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100) to a final concentration of 1% (v/v) to stabilize Hg. Urine was stored frozen at -80°C pending further preparation. The urine was thawed at room temperature and precipitated salts removed by centrifugation. Urine supernatants were combined into five separate pools. Each urine pool was supplemented with arsenic (As), barium (Ba), beryllium (Be), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), thallium (Tl), uranium (U), aluminum (Al), cesium (Cs), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), strontium (Sr), tellurium (Te), titanium (Ti), vanadium (V), tungsten (W), and zinc (Zn). PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Eleven elements in urine are formally graded: As, Ba, Be, Cd, Co, Cr, Hg, Mn, Pb, Tl, and U. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 23 elements were reported by at least one participant: Ag, Al, B, Bi, Cs, Cu, Fe, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



### Results for Event #3, 2023: Summary Statistics

	Urine As (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	2.6	48.2	77	10.9	26.3
<b>Upper Limit</b>	8.6	57.8	92	16.9	32.3
<b>Lower Limit</b>	0.0	38.6	62	4.9	20.3
<b>Robust SD (s*)</b>	0.3	1.9	4	0.7	0.9
<b>Robust RSD (%)</b>	10	3.9	5.6	6.4	3.4
<b>Number of Sample Measurements (N)</b>	12	14	14	14	14
<b>Standard Uncertainty (u)</b>	0.1	0.6	1	0.2	0.3

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #3, 2023: Performance of Participating Laboratories






Urine As (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
	Target	2.6	48.2	77	10.9	26.3
103	ICP-MS/MS	2.44	48.0	75.9	10.5	25.7
107	DRC/CC-ICP-MS	2.29	42.42	73.91	9.03	24.81
110	DRC/CC-ICP-MS	2.87	52.9	84.9	11.8	28.7
116	ICP-MS/MS	<6.00	48.4	79.4	10.7	26.8
147	ICP-MS	2.53	48.5	76.4	10.7	26.3
220	DRC/CC-ICP-MS	2.97	54.4	85.0	11.9	28.6
264	ICP-MS	2.94	51.50	83.55	11.23	28.22
293	DRC/CC-ICP-MS	2.36	45.97	70.49	10.43	25.21
399	DRC/CC-ICP-MS	2.55	48.4	74.4	10.6	26.5
442	ICP-MS/MS	2.32	47.1	73.7	10.2	26.2
597	ICP-MS/MS	2.63	46.4	74.5	10.4	25.4
605	ICP-MS	2.76	47.5	76.7	11.7	26.2
606	ICP-MS/MS	2.52	47.9	80.6	12.1	25.8
686	DRC/CC-ICP-MS	<6.00	49.4	77.7	11.1	26.3

Based on the grading criteria for As in Urine, 100% of results were satisfactory, with 0 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine As



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±6 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±6 µg/L at concentrations less than or equal to 30 µg/L.



## Results for Event #3, 2023: Summary Statistics

	Urine Ba (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	0.73	8.1	5.8	2.99	3.9
<b>Upper Limit</b>	1.73	9.7	7.0	3.99	4.9
<b>Lower Limit</b>	0.00	6.5	4.6	1.99	2.9
<b>Robust SD (s*)</b>	0.08	0.4	0.4	0.21	0.5
<b>Robust RSD (%)</b>	11	4.9	7.1	7.0	13
<b>Number of Sample Measurements (N)</b>	12	12	12	12	12
<b>Standard Uncertainty (u)</b>	0.03	0.2	0.1	0.08	0.2

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #3, 2023: Performance of Participating Laboratories






		Urine Ba (µg/L)				
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
Target		0.73	8.1	5.8	2.99	3.9
107	ICP-MS	0.672	7.978	5.670	2.892	3.551
110	ICP-MS	0.90	9.36	6.78	3.43	4.43
116	ICP-MS/MS	0.736	8.21	5.96	3.15	4.2
147	ICP-MS	0.740	8.45	5.77	2.95	4.01
220	ICP-MS	0.801	8.52	8.91 ↑	3.08	4.08
264	ICP-MS	0.62	7.18	4.93	2.56	3.21
293	DRC/CC-ICP-MS	0.72	8.51	6.44	3.03	4.64
399	ICP-MS/MS	0.831	8.22	5.76	3.14	3.85
597	ICP-MS/MS	0.782	7.80	5.54	2.82	3.74
605	ICP-MS	0.673	7.84	5.36	2.85	3.44
606	ICP-MS/MS	0.669	7.76	5.82	3.17	4.15
686	ICP-MS	0.689	8.07	5.51	2.81	3.58

Based on the grading criteria for Ba in Urine, 98% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine Ba



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 5 µg/L.



### Results for Event #3, 2023: Summary Statistics

	Urine Be (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	2.33	0.98	0.60	1.25	4.41
<b>Upper Limit</b>	3.33	1.98	1.60	2.25	5.41
<b>Lower Limit</b>	1.33	0.00	0.00	0.25	3.41
<b>Robust SD (s*)</b>	0.13	0.05	0.03	0.04	0.23
<b>Robust RSD (%)</b>	5.6	5.1	5.5	2.9	5.2
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.05	0.02	0.01	0.01	0.09

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #3, 2023: Performance of Participating Laboratories






		Urine Be (µg/L)				
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
Target		2.33	0.98	0.60	1.25	4.41
107	ICP-MS	2.176	0.940	0.602	1.201	4.204
110	ICP-MS	2.47	1.00	0.632	1.27	4.64
116	ICP-MS/MS	2.44	0.966	0.649	1.32	4.57
147	ICP-MS	2.51	1.06	0.571	1.33	4.91
220	ICP-MS	2.40	1.01	0.608	1.27	4.50
264	ICP-MS	2.05	0.86	0.56	1.13	3.90
293	ICP-MS	2.33	0.98	0.64	1.22	2.27 ↓
399	ICP-MS/MS	2.28	0.929	0.617	1.23	4.43
597	ICP-MS/MS	2.28	1.05	0.590	1.24	4.31
605	ICP-MS	2.32	0.982	0.580	1.25	4.43
686	ICP-MS	2.27	0.977	0.591	1.25	4.46

Based on the grading criteria for Be in Urine, 98% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine Be



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



## Results for Event #3, 2023: Summary Statistics

	Urine Cd ( $\mu\text{g/L}$ )				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (<math>x^*</math>))</b>	2.18	3.47	0.384	0.96	5.7
<b>Upper Limit</b>	3.18	4.47	1.384	1.96	6.7
<b>Lower Limit</b>	1.18	2.47	0.000	0.00	4.7
<b>Robust SD (<math>s^*</math>)</b>	0.09	0.16	0.026	0.06	0.3
<b>Robust RSD (%)</b>	4.1	4.6	6.8	6.3	5.3
<b>Number of Sample Measurements (N)</b>	15	15	14	14	15
<b>Standard Uncertainty (<math>u</math>)</b>	0.03	0.05	0.009	0.02	0.1

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #3, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine Cd (µg/L)				
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
	<b>Target</b>	<b>2.18</b>	<b>3.47</b>	<b>0.384</b>	<b>0.96</b>	<b>5.7</b>
103	ICP-MS/MS	2.17	3.55	0.391	0.941	5.61
107	DRC/CC-ICP-MS	2.317	3.603	0.411	1.008	6.143
110	ICP-MS	2.43	3.80	0.414	1.04	6.10
116	ICP-MS/MS	2.22	3.38	0.406	0.864	5.77
147	ICP-MS	2.17	3.64	0.378	0.979	5.99
220	ICP-MS	2.21	3.43	0.374	0.924	5.60
264	ICP-MS	2.24	3.57	0.41	1.03	6.05
293	DRC/CC-ICP-MS	2.14	3.47	0.39	1	5.47
324	ICP-MS	2.178	3.302	<1	<1	5.505
399	DRC/CC-ICP-MS	2.10	3.38	0.370	0.918	5.52
442	ICP-MS/MS	2.24	3.66	0.405	0.960	5.83
597	ICP-MS/MS	2.08	3.32	0.350	0.892	5.36
605	ICP-MS	2.13	3.47	0.351	0.963	5.80
606	ICP-MS/MS	2.00	3.19	0.359	0.966	5.18
686	ICP-MS	2.10	3.39	0.368	0.921	5.48

Based on the grading criteria for Cd in Urine, 100% of results were satisfactory, with 0 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine Cd



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ .



### Results for Event #3, 2023: Summary Statistics

	Urine Co (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	0.49	25.3	2.30	8.3	8.5
<b>Upper Limit</b>	1.99	29.1	3.80	9.8	10.0
<b>Lower Limit</b>	0.00	21.5	0.80	6.8	7.0
<b>Robust SD (s*)</b>	0.04	1.0	0.14	0.5	0.4
<b>Robust RSD (%)</b>	8.2	4.0	6.0	6.2	4.7
<b>Number of Sample Measurements (N)</b>	11	12	12	12	12
<b>Standard Uncertainty (u)</b>	0.02	0.4	0.05	0.2	0.2

The acceptable range is based on quality specifications: ±1.5 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 10 µg/L. These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #3, 2023: Performance of Participating Laboratories






		Urine Co (µg/L)				
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
Target		0.49	25.3	2.30	8.3	8.5
103	ICP-MS/MS	0.440	24.6	2.22	7.86	8.15
107	ICP-MS	0.451	22.535	2.135	7.572	7.616
110	ICP-MS	0.519	26.9	2.43	8.77	9.09
147	ICP-MS	0.464	25.1	2.20	8.01	8.37
220	ICP-MS	0.535	26.5	2.43	8.66	8.87
264	ICP-MS	0.48	25.78	2.33	8.32	8.60
293	DRC/CC-ICP-MS	0.5	26.1	2.28	8.64	9.01
324	ICP-MS	<1	25.703	2.511	8.333	8.670
399	DRC/CC-ICP-MS	0.457	23.2	2.15	7.66	8.16
597	ICP-MS/MS	0.559	25.2	2.26	8.25	8.51
605	ICP-MS	0.494	25.3	2.27	8.21	8.49
606	ICP-MS/MS	0.529	24.7	2.45	9.18	8.09

Based on the grading criteria for Co in Urine, 100% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine Co



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



### Results for Event #3, 2023: Summary Statistics

	Urine Cr (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	7.4	0.81	4.22	4.8	6.33
<b>Upper Limit</b>	10.4	3.81	7.22	7.8	9.33
<b>Lower Limit</b>	4.4	0.00	1.22	1.8	3.33
<b>Robust SD (s*)</b>	0.3	0.15	0.22	0.3	0.21
<b>Robust RSD (%)</b>	4.0	19	5.2	6.9	3.3
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (u)</b>	0.1	0.06	0.09	0.1	0.08

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $15 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



## Results for Event #3, 2023: Performance of Participating Laboratories






		Urine Cr ( $\mu\text{g/L}$ )				
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
Target		7.4	0.81	4.22	4.8	6.33
103	ICP-MS/MS	7.50	0.672	4.51	4.52	6.32
107	DRC/CC-ICP-MS	6.49	0.79	4.16	3.81	5.93
110	DRC/CC-ICP-MS	8.08	1.28	5.19	5.75	7.11
116	ICP-MS/MS	7.7	0.78	4.1	4.58	6.22
147	DRC/CC-ICP-MS	6.86	0.733	4.00	4.53	5.88
264	ICP-MS	7.53	0.84	4.42	4.89	6.42
293	DRC/CC-ICP-MS	7.37	2.74	4.11	4.92	6.47
324	ICP-MS	7.628	1.117	4.933	5.294	6.462
597	ICP-MS/MS	7.18	0.561	3.99	4.56	6.15
605	ICP-MS	7.47	0.737	4.10	4.88	6.42

Based on the grading criteria for Cr in Urine, 100% of results were satisfactory, with 0 of the 10 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine Cr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±3 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±3 µg/L at concentrations less than or equal to 15 µg/L.



## Results for Event #3, 2023: Summary Statistics

	Urine Hg (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	5.5	3.5	25.3	2.2	0.25
<b>Upper Limit</b>	8.5	6.5	32.9	5.2	3.25
<b>Lower Limit</b>	2.5	0.5	17.7	0.0	0.00
<b>Robust SD (s*)</b>	0.6	0.4	2.1	0.3	0.05
<b>Robust RSD (%)</b>	11	12	8.3	14	21
<b>Number of Sample Measurements (N)</b>	12	12	12	12	8
<b>Standard Uncertainty (u)</b>	0.2	0.1	0.8	0.1	NA

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.

An arithmetic mean, SD, RSD and n are provided for sample UE23-15.



### Results for Event #3, 2023: Performance of Participating Laboratories






		Urine Hg (µg/L)				
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
Target		5.5	3.5	25.3	2.2	0.25
103	ICP-MS/MS	5.67	3.64	26.6	2.39	0.242
107	DRC/CC-ICP-MS	4.77	2.67	20.19	1.69	0.24
110	ICP-MS	5.07	3.30	24.9	2.00	0.162
116	ICP-MS/MS	4.70	3.04	23.2	1.79	<0.500
147	ICP-MS	5.32	3.27	25.1	2.11	0.223
264	ICP-MS	6.59	4.49	26.46	2.52	0.23
293	DRC/CC-ICP-MS	5.59	3.48	24.24	2.21	0.31
442	ICP-MS/MS	5.02	3.40	25.2	2.06	0.238
597	ICP-MS/MS	5.29	3.16	23.5	2.07	0.322
605	ICP-MS	6.51	3.91	26.9	2.74	<1.00
606	ICP-MS/MS	6.19	3.95	29.5	2.46	<1.00
686	ICP-MS	5.52	3.42	27.3	2.18	<1.00

Based on the grading criteria for Hg in Urine, 100% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #3, 2023: Summary Figures

### Urine Hg



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



### Results for Event #3, 2023: Summary Statistics

	Urine Mn (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	4.90	8.5	3.51	1.58	0.26
<b>Upper Limit</b>	6.13	10.6	4.39	2.13	0.81
<b>Lower Limit</b>	3.68	6.4	2.63	1.03	0.00
<b>Robust SD (s*)</b>	0.20	0.4	0.18	0.12	0.05
<b>Robust RSD (%)</b>	4.1	4.3	5.1	7.6	20
<b>Number of Sample Measurements (N)</b>	13	13	13	13	9
<b>Standard Uncertainty (u)</b>	0.07	0.1	0.06	0.04	NA

The acceptable range is based on quality specifications:  $\pm 0.55 \mu\text{g/L}$  or  $\pm 25\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.55 \mu\text{g/L}$  at concentrations less than or equal to  $2.2 \mu\text{g/L}$ . Quality specifications for Mn are consistent with those used by other External Quality Assessment Schemes for trace elements. (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry and Laboratory Medicine.2016; 54(12): 1921-1928).

An arithmetic mean, SD, RSD and n are provided for sample UE23-15.



### Results for Event #3, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine Mn (µg/L)				
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
	<b>Target</b>	<b>4.90</b>	<b>8.5</b>	<b>3.51</b>	<b>1.58</b>	<b>0.26</b>
103	ICP-MS/MS	4.89	8.55	3.48	1.46	0.257
107	DRC/CC-ICP-MS	5.414	9.332	3.722	1.668	0.249
110	DRC/CC-ICP-MS	5.16	8.96	3.59	1.53	0.181
116	ICP-MS/MS	4.83	8.39	3.31	1.22	<0.300
147	DRC/CC-ICP-MS	4.62	8.19	3.42	1.57	0.272
220	DRC/CC-ICP-MS	4.59	7.86	3.20	1.38	0.25
264	ICP-MS	5.25	8.83	3.68	1.64	0.35
293	DRC/CC-ICP-MS	5	9.5	3.43	19 ↑	<0.27
324	ICP-MS	4.950	8.788	3.772	1.670	<1
399	DRC/CC-ICP-MS	4.88	8.21	3.40	1.59	0.315
597	ICP-MS/MS	4.98	8.41	3.42	1.54	0.259
605	ICP-MS	4.68	8.27	3.51	1.61	<0.300
606	ICP-MS/MS	4.78	8.31	3.68	1.69	0.194

Based on the grading criteria for Mn in Urine, 98% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine Mn



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±0.55 µg/L or ±25% around the target value, whichever is greater; thus, it is fixed at ±0.55 µg/L at concentrations less than or equal to 2.2 µg/L.



### Results for Event #3, 2023: Summary Statistics

	Urine Pb (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	9.02	1.53	0.79	5.14	2.57
<b>Upper Limit</b>	10.82	2.53	1.79	6.17	3.57
<b>Lower Limit</b>	7.22	0.53	0.00	4.11	1.57
<b>Robust SD (s*)</b>	0.29	0.05	0.04	0.12	0.11
<b>Robust RSD (%)</b>	3.2	3.3	5.2	2.3	4.3
<b>Number of Sample Measurements (N)</b>	15	15	14	15	15
<b>Standard Uncertainty (u)</b>	0.09	0.02	0.01	0.04	0.04

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #3, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine Pb (µg/L)				
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
	<b>Target</b>	<b>9.02</b>	<b>1.53</b>	<b>0.79</b>	<b>5.14</b>	<b>2.57</b>
103	ICP-MS/MS	9.03	1.49	0.714	5.09	2.48
107	ICP-MS	9.240	1.560	0.809	5.209	2.621
110	ICP-MS	9.28	1.57	0.80	5.15	2.65
116	ICP-MS/MS	9.35	1.50	0.855	5.22	2.96
147	ICP-MS	8.91	1.54	0.847	5.18	2.55
220	ICP-MS	8.94	1.56	0.797	4.98	2.59
264	ICP-MS	9.40	1.54	0.78	5.53	2.62
293	DRC/CC-ICP-MS	8.77	1.53	0.7	5.1	2.5
324	ICP-MS	8.758	1.535	<1	4.952	2.507
399	ICP-MS/MS	8.72	1.48	0.766	4.87	2.45
442	ICP-MS/MS	8.95	1.65	0.760	5.09	2.57
597	ICP-MS/MS	9.31	1.60	0.814	5.25	2.74
605	ICP-MS	8.85	1.47	0.729	5.04	2.51
606	ICP-MS/MS	8.59	1.43	0.772	5.32	2.42
686	ICP-MS	9.15	1.54	0.802	5.16	2.69

Based on the grading criteria for Pb in Urine, 100% of results were satisfactory, with 0 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine Pb



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



### Results for Event #3, 2023: Summary Statistics

	Urine TI (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	2.23	1.70	3.71	0.938	0.367
<b>Upper Limit</b>	2.68	2.04	4.45	1.138	0.567
<b>Lower Limit</b>	1.78	1.36	2.97	0.738	0.167
<b>Robust SD (s*)</b>	0.06	0.03	0.06	0.016	0.010
<b>Robust RSD (%)</b>	2.7	1.9	1.6	1.7	2.7
<b>Number of Sample Measurements (N)</b>	12	12	12	12	12
<b>Standard Uncertainty (u)</b>	0.02	0.01	0.02	0.006	0.003

The acceptable range is based on quality specifications: ±0.2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.2 µg/L at concentrations less than or equal to 1 µg/L. These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #3, 2023: Performance of Participating Laboratories






		Urine TI (µg/L)				
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
	<b>Target</b>	<b>2.23</b>	<b>1.70</b>	<b>3.71</b>	<b>0.938</b>	<b>0.367</b>
103	ICP-MS/MS	2.25	1.70	3.74	0.935	0.359
107	ICP-MS	2.321	1.777	3.897	0.977	0.373
110	ICP-MS	2.30	1.71	3.75	0.950	0.383
147	ICP-MS	2.23	1.71	3.72	0.930	0.362
220	ICP-MS	2.20	1.70	3.60	0.924	0.373
264	ICP-MS	2.25	1.72	3.71	0.94	0.37
293	DRC/CC-ICP-MS	2.28	1.7	3.67	0.94	0.37
399	ICP-MS/MS	2.15	1.65	3.57	0.887	0.358
597	ICP-MS/MS	2.22	1.67	3.73	0.940	0.356
605	ICP-MS	2.14	1.64	3.61	0.923	0.360
606	ICP-MS/MS	2.20	1.68	3.77	0.970	0.362
686	ICP-MS	2.26	1.73	3.73	0.934	0.378

Based on the grading criteria for TI in Urine, 100% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine TI



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±0.2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.2 µg/L at concentrations less than or equal to 1 µg/L.



### Results for Event #3, 2023: Summary Statistics

	Urine U (µg/L)				
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Target (Robust Mean (x*))</b>	0.108	0.168	0.0396	0.058	0.005
<b>Upper Limit</b>	0.138	0.202	0.0696	0.088	0.035
<b>Lower Limit</b>	0.078	0.134	0.0096	0.028	0.000
<b>Robust SD (s*)</b>	0.005	0.005	0.0020	0.003	0.001
<b>Robust RSD (%)</b>	4.6	2.8	5.1	5.2	18
<b>Number of Sample Measurements (N)</b>	13	13	13	13	5
<b>Standard Uncertainty (u)</b>	0.002	0.002	0.0007	0.001	NA

The acceptable range is based on quality specifications:  $\pm 0.03 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.03 \mu\text{g/L}$  at concentrations less than or equal to  $0.15 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.

An arithmetic mean, SD, RSD and n are provided for sample UE23-15.



### Results for Event #3, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine U (µg/L)				
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
	<b>Target</b>	<b>0.108</b>	<b>0.168</b>	<b>0.0396</b>	<b>0.058</b>	<b>0.005</b>
103	ICP-MS/MS	0.109	0.171	0.0470	0.0679	<0.0200
107	ICP-MS	0.1057	0.1689	0.0374	0.0569	0.0038
110	ICP-MS	0.113	0.174	0.0403	0.0600	0.0049
116	ICP-MS/MS	0.113	0.172	0.0412	0.0611	<0.0150
147	ICP-MS	0.115	0.168	0.0417	0.0569	<0.00810
220	ICP-MS	0.109	0.171	0.039	0.057	<0.005
264	ICP-MS	0.11	0.17	0.04	0.06	*0.01
324	ICP-MS	0.101	0.160	0.037	0.056	0.004
399	ICP-MS/MS	0.103	0.157	0.037	0.051	0.005
597	ICP-MS/MS	0.109	0.153	0.0409	0.0550	0.00588
605	ICP-MS	0.104	0.166	0.0380	0.0560	<0.00500
606	ICP-MS/MS	0.107	0.165	0.040	0.064	<0.005
686	ICP-MS	0.110	0.172	0.0393	0.0588	<0.0150

Based on the grading criteria for U in Urine, 100% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Urine U



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±0.03 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.03 µg/L at concentrations less than or equal to 0.15 µg/L.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics






Urine AI (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
147	ICP-MS	24.9	<13.8	<13.8	19.1	<13.8
264	ICP-MS	20.71	9.18	11.42	16.73	6.63
293	DRC/CC-ICP-MS	22.93	12.41	14.3	18.35	7.82
324	ICP-MS	17.604	9.166	13.156	16.006	6.186
597	ICP-MS/MS	23.0	9.26	12.2	15.8	8.32
Summary Statistics						
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		21.8	10.0	12.8	17.2	7.2
<b>Arithmetic SD (s)</b>		2.8	1.6	1.2	1.5	1.0
<b>Arithmetic RSD (%)</b>		13	16	9.4	8.7	14
<b>Number of Sample Measurements (N)</b>		5	4	4	5	4

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Urine AI



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics






Urine Cs (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
107	ICP-MS	9.01	1.79	3.56	1.21	4.02
110	ICP-MS	11.0	2.07	4.17	1.40	4.89
147	ICP-MS	9.50	1.77	3.51	1.20	4.28
220	ICP-MS	9.56	1.84	3.58	1.21	4.21
264	ICP-MS	7.10	1.36	2.71	0.91	3.21
399	ICP-MS/MS	9.06	1.77	3.48	1.20	4.14
597	ICP-MS/MS	8.98	1.67	3.30	1.13	3.98
605	ICP-MS	8.57	1.67	3.22	1.09	3.95
606	ICP-MS/MS	9.08	1.70	3.67	1.33	3.96
Summary Statistics						
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		9.1	1.74	3.5	1.19	4.1
<b>Arithmetic SD (s)</b>		1.0	0.19	0.4	0.14	0.4
<b>Arithmetic RSD (%)</b>		11	11	11	12	11
<b>Number of Sample Measurements (N)</b>		9	9	9	9	9

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Urine Cs



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics






Urine Cu (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
110	ICP-MS	105	19.4	70.0	45.6	5.84
116	ICP-MS/MS	98.3	14.2	61.7	37.5	<6.00
147	ICP-MS	109	20.0	71.2	45.9	4.70
264	ICP-MS	107.6	16.2	66.3	41.5	3.73
293	DRC/CC-ICP-MS	95.99	13.99	59.76	36.87	3.18
324	ICP-MS	98.226	14.903	61.931	38.488	3.681
597	ICP-MS/MS	102	14.7	63.3	39.5	3.45
Summary Statistics						
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		102	16.2	65	41	4.1
<b>Arithmetic SD (s)</b>		5	2.5	4	4	1.0
<b>Arithmetic RSD (%)</b>		4.9	15	6.2	9.1	24
<b>Number of Sample Measurements (N)</b>		7	7	7	7	6

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Urine Cu



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Urine Mo (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
103	ICP-MS/MS	77.9	45.0	15.3	31.4	113
107	ICP-MS	77.99	44.93	16.00	31.73	113.00
110	ICP-MS	82.6	47.7	16.9	33.5	123
147	ICP-MS	75.3	43.8	15.5	30.7	111
220	ICP-MS	80.3	45.7	16.1	32.1	116
264	ICP-MS	58.26	33.66	10.11	22.20	85.71
293	DRC/CC-ICP-MS	79.66	46.05	16.95	33.4	127.37
324	ICP-MS	74.977	42.804	15.078	29.838	108.987
399	ICP-MS/MS	75.6	43.8	15.2	30.2	112
597	ICP-MS/MS	72.1	40.7	14.3	28.4	103
605	ICP-MS	72.9	42.6	15.4	30.2	109
606	ICP-MS/MS	75.1	42.7	16.0	33.5	107

### Summary Statistics






	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Robust Mean (x*)</b>	76	43.8	15.5	31.1	111
<b>Robust SD (s*)</b>	4	2.1	0.9	2.0	6
<b>Robust RSD (%)</b>	5.0	4.8	5.8	6.4	5.4
<b>Number of Sample Measurements (N)</b>	12	12	12	12	12
<b>Standard Uncertainty (u)</b>	1	0.7	0.3	0.7	2

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Urine Mo



### Legend:

- HHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Urine Ni (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
103	ICP-MS/MS	1.57	13.2	4.50	3.95	2.82
107	DRC/CC-ICP-MS	1.58	12.41	4.45	3.72	3.01
110	ICP-MS	2.70	15.2	5.62	5.31	4.07
147	ICP-MS	1.89	14.7	4.80	4.51	3.15
264	ICP-MS	0.63	12.89	3.53	3.13	2.00
293	DRC/CC-ICP-MS	1.53	13.4	4.39	4.18	3.63
324	ICP-MS	1.892	13.431	4.837	4.499	3.211
442	DRC/CC-ICP-MS	1.75	14.6	4.66	4.34	3.18
597	ICP-MS/MS	3.57	13.9	4.94	5.02	3.40
605	ICP-MS	1.61	13.2	4.64	4.08	2.87

### Summary Statistics






	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Robust Mean (x*)</b>	1.7	13.6	4.7	4.3	3.2
<b>Robust SD (s*)</b>	0.3	0.8	0.3	0.5	0.4
<b>Robust RSD (%)</b>	17	5.9	6.5	12	13
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (u)</b>	0.1	0.3	0.1	0.2	0.2

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Urine Ni



### Legend:

- HHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics






Urine Pt (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
107	ICP-MS	4.0776	1.2594	0.1819	0.3750	0.9687
110	ICP-MS	4.15	1.25	0.191	0.385	0.965
220	ICP-MS	3.97	1.21	0.19	0.37	0.91
264	ICP-MS	3.81	1.16	0.17	0.33	0.92
293	DRC/CC-ICP-MS	4.07	1.2	0.2	0.37	1
399	ICP-MS/MS	3.98	1.22	0.199	0.352	0.931
605	ICP-MS	3.72	1.14	0.175	0.348	0.886
Summary Statistics						
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		3.97	1.21	0.187	0.361	0.94
<b>Arithmetic SD (s)</b>		0.15	0.04	0.012	0.019	0.04
<b>Arithmetic RSD (%)</b>		3.8	3.3	6.4	5.3	4.1
<b>Number of Sample Measurements (N)</b>		7	7	7	7	7

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Summary Figures

### Urine Pt



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Urine Sb (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
103	ICP-MS/MS	1.15	0.298	0.961	0.113	2.22
107	ICP-MS	1.195	0.345	1.002	0.124	2.196
110	ICP-MS	1.36	0.653	1.09	0.143	2.49
147	ICP-MS	1.19	0.360	0.959	0.135	2.23
220	ICP-MS	1.22	0.361	1.01	0.137	2.28
264	ICP-MS	0.79	0.23	0.66	0.07	1.57
293	DRC/CC-ICP-MS	1.3	0.36	1.06	0.13	2.53
324	ICP-MS	1.148	<1	<1	<1	2.098
399	ICP-MS/MS	1.16	0.334	0.962	0.132	2.20
597	ICP-MS/MS	1.09	0.329	0.930	0.113	2.03
605	ICP-MS	1.26	<0.800	0.876	<0.800	2.25
606	ICP-MS/MS	1.25	0.374	1.02	0.115	2.13

### Summary Statistics






	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Robust Mean (x*)</b>	1.20	0.35	0.97	0.124	2.20
<b>Robust SD (s*)</b>	0.08	0.03	0.08	0.015	0.13
<b>Robust RSD (%)</b>	6.7	9.2	8.2	12	5.9
<b>Number of Sample Measurements (N)</b>	12	10	11	10	12
<b>Standard Uncertainty (u)</b>	0.03	0.01	0.03	0.006	0.05

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Urine Sb



### Legend:

- HHEAR Labs
- Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Urine Se (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
103	ICP-MS/MS	135	54.3	92.1	180	221
110	DRC/CC-ICP-MS	133	53.0	89.6	180	225
147	ICP-MS	135	53.4	90.8	183	231
293	DRC/CC-ICP-MS	120.85	47.39	78.99	160.35	202.21
597	ICP-MS/MS	123	50.8	84.6	166	209

### Summary Statistics






	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
Arithmetic Mean ( $\bar{x}$ )	129	51.8	87	174	218
Arithmetic SD (s)	7	2.8	5	10	12
Arithmetic RSD (%)	5.4	5.4	5.7	5.7	5.5
Number of Sample Measurements (N)	5	5	5	5	5

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Summary Figures

### Urine Se



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics

Urine Sn (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
107	ICP-MS	2.73	0.78	5.55	2.21	3.13
110	ICP-MS	3.31	0.95	6.79	2.85	3.65
147	ICP-MS	2.41	0.664	5.15	2.04	2.81
220	ICP-MS	2.76	0.714	5.22	2.23	3.31
264	ICP-MS	1.49	*0.23	2.11	0.70	1.73
293	DRC/CC-ICP-MS	2.88	0.78	4.75	2.19	3.11
324	ICP-MS	2.461	<1	4.942	1.962	2.869
399	ICP-MS/MS	2.58	0.729	5.24	2.31	3.24
597	ICP-MS/MS	2.17	0.493	3.43	0.994	2.30
605	ICP-MS	2.53	<0.900	5.51	2.24	2.94

Summary Statistics					
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Robust Mean (x*)</b>	2.6	0.73	5.1	2.14	3.0
<b>Robust SD (s*)</b>	0.3	0.14	0.6	0.21	0.4
<b>Robust RSD (%)</b>	13	19	12	9.8	12
<b>Number of Sample Measurements (N)</b>	10	7	10	10	10
<b>Standard Uncertainty (u)</b>	0.1	NA	0.2	0.08	0.1






\*Denotes a statistical Outlier.

An arithmetic mean, SD, RSD and n are provided for sample UE23-12.



## Results for Event #3, 2023: Summary Figures

### Urine Sn



#### Legend:

- HHEAR Labs
- Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics

#### Urine Sr (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
103	ICP-MS/MS	34.7	135	274	72.2	131
107	ICP-MS	31.6	118.7	288.0	65.7	114.2
200	ICP-MS	*67.5	122.7	264.6	67.5	117.4
220	ICP-MS	35.2	134	281	72.5	129
264	ICP-MS	29.51	112.7	240.0	61.9	107.2
399	DRC/CC-ICP-MS	32.5	122	264	66.9	121
597	ICP-MS/MS	32.7	120	254	65.7	117
605	ICP-MS	32.9	126	266	70.6	124

#### Summary Statistics






	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
Arithmetic Mean ( $\bar{x}$ )	32.7	124	266	67.9	120
Arithmetic SD (s)	1.9	8	15	3.7	8
Arithmetic RSD (%)	5.8	6.5	5.6	5.4	6.7
Number of Sample Measurements (N)	7	8	8	8	8

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Urine Sr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics






Urine V (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
116	ICP-MS/MS	0.916	3.33	1.37	0.436	6.04
147	DRC/CC-ICP-MS	0.944	3.46	1.36	0.466	6.38
293	DRC/CC-ICP-MS	0.93	3.46	*1.48	0.51	6.54
597	ICP-MS/MS	0.911	3.30	1.36	0.411	6.15
605	ICP-MS	*0.802	3.31	1.33	0.383	5.99
Summary Statistics						
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.925	3.37	1.36	0.44	6.2
<b>Arithmetic SD (s)</b>		0.015	0.08	0.02	0.05	0.2
<b>Arithmetic RSD (%)</b>		1.6	2.4	1.3	11	3.7
<b>Number of Sample Measurements (N)</b>		4	5	4	5	5

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Summary Figures

### Urine V



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Urine W (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
107	ICP-MS	1.166	4.144	0.855	2.268	0.507
110	ICP-MS	1.20	4.22	0.872	2.23	0.519
147	ICP-MS	1.05	3.97	0.859	2.19	0.474
200	ICP-MS	0.92	3.71	0.88	2.24	0.44
220	ICP-MS	1.20	4.25	0.852	2.32	0.519
264	ICP-MS	1.08	3.93	0.78	2.08	0.44
324	ICP-MS	1.130	3.881	<1	2.044	<1
399	ICP-MS/MS	1.16	4.09	0.821	2.22	0.523
597	ICP-MS/MS	1.20	3.98	0.826	2.20	0.521
605	ICP-MS	1.06	3.99	0.806	2.14	0.465
606	ICP-MS/MS	1.12	3.89	0.878	2.38	0.487

### Summary Statistics






	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Robust Mean (x*)</b>	1.13	4.00	0.84	2.21	0.49
<b>Robust SD (s*)</b>	0.08	0.16	0.03	0.08	0.04
<b>Robust RSD (%)</b>	7.1	4.0	4.0	3.6	7.3
<b>Number of Sample Measurements (N)</b>	11	11	10	11	10
<b>Standard Uncertainty (u)</b>	0.03	0.06	0.01	0.03	0.01

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Urine W



### Legend:

- HHEAR Labs
- Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics






Urine Zn (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
110	ICP-MS	476	121	211	741	210
147	ICP-MS	541	*140	238	830	246
264	ICP-MS	518.5	122.5	228.5	808.8	220.6
293	DRC/CC-ICP-MS	465.36	119.61	207.84	720.26	223.53
324	ICP-MS	474.964	115.610	209.825	745.595	207.073
597	ICP-MS/MS	500	116	218	784	214
Summary Statistics						
		UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		496	119	219	770	220
<b>Arithmetic SD (s)</b>		29	3	12	40	14
<b>Arithmetic RSD (%)</b>		5.8	2.5	5.5	5.2	6.4
<b>Number of Sample Measurements (N)</b>		6	5	6	6	6

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Summary Figures

### Urine Zn



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Urine I (µg/L)						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
110	ICP-MS	106	156	147	143	74.6
147	ICP-MS	92.7	146	144	135	75.9

Summary Statistics						
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15	
Arithmetic Mean ( $\bar{x}$ )	99	151	146	139	75.3	
Arithmetic SD (s)	9	7	2	6	0.9	
Arithmetic RSD (%)	9.1	4.6	1.4	4.3	1.2	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Urine Te ( $\mu\text{g/L}$ )						
Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
110	ICP-MS	0.257	1.37	0.762	0.664	3.10
147	ICP-MS	0.222	1.34	0.676	0.666	3.25

Summary Statistics						
	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15	
Arithmetic Mean ( $\bar{x}$ )	0.24	1.36	0.72	0.665	3.17	
Arithmetic SD (s)	0.02	0.02	0.06	0.001	0.11	
Arithmetic RSD (%)	8.3	1.5	8.3	0.21	3.5	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Additional Elements in Urine

### Urine Ag (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
147	ICP-MS	<0.151	<0.151	<0.151	<0.151	<0.151

### Urine B (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
200	ICP-MS	216	400	400	367	162

### Urine Bi (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
147	ICP-MS	<0.0794	<0.0794	<0.0794	<0.0794	<0.0794
597	ICP-MS/MS	<0.0196	<0.0196	<0.0196	<0.0196	<0.0196

### Urine Fe (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
324	ICP-MS	15.463	56.207	15.009	73.875	11.200

### Urine Li (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
147	ICP-MS	5.36	8.74	9.37	8.88	4.14

### Urine Mg (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
597	ICP-MS/MS	12300	20000	24200	22300	11700

### Urine Th (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
147	ICP-MS	<0.102	<0.102	<0.102	<0.102	<0.102
597	ICP-MS/MS	0.0257	0.0310	0.0220	<0.0169	<0.0169

### Urine Ti (µg/L)

Lab Code	Method	UE23-11	UE23-12	UE23-13	UE23-14	UE23-15
597	ICP-MS/MS	4.68	2.26	6.76	1.79	2.73



**Department  
of Health**

**Wadsworth  
Center**

**Event #3, 2023**

**Trace Elements in  
Serum**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #3, 2023: Trace Elements in Serum

### PT Materials

Test materials were prepared from human serum obtained from Zen-Bio, Inc. The company certifies that these materials were tested by FDA approved methods and found to be negative for HIV 1Z2 and HIV-1 RNA, and non-reactive to HBsAg, HCV3 and STS. Units of serum were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with aluminum (Al), cobalt (Co), chromium (Cr), copper (Cu), selenium (Se), zinc (Zn), arsenic (As), beryllium (Be), cadmium (Cd), mercury (Hg), manganese (Mn), molybdenum (Mo), nickel (Ni), lead (Pb), platinum (Pt), antimony (Sb), tin (Sn), strontium (Sr), titanium (Ti), thallium (Tl), uranium (U), vanadium (V) and tungsten (W). PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Six elements in serum are formally graded: Al, Co, Cr, Cu, Se, and Zn. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 27 were reported by at least one participant: As, B, Ba, Be, Bi, Cd, Cs, Fe, Hg, I, Li, Mg, Mn, Mo, Ni, Pb, Pt, Sb, Sn, Sr, Te, Th, Ti, Tl, U, V, and W. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



### Results for Event #3, 2023: Summary Statistics

	Serum AI (µg/L)				
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	23.0	35.1	11.9	65.5	9.1
<b>Upper Limit</b>	28.0	42.1	16.9	78.6	14.1
<b>Lower Limit</b>	18.0	28.1	6.9	52.4	4.1
<b>Arithmetic SD (s)</b>	0.5	1.7	2.4	3.2	2.5
<b>Arithmetic RSD (%)</b>	2.2	4.8	20	4.9	27
<b>Number of Sample Measurements (N)</b>	5	6	6	6	6

The acceptable range is based on quality specifications:  $\pm 5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 5 \mu\text{g/L}$  at concentrations less than or equal to  $25 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #3, 2023: Performance of Participating Laboratories






Lab Code	Method	Serum AI (µg/L)				
		SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
	<b>Target</b>	<b>23.0</b>	<b>35.1</b>	<b>11.9</b>	<b>65.5</b>	<b>9.1</b>
147	ETAAS-Z	23.5	35.3	11.6	67.6	5.86
264	ICP-MS	22.58	36.90	8.96	65.10	8.62
293	DRC/CC-ICP-MS	22.31	35.22	13.17	68.82	8.33
391	ETAAS-Z	*27.19	32.41	15.84	59.64	13.32
485	HR-ICP-MS	22.9	36.7	10.5	65.7	8.18
597	ICP-MS/MS	23.5	34.0	11.5	66.2	10.1

Based on the grading criteria for AI in Serum, 100% of results were satisfactory, with 0 of the 6 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Serum AI



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±5 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±5 µg/L at concentrations less than or equal to 25 µg/L.



### Results for Event #3, 2023: Summary Statistics

	Serum Co (µg/L)				
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	3.34	8.6	5.25	1.99	0.87
<b>Upper Limit</b>	4.84	10.1	6.75	3.49	2.37
<b>Lower Limit</b>	1.84	7.1	3.75	0.49	0.00
<b>Arithmetic SD (s)</b>	0.08	0.4	0.25	0.08	0.06
<b>Arithmetic RSD (%)</b>	2.4	4.5	4.8	4.0	6.9
<b>Number of Sample Measurements (N)</b>	6	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #3, 2023: Performance of Participating Laboratories






		Serum Co (µg/L)				
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
	<b>Target</b>	<b>3.34</b>	<b>8.6</b>	<b>5.25</b>	<b>1.99</b>	<b>0.87</b>
103	ICP-MS/MS	*2.95	7.98	4.87	1.85	0.818
110	ICP-MS/MS	3.45	8.98	5.51	2.10	0.93
147	DRC/CC-ICP-MS	3.30	8.43	5.10	1.97	0.925
264	ICP-MS	3.33	8.51	5.18	2.01	0.80
293	DRC/CC-ICP-MS	3.4	8.90	5.59	1.99	0.92
485	HR-ICP-MS	3.33	9.08	5.3	2.06	0.828
597	ICP-MS/MS	3.23	8.46	5.19	1.96	0.863

Based on the grading criteria for Co in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Serum Co



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories. Gray area = acceptable range based on quality specifications:

±1.5 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 10 µg/L.



### Results for Event #3, 2023: Summary Statistics

	Serum Cr (µg/L)				
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	0.65	11.2	1.33	3.79	4.6
<b>Upper Limit</b>	2.65	13.4	3.33	5.79	6.6
<b>Lower Limit</b>	0.00	9.0	0.00	1.79	2.6
<b>Arithmetic SD (s)</b>	0.08	0.6	0.07	0.13	0.3
<b>Arithmetic RSD (%)</b>	12	5.4	5.3	3.4	6.1
<b>Number of Sample Measurements (N)</b>	6	7	7	6	7

The acceptable range is based on quality specifications:  $\pm 2$  µg/L or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2$  µg/L at concentrations less than or equal to 10 µg/L. These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #3, 2023: Performance of Participating Laboratories






		Serum Cr (µg/L)				
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
	<b>Target</b>	<b>0.65</b>	<b>11.2</b>	<b>1.33</b>	<b>3.79</b>	<b>4.6</b>
103	ICP-MS/MS	0.530	10.8	1.30	3.73	4.11
110	ICP-MS/MS	*1.45	11.8	1.34	*4.57	4.74
147	DRC/CC-ICP-MS	0.593	10.3	1.22	3.62	4.41
264	ICP-MS	0.60	11.61	1.46	3.99	4.87
293	DRC/CC-ICP-MS	0.72	11.71	1.34	3.75	4.87
485	HR-ICP-MS	0.68	11.6	1.29	3.84	4.43
597	ICP-MS/MS	0.747	10.9	1.36	3.83	4.59

Based on the grading criteria for Cr in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #3, 2023: Summary Figures

### Serum Cr



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #3, 2023: Summary Statistics

	Serum Cu (µg/L)				
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	1050	1170	1420	790	1720
<b>Upper Limit</b>	1210	1350	1630	909	1980
<b>Lower Limit</b>	890	990	1210	672	1460
<b>Arithmetic SD (s)</b>	50	50	60	34	80
<b>Arithmetic RSD (%)</b>	4.8	4.3	4.2	4.3	4.7
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #3, 2023: Performance of Participating Laboratories






		Serum Cu (µg/L)				
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Target		1050	1170	1420	790	1720
107	DRC/CC-ICP-MS	1023	1165	1424	784	1754
110	ICP-MS/MS	1120	1230	1480	840	1820
147	DRC/CC-ICP-MS	1048	1163	1391	769	1684
264	ICP-MS	1056.2	1175.7	1397.3	791.1	1703.9
293	DRC/CC-ICP-MS	966	1087	1322	731	1570
436	FAAS	1080	1207	1509	800	1735
597	ICP-MS/MS	1080	1190	1430	814	1760

Based on the grading criteria for Cu in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #3, 2023: Summary Figures

### Serum Cu



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ .



## Results for Event #3, 2023: Summary Statistics

	Serum Se (µg/L)				
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	107	140	238	201	123
<b>Upper Limit</b>	128	168	286	241	148
<b>Lower Limit</b>	86	112	190	161	98
<b>Arithmetic SD (s)</b>	4	5	8	6	3
<b>Arithmetic RSD (%)</b>	3.5	3.6	3.4	3.1	2.8
<b>Number of Sample Measurements (N)</b>	7	8	8	8	8

The acceptable range is based on quality specifications:  $\pm 2$  µg/L or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2$  µg/L at concentrations less than or equal to 10 µg/L. These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #3, 2023: Performance of Participating Laboratories






		Serum Se (µg/L)				
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
	<b>Target</b>	<b>107</b>	<b>140</b>	<b>238</b>	<b>201</b>	<b>123</b>
103	ICP-MS/MS	106	145	241	203	123
107	DRC/CC-ICP-MS	103.2	138.1	236.4	198.2	124.1
110	ICP-MS/MS	111	147	251	208	129
147	DRC/CC-ICP-MS	104	139	232	191	120
264	ICP-MS	108.49	142.02	233.59	200.07	125.81
293	DRC/CC-ICP-MS	102	137	234	194	122
436	ETAAS-Z	*90.00	131.00	227.00	202.00	118.00
597	ICP-MS/MS	111	143	248	209	123

Based on the grading criteria for Se in Serum, 100% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #3, 2023: Summary Figures

## Serum Se



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±2 µg/L at concentrations less than or equal to 10 µg/L.



## Results for Event #3, 2023: Summary Statistics

	Serum Zn (µg/L)				
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	860	1030	1310	660	1170
<b>Upper Limit</b>	990	1180	1510	760	1350
<b>Lower Limit</b>	730	880	1110	560	990
<b>Arithmetic SD (s)</b>	50	50	80	40	70
<b>Arithmetic RSD (%)</b>	5.8	4.9	6.1	6.1	6.4
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

The acceptable range is based on quality specifications:  $\pm 15 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 15 \mu\text{g/L}$  at concentrations less than or equal to  $100 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #3, 2023: Performance of Participating Laboratories






		Serum Zn (µg/L)				
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
	<b>Target</b>	<b>860</b>	<b>1030</b>	<b>1310</b>	<b>660</b>	<b>1170</b>
107	DRC/CC-ICP-MS	863	1055	1340	667	1221
110	ICP-MS/MS	890	1080	1360	687	1230
147	DRC/CC-ICP-MS	843	1007	1261	632	1118
264	ICP-MS	836.2	1001.0	1248.5	633.2	1107.5
293	DRC/CC-ICP-MS	784	954	1216	608	1078
597	ICP-MS/MS	919	1100	1410	707	1250

Based on the grading criteria for Zn in Serum, 100% of results were satisfactory, with 0 of the 6 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #3, 2023: Summary Figures

### Serum Zn



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±15 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±15 µg/L at concentrations less than or equal to 100 µg/L.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics






Serum Mn (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	3.60	17.0	6.74	10.4	6.60
110	ICP-MS/MS	4.29	18.5	7.35	11.7	7.18
147	DRC/CC-ICP-MS	3.90	15.6	6.87	10.3	6.65
264	ICP-MS	*5.66	21.44	*8.82	13.15	*8.27
293	DRC/CC-ICP-MS	4.050	18.15	7.31	11.47	6.960
597	ICP-MS/MS	3.96	17.6	7.21	11.2	6.90
Summary Statistics						
		SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		4.0	18.0	7.1	11.4	6.9
<b>Arithmetic SD (s)</b>		0.3	1.9	0.3	1.0	0.2
<b>Arithmetic RSD (%)</b>		6.3	11	3.8	8.8	3.5
<b>Number of Sample Measurements (N)</b>		5	6	5	6	5

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Serum Mn



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics






Serum Mo (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	1.42	0.852	5.80	3.37	2.04
110	ICP-MS/MS	1.73	0.91	6.05	3.63	2.16
147	DRC/CC-ICP-MS	1.63	0.953	5.56	3.26	1.91
293	DRC/CC-ICP-MS	1.710	0.920	6.380	3.540	2.310
485	HR-ICP-MS	1.69	1.02	5.92	3.38	2.01
597	ICP-MS/MS	1.48	0.801	5.56	3.27	1.91
Summary Statistics						
		SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		1.61	0.91	5.9	3.41	2.06
<b>Arithmetic SD (s)</b>		0.13	0.08	0.3	0.15	0.16
<b>Arithmetic RSD (%)</b>		8.1	8.8	5.3	4.4	7.8
<b>Number of Sample Measurements (N)</b>		6	6	6	6	6

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Serum Mo



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics






Serum Ni (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	4.85	6.57	1.14	2.68	0.625
110	ICP-MS/MS	5.58	7.93	1.54	*3.38	1.22
147	DRC/CC-ICP-MS	5.11	7.16	1.35	2.82	0.722
293	DRC/CC-ICP-MS	5.50	7.13	1.33	2.85	0.96
442	DRC/CC-ICP-MS	5.36	7.44	1.35	2.98	0.713
485	HR-ICP-MS	4.84	7.10	1.26	2.69	0.730
597	ICP-MS/MS	5.10	6.99	1.69	2.88	0.912
Summary Statistics						
		SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		5.2	7.2	1.38	2.82	0.84
<b>Arithmetic SD (s)</b>		0.3	0.4	0.18	0.12	0.21
<b>Arithmetic RSD (%)</b>		5.8	5.6	13	4.3	25
<b>Number of Sample Measurements (N)</b>		7	7	7	6	7

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Summary Figures

### Serum Ni



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum V (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
110	ICP-MS/MS	2.03	3.77	0.87	3.09	1.12
147	DRC/CC-ICP-MS	1.85	3.45	0.745	2.72	0.954
293	DRC/CC-ICP-MS	2.06	3.67	0.97	3.10	1.3
485	HR-ICP-MS	1.80	3.48	0.748	2.75	1.00
597	ICP-MS/MS	1.83	3.44	0.797	2.84	1.03






Summary Statistics					
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	1.91	3.56	0.83	2.90	1.08
Arithmetic SD (s)	0.12	0.15	0.10	0.18	0.14
Arithmetic RSD (%)	6.3	4.2	12	6.2	13
Number of Sample Measurements (N)	5	5	5	5	5

\*Denotes a statistical Outlier.



# Results for Event #3, 2023: Summary Figures

## Serum V



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum As (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	19.5	2.86	9.06	1.18	4.39
110	ICP-MS/MS	20.3	2.92	9.43	1.25	4.59
147	DRC/CC-ICP-MS	19.2	2.76	8.91	1.18	4.23
597	ICP-MS/MS	18.8	2.73	9.00	1.17	4.37
Summary Statistics						
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15	
Arithmetic Mean ( $\bar{x}$ )	19.5	2.82	9.1	1.20	4.39	
Arithmetic SD (s)	0.6	0.09	0.2	0.04	0.15	
Arithmetic RSD (%)	3.1	3.2	2.5	3.3	3.4	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Ba (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
110	ICP-MS/MS	3.64	3.23	2.87	3.78	4.21
147	ICP-MS	3.87	3.20	2.69	3.45	3.68
597	ICP-MS/MS	3.44	3.17	2.83	3.55	3.69
Summary Statistics						
		SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )		3.7	3.20	2.80	3.6	3.9
Arithmetic SD (s)		0.2	0.03	0.09	0.2	0.3
Arithmetic RSD (%)		5.9	0.94	3.2	4.7	7.7
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Be (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
110	ICP-MS/MS	0.47	5.67	3.44	1.31	0.78
147	ICP-MS	0.536	5.11	3.33	1.25	0.784
293	ICP-MS	0.52	5.39	3.40	1.19	0.79
597	ICP-MS/MS	0.437	5.08	3.36	1.19	0.838
Summary Statistics						
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15	
Arithmetic Mean ( $\bar{x}$ )	0.49	5.3	3.38	1.24	0.80	
Arithmetic SD (s)	0.05	0.3	0.05	0.06	0.03	
Arithmetic RSD (%)	10	5.3	1.5	4.8	3.4	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Cd (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	0.447	1.47	2.96	5.72	0.865
110	ICP-MS/MS	0.45	1.48	3.14	5.90	0.91
147	ICP-MS	0.418	1.39	2.81	5.29	0.816
597	ICP-MS/MS	0.426	1.40	2.95	5.56	0.841

Summary Statistics					
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	0.435	1.44	2.96	5.6	0.86
Arithmetic SD (s)	0.016	0.05	0.14	0.3	0.04
Arithmetic RSD (%)	3.7	3.5	4.7	4.6	4.7
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Cs ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
110	ICP-MS/MS	0.82	0.91	1.99	0.68	0.34
597	ICP-MS/MS	0.772	0.862	1.88	0.640	0.332

Summary Statistics						
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15	
Arithmetic Mean ( $\bar{x}$ )	0.80	0.89	1.94	0.66	0.336	
Arithmetic SD (s)	0.03	0.03	0.08	0.03	0.006	
Arithmetic RSD (%)	3.8	3.4	4.1	4.5	1.8	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Hg (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	8.91	4.19	2.12	5.08	1.14
110	ICP-MS/MS	9.57	4.35	2.29	5.45	1.21
597	ICP-MS/MS	8.40	3.90	2.14	4.80	1.09
Summary Statistics						
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15	
Arithmetic Mean ( $\bar{x}$ )	9.0	4.1	2.18	5.1	1.15	
Arithmetic SD (s)	0.6	0.2	0.09	0.3	0.06	
Arithmetic RSD (%)	6.7	5.5	4.1	5.9	5.2	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Serum I (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
147	ICP-MS	58.1	58.9	66.7	57.3	48.5
442	ICP-MS	53.9	57.5	68.1	57.6	48.4

### Summary Statistics

	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	56.0	58.2	67.4	57.5	48.5
Arithmetic SD (s)	3.0	1.0	1.0	0.2	0.1
Arithmetic RSD (%)	5.4	1.7	1.5	0.35	0.14
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Mg (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
264	ICP-MS	20266.0	21702.0	21702.0	19225.0	19893.5
597	ICP-MS/MS	20800	22300	23100	20400	20000

Summary Statistics						
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15	
Arithmetic Mean ( $\bar{x}$ )	20500	22000	22400	19800	19950	
Arithmetic SD (s)	400	400	1000	800	80	
Arithmetic RSD (%)	1.8	1.8	4.5	4.2	0.38	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Pb ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	4.74	1.08	2.26	3.37	9.54
110	ICP-MS/MS	4.97	1.16	2.38	3.73	14.5
597	ICP-MS/MS	4.93	1.21	2.26	3.41	9.57
Summary Statistics						
		SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )		4.88	1.15	2.30	3.5	11
Arithmetic SD (s)		0.12	0.07	0.07	0.2	3
Arithmetic RSD (%)		2.5	6.1	3.0	5.7	27
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Pt (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
110	ICP-MS/MS	0.398	0.940	1.88	1.24	0.610
264	ICP-MS	0.48	1.09	2.16	1.37	0.64
293	DRC/CC-ICP-MS	0.38	0.95	1.98	1.19	0.62
Summary Statistics						
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15	
Arithmetic Mean ( $\bar{x}$ )	0.42	0.99	2.01	1.27	0.623	
Arithmetic SD (s)	0.05	0.08	0.14	0.09	0.015	
Arithmetic RSD (%)	12	8.1	7.1	7.1	2.4	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Sb (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	1.11	2.77	2.02	3.97	5.99
110	ICP-MS/MS	1.09	2.72	1.97	3.80	5.83
147	ICP-MS	1.09	2.53	1.81	3.53	5.42
597	ICP-MS/MS	0.959	2.32	1.69	3.31	5.11

Summary Statistics					
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	1.06	2.6	1.87	3.7	5.6
Arithmetic SD (s)	0.07	0.2	0.15	0.3	0.4
Arithmetic RSD (%)	6.6	7.7	8.1	7.9	7.1
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Serum Sn (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
110	ICP-MS/MS	6.75	4.04	2.52	0.95	1.79
597	ICP-MS/MS	5.80	3.65	2.17	0.793	1.53

### Summary Statistics

	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	6.3	3.8	2.3	0.87	1.7
Arithmetic SD (s)	0.7	0.3	0.2	0.11	0.2
Arithmetic RSD (%)	11	7.9	8.7	13	11
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Sr ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	101	116	85.5	117	116
200	ICP-MS	105.1	120.9	87.6	114.8	128.8
597	ICP-MS/MS	101	114	85.8	120	115
Summary Statistics						
		SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )		102	117	86.3	117	120
Arithmetic SD (s)		2	4	1.1	3	8
Arithmetic RSD (%)		2.3	3.4	1.3	2.6	6.7
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum Ti ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
200	DRC/CC-ICP-MS	5.3	3.2	5.3	4.1	6.9
442	ICP-MS/MS	7.80	3.60	4.43	5.93	11.2
485	HR-ICP-MS	7.16	3.09	3.87	5.33	9.83
597	ICP-MS/MS	9.42	5.29	6.22	8.16	12.3

Summary Statistics					
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	7.4	3.8	5.0	5.9	10
Arithmetic SD (s)	1.7	1.0	1.0	1.7	2
Arithmetic RSD (%)	23	26	20	29	23
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

Serum TI (µg/L)						
Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	3.04	1.55	2.11	0.899	0.683
110	ICP-MS/MS	3.15	1.56	2.12	0.93	0.69
147	ICP-MS	3.09	1.50	1.97	0.861	0.656
597	ICP-MS/MS	3.07	1.51	2.11	0.910	0.684

Summary Statistics					
	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	3.09	1.53	2.08	0.90	0.678
Arithmetic SD (s)	0.05	0.03	0.07	0.03	0.015
Arithmetic RSD (%)	1.6	1.9	3.4	3.2	2.2
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Laboratory Data and Summary Statistics

### Serum U (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
103	ICP-MS/MS	0.146	0.121	0.166	0.101	0.0606
110	ICP-MS/MS	0.158	0.123	0.183	0.106	0.0616
597	ICP-MS/MS	0.144	0.118	0.177	0.0994	0.0554

### Summary Statistics

	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	0.149	0.121	0.175	0.102	0.059
Arithmetic SD (s)	0.008	0.003	0.009	0.003	0.003
Arithmetic RSD (%)	5.4	2.1	5.1	2.9	5.1
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



### Results for Event #3, 2023: Laboratory Data and Summary Statistics

#### Serum W (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
110	ICP-MS/MS	1.64	0.54	2.42	1.07	0.24
200	ICP-MS	1.69	0.57	2.67	1.18	0.26
597	ICP-MS/MS	1.48	0.515	2.31	1.00	0.240

#### Summary Statistics

	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
Arithmetic Mean ( $\bar{x}$ )	1.60	0.54	2.5	1.08	0.247
Arithmetic SD (s)	0.11	0.03	0.2	0.09	0.012
Arithmetic RSD (%)	6.9	5.6	7.3	8.3	4.9
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #3, 2023: Additional Elements in Serum

### Serum B (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
200	ICP-MS	49	53	62	49	56

### Serum Bi (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
147	ICP-MS	<0.0397	<0.0397	<0.0397	<0.0397	<0.0397
597	ICP-MS/MS	<0.0278	<0.0278	0.0500	<0.0278	0.0298

### Serum Fe (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
264	ICP-MS	1005.5	831.1	609.9	738.9	1309.6

### Serum Li (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
147	ICP-MS	1.17	0.756	1.07	0.951	1.60

### Serum Te (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
110	ICP-MS/MS	0.00	0.00	0.00	0.01	0.00

### Serum Th (µg/L)

Lab Code	Method	SE23-11	SE23-12	SE23-13	SE23-14	SE23-15
597	ICP-MS/MS	<0.0155	<0.0155	<0.0155	<0.0155	<0.0155



## References

1. ISO/FDIS-13528 (2005) Statistical methods for use in proficiency testing by interlaboratory comparisons. International Organization for Standardization, Geneva.
2. Taylor A, Angerer J, Arnaud J, Claeys F, Jones RL, Mazarrasa O, Mairiaux E, Menditto A, Parsons PJ, Patriarca M, Pineau A, Valkonen S, Weber J-P, Weykamp C. Occupational and environmental laboratory medicine: A network of EQAS organisers. Accreditation and Quality Assurance. 2006;11(8-9):435-9. PubMed PMID: 086NJ-0011.