

Immunohematology Proficiency Testing Program
New York State Department of Health
Wadsworth Center
Empire State Plaza, PO Box 509
Albany, New York 12201-0509

Performance Summary for the October 26, 2015 Proficiency Test Event

Number of Participating Laboratories

Satisfactory 248 Unsatisfactory 8

Specimen #	Number (%) of acceptable responses									
	ABO Group	Subgroups of A	Direct Antiglobulin Test			Rh _o (D) Group	Rh Phenotype			
			Polyspecific	IgG	Complement		C (rh')	E (rh'')	c (hr')	e (hr'')
NYS-15301	250 (99%)	190 (99%)	192 (100%)	210 (100%)	152 (99%)	253 (100%)	195 (100%)	195 (100%)	195 (100%)	190 (99%)
NYS-15302	251 (100%)	191 (100%)	192 (100%)	210 (100%)	154 (100%)	252 (99%)	195 (100%)	195 (100%)	195 (100%)	191 (99%)
NYS-15303	251 (100%)	189 (99%)	190 (99%)	210 (100%)	154 (100%)	253 (100%)	193 (99%)	195 (100%)	195 (100%)	192 (100%)
NYS-15304	251 (100%)	189 (99%)	192 (100%)	210 (100%)	154 (100%)	252 (99%)	195 (100%)	195 (100%)	195 (100%)	192 (100%)
NYS-15305	251 (100%)	190 (99%)	192 (100%)	210 (100%)	154 (100%)	253 (100%)	195 (100%)	195 (100%)	195 (100%)	191 (99%)
Total labs	251	191	192	210	154	253	195	195	195	192

Specimen #	Number (%) of acceptable responses		
	Unexpected Antibody Detection	Antibody Identification	Compatibility Testing
NYS-15301	237 (100%)	192 (100%)	202 (99%)
NYS-15302	237 (100%)	191 (99%)	203 (100%)
NYS-15303	236 (99%)	192 (100%)	203 (100%)
NYS-15304	237 (100%)	192 (100%)	203 (100%)
NYS-15305	237 (100%)	192 (100%)	203 (100%)
Total labs	237	192	203

* Satisfactory – No or minor error(s) in interpretation

** Unsatisfactory – Significant error(s) in interpretation or failure to participate