

Nirav R. Shah, M.D., M.P.H.
Commissioner

NEW YORK
state department of
HEALTH

Sue Kelly
Executive Deputy Commissioner

Statistical Report
Urinalysis/Urine Pregnancy Testing
Mail out date: November 26, 2012

This report summarizes data from the urinalysis/urine pregnancy proficiency testing of November 26, 2012.

If you have any questions regarding this report, please contact Regina Saulsbery at 518-474-6156. You may also contact us by e-mail: clinchem@wadsworth.org

Results are shown for each analyte arranged by instrument and reagent system. Results for specific gravity and pH are listed as mean \pm 1SD for each of the instruments and reagents. All other results are shown as the number of participant results for that instrument or reagent.

The use of brand and/or trade names in this report does not constitute an endorsement of the products on the part of the Wadsworth Center or the New York State Department of Health.

PLEASE NOTE: The evaluation report and summary of the educational quantitative urine clinical chemistry proficiency test will be mailed separately if your laboratory participated in that proficiency test.

HEALTH.NY.GOV
[facebook.com/NYSDOH](https://www.facebook.com/NYSDOH)
twitter.com/HealthNYGov

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Specific Gravity

Specimen: U06	Specimen: U07	Specimen: U08	Specimen: U09	Specimen: U10	Number	Instrument or Reagent System
1.018 ± 0.004	1.014 ± 0.004	1.020 ± 0.006	1.015 ± 0.004	1.022 ± 0.004	n = 375	All Methods & Instruments
1.020	1.015	1.020	1.015	1.020	n = 375	Median value, All Methods & Instruments
<Instruments>						
1.013 ± 0.003	1.014 ± 0.004	1.022 ± 0.003	1.013 ± 0.005	1.022 ± 0.003	n = 26	Visual examination
1.020 ± 0.002	1.012 ± 0.002	1.016 ± 0.002	1.014 ± 0.002	1.020 ± 0.001	n = 92	Siemens Atlas
1.021 ± 0.001	1.019 ± 0.002	1.026 ± 0.002	1.020 ± 0.001	1.025 ± 0.002	n = 37	Siemens Clinitek Status
1.016 ± 0.004	1.016 ± 0.004	1.028 ± 0.003	1.017 ± 0.003	1.026 ± 0.004	n = 8	Siemens Clinitek 50
1.022 ± 0.003	1.020 ± 0.000	1.026 ± 0.002	1.021 ± 0.002	1.026 ± 0.002	n = 5	Siemens Clinitek 100
1.018 ± 0.003	1.017 ± 0.003	1.026 ± 0.002	1.018 ± 0.003	1.025 ± 0.004	n = 45	Siemens Clinitek 500
1.018 ± 0.002	1.017 ± 0.003	1.026 ± 0.002	1.019 ± 0.002	1.025 ± 0.002	n = 47	Siemens Clinitek Advantus
1.012 ± 0.003	1.017 ± 0.003	1.020 ± 0.000	1.015 ± 0.005	1.018 ± 0.003	n = 3	Roche UA Reader/Super UA Reader
1.013 ± 0.004	1.014 ± 0.003	1.020 ± 0.001	1.013 ± 0.002	1.018 ± 0.002	n = 13	Roche Chemstrip Criterion
1.017 ± 0.003	1.013 ± 0.002	1.018 ± 0.002	1.014 ± 0.002	1.019 ± 0.001	n = 34	Roche Urisys
1.015 ± 0.001	1.009 ± 0.001	1.013 ± 0.001	1.009 ± 0.001	1.017 ± 0.001	n = 37	IRIS AUTION MAX
1.015 ± 0.001	1.010 ± 0.000	1.013 ± 0.001	1.009 ± 0.001	1.018 ± 0.000	n = 7	IRIS iChem Velocity
1.020 ± 0.002	1.011 ± 0.001	1.015 ± 0.001	1.013 ± 0.001	1.019 ± 0.001	n = 17	Refractometer
<Reagents>						
1.018 ± 0.004	1.017 ± 0.004	1.025 ± 0.003	1.018 ± 0.004	1.025 ± 0.004	n = 167	Siemens Reagent Strips
1.020 ± 0.002	1.012 ± 0.002	1.016 ± 0.002	1.014 ± 0.002	1.020 ± 0.001	n = 87	Siemens Atlas Reagent
1.014 ± 0.003	1.014 ± 0.002	1.020 ± 0.001	1.014 ± 0.003	1.019 ± 0.002	n = 43	Roche Chemstrips
1.020 ± 0.001	1.012 ± 0.001	1.015 ± 0.000	1.013 ± 0.001	1.020 ± 0.001	n = 13	Roche Urisys Cassette
1.015 ± 0.001	1.009 ± 0.000	1.012 ± 0.001	1.009 ± 0.001	1.017 ± 0.001	n = 22	ARKRAY Aution Sticks
1.015 ± 0.001	1.010 ± 0.000	1.013 ± 0.001	1.009 ± 0.001	1.018 ± 0.001	n = 17	IRIS Reagent Strips
1.019 ± 0.002	1.011 ± 0.001	1.014 ± 0.001	1.012 ± 0.002	1.018 ± 0.001	n = 22	None, no reagent or strip used
1.014 ± 0.009	1.011 ± 0.006	1.016 ± 0.010	1.011 ± 0.010	1.017 ± 0.006	n = 4	Other

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

pH

Specimen: U06	Specimen: U07	Specimen: U08	Specimen: U09	Specimen: U10	Number	Instrument or Reagent System
7.08 ± 0.20 7.00	6.73 ± 0.28 6.50	5.99 ± 0.17 6.00	7.05 ± 0.18 7.00	5.92 ± 0.26 6.00	n = 380 n = 380	All Methods & Instruments Median value, All Methods & Instruments
<Instruments>						
7.25 ± 0.33	6.59 ± 0.34	5.88 ± 0.39	7.31 ± 0.35	5.79 ± 0.45	n = 34	Visual examination
7.12 ± 0.21	6.82 ± 0.24	6.01 ± 0.05	7.01 ± 0.07	6.00 ± 0.00	n = 92	Siemens Atlas
7.01 ± 0.08	7.00 ± 0.00	6.06 ± 0.17	7.00 ± 0.00	5.95 ± 0.19	n = 41	Siemens Clinitek Status
7.00 ± 0.00	6.56 ± 0.17	6.00 ± 0.00	7.00 ± 0.00	5.94 ± 0.17	n = 9	Siemens Clinitek 50
7.42 ± 0.20	6.92 ± 0.20	6.08 ± 0.20	7.42 ± 0.20	5.92 ± 0.20	n = 6	Siemens Clinitek 100
7.07 ± 0.21	6.61 ± 0.31	5.97 ± 0.12	7.05 ± 0.19	5.98 ± 0.10	n = 47	Siemens Clinitek 500
7.10 ± 0.20	6.66 ± 0.28	6.00 ± 0.00	7.06 ± 0.17	5.96 ± 0.14	n = 49	Siemens Clinitek Advantus
7.00 ± 0.00	6.67 ± 0.29	6.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	n = 3	Roche UA Reader/Super UA Reader
7.00 ± 0.00	6.61 ± 0.21	6.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	n = 14	Roche Chemstrip Criterion
7.00 ± 0.00	6.78 ± 0.25	5.96 ± 0.26	7.00 ± 0.00	5.56 ± 0.50	n = 34	Roche Urisys
7.00 ± 0.00	6.53 ± 0.11	6.00 ± 0.00	7.00 ± 0.00	5.91 ± 0.20	n = 38	IRIS AUTION MAX
7.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	6.00 ± 0.00	n = 8	IRIS iChem Velocity
<Reagents>						
7.12 ± 0.24	6.73 ± 0.29	6.01 ± 0.14	7.10 ± 0.23	5.97 ± 0.16	n = 181	Siemens Reagent Strips
7.11 ± 0.21	6.82 ± 0.24	6.01 ± 0.07	7.01 ± 0.07	6.00 ± 0.00	n = 88	Siemens Atlas Reagent
7.00 ± 0.00	6.60 ± 0.29	5.88 ± 0.30	7.00 ± 0.00	5.68 ± 0.46	n = 47	Roche Chemstrips
7.00 ± 0.00	6.92 ± 0.19	5.96 ± 0.32	7.00 ± 0.00	5.46 ± 0.52	n = 13	Roche Urisys Cassette
7.00 ± 0.00	6.50 ± 0.00	6.00 ± 0.00	7.00 ± 0.00	5.91 ± 0.19	n = 29	ARKRAY Aution Sticks
7.00 ± 0.00	6.79 ± 0.25	6.00 ± 0.00	7.00 ± 0.00	5.94 ± 0.17	n = 17	IRIS Reagent Strips
7.05 ± 0.26	6.72 ± 0.36	6.00 ± 0.29	6.98 ± 0.47	5.90 ± 0.18	n = 4	Other

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Bilirubin	Specimen: U06	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		58	0	0	0	0
Siemens Atlas		91	0	0	0	0
Siemens Clinitek Status		39	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0
Siemens Clinitek 500		46	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		33	0	0	0	0
Roche Cobas U411		1	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0
<Reagents>						
Siemens Reagent Strips		176	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	0
Roche Chemstrips		43	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0
Siemens Ictotest		29	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Hitachi/Roche Reagents		1	0	0	0	0
Other		2	0	0	0	0
All Instruments and Reagents		393	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Bilirubin	Specimen: U07	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		58	0	0	0	0
Siemens Atlas		91	0	0	0	0
Siemens Clinitek Status		39	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0
Siemens Clinitek 500		46	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		33	0	0	0	0
Roche Cobas U411		1	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0
<Reagents>						
Siemens Reagent Strips		176	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	0
Roche Chemstrips		43	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0
Siemens Ictotest		29	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Hitachi/Roche Reagents		1	0	0	0	0
Other		2	0	0	0	0
All Instruments and Reagents		393	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Bilirubin	Specimen: U08	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		58	0	0	0	0
Siemens Atlas		91	0	0	0	0
Siemens Clinitek Status		39	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0
Siemens Clinitek 500		46	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		33	0	0	0	0
Roche Cobas U411		1	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0
<Reagents>						
Siemens Reagent Strips		176	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	0
Roche Chemstrips		43	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0
Siemens Ictotest		29	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Hitachi/Roche Reagents		1	0	0	0	0
Other		2	0	0	0	0
All Instruments and Reagents		393	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Bilirubin	Specimen: U09	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		58	0	0	0	0
Siemens Atlas		91	0	0	0	0
Siemens Clinitek Status		39	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0
Siemens Clinitek 500		46	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0
Roche Chemstrip Criterion		13	0	0	0	0
Roche Urisys		33	0	0	0	0
Roche Cobas U411		1	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0
IRIS iChem Velocity		7	0	0	0	0
<Reagents>						
Siemens Reagent Strips		176	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	0
Roche Chemstrips		43	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0
Siemens Ictotest		29	0	0	0	0
Siemens Hemastix		1	0	0	0	0
Hitachi/Roche Reagents		1	0	0	0	0
Other		2	0	0	0	0
All Instruments and Reagents		393	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Bilirubin	Specimen: U10	Negative	Small 1+ 1mg/dL	Moderate 2+ 3mg/dL	Large 3+ 6mg/dL	Positive
<Instruments>						
Visual examination		1	0	16	28	90
Siemens Atlas		0	0	83	8	0
Siemens Clinitek Status		1	1	30	6	1
Siemens Clinitek 50		0	0	4	4	1
Siemens Clinitek 100		0	0	1	5	0
Siemens Clinitek 500		1	0	44	0	2
Siemens Clinitek Advantus		0	0	48	0	1
Roche UA Reader/Super UA Read		0	2	1	0	0
Roche Chemstrip Criterion		0	11	2	0	1
Roche Urisys		1	27	4	0	1
Roche Cobas U411		0	0	1	0	0
IRIS AUTION MAX		1	3	33	0	1
IRIS iChem Velocity		7	0	0	0	0
<Reagents>						
Siemens Reagent Strips		2	1	135	34	4
Siemens Atlas Reagent		0	1	76	7	0
Roche Chemstrips		0	29	10	5	1
Roche Urisys Cassette		1	10	1	0	1
ARKRAY Aution Sticks		0	2	24	0	1
IRIS Reagent Strips		8	1	8	0	0
Siemens Ictotest		1	0	10	4	90
Siemens Hemastix		0	0	1	0	0
Hitachi/Roche Reagents		0	0	0	1	0
Other		0	0	2	0	1
All Instruments and Reagents		12	44	267	51	98

**New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012**

Glucose	Specimen: U06	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		1	32	27	47	12	8	1
Siemens Atlas		1	0	41	51	0	0	0
Siemens Clinitek Status		0	38	4	0	0	0	0
Siemens Clinitek 50		0	6	3	0	0	0	0
Siemens Clinitek 100		0	6	0	0	0	0	0
Siemens Clinitek 500		1	40	7	1	0	0	0
Siemens Clinitek Advantus		0	41	9	0	0	0	0
Roche UA Reader/Super UA Read		0	0	0	0	0	3	0
Roche Chemstrip Criterion		0	0	8	3	0	3	0
Roche Urisys		0	0	22	2	0	11	0
Roche Cobas U411		0	0	1	0	0	0	0
IRIS AUTION MAX		0	0	8	23	3	4	0
IRIS iChem Velocity		0	0	0	7	0	1	0
Hitachi/Roche Analyzer		0	0	1	1	0	0	0
Other Chemistry Analyzer		0	0	1	0	0	0	0
<Reagents>								
Siemens Reagent Strips		2	152	27	4	0	0	0
Siemens Atlas Reagent		1	1	35	49	0	0	0
Sulfosalicylic Acid		0	1	0	0	0	0	0
Roche Chemstrips		0	0	25	11	0	15	0
Roche Urisys Cassette		0	0	7	0	0	6	0
ARKRAY Aution Sticks		0	0	6	17	3	3	0
IRIS Reagent Strips		0	0	2	13	0	2	0
Siemens Acetest		0	1	0	0	0	0	0
Bayer Clinitest		0	4	29	40	12	4	1
Siemens Uristix		0	3	0	0	0	0	0
Hitachi/Roche Reagents		0	0	1	1	0	0	0
Other		0	1	0	0	0	0	0
All Instruments and Reagents		3	163	132	135	15	30	1

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Glucose	Specimen: U07	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		88	1	1	0	0	0	0
Siemens Atlas		92	0	1	0	0	0	0
Siemens Clinitek Status		42	0	0	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0	0
Siemens Clinitek 500		47	0	0	0	0	0	0
Siemens Clinitek Advantus		48	1	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		35	0	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0	0
IRIS iChem Velocity		8	0	0	0	0	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		183	1	0	0	0	0	0
Siemens Atlas Reagent		85	0	1	0	0	0	0
Sulfosalicylic Acid		1	0	0	0	0	0	0
Roche Chemstrips		46	0	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0	0
Bayer Clinitest		54	1	1	0	0	0	0
Siemens Uristix		2	0	0	0	0	0	0
Hitachi/Roche Reagents		2	0	0	0	0	0	0
Other		1	0	0	0	0	0	0
All Instruments and Reagents		433	2	2	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Glucose	Specimen: U08	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		89	0	0	0	0	0	0
Siemens Atlas		93	0	0	0	0	0	0
Siemens Clinitek Status		42	0	0	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0	0
Siemens Clinitek 500		47	0	0	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0	0
Roche Urisys		35	0	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0	0
IRIS iChem Velocity		8	0	0	0	0	0	0
Hitachi/Roche Analyzer		2	0	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		184	0	0	0	0	0	0
Siemens Atlas Reagent		86	0	0	0	0	0	0
Sulfosalicylic Acid		1	0	0	0	0	0	0
Roche Chemstrips		46	0	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0	0
Bayer Clinitest		55	0	0	0	0	0	0
Siemens Uristix		2	0	0	0	0	0	0
Hitachi/Roche Reagents		2	0	0	0	0	0	0
Other		1	0	0	0	0	0	0
All Instruments and Reagents		436	0	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Glucose	Specimen: U09	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		1	6	67	46	1	3	2
Siemens Atlas		0	1	90	2	0	0	0
Siemens Clinitek Status		0	1	27	10	0	3	1
Siemens Clinitek 50		0	1	3	4	0	1	0
Siemens Clinitek 100		0	0	3	3	0	0	0
Siemens Clinitek 500		0	0	33	16	0	0	0
Siemens Clinitek Advantus		1	1	28	20	0	0	0
Roche UA Reader/Super UA Read		0	0	3	0	0	0	0
Roche Chemstrip Criterion		0	0	13	1	0	0	0
Roche Urisys		0	0	28	1	0	6	0
Roche Cobas U411		0	0	1	0	0	0	0
IRIS AUTION MAX		0	0	28	3	4	3	0
IRIS iChem Velocity		0	0	0	7	0	1	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
Other Chemistry Analyzer		0	0	1	0	0	0	0
<Reagents>								
Siemens Reagent Strips		0	3	113	65	0	3	1
Siemens Atlas Reagent		0	1	83	2	0	0	0
Sulfosalicylic Acid		0	0	0	1	0	0	0
Roche Chemstrips		0	0	45	5	0	0	0
Roche Urisys Cassette		0	0	6	1	0	6	0
ARKRAY Aution Sticks		0	0	22	1	4	2	0
IRIS Reagent Strips		0	0	6	9	0	2	0
Siemens Acetest		0	1	0	0	0	0	0
Bayer Clinitest		2	5	50	26	1	3	2
Siemens Uristix		0	0	0	3	0	0	0
Hitachi/Roche Reagents		0	0	2	0	0	0	0
Other		0	0	0	0	0	1	0
All Instruments and Reagents		2	10	327	113	5	17	3

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Glucose	Specimen: U10	Negative	Trace 50-100mg/dL	1/4% 150-300mg/dL	1/2% 500mg/dL	3/4% 750mg/dL	1% >=1000mg/dL	2% >=2000mg/dL
<Instruments>								
Visual examination		76	21	4	0	0	0	0
Siemens Atlas		93	0	0	0	0	0	0
Siemens Clinitek Status		16	26	0	0	0	0	0
Siemens Clinitek 50		2	7	0	0	0	0	0
Siemens Clinitek 100		5	1	0	0	0	0	0
Siemens Clinitek 500		25	24	0	0	0	0	0
Siemens Clinitek Advantus		31	19	0	0	0	0	0
Roche UA Reader/Super UA Read		1	2	0	0	0	0	0
Roche Chemstrip Criterion		4	10	0	0	0	0	0
Roche Urisys		11	24	0	0	0	0	0
Roche Cobas U411		0	1	0	0	0	0	0
IRIS AUTION MAX		33	5	0	0	0	0	0
IRIS iChem Velocity		0	7	1	0	0	0	0
Hitachi/Roche Analyzer		1	1	0	0	0	0	0
<Reagents>								
Siemens Reagent Strips		103	81	0	0	0	0	0
Siemens Atlas Reagent		86	0	0	0	0	0	0
Sulfosalicylic Acid		0	1	0	0	0	0	0
Roche Chemstrips		11	38	0	0	0	0	0
Roche Urisys Cassette		6	7	0	0	0	0	0
ARKRAY Aution Sticks		25	4	0	0	0	0	0
IRIS Reagent Strips		8	8	1	0	0	0	0
Siemens Acetest		1	0	0	0	0	0	0
Bayer Clinitest		56	5	4	0	0	0	0
Siemens Uristix		0	3	0	0	0	0	0
Hitachi/Roche Reagents		1	1	0	0	0	0	0
Other		1	0	0	0	0	0	0
All Instruments and Reagents		298	148	5	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Hemoglobin	Specimen: U06	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		0	0	0	1	0	39
Siemens Atlas		1	0	0	0	0	91
Siemens Clinitek Status		0	0	0	2	0	39
Siemens Clinitek 50		0	0	0	0	0	9
Siemens Clinitek 100		0	0	0	1	0	5
Siemens Clinitek 500		0	0	0	0	0	48
Siemens Clinitek Advantus		0	0	0	0	0	49
Roche UA Reader/Super UA Read		0	0	0	1	0	2
Roche Chemstrip Criterion		0	0	0	0	0	14
Roche Urisys		0	0	0	2	0	32
Roche Cobas U411		0	0	0	0	0	1
IRIS AUTION MAX		0	0	0	25	0	13
IRIS iChem Velocity		0	0	0	2	0	6
Hitachi/Roche Analyzer		0	0	0	0	0	1
<Reagents>							
Siemens Reagent Strips		0	0	0	4	0	181
Siemens Atlas Reagent		1	0	0	0	0	87
Roche Chemstrips		0	0	0	3	0	45
Roche Urisys Cassette		0	0	0	0	0	13
ARKRAY Aution Sticks		0	0	0	21	0	8
IRIS Reagent Strips		0	0	0	6	0	11
Siemens Hemastix		0	0	0	0	0	1
None, no reagent or strip use		0	0	0	0	0	1
Other		0	0	0	0	0	2
All Instruments and Reagents		1	0	0	34	0	349

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Hemoglobin	Specimen: U07	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		38	0	0	0	0	0
Siemens Atlas		90	1	0	0	0	1
Siemens Clinitek Status		39	2	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		46	1	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
IRIS iChem Velocity		8	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		180	3	0	0	0	0
Siemens Atlas Reagent		86	1	0	0	0	1
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		2	0	0	0	0	0
All Instruments and Reagents		376	4	0	0	0	1

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Hemoglobin	Specimen: U08	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		33	5	0	0	0	0
Siemens Atlas		89	3	0	0	0	0
Siemens Clinitek Status		13	28	0	0	0	0
Siemens Clinitek 50		2	7	0	0	0	0
Siemens Clinitek 100		3	3	0	0	0	0
Siemens Clinitek 500		8	38	1	0	0	0
Siemens Clinitek Advantus		6	43	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
IRIS iChem Velocity		3	0	5	0	0	0
Hitachi/Roche Analyzer		0	1	0	0	0	0
<Reagents>							
Siemens Reagent Strips		59	123	1	0	0	0
Siemens Atlas Reagent		85	3	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		12	0	5	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		0	2	0	0	0	0
All Instruments and Reagents		247	128	6	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Hemoglobin	Specimen: U09	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		2	0	1	0	0	36
Siemens Atlas		0	0	0	0	0	92
Siemens Clinitek Status		0	0	0	2	0	39
Siemens Clinitek 50		0	0	0	1	0	8
Siemens Clinitek 100		0	0	0	0	0	6
Siemens Clinitek 500		0	0	0	1	0	47
Siemens Clinitek Advantus		0	0	0	0	0	49
Roche UA Reader/Super UA Read		0	0	0	1	0	2
Roche Chemstrip Criterion		0	0	0	0	0	14
Roche Urisys		0	0	0	2	1	31
Roche Cobas U411		0	0	0	0	0	1
IRIS AUTION MAX		0	0	0	25	0	13
IRIS iChem Velocity		0	0	1	0	0	7
Hitachi/Roche Analyzer		0	0	0	0	0	1
<Reagents>							
Siemens Reagent Strips		2	0	1	4	0	178
Siemens Atlas Reagent		0	0	0	0	0	88
Roche Chemstrips		0	0	0	3	1	43
Roche Urisys Cassette		0	0	0	0	0	13
ARKRAY Aution Sticks		0	0	0	19	0	10
IRIS Reagent Strips		0	0	1	6	0	10
Siemens Hemastix		0	0	0	0	0	1
None, no reagent or strip use		0	0	0	0	0	1
Other		0	0	0	0	0	2
All Instruments and Reagents		2	0	2	32	1	346

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Hemoglobin	Specimen: U10	Negative	Trace 5-10ery/uL	Small 1+ 25ery/uL	Moderate 2+ 50ery/uL	150ery/uL	Large 3+ 250ery/uL
<Instruments>							
Visual examination		38	0	0	0	0	0
Siemens Atlas		91	1	0	0	0	0
Siemens Clinitek Status		36	5	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		45	2	0	0	0	0
Siemens Clinitek Advantus		46	2	0	0	0	1
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
IRIS iChem Velocity		8	0	0	0	0	0
Hitachi/Roche Analyzer		1	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		173	9	0	0	0	1
Siemens Atlas Reagent		87	1	0	0	0	0
Roche Chemstrips		47	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		29	0	0	0	0	0
IRIS Reagent Strips		17	0	0	0	0	0
Siemens Hemastix		1	0	0	0	0	0
None, no reagent or strip use		1	0	0	0	0	0
Other		2	0	0	0	0	0
All Instruments and Reagents		370	10	0	0	0	1

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Ketones	Specimen: U06	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		0	0	0	11	49	3
Siemens Atlas		1	0	0	7	85	0
Siemens Clinitek Status		0	0	0	0	41	0
Siemens Clinitek 50		0	0	0	1	8	0
Siemens Clinitek 100		0	0	0	0	6	0
Siemens Clinitek 500		0	0	0	16	32	0
Siemens Clinitek Advantus		0	0	1	12	36	0
Roche UA Reader/Super UA Read		0	0	0	1	2	0
Roche Chemstrip Criterion		0	0	0	2	12	0
Roche Urisys		0	0	0	3	31	0
Roche Cobas U411		0	0	0	0	1	0
IRIS AUTION MAX		0	0	0	2	36	0
IRIS iChem Velocity		0	0	0	4	4	0
<Reagents>							
Siemens Reagent Strips		0	0	1	30	155	0
Siemens Atlas Reagent		1	0	0	7	77	0
Roche Chemstrips		0	0	0	10	39	0
Roche Urisys Cassette		0	0	0	0	13	0
ARKRAY Aution Sticks		0	0	0	1	26	0
IRIS Reagent Strips		0	0	0	5	13	0
Siemens Acetest		0	0	0	6	14	3
Siemens Ketostix		0	0	0	0	3	0
Other		0	0	0	0	3	0
All Instruments and Reagents		1	0	1	59	343	3

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Ketones	Specimen: U07	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		47	0	0	0	0	0
Siemens Atlas		92	0	0	0	1	0
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		47	0	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
IRIS iChem Velocity		8	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		185	0	0	0	0	0
Siemens Atlas Reagent		84	0	0	0	1	0
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0	0
IRIS Reagent Strips		18	0	0	0	0	0
Siemens Acetest		10	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		3	0	0	0	0	0
All Instruments and Reagents		389	0	0	0	1	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Ketones	Specimen: U08	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		47	0	0	0	0	0
Siemens Atlas		93	0	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		47	0	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
IRIS iChem Velocity		8	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		185	0	0	0	0	0
Siemens Atlas Reagent		85	0	0	0	0	0
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0	0
IRIS Reagent Strips		18	0	0	0	0	0
Siemens Acetest		10	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		3	0	0	0	0	0
All Instruments and Reagents		390	0	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Ketones	Specimen: U09	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		46	1	0	0	0	0
Siemens Atlas		93	0	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		47	0	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
IRIS iChem Velocity		8	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		184	1	0	0	0	0
Siemens Atlas Reagent		85	0	0	0	0	0
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0	0
IRIS Reagent Strips		18	0	0	0	0	0
Siemens Acetest		10	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		3	0	0	0	0	0
All Instruments and Reagents		389	1	0	0	0	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Ketones	Specimen: U10	Negative	Trace 5mg/dL	Small 1+ 15mg/dL	Moderate 2+ 40-50mg/dL	Large 3+ >=80mg/dL	Positive
<Instruments>							
Visual examination		47	0	0	0	0	0
Siemens Atlas		93	0	0	0	0	0
Siemens Clinitek Status		41	0	0	0	0	0
Siemens Clinitek 50		9	0	0	0	0	0
Siemens Clinitek 100		6	0	0	0	0	0
Siemens Clinitek 500		47	0	0	0	0	0
Siemens Clinitek Advantus		49	0	0	0	0	0
Roche UA Reader/Super UA Read		3	0	0	0	0	0
Roche Chemstrip Criterion		14	0	0	0	0	0
Roche Urisys		34	0	0	0	0	0
Roche Cobas U411		1	0	0	0	0	0
IRIS AUTION MAX		38	0	0	0	0	0
IRIS iChem Velocity		8	0	0	0	0	0
<Reagents>							
Siemens Reagent Strips		185	0	0	0	0	0
Siemens Atlas Reagent		85	0	0	0	0	0
Roche Chemstrips		48	0	0	0	0	0
Roche Urisys Cassette		13	0	0	0	0	0
ARKRAY Aution Sticks		27	0	0	0	0	0
IRIS Reagent Strips		18	0	0	0	0	0
Siemens Acetest		10	0	0	0	0	0
Siemens Ketostix		1	0	0	0	0	0
Other		3	0	0	0	0	0
All Instruments and Reagents		390	0	0	0	0	0

**New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012**

Protein	Specimen: U06	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	8	63	8	0	17
Siemens Atlas		0	0	1	92	0	0	0
Siemens Clinitek Status		0	0	0	25	15	0	1
Siemens Clinitek 50		0	0	0	2	7	0	0
Siemens Clinitek 100		0	0	0	4	2	0	0
Siemens Clinitek 500		0	0	0	48	0	0	0
Siemens Clinitek Advantus		0	0	0	50	0	0	0
Roche UA Reader/Super UA Read		0	0	0	3	0	0	0
Roche Chemstrip Criterion		0	0	1	13	0	0	0
Roche Urisys		0	0	3	31	0	0	0
Roche Cobas U411		0	0	0	1	0	0	0
IRIS AUTION MAX		0	0	1	37	0	0	0
IRIS iChem Velocity		0	0	0	8	0	0	0
Hitachi/Roche Analyzer		0	0	0	2	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	1	157	26	0	0
Siemens Atlas Reagent		0	0	1	88	0	0	0
Sulfosalicylic Acid		0	0	6	31	3	0	18
Roche Chemstrips		0	0	4	40	3	0	0
Roche Urisys Cassette		0	0	1	12	0	0	0
ARKRAY Aution Sticks		0	0	1	28	0	0	0
IRIS Reagent Strips		0	0	0	17	0	0	0
Siemens Uristix		0	0	0	3	0	0	0
Hitachi/Roche Reagents		0	0	0	1	0	0	0
Other		0	0	0	2	0	0	0
All Instruments and Reagents		0	0	14	379	32	0	18

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Protein	Specimen: U07	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		2	3	34	41	0	0	17
Siemens Atlas		0	0	7	86	0	0	0
Siemens Clinitek Status		0	0	1	39	0	0	1
Siemens Clinitek 50		0	0	0	9	0	0	0
Siemens Clinitek 100		0	0	0	6	0	0	0
Siemens Clinitek 500		0	0	43	5	0	0	0
Siemens Clinitek Advantus		0	0	45	5	0	0	0
Roche UA Reader/Super UA Read		0	0	0	3	0	0	0
Roche Chemstrip Criterion		0	0	12	2	0	0	0
Roche Urisys		0	1	5	28	0	0	0
Roche Cobas U411		0	0	0	1	0	0	0
IRIS AUTION MAX		0	0	35	3	0	0	0
IRIS iChem Velocity		0	0	1	7	0	0	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
<Reagents>								
Siemens Reagent Strips		1	1	96	86	0	0	0
Siemens Atlas Reagent		0	0	7	82	0	0	0
Sulfosalicylic Acid		1	2	22	16	0	0	18
Roche Chemstrips		0	0	19	28	0	0	0
Roche Urisys Cassette		0	1	0	12	0	0	0
ARKRAY Aution Sticks		0	0	28	1	0	0	0
IRIS Reagent Strips		0	0	8	9	0	0	0
Siemens Uristix		0	0	3	0	0	0	0
Hitachi/Roche Reagents		0	0	1	0	0	0	0
Other		0	0	1	1	0	0	0
All Instruments and Reagents		2	4	185	235	0	0	18

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Protein	Specimen: U08	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	3	56	20	0	17
Siemens Atlas		1	0	0	87	5	0	0
Siemens Clinitek Status		0	0	0	17	23	0	1
Siemens Clinitek 50		0	0	0	2	7	0	0
Siemens Clinitek 100		0	0	0	3	3	0	0
Siemens Clinitek 500		0	0	0	48	0	0	0
Siemens Clinitek Advantus		0	0	0	50	0	0	0
Roche UA Reader/Super UA Read		0	0	0	3	0	0	0
Roche Chemstrip Criterion		0	0	0	14	0	0	0
Roche Urisys		0	0	1	32	1	0	0
Roche Cobas U411		0	0	0	1	0	0	0
IRIS AUTION MAX		0	0	0	38	0	0	0
IRIS iChem Velocity		0	0	0	8	0	0	0
Hitachi/Roche Analyzer		0	0	0	2	0	0	0
<Reagents>								
Siemens Reagent Strips		0	0	0	141	43	0	0
Siemens Atlas Reagent		1	0	0	83	5	0	0
Sulfosalicylic Acid		0	0	3	30	7	0	18
Roche Chemstrips		0	0	0	43	4	0	0
Roche Urisys Cassette		0	0	1	12	0	0	0
ARKRAY Aution Sticks		0	0	0	29	0	0	0
IRIS Reagent Strips		0	0	0	17	0	0	0
Siemens Uristix		0	0	0	3	0	0	0
Hitachi/Roche Reagents		0	0	0	1	0	0	0
Other		0	0	0	2	0	0	0
All Instruments and Reagents		1	0	4	361	59	0	18

**New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012**

Protein	Specimen: U09	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		0	0	1	3	65	11	17
Siemens Atlas		0	0	0	0	85	9	0
Siemens Clinitek Status		0	0	0	0	40	0	1
Siemens Clinitek 50		0	0	0	0	9	0	0
Siemens Clinitek 100		0	0	0	0	6	0	0
Siemens Clinitek 500		0	0	0	16	32	0	0
Siemens Clinitek Advantus		0	0	0	8	42	0	0
Roche UA Reader/Super UA Read		0	0	0	0	3	0	0
Roche Chemstrip Criterion		0	0	0	0	14	0	0
Roche Urisys		0	0	0	0	34	0	0
Roche Cobas U411		0	0	0	0	1	0	0
IRIS AUTION MAX		0	0	0	0	38	0	0
IRIS iChem Velocity		0	0	0	1	7	0	0
Hitachi/Roche Analyzer		0	0	0	0	2	0	0
<Reagents>								
Siemens Reagent Strips		0	0	1	24	156	3	0
Siemens Atlas Reagent		0	0	0	0	80	9	0
Sulfosalicylic Acid		0	0	0	2	31	8	18
Roche Chemstrips		0	0	0	1	46	0	0
Roche Urisys Cassette		0	0	0	0	13	0	0
ARKRAY Aution Sticks		0	0	0	0	29	0	0
IRIS Reagent Strips		0	0	0	1	16	0	0
Siemens Uristix		0	0	0	0	3	0	0
Hitachi/Roche Reagents		0	0	0	0	1	0	0
Other		0	0	0	0	3	0	0
All Instruments and Reagents		0	0	1	28	378	20	18

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Protein	Specimen: U10	Negative	Trace 10-15mg/dL	1+ 25-30mg/dL	2+ 75-150mg/dL	3+ 300-600mg/dL	4+ >=1000mg/dL	Positive
<Instruments>								
Visual examination		4	18	56	3	0	0	16
Siemens Atlas		0	0	93	0	0	0	0
Siemens Clinitek Status		0	0	25	15	0	0	1
Siemens Clinitek 50		0	0	7	2	0	0	0
Siemens Clinitek 100		0	0	5	1	0	0	0
Siemens Clinitek 500		0	20	28	0	0	0	0
Siemens Clinitek Advantus		0	18	32	0	0	0	0
Roche UA Reader/Super UA Read		0	0	2	1	0	0	0
Roche Chemstrip Criterion		0	2	12	0	0	0	0
Roche Urisys		0	7	15	12	0	0	0
Roche Cobas U411		0	0	0	1	0	0	0
IRIS AUTION MAX		0	4	34	0	0	0	0
IRIS iChem Velocity		0	0	8	0	0	0	0
Hitachi/Roche Analyzer		0	0	2	0	0	0	0
<Reagents>								
Siemens Reagent Strips		0	42	124	18	0	0	0
Siemens Atlas Reagent		0	0	89	0	0	0	0
Sulfosalicylic Acid		4	10	27	1	0	0	17
Roche Chemstrips		0	8	27	12	0	0	0
Roche Urisys Cassette		0	3	6	4	0	0	0
ARKRAY Aution Sticks		0	2	27	0	0	0	0
IRIS Reagent Strips		0	2	15	0	0	0	0
Siemens Uristix		0	2	1	0	0	0	0
Hitachi/Roche Reagents		0	0	1	0	0	0	0
Other		0	0	2	0	0	0	0
All Instruments and Reagents		4	69	319	35	0	0	17

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Qualitative hCG report:

Beckman Coulter ICON II	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	2	2	0	0	2
Indeterminant	0	0	0	0	0
Positive	0	0	2	2	0
Beckman Coulter ICON 25	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	25	25	0	0	25
Indeterminant	0	0	0	0	0
Positive	0	0	25	25	0
Beckman Coulter ICON 20	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	23	23	0	0	23
Indeterminant	0	0	0	0	0
Positive	0	0	23	23	0
Cardinal Health hCG Combo	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	17	17	1	0	17
Indeterminant	0	0	0	0	0
Positive	0	0	16	17	0
Wampole Clearview hCG Combo	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	2	2	0	0	2
Indeterminant	0	0	0	0	0
Positive	0	0	2	2	0
Cen Med Elite	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	19	19	1	0	19
Indeterminant	0	0	0	0	0
Positive	0	0	18	19	0
Wampole Clearview hCG II	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	6	6	0	0	6
Indeterminant	0	0	0	0	0
Positive	0	0	6	6	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Sekisui OSOM Combo (Genzyme)	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	20	20	5	1	20
Indeterminant	0	0	1	0	0
Positive	0	0	14	19	0
Henry Schein One Step	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	2	2	0	0	2
Indeterminant	0	0	0	0	0
Positive	0	0	2	2	0
Inverness Medical BioStar Acceava	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	2	2	0	0	2
Indeterminant	0	0	0	0	0
Positive	0	0	2	2	0
Lifesign Status (serum/urine)	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	1	1	0	0	1
Indeterminant	0	0	0	0	0
Positive	0	0	1	1	0
McKesson hCG Combo	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	1	1	0	0	1
Indeterminant	0	0	0	0	0
Positive	0	0	1	1	0
Polymedco Poly stat	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	6	6	0	0	6
Indeterminant	0	0	0	0	0
Positive	0	0	6	6	0
Quidel QuickVue	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	77	77	5	0	77
Indeterminant	0	0	1	0	0
Positive	0	0	71	77	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

SA Scientific	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	2	2	1	0	2
Indeterminant	0	0	0	0	0
Positive	0	0	1	2	0
Stanbio QuPID	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	18	18	0	0	18
Indeterminant	0	0	0	0	0
Positive	0	0	18	18	0
Fisher Sure-Vue serum/urine	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	38	38	1	0	38
Indeterminant	0	0	0	0	0
Positive	0	0	37	38	0
Fisher Sure-Vue STAT serum/urine	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	26	26	1	0	26
Indeterminant	0	0	0	0	0
Positive	0	0	25	26	0
Signify hCG Urine/Serum	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	5	5	1	0	5
Indeterminant	0	0	0	0	0
Positive	0	0	4	5	0
Sacks Medical RefuAH	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	2	2	0	0	2
Indeterminant	0	0	0	0	0
Positive	0	0	2	2	0
Fisher Sure-Vue (urine only)	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	11	11	0	0	11
Indeterminant	0	0	0	0	0
Positive	0	0	11	11	0
True Medix	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	4	4	1	0	4
Indeterminant	0	0	0	0	0
Positive	0	0	3	4	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

Method not listed	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	21	21	1	0	21
Indeterminant	0	0	0	0	0
Positive	0	0	20	21	0

All Methods & Instruments	Specimen:U06	Specimen:U07	Specimen:U08	Specimen:U09	Specimen:U10
Negative	330	330	18	1	330
Indeterminant	0	0	2	0	0
Positive	0	0	310	329	0

New York State Department of Health - Wadsworth Center
Urinalysis/Urine Pregnancy Proficiency Testing - 26 November 2012

hCG Summary:

[Code], hCG Kit name	U06			U07			U08			U09			U10		
	N	I	P	N	I	P	N	I	P	N	I	P	N	I	P
[BC] Beckman Coulter ICON II	2	0	0	2	0	0	0	0	2	0	0	2	2	0	0
[BN] Beckman Coulter ICON 25	25	0	0	25	0	0	0	0	25	0	0	25	25	0	0
[BT] Beckman Coulter ICON 20	23	0	0	23	0	0	0	0	23	0	0	23	23	0	0
[CA] Cardinal Health hCG Combo	17	0	0	17	0	0	1	0	16	0	0	17	17	0	0
[CC] Wampole Clearview hCG Combo	2	0	0	2	0	0	0	0	2	0	0	2	2	0	0
[CM] Cen Med Elite	19	0	0	19	0	0	1	0	18	0	0	19	19	0	0
[CV] Wampole Clearview hCG II	6	0	0	6	0	0	0	0	6	0	0	6	6	0	0
[GO] Sekisui OSOM Combo (Genzyme)	20	0	0	20	0	0	5	1	14	1	0	19	20	0	0
[HS] Henry Schein One Step	2	0	0	2	0	0	0	0	2	0	0	2	2	0	0
[IM] Inverness Medical BioStar Acceava	2	0	0	2	0	0	0	0	2	0	0	2	2	0	0
[LS] Lifesign Status (serum/urine)	1	0	0	1	0	0	0	0	1	0	0	1	1	0	0
[MK] McKesson hCG Combo	1	0	0	1	0	0	0	0	1	0	0	1	1	0	0
[PM] Polymedco Poly stat	6	0	0	6	0	0	0	0	6	0	0	6	6	0	0
[QQ] Quidel QuickVue	77	0	0	77	0	0	5	1	71	0	0	77	77	0	0
[SA] SA Scientific	2	0	0	2	0	0	1	0	1	0	0	2	2	0	0
[SB] Stanbio QuPID	18	0	0	18	0	0	0	0	18	0	0	18	18	0	0
[SE] Fisher Sure-Vue serum/urine	38	0	0	38	0	0	1	0	37	0	0	38	38	0	0
[SF] Fisher Sure-Vue STAT serum/urine	26	0	0	26	0	0	1	0	25	0	0	26	26	0	0
[SG] Signify hCG Urine/Serum	5	0	0	5	0	0	1	0	4	0	0	5	5	0	0
[SM] Sacks Medical RefuAH	2	0	0	2	0	0	0	0	2	0	0	2	2	0	0
[SU] Fisher Sure-Vue (urine only)	11	0	0	11	0	0	0	0	11	0	0	11	11	0	0
[TX] True Medix	4	0	0	4	0	0	1	0	3	0	0	4	4	0	0
[90] Method not listed	21	0	0	21	0	0	1	0	20	0	0	21	21	0	0
[99] All hCG Methods	330	0	0	330	0	0	18	2	310	1	0	329	330	0	0