


Air and Emissions Fibers

Shipment: 389 **Matrix:** Air and Emissions
Analyte: Fibers
Sample: 1342
Technology: PCM
Prep Method:
Target: 272.7 Fibers per mm2
Limits: 24.5 to 418 Fibers per mm2
Basis: Robust Statistics (Normal Distribution)
Robust Mean: 223.7 Fibers per mm2
Std Dev: 67.6109 Fibers per mm2
Median: 214 Fibers per mm2

Totals for Matrix/Analyte/Technology:

Labs: Pass: 83 **Fail:** 1 **Total:** 84

Totals for Matrix/Analyte: **Labs:** Pass: 83 **Fail:** 1 **Total:** 84




Air and Emissions Fibers

Shipment: 389 **Matrix:** Air and Emissions
Analyte: Fibers
Sample: 4220
Technology: PCM
Prep Method:
Target: 133.72 Fibers per mm2
Limits: 23.4 to 223 Fibers per mm2
Basis: Robust Statistics (Normal Distribution)
Robust Mean: 124.1 Fibers per mm2
Std Dev: 38.7696 Fibers per mm2
Median: 123.55 Fibers per mm2

Totals for Matrix/Analyte/Technology:

Labs: Pass: 81 **Fail:** 3 **Total:** 84

Totals for Matrix/Analyte: **Labs:** Pass: 81 **Fail:** 3 **Total:** 84




Air and Emissions Fibers

Shipment: 389 **Matrix:** Air and Emissions
Analyte: Fibers
Sample: 7112
Technology: PCM
Prep Method:
Target: 521.76 Fibers per mm2
Limits: 111 to 771 Fibers per mm2
Basis: Robust Statistics (Normal Distribution)
Robust Mean: 442 Fibers per mm2
Std Dev: 108.0667 Fibers per mm2
Median: 433.8 Fibers per mm2

Totals for Matrix/Analyte/Technology:

Labs: Pass: 84 **Fail:** 0 **Total:** 84

Totals for Matrix/Analyte: **Labs:** Pass: 84 **Fail:** 0 **Total:** 84




Air and Emissions Fibers

Shipment: 389 **Matrix:** Air and Emissions
Analyte: Fibers
Sample: 7296
Technology: PCM
Prep Method:
Target: 521.76 Fibers per mm2
Limits: 105 to 921 Fibers per mm2
Basis: Robust Statistics (Normal Distribution)
Robust Mean: 511.7 Fibers per mm2
Std Dev: 132.1151 Fibers per mm2
Median: 527.68 Fibers per mm2

Totals for Matrix/Analyte/Technology:

Labs: Pass: 84 **Fail:** 0 **Total:** 84


Totals for Matrix/Analyte: **Labs:** Pass: 84 **Fail:** 0 **Total:** 84



Solid and Hazardous Waste Asbestos in Friable Material

Solid and Hazardous Waste Asbestos in Friable Material

Shipment: 389 Matrix: Solid and Hazardous Waste
Analyte: Asbestos in Friable Material
Sample: 4560
Technology: PLM
Prep Method:
Target: 5.4 %
Limits: 1 to 5.9 %
Basis: Swa scoring
Robust Mean: 1.8 %
Std Dev: 0.9661 %
Median: 1.8 %




Totals for Matrix/Analyte/Technology:

Labs: Pass: 58 Fail: 12 Total: 70

Totals for Matrix/Analyte: Labs: Pass: 58 Fail: 12 Total: 70

Solid and Hazardous Waste Asbestos in Friable Material

Shipment: 389 Matrix: Solid and Hazardous Waste
Analyte: Asbestos in Friable Material
Sample: 7267
Technology: PLM
Prep Method:
Target: 5 %
Limits: 1 to 6.3 %
Basis: Swa scoring
Robust Mean: 3.1 %
Std Dev: 1.6073 %
Median: 3.175 %




Totals for Matrix/Analyte/Technology:

Labs: Pass: 61 Fail: 9 Total: 70

Totals for Matrix/Analyte: Labs: Pass: 61 Fail: 9 Total: 70

Solid and Hazardous Waste Asbestos in Friable Material

Shipment: 389 Matrix: Solid and Hazardous Waste
Analyte: Asbestos in Friable Material
Sample: 8722
Technology: PLM
Prep Method:
Target: 10.2 %
Limits: 1 to 11 %
Basis: Swa scoring
Robust Mean: 4.7 %
Std Dev: 2.2493 %
Median: 4.725 %




Totals for Matrix/Analyte/Technology:

Labs: Pass: 67 Fail: 3 Total: 70

Totals for Matrix/Analyte: Labs: Pass: 67 Fail: 3 Total: 70

Solid and Hazardous Waste Asbestos in Friable Material

Shipment: 389 Matrix: Solid and Hazardous Waste
Analyte: Asbestos in Friable Material
Sample: 9496
Technology: PLM
Prep Method:
Target: 9 %
Limits: 1 to 9.9 %
Basis: Swa scoring
Robust Mean: 4 %
Std Dev: 2.6806 %
Median: 3.3 %



Totals for Matrix/Analyte/Technology:

Labs: Pass: 55 Fail: 15 Total: 70

Totals for Matrix/Analyte: Labs: Pass: 55 Fail: 15 Total: 70