


Air and Emissions Fibers

Shipment: 399 **Matrix:** Air and Emissions
Analyte: Fibers
Sample: 3041
Technology: PCM
Prep Method:
Target: 77.62 Fibers per mm2
Limits: 10.3 to 121 Fibers per mm2
Basis: Robust Statistics (Normal Distribution)
Robust Mean: 65.3 Fibers per mm2
Std Dev: 18.2606 Fibers per mm2
Median: 66.9 Fibers per mm2

Totals for Matrix/Analyte/Technology:

Labs: Pass: 84 **Fail:** 0 **Total:** 84

Totals for Matrix/Analyte: **Labs:** Pass: 84 **Fail:** 0 **Total:** 84




Air and Emissions Fibers

Shipment: 399 **Matrix:** Air and Emissions
Analyte: Fibers
Sample: 7002
Technology: PCM
Prep Method:
Target: 61.2 Fibers per mm2
Limits: 4.9 to 118 Fibers per mm2
Basis: Robust Statistics (Normal Distribution)
Robust Mean: 61.5 Fibers per mm2
Std Dev: 21.2035 Fibers per mm2
Median: 62 Fibers per mm2

Totals for Matrix/Analyte/Technology:

Labs: Pass: 81 **Fail:** 3 **Total:** 84

Totals for Matrix/Analyte: **Labs:** Pass: 81 **Fail:** 3 **Total:** 84




Air and Emissions Fibers

Shipment: 399 **Matrix:** Air and Emissions
Analyte: Fibers
Sample: 7332
Technology: PCM
Prep Method:
Target: 133.86 Fibers per mm2
Limits: 28.4 to 210 Fibers per mm2
Basis: Robust Statistics (Normal Distribution)
Robust Mean: 120.1 Fibers per mm2
Std Dev: 33.0006 Fibers per mm2
Median: 117.65 Fibers per mm2

Totals for Matrix/Analyte/Technology:

Labs: Pass: 83 **Fail:** 1 **Total:** 84

Totals for Matrix/Analyte: **Labs:** Pass: 83 **Fail:** 1 **Total:** 84




Air and Emissions Fibers

Shipment: 399 **Matrix:** Air and Emissions
Analyte: Fibers
Sample: 8787
Technology: PCM
Prep Method:
Target: 834.04 Fibers per mm2
Limits: 194 to 1270 Fibers per mm2
Basis: Robust Statistics (Normal Distribution)
Robust Mean: 727.9 Fibers per mm2
Std Dev: 215.3703 Fibers per mm2
Median: 724.11 Fibers per mm2

Totals for Matrix/Analyte/Technology:

Labs: Pass: 82 **Fail:** 2 **Total:** 84


Totals for Matrix/Analyte: **Labs:** Pass: 82 **Fail:** 2 **Total:** 84



Solid and Hazardous Waste Asbestos in Friable Material

Solid and Hazardous Waste Asbestos in Friable Material

Shipment: 399 **Matrix:** Solid and Hazardous Waste
Analyte: Asbestos in Friable Material
Sample: 4312
Technology: PLM
Prep Method:
Target: 64.3 %
Limits: 26 to 87 %
Basis: Swa scoring
Robust Mean: 55.6 %
Std Dev: 14.8956 %
Median: 57 %




Totals for Matrix/Analyte/Technology:

Labs: Pass: 69 **Fail:** 2 **Total:** 71

Totals for Matrix/Analyte: **Labs: Pass:** 69 **Fail:** 2 **Total:** 71

Solid and Hazardous Waste Asbestos in Friable Material

Shipment: 399 **Matrix:** Solid and Hazardous Waste
Analyte: Asbestos in Friable Material
Sample: 8677
Technology: PLM
Prep Method:
Target: 3.8 %
Limits: 1 to 13 %
Basis: Swa scoring
Robust Mean: 7.6 %
Std Dev: 3.8873 %
Median: 7.3 %




Totals for Matrix/Analyte/Technology:

Labs: Pass: 61 **Fail:** 10 **Total:** 71

Totals for Matrix/Analyte: **Labs: Pass:** 61 **Fail:** 10 **Total:** 71

Solid and Hazardous Waste Asbestos in Friable Material

Shipment: 399 **Matrix:** Solid and Hazardous Waste
Analyte: Asbestos in Friable Material
Sample: 8747
Technology: PLM
Prep Method:
Target: 7.8 %
Limits: 1 to 12 %
Basis: Swa scoring
Robust Mean: 6.7 %
Std Dev: 3.1612 %
Median: 6.2 %




Totals for Matrix/Analyte/Technology:

Labs: Pass: 64 **Fail:** 7 **Total:** 71

Totals for Matrix/Analyte: **Labs: Pass:** 64 **Fail:** 7 **Total:** 71

Solid and Hazardous Waste Asbestos in Friable Material

Shipment: 399 **Matrix:** Solid and Hazardous Waste
Analyte: Asbestos in Friable Material
Sample: 9475
Technology: PLM
Prep Method:
Target: 13 %
Limits: 1 to 14 %
Basis: Swa scoring
Robust Mean: 7.1 %
Std Dev: 3.782 %
Median: 6.78 %



Totals for Matrix/Analyte/Technology:

Labs: Pass: 58 **Fail:** 13 **Total:** 71

Totals for Matrix/Analyte: **Labs: Pass:** 58 **Fail:** 13 **Total:** 71